## **RSC Advances**



PAPER

View Article Online

View Journal | View Issue



Cite this: RSC Adv., 2017, 7, 48077

# Selective detection of glufosinate using CuInS<sub>2</sub> quantum dots as a fluorescence probe

Fengyuan Zhu, Da Jing Zhu\*b and Zhengzhu Zhang \*\*

We designed a novel method for rapid detection of glufosinate, based on quantum dots (QDs) as a fluorescence probe. To date, no studies have been published on the detection of glufosinate using QDs. The innovation of this fluorescence system, which was constructed using  $CuInS_2$  QDs and  $Cu(IDA) \cdot 2H_2O$ , is based on the use of intermolecular hydrogen bonds between  $Cu(IDA) \cdot 2H_2O$  and carboxyl group (-COOH) to detect glufosinate. In the fluorescence "turn-on" step, the presence of glufosinate induces the release of  $Cu(IDA) \cdot 2H_2O$  from the surface of the QDs, resulting in fluorescence intensity recovery. This method was performed in relatively clean aqueous solutions. Under optimal conditions, the calibration curve of the method showed good linearity for glufosinate, with a correlation coefficient ( $R^2$ ) of 0.99597. The method was employed to detect glufosinate on fresh tea-leaves, with satisfactory results

Received 10th August 2017 Accepted 28th September 2017

DOI: 10.1039/c7ra08848e

rsc.li/rsc-advances

### Introduction

Herbicides have been widely used in agriculture; consequently, their pollutants have become a severe problem worldwide. In 1972, a German laboratory isolated an amino acid from one species of soil bacterium and named it phosphinothricin, which was later renamed glufosinate.1 Glufosinate, registered for use in more than 40 countries, is a broad-spectrum and postemergence organophosphorus herbicide that is applied in various applications involving annual and perennial dicotyledonous weeds. Its residues can accumulate in environmental samples and destroy the natural equilibria among soil microorganisms. Since the entry of genetically engineered herbicideresistant crops into the market, the use of glufosinate has increased drastically. Consequently, glufosinate's lethal toxicity, teratogenicity, and potential reproductive toxicity have attracted increased research attention. Recent studies have shown that glufosinate can cause oxidative damage to human DNA.2-5 In addition, glufosinate can lead to decreased concentrations of glutamine in cerebrospinal fluid.6 With the increase in glufosinate usage, especially in tea growing areas, the monitoring of residues has attracted more and more interest.

To date, several methods have been developed to detect herbicides. However, the target samples must be extracted and purified before the herbicides can be detected using high performance liquid chromatography (HPLC),<sup>7-11</sup> gas chromatography (GC),<sup>12-16</sup> capillary electrophoresis,<sup>17-21</sup>

ion chromatography22,23 or enzyme-linked immunosorbent assays.24-26 Druart11 used HPLC on a 300 mm C18 column to detected glufosinate and the detection limits were in 0.015 to 0.13 ppm. After protein precipitation using acetonitrile and solid-phase extraction, serum samples were derivatized and analyzed by GC-MS (gas chromatography-mass spectroscopy), Tseng12 detected glufosinate in a concentration range of 3-100 ppm. To avoid and minimize potential interfering substances, most preferred analytical systems generally require expensive instrumentation, complicated sample pretreatments, and purification procedures to remove polysaccharides, protein, and peptides through precipitation, filtration, or centrifugation. Analyzing the sample and spectral data are also time-consuming. For simplicity, photometry using a diode array detector, fluorescence detection, and MS27-30 have also been adopted. Generally, glufosinate's lack of chromophores means that it must be derivatized before fluorescence detection, mostly using fluorenylmethoxycarbonyl chloride (FMOC-Cl) to react with the amino group as a fluorogenic reagent. This process is usually completed overnight, before GC analysis. Glufosinate has many unusual physical-chemical characteristics, such as high solubility in water (1370 g L<sup>-1</sup>), insolubility in organic solvents, low volatility, and high polarity; these characteristics make it difficult to establish methods for the analysis of glufosinate at trace levels. Therefore, it is necessary to develop fast, simple, and cost-effective methods to detect glufosinate.

To this end, quantum dots (QDs)—semiconductor nanoparticles confined to 1–10 nm in three dimensions—are ideal nanomaterials to detect pesticides because of their unique optical properties. Ternary I–III–VI CuInS<sub>2</sub> QDs have been intensively investigated for their potential to replace commonly available colloidal nanocrystals containing toxic elements in

<sup>&</sup>quot;State Key Laboratory of Tea Plant Biology and Utilization, Anhui Agricultural University, Hefei 230036, Anhui, People's Republic of China. E-mail: 2408337848@ qq.com; zzz@ahau.edu.cn

<sup>&</sup>lt;sup>b</sup>Department of Applied Chemistry, Anhui Agricultural University, Hefei 230036, Anhui, People's Republic of China. E-mail: zhujing@ahau.edu.cn

**RSC Advances** Paper

light-emitting and solar-harvesting applications. This is because of their pronounced quantum confinement effect, stable photoluminescence, and the absence of toxic elements (such as Cd, Hg or Pb).31-33 In particular, several pesticides have been detected using QD sensor systems. Some of these experiments required the use of acetylcholinesterase (AChE) enzymatic reaction for the specific selection of dichlorvos, 34,35 paraoxon,36 and methyl parathion.37 Other studies used the specificity of organophosphorus hydrolase (OPH) to detect parathion,38 and parathion-methyl.39 Specific antibodies or aptamers have also been used to identify paraquat,40 and acetamiprid.41 These substances increase the costs and affect device stability and practicability. Currently, most studies have focused on the fluorescence "off" patterns; however, these sensors have low sensitivity. Therefore, new "switch-on" models, with high sensitivity and selectivity for pesticides detection, are becoming more popular. These models use surface-modified reagents to confer variable affinity and specificity of QDs towards different targets. However, no studies have vet reported a sensor system based on QDs for the highly specificity and sensitive detection of glufosinate.

In the present work, we developed an ecologically-friendly, facile, and sensitive fluorescence sensor for glufosinate detection. This fluorescence system was constructed using CuInS2 QDs and the metal-chelate, Cu(IDA) · 2H2O. The schematic illustration of the measurement principle is shown in Fig. 1. L-Cysteine was modified on the surface of QDs by coordination between sulfhydryl and the metal. The amino groups (-NH<sub>2</sub>) of the surfacemodified substance L-cysteine can interact with H2O molecules via hydrogen bonding, which leads to quenching of the fluorescence intensity of CuInS2 QDs.42 The fluorescence intensity of the CuInS<sub>2</sub> QDs was quenched when Cu(IDA)·2H<sub>2</sub>O attached to the QDs through coordination with the surface-modified substance L-cysteine via hydrogen bonding between -NH2 and H2O. Glufosinate has a stronger coordinative interaction with Cu(IDA) · 2H<sub>2</sub>O than L-cysteine because the hydrophilicity of -COOH is stronger than that of -NH<sub>2</sub>. In fluorescence "turn-on," after the addition of glufosinate, Cu(IDA)·2H<sub>2</sub>O is released from the surface of QDs because intermolecular hydrogen bonds form between Cu(IDA)· 2H<sub>2</sub>O and the -COOH of glufosinate. Subsequently, the fluorescence intensity recovers. The proposed method, which mostly used relatively clean aqueous solutions, was applied to detect glufosinate in fresh tea-leaves.

## **Experimental**

#### Materials and devices

All chemicals and reagents used in this work were of analytical grade. Sodium hydrate (NaOH), potassium hydrate (KOH), copper(II) chloride dehydrate (CuCl<sub>2</sub>·2H<sub>2</sub>O), copper(II) hydroxide (Cu(OH)<sub>2</sub>), sulfourea (CS(NH<sub>2</sub>)<sub>2</sub>), and indium(III) chloride tetrahydrate (InCl<sub>3</sub>·4H<sub>2</sub>O) were purchased from Sinopharm Chemical Reagent Co. Ltd. Ethanol was purchased from Wuxi Zhanwang Chemical Works. The resistivity of the test water was greater than 18 M $\Omega$  cm<sup>-1</sup>. L-Cysteine (Cys) was purchased from Beijing Solarbio Science & Technology Co. Ltd. Iminodiacetic acid (IDA) was purchased from Shanghai Ryon Biological Technology Co. Ltd. L-Serine (Ser) was purchased from Sigma-Aldrich Corporation. Glufosinate and (aminomethyl)phosphonic acid were obtained from Dr Ehrenstorfer (Reference Materials for Residue Analysis). Fluorescence measurements were performed on a Carry Eclipse Fluorescence Spectrophotometer (Agilent Technologies) with a quartz cuvette path-length of 1 cm. The pH was measured using by an FE20 FiveEasy Plus pH meter (Mettler-Toledo). UV-Vis absorption spectra were recorded on a PerkinElmer UV/Vis Spectrometer Lambda 35. The centrifuge used was an Eppendorf D3024R.

#### Syntheses of CuInS2 QDs

Hydrothermal synthesis has become a hot research topic because of its convenience and simplicity. We followed published reports42 to synthetize CuInS2 QDs modified using L-cysteine. First, L-cysteine (3.60 mmol) was added to ultrapure water (7.5 mL) and mixed to form a clear solution. Then, CuCl<sub>2</sub>·2H<sub>2</sub>O (0.15 mmol) was added to the solution with stirring for 5 min. Next, InCl<sub>3</sub>·4H<sub>2</sub>O (0.15 mmol) was added to the solution with stirring for 10 min, and the pH value was adjusted to 11.3 using an NaOH solution (4 mol  $L^{-1}$ ). Ten minutes later, CS(NH<sub>2</sub>)<sub>2</sub> (0.30 mmol) was mixed in. Finally, the fully stirred solution was transferred into a Teflon-lined stainless steel autoclave (15 mL) and maintained at 150 °C. After 23 h, the autoclave was cooled to room temperature. The solution was purified by adding ethanol and centrifuging, and the centrifugal precipitation products were dried to obtain the CuInS2 QDs powder. The powder was dissolved in ultrapure water to obtain the CuInS $_2$  QDs solution (182.3  $\mu$ mol L $^{-1}$ ) and stored it at 4  $^{\circ}$ C for further research.




Fig. 1 Schematic illustration of the measurement principle of CuInS<sub>2</sub> QDs-Cu(IDA)·2H<sub>2</sub>O sensor for glufosinate detection.

Paper RSC Advances

#### Synthesis of Cu(IDA)·2H2O

 ${\rm Cu(IDA)\cdot 2H_2O}$  was synthetized in an aqueous solution. First,  ${\rm Cu(OH)_2}$  (0.25 mol) and IDA (0.25 mol) were successively dissolved in 200 mL of distilled water with stirring for 60 min. The  ${\rm Cu(OH)_2}$  precipitates were dissolved, and the solution became light blue after the addition of IDA. The solution was evaporated, crystallized, and dried in a vacuum to obtain a blue powder. The powder was then dissolved in distilled water at a concentration of 1 mmol  ${\rm L^{-1}}$  and stored for further experimentation.

#### Fluorescence quenching by Cu(IDA) · 2H2O

For the fluorescence quenching experiment,  $100~\mu L$  of  $CuInS_2$  QDs solution and different amounts of  $Cu(IDA) \cdot 2H_2O$ , without buffer solutions, were mixed in calibrated test tubes, and the mixed solution was diluted to the marker of 4 mL with ultrapure water. Ultrasonic concussion was used to thoroughly shake the test tube, and water was circulated to keep the temperature constant. Fluorescence was then measured using the Carry Eclipse Fluorescence Spectrophotometer.

#### **Detection of glufosinate**

Different concentrations of glufosinate were prepared by diluting a glufosinate sample (1000 ppm) in ultrapure water and shaking thoroughly. Then, 100  $\mu L$  of CuInS $_2$  QD solution, 300  $\mu L$  of Cu(IDA)·2H $_2$ O, and different concentrations of glufosinate solution (1–140 ppm) were admixed separately in calibrated test tubes, and ultrapure water was used to dilute to the marker of 4 mL. Subsequently, the mixtures were shaken thoroughly for 20 min using ultrasonic concussion before detection using the Carry Eclipse Fluorescence Spectrophotometer.

#### Real sample detection

The proposed method was used to detect glufosinate on the surface of fresh tea leaves. The leaves were cut off from a tea plant and cleaned with ultrapure water, and then glufosinate solutions of various concentrations were placed on the leaves and allowed to dry in a natural environment. The leaves were then washed using ultrapure water and the solution collected for analysis.

#### Results and discussion

#### Choosing an intermediate

The fluorescence and absorption spectra of the CuInS<sub>2</sub> QDs are obtained and shown in Fig. 2(a). The photoluminescence quantum yield we measured is 19.6%.

Previous studies have reported the special binding capacity of  $Cu(\pi)^{43}$  and the metal-chelate Cu-IDA.<sup>44</sup> We synthesized another substance,  $KCu(IDA)(Ser) \cdot H_2O$  using serine, according to the main toxicity of organophosphorus pesticides resulted from covalent binding to a serine OH group. In this study, we compared the selectivity of  $Cu(\pi)$ ,  $Cu(IDA) \cdot 2H_2O$ , and  $KCu(IDA)(Ser) \cdot H_2O$  to glufosinate.




Fig. 2 (a) Absorption and fluorescence spectra of CulnS<sub>2</sub> QDs. (b) CulnS<sub>2</sub> QDs quenched by intermediates and recovered by adding qlufosinate.

Cu(IDA)·2H<sub>2</sub>O had a stronger and more sensitive coordinated interaction with glufosinate compared with the other substances (Fig. 2(b)). The fluorescence intensity of CuInS<sub>2</sub> QDs decreased to varying degrees with the addition of 100  $\mu mol\ L^{-1}$  of Cu(II), Cu(IDA)·2H<sub>2</sub>O, and KCu(IDA)(Ser)·H<sub>2</sub>O, respectively. After the addition of glufosinate, the fluorescence intensity of the CuInS<sub>2</sub> QDs-Cu(IDA)·2H<sub>2</sub>O system recovered to a higher level than the other substances. Therefore, Cu(IDA)·2H<sub>2</sub>O had a suitable affinity for glufosinate in this study.

#### Quenching effect of Cu(IDA) · 2H<sub>2</sub>O on CuInS<sub>2</sub>

Considering the adsorption and stability of  $Cu(IDA) \cdot 2H_2O$ , ultrapure water was used in this work to formulate the solution and dilute to the marker in all experiment. We did not add any buffer solutions or other expensive biomolecular reagents. All the required substances are easily synthesized and do not harm the environment. Yan's research<sup>39</sup> showed that in an acid environment, the structure of  $CuInS_2$  QDs would be destroyed, and in an alkaline environment,  $Cu(OH)_2$  would precipitate. Therefore, ultrapure water, whose pH value has not been adjusted, was the most suitable choice to perform the experiments conveniently and efficiently.

Cu(IDA)·2H<sub>2</sub>O can compete with L-cysteine modified on QDs. It changed the photophysical properties of the QDs, resulting in the fluorescence intensity being quenched. The fluorescence intensity of CuInS<sub>2</sub> QDs at different concentrations of Cu(IDA)·2H<sub>2</sub>O (0, 25, 50, 75, 100, 125, 150, 175  $\mu$ mol L<sup>-1</sup>) are shown in Fig. 3. The decreasing trend began to ease when the concentration was greater than 75  $\mu$ mol L<sup>-1</sup>, indicating that Cu(IDA)·2H<sub>2</sub>O gradually reached saturation. Therefore, we chose the working concentration of Cu(IDA)·2H<sub>2</sub>O as 75  $\mu$ mol L<sup>-1</sup>.

#### Selectivity of the CuInS<sub>2</sub> QDs-Cu(IDA) · 2H<sub>2</sub>O system

(Aminomethyl)phosphonic acid (AMPA), the major metabolite of glyphosate degradation under biological conditions, and dimethoate (DIM) a moderately toxic organophosphorus pesticide, were chosen to test the selectivity of the designed method. The fluorescence intensity of the CuInS $_2$  QDs-Cu(IDA)  $\cdot$  2H $_2$ O system did not show a good linear correlation with the AMPA and dimethoate concentration (Fig. 4(a and b)). The chemical structures of these three substances are shown in Fig. 4(c).




Fig. 3 The fluorescence spectra upon the addition of Cu(IDA)  $\cdot$  2H<sub>2</sub>O (0 to 175  $\mu$ mol L<sup>-1</sup>).

These compounds share structures involving an amino group and phosphate group, but not a carboxyl group. Thus, the carboxyl group on glufosinate may be the selective group. A review of relevant literature<sup>45</sup> revealed that the metal-chelate Cu-IDA is always applied in the biomolecular field to combine protein and amino acids because of strong affinity of the metalchelate Cu-IDA for proteins and amino acids. The Cu(II) complex Cu(IDA)·2H2O is an independent substance that can be synthesized without involving other elements or ions. The H<sub>2</sub>O molecule on the axial copper site can form hydrogen bonds with the oxygen atoms of the carboxyl groups. 46 The presence of intermolecular and intramolecular hydrogen bonds increased the stability of the material's morphology. Consequently, glufosinate, with a carboxyl group, was selectively detected by the CuInS<sub>2</sub> QDs-Cu(IDA)·2H<sub>2</sub>O system. The H<sub>2</sub>O molecule and carboxyl group formed intermolecular hydrogen bonds between the oxygen atoms of water molecules and the oxygen atoms of the carboxyl groups. Compared with the recovery effect of the KCu(IDA)(Ser) · 2H<sub>2</sub>O complex shown in Fig. 2(b), the structure of the KCu(IDA)(Ser)·2H<sub>2</sub>O affected the affinity of H<sub>2</sub>O




Fig. 4  $F_1/F_0$  of CuInS<sub>2</sub> QDs-Cu(IDA)·2H<sub>2</sub>O system upon the addition of (a) (aminomethyl)phosphonic acid and (b) dimethoate. (c) Chemical structure of glufosinate, dimethoate and, (aminomethyl)phosphonic acid.

molecules. Hence, IDA has an influential role in stimulating the affinity of  $H_2O$  molecules.

Although AMPA and DIM, could also cause slight changes to the system, their groups have very weak coordinative interaction with the QDs. Therefore, the addition of AMPA, and DIM caused little change in the intensity of the CuInS $_2$  QDs-Cu(IDA) $_2$ CD system. Thus, the selectivity of the proposed method for glufosinate detection was confirmed using the two pesticides mentioned above.

#### **Detection of glufosinate**

The effect of reaction time on the fluorescence intensity of the CuInS<sub>2</sub> QDs-Cu(IDA)·2H<sub>2</sub>O system after the addition of glufosinate is shown in Fig. 5(a), all the data were averaged after 5 measurements. The recovery effect of glufosinate appears stable in 20 min, and the fluorescence intensity remained stable for more than 50 min. Fig. 5(b) shows the fluorescence spectra of the CuInS2 QDs-Cu(IDA)·2H2O system upon addition of different concentrations of glufosinate. The fluorescence intensity of the CuInS2 QDs-Cu(IDA)·2H2O system recovered gradually with increasing concentrations of glufosinate. Furthermore, there was good linearity between the recovered fluorescence and glufosinate concentrations between 1 and 25 ppm (*i.e.*, 4.65–116.18  $\mu$ mol L<sup>-1</sup>). The calibration curve was  $F_1/F_0 = 1.00953 + 0.00825C_{[glufosinate]}$  ppm (where  $F_1$  and  $F_0$  are the fluorescence intensity in the presence and absence of glufosinate, respectively). The corresponding regression coefficient was 0.99597 and the lower limit of detection (LOD) for glufosinate was 0.01 ppm. The detection limit is defined by the equation LOD =  $3\sigma/s$ , where  $\sigma$  is the




Fig. 5 (a) The effect of reaction time after the addition of glufosinate. (b) The effect of different concentrations of glufosinate on the fluorescence maxima. (c) The effect of different concentrations of glufosinate on total fluorescence.

Paper

standard deviation of the corrected blank signals of the CuInS<sub>2</sub> QDs-Cu(IDA)  $\cdot$  2H<sub>2</sub>O system and s is the slope of the calibration curve. In the 30–120 ppm range (i.e., 139–464.71  $\mu$ mol L<sup>-1</sup>), the calibration curve was  $F_1/F_0 = 1.16951 + 0.00188C_{\text{[glufosinate]}}$ ppm and  $R^2 = 0.98934$ . The reason for generating two detection ranges is that glufosinate possesses a strong affinity towards QDs, and it can create more radiative centers and stimulate blocking of nonradiative electron-hole recombination defect sites on the surface of the QDs.47 This combined action produced the first calibration curve. With the increase of glufosinate, this status reached saturation, and the second calibration curve appeared. The results showed that the CuInS<sub>2</sub> QDs-Cu(IDA)·2H<sub>2</sub>O system can detect glufosinate in a relatively clean aqueous solution.

Most previous studies of the detection of glufosinate were based on HPLC, GC, and MS, which require derivatization. Other methods for detecting pesticides and herbicides required the corresponding antibodies, bi-enzymes, or electrochemical potential. The proposed method is more sensitive than previously reported methods, whose detection limits were in the range of 3-100 ppm used GC-MS12 or the LOD of 0.017 ppm based on antibodies.26 Our method is appreciably simpler and faster than the others, lacking steps of derivatization, antibodies, or specific conditions.

#### Detection of glufosinate on tea leaves

When applied to plants, glufosinate is absorbed through the leaf surface, and almost no absorption occurs through the roots or stems. Therefore, the proposed method of detecting pesticides on the surface of tea-leaves was potentially significant for practical detection. We selected several concentrations (5, 10, 40, 60, 80 ppm) to carry out the experiment. The results of the sample detection were 4.99, 10.00, 34.71, 60.80, and 78.60 ppm. These results suggested the possibility of actual detection of glufosinate on leaves.

#### Conclusions

We designed a rapid method to detect glufosinate in pure aqueous solution, based on a CuInS<sub>2</sub> QDs-Cu(IDA)·2H<sub>2</sub>O fluorescence sensor. No studies have yet reported on the detection of glufosinate using QDs. The H<sub>2</sub>O molecule on the axial copper site can form hydrogen bonds with -NH2 and -COOH. The hydrophilicity of -COOH is stronger than that of -NH2. Consequently, glufosinate, with -COOH, led to the recovery of the system's fluorescence intensity after it had been quenched by the Cu(IDA)·2H<sub>2</sub>O complex. Under optimum conditions, this sensor exhibited good linearity and practical range without requiring expensive antibodies, or complicated surface modification procedures. However, our method faces many challenges in the detection of complex samples, such as residues in plant-based foods with many interfering substances. Overall, this work provided a new strategy for detecting glufosinate herbicide residues in relatively clean aqueous solutions.

#### Conflicts of interest

There are no conflicts to declare.

## Acknowledgements

This study was financially supported by the Program for Changjiang Scholars and Innovative Research Team in University, no. IRT1101, the National Natural Science Foundation of China (NSFC, 51702004), the Scientific Research Foundation of Anhui Agricultural University (Grant No. yj2017-06), and the Youth Research Foundation of Anhui Agricultural University (2016ZR003).

#### Notes and references

- 1 G. Hoerlein, Rev. Environ. Contam. Toxicol., 1994, 138, 73-
- 2 M. Koureas, A. Tsezou, A. Tsakalof, T. Orfanidou and C. Hadjichristodoulou, Sci. Total Environ., 2014, 496, 358-
- 3 L. Hoffman and D. Hardej, Environ. Toxicol. Pharmacol., 2012, 34, 556-573.
- 4 R. Paro, G. M. Tiboni, R. Buccione, G. Rossi, V. Cellini, R. Canipari and S. Cecconi, Toxicol. Appl. Pharmacol., 2012, 260, 155-161.
- 5 A. K. Srivastava, W. Ali, R. Singh, K. Bhui, S. Tyagi, A. A. Al-Khedhairy, P. K. Srivastava, J. Musarrat and Y. Shukla, Life Sci., 2012, 90, 815-824.
- 6 T. Ohtake, H. Yasuda, H. Takahashi, T. Goto, K. Suzuki, K. Yonemura and A. Hishida, Hum. Exp. Toxicol., 2001, 20, 429-434.
- 7 G. R. Todorovic, A. Mentler, M. Popp, S. Hann, G. Kollensperger, N. Rampazzo and W. E. H. Blum, Soil Sediment Contam., 2013, 22, 332-350.
- 8 N. Ngaosi, K. Seebunrueng and S. Srijaranai, Anal. Methods, 2016, 8, 4254-4262.
- 9 L. J. Liu, L. A. Xia, C. Y. Guo, C. X. Wu, G. Chen, G. L. Li, Z. W. Sun and J. M. You, Anal. Methods, 2016, 8, 3536-3544.
- 10 M. P. Yurkova, I. P. Pozdnyakov, V. F. Plyusnin, V. P. Grivin, N. M. Bazhin, A. I. Kruppa and T. A. Maksimova, *Photochem.* Photobiol. Sci., 2013, 12, 684-689.
- 11 C. Druart, O. Delhomme, A. de Vaufleury, E. Ntcho and M. Millet, Anal. Bioanal. Chem., 2011, 399, 1725-1732.
- 12 S. H. Tseng, Y. W. Lo, P. C. Chang, S. S. Chou and H. M. Chang, J. Agric. Food Chem., 2004, 52, 4057-4063.
- 13 M. Motojyuku, T. Saito, K. Akieda, H. Otsuka, I. Yamamoto and S. Inokuchi, J. Chromatogr. B, 2008, 875, 509-514.
- 14 E. Ueno, J. Pestic. Sci., 2011, 36, 554-558.
- 15 G. F. Xia, X. Y. Fang, Y. R. Wang and X. Y. Yang, Anal. Lett., 2017, 50, 787-796.
- 16 P. Moreira, P. G. Pinho, M. T. Baltazar, M. L. Bastos, F. Carvalho and R. J. Dinis-Oliveira, Toxicol. Lett., 2011,
- 17 E. Orejuela and M. Silva, Electrophoresis, 2005, 26, 4478-4485.

**RSC Advances** 

- 18 S. Y. Chang and C. H. Liao, J. Chromatogr. A, 2002, 959, 309-315.
- 19 R. Lanaro, J. L. Costa, L. C. R. Fernandes, R. R. Resende and M. F. M. Tavares, J. Anal. Toxicol., 2011, 35, 274-279.
- 20 M. Chicharro, E. Bermejo, S. Ongay and A. Zapardiel, Electroanalysis, 2008, 20, 534-541.
- 21 K. Islam, S. K. Jha, R. Chand, D. Han and Y. S. Kim, Microelectron. Eng., 2012, 97, 391-395.
- 22 K. C. Wang, S. M. Chen, J. F. Hsu, S. G. Cheng and C. K. Lee, J. Chromatogr. B, 2008, 876, 211-218.
- 23 H. Kidwell, J. J. Jones and D. E. Games, Rapid Commun. Mass Spectrom., 2001, 15, 1181-1186.
- 24 Z. Q. Lv, C. H. Wang, T. T. Wang, C. C. Chen, Y. Wang, B. A. Ning, M. Liu, J. Q. Liu, J. L. Bai, Y. Peng and Z. X. Gao, Biomed. Environ. Sci., 2013, 26, 398-402.
- 25 A. Abad-Fuentes, F. A. Esteve-Turrillas, C. Agullo, A. Abad-Somovilla and J. V. Mercader, Food Chem., 2012, 135, 276-284.
- 26 M. G. Weller, M. Diemer, C. Wersching, R. Niessner and H. Sochor, J. Agric. Food Chem., 2003, 51, 6668-6675.
- 27 L. Goodwin, J. R. Startin, B. J. Keely and D. M. Goodall, J. Chromatogr. A, 2003, 1004, 107-119.
- 28 M. Corbera, B. M. Simonet, V. Salvado and M. Hidalgo, Rapid Commun. Mass Spectrom., 2010, 24, 2931-2937.
- 29 X. F. Dong, S. X. Liang, Z. H. Shi and H. W. Sun, Food Chem., 2016, 192, 432-440.
- 30 J. Fenoll, P. Hellin, P. Sabater, P. Flores and S. Navarro, Talanta, 2012, 101, 273-282.
- 31 H. Z. Zhong, Z. L. Bai and B. S. Zou, J. Phys. Chem. Lett., 2012, 3, 3167-3175.
- 32 T. L. Li, Y. L. Lee and H. Teng, Energy Environ. Sci., 2012, 5, 5315-5324.

- 33 M. Michalska, A. Florczak, H. Dams-Kozlowska, J. Gapinski, S. Jurga and R. Schneider, Acta Biomater., 2016, 35, 293-304.
- 34 X. W. Meng, J. F. Wei, X. L. Ren, J. Ren and F. Q. Tang, Biosens. Bioelectron., 2013, 47, 402-407.
- 35 Z. Z. Zheng, X. Y. Li, Z. F. Dai, S. Q. Liu and Z. Y. Tang, J. Mater. Chem., 2011, 21, 16955-16962.
- 36 Z. Z. Zheng, Y. L. Zhou, X. Y. Li, S. Q. Liu and Z. Y. Tang, Biosens. Bioelectron., 2011, 26, 3081-3085.
- 37 H. Liang, D. D. Song and J. M. Gong, Biosens. Bioelectron., 2014, 53, 363-369.
- 38 D. Du, W. J. Chen, W. Y. Zhang, D. L. Liu, H. B. Li and Y. H. Lin, Biosens. Bioelectron., 2010, 25, 1370-1375.
- 39 X. Yan, H. X. Li, Y. Yan and X. G. Su, Food Chem., 2015, 173, 179-184.
- 40 E. Valera, R. Garcia-Febrero, M. I. Pividori, F. Sanchez-Baeza and M. P. Marco, Sens. Actuators, B, 2014, 194, 353-360.
- 41 J. J. Guo, Y. Li, L. K. Wang, J. Y. Xu, Y. J. Huang, Y. L. Luo, F. Shen, C. Y. Sun and R. Z. Meng, Anal. Bioanal. Chem., 2016, 408, 557-566.
- 42 Z. P. Liu, O. Ma, X. Y. Wang, Z. H. Lin, H. Zhang, L. L. Liu and X. G. Su, Biosens. Bioelectron., 2014, 54, 617-622.
- 43 Z. Q. Liu, S. P. Liu, P. F. Yin and Y. Q. He, Anal. Chim. Acta, 2012, 745, 78-84.
- 44 Z. el Rassi and C. Horvath, J. Chromatogr., 1986, 359, 241-
- 45 G. C. Serafica, G. Belfort and J. Pimbley, Biotechnol. Bioeng., 1994, 43, 21-36.
- 46 M. Tribet, B. Covelo, D. Choquesillo-Lazarte, J. M. Gonzalez-Perez, A. Castineiras and J. Niclos-Gutierrez, Inorg. Chem. Commun., 2003, 6, 343-345.
- 47 T. Khantaw, C. Boonmee, T. Tuntulani and W. Ngeontae, Talanta, 2013, 115, 849-856.