



Cite this: DOI: 10.1039/d5tc03054d

PEDOT:PSS conducting eutectogel for enhanced electrical recording and stimulation in implantable neural interfaces

Martin Kasavetov,^a Ruben Ruiz-Mateos Serrano,^b Antonio Dominguez-Alfaro,^{id c} Chaeyeon Lee,^b Matias L. Picchio,^{id de} David Mecerreyes,^{id de} George G. Malliaras,^{id b} Viviana Rincón Montes^{id †*ab} and Amparo Güemes^{id †*b}

Conductive polymers such as PEDOT:PSS are widely used in bioelectronic interfaces due to their mixed ionic–electronic conductivity and biocompatibility. However, their mechanical fragility and limited processability constrain their performance in implantable devices. Deep eutectic solvents (DES), when combined with PEDOT:PSS, form eutectogels that enable thick, soft coatings. Here, we present a PEDOT:PSS-based eutectogel incorporating choline chloride, lactic acid and GOPS, integrated into flexible thin-film electrode arrays for sciatic nerve interfacing. These implants feature an array of electrodes and a pre-formed spiral geometry to conformally wrap small-diameter nerves. Devices were fabricated using standard photolithography and reactive ion etching techniques, allowing side-by-side comparison of PEDOT:PSS/DES with conventional PEDOT:PSS electrodes. PEDOT:PSS/DES enabled single-layer films up to 800 nm thick, significantly greater than PEDOT:PSS, and yielding over three-fold improvements in impedance and charge injection capacity *in vitro*. Acute *in vivo* electrophysiology in rats confirmed enhanced neural recording and stimulation capabilities, with lower impedance, higher capacitance, and reduced motor activation thresholds. While PEDOT:PSS/DES more reliably elicited motor responses at lower stimulation currents, electromyogram signal amplitudes from the tibialis anterioris at matched stimulation levels were comparable between materials. These results suggest that while superior electrochemical properties improve neural interface performance, local electrode–tissue interactions remain critical. Overall, this work establishes DES-modified PEDOT:PSS as a promising electrode material for soft neural interfaces and highlights its potential for advancing implantable bioelectronics.

Received 12th August 2025,
Accepted 27th October 2025

DOI: 10.1039/d5tc03054d

rsc.li/materials-c

Introduction

Implantable bioelectronic devices have shown significant promise in treating a variety of pathological conditions, including arrhythmias, deafness, drug-resistant epilepsy, and Parkinson's disease, among others.^{1–5} In particular, thin-film microelectrode

arrays have enabled detailed study and monitoring of both the central and peripheral nervous systems.^{6–13} The principal advantage of these devices lies in their ultra-thin profiles, often just a few microns thick, which allows for excellent conformability, minimal foreign body reaction, and close contact between the electrodes and neural tissue.^{6,14} Poly(3,4-ethylenedioxythiophene) doped with polystyrene sulfonate (PEDOT:PSS) is a material that has gained popularity in the last decades as an electrode coating for thin-film electrode arrays due to its exceptional mixed ionic–electronic conductivity and volumetric capacitance.¹⁵ These properties enable the fabrication of small, high-performance electrodes with low impedance and mechanical flexibility, while maintaining biocompatibility and stability during electrical stimulation.^{16–18}

Despite its improved characteristics, its application in implantable neurotechnology remains constrained by several critical limitations. Thin-film PEDOT:PSS coatings, especially

^a Institute of Biological Information Processing (IBI-3) – Bioelectronics, Forschungszentrum Jülich, Leo-Brandt-Str., D-52425 Jülich, Germany. E-mail: v.rincon.montes@fz-juelich.de

^b Electrical Engineering Division, Department of Engineering, University of Cambridge, 9 JJ Thomson Ave, CB3 0FA, Cambridge, UK. E-mail: ag2239@cam.ac.uk

^c Instituto de Microelectrónica de Sevilla, IMSE-CNM (CSIC, Universidad de Sevilla), Av. Américo Vespucio 28, 41092, Sevilla, Spain

^d POLYMAT, Applied Chemistry Department, University of the Basque Country UPV/EHU, Paseo Manuel de Lardizábal 3, Donostia-San Sebastián 20018, Spain

^e IKERBASQUE, Basque Foundation for Science, Plaza Euskadi 5, Bilbao 48009, Spain

† These authors contributed equally to this work.



when electrodeposited or spin-coated, are mechanically fragile and prone to delamination,^{19–21} particularly under mechanical stress or during chronic implantation. This is a major concern for long-term stability *in vivo*, where micromotion between the implant and tissue is inevitable. Recent studies have demonstrated that electrodes based on conventional PEDOT:PSS formulations, including additives such as ethylene glycol (EG) and dodecylbenzene sulfonic acid (DBSA), exhibit superior stability compared to both gold (Au) and PEDOT:PSS-coated Au electrodes under accelerated aging conditions involving electrical stimulation, oxidative stress, and mechanical agitation.²² These findings highlight the potential of PEDOT:PSS as a standalone material for chronic neuromodulation applications. PEDOT:PSS also offers significantly higher charge injection capacity (CIC) than conventional metal coatings,^{20,23,24} making it attractive for electrical stimulation. However, further enhancement is needed to meet the demands of high-resolution stimulation and the ongoing miniaturisation of neural probes. Despite these advantages, PEDOT:PSS is typically processed as a thin film. Single spin-coated layers typically achieve thicknesses of 80–300 nm,^{21,25,26} and while thicker films can be fabricated by stacking multiple layers, such multilayer structures often suffer from poor mechanical integrity and increased risk of delamination.^{21,26} Its limited processability into thicker films therefore restricts the performance of electrodes. This is suboptimal for interfacing with soft, curved, and dynamic neural tissues, such as small peripheral nerves and the enteric nervous system, where conformability and mechanical integration are essential for stable, long-term performance.^{9,11,27}

To overcome these limitations, deep eutectic solvents (DES) have emerged as a promising additive to impart printability of three-dimensional (3D) coatings and enhance properties of PEDOT:PSS.²⁸ DES are a novel class of ionic additives, similar to ionic liquids, formed by mixing two or more components,

typically a hydrogen bond donor (HBD) and a hydrogen bond acceptor (HBA), that interact through hydrogen bonding to produce an eutectic mixture with an abnormal depression in its melting point.^{29–32} These mixtures are liquid in a broad temperature range, can be easily synthesised through green, low-cost processes, and exhibit biocompatibility, low toxicity, and ability to solubilize both hydrophilic and lipophilic compounds,^{33,34} making them attractive for biomedical applications. When mixed with PEDOT:PSS, DES act as gelators, forming supramolecular networks denoted as conducting eutectogel. These gels possess shear-thinning properties suitable for extrusion-based 3D printing,³⁵ enabling the fabrication of soft, conformable electrode architectures. In addition to enhancing processability, DES significantly improve both ionic and electronic conductivity compared to pristine PEDOT:PSS and plasticise the polymer at low concentrations, increasing its self-standability.^{36–40} Altogether, DES incorporation enables the fabrication of mechanically robust, self-supporting 3D electrode architectures, representing a significant step forward in addressing the structural fragility and integration challenges of thick conventional PEDOT:PSS coatings.^{41–48} Nonetheless, the rapid gelation of such materials³⁵ could pose challenges for their applicability in thin film technology. To the best of our knowledge, conducting eutectogels based on PEDOT:PSS and DES have never been reported in the literature for implantable bioelectronics.

In this work, we build upon a previously reported formulation of PEDOT:PSS combined with the DES formed by choline chloride and lactic acid (ChCl:LAC), by introducing an optimised concentration of (3-glycidyloxypropyl)trimethoxysilane (GOPS) as crosslinker. The addition of GOPS is critical to ensure the mechanical and chemical robustness of the material during processing, thereby enabling the microfabrication of thin-film electrode arrays. Similar choline-based ionic systems have also been used to create biocompatible solid-state electrolytes for bioelectronic devices.⁴⁹ In particular, conductive eutectogels formulated with PEGDA–PEDOT:PSS/ChCl:LAC have previously shown excellent biocompatibility with NIH/3T3 fibroblast cells, exhibiting no cytotoxic effects after 24 hours of culture.⁴⁶ Moreover, similar eutectogel systems have been successfully employed in our previous studies on wearable electrophysiology, demonstrating outstanding compatibility for skin-contact applications.^{41,50} We extensively evaluated the resulting PEDOT:PSS/ChCl:LAC/GOPS electrodes, hereafter referred to as PEDOT:PSS/DES, and compared against PEDOT:PSS electrodes using electrochemical and electrical characterizations in an *in vitro* setting to validate the performance of the materials for neural recording and stimulation. Finally, we validated the functionality and overall improved capabilities of the electrodes in acute *in vivo* experiments by electrically recording and stimulating the sciatic nerve in rats. Our results demonstrate that the PEDOT:PSS/DES formulation yields thicker films, resulting in enhanced electrochemical properties than conventional PEDOT:PSS formulations under identical spin-coating conditions, supporting it as a high-performance material platform for implantable bioelectronic interfaces.



Amparo Güemes

Amparo Güemes is a Royal Academy of Engineering and Rosetrees Research Fellow at the University of Cambridge, developing conductive polymer-based bioelectronic devices for neurotechnology and metabolic control. Her work combines materials science, engineering, and physiology to create soft electronics systems for real-time monitoring and modulation of neural circuits and metabolic processes. She received her BSc

in Biomedical Engineering from the Polytechnic University of Madrid, and her MSc and PhD in Biomedical and Electrical Engineering, respectively, from Imperial College London. Alongside her research and academic activities, she actively engages in outreach, including PPIE initiatives and student mentoring, to connect with the wider community.



Experimental

DES synthesis

DES were synthesised using the thermal mixing method. ChCl and LAC were combined at a molar ratio of 1:2 and heated to 60 °C under constant stirring until a clear, homogeneous liquid was formed.

Preparation of PEDOT:PSS/ChCl:LAC/GOPS eutectogels

PEDOT:PSS (Clevios™ PH1000) and the synthesised DES (ChCl:LAC) were mixed at room temperature, resulting in an estimated DES concentration of 1.3% (w/w) in PH1000. The mixture was vortexed for approximately 20 seconds to ensure homogeneity. Next, GOPS was added at a concentration of 0.65% (w/w). This concentration was selected as the minimum amount required to guarantee film integrity during photolithographic processing, while maintaining electrical conductivity and mechanical compliance. The solution was vortexed again for 20 seconds. Upon complete mixing, the composite exhibited a noticeable increase in viscosity, forming a gel-like consistency. The resulting eutectogel was loaded into a syringe, filtered using a polyvinylidene difluoride (PVDF) filter with a pore size of 0.4 µm, and deposited onto wafers for further processing.

Preparation of conventional PEDOT:PSS

The conventional PEDOT:PSS formulation was prepared similar to previously published protocols.⁵¹ Briefly, PEDOT:PSS (Clevios™ PH1000) was mixed with 5% (v/v) of EG and 0.25% (w/w) of DBSA and subjected to ultrasonication for 10 minutes. Next, 1% (v/v) of GOPS was added to the mixture, followed by an additional 1 minute of ultrasonication. Prior to deposition of the wafer, the solution was filtered using a PVDF filter with a pore size of 0.4 µm.

Device fabrication

The microfabrication was carried out at the Helmholtz Nano Facility at the *Forschungszentrum Jülich*.⁵² The devices were fabricated using conventional microfabrication techniques similar to previously reported protocols⁵³ and feature a layer of metal and a conductive polymer film (PEDOT:PSS/DES or PEDOT:PSS) embedded between two insulating layers of parylene-C (PaC).

As a first step, 2 µm of PaC were deposited on a 4-inch silicon host wafer *via* chemical vapor deposition (CVD) using the PDS 2010 Labcoater 2 tool (Specialty Coating Systems Inc., USA), followed by deposition of the metal layer. The latter was patterned using a bi-layer lift-off mask that comprised a LOR3B (Micro Resist Technology, Germany) bottom layer and an AZ nLOF2020 (MicroChemicals GmbH, Germany) top layer as image resist. Prior to spin-coating, the PaC substrate was dehydrated for 2 minutes at 120 °C on a direct-contact hotplate. The LOR3B layer was spin-coated at 3000 rpm for 30 seconds with a ramp of 500 rpm s⁻¹ and soft-baked at 150 °C for 5 minutes on a direct contact hotplate. After cooling down, the AZ nLOF2020 layer was spin-coated using the same

parameters, followed by a final soft bake at 120 °C for 1 minute. The photoresist was exposed using a maskless aligner (MLA-150, Heidelberg Instruments, Germany) with a dose of 260 mJ cm⁻² and a defocus distance (defoc) of 0 at a wavelength of 375 nm, followed by a post-exposure bake (PEB) at 110 °C for 2 minutes. The lift-off mask was finally developed in AZ 326 MIF (MicroChemicals GmbH, Germany) for 35 seconds and rinsed in a deionized water cascade.

The metal layer, comprising 20 nm of titanium (Ti) and 100 nm of Au, was deposited *via* electron-beam assisted evaporation using the Univex 400 physical vapor deposition tool (Leybold GbmH, Germany) with a deposition rate of 0.1 nm s⁻¹ for Ti and 0.5 nm s⁻¹ for Au. For the lift-off step, the wafers were submerged in acetone for 2–3 hours until the metal residues were lifted and rinsed in isopropanol. Finally, the solvent-resistant LOR3B layer was stripped in AZ 326 MIF for 5 minutes, followed by a rinse in a water cascade.

The conductive polymer film was deposited after patterning the metal layer *via* spin-coating. Prior to spin-coating, the wafers were dehydrated for 2 minutes at 120 °C and subjected to oxygen plasma activation at 0.8 mbar O₂ with a power of 80 W for 3 minutes (Gigabatch 310M, PVA TePla AG, Germany). For each wafer, 10 ml of PEDOT:PSS or PEDOT:PSS/DES was prepared as described previously and spin-coated at 500 rpm with a ramp of 5000 rpm s⁻¹ for 60 seconds. The resulting films were hard-baked at 120 °C for 1 hour on a direct contact hotplate. The PEDOT:PSS film was additionally soaked in deionized water overnight (~12 hours) to induce swelling and remove the volatile components.

Both the PEDOT:PSS and PEDOT:PSS/DES films were patterned *via* reactive ion etching (RIE) using a CF₄:O₂ (5:50 sccm) gas mix and a radio-frequency (RF) power of 150 W (Oxford Plasmalab 100-ICP 180, Oxford Instruments, United Kingdom) using AZ5214E as a positive photoresist etch mask. The latter was prepared by spin-coating at 3000 rpm (PEDOT:PSS) or 2500 rpm (PEDOT:PSS/DES) for 30 seconds after initial 5 seconds at 500 rpm with an acceleration of 500 rpm s⁻¹, followed by soft-baking at 110 °C for 1 min. The positive photoresist was exposed with a dose of 100 mJ cm⁻² and a defoc of 0 using the MLA-150 and developed in AZ 326 MIF for 70 seconds. In the case of PEDOT:PSS, the photoresist was spin-coated directly on top of the conductive polymer film, while PEDOT:PSS/DES films were first coated with a 100 nm-thick layer of PaC to promote the adhesion of the etch mask. This interlayer was etched through during the PEDOT:PSS/DES etching step, facilitating the structuring of the conductive polymer film. Lastly, the etch mask was stripped in acetone using gentle ultrasonication and rinsed in isopropanol.

To passivate the devices, an additional 2 µm of PaC were deposited as described before.⁵¹ Silane A127 adhesion promoter was applied in the chamber using the swab method to facilitate the adhesion of the passivation layer. Afterwards, the outline of the devices, as well as the openings of the electrodes and contact pads were patterned using RIE in two steps. To this end, a 15 µm-thick positive photoresist layer of AZ12XT (MicroChemicals GmbH, Germany) was spin-coated at 1000 rpm for



180 seconds with an acceleration of 200 rpm s^{-1} and soft-baked at 100°C for 4 minutes. The resist was subsequently exposed with a dose of 350 mJ cm^{-2} and a defoc of 2 with maskless lithography (MLA-150), subjected to a PEB (1 min at 90°C) and developed in AZ 326 MIF for 2 minutes. The outline of the probes, including the openings of the metal contact pads, was etched in the first step using a $\text{CF}_4:\text{O}_2$ (4:36 sccm) gas mix, RF power of 50 W, and ICP power of 500 W. The etch mask was stripped using two baths of AZ 100 Remover, followed by one bath of acetone and one bath of isopropanol, each comprising two to three minutes. Gentle ultrasonication was applied in the first bath with AZ 100 Remover to aid the dissolution of the photoresist. The wafers were finally dried using a compressed nitrogen gun. After stripping the resist, a new layer of AZ12XT was spin-coated, and the aforementioned process was repeated to etch the electrode openings.

This process flow was used to fabricate probes for both *in vitro* and *in vivo* settings. After microfabrication, the *in vitro* probes were released from the host substrate using drops of deionised water. They were then flip-chip bonded manually on custom-made printed-circuit boards (PCBs) using a low-temperature solder paste alloy (SMDLTB35T4, Chip Quik, Sn60/Bi40) at 180°C and cooled down to room temperature. Conversely, the *in vivo* probes were bonded directly on the host wafer. A customised flat flexible cable (FFC)⁵⁴ was bonded using anisotropic conductive film (ACF, TESA, Germany) and a FINEPLACER pico2 (Finetech GmbH & Co. KG, Germany), as previously described.⁵¹ After bonding, the flexible probes were released from the wafer using drops of deionised water.

Electrochemical and electrical characterization

All electrochemical measurements were carried out in a bath of saline solution ($1\times\text{PBS}$). Electrochemical impedance spectroscopy (EIS) was performed with the VSP-300 (BioLogic Science Instruments, France) potentiostat using a 10-mV sinusoidal signal and a frequency sweep from 1 Hz to 1 MHz. A 3-electrode setup was utilised comprising a Ag/AgCl reference electrode, a platinum (Pt) wire as a counter electrode, and the device electrodes serving as individual working electrodes. To further evaluate the performance of the electrodes, the impedance magnitude ($|Z|$) was extracted at 1 kHz as well as the specific capacitance (C_s) from the EIS data. The specific capacitance was estimated according to eqn (1):

$$C_s = \frac{C}{\text{GSA}} \quad (1)$$

where GSA is the geometric surface area of the electrodes and C is the equivalent capacitance, extracted by performing a Z fit at the low-frequency regime (1–100 Hz) using a simple Randles circuit.⁵⁵ In addition, we estimated the thermal noise level (ν_n) in the relevant frequency bands according to eqn (2):⁵⁶

$$\nu_m = \sqrt{4 \cdot k_B \cdot T \cdot \int_{f_1}^{f_2} \text{Re}\{Z(f)\} df} \quad (2)$$

where k_B is the Boltzmann constant, f_1 and f_2 are the lower and upper limits of the frequency band of interest ($f_1 = 300 \text{ Hz}$ and

$f_2 = 3 \text{ kHz}$ for the higher frequency band, and $f_1 = 1 \text{ Hz}$ and $f_2 = 300 \text{ Hz}$ for the lower frequency band), $T = 300 \text{ K}$ is the temperature, and $\text{Re}\{Z(f)\}$ is the real component of the impedance.

To estimate the cathodic charge storage capacity (CSC_c), we performed cyclic voltammetry between the potential limits of -0.9 V and 0.6 V , corresponding to the water window for conductive polymers.⁵⁷ The measurements were carried out on a multichannel potentiostat (CH Instruments Inc., USA) in a 3-electrode setup as described above with a scan rate of 100 mV s^{-1} and a current sensitivity level of 100 nA . At least 5 cycles were carried out for each electrode and only the last cycle was taken for further analysis. The cathodic CSC_c was estimated by integrating the cathodic current I_E over the potential limits ($E_a = 0.6 \text{ V}$ and $E_c = -0.9 \text{ V}$) according to eqn (3), where ν is the scan rate:

$$\text{CSC}_c = \frac{1}{\nu \cdot \text{GSA}} \int_{E_c}^{E_a} I(E) dE \quad (3)$$

To estimate the CIC, we utilized the current-controlled stimulator of the ME2100 system (Multi Channel Systems MCS GmbH, Germany) to deliver current pulses of defined amplitude and duration to the electrodes, while recording the resulting voltage transients with a 2208 Pico-Scope oscilloscope (Pico Technology, UK). Measurements were performed in saline solution using a two-electrode setup, with an Ag/AgCl pellet serving as both reference and counter electrode, and the oscilloscope connected in parallel to a working electrode. We utilized square biphasic charge-balanced pulses with a leading cathodic phase, followed by a symmetrical anodic phase with an interphase gap of $20 \mu\text{s}$. For the $100 \mu\text{m}$ size electrodes, which were implemented in the *in vivo* design, we tested four different phase durations (0.1 ms, 0.5 ms, 1 ms, and 5 ms) in order to gain insights into the time-dependent properties of the charge injection mechanism. To account for size-dependent scaling effects, we additionally measured the CIC for smaller electrodes (25 and $50 \mu\text{m}$) at a fixed phase duration of 0.5 ms. In each measurement, the phase duration (T_{ph}) and pulse shape were kept constant, and the current amplitude was gradually increased until the polarization limit of the electrodes was reached. As explained earlier, we assumed polarization limits of -0.9 V and 0.6 V for the cathodic and anodic regimes, respectively. The maximum voltage excursion during each phase of the pulse was extracted by subtracting the access voltage (V_A) and compared with the anodic and cathodic potential limits of the electrodes. In all measurements, the cathodic polarization limit was reached first. After extracting the maximum current at the polarization limit ($I_{\text{inj,max}}$), the CIC of each electrode was calculated according to eqn (4):

$$\text{CIC} = \frac{I_{\text{inj,max}} \cdot T_{\text{ph}}}{\text{GSA}} \quad (4)$$

The sheet resistance of the films was measured in dry conditions at room temperature with the Van der Pauw (VDP) method using a Keithley 4200 with probe station MPI-TS200 (Tektronix Inc., USA) and a 4-probe setup using tungsten



probes (Picoprobes ST-20-2). To facilitate the measurements, we fabricated specialised test structures to aid the contacting of the conductive polymer film. The test structures comprised metal pads (Ti/Au) contacting the edges of a 500 μm wide square of conductive polymer film. The structures were fabricated on a PaC-coated wafer with the process described earlier, and the conductive polymer squares were structured *via* dry etching. The conductivity of the films was estimated from the sheet resistance and the film thickness using eqn (5):

$$\sigma = \frac{1}{R_s \cdot t} \quad (5)$$

where R_s is the sheet resistance and t is the film thickness.

Film thickness measurements

The thickness of the spin-coated films was measured using a Dektak XT profilometer (Bruker, USA) with a vertical resolution of 2 nm. To facilitate the measurements, the films were mechanically scratched from the substrate at multiple positions of the wafer, allowing the thickness to be estimated by the resulting step height. For the conductivity measurement structures, the film thickness was determined separately for each structure *via* profilometry in a similar fashion by using the metal pads as a reference and considering the thickness of the metal layer. The thickness in the finalized devices was estimated by focused ion-beam (FIB) cuts and scanning electron microscopy (SEM) of the electrode cross section using an FEI TFS Helios 600i (G3) FIB/SEM dual-beam instrument. In this case, the thickness was measured directly under SEM by utilizing the material contrast between the conductive polymer films and the PaC passivation layers.

Statistical analysis for *in vitro* comparisons

Statistical analysis was performed using self-written scripts in Python utilizing the statistics package. Unless stated otherwise, the data is depicted as mean $\pm$ standard deviation. Normality of the data was tested using a Shapiro–Wilk test. If normality was rejected, a Mann–Whitney U test was employed to compare the data.

Surgery & *in vivo* electrophysiology measurements

All experimental procedures were performed in accordance with the UK Animals (Scientific Procedures) Act 1986 and were approved by the animal welfare ethical review body at the University of Cambridge. These procedures were performed under a project license (PP5478947) by A. Güemes (personal licence I10076024), issued by the UK Home Office. Non-recovery surgeries were performed in four female Sprague–Dawley rats (purchased at 200–250 g) (Charles River Laboratories, Kent, UK) with bilateral sciatic nerve implantation, resulting in a total of eight nerves ($N = 8$). Animals were anaesthetized with urethane (1.2–1.5 g kg^{−1}, intraperitoneally), and anaesthesia depth was monitored throughout. Body temperature was maintained using a heating pad, and the exposed nerve was kept moist with sterile saline throughout the experiment.

A mid-thigh incision was made to expose the sciatic nerve. Two types of cuff electrode coatings, PEDOT:PSS/DES and PEDOT:PSS, were implanted side by side on each nerve prior to its bifurcation, with their positions alternated across animals to control for proximal-distal effects. Care was taken to ensure consistent placement of corresponding electrodes from both cuffs around the nerve to enable a fair comparison. Neural and EMG signals were acquired using the Intan RHS system (RHS2000/Stim, Intan Technologies) at a sampling rate of 30 kHz.

The PEDOT/DES and PEDOT:PSS cuffs were connected to two separate ports to allow simultaneous recording, and the EMG needle was connected to an additional port. Two ground electrodes, connected to the respective GND/reference input of the headstage, were placed subcutaneously in two small pockets, one for the neural electrodes and another for the EMG electrode. For EMG recordings, the skin over the tibialis anterior muscle was incised to expose the muscle belly, and a 30G needle electrode was inserted directly into the muscle.

Each nerve underwent an electrophysiology protocol lasting approximately 1 hour. Spontaneous and pain-evoked activity was recorded first, beginning with a 5-minute baseline followed by repeated mechanical stimulation of the tibialis anterior muscle using a needle press to evoke nociceptive responses. This was followed by electrical stimulation through the implanted cuffs in monopolar configuration. Stimulation was delivered using single pulses (100 μs pulse width) at increasing amplitudes: from 5 μA to 10 μA in 1 μA steps, and then at 15, 20, and 30 μA . Each stimulation trial consisted of a 10-second baseline followed by 10 pulses at 1 Hz (1-second inter-pulse interval, pre-set in Intan) stimulating with one cuff, followed by 10 pulses stimulating with the other cuff at 1 Hz, alternating between cuffs to compare responses. Channels with impedance values exceeding 10 k Ω were excluded from further analysis.

Analysis of electrophysiology signals

Impedance magnitude and phase values were extracted from the Intan file metadata corresponding to the baseline recording for each available circular channel (100 μm diameter). From these values, the equivalent series resistance (R) and capacitance (C) of each electrode were computed using standard impedance modelling techniques, assuming a series RC circuit approximation.

Offline data processing was performed in Python, where signals were bandpass filtered between 200–3000 Hz for high-frequency activity analysis. For the analysis of pain-evoked activity, a 60-second segment of baseline activity and a 60-second segment during mechanical stimulation were selected. The root mean square (RMS) of the signal was computed for each channel, and the values were averaged across channels. Signal-to-noise ratio (SNR) was calculated as the ratio of RMS during evoked activity to RMS during baseline. Spikes elicited by the stimulation were detected using a negative threshold of $-12 \mu\text{V}$. Waveforms were extracted using a 4-ms window centred around the negative peak and averaged to characterize the evoked response. The same



procedure was applied to the capsaicin and nerve-press recordings.

For electrical stimulation analysis, stimulation artifacts were detected using an adaptive threshold, with a minimum inter-artifact interval of 1 second to match the stimulation frequency. A post-stimulus window from 11 ms to 200 ms after each artifact was extracted to capture evoked neural activity while avoiding contamination from the artifact itself, based on expected conduction velocities. RMS values were computed within this window and averaged across trials. Baseline RMS was also calculated using a window of the same duration immediately preceding the onset of stimulation. The increase in RMS relative to baseline was compared between the two electrode materials. Additionally, the stimulation threshold required to elicit a visible muscle twitch was noted during experiments and compared across materials.

Population-level analysis was performed using paired statistical comparisons within each animal. Assumptions were checked prior to applying parametric *t*-tests by testing whether the differences between paired values were normally distributed using the Shapiro–Wilk test. If normality was violated, the non-parametric Wilcoxon signed-rank test was used as an alternative. The number of available channels for each material varied between 4 and 9 in all trials, but was always kept consistent between the two recording ports within each trial to ensure fair pairwise comparison between material types. For the analysis of impedance and equivalent RC, values were averaged across available channels for each nerve, and paired statistical comparisons were made between the two materials. To visualize variability and distribution, boxplots were also generated for all individual channel impedance and RC equivalent values across all trials and animals. For the recording and stimulation-related metrics (RMS, SNR, muscle stimulation threshold), the difference between the two materials (Δ) was computed for comparison purposes, and paired statistical comparisons were also assessed.

Results and discussion

Material formulation, device design and fabrication

To explore the potential of PEDOT:PSS eutectogel enhanced with DES, we selected a composition using LAC as the HBD and ChCl as the HBA (Fig. 1(a), left). This specific DES (ChCl:LAC) was chosen based on prior studies demonstrating improved conductivity in thick pristine PEDOT:PSS films deposited *via* drop-casting.⁵⁸ We formulated a PEDOT:PSS/DES blend by mixing 1.3% (w/w) of DES with PH1000, followed by the addition of 0.65% (w/w) of GOPS, resulting in a PEDOT:DES:GOPS weight-ratio of 2:2:1. The addition of GOPS to the blend results in cross-linking of the film during the hard baking step, enhancing the stability of the material and thereby ensuring compatibility with standard microfabrication techniques.

Conventional PEDOT:PSS formulations often contain additives (Fig. 1(a), right) such as EG and DBSA to enhance conductivity *via* conformational changes in PEDOT chains,

from coiled to linear structures, driven by reduced water evaporation and a screening effect between the PSS-rich polymer and its dopant.^{59,60} Such PEDOT:PSS/EG/DBSA blends represent the standard benchmark for high-conductivity films in the literature, typically reaching conductivities of 10^2 – 10^3 S cm^{−1},^{61,62} whereas pristine or GOPS-modified PEDOT:PSS exhibits much lower values (<1 S cm^{−1}). Therefore, this formulation was selected as a relevant and practically representative control for comparison with the PEDOT:PSS/DES system. In contrast, DES-based additives achieve similar or superior performance. During evaporation, the negatively charged LAC molecules in the DES displace PSS anions, promoting rearrangement of positively charged PEDOT chains. This interaction was confirmed in our previous work⁵⁸ through a red shift in the symmetric SO₃[−] stretching vibration from 1099 cm^{−1} to 1014 cm^{−1}, consistent with PSS displacement, a phenomenon also reported in PEDOT:PSS formulations with cholinium lactate-based ionic liquids.^{39,63}

The PEDOT:PSS/DES formulation forms a supramolecular gel that increases viscosity without phase separation.^{35,64} When processed *via* spin coating, it enables significantly thicker films (451–818 nm, Fig. 1(b)) compared to conventional PEDOT:PSS/EG/DBSA formulations (hereafter referred to as PEDOT:PSS). To evaluate this advantage, we optimised spin-coating parameters and deposition timing to avoid premature gelation. Using a spin speed of 500 rpm and a high-acceleration ramp (5000 rpm s^{−1}), we achieved PEDOT:PSS/DES film thicknesses higher than 800 nm in a single step, substantially exceeding the 117–199 nm range of PEDOT:PSS films under the same conditions. Thickness measurements at three wafer locations (centre, mid-radius, and edge; Fig. 1(b)) revealed increased variability for PEDOT:PSS/DES, attributed to the gel's limited flow during spinning. Similar to the levelling effects achieved by surfactants such as DBSA in conventional PEDOT:PSS systems, introducing compatible interfacial modifiers could help improve film uniformity and scalability.

Despite the simpler additive composition (13 mg DES *vs.* 100 mg EG + 2.5 mg DBSA in the standard blend), PEDOT:PSS/DES yielded films 3–5 times thicker, highlighting the importance of supramolecular interactions and secondary doping effects. This enhanced thickness is particularly beneficial in bioelectronic interfaces, where the volumetric capacitance of PEDOT:PSS scales with film thickness, improving electrochemical performance.^{26,65} While thicker PEDOT:PSS films can be achieved *via* sequential layering, weak interlayer adhesion may compromise long-term stability.^{21,26} In contrast, PEDOT:PSS/DES enables thick, stable films in a single step, making it attractive for applications requiring both conformability and robust electrochemical properties.

To demonstrate applicability, we integrated both PEDOT:PSS/DES and conventional PEDOT:PSS into flexible PaC-based cuff implants for sciatic nerve interfacing in rodents (Fig. 1(c)). The electrode array consisted of 18 circular electrodes (100 μ m diameter, 175 μ m inter-electrode spacing) and six rectangular electrodes (800 $\times$ 120 μ m) arranged symmetrically. A 30° pre-formed bend allowed the array to warp






Fig. 1 Material formulation, device design and fabrication (a) Material composition of the PEDOT:PSS/DES blend proposed in this work (left). The formulation combines PEDOT:PSS, DES and GOPS at a weight ratio of 2 : 2 : 1. The performance of the new blend is compared to a standard PEDOT:PSS formulation (right) featuring EG, GOPS, and DBSA. (b) Violin plots illustrating the comparison of the wafer distribution of the film thickness of spin-coated films with the PEDOT:PSS/DES blend (blue) and the conventional PEDOT:PSS blend (orange). PEDOT:PSS/DES films exhibit a reduction in thickness towards the edge of the wafer. (c) Schematic of the *in vivo* probe used in this work. The design is adapted for interfacing with the rat sciatic nerve. (d) Process flow describing the fabrication of the devices. The process comprises the following steps: (1) CVD deposition of a PaC substrate layer (2 μm) on a 4" Si carrier wafer, (2) E-beam evaporation of the metal layer (Ti/Au) and patterning via lift-off. The metal layer comprises the feedlines as well as the contact pads for the front-end connector. (3) Spin-coating of the conductive polymer film and patterning of the film via RIE (4 and 5) to form metal-free electrodes. Deposition of a PaC encapsulation layer (6) with a thickness of 2 μm. Patterning of the shape and electrode openings via RIE (7 and 8) followed by etching of the electrode openings (9 and 10). (e), (f) Optical microscopy pictures of finalized *in vivo* probes, highlighting the sensing area of the devices. PEDOT:PSS/DES (e) devices feature an enhanced film thickness compared to conventional PEDOT:PSS devices (f). Insets show a close-up of the metal free polymeric electrodes. Scale bars correspond to 200 μm (main images) and 50 μm (insets). (g), (h) SEM images of the cross section of 25 μm PEDOT:PSS/DES (g) and PEDOT:PSS (h) electrodes obtained via FIB cuts. The estimated film thickness at the edge of the electrodes is 550 nm for the PEDOT:PSS/DES and 171 nm for the PEDOT:PSS electrode.

around the nerve up to three full turns, accommodating a range of rat nerve diameters (from 200 μm to 1 mm), thereby enabling a generalizable and anatomically adaptive interface design.

The devices were fabricated using standard microfabrication protocols⁵³ (Fig. 1(d)) with a Ti/Au metal stack for contact pads and feedlines, and a spin-coated conductive polymer layer for the electrodes. These layers were sandwiched between 2 μm-thick PaC substrate and passivation layers, resulting in a total device thickness of 4 μm. The thin PaC layers ensure that the resulting devices are highly flexible, facilitating efficient wrapping around the nerve. Aiming for maximal conformability and stability, we designed the devices with metal-free electrodes made either from spin-coated PEDOT:PSS/DES or PEDOT:PSS. While the metallic feedlines were patterned using a lift-off technique, the conductive polymer film was spin-coated and patterned *via* dry etching. Since the photoresist etch mask

exhibited poor adhesion when spin-coated directly on top of PEDOT:PSS/DES films, an additional 100 nm thick PaC interlayer was deposited to improve photoresist adhesion. In contrast, this step was not required for conventional PEDOT:PSS films, which exhibited good photoresist adhesion. After patterning the conductive polymer layer, the devices were passivated with PaC and structured using a two-step RIE process. The outline of the devices and the contact pad openings were etched first, followed by the etching of the electrode openings.

Final devices (Fig. 1(e) and (f)) showed a clear visual difference between the thicker PEDOT:PSS/DES (darker blue) and thinner PEDOT:PSS films. Finally, FIB cross-sections confirmed thicknesses of 550 nm for PEDOT:PSS/DES and 171 nm for PEDOT:PSS at the edges of the wafer (Fig. 1(g) and (h)), validating the enhanced deposition performance of the DES formulation.



Enhanced electrochemical and electrical performance of PEDOT:PSS/DES electrodes

We carried out an extensive electrochemical and electrical characterisation to evaluate the suitability of the PEDOT:PSS/DES blend for neural recording and stimulation, comprising EIS, estimation of the CSC and CIC, as well as resistivity measurements using the Van der Pauw method. To account for scaling effects in the electrochemical performance, we fabricated test probes featuring different electrode diameters, ranging from 25 μm to 100 μm (Fig. S1).

Fig. 2(a) depicts a Bode plot comparing the impedance and phase of PEDOT:PSS/DES and PEDOT:PSS electrodes with a diameter of 100 μm . Overall, PEDOT:PSS/DES demonstrates an excellent impedance with a typical behaviour featuring a resistive regime at high and intermediate frequencies (100 Hz–100 kHz) where the phase is close to 0° and the impedance is dominated by the access resistance, influenced by the resistive contribution of the electrolyte at the electrode–electrolyte interface. Subsequently, a transition to a capacitive regime at lower frequencies occurs.⁶⁶ Interestingly, the PEDOT:PSS electrodes exhibit an additional peak in phase in the frequency range between 100–300 Hz. This nontypical behaviour likely originates from oxidation of the top layer of the films during RIE of the contact openings, which was performed with strong oxygen-containing plasma. Due to the larger thickness of the PEDOT:PSS/DES the relative effect of the oxidation is not as prominent and does not result in the formation of a distinct peak. Nevertheless, for both PEDOT:PSS-based materials the phase of the electrodes approaches -90° at low frequencies (below 100 Hz), approximating a perfectly capacitive behaviour as reported before.^{53,55,67}

PEDOT:PSS/DES features a significantly reduced impedance compared to PEDOT:PSS for frequencies below 10 kHz (Fig. 2(b)). The relevant frequency spectrum in the context of neural interfacing depends on the type of recorded signals. While slow signals such as local field potentials (LFPs) are typically analysed at frequencies under 300 Hz, the dominant frequency components of faster action potential signals lie in the range of 300 Hz to 3 kHz.⁶⁸ Therefore, neural implants are often benchmarked considering their total impedance value at 1 kHz.⁶⁶ The PEDOT:PSS/DES electrodes significantly outperform PEDOT:PSS with respect to this metric, featuring an impedance of $8.3 \pm 1.6 \text{ k}\Omega$ compared to $20.3 \pm 8.4 \text{ k}\Omega$ for PEDOT:PSS, considering a diameter of 100 μm ($p < 0.001$, $N = 19$). Normalizing the impedance to the GSA yields a specific impedance of $0.652 \pm 0.126 \Omega \text{ cm}^2$ and $1.59 \pm 0.66 \Omega \text{ cm}^2$ for PEDOT:PSS/DES and PEDOT:PSS, respectively. In addition, PEDOT:PSS/DES electrodes retain their resistive behaviour at lower frequencies, which can facilitate high-fidelity recording of slower signals with better SNR and less signal distortion due to phase shifts. For instance, the reduced impedance would translate to a thermal noise level of $2.02 \pm 0.17 \mu\text{V}$ for PEDOT:PSS/DES and $3.74 \pm 0.62 \mu\text{V}$ for PEDOT:PSS in the low frequency band of 1–300 Hz, and $1.79 \pm 0.182 \mu\text{V}$ and $2.37 \pm 0.54 \mu\text{V}$ in the higher frequency band of 300–3 kHz.




Fig. 2 Electrochemical and electrical characterization. (a) Bode plots showing the impedance of 100 μm PEDOT:PSS/DES and PEDOT:PSS electrodes obtained via EIS. The shaded region represents the standard deviation ($N = 19$). (b) Impedance magnitude at 1 kHz for PEDOT:PSS/DES and PEDOT:PSS electrodes with different diameters (25, 50, and 100 μm). The quality of the fits (a/D^b , where D is the electrode diameter) is $r^2 = 0.985$ (blue) and $r^2 = 0.999$ (red). (c) Cyclic voltammetry of 100 μm PEDOT:PSS/DES ($N = 14$) and PEDOT:PSS ($N = 17$) electrodes with standard potential limits of -0.9 V and 0.6 V . (d) CSC_c scaling dependency (a/D^b) for PEDOT:PSS/DES ($r^2 = 0.983$) and PEDOT:PSS ($r^2 = 0.999$) electrodes of different sizes. (e) Exemplary voltage transient measurements for a PEDOT:PSS and a PEDOT:PSS/DES electrode (100 μm diameter) after applying a squared biphasic current pulse with an amplitude of 36 μA and a pulse width of 0.5 ms. The cathodic voltage excursion of the electrodes (V_{E-C}) is determined after subtracting the access voltage V_A . The transient curves were averaged over 10 pulse repetitions (the standard deviation depicted as shaded region). (f) Averaged voltage transient of the 100 μm PEDOT:PSS/DES electrodes ($N = 14$) for a pulse width of 0.5 ms at the cathodic excursion limit ($V_{E-C} = -0.9 \text{ V}$). The shaded region depicts the standard deviation. (g) CIC pulse width dependency (a/T^b , where T is the pulse width) for PEDOT:PSS/DES ($r^2 = 0.971$) and PEDOT:PSS ($r^2 = 0.999$) electrodes (100 μm diameter). (h) CIC electrode diameter scaling dependency (a/D^b) for PEDOT:PSS/DES ($r^2 = 0.998$) and PEDOT:PSS ($r^2 = 0.992$) determined for a fixed pulse width of 0.5 ms. (i) Optical microscopy image of the Van der Pauw structure for conductivity measurements. Metal pads (Ti/Au, yellow) are contacting the edges of the PEDOT:PSS square (dark blue) with a side width of 500 μm . (j) Sheet resistance and conductivity of the PEDOT:PSS/DES and PEDOT:PSS films obtained from multiple measurements ($N = 10$).



The enhanced recording capabilities of PEDOT:PSS/DES electrodes are also maintained for smaller electrode dimensions (Fig. 2(b) and Fig. S2). For instance, 25 μm PEDOT:PSS/DES electrodes feature an impedance of $31.5 \pm 10.6 \text{ k}\Omega$, which is still comparable to the impedance of the large PEDOT:PSS electrodes. These results highlight the potential of PEDOT:PSS/DES for enhancing the resolution of neural interfaces, as it can enable the utilisation of smaller electrodes without sacrificing the quality of the recording.

The lower impedance of PEDOT:PSS/DES originates from the high specific capacitance of the electrodes. To estimate the capacitance, we utilized the low frequency spectrum of the EIS data and fitted a simple Randles circuit to model the electrode-electrolyte interface.²⁶ This resulted in a specific capacitance of $2.22 \pm 0.05 \text{ mF cm}^{-2}$ for PEDOT:PSS/DES electrodes and $0.65 \pm 0.09 \text{ mF cm}^{-2}$ for PEDOT:PSS. Taking into account the film thickness as estimated from the FIB cuts, the equivalent specific volumetric capacitance C^* amounts to 40 F cm^{-3} and 38 F cm^{-3} for PEDOT:PSS/DES and PEDOT:PSS, respectively. These values are on the lower range compared with existing literature on PEDOT:PSS. For instance, Rivnay *et al.*⁶⁹ reported a C^* of 39 F cm^{-3} while the value estimated by Biachni *et al.*⁶⁵ reached 170 F cm^{-3} . However, it should be noted that the calculated values likely underestimate the actual volumetric capacitance, since the electrode coating was thinned down during RIE, especially in the centre of the electrodes. This effect is difficult to quantify, as the amount of over-etching depends on the position of the devices on the wafer. Nevertheless, these results suggest that the volumetric capacitance of PEDOT:PSS/DES films is comparable to conventional PEDOT:PSS, and the enhanced electrochemical performance rather stems from its capability of creating thicker electrode films. Electrochemical enhancement dependence on film thickness was further confirmed when characterising PEDOT:PSS/DES films of varying thicknesses, which exhibited a consistent scaling behaviour of specific capacitance and CSC from approximately 500 nm up to up to 1.4 μm (Fig. S3).

To further elucidate the electrochemical characteristics of the materials, we performed CV with potential limits of -0.9 V and 0.6 V , representing the water-window for PEDOT-based electrodes⁷⁰ (Fig. 2(c) and Fig. S4). Both PEDOT:PSS/DES and PEDOT:PSS electrodes feature a nearly rectangular shape of the voltammogram, highlighting the pseudocapacitive property of the films. A prominent cathodic peak can be observed for PEDOT:PSS at a potential of roughly -0.7 V , which has been attributed to the injection of electrons in the PEDOT coupled with incorporation of cations into the film.⁷¹ Notably, this peak is shifted to a lower potential of -0.4 V in the case of PEDOT:PSS/DES. The cathodic CSC_c was calculated for both materials by integrating the cathodic charge in the CV curve (Fig. 2(d)), revealing an increase in the CSC_c by a factor of 2.3 for 100 μm PEDOT:PSS/DES ($2.98 \pm 0.19 \text{ mC cm}^{-2}$) electrodes compared to PEDOT:PSS ($1.27 \pm 0.21 \text{ mC cm}^{-2}$) electrodes. The CSC_c furthermore increased significantly with the reduction in electrode diameter, reaching $6.48 \pm 1.43 \text{ mC cm}^{-2}$ for PEDOT:PSS/DES and $3.83 \pm 1.01 \text{ mC cm}^{-2}$ for PEDOT:PSS, comparable to other works.⁵⁷

Since neural stimulation employs fast pulses in the micro- to millisecond range, only a small fraction of the CSC can be safely delivered during a stimulation pulse.⁷² We therefore determined the CIC of the materials by recording the voltage excursions of the electrodes during biphasic current stimulation using symmetric square charge-balanced pulses with a leading cathodic polarity (Fig. S5 and S6). Fig. 2(e) depicts the voltage transient of a 100 μm PEDOT:PSS electrode while applying a 0.5 ms current pulse with an amplitude of 36 μA . The maximum cathodic voltage excursion reaches the cathodic limit of -0.9 V , indicating that the electrode is already operating at the current injection limit. In contrast, a PEDOT:PSS/DES electrode with the same size is polarized only up to -0.25 V while injecting the same amount of current. This suggests that PEDOT:PSS/DES electrodes can deliver the same amount of charge at reduced potentials which translates into improved safety and lower energy requirements for neural stimulation.

On average, the PEDOT:PSS/DES electrodes demonstrated a limiting current of $77.3 \pm 9.7 \mu\text{A}$ before reaching the cathodic polarization limit (Fig. 2(f)) which corresponds to a CIC of $0.49 \pm 0.06 \text{ mC cm}^{-2}$, about 2.5 times higher compared to PEDOT:PSS ($0.19 \pm 0.05 \text{ mC cm}^{-2}$). The CIC increased consistently for longer phase durations, exceeding 0.7 mC cm^{-2} at 5 ms for 100 μm PEDOT:PSS/DES, compared to just over 0.3 mC cm^{-2} for a 100 μs pulse width (Fig. 2(g)). This behaviour is consistent with previous findings⁵⁷ and suggests that prolonged phase durations permit higher charge densities. Similarly to the CSC, the CIC also exhibits a strong dependency on the electrode size, reaching $1.46 \pm 0.26 \text{ mC cm}^{-2}$ and $0.6 \pm 0.04 \text{ mC cm}^{-2}$ at 25 μm for PEDOT:PSS/DES and PEDOT:PSS, respectively (Fig. 2(h)). It should be noted that the CIC values for PEDOT:PSS are relatively low compared to other reports,⁵⁷ however these discrepancies can be attributed to a combination of factors, such as variations in the film thickness, differences in the processing of the film (*i.e.*, dry etching *vs.* stencil patterning), over-etching of the film, as well as to the absence of metal contacts under the electrodes. For instance, Cho *et al.*⁷³ recently reported a CIC of 0.7 mC cm^{-2} for 0.4 ms pulses using bare PEDOT:PSS electrodes, which closely resembles the values found in this work. The size-dependent CIC, on the other hand, can be related to edge effects,⁵⁷ which are likely exacerbated by the absence of a highly conductive metallic surface under the electrodes, as well as to the contribution of the PEDOT:PSS feedline, which was kept at a constant width for all electrode sizes.

Compared to typical inorganic electrode materials used in implantable neural probes, PEDOT:PSS/DES electrodes in this work outperform significantly planar metal electrodes such as those based on Pt and platinum-iridium (PtIr) alloys, which typically exhibit a CIC on the order of $0.023\text{--}0.15 \text{ mC cm}^{-2}$, recorded for pulse widths of 0.2 and 0.1 ms.^{74,75} In contrast, the CIC of titanium nitride (TiN) electrodes has been reported to be around 0.9 mC cm^{-2} for 0.5 ms pulses,⁷⁶ whereas values for iridium oxide electrodes vary widely from 0.75 up to 9 mC cm^{-2} ,^{77–80} with a CIC of 1.9 mC cm^{-2} reported for 20 μm sputtered iridium oxide (SIROF) electrodes upon application of 0.2 ms pulses without an interphase gap.⁸⁰



Furthermore, reduced graphene oxide (rGO) electrodes were recently reported, featuring a CIC of 1.17 mC cm^{-2} for $100 \mu\text{m}$ and 3.94 mC cm^{-2} for $25 \mu\text{m}$ electrodes using 1 ms cathodic pulses.⁸¹

The highest CIC recorded in this work was roughly 1.5 mC cm^{-2} , obtained for $25 \mu\text{m}$ electrodes with a pulse duration of 0.5 ms , demonstrating a competitive performance relative to other state-of-the-art electrode materials. Further optimisation of the fabrication methods, particularly addressing the thinning and oxidation of the PEDOT:PSS/DES coating during the RIE step, for instance by utilizing an etch-stop or a photo patternable passivation layer, could lead to additional improvements in device performance. In addition, it should be noted that in this work we utilized the initial segment of the voltage transients to estimate the potential drop over the access resistance,^{66,82} which is essential for calculating the electrode polarization and the CIC. However, this method can potentially underestimate the true CIC, as the access resistance can increase within the duration of the cathodic pulse due to reduction of the PEDOT, resulting in an increased resistivity of the electrodes.⁷⁰ Further experiments are thus needed to evaluate the contribution of this effect for the case of PEDOT:PSS/DES electrodes.

Finally, the conductivity of the films was estimated using the Van der Pauw method. To this end, we utilized dedicated structures featuring gold pads contacting the edges of a $500 \mu\text{m}$ square of conductive film (Fig. 2(i)). The measurements yielded a conductivity of $231.8 \pm 8.83 \text{ S cm}^{-1}$ for the standard PEDOT:PSS blend (Fig. 2(j)), in good agreement with other reports on PEDOT:PSS films with a similar composition.⁵³ In this regard, the PEDOT:PSS outperformed the novel PEDOT:PSS/DES blend, which proved to be roughly 3 times less conductive ($78.4 \pm 40.9 \text{ S cm}^{-1}$). The latter is expected, given that additives such as EG and DBSA have been shown to increase the conductivity of PEDOT:PSS films, whereas GOPS has the opposite effect,⁸³ resulting in films that are less conductive than PEDOT:PSS/DES blends without GOPS, as reported previously.⁵⁸ Additionally, we observed a large variability in the film conductivity across the wafer, suggesting that some inhomogeneities in the deposited film might be present. Nevertheless, the sheet resistance of the PEDOT:PSS/DES was comparable to PEDOT:PSS ($211.3 \pm 91.9 \Omega \text{ sq}^{-1}$ vs. $172.8 \pm 2.44 \Omega \text{ sq}^{-1}$), indicating that the increased thickness of the PEDOT:PSS/DES film is sufficient to compensate for the reduced conductivity.

Enhanced electrochemical properties of PEDOT:PSS/DES translate to superior signal recording and stimulation efficiency

To evaluate the functional relevance of the enhanced electrochemical properties observed *in vitro*, we performed *in vivo* experiments using the cuff multielectrode arrays described in the previous section, comparing PEDOT:PSS/DES and conventional PEDOT:PSS electrodes for neural recording and stimulation of the rat sciatic nerve in non-recovery implantations (Fig. 3(a)). PEDOT:PSS/DES electrodes demonstrated superior performance compared to PEDOT:PSS across multiple

electrophysiological metrics, including impedance characteristics, signal acquisition, and stimulation efficiency.

Impedance magnitude and phase values extracted from baseline recordings did not differ significantly between materials (Fig. 3(b)). However, PEDOT:PSS/DES consistently exhibited lower impedance magnitude across most animals. When modelled using a series RC circuit, implanted PEDOT:PSS/DES electrodes showed significantly lower resistance and higher capacitance than PEDOT:PSS electrodes (Fig. 3(c)), indicating improved charge transfer and capacitive behaviour. The influence of stray capacitance from the experimental setup cannot be excluded and may affect impedance measurements. As a result, direct comparisons between *in vitro* and *in vivo* values are not encouraged. Here, all *in vivo* comparisons were performed between paired cuffs implanted simultaneously in the same animal to ensure consistency and control for inter-animal variability.

The improved electrochemical and conformational properties of PEDOT:PSS/DES translated into enhanced neural signal acquisition (Fig. 3(d)). SNR was higher for PEDOT:PSS/DES in the majority of trials, suggesting greater neural recording quality (Fig. 3(d) and (e)), although this trend did not reach statistical significance using a paired *t*-test. Only two trials showed superior performance for PEDOT:PSS, one of which exhibited similar impedance and capacitance across materials. This suggests that while the materials' electrochemical and mechanical properties are important, they are not the sole determinant of the recording and stimulation efficacy. Other factors, such as the proximity of the electrodes to motor and sensory fascicles, play a significant role in shaping functional outcomes. Additionally, spike amplitude was generally greater, although not significantly in statistical terms, with PEDOT:PSS/DES in five out of seven trials, further supporting its enhanced recording capability (Fig. 3(e)).

Stimulation responses, as assessed by EMG RMS amplitude across increasing current levels, were broadly similar between PEDOT:PSS/DES and PEDOT:PSS electrodes. However, PEDOT:PSS/DES more reliably elicited visible motor responses at lower current thresholds, indicating improved recruitment of the tibialis anterior muscle (Fig. 3(f)). This advantage was generally not reflected in EMG signal magnitude at matched currents (Fig. S7), where statistical comparisons at $10 \mu\text{A}$ and $30 \mu\text{A}$ stimulation currents using paired *t*-tests did not reveal significant differences between materials ($p = 0.05$). However, clearer distinctions emerged in some cases at lower current levels (Fig. 3(g), corresponding to Rat 1-RN).

The improved stimulation efficacy of PEDOT:PSS/DES is likely related to its enhanced electrochemical properties, including lower series resistance and higher capacitance, which support more efficient charge delivery and better coupling at the electrode–tissue interface. These characteristics may reduce the current required to reach activation thresholds, even when EMG output remains comparable.

Notably, the only case in which PEDOT:PSS outperformed PEDOT:PSS/DES corresponded to a situation where PEDOT:PSS had clearly superior electrochemical properties (higher capacitance






Fig. 3 *In vivo* evaluation of PEDOT:PSS and PEDOT:PSS/DES devices. (a) Schematic of the experimental setup showing surgical implantation at the sciatic nerve with side-by-side placement of the two cuffs. Inset: Magnified view of the cuff placement on the sciatic nerve. Bottom: Photograph of the bonded electrode assembly with FFC connector interfacing with a custom-made PCB, which connects to the Intan RHS headstage. EMG needle electrodes were placed in the tibialis anterior muscle. All signals were recorded via the Intan RHS system. (b) Impedance magnitude and phase at 1 kHz for PEDOT:PSS/DES (green) and PEDOT:PSS (blue). (c) Resistance and capacitance values extracted from equivalent series RC circuit modeling, revealing statistically significant differences between materials (paired *t*-test, *p* < 0.05). (d) Trial-wise difference in signal-to-noise ratio (SNR) between materials (PEDOT:PSS/DES – PEDOT:PSS) for both right and left sciatic nerves across four rats (8 independent trials). (e) Representative neural recordings from PEDOT:PSS/DES and PEDOT:PSS during evoked pain activity, with corresponding average spike waveforms. (f) Differences in stimulation threshold (minimum current to evoke a visible motor twitch) between materials. (g) Representative EMG RMS responses at increasing stimulation currents for each device. Each point represents the average RMS over 20 stimulation pulses (pulse duration: 100 μs). Red dashed line represents EMG RMS at baseline prior to stimulation. Vertical dashed lines represent the currents at which a motor response was visually observed.

and lower impedance), consistent with expected trends. However, in the case where PEDOT:PSS/DES showed the largest functional advantage, the difference in electrochemical performance compared to PEDOT:PSS was relatively modest. In contrast, other recordings with more dramatic electrochemical differences showed smaller functional effects, indicating, once again, that local interface conditions may influence performance. Comparing the magnitude of differences across trials is not advised, as each trial may involve slightly different electrodes in terms of number and placement. It is more appropriate to focus on differences between materials within a single trial, where electrode configuration and placement remain more consistent.

Conclusions

This study demonstrates that incorporating DES (ChCl:LAC) into PEDOT:PSS formulations offers a powerful strategy to enhance the electrochemical performance of implantable bioelectronic interfaces. We developed a PEDOT:PSS/DES formulation that facilitates the processing of thick, conductive polymer, gel-like film electrodes, compatible with established microfabrication workflows based on thin-film technology. The enhanced

thickness of PEDOT:PSS/DES electrodes resulted in exceptional electrochemical characteristics, paving the way for miniaturized, higher-density neural interfaces without sacrificing functionality.

The use of eutectogels opens a new venue in the tailored modification of PEDOT:PSS, such as anticorrosion resistance, mixed ionic–electronic conductivity, and long-term mechanical stability, features that are critical for chronic implantation and multi-modal bioelectronic therapies. Future work will focus on evaluating these properties under long-term electrophysiology setups, and the use of alternative DES chemistries to further expand the functional landscape of PEDOT-based materials for implantable technologies.

Author contributions

M. K., R. R. M., A. D. A., V. R. M., and A. G. G. conceived and planned the study. R. R. M. and A. D. A. developed the PEDOT:PSS/DES formulation. A. D. A., M. L. P., and D. M. provided the DES. M. K. and V. R. M. established the micro-fabrication process-flow, fabricated the probes, and carried out the electrochemical and electrical characterisations. M. K. processed the data of *in vitro* characterisations with the support of V. R. M. C. L. and A. G. G. carried out the *in vivo* validation.



A. G. G. developed the custom code and analysed the *in vivo* data. V. R. M. and A. G. G. provided overall guidance and direction for the study. D. M. and G. G. M. contributed laboratory facilities, equipment, and other resources required for the experimental work. M. K., R. R. M., A. D. A., V. R. M., and A. G. G. wrote the initial draft of the manuscript. All authors reviewed and edited the manuscript.

Conflicts of interest

There are no conflicts to declare.

Data availability

The data supporting the electrochemical characterisation, including impedance spectroscopy, cyclic voltammetry, voltage transients, and stimulation efficacy, have been included as part of the supplementary information (SI). Supplementary information is available. See DOI: <https://doi.org/10.1039/d5tc03054d>.

All data collected for the animal work, including the recording and stimulation electrophysiology datasets, are available at Zenodo public repository with DOI: <https://doi.org/10.5281/zenodo.16746532>.⁸⁴ The code for analysing the electrophysiology data can be found at Zenodo with DOI: <https://doi.org/10.5281/zenodo.16763365>.⁸⁵ The version of the code employed for this study is version V1.

Acknowledgements

The authors thank the Helmholtz Nano Facility (HNF) at Forschungszentrum Jülich for facilitating the microfabrication and S. Decke, E. Yilmaz, and R. Stockmann for providing support during microfabrication. The authors also thank E. Brauweiler-Reuters for carrying out FIB sectioning and SEM. The authors thank A. Offenhäusser for infrastructural and scientific support. R. R. M. acknowledges support from EPSRC grant [EP/S022139/1]. G. G. M. and A. D. A. acknowledge support from Horizon Europe UKRI Underwrite Innovate grant COPE-Nano under project number 10078978. M. L. P. and D. M. gratefully acknowledge the financial support from IKERBASQUE-Basque Foundation for Science and the Marie Skłodowska-Curie Research and Innovation Staff Exchanges program under grant agreement IONBIKE 2.0 MSCA-SE 101129945. A. G. acknowledges support from the Royal Academy of Engineering and Rosetrees Trust through grant no. RF-2324-23-284. This work was funded by UKRI grants and was supported by the Deutsche Forschungsgemeinschaft (DFG, German Research Foundation; GRK2610 (project number 424556709)). For the purpose of open access, the author has applied a Creative Commons Attribution (CC BY) licence to any Author Accepted Manuscript version arising.

References

- 1 M. Cook, M. Murphy, K. Bulluss, W. D'Souza, C. Plummer, E. Priest, C. Williams, A. Sharan, R. Fisher, S. Pincus, E. Distad, T. Anchordoquy and D. Abrams, *EClinicalMedicine*, 2020, **22**, 100326.
- 2 T. Lenarz, H.-W. Pau and G. Paasche, *Curr. Pharm. Biotechnol.*, 2013, **14**, 112–123.
- 3 S. Boveda, S. Garrigue and P. Ritter, in *Dawn and Evolution of Cardiac Procedures: Research Avenues in Cardiac Surgery and Interventional Cardiology*, ed. M. Picichè, Springer, Milan, Milano, 2013, pp. 253–264.
- 4 M. Arlotti, M. Colombo, A. Bonfanti, T. Mandat, M. M. Lanotte, E. Pirola, L. Borellini, P. Rampini, R. Eleopra, S. Rinaldo, L. Romito, M. L. F. Janssen, A. Priori and S. Marceglia, *Front. Neurosci.*, 2021, **15**, 763235.
- 5 H. C. Koydemir and A. Ozcan, *Annu Rev. Anal. Chem.*, 2018, **11**, 127–146.
- 6 A. J. Boys, A. Carnicer-Lombarte, A. Güemes-Gonzalez, D. C. van Niekerk, S. Hilton, D. G. Barone, C. M. Proctor, R. M. Owens and G. G. Malliaras, *Adv. Mater.*, 2023, **35**(8), e2207847.
- 7 L. Koschinski, T. Grap, E. Yilmaz, M. Kleutgens, S. Decke, M. Kasavetov, M. Jung, A. Carnicer-Lombarte, G. Malliaras, A. Offenhäusser and V. Rincón Montes, *Adv. Electron. Mater.*, 2025, **11**(17), 2500088.
- 8 S. Middya, A. Carnicer-Lombarte, S. Sawiak, S. Hilton, V. F. Curto, D. G. Barone, G. S. K. Schierle and G. G. Malliaras, *Adv. Funct. Mater.*, 2025, **35**, 2417312.
- 9 A. Carnicer-Lombarte, A. J. Boys, A. Güemes, J. Gurke, S. Velasco-Bosom, S. Hilton, D. G. Barone and G. G. Malliaras, *Nat. Commun.*, 2024, **15**, 7523.
- 10 B. J. Woodington, J. Lei, A. Carnicer-Lombarte, A. Güemes-González, T. E. Naegele, S. Hilton, S. El-Hadwe, R. A. Trivedi, G. G. Malliaras and D. G. Barone, *Sci. Adv.*, 2024, **10**(19), DOI: [10.1126/sciadv.adl1230](https://doi.org/10.1126/sciadv.adl1230).
- 11 A. J. Boys, A. Güemes, L. Ma, R. A. Gupta, Z. Lu, C. Lee, S. El-Hadwe, A. Carnicer-Lombarte, T. E. Naegele, F. Uhlig, D. G. Barone, D. C. Bulmer, J. N. Gelinas, N. P. Hyland, D. Khodagholi, G. G. Malliaras and R. M. Owens, *bioRxiv*, 2025, preprint, DOI: [10.1101/2024.03.22.586292](https://doi.org/10.1101/2024.03.22.586292).
- 12 M. Jung, J. Abu Shihada, S. Decke, L. Koschinski, P. S. Graff, S. Maruri Pazmino, A. Höllig, H. Koch, S. Musall, A. Offenhäusser and V. Rincón Montes, *Adv. Mater.*, 2025, **37**, 2418524.
- 13 C. Orlemann, C. Boehler, R. N. Kooijmans, B. Li, M. Asplund and P. R. Roelfsema, *Adv. Healthcare Mater.*, 2024, **13**, 2304169.
- 14 S. Middya, A. Carnicer-Lombarte, V. F. Curto, S. Hilton, A. Genewsky, A. L. Rutz, D. G. Barone, G. S. Kaminski Schierle, A. Sirota and G. G. Malliaras, *Adv. Electron. Mater.*, 2023, **9**, 2200883.
- 15 J. Rivnay, S. Inal, B. A. Collins, M. Sessolo, E. Stavrinidou, X. Strakosas, C. Tassone, D. M. Delongchamp and G. G. Malliaras, *Nat. Commun.*, 2016, **7**, 11287.
- 16 S. Roubert Martinez, P. Le Floch, J. Liu and R. D. Howe, *Adv. Healthcare Mater.*, 2023, **12**(17), e2202661, DOI: [10.1002/adhm.202202661](https://doi.org/10.1002/adhm.202202661), (accessed August 7, 2025).
- 17 J. Heikenfeld, A. Jajack, J. Rogers, P. Gutruf, L. Tian, T. Pan, R. Li, M. Khine, J. Kim, J. Wang and J. Kim, *Lab Chip*, 2018, **18**, 217–248.



- 18 A. Dominguez-Alfaro, E. Mitoudi-Vagourdi, I. Dimov, M. L. Picchio, N. Lopez-Larrea, J. L. de Lacalle, X. Tao, R. R.-M. Serrano, A. Gallastegui, N. Vassardanis, D. Mecerreyes and G. G. Malliaras, *Adv. Sci.*, 2024, **11**, 2306424.
- 19 P. Oldroyd and G. G. Malliaras, *Acta Biomater.*, 2022, **139**, 65–81.
- 20 G. Dijk, J. Pas, K. Markovic, J. Scancar and R. P. O'Connor, *APL Bioeng.*, 2023, **7**, 046117.
- 21 K. Srikantharajah, RWTH Aachen University.
- 22 P. Oldroyd, J. Gurke and G. G. Malliaras, *Adv. Funct. Mater.*, 2023, **33**, 2208881.
- 23 M. Ganji, A. Tanaka, V. Gilja, E. Halgren and S. A. Dayeh, *Adv. Funct. Mater.*, 2017, **27**, 1703019.
- 24 A. S. Pranti, A. Schander, A. Bödecker and W. Lang, *Proceedings*, 2017, **1**, 492.
- 25 B. Charlot, G. Sassine, A. Garraud, B. Sorli, A. Giani and P. Combette, *Microsyst. Technol.*, 2013, **19**, 895–903.
- 26 G. Dijk, H. J. Ruigrok and R. P. O'Connor, *Adv. Mater. Interfaces*, 2020, **7**, 2000675.
- 27 A. Güemes Gonzalez, A. Carnicer-Lombarte, S. Hilton and G. Malliaras, *J. Neural Eng.*, 2023, **20**, 056033.
- 28 E. L. Smith, A. P. Abbott and K. S. Ryder, *Chem. Rev.*, 2014, **114**(21), 11060–11082, DOI: [10.1021/cr300162p](https://doi.org/10.1021/cr300162p), (accessed August 7, 2025).
- 29 L. C. Tomé and D. Mecerreyes, *J. Phys. Chem. B*, 2020, **124**(39), 8465–8478, DOI: [10.1021/acs.jpcc.0c04769](https://doi.org/10.1021/acs.jpcc.0c04769), (accessed August 11, 2025).
- 30 I. Abdel Aziz and D. Mecerreyes, *Prog. Polym. Sci.*, 2025, **167**, 101994.
- 31 D. O. Abranches and J. A. P. Coutinho, *Annu Rev. Chem. Biomol. Eng.*, 2023, **14**, 141–163, DOI: [10.1146/annurev-chembioeng-101121-085323](https://doi.org/10.1146/annurev-chembioeng-101121-085323), (accessed August 11, 2025).
- 32 P. A. Mercadal, A. González, A. Beloqui, L. C. Tomé, D. Mecerreyes, M. Calderón and M. L. Picchio, *JACS Au*, 2024, **4**(10), 3744–3758, DOI: [10.1021/jacsau.4c00677](https://doi.org/10.1021/jacsau.4c00677), (accessed August 11, 2025).
- 33 N. Nining, Y. W. Wardhana, T. Rusdiana and H. Hariyanti, *ChemistryOpen*, 2025, 2500332.
- 34 R. L. Keate, J. Tropp, C. P. Collins, H. O. T. Ware, A. J. Petty II, G. A. Ameer, C. Sun and J. Rivnay, *Macromol. Biosci.*, 2022, **22**, 2200103.
- 35 A. Aguzin, A. Dominguez-Alfaro, M. Criado-Gonzalez, S. Velasco-Bosom, M. L. Picchio, N. Casado, E. Mitoudi-Vagourdi, R. J. Minari, G. G. Malliaras and D. Mecerreyes, *Mater. Horiz.*, 2023, **10**, 2516–2524.
- 36 Z. Zhu, C. Liu, Q. Jiang, H. Shi, J. Xu, F. Jiang, J. Xiong and E. Liu, *Synth. Met.*, 2015, **209**, 313–318.
- 37 Z. Zhu, C. Liu, H. Shi, Q. Jiang, J. Xu, F. Jiang, J. Xiong and E. Liu, *J. Polym. Sci., Part B: Polym. Phys.*, 2015, **53**, 885–892.
- 38 Y. Lu, R. Liu, X.-C. Hang and D. J. Young, *Polym. Chem.*, 2021, **12**, 2115–2121.
- 39 N. Casado, S. Zendegi, L. C. Tomé, S. Velasco-Bosom, A. Aguzin, M. Picchio, M. Criado-Gonzalez, G. G. Malliaras, M. Forsyth and D. Mecerreyes, *J. Mater. Chem. C*, 2022, **10**, 15186–15193.
- 40 S. Velasco-Bosom, N. Karam, A. Carnicer-Lombarte, J. Gurke, N. Casado, L. C. Tomé, D. Mecerreyes and G. G. Malliaras, *Adv. Healthcare Mater.*, 2021, **10**, 2100374.
- 41 A. Aguzin, A. Dominguez-Alfaro, M. Criado-Gonzalez, S. Velasco-Bosom, M. L. Picchio, N. Casado, E. Mitoudi-Vagourdi, R. J. Minari, G. G. Malliaras and D. Mecerreyes, *Mater. Horiz.*, 2023, **10**, 2516–2524.
- 42 B. Joos, T. Vranken, W. Marchal, M. Safari, M. K. Van Bael and A. T. Hardy, *Chem. Mater.*, 2018, **30**, 655–662.
- 43 M. L. Picchio, A. Gallastegui, N. Casado, N. Lopez-Larrea, B. Marchiori, I. del Agua, M. Criado-Gonzalez, D. Mantione, R. J. Minari and D. Mecerreyes, *Adv. Mater. Technol.*, 2022, **7**(10), 2101680, DOI: [10.1002/admt.202101680](https://doi.org/10.1002/admt.202101680), (accessed August 7, 2025).
- 44 M. Criado-Gonzalez, N. Alegret, A. M. Fracaroli, D. Mantione, G. Guzmán-González, R. Del Olmo, K. Tashiro, L. C. Tomé, M. L. Picchio and D. Mecerreyes, *Angew. Chem., Int. Ed.*, 2023, **62**(26), e202301489, DOI: [10.1002/anie.202301489](https://doi.org/10.1002/anie.202301489), (accessed August 7, 2025).
- 45 J. L. de Lacalle, M. L. Picchio, A. Dominguez-Alfaro, R. R. Serrano, B. Marchiori, I. D. Agua, N. Lopez-Larrea, M. Criado-Gonzalez, G. G. Malliaras and D. Mecerreyes, *ACS Mater. Lett.*, 2023, **5**(12), 3340–3346, DOI: [10.1021/acsmaterialslett.3c00938](https://doi.org/10.1021/acsmaterialslett.3c00938), (accessed August 7, 2025).
- 46 R. Ruiz-Mateos Serrano, A. Aguzin, E. Mitoudi-Vagourdi, X. Tao, T. E. Naegle, A. T. Jin, N. Lopez-Larrea, M. L. Picchio, M. Vinicio Alban-Paccha, R. J. Minari, D. Mecerreyes, A. Dominguez-Alfaro and G. G. Malliaras, *Biomaterials*, 2024, **310**, 122624.
- 47 C. Slaughter, S. Velasco-Bosom, X. Tao, R. R.-M. Serrano, S. Kissovsky, R. Mizuta, D. Mantione, S. T. Keene, G. G. Malliaras and A. Dominguez-Alfaro, *J. Mater. Chem. C*, 2024, **12**, 14944–14954.
- 48 S. Locatelli, G. C. Luque, R. Ruiz-Mateos Serrano, A. Dominguez-Alfaro, G. Reniero, M. L. Picchio, J. Leiva, L. M. Gugliotta, G. G. Malliaras, D. Mecerreyes, L. I. Ronco and R. J. Minari, *ACS Appl. Polym. Mater.*, 2025, **7**, 2945–2954.
- 49 Y. J. Jo, H. Kim, J. Ok, Y.-J. Shin, J. H. Shin, T. H. Kim, Y. Jung and T. Kim, *Adv. Funct. Mater.*, 2020, **30**, 1909707.
- 50 R. R.-M. Serrano, S. Velasco-Bosom, A. Dominguez-Alfaro, M. L. Picchio, D. Mantione, D. Mecerreyes and G. G. Malliaras, *Adv. Sci.*, 2024, **11**, 2301176.
- 51 P. Oldroyd, S. Velasco-Bosom, S. L. Bidinger, T. Hasan, A. J. Boys and G. G. Malliaras, *Nat. Protoc.*, 2025, **20**, 2100–2124.
- 52 W. Albrecht, J. Moers and B. Hermanns, *J. Large-Scale Res. Facil.*, 2017, **3**, A112.
- 53 L. Koschinski, B. Lenyk, M. Jung, I. Lenzi, B. Kampa, D. Mayer, A. Offenhäusser, S. Musall and V. Rincón Montes, *J. Mater. Chem. B*, 2023, **11**, 9639–9657.
- 54 A. J. Boys, A. Güemes, L. Ma, R. A. Gupta, Z. Lu, C. Lee, S. El-Hadwe, A. Carnicer-Lombarte, T. E. Naegle, F. Uhlig, D. G. Barone, D. C. Bulmer, J. N. Gelinis, D. Khodagholy, G. G. Malliaras and R. M. Owens.
- 55 M. Jung, A. Willuweit and V. Rincón Montes, *Sensors*, 2025, **25**, 3765.



- 56 M. Mierzejewski, H. Steins, P. Kshirsagar and P. D. Jones, *J. Neural Eng.*, 2020, **17**, 052001.
- 57 M. Ganji, A. Tanaka, V. Gilja, E. Halgren and S. A. Dayeh, *Adv. Funct. Mater.*, 2017, **27**, 1703019.
- 58 A. Aguzin, A. Dominguez-Alfaro, M. Criado-Gonzalez, S. Velasco-Bosom, M. L. Picchio, N. Casado, E. Mitoudi-Vagourdi, R. J. Minari, G. G. Malliaras and D. Mecerreyes, *Mater. Horiz.*, 2023, **10**, 2516–2524.
- 59 X. Crispin, F. L. E. Jakobsson, A. Crispin, P. C. M. Grim, P. Andersson, A. Volodin, C. van Haesendonck, M. Van der Auweraer, W. R. Salaneck and M. Berggren, *Chem. Mater.*, 2006, **18**(18), 4354–4360, DOI: [10.1021/cm061032%2B](https://doi.org/10.1021/cm061032%2B), (accessed August 7, 2025).
- 60 J. Ouyang, Q. Xu, C.-W. Chu, Y. Yang, G. Li and J. Shinar, *Polymer*, 2004, **45**, 8443–8450.
- 61 G.-H. Kim, L. Shao, K. Zhang and K. P. Pipe, *Nat. Mater.*, 2013, **12**, 719–723.
- 62 J. Ouyang, C.-W. Chu, F.-C. Chen, Q. Xu and Y. Yang, *Adv. Funct. Mater.*, 2005, **15**, 203–208.
- 63 M. N. Gueye, A. Carella, J. Faure-Vincent, R. Demadrille and J.-P. Simonato, *Prog. Mater. Sci.*, 2020, **108**, 100616.
- 64 L. C. Tomé and D. Mecerreyes, *J. Phys. Chem. B*, 2020, **124**, 8465–8478.
- 65 M. Bianchi, S. Carli, M. Di Lauro, M. Prato, M. Murgia, L. Fadiga and F. Biscarini, *J. Mater. Chem. C*, 2020, **8**, 11252–11262.
- 66 C. Boehler, S. Carli, L. Fadiga, T. Stieglitz and M. Asplund, *Nat. Protoc.*, 2020, **15**, 3557–3578.
- 67 M. Ganji, A. T. Elthakeb, A. Tanaka, V. Gilja, E. Halgren and S. A. Dayeh, *Adv. Funct. Mater.*, 2017, **27**, 1703018.
- 68 H. G. Rey, C. Pedreira and R. Quiñan Quiroga, *Brain Res. Bull.*, 2015, **119**, 106–117.
- 69 J. Rivnay, P. Leleux, M. Ferro, M. Sessolo, A. Williamson, D. A. Koutsouras, D. Khodagholy, M. Ramuz, X. Strakosas, R. M. Owens, C. Benar, J.-M. Badier, C. Bernard and G. G. Malliaras, *Sci. Adv.*, 2015, **1**(4), e1400251.
- 70 S. Cogan, *Annu. Rev. Biomed. Eng.*, 2008, **10**, 275–309.
- 71 A. V. Volkov, K. Wijeratne, E. Mitraka, U. Ail, D. Zhao, K. Tybrandt, J. W. Andreasen, M. Berggren, X. Crispin and I. V. Zozoulenko, *Adv. Funct. Mater.*, 2017, **27**, 1700329.
- 72 Z. Aqrawe, B. Wright, N. Patel, Y. Vyas, J. Malmstrom, J. M. Montgomery, D. Williams, J. Travas-Sejdic and D. Svirskis, *Sens. Actuators, B*, 2019, **281**, 549–560.
- 73 Y. U. Cho, K. Kim, A. Dutta, S. H. Park, J. Y. Lee, H. W. Kim, J. Park, J. Kim, W. K. Min, C. Won, J. Park, Y. Kim, G. Nan, J. Y. Kim, T. Lee, H. J. Kim, D. Kim, J. E. Lee, B.-W. Min, I.-J. Cho, B. H. Lee, H. Cheng, M. Cha and K. J. Yu, *Adv. Funct. Mater.*, 2024, **34**, 2310908.
- 74 T. L. Rose and L. S. Robblee, *IEEE Trans. Biomed. Eng.*, 1990, **37**, 1118–1120.
- 75 E. K. Brunton, B. Winther-Jensen, C. Wang, E. B. Yan, S. Hagh Gooie, A. J. Lowery and R. Rajan, *Front. Neurosci.*, 2015, **9**, 265.
- 76 J. D. Weiland, D. J. Anderson and M. S. Humayun, *IEEE Trans. Biomed. Eng.*, 2002, **49**, 1574–1579.
- 77 S. F. Cogan, J. Ehrlich, T. D. Plante, A. Smirnov, D. B. Shire, M. Gingerich and J. F. Rizzo, *J. Biomed. Mater. Res., Part B*, 2009, **89**, 353–361.
- 78 S. F. Cogan, T. D. Plante and J. Ehrlich, in *The 26th Annual International Conference of the IEEE Engineering in Medicine and Biology Society*, 2004, vol. 2, pp. 4153–4156.
- 79 Y. Lu, Z. Cai, Y. Cao, H. Yang and Y. Y. Duan, *Electrochem. Commun.*, 2008, **10**, 778–782.
- 80 A. Ghazavi, J. Maeng, M. Black, S. Salvi and S. F. Cogan, *J. Neural Eng.*, 2020, **17**, 016022.
- 81 F. T. Duvan, M. Cunquero, E. Masvidal-Codina, S. T. Walston, M. Marsal, J. M. de la Cruz, D. Viana, D. Nguyen, J. Degardin, X. Illa, J. M. Zhang, M. del Pilar Bernícola, J. G. Macias-Montero, C. Puigdemgole, G. Castro-Olvera, E. del Corro, S. Dokos, M. Chmeissani, P. Loza-Alvarez, S. Picaud and J. A. Garrido, *Nanoscale Horiz.*, 2024, **9**, 1948–1961.
- 82 A. Cisnal, J.-C. Fraile, J. Pérez-Turiel, V. Muñoz-Martínez, C. Müller and F. R. Ihmig, *Sensors*, 2018, **18**(12), 4152.
- 83 S. Zhang, P. Kumar, A. S. Nouas, L. Fontaine, H. Tang and F. Cicoira, *APL Mater.*, 2015, **3**, 014911.
- 84 A. Güemes, Electrophysiology datasets, Zenodo, 2025, DOI: [10.5281/zenodo.16746532](https://doi.org/10.5281/zenodo.16746532).
- 85 A. Güemes, Code Version V1, Zenodo, 2025, DOI: [10.5281/zenodo.16763365](https://doi.org/10.5281/zenodo.16763365).

