


# Journal of Materials Chemistry C

Materials for optical, magnetic and electronic devices

[rsc.li/materials-c](https://rsc.li/materials-c)



ISSN 2050-7526

**PAPER**

Cinzia Di Franco, Luisa Torsi, Gaetano Scamarcio *et al.*  
Kelvin probe force microscopy on patterned large-area  
biofunctionalized surfaces: a reliable ultrasensitive platform  
for biomarker detection

Cite this: *J. Mater. Chem. C*,  
2024, 12, 73Kelvin probe force microscopy on patterned  
large-area biofunctionalized surfaces: a reliable  
ultrasensitive platform for biomarker detection†Cinzia Di Franco,<sup>ib</sup> ‡<sup>a</sup> Matteo Piscitelli,<sup>‡</sup><sup>ab</sup> Eleonora Macchia,<sup>ib</sup> <sup>cd</sup>  
Cecilia Scandurra,<sup>ib</sup> <sup>e</sup> Michele Catacchio,<sup>e</sup> Luisa Torsi<sup>ib</sup> <sup>\*e</sup> and  
Gaetano Scamarcio<sup>ib</sup> <sup>\*ab</sup>

Kelvin probe force microscopy (KPFM) allows the detection of single binding events between immunoglobulins (IgM, IgG) and their cognate antibodies (anti-IgM, anti-IgG). Here an insight into the reliability and robustness of the methodology is provided. Our method is based on imaging the surface potential shift occurring on a dense layer of  $\sim 5 \times 10^7$  antibodies physisorbed on a  $50 \mu\text{m} \times 90 \mu\text{m}$  area when assayed with increasing concentrations of antigens in phosphate buffer saline (PBS) standard solutions, in air and at a fixed scanning location. A comprehensive investigation of the influence of the main experimental parameters that may interfere with the outcomes of KPFM immune-assay is provided, showing the robustness and reliability of our approach. The data are supported also by a thorough polarization modulation infrared reflection-absorption spectroscopy (PM-IRRAS) analysis of the physisorbed biolayer, in the spectral region of the amide I, amide II and amide A bands. Our findings demonstrate that a 10 min incubation in 500  $\mu\text{L}$  PBS encompassing  $\approx 30$  antigens (100 zM) triggers an extended surface potential shift that involves the whole investigated area. Such a shift quickly saturates at increasing ligand concentration, showing that the developed sensing platform works as an OFF/ON detector, capable of assessing the presence of a few specific biomarkers in a given assay volume. The reliability of the developed methodology KPFM is an important asset in single molecule detections at a wide electrode interface.

Received 29th August 2023,  
Accepted 18th November 2023

DOI: 10.1039/d3tc03110a

rsc.li/materials-c

## Introduction

Kelvin probe force microscopy (KPFM) is a powerful analytical technique that allows surface potential (SP) mapping with high spatial resolution by measuring the contact potential difference (CPD) between the metal coated tip of an atomic force microscope (AFM) operated in tapping mode and the sample surface.<sup>1</sup> It allows the simultaneous high-resolution morphological imaging of the sample surface and the mapping of the

electrostatic configuration, originating from the local distribution of surface dipoles. While KPFM is inherently a non-invasive technique, it is extremely sensitive to changes occurring in the top-most atomic layer structures, as caused by terminal group modifications, molecular bindings or deposition of adlayers.<sup>2–4</sup>

Several reports have shown the possibility to use KPFM to study biorecognition processes on biofunctionalized surfaces. For example, DNA either immobilized or physisorbed on Au coated Si substrates has been detected after exposure to dilutions in the  $\mu\text{M}$  range.<sup>2,5–7</sup> ATP-binding on protein kinase has been detected with sensitivity of 10 pM and DNA-bound  $\text{Hg}^+$  ions with sensitivity in the fM range.<sup>8,9</sup> KRAS and EGFR oncogene mutant DNA, bonded to Au nanoparticles, have been revealed with sensitivity in the pM range.<sup>10</sup> Recently, we have demonstrated that KPFM can be used to assess the surface potential shift<sup>11</sup> triggered by affinity binding on the biofunctionalized gate of a single molecule with a large transistor (SiMoT) sensor<sup>12,13</sup> incubated with ligand concentrations as low as  $10^{-19}$  mole  $\text{L}^{-1} \equiv 10^2$  zM, corresponding to the ultimate limit of single molecule detection.

<sup>a</sup> CNR – Institute of Photonics and Nanotechnologies, Via Amendola 173, 70126 Bari, Italy. E-mail: cinzia.difranco@cnr.it<sup>b</sup> Dipartimento Interateneo di Fisica “M. Merlin”, Università degli Studi di Bari Aldo Moro, 70126 Bari, Italy. E-mail: gaetano.scamarcio@uniba.it<sup>c</sup> Dipartimento di Farmacia-Scienze del Farmaco, Università degli Studi di Bari Aldo Moro, 70126 Bari, Italy<sup>d</sup> The Faculty of Science and Engineering, Åbo Akademi University, 20500 Turku, Finland<sup>e</sup> Dipartimento di Chimica, Università degli Studi di Bari Aldo Moro, Via E. Orabona 4, 70125 Bari, Italy. E-mail: luisa.torsi@uniba.it† Electronic supplementary information (ESI) available. See DOI: <https://doi.org/10.1039/d3tc03110a>

‡ Cinzia Di Franco and Matteo Piscitelli equally contributed to this work.

This work is focused on the development and optimization of a KPFM based sensing platform for the detection of immunoglobulin biomarkers on large-area biofunctionalized surfaces. Specifically, we demonstrate the detection of a few immunoglobulins (IgM, IgG) by mapping the surface potential of Au/anti-immunoglobulin M or G (anti-IgM, anti-IgG) interfaces with unprecedented sensitivity in the zeptomole range. Our method is based on the comparison of KPFM images taken in air at a fixed scanning location across the Au/biolyer interface, after 10 minute long sequential incubations at increasing concentrations of analyte in phosphate buffer saline (PBS) standard solutions followed by a drying procedure. This procedure inherently causes reproducibility issues, mainly due to surface contamination by adventitious carbon.<sup>14,15</sup> Hence, we have carried out a comprehensive investigation of the influence of the main experimental parameters that may interfere with the outcome of a conventional KPFM analysis, showing the efficacy and reliability of a method based on the simultaneous SP mapping of both the bare and the bio-functionalized Au areas. Our findings reveal that the formation of the immune-complex reliably triggers an extended SP shift due to a surface amplification effect. This approach paves the way for a fast, reliable, and qualitative analysis of biomolecular interactions.

## Materials and methods

### Materials

HPLC-grade water, hydrogen peroxide 30%, VLSI grade, and sulfuric acid 96% VLSI Selectipur grade were purchased from Avantor. Polyclonal antibodies: anti-human immunoglobulin M (anti-IgM), anti-human immunoglobulin G (anti-IgG), human IgM (~950 kDa), human IgG (~150 kDa) (Sigma-Aldrich), were used without additional purifications. Phosphate-buffered saline (PBS, Sigma-Aldrich) solution had ionic strength of 162 mM, pH 7.4, at room temperature.

### Biofunctionalization protocol

The biofunctionalized surfaces were prepared starting from 10 mm × 5 mm substrates cut from As-doped silicon wafers with a 300 nm SiO<sub>2</sub> layer on top. The substrates were preliminary cleaned by ultrasonic baths in acetone and 2-propanol for 10 min and dried with nitrogen. Afterwards, a 5 nm thick layer of Ti and a 50 nm Au layer were deposited by electron-beam evaporation. Before the biofunctionalization, each substrate was cleaned in a piranha solution (3 : 1 mixture of sulfuric acid and hydrogen peroxide), rinsed with HPLC water, and dried with nitrogen. The flowchart of the gold biofunctionalization protocol is sketched in Fig. S1 (ESI<sup>†</sup>). A suitable polymeric mask shadowed one half of the gold substrate. Subsequently, the sample was submerged in a PBS solution containing 0.1 mg mL<sup>-1</sup> of anti-IgM or anti-IgG for 150 minutes at room temperature. The substrates were then extensively rinsed sequentially in PBS and in HPLC water and dried by spinning at 3000 rpm for 60 s before any characterization. The surface coverage is

$1.2 \times 10^{12} \text{ cm}^{-2}$  antibodies as measured by surface plasmon resonance experiments.<sup>11</sup>

### Kelvin probe force microscopy

KPFM measurements were performed using an AFM system (mod. NTEGRA Spectra, NT-MDT, Moscow, Russia) equipped with a platinum-iridium coated tip FMG01/Pt (TipsNano) with an apex size of 35 nm, a resonance frequency  $f = 77.7 \text{ kHz}$ , and a quality factor  $Q = 242$ . The sample was grounded using a metallic clamp. Morphology, phase, and surface potential images have been acquired simultaneously using the two-pass mode in which each line is sequentially scanned 2 times. In the first pass, the morphology is recorded in semi-contact mode, while the AFM cantilever is mechanically excited at frequency  $f$ . In the second pass, the tip is retraced at a set lift height ( $h_{\text{lift}}$ ) from the sample surface to measure the surface potential. During the second pass the cantilever is no longer excited mechanically but electrically by applying  $V_{\text{tip}}$ . During this phase, the tip and the sample behave like the opposing plates of a capacitor. Because of their different work functions, a force between the plates exists, which can be modulated by applying a voltage  $V_{\text{tip}}$  to the tip.<sup>1</sup> In the vibrating capacitor model,  $V_{\text{tip}}$  consists of a constant ( $V_{\text{DC}}$ ) and an alternated ( $V_{\text{AC}}$ ) component with a frequency matching the cantilever resonant one. The force felt by the vibrating capacitor is:


$$F = -\frac{dC}{dz}(V_{\text{DC}} - \text{CPD}_{\text{tip-s}})V_{\text{AC}}\sin(2\pi ft) \quad (1)$$

where  $C$  is the capacitance,  $Z$  is the separation, and  $\text{CPD}_{\text{tip-s}}$  is the contact potential difference between the tip and the sample surface.  $V_{\text{DC}}$  is constantly adjusted by the AFM feedback electronics to nullify the force, and thus it corresponds to  $\text{CPD}_{\text{tip-s}}$ . The latter follows the same spatial variation of the sample surface potential and hence, in the rest of the paper, whenever we mention experimental SP values, we refer to the relative surface potential with respect to the tip. All images have been processed with the Image Analysis software. A scheme of the KPFM is reported in Fig. 1a.

Fig. 1 shows the morphology (b) of a representative  $90 \mu\text{m} \times 90 \mu\text{m}$  KPFM substrate, along with the corresponding surface potential image (c). Notably, the morphology image and the SP map exhibit a sharp nanometric boundary, demarcating the bio-functionalized region (leftmost side) and the gold region (rightmost side). The anti-IgM biofunctionalized area consists of a uniform coating with a significant different contrast with respect to the adjacent gold region, with a SP value  $150 \pm 11 \text{ mV}$  higher, as shown in Fig. 1d. The histograms of the SP distributions are given in Fig. 1e. We chose as an analytical parameter the surface potential difference (SPD) calculated between the peaks of the SP distributions corresponding to the anti-IgM and the Au areas, respectively. The immunological assays were performed by sequentially recording the SP images, at a fixed scanning location, after incubation for 10 min, in 0.5 mL of phosphate buffer saline (PBS) standard solutions of immunoglobulins (IgM and IgG) with concentrations ranging from  $10^2 \text{ zM}$  to  $10^6 \text{ zM}$ . After incubation in each of the PBS standard-







**Fig. 1** (a) Schematics of two pass-mode Kelvin probe force microscopy. Each line is sequentially scanned two times. In the first pass the morphology is recorded via semi-contact AFM mode, in the second pass the tip is re-traced at a set lift-height ( $h_{\text{lift}}$ ) from the sample surface to detect the surface potential. (b) Representative morphology ( $90\ \mu\text{m} \times 90\ \mu\text{m}$  area) of a patterned anti-IgM layer at the edge between the gold area and the protein region. A sharp edge between the bio-functionalized (leftmost side) area and gold (rightmost side) region is visible, highlighted by the dashed blue line. (c) Representative surface potential (SP) image. (d) Average surface potential profile. (e) Histogram of the surface potential distribution.

solutions, the substrates were washed thoroughly with HPLC water to remove the unreacted ligands, dried by spinning and measured by KPFM.

### Polarization modulation infrared reflection-absorption spectroscopy

Polarization modulation infrared reflection-absorption spectroscopy (PM-IRRAS) measurements were performed using a Nicolet iS50 Fourier transform spectrometer equipped with a research module containing a wire-grid polarizer, a photo-elastic polarization modulator (PEM) operating at 50 kHz, a variable angle sample holder and a liquid-nitrogen-cooled mercury-cadmium-telluride (MCT) detector. An air purging system kept the relative humidity below 17% and a low level of carbon dioxide in the system compartments.

The incident and reflected IR beams propagate at angles of  $82^\circ$  with respect to the sample surface normal. The PEM phase was set to have a maximum PM-IRRAS signal in the spectral regions ( $1000\text{--}1800\ \text{cm}^{-1}$ ) and ( $2600\text{--}3500\ \text{cm}^{-1}$ ). Each spectrum was the average of 1000 scans with a resolution of  $4\ \text{cm}^{-1}$ . Two PM-IRRAS reflective substrates were prepared depositing over a glass slide a 5 nm-thick Ti adhesion promoter layer and a 50 nm-thick Au layer, *via* electron beam evaporation. They were cleaned with a sequential ultrasonic bath in water, acetone, and 2-propanol for 10 minutes and dried under nitrogen flux. One substrate was further treated for 10 minutes in an ozone cleaner before submerging it into  $0.1\ \text{mg mL}^{-1}$  anti-IgM PBS (pH = 7.4, ionic strength = 162 mM) solution for 2 hours at  $25^\circ\text{C}$  to allow the formation of a stable layer of physisorbed antibodies. Then, it was thoroughly rinsed with HPLC water, and


dried using a spinner (1500 rpm, 3 minutes) before PM-IRRAS analysis. The PM-IRRAS spectra of the two bare substrates were first measured to check for the same starting chemical composition of their surface. A layer of anti-IgM was prepared on one substrate, as described above. Then, it was thoroughly rinsed with deionized water, dried by spinning and the PM-IRRAS spectrum recorded. Afterwards, both substrates were stored at  $4^\circ\text{C}$ , submerged in PBS solution for 24 h. The day after they were rinsed with water and dried, the PM-IRRAS spectra were measured, and they were stored again in buffer solution at  $4^\circ\text{C}$ . This protocol was repeated to measure the spectra of both the bare gold substrate and that of the protein layer after 48 h, 72 h, and 144 h. The baseline correction on the original PM-IRRAS signal was calculated in the ranges  $950\text{--}1780\ \text{cm}^{-1}$  and  $2700\text{--}3500\ \text{cm}^{-1}$  using a least-square third-order polynomial fitting method.<sup>16</sup>

## Results and discussion

### Experimental data and method optimization

The SP values extracted by KPFM are inherently affected by the surface conditions or experimental parameters. In particular, the strongest influence is due to the chosen tip-sample distance  $h_{\text{lift}}$  during the second pass, or the surface contamination by adventitious carbon in ambient air.<sup>14,15</sup> On the other hand, while the SP values may vary as a function of extrinsic parameters, the average value of the relative contact potential differences (SPD) between the bio-functionalized and the Au areas is less sensitive to the external conditions and can be






**Fig. 2** (a) Measured SP of Au (black square symbols) and anti-IgM areas (red circles) and SPD (b) as a function of the tip-sample distance in the 2nd pass of KPFM. The tip oscillation amplitude has been kept constant at 70 nm. The error bars have been calculated from the standard deviation of SP distributions. Increasing the tip-sample distance affects the SP values, however the SPD becomes approximately constant for  $h_{\text{lift}} > 200$  nm.

reliably used as a suitable analytical quantity to assess the surface transformation induced by affinity binding events.

To confirm this claim, we first consider the effect of varying  $h_{\text{lift}}$ . Controlling the tip-sample distance involves carefully calibrating the AFM setup, minimizing thermal drift, and accounting for any vibration-induced perturbations. Fig. 2 shows the peak values in the SP histograms and the corresponding surface potential differences of a representative anti-IgM/Au substrate, as a function of the tip-sample distance. While for small distances SP and SPD vary substantially, for  $h_{\text{lift}} > 200$  nm they reach a constant value. Accordingly, all KPFM measurements have been performed setting  $h_{\text{lift}} = 250$  nm.

Surface contamination by hydrocarbons and carbo-oxides<sup>17</sup> typically increases with storage time and becomes noticeable in a matter of minutes. With the aim of investigating the role of contaminants on the Au/anti-IgM substrates, we have plotted in Fig. 3 the measured SP and the corresponding SPD values as a function of time over several days. While the SP of both regions varies with time, probably due to the increased surface contamination by airborne adventitious carbon, it is worth noting that the SPD values remain approximately constant. Fig. 4b shows the time evolution of the PM-IRRAS spectra over a 6 day period. The largest modification occurs for the infrared spectrum of the bare gold substrate in which three bands centred at  $1540\text{ cm}^{-1}$ ,  $1660\text{ cm}^{-1}$  and  $1740\text{ cm}^{-1}$ , ascribed to the C–N, C–C and C=O stretching vibrations of carbonaceous molecules appear, respectively.<sup>18,19</sup> The intensities of these bands monotonically increase with time. Also, the Amide I and Amide II bands measured on anti-IgM coated samples change with time, though to a lesser extent. These findings demonstrate the progressive contamination of substrates due to the procedure of storage, drying and exposure to ambient air, as described in the method paragraph. On the other hand, the contamination effects can be effectively tackled by considering the relative surface potential shift SPD, so that we can use KPFM



**Fig. 3** Time variation of the surface potential measured on the Au (black squares) and the anti-IgM (red circles) areas of an Au/anti-IgM patterned substrate. The SPD values are plotted as blue triangles. The tip-sample distance has been set at 250 nm. The error bars have been calculated from the standard deviation of the distributions.

measurements as a reliable analytical platform, as demonstrated in the following.

### Method validation

Two antigen/antibody couples were used to validate our sensing approach, as well as to test its performance, along with levels of noise and limit of detections (LOD). The selected systems were the IgM/anti-IgM and IgG/anti-IgG couples. We have used KPFM to assess the occurrence of affinity binding interactions occurring at the biofunctionalized substrate after the incubation with increasing concentration of antigens in PBS standard solutions. The SP images were iteratively recorded at the same  $90\text{ }\mu\text{m} \times 90\text{ }\mu\text{m}$  area, across the interface between the Au region and the antibody one. Fig. 5 shows the SPD variation ( $\Delta\text{SPD}$ )






Fig. 4 (a) PM-IRRAS spectra of the anti-IgM layer physisorbed on the Au substrate (solid red line) and bare gold sample (dashed blue line). Background subtraction was accomplished by dividing the recorded PM-IRRAS signal with a third-order polynomial curve. It has been obtained using a modified least square minimization algorithm implemented in MATLAB. The spectrum of the anti-IgM sample shows characteristic features of a protein layer, including: (i) Amide I ( $1600\text{--}1710\text{ cm}^{-1}$ ), (ii) Amide II ( $1500\text{--}1570\text{ cm}^{-1}$ ), and (iii) Amide A ( $3200\text{--}3400\text{ cm}^{-1}$ ). (b) PM-IRRAS spectra of the anti-IgM layer (solid lines) and Au substrate (dashed lines) over a span of 6 days. The initial spectra of both pristine samples are shown as blue lines. Subsequent spectra were recorded after 24 hours (represented by solid and dashed violet lines), 48 hours (solid and dashed green lines), 72 hours (solid and dashed black lines), and 144 hours (solid and dashed red lines). Throughout the 6-day period, the samples were stored in a buffer solution of HEPES with a pH of 7.4 and an ionic strength of 150 mM, in a refrigerated environment, between successive measurements. As the measurement time progressed, new spectral features emerged on both the bare gold and protein spectra, centered at approximately  $1540\text{ cm}^{-1}$ ,  $1660\text{ cm}^{-1}$ , and  $1740\text{ cm}^{-1}$ . These spectral changes are attributed to the adsorption of adventitious layers on the sample surfaces.




Fig. 5 The SPD shift ( $\Delta\text{SPD}$ ) measured on the Au biofunctionalized substrate, as a function of the ligand concentration as compared to the SPD measured on pristine Au/biolyer substrates (reference). (a) Assay of IgM. Black squares: incubations with standard solutions of IgMs. Red circles: IgG incubations (negative control). The proteins are assayed from standard solutions in phosphate saline buffer solution (PBS, pH = 7.4, ionic strength: 162 mM), starting from the more diluted solution. The analytical parameter is the SPD shift ( $\Delta\text{SPD}$ ) at different analyte concentrations as compared to the SPD measured on the pristine sample exposed to a PBS solution (baseline). The values have been calculated from the peak values of the SP distributions. The binding and the negative control experiments have been performed on different anti-IgM/Au samples. The formation of the anti-IgM/IgM complex triggers an overall decrease by  $-74\text{ mV}$  of the investigated anti-IgM surface potential area. (b) Assay of IgG. Black squares: incubations with standard solutions of IgGs. Red circle: IgM  $1\text{ fM}$  in PBS (negative control). The affinity binding and the negative control experiments have been performed on the same anti-IgG/Au sample. Upon binding, an overall decrease by  $-90\text{ mV}$  of the investigated anti-IgG surface potential was found. The error bars are taken as one standard deviation.



when a pristine Au/biolyer substrate is exposed to increasing concentrations of specific antigens (recognition experiments) or un-binding ligands (negative controls). The IgM/anti-IgM findings are given in Fig. 5a, and the corresponding KPFM images and histograms of the SP distributions are shown in Fig. S2 (ESI<sup>†</sup>). Taking the average value ( $\Delta\text{SPD} = -15$  mV) and the standard deviation ( $\text{SD}_{\text{NC}} = 21.5$  mV) of the negative control data as a reference, we note that the recognition events cause an average  $\Delta\text{SPD}$  shift of  $-74 \pm 9$  mV, which is significantly larger than the shift at the limit of detection, taken as  $3 \text{ SD}_{\text{NC}}$ . Hence, with 99.7% confidence, we ascribe the observed  $\Delta\text{SPD}$  shift to the extended modification of the surface dipole distribution in the anti-IgM recognition layer, which is triggered by a relatively small number of affinity binding events. Notably, the surface potential change occurs in the entire scanned surface, which is encompassed by  $\sim 6 \times 10^7$  molecules, whereas the number of IgM ligands within the 500  $\mu\text{L}$  incubation volume varies in the range of  $3 \times 10^1$ – $3 \times 10^5$ . The nominal numbers of ligands hosted at each concentration are estimated as  $C \cdot V \cdot N_A$ , where  $C$  is the analyte concentration,  $V$  is the volume of the standard PBS solution in which the sample is incubated, and  $N_A$  is the Avogadro number. The error associated with the sampling procedure can be estimated according to the Poisson's distribution. The KPFM assay performance was tested also in the recognition process between human-IgG and its cognate antibody anti-IgG. To this end, a patterned anti-IgG/Au sample was used. The representative morphology and the corresponding average profile are reported in Fig. S4 (ESI<sup>†</sup>).

Fig. 5b shows the  $\Delta\text{SPD}$  induced by the exposure of a pristine Au/anti-IgG substrate to increasing concentrations of IgG (affinity binding) ligands. A negative control experiment was performed by incubating the substrate with 1 fM IgM in PBS. No significant SPD change was found with respect to the baseline, confirming the high selectivity of the sensing platform. The same probe was sequentially incubated with increasing concentrations of IgG. A mean surface potential shift of  $-94 \pm 10$  mV was found, due to the immunorecognition events. The corresponding KPFM images and surface potential distributions are shown in Fig. S6 (ESI<sup>†</sup>). A qualitative agreement with the data of Fig. 5a is evident, confirming that KPFM imaging is a reliable and ultrasensitive platform for biomarker detection on large-area biofunctionalized surfaces.

It is worth noting that increasing the ligand concentration does not produce a further increase of the  $\Delta\text{SPD}$  shift. This confirms the notion that the surface potential transition, locally triggered by a few affinity binding events, propagates over large areas, probably funnelled by the hydrogen bonding network.<sup>13</sup> Accordingly, our sensing platform has the tendency to operate as an OFF/ON detector, capable of assessing the presence of a single specific biomarker in a given assay volume. This is analogous to similar findings obtained by us using a transistor-like detecting platform.<sup>20–25</sup>

## Conclusion

We have developed and optimized an ultrasensitive detection platform based on KPFM and patterned surfaces

biofunctionalized with dense physisorbed antibody layers, suitable for assessing the presence of a few ligands in an OFF/ON mode. We estimate that a small number of affinity binding events ( $30 \pm 10$  molecules) cause extended surface potential shifts over areas larger than hundreds of squared micrometres, hosting  $>10^7$  highly packed capturing sites. This implies the existence of a propagation phenomenon of the electrostatic change occurring on a few capturing antibody sites. We postulated a domino-like surface phenomenon, ultimately affecting the densely packed neighbouring antibodies, and thus a single antibody–antigen binding occurrence could induce an extended electrostatic reorientation involving millions of antibody dipoles. While the detailed modelling of the observed amplification effect needs further studies, we have addressed here some reproducibility issues. PM-IRRAS analysis shows that adventitious contaminants actually affect the spectral features and chemical composition of both the biolayer and bare gold sample. Changes in the SP extracted by KPFM due to tip-sample distance, aging and storage effects cannot be neglected, nevertheless the simultaneous SP mapping of the bare and the biofunctionalized Au areas is a reliable tool to investigate the SP variation induced by antigen/antibody complex formation.

Our findings show that a change of  $\sim -70$  mV ( $\sim -90$  mV) is induced by the interaction of anti-IgM (anti-IgG) with the cognate IgM (IgG) ligands. The process is also proved to be irreversible, and thus no significant further surface potential change has been measured, if other affinity events occur. The presented data demonstrate a proof of concept and are relevant for future applications in the selective detection at the physical limit of immunoglobulin traces in real fluids.

## Author contributions

C. D. F. and G. S. conceived the paper and designed the experiments. C. D. F. and M. P. performed the KPFM and PM-IRRAS measurements and analysis. C. S. and M. C. contributed to sample preparation and characterization, E. M. and L. T. contributed to the interpretation of the results and the funding project supervision. C. D. F., M. P. and G. S. wrote the manuscript that was approved by all authors.

## Conflicts of interest

The authors declare no conflict of interest.

## Acknowledgements

The following projects are acknowledged for partial financial support: “A binary sensor with single-molecule digit to discriminate biofluids enclosing zero or at least one biomarker (NoOne)” ERC Starting Grant 2021 (GA 101040383); “Tecnologie portatili e protocolli innovativi per la diagnosi ultrasensibile di Xylella fastidiosa direttamente in piante e vettori” (1LIVEX-YLELLA) Ministero dell'agricoltura, della sovranità alimentare e delle foreste – MIPAAF D.M. n.419161 del 13/09/2022; Centro di





Innovazione Regionale Digital Assay, Regione PUGLIA DR n.702 del 08/11/2022; Italian network of excellence for advanced diagnosis (INNOVA), Ministero della Salute -code PNC-E3-2022-23683266 PNC-HLS-DA, CUP: C43C22001630001; Complementary National Plan PNC-I.1 "Research initiatives for innovative technologies and pathways in the health and welfare sector" D.D. 931 of 06/06/2022, DARE - DigitAl lifelong pRevEntion initiative, code PNC0000002, CUP: B53C22006420001, PNRR MUR project PE0000023-NQSTI; MUR – Dipartimenti di Eccellenza 2023–2027 – Quantum Sensing and Modelling for One-Health (QuaSiModO).

## References

- 1 M. Nonnenmacher, M. O'Boyle and H. K. Wickramasinghe, *Ultramicroscopy*, 1992, **42–44**, 268–273.
- 2 D. C. Hansen, K. M. Hansen, T. L. Ferrell and T. Thundat, *Langmuir*, 2003, **19**, 7514–7520.
- 3 G. Gauglitz, *Anal. Bioanal. Chem.*, 2020, **412**, 7043–7045.
- 4 M. Salerno and S. Dante, *Materials*, 2018, **11**, 951–956.
- 5 L. E. Cheran, M. Chacko, M. Zhang and M. Thompson, *Analyst*, 2004, **129**, 161–168.
- 6 L. E. Cheran, M. E. McGovern and M. Thompson, *Faraday Discuss.*, 2000, **116**, 23–34.
- 7 L.-E. Cheran, S. Johnstone, S. Sadeghi and M. Thompson, *Meas. Sci. Technol.*, 2007, **18**, 567–578.
- 8 J. Park, S. Lee, K. Jang and S. Na, *Biosens. Bioelectron.*, 2014, **60**, 299–304.
- 9 J. Park, J. Yang, G. Lee, C. Y. Lee, S. Na, S. W. Lee, S. Haam, Y.-M. Huh, D. S. Yoon, K. Eom and T. Kwon, *ACS Nano*, 2011, **5**, 6981–6990.
- 10 K. Jang, J. Choia, C. Park and S. Na, *Biosens. Bioelectron.*, 2017, **87**, 222–228.
- 11 C. Di Franco, E. Macchia, L. Sarcina, N. Ditaranto, A. Khaliq, L. Torsi and G. Scamarcio, *Adv. Mater. Interfaces*, 2022, **10**, 2201829.
- 12 E. Macchia, K. Manoli, B. Holzer, C. Di Franco, M. Ghittorelli, F. Torricelli, D. Alberga, G. F. Mangiatordi, G. Palazzo, G. Scamarcio and L. Torsi, *Nat. Commun.*, 2018, 3223.
- 13 E. Macchia, F. Torricelli, P. Bollella, L. Sarcina, A. Tricase, C. Di Franco, R. Österbacka, Z. M. Kovács-Vajna, G. Scamarcio and L. Torsi, *Chem. Rev.*, 2022, **122**, 4636–4699.
- 14 A. Liscio, V. Palermo, K. Müllen and P. Samorì, *J. Phys. Chem. C*, 2008, **112**, 17368–17377.
- 15 N. Turetta, F. Sedona, A. Liscio, M. Sambì and P. Samorì, *Adv. Mater. Interfaces*, 2021, **10**, 2100068.
- 16 C. A. Lieber and A. Mahadevan-Jansen, *Appl. Spectrosc.*, 2003, **57**, 1363–1367.
- 17 P. Swift, *Surf. Interface Anal.*, 1982, **4**, 47–51.
- 18 A. Barth, *Biochim. Biophys. Acta, Bioenerg.*, 2007, **1767**, 1073–1101.
- 19 V. A. Lorenz-Fonfria, *Chem. Rev.*, 2020, **120**, 3466–3576.
- 20 E. Macchia, Z. M. Kovács-vajna, D. Loconsole, L. Sarcina, M. Redolfi, M. Chironna, F. Torricelli, L. Torsi, D. Chimica and B. Aldo, *Sci. Adv.*, 2018, **8**, eabo0881.
- 21 L. Sarcina, F. Viola, F. Modena, R. A. Picca, P. Bollella, C. Di Franco, N. Cioffi, M. Caironi, R. Österbacka, I. Esposito, G. Scamarcio, L. Torsi, F. Torricelli and E. Macchia, *Anal. Bioanal. Chem.*, 2022, **414**, 5657–5669.
- 22 E. Macchia, L. De Caro, F. Torricelli, C. Di Franco, G. F. Mangiatordi, G. Scamarcio and L. Torsi, *Adv. Sci.*, 2022, **9**, 2104381.
- 23 L. Sarcina, E. Macchia, G. Loconsole, G. D'Attoma, P. Bollella, M. Catacchio, F. Leonetti, C. Di Franco, V. Elicio, G. Scamarcio, G. Palazzo, D. Boscia, P. Saldarelli and L. Torsi, *Adv. Sci.*, 2022, **9**, e2203900.
- 24 C. Scandurra, K. Björkström, L. Sarcina, A. Imbriano, C. Di Franco, R. Österbacka, P. Bollella, G. Scamarcio, L. Torsi and E. Macchia, *Adv. Mater. Technol.*, 2023, **8**, 2201910.
- 25 L. Sarcina, C. Scandurra, C. Di Franco, M. Caputo, M. Catacchio, P. Bollella, G. Scamarcio, E. Macchia and L. Torsi, *J. Mater. Chem. C Mater.*, 2023, **11**, 9093–9106.

