



Cite this: *Lab Chip*, 2021, 21, 2901

## High-throughput optofluidic screening for improved microbial cell factories *via* real-time micron-scale productivity monitoring†

Matthew Rienzo,<sup>‡a</sup> Ke-Chih Lin,<sup>‡b</sup> Kellen C. Mobilia,<sup>b</sup> Eric K. Sackmann,<sup>b</sup> Volker Kurz,<sup>b</sup> Adam H. Navidi,<sup>a</sup> Jarett King,<sup>a</sup> Robert M. Onorato,<sup>b</sup> Lawrence K. Chao,<sup>a</sup> Tony Wu,<sup>a</sup> Hanxiao Jiang,<sup>a</sup> Justin K. Valley,<sup>a</sup> Troy A. Lionberger<sup>‡\*b</sup> and Michael D. Leavell<sup>‡\*a</sup>

The industrial synthetic biology sector has made huge investments to achieve relevant miniaturized screening systems for scalable fermentation. Here we present the first example of a high-throughput ( $>10^3$  genotypes per week) perfusion-based screening system to improve small-molecule secretion from microbial strains. Using the Berkeley Lights Beacon® system, the productivity of each strain could be directly monitored in real time during continuous culture, yielding phenotypes that correlated strongly ( $r^2 > 0.8$ ,  $p < 0.0005$ ) with behavior in industrially relevant bioreactor processes. This method allows a much closer approximation of a typical fed-batch fermentation than conventional batch-like droplet or microplate culture models, in addition to rich time-dependent data on growth and productivity. We demonstrate these advantages by application to the improvement of high-productivity strains using whole-genome random mutagenesis, yielding mutants with substantially improved (by up to 85%) peak specific productivities in bioreactors. Each screen of  $\sim 5 \times 10^3$  mutants could be completed in under 8 days (including 5 days involving user intervention), saving  $\sim 50$ – $75\%$  of the time required for conventional microplate-based screening methods.

Received 4th May 2021,  
Accepted 9th June 2021

DOI: 10.1039/d1lc00389e

rsc.li/loc

## Introduction

Over the past several decades, metabolic engineering has developed microbial cell factories as sustainable alternatives to chemical synthesis from petroleum feedstocks or harvesting of animals and plants.<sup>1–9</sup> These fermentation processes employ natural and engineered enzymes to give elegant one-pot alternatives to conventional manufacturing. The gold-standard method for selection of a microbial strain for manufacturing is the lab-scale bioreactor, which represents a costly and labor-intensive commitment, as thousands or more candidate genotypes are often considered. Most high-throughput strain improvement campaigns have heavily relied on hit selection from microplate-based culture models,<sup>10,11</sup> growth-coupled selection schemes,<sup>12</sup> and more recently droplet microfluidics,<sup>13</sup> with varying success in the prediction of scalable bioreactor phenotypes.

A shortcoming of microplate- and droplet-based culture models is that they often do not replicate the conditions observed in feedback-regulated bioreactor systems.<sup>10</sup> Indeed, it is infeasible to individually and dynamically control the chemical composition of thousands of wells or droplets in a typical screening scheme, as is done in low-throughput bioreactors. The Beacon® optofluidic system developed by Berkeley Lights is unique in that it addresses this challenge by spatially isolating strains in incubation chambers (“NanoPen™ chambers” or “pens”) within a microfluidic device, the OptoSelect™ chip, while continuously controlling the extracellular environment by perfusion of fresh media.<sup>14</sup> This approach has the potential to more closely mimic a bioreactor by tightly controlling nutrient levels, oxygenation, and pH balance during growth. Previous reports of similar “microchemostat” systems have established a proof of concept for this style of microbial culture and *in situ* assay,<sup>15–19</sup> but such approaches have not demonstrated the throughput, automation, robustness, and operational feasibility that are possible with the Beacon system and/or are not intended to screen for improved metabolic flux.

The operation of the Beacon system is illustrated in Fig. 1. Briefly, the system can hold up to four chips, each of which includes a set of channels that traverse the full area of the

<sup>a</sup> Research and Development, Amyris, Inc., 5885 Hollis St., Suite 100, Emeryville, CA 94608, USA. E-mail: leavell@amyris.com

<sup>b</sup> Technology and Business Development, Berkeley Lights, Inc., 5858 Horton St., Unit 320, Emeryville, CA 94608, USA. E-mail: troy.lionberger@berkeleylights.com

† Electronic supplementary information (ESI) available. See DOI: 10.1039/d1lc00389e

‡ These authors contributed equally to this work.





**Fig. 1** The Beacon® optofluidic system. (a) Bright-field image of an entire OptoSelect™ chip with channels indicated in blue. (b) Magnified image of a portion of a channel following injection of suspended cells by syringe pump. Single cells (green) and multiples (red) are highlighted. (c) Each channel branches off into  $10^2$ – $10^3$  “NanoPen™ chambers” of volume  $\sim 0.25$ – $1.7$  nL, depending on chip design. (d) Fluorescence images can be collected periodically in addition to bright-field images, allowing identification of clones with desirable phenotypes. (e) Desired clones are exported into microplates for propagation and further study using Opto-ElectroPositioning (OEP) force, followed by media perfusion.

chip in a serpentine fashion. Cells are loaded into the chip by syringe pump from a sample vessel and allowed to collect throughout the chip within the channels.  $10\times$  bright-field imaging is used to visualize the cells, and a proprietary algorithm is used to select cells of appropriate roundness, size, and distance from other cells for seeding into pens. The device uses light-actuated cell positioning, termed OptoElectroPositioning (OEP™), which entails irradiating the silicon base of the chip with visible light while applying an electrical potential across the device. This induces a negative dielectrophoretic force that can be used to individually manipulate single cells by moving a square “light cage” surrounding each targeted single cell,<sup>20</sup> a process that has been automated at Berkeley Lights using image recognition software. The cells are moved into separate pens to be cultured as individual colonies. The chip is perfused with culture media as cells multiply. Cells within pens are not disturbed by media flow, but passive diffusion causes rapid exchange with fresh media such that nutrient levels are kept constant and waste is carried away rapidly. The phenotype of each colony can then be assessed by monitoring the growth and secretions using bright-field imaging and fluorescence-based assays. In this manner, the Beacon system can allow simultaneous culture of  $\sim 10^4$ – $10^5$  segregated colonies of differing genotypes. Once colonies bearing desirable phenotypes are identified, OEP can be used to reposition clumps of cells from a colony of interest into the channel using sequences of moving light bars, as shown in Fig. 1e. Subsequently, cells in the channels can be exported into microplates for further study through media perfusion. Charts illustrating some typical screening workflows are shown in Fig. S1 and S2.†

Here, we describe two novel methods for identifying improved small molecule-producing strains using the Beacon system. The methods rely on fluorescence-based detection of secreted product, from which per cell productivities can be inferred. Although similar assays have previously been used to identify mammalian cell lines for antibody production, this is the first demonstration that the Beacon system can be applied to screening microbial strains for efficient small-molecule production. To our knowledge, it is also the first example of high-throughput ( $>10^3$  strains per week) strain screening based on small-molecule secretion rate under steady-state culture conditions. Using strains that produce a fluorescent product as a test case, we show that Beacon system's assay scores correlate well with peak strain performance in 0.5- and 2.0 L pulse-fed bioreactors. We also demonstrate the ability of the platform to rapidly isolate strains from random whole-genome mutagenesis libraries that exhibit improved specific productivities in ambr250 bioreactors. The optimized two-tiered screening workflow reduced the timeline of library screening by 2-fold, obviated solid culture and colony-picking, and eliminated  $>95\%$  of material waste produced in conventional microplate screening.

## Results

### Bioreactor performance metrics and screening strategy

In metabolic engineering, specific productivity ( $q_p$ ; mol product per mol biomass per hour) and growth rate ( $\mu$ ; per hour) are two key variables intrinsic to the genotype of a strain under a set of fixed culture conditions. If a



scaled-down screening system effectively models strain physiology in a lab-scale bioreactor, estimates of  $q_p$  and  $\mu$  can be used to predict aspects of bioreactor performance. In this study, microfluidic culture conditions were empirically selected by optimizing the correlation between strain  $q_p$  on the Beacon system and in a gold-standard lab-scale fermentation platform. The assays used allow measurement of biomass and product secretion, which together enable inference of both  $q_p$  and  $\mu$ . Note that in this work, conditions were *not* selected to optimize correlation of  $\mu$  between the Beacon system's culture model and the lab-scale bioreactor.

The values of  $q_p$  and  $\mu$  for a strain are industrially relevant because they are determinants of yield ( $Y$ ; mol product per mol sugar fed) and volumetric productivity ( $P$ ; mol product per L per hour), which are two key factors in the cost of production during manufacturing. In this work, Beacon assays were evaluated by comparing assay values for strains to  $q_p$ ,  $Y$ , and  $P$  values observed from lab-scale bioreactor fermentations. These fermentations typically begin with a period of rapid growth, followed by a later stage of pseudo-steady-state production when nutrients are more restricted. The Beacon system's assays attempt to model the latter stage. Thus, the relevant performance metrics were taken from the interval 24–48 hours following bioreactor inoculation, after the end of the rapid growth phase.

### Bright-field intensity integration for biomass measurement

Most previously reported applications of the Beacon system have used mammalian cells, which are large enough to count explicitly *via* digital image processing. Because microbial cells are generally smaller and can grow in multilayers on the OptoSelect chip, counting is not feasible for hosts like *S. cerevisiae*. Instead, bright-field images were automatically processed to determine colony area and cell density. For the latter method, pixel intensity was integrated over the area of each NanoPen chamber, using reference measurements from nearby empty pens to calibrate the local background signal.

The resulting background-corrected assay value is hereafter termed “OD score”.

To validate OD score as a biomass measurement, NanoPen chambers in an OptoSelect chip were seeded with varying numbers of *S. cerevisiae* cells from a mixture of strains, which were then cultured on the Beacon system until a range of colony sizes was visible. After imaging the chip, cells were packed tightly into the NanoPen chambers *via* centrifugation to achieve uniform density before reimaging (Fig. 2). The area of the resulting packed colony was assumed to be a linear reflection of biomass. The OD score of the colonies measured before centrifugation was proportional to this packed colony area, indicating that it was linear over a wide range of colony sizes (Fig. 2), with typical CV < 15% for colonies > 3500  $\mu\text{m}^2$  in area. In contrast, unpacked colony area did not linearly predict packed colony area, likely because larger colonies accumulated biomass in multilayers rather than strictly by increasing in area (Fig. S3†).

### Sealed-pen productivity and yield inference

The Beacon system is configured to acquire fluorescence images of OptoSelect chips during strain cultivation. Although there are many ways to transduce a fluorescent signal in response to the local concentration of an analyte, detection is easiest when the product is intrinsically fluorescent.<sup>21,22</sup> A set of *S. cerevisiae* strains engineered to produce a fluorescent small molecule was thus chosen to demonstrate the feasibility of screening for a secretion phenotype.

Previous work on the Beacon system focused on secretion of macromolecules, which diffuse slowly enough to allow accumulation and visualization using fluorescent readouts.<sup>23</sup> For low-molecular-weight compounds (<1 kD) like that tested here, rapid diffusion from the NanoPen chamber was initially a concern. A method was therefore devised to seal the NanoPen chambers using a hydrophobic oil with high respiratory gas solubility, but low product and carbohydrate solubility. After sealing, each pen effectively becomes a



**Fig. 2** Bright-field density biomass measurement. (a) Representative image of a chip before and after centrifugation packing, illustrating the effect of packing on colony area and bright-field image density. (b) Top, relationship between unpacked colony OD score and packed colony area, illustrating assay linearity for the corrected biomass measurement. Bottom, relative variability (coefficient of variation, CV) of biomass sample measurement as a function of colony size, assuming perfect linearity of post-packing colony area with biomass.





miniature batch reactor with fixed carbon source, accumulating product, and continuous gas exchange. When the assay is complete, the oil can be replaced with aqueous media to allow export of desired strains.


To determine whether strain performance in the sealed-pen assay aligns with lab-scale bioreactor performance, six strains previously tested in 0.5 L bioreactors were selected. The strains were sequentially seeded into NanoPen chambers on OptoSelect chips on the Beacon system in high replication ( $n = 40\text{--}50$ ). Strains were cultured for  $\sim 18$  hours before sealing and then incubated for an additional 20 minutes with periodic fluorescence imaging (Fig. 3a), allowing inference of the relative  $P$  for each pen from the slope of the intensity curve for each colony during the 20 minute interval. To compare the per cell metabolic fluxes generated by each strain, relative specific productivities for the cells in each pen were calculated by normalizing the slope of the fluorescence intensity curve to the measured OD score, giving what is hereafter referred to as the “ $q_p$  score”. Note that the precise definition of  $q_p$  includes generation of both secreted and intracellular product; however, the strains tested in this work secrete the vast majority of the fluorescent product (Jiang *et al.*, unpublished results), and so rough relative inferences of  $q_p$  can be made from the biomass-normalized measurement of secreted product in this context.

The Beacon  $q_p$  scores clearly distinguished the three top performing strains from the three weaker producers (Fig. 3c). Mean  $q_p$  scores from the Beacon assay showed positive correlation ( $r^2 > 0.85$ ) with the  $q_p$  measured in bioreactor experiments in the 24–48 hour interval. This encouraging result indicated that the Beacon system could provide a relevant culture model for the selection of the most promising strains for bioreactor testing.

In addition to the 20 minute productivity measurement, the sealed-pen method also has the potential to assess the average yield of product with respect to feedstock. In early experiments, we demonstrated that 24 hour sealed-pen cultures of four strains gave endpoint fluorescence values that ranked them consistently with the relative titers attained in microplate models after carbon sources had been exhausted (Fig. S6†). Although the sealed-pen assays were not pursued as a primary strategy for rapid screening, this preliminary result suggests a complementary approach for assessing strain resilience to product and byproduct accumulation.


### Real-time productivity monitoring *via* steady-state product gradient assay

The rapid development of fluorescence signal in the sealed-pen assay prompted investigation of an alternative method, in which pens are left unsealed throughout the experiment. In this assay, local product concentration within each pen is mainly governed by the rate of production by the colony and diffusion away from the colony. As a result of the effective boundary condition just outside of each pen (product concentration  $\approx 0$ ), the concentration gradient in the cell-free gradient measurement area at steady state will be proportional to the rate of secretion of the analyte (Fig. S8†). This allows for inference of the productivity in each pen in the few seconds preceding a single fluorescence image of the chip by simply extracting the slope of the linear fit of signal intensity with respect to position in the pen (Fig. 4). If images are acquired periodically, one can easily collect time-dependent productivity data for every colony throughout the experiment.



**Fig. 3** Sealed-pen productivity assay. (a) Illustration of the sealed-pen assay, showing bright-field and time-lapse fluorescence images. (b) Examples of biomass-normalized time-course plots of NanoPen chamber fluorescence during the assay period for three strains (AMRS004, AMRS005, and AMRS006). Each faint line represents an individual pen, with the average value for each strain indicated with a dark line. (c) Correlation between average sealed-pen assay  $q_p$  scores and average lab-scale bioreactor  $q_p$  measured during the 24–48 hour interval. The Beacon system  $q_p$  score for each pen was calculated by normalizing the average rate of increase in fluorescence to the OD score (corresponding to the slopes of the faint lines in b). Bioreactor data plotted are the mean of several biological replicates per strain ( $n = 2\text{--}5$ ). Error bars in both directions show the 95% confidence interval of the mean. The least-squares linear regression line (blue) and its 95% confidence band (gray) are also shown ( $r^2 = 0.87$ ,  $p = 0.006$ ).





**Fig. 4** Open-pen gradient productivity assay. (a) Illustration of the diffusion gradient productivity assay, with the fluorescence image superimposed over the bright-field image. Images are periodically acquired in the bright field and fluorescence channels while the chips are continuously perfused with fresh culture media; the open pens allow product to diffuse out of the pens and into the channel, where it is rapidly carried away by the flowing media. Raw assay score is calculated by taking the average slope of the fluorescence intensity in the pen with respect to distance from the channel, after background correction. Refer to Materials and methods section for details on image processing. (b) Example data illustrating average time-dependent biomass measurements, raw assay scores (*i.e.* fluorescence slope), and  $q_p$  scores within each genotype. Data shown in color were included in calculation of average  $q_p$ . Data shown in gray were collected from colonies exceeding OD score of 0.25 and were excluded from average  $q_p$  calculations. NanoPen chambers above this threshold OD score frequently contained biomass within the measurement area of the pen, resulting in spurious spikes in the fluorescence gradient measured. In the  $q_p$  plots, average  $q_p$  across all pens within the accepted OD score range is plotted in black. (c) Correlation between open-pen Beacon assay  $q_p$  score at 18 hours of incubation and peak bioreactor  $q_p$ . Error bars in both directions show the 95% confidence interval of the mean. The least-squares linear regression line (blue) and its 95% confidence band (gray) are also shown ( $r^2 = 0.82$ ,  $p = 0.00012$ ). (d) Assay performance as a function of culture duration. Each line plot shows the time dependence of the  $R^2$  values for linear correlations like those shown in c, between mean bioreactor performance metrics ( $q_p$ , volumetric productivity, yield) from the 24–48 hour interval for each strain and mean Beacon  $q_p$  score. Points show the distribution of coefficients of variation for each of the strains tested over the course of the experiment.

Strain productivity was high enough to enable quantitation of productivity within hours of initiating incubation. To check the comparability of the sealed- and open-pen assays, the same set of six strains tested in the sealed-pen assay were subjected to the new format using identical culture media. The mean  $q_p$  scores for each strain from the two assays correlated extremely well ( $r^2 > 0.98$ , Fig. S9†) and the variability was comparable, with most strains showing CVs around 15–20%.

Next, an expanded set of 11 strains that had exhibited varying performance in 0.5 L bioreactors was selected to help identify optimal chip culture conditions for screening (Fig. S10, Table S1†). Using the real-time data made available by the gradient productivity assay, several media compositions and assay durations were evaluated to maximize correlation of  $q_p$  score to values of  $Y$ ,  $P$ , and  $q_p$  measured from bioreactor fermentations. In each experiment, an  $r^2$  value was computed

for the linear regression of the bioreactor metric with respect to the mean chip score at a given time. A plot was then constructed to show the evolution of the correlation strength over time to aid in the selection of assay duration (Fig. 4). Ultimately, the conditions that optimized prediction of bioreactor performance resulted in a strong positive correlation ( $r^2 = 0.82$ ) between mean  $q_p$  score and 24–48 hour interval bioreactor  $q_p$  for a set of 11 diverse engineered strains. As a comparison, endpoint titers in microplate culture model tests of the same strains yielded at best a correlation of  $r^2 = 0.73$ , also with respect to bioreactor  $q_p$  (Fig. S11†).

There are several advantages offered by the open-pen assay when compared with the sealed-pen assay: (1) the perfusion-based culture system likely provides a closer approximation of common fed-batch commercial processes, wherein nutrients are constantly provided and waste is



removed; (2) low concentrations of feed source and other nutrients, which might be quickly depleted in a sealed assay, can be used in order to avoid the problem of overflow metabolism often observed in batch processes; (3) the export of desired clones is simplified and accelerated; (4) the open-pen assay gives a more direct readout for rate of conversion of sugar into product, since secreted intermediates like ethanol may diffuse out of the pen before being reconsumed; (5) productivity data can be aggregated based on colony size at different time points, rather than requiring all data to be collected at a single time point with possibly wide-ranging OD scores.

### Productivity-based screening identified hits exhibiting increased peak $q_p$ in lab-scale bioreactors


Following the optimization described above, four strain libraries were generated by random whole-genome chemical mutagenesis of strains containing proprietary engineering for overproduction of the fluorescent target. As a point of reference, these parent strains had previously been shown to produce the target at titers reaching  $15\text{--}22\text{ g L}^{-1}$  in lab-scale bioreactors. Libraries were screened using a two-tiered scheme. In tier 1, four chips were dedicated to screening each clone at  $n = 1$ , with a total throughput of  $\sim 4000$  clones per library (Fig. 5 and S12 and S13†). After selection and export of 44 high-scoring clones (hits), strains were cultured to accumulate biomass in a 96-well plate. Hits were then submitted to tier 2 screening on two additional chips, in which each strain was cultured in higher replication (typically  $n \sim 50\text{--}100$ ) to identify strains for promotion to bioreactors

with higher confidence. Details on the throughput and data variability for each screen are available in Table S2†. These workflows can be operationalized highly efficiently, allowing at least two tier-1 or tier-2 screens to be scheduled on each Beacon system within each 5 day work week (Fig. S1 and S2†).

Each tier-2 screen identified at least one mutant with mean  $q_p$  score improved over the parent strain by  $>20\%$ . Seven mutants improved by varying extents were then tested alongside their parent strains in ambr250 bioreactors to assess performance at lab scale. Excitingly, four of the seven strains exhibited peak bioreactor  $q_p$  values improved over parent by 10–85% (Fig. 6 and S14†). These gains were aligned with or even larger than the magnitude of increases observed in the nanoliter culture model. The strain with the largest bioreactor  $q_p$  improvement (AMRS046\_12) additionally achieved a 20% increase in average  $P$  over a 6 day fermentation, making it a lead candidate for further engineering.

## Conclusions


The Beacon system represents an extremely promising opportunity for the high-throughput screening of microbial cell factories. Remarkably, the platform can identify strains with improved  $q_p$  at a volume  $10^9$ -fold smaller than the bioreactor it models. Strain productivity on the Beacon system correlated with peak bioreactor performance parameters as well as or better than data from conventional microplate culture models. The improvement in data quality may be partly attributed to the direct readout for  $q_p$ , as



**Fig. 5** Hit selection in tiered Beacon system screening. (a) Representative tier 1 assay data, showing normalized  $q_p$  scores from AMRS038 and its mutants on four OptoSelect chips. Strains promoted to tier 2 are highlighted in blue and were selected based on score magnitude and consistency over several time points. Error bars show one and two standard deviations above and below the mean for the parent strain on each chip. (b) Representative tier 2 assay data, showing  $q_p$  scores from biological replicate colonies ( $n = 20\text{--}100$ ) of strains promoted from tier 1. Error bars (dashed lines) represent the 95% confidence interval of the population mean. Strains highlighted in blue were statistically improved over AMRS038 by a two-tailed Dunnett's test of multiple comparisons ( $p < 0.05$ ).







**Fig. 6** Translation of improved flux phenotype into lab-scale bioreactors. Performance of mutants selected for analysis in lab-scale bioreactors from 2 of the 4 library screens. (a and c) Comparison of the  $q_p$  scores measured in the Beacon instrument's open-pen assay and the peak (interval 2)  $q_p$  measured from each ambr250 bioreactor run. Strains are color-coded, with black representing the parent strain in each screen. Hits in the screens shown were promoted based on performance on two separate OptoSelect chips within the screen. Beacon scores for mutants were normalized to the parent strain analyzed on the same chip, shown to their left in black. Bioreactor experiments were performed in duplicate, with each biological replicate shown explicitly and each point normalized to the mean of the peak  $q_p$  values measured for the parent strain. (b and d) Time-course data for ambr250 fermentations run with each parent strain and mutant. Each point indicates the performance metric (specific productivity  $q_p$ , volumetric productivity  $P$ , or yield with respect to sugar  $Y$ ) calculated over a 24 hour interval. The bioreactor measurements are normalized to the highest value shown in each graph.

opposed to endpoint yield measurements from carbon-exhausted cultures or error-prone plate-based productivity measurements. It is likely also a consequence of maintaining a roughly constant chemical environment, in which strains are held in a stable metabolic state throughout the experiment.

The two-tiered libraries above were processed without microplates, agar trays, liquid handling robotics, or extraction solvents. Complete results for both tiers were generated in under ten days – approximately half of the time typically used for a microplate library screen of comparable size with sophisticated automation. The importance of reducing this cycle time cannot be overstated, as each round of strain design in an engineering campaign depends on results from the previous round. Together, the rapid data turnaround and high data quality showcased here could reduce time-to-market for a metabolic engineering product by months or more. We note that some alternative microfluidic approaches, such as culture within droplets in an emulsion, can encapsulate and sort based on fluorescence readouts at much higher throughputs ( $>10^4$  droplets per second). However, the flexible feed style and productivity monitoring described here are important advancements that are not yet attainable through droplet-based screening.

As stated above, strain improvement efforts are motivated by the need to reduce manufacturing costs, which are influenced by several factors. Here we have focused on screening for strains with elevated peak  $q_p$ , but the same assays could be used to improve other phenotypes as well. For example, screens could be designed to find strains more likely to sustain peak  $q_p$  throughout a longer fermentation by

- (1) choosing strains with less selective pressure for mutation,
- (2) identifying strains with resistance to feedback inhibition by products or intermediates, or
- (3) uncoupling growth rate from metabolic pathway flux.<sup>24</sup>

As can be seen in Fig. 6, some strains with highly improved bioreactor  $q_p$  values exhibit relatively smaller improvements in  $P$  and  $Y$ , which must be increased to reduce manufacturing cost. The Beacon system's assays offer the exciting opportunity to monitor  $q_p$  and  $\mu$  simultaneously, which can in theory allow one to screen for strains with improved  $P$  and  $Y$  as well. These variables can be written as:

$$P = q_p c_x \quad Y = \frac{q_p}{q_s} = \frac{q_p}{a\mu + bq_p + cq_{CO_2} + \dots}$$

where  $c_x$  is the biomass density,  $q_s$  is the specific rate of sugar consumption,  $q_{CO_2}$  is the specific rate of  $CO_2$  release,  $a$ ,  $b$ , and  $c$  are stoichiometric coefficients, and “...” indicates any other byproducts formed. Because  $c_x$  is a direct consequence of  $\mu$  and dilution rate, these relationships show that in addition to  $q_p$ ,  $\mu$  is also an important determinant of both  $P$  and  $Y$ . However, the scope of this work was limited to optimizing the correlation of Beacon  $q_p$  score to lab-scale bioreactor performance metrics;  $\mu$  observed in the final Beacon system's culture model did not correlate well with  $\mu$  measured in lab-scale bioreactors. One probable reason for this gap is the high carbon content of the culture medium, which was used due to the advantage of rapid growth for a short assay duration. Further refinement of the Beacon system and assay will likely lead to improved predictions of bioreactor performance in many aspects.



This journal is © The Royal Society of Chemistry 2021



For spectroscopic analysis of titers of the fluorescent product made by the engineered strains, production plates were shaken at 1000 rpm for 30 s before addition of DMSO (900  $\mu\text{L well}^{-1}$ ) and then again shaken at 1000 rpm for 30 s. Plates were centrifuged at  $500 \times g$  for 30 seconds at ambient temperature, and then 10  $\mu\text{L well}^{-1}$  of the supernatant was transferred for dilution into 90  $\mu\text{L well}^{-1}$  of DMSO in a 300  $\mu\text{L}$  flat-bottom black fluorescence assay plate (Corning). Assay plates were shaken at 1500 rpm for 30 s and then analyzed for full-well fluorescence using an M5 plate reader (Molecular Devices).

### Random whole-genome mutagenesis

Single colonies were transferred using a sterile wooden rod into rich liquid growth media (3 mL) in a 14 mL polypropylene culture tube with a loose-fitting cap (Corning). Cultures were shaken at an angle at 28 °C for 18 h until reaching stationary phase. A 1 mL aliquot of the starter culture was then diluted into a 250 mL baffled PETG flask (ThermoFisher) containing rich growth media (40 mL) followed by an additional 24 h incubation at 28 °C with agitation.

After outgrowth, cells were pelleted by centrifugation ( $3000 \times g$ , 3.5 min, room temperature) and washed once in 45 mL phosphate-buffered saline (PBS). Cell density was then determined by counting and a 1 mL aliquot was prepared at  $6 \times 10^7$  cells  $\text{mL}^{-1}$  in PBS. To this suspension was added the chemical mutagen *N*-ethyl-*N*-nitrosourea to a final concentration of 20 mM. The reaction was incubated for an exposure time of 30 min with agitation before neutralization of the mutagen by addition of 15% sodium thiosulfate (500  $\mu\text{L}$ ). Cells were then pelleted again by centrifugation and washed twice with 1 mL PBS before resuspending in 2 mL rich growth medium. The cells were recovered overnight ( $\sim 18$  h) at 28 °C with agitation in a 14 mL polypropylene culture tube with a loose-fitting cap, then stored at 4 °C and submitted for Beacon system screening within 48 h. Strains with names notated in the format “AMRS###\_##” were generated using this procedure and identified through Beacon system screening, where the first 3 digits of the number indicate the mutagenized parent strain and the last two digits are used to index the hits. Strains AMRS042-AMRS050 were also generated through random mutagenesis but were selected through conventional microplate screening methods.

### Beacon system experiments

**System preparation.** Prior to initiating the workflow, the Beacon system was sterilized by purging all fluidic lines with SporGon decontamination reagent (Decon Labs, Inc.). After soaking for 2 hours, lines were rinsed with sterile water for 1 hour to eliminate residual SporGon.

Following sterilization, 4 OptoSelect chips were loaded onto the Beacon system for a sequence of pre-workflow operations. The first step was chip wetting and priming. The

Wetting Solution was introduced to the chip to eliminate air pockets and was then incubated for surface functionalization. DI water was then flushed through each chip to remove the wetting solution.

To enable the sealed-pen productivity assay (see section below), a specialized differential wetting procedure was implemented in place of the typical wetting procedure. In this case, a functionalization that renders surfaces hydrophobic was implemented to wet the surface along the channels exclusively, while the standard wetting solution was used to wet the surface in the NanoPen chambers. Differential wetting allowed for different surface properties between the channels and the NanoPen chambers, enhancing the sealing performance in the sealed-pen productivity assay.

After wetting and priming, the Beacon system automatically located the fiducial markers on the chips for *x-y* stage and focus calibration. Lastly, two types of reference imaging were performed sequentially. The bright-field high-magnification (10 $\times$  objective) images were taken across the chips as a reference for the quantification of colony size during cell culture. Subsequently, the fluorescence images (4 $\times$  objective, FITC channel) were captured before and after equilibrating 50  $\text{mg L}^{-1}$  of the final product throughout all OptoSelect chips. The fluorescence images taken before and after equilibration were used as the background reference and the normalization reference, respectively, for the quantification of product concentration during the productivity assay.

**Single-cell loading.** Broth from microplate precultures or mutagenesis recovery was diluted in PBS to a target OD600 of 0.1 and transferred to a 300  $\mu\text{L}$  96 well plate for import (Corning). The import plate was then loaded into the well-plate incubator on the Beacon system for single cell loading. During the loading process, the temperatures of the well-plate incubator and the OptoSelect chips were set at 4 °C and 18 °C, respectively, to reduce cell proliferation during the loading period.

For each strain to be loaded into the NanoPen chambers, 25  $\mu\text{L}$  of cell suspension was imported from the well-plate incubator to the channels of the OptoSelect chips. Single cells were identified automatically by the Beacon instrument control software using a convolutional neural network algorithm optimized for yeast cells. A positioning strategy was then automatically implemented to maximize loading throughput using the OEP technology. Residual cells in the channels were flushed to waste after loading. In general, it took roughly 4–5 hours to load the cells across all 4 chips in a single library screen workflow.

**Culture and biomass measurement.** After cell loading, the cells were cultured on-chip at 33.5 °C with constant perfusion of fresh media (BSM + carbon source) at 0.1  $\mu\text{L s}^{-1}$ . Time-lapse bright-field images were taken every hour to keep track of cell growth for each colony. Depending on the experiment, one of two types of productivity measurements was performed within the culture period: the sealed-pen assay or the open-pen assay.



**Sealed-pen productivity assay.** After incubation on the chip for the desired time, fluorinated oil (HFE-7500) was imported into the main channel to seal each pen and thereby mitigate diffusion of the secreted product into the main channel. The increase in fluorescence of the secreted product was tracked over a short period (~20 minutes) to determine the productivity of each colony. After the assay, the oil was flushed out of the channels and a surfactant (0.25% Capstone-30 in PBS) was used to clear the residual oil along the fluidic path.

**Open-pen real-time productivity assay.** The open-pen assay measured the steady-state position-dependent fluorescence gradient in each pen during media perfusion. In this gradient, the cells were the source and the channels the sink for the fluorescent product, maintaining a gradient proportional to the secretion rate of the cells. During the open-pen assay, the media perfusion rate was raised to 0.3  $\mu\text{L s}^{-1}$  for 10 minutes to establish steady chemical gradients from the pens to the channels before taking the fluorescence images. The only requirements for this assay were to set a sufficiently high flow rate to maintain a uniform sink in the channel and sufficient time to reach a steady state. The mechanics of this assay were such that measurements could be captured at a 12 minute interval on 1 chip without significant interruption to culture. For routine experiments, the open-pen assay was performed at 1 hour intervals to measure the productivity of each colony across 4 chips running in parallel.

**Chip image analysis.** The bright-field biomass measurements and the fluorescence productivity assays were analyzed using the Cell Analysis Suite (CAS<sup>TM</sup>) 2.1 Software from Berkeley Lights. The biomass (OD), total productivity, and relative specific productivity (“ $q_p$  score”) were quantified per pen at each time point.

The OD score of each colony  $i$  at a given time  $t$  ( $\text{OD}_{i,t}$ ) was quantified by comparing the intensity of the bright-field images within each pen ( $I_{i,t}$ ) to that of the reference images taken during the system preparation process ( $I_{i,\text{Ref}}$ ), with the same illumination conditions. The percentage difference in brightness was proportional to the amount of biomass in the pen. Brightness values were normalized to the average pixel brightness of all empty pens in same image to correct bright field intensity fluctuation from image to image. The OD score was calculated from the formula:

$$\text{OD}_{i,t} = 1 - \frac{\bar{I}_{i,t}}{\bar{I}_{i,\text{Ref}}}$$

where  $\text{OD}_{i,t}$  is the score for pen  $i$  at time  $t$ ;  $\bar{I}_{i,t}$  is the normalized brightness of pen  $i$  at time  $t$ ; and  $\bar{I}_{i,\text{Ref}}$  is the normalized brightness of pen  $i$  in a reference image.  $\bar{I}_{i,t}$  and  $\bar{I}_{i,\text{Ref}}$  are defined as:

$$\bar{I}_{i,t} = \frac{I_{i,t}}{\frac{1}{|E_v|} \sum_{x \in E_v} I_{x,t}}$$

$$\bar{I}_{i,\text{Ref}} = \frac{I_{i,\text{Ref}}}{\frac{1}{|E_v|} \sum_{x \in E_v} I_{x,\text{Ref}}}$$

where  $E_v$  is the set of empty pens in image  $v$ , and the denominators represent the average raw brightness of empty pens.

The product concentration at pixel  $i$  at time  $t$  ( $C_{i,t}$ ) was calculated from the following formula:

$$C_{i,t} = \frac{F_{i,t} - F_{i,\text{Bg}}}{F_{i,\text{Ref}} - F_{i,\text{Bg}}}$$

where  $F_{i,t}$  is the fluorescence intensity at pixel  $i$  at a given time  $t$ ;  $F_{i,\text{Bg}}$  is the fluorescence intensity of the background reference at pixel  $i$ ; and  $F_{i,\text{Ref}}$  is the fluorescence intensity of the 50  $\text{mg L}^{-1}$  reference at pixel  $i$ . In the sealed-pen assay, the total relative productivity was quantified by measuring the increase of product concentration in each pen divided by the assay duration. In the open-pen assay, the total relative productivity was quantified by measuring the gradient of product concentration 20  $\mu\text{m}$  away from the opening of each pen. The  $q_p$  score of each colony was then calculated as the total relative productivity divided by OD score.

In a library screen workflow, the top producers were selected based on assay scores with customizable criteria. For instance, in the library screens for the mutagenesis of AMRS038 and AMRS043, the exported colonies maintained the highest  $q_p$  scores over 3 consecutive assay time points, equivalent to a 3 hour duration.

**Colony export.** The colony export process included a series of operations prior to OEP unloading – colony unpacking, pruning, and extended PBS flush, in order to reduce cross-pen contamination from the overgrowing colonies. Throughout the export process, chip temperature was set at 18  $^{\circ}\text{C}$  to reduce metabolic activity of the cells. We first perfused 500  $\mu\text{L}$  of DI water across the chips for 10 min. The osmotic pressure led to the swelling of cells, unpacking the tightly aggregated colonies. Immediately afterwards, an OEP sequence was performed to prune away excessive cells from the NanoPen chambers. The Beacon system software utilized a convolutional neural network-based cell detection algorithm to keep track of pens with potential clonality risk. PBS was perfused through the chips continuously at 0.5  $\mu\text{L s}^{-1}$  during the pruning process to flush out excessive cells into the waste bottle. After the pruning process, an extended PBS flushing operation was performed at 1  $\mu\text{L s}^{-1}$  for an hour to minimize potential sources of contamination from cells trapped along the fluidic path.

After the series of clonality de-risking operations, the OEP unload sequence was performed to export the chosen best clones identified in the assay. Each colony of interest was unpenned by OEP into the channels and delivered into the 96 well export plate (Corning), where 150  $\mu\text{L}$  of rich growth media was loaded in each well. Blank exports and colony



exports were performed alternately to assess clonality control. The 96 well export plate was kept at 4 °C in the well-plate incubator during the export process and switched to room temperature at the end of the workflow. The entire colony export process took about 15 hours for 4 chips with 48 colonies retrieved in total, but can be easily optimized by a wide margin for higher efficiency. The export colonies were then transferred to a 1.6 mL round-well 96 well plate (Axygen) for outgrowth.

**Statistical analysis.** Microbial cell metabolic activity is known to vary among cells with a common genotype and fluctuate stochastically in time within single cells.<sup>31,32</sup> For colonies containing small numbers of cells, such heterogeneity can contribute meaningfully to variation in measurements made on samples of a single strain. Screening efforts in this work seek to identify strains with improved *ensemble* performance in a bioreactor; therefore, the summary statistic of interest for distinguishing strain performance in Tier-2 screens was the 95% confidence interval of the mean  $q_p$  score, which is represented by error bars in plots of Tier-2 data. Additional information on sources of variability is discussed in the Supporting Information (Fig. S3–S5†).

### Lab-scale bioreactor experiments

**Seed build for fermentation experiments.** Colonies were plated onto solid media from frozen yeast glycerol stocks and incubated at 28 °C for 72 hours. Three 10 µL loops of colonies were then picked into a 50 mL baffled conical tube containing 8 mL of yeast seed media. This preculture was grown at 28 °C for 8 hours to a target optical density (OD) of 1–2, and then 5 mL was sub-cultured into seed flask media (120 mL) in a 500 mL baffled flask. Flasks were grown at 28 °C for 16 hours to a target OD of 4–6. Seed culture was used to inoculate bioreactors at 20% v/v with yeast bioreactor media. Bioreactors were inoculated to 50% of their total vessel volume.

**Bioreactor control.** Lab-scale fermentations were carried out at the 0.5- and 2.0 L scales using BioStat Q Plus (Sartorius) and at the 0.25 L scale using the ambr250 high-throughput fermentation system (Sartorius).

For the 0.5- and 2.0 L fermentations, the glass bioreactors were configured with dual Rushton impellers and autoclaved before use. Air flow was kept constant at 0.5 L min<sup>−1</sup> during each run. Temperature began at 28 °C for the first 12 hours and increased at 0.5 °C hour<sup>−1</sup> to a setpoint of 30 °C. Dissolved oxygen was measured by submerged probes and maintained at 30% through agitation control for the first 8 feed pulses. The target pH was set at 5.0 with one-sided control using ammonium hydroxide (20% v/v). Load cells were in place to record weights of vessels, feed, and base in real time. L81 PEG-PPG-PEG block copolymer was used to control foaming.

The ambr250 system uses filterless microbial disposable bioreactors with dual impellers. Temperature was kept at 28

°C for first 12 hours and increased at 0.5 °C hour<sup>−1</sup> to a setpoint of 30 °C. Dissolved oxygen is measured with an optode and drives the same control scheme as described for the 0.5 L systems. Off-gas was measured using a gas analyzer custom-built for the ambr250 system (BlueSens 001-0GSR52). The target pH was set at 5.0 with one-sided control using ammonium hydroxide (20% v/v). The ambr250 system was also configured with an integrated ViCell XR cell counter (Beckmann) for biomass quantification. L81 PEG-PPG-PEG block copolymer was used to control foaming. Feed consumption was recorded by tracking added volume.

The 0.5 L, 2.0 L, and 0.25 L fermentations all used an identical fed-batch process with daily drawdowns (removal of culture broth to avoid overflows). To begin the initial growth phase, the fed-batch process delivered a fixed bolus of 10 g L<sup>−1</sup> of syrup at a starting rate of 6 g L<sup>−1</sup> hour<sup>−1</sup>. This early phase was marked by rapid biomass production and ethanol accumulation. Following sugar depletion and re-consumption of ethanol produced, a rapid rise in dissolved oxygen triggered the system to deliver another 10 g L<sup>−1</sup> bolus of syrup. The feed rate automatically increased by 2.5% if the total time taken to consume the previous bolus was less than 30 minutes. This feed strategy was repeated for 8 boluses in the growth phase. During the growth phase, an agitation ramp was enabled to keep the dissolved oxygen above 30% saturation. Once 8 boluses were fed, the bolus size was increased to 15 g L<sup>−1</sup> and two-sided feed control was enabled. Consumption times less than 10 minutes triggered a 2.5% increase; greater than 70 minutes, a decrease. Samples were collected every 24 hours for measurement of titer and biomass. Broth was removed at collection times until 50% of the vessel volume remained, and excess broth was discarded. Following broth removal, a volume of PSA was added equal to the volume of broth removed × 0.054.

**Yield, productivity, and  $q_p$  calculations.** Biomass in 0.5 L and 2.0 L bioreactors was measured at 24 hour intervals using a calibrated capillary packed-cell volume assay. Biomass in ambr250 bioreactors was assessed using a Vi-CELL cell counter (Beckmann Coulter). Product titers were measured in all bioreactors at 24 hour intervals by adding 125 µL whole broth to 18 mL of diluent (98.9% DMSO/1% water/0.1% formic acid with 100 mg kg<sup>−1</sup> fluorescein internal standard), homogenizing, and then analyzing *via* UPLC-UV.

Yield ( $Y$ ), volumetric productivity ( $P$ ), and specific productivity ( $q_p$ ) in any interval  $i$  were calculated respectively from the equations:

$$Y_i = 100\% \times \frac{M_{p,i}}{M_{s,i}} \quad P_i = \frac{M_{p,i}}{\bar{V}_i t_i} \quad q_{p,i} = \frac{M_{p,i}}{\bar{X}_i t_i}$$

where  $M_{p,i}$  and  $M_{s,i}$  are the respective masses of product generated and sugar consumed during interval  $i$ ,  $\bar{V}_i$  and  $\bar{X}_i$  are the time averages of, respectively, the liquid volume and the dry biomass content of the bioreactor over interval  $i$ , and  $t_i$  is the duration of interval  $i$ . For each bioreactor run, the





values of  $Y$ ,  $P$ , and  $q_p$  calculated from the 24–48 hour interval were used for correlation analyses.

## Conclusions

The conclusions section should come in this section at the end of the article, before the acknowledgements.

## Author contributions

MDL, TAL, AHN, and JKV conceived the study. TW and HJ designed and constructed the parent strains. RMO and VK wrote the image-processing software. EKS, VK, MR, KCL, KCM, and LKC designed the experiments. KCL, KCM, VK, MR, JK, and LKC performed the experiments. MR, KCL, KCM, VK, JK, and LKC performed the data analysis. MR, KCL, LKC, and JK wrote the manuscript.

## Conflicts of interest

Authors are employees of Amyris, Inc., or Berkeley Lights, Inc.

## Acknowledgements

We thank Sunil Chandran, Kirsten Benjamin, Dana Nadler, and Stefan Moser for helpful discussion.

## References

- 1 J. Nielsen and J. D. Keasling, *Cell*, 2016, **164**, 1185–1197.
- 2 B. M. Woolston, S. Edgar and G. Stephanopoulos, *Annu. Rev. Chem. Biomol. Eng.*, 2013, **4**, 259–288.
- 3 P. Hill, *et al.*, *J. Ind. Microbiol. Biotechnol.*, 2020, **47**, 965–975.
- 4 A. Burgard, M. J. Burk, R. Osterhout, S. Van Dien and H. Yim, *Curr. Opin. Biotechnol.*, 2016, **42**, 118–125.
- 5 D. Julleson, F. David, B. Pfleger and J. Nielsen, *Biotechnol. Adv.*, 2015, **33**, 1395–1402.
- 6 V. Chubukov, A. Mukhopadhyay, C. J. Petzold, J. D. Keasling and H. G. Martín, *npj Syst. Biol. Appl.*, 2016, **2**, 1–11.
- 7 Y.-J. Wee, J.-N. Kim and H.-W. Ryu, *Food Technol. Biotechnol.*, 2006, **44**, 163–172.
- 8 National Science Foundation, Catalyzing Commercialization: Fermentation Process Makes Malonates from Sugar, Not Cyanide, *Chem. Eng. Prog.*, 2016, 16.
- 9 B. Dumas, C. Brocard-Masson, K. Assemet-Lebrun and T. Achstetter, *Biotechnol. J.*, 2006, **1**, 299–307.
- 10 M. Rienzo, *et al.*, *Metab. Eng.*, 2020, **63**, 102–125.
- 11 M. Wehrs, *et al.*, *J. Ind. Microbiol. Biotechnol.*, 2020, **47**, 913–927.
- 12 J. A. Dietrich, A. E. McKee and J. D. Keasling, *Annu. Rev. Biochem.*, 2010, **79**, 563–590.
- 13 E. K. Bowman and H. S. Alper, *Trends Biotechnol.*, 2020, **38**, 701–714.
- 14 A. Mocciaro, *et al.*, *Commun. Biol.*, 2018, **1**, 41.
- 15 N. Dénervaud, *et al.*, *Proc. Natl. Acad. Sci.*, 2013, **110**, 15842–15847.
- 16 J. Park, J. Wu, M. Polymenis and A. Han, *Lab Chip*, 2013, **13**, 4217–4224.
- 17 M. Kim, J. Bae and T. Kim, *Anal. Chem.*, 2017, **89**, 9676–9684.
- 18 F. K. Balagaddé, L. You, C. L. Hansen, F. H. Arnold and S. R. Quake, *Science*, 2005, **309**, 137–140.
- 19 A. E. Vasdekis and G. Stephanopoulos, *Metab. Eng.*, 2015, **27**, 115–135.
- 20 P. Y. Chiou, A. T. Ohta and M. C. Wu, *Nature*, 2005, **436**, 370–372.
- 21 G. H. An, J. Bielich, R. Auerbach and E. A. Johnson, *Bio/Technology*, 1991, **9**, 70–73.
- 22 J. M. Wagner, *et al.*, *Metab. Eng.*, 2018, **47**, 346–356.
- 23 K. Le, *et al.*, *Biotechnol. Prog.*, 2018, **34**, 1438–1446.
- 24 P. Rugbjerg and L. Olsson, *J. Ind. Microbiol. Biotechnol.*, 2020, **47**, 993–1004.
- 25 R. Metzenberg, *Fungal Genet. Rep.*, 2004, **51**, 19–20.
- 26 J. F. Nijkamp, *et al.*, *Microb. Cell Fact.*, 2012, **11**, 36.
- 27 S. de Kok, *et al.*, *ACS Synth. Biol.*, 2014, **3**, 97–106.
- 28 Z. Serber, R. Lowe, J. A. Ubersax and S. S. Chandran, Compositions and methods for the assembly of polynucleotides, *US Pat. No. 8221982 B2*, US Pat. and Trademark Office, 2012.
- 29 C. Guthrie and G. R. Fink, *Methods in enzymology: guide to yeast genetics and molecular biology*, Academic Press, 1991.
- 30 M. D. Rose, F. Winston and P. Hieter, *Open Biol.*, 1990, **6**, 14.
- 31 J. Ansel, *et al.*, *PLoS Genet.*, 2008, **4**, e1000049.
- 32 M. Ackermann, *Nat. Rev. Microbiol.*, 2015, **13**, 497–508.

