



Cite this: *Biomater. Sci.*, 2024, **12**, 252

## Research progress and perspective of metallic implant biomaterials for craniomaxillofacial surgeries

Huafang Li, \*<sup>a</sup> Jiaqi Hao<sup>a</sup> and Xiwei Liu<sup>b</sup>

Craniomaxillofacial bone serves a variety of functions. However, the increasing number of cases of craniomaxillofacial bone injury and the use of selective rare implants make the treatment difficult, and the cure rate is low. If such a bone injury is not properly treated, it can lead to a slew of complications that can seriously disrupt a patient's daily life. For example, premature closure of cranial sutures or skull fractures can lead to increased intracranial pressure, which can lead to headaches, vomiting, and even brain hernia. At present, implant placement is one of the most common approaches to repair craniomaxillofacial bone injury or abnormal closure, especially with biomedical metallic implants. This review analyzes the research progress in the design and development of degradable and non-degradable metallic implants in craniomaxillofacial surgery. The mechanical properties, corrosion behaviours, as well as *in vitro* and *in vivo* performances of these materials are summarized. The challenges and future research directions of metallic biomaterials used in craniomaxillofacial surgery are also identified.

Received 31st August 2022,  
Accepted 4th October 2023

DOI: 10.1039/d2bm01414a

rsc.li/biomaterials-science

### 1. Introduction

For human beings, in addition to maintaining the facial contour and shape, the maxillofacial bone is also closely related to the functions of chewing, pronunciation and speech. Injury or diseases (such as tumors, functional atrophy, congenital diseases, periodontitis, iatrogenic injuries, *etc.*, which can lead to the loss of bone tissues in the oral cavity) will affect the normal life and mental state of patients.<sup>1</sup> Cranial problems are also very complex, such as craniosynostosis, traumatic skull defects and deformity. These diseases commonly need the intervention of implant materials to cure, so it promotes the development of implant materials to a certain extent. Currently, bone substitutions including autografts, allografts, xenografts, and artificial materials are used to reconstruct cranio-maxillofacial regions. Autologous bone grafts are considered the gold standard of implants because they contain growth factors for osteoinduction, cells for osteogenesis and the framework for osteoconduction,<sup>2</sup> which are thought to favor bone formation. The disadvantages of autologous bone grafts, allografts and xenografts, however, are limited sources, high surgical risks, and high psychological stress.<sup>3</sup>

Aiming at natural bone grafted defects, researchers have gradually developed various artificial materials. Since hard tissues are responsible for the mechanical stability of the body, materials required to repair, replace and/or restore hard tissues must possess strength, an anti-corrosion/degradation ability, good biocompatibility and good wear resistance, while biomedical metallic materials happen to possess high yield strength (YS), high ultimate tensile strength (UTS), and high fatigue resistance and fracture toughness, and thus have become one of the most widely used materials in craniomaxillofacial surgery;<sup>4</sup> particularly, biomedical nondegradable metals such as titanium (Ti) have been widely used in the treatment of calvarial defects, reducing the source of infection after craniotomy<sup>5</sup> and repairing the maxillofacial defects of patients with common diseases in craniomaxillofacial surgery.<sup>6</sup> But biomedical nondegradable metals also have limitations. Taking Ti plates as an example, due to their excellent corrosion resistance, they will not disappear with the patient's recovery after implantation in the human body, which means that they need to be removed by secondary surgery after healing, and due to the complex craniomaxillofacial structure, reoperation means a greater risk; in addition, Ti plates will cause an obstacle to child growth and produce hypersensitivity to cold stimuli and stress shielding.<sup>7,8</sup> Therefore, while continuing to optimize the performance of nondegradable metal materials for biomedical applications, researchers are actively developing degradable metal materials, such as iron (Fe), magnesium (Mg), zinc (Zn), and their alloys

<sup>a</sup>School of Materials Science and Engineering, University of Science and Technology Beijing, Beijing, 100083, China. E-mail: huafangli@ustb.edu.cn

<sup>b</sup>Lepu Medical Technology Co., Ltd, Beijing 102200, China

which are representative of the current situation at a certain scale, but still have various deficiencies, such as the mechanical properties being relatively poor, the degradation rate not matching the rate of wound recovery and so on. Furthermore, there is a lack of up-to-date and systematic summary of the current research progress of biomedical metal materials in craniomaxillofacial surgery, which may hinder the continuous optimization design of the performance of biomedical metal materials. Therefore, this article summarizes the recent research progress in conventional biomedical metal materials including Ti and its alloys, stainless steel (SS), tantalum (Ta) and biomedical degradable metal materials including iron (Fe), magnesium (Mg), and zinc (Zn), and their alloys in craniomaxillofacial surgery including the skull, maxilla, and mandible. In addition, the challenges and future research directions on metallic biomaterials that can be used in craniomaxillofacial surgery are discussed.

## 2. Conventional nondegradable metallic biomaterials in craniomaxillofacial reconstruction

As one of the first artificial materials to be implanted in craniomaxillofacial bone, biomedical non-degradable metals have been extensively investigated and are widely used in clinical practice due to their excellent mechanical properties, biocompatibility, and corrosion resistance.<sup>4,9</sup> In particular, Ti and its alloys have been developed in various forms for use in craniomaxillofacial surgery such as titanium–nickel (Ti–Ni) springs for the treatment of sagittal cranial suture premature closure,<sup>10</sup> Ti mesh for the treatment of skull defects,<sup>5</sup> and Ti plates and screws for the treatment of mandibular fractures.<sup>6</sup> This section will focus on the progress of research on the use of Ti-based alloys in craniomaxillofacial surgery. Other commonly used conventional nondegradable biomedical metal materials are also presented, including SS, cobalt–chromium (Co–Cr) alloys, and Ta and their main mechanical and biological performances in craniomaxillofacial reconstruction applications are summarized and compared in Table 1.

### 2.1. SS for craniomaxillofacial reconstruction

The use of SS for craniofacial repair can be traced back to 1886, but the effect was unsatisfactory until 1967, when Luhr<sup>29</sup> performed a maxillofacial surgery using a compression plate, and then evolved the experience of craniofacial reconstruction. Examples include the use of dynamic compression plates,<sup>30</sup> self-healing plates for mandibular splits,<sup>31,32</sup> and the use of monocortical microplates for mandibular and midface fractures without axial compression.<sup>33</sup>

SS springs are also used in the field of craniofacial repair. Gewalli *et al.*<sup>11</sup> implanted a dynamic spring made of SS into the rabbit's skull to verify the treatment of premature fusion of the cranial sutures by the dynamic spring (in this experiment, the titanium–molybdenum (Ti–Mo) alloy was used as the

control group). The results showed that the force exerted by the SS spring on the skull was stronger than that of the Ti-based shape memory alloy, both *in vivo* and *in vitro*. In addition, relatively long-term (24 weeks) *in vivo* experiments using rabbits also showed that this spring had good biocompatibility, which was manifested in the formation of tissue in the sutures. These results were later confirmed by David *et al.*<sup>12</sup> Lauritzen *et al.*<sup>13</sup> were more concerned about the role of SS springs in children's skull remodeling. Through the implantation of SS springs, the shape of the patient's skull was restored to the normal range within three months. At the same time, the removed spring remains in its original shape. These experiences all brought revolutionary changes to the high corrosion-resistant metallic implants in the field of maxillofacial correction. Since then, many researchers have indirectly tested the relevant properties of non-spring SS by choosing stainless steel as the experimental reference. The results showed that SS exhibited lower biocompatibility and corrosion resistance than Ti and its alloys and lower wear resistance than Co–Cr alloys.<sup>4</sup>

### 2.2. Ti and Ti-based biomedical metals for craniomaxillofacial reconstruction

Ti is one of the few materials that naturally meet the requirements of human implantation due to its light weight, high specific strength, low toxicity, and high corrosion resistance.<sup>34</sup> Despite the significant developments in complete tissue-engineered cranial and maxillofacial structures, Ti is still the preferred artificial material for the repair of defects in load-bearing areas, including the mandible, maxilla and craniofacial.<sup>35,36</sup>

**2.2.1. Pure Ti.** Brem *et al.*<sup>14</sup> found that commercially pure Ti (CP-Ti; Grade 2) could combine hardness and maximum strength with sufficient malleability to avoid brittle fracture, so suitable plates and screws were prepared by this method. Eichner<sup>37</sup> and Schwickerath<sup>38</sup> made a mechanical comparison of the Ti plate formed in this way, and found that the bearable force of the plate depends on the YS and geometric shape of the plate-screw system (the size and notch effect caused by screws and screw holes), in which maxillary plate < mandibular plate < SS (590 N < 1250 N < 1250 N). These forces are much greater than the force required for physical chewing (100 N), which can meet the requirements of daily chewing. Animal experiments showed that the abovementioned materials accelerated new bone formation, while restrained connective tissue appearance in mini-pigs. Kuttenger *et al.*<sup>15</sup> conducted a long-term follow-up study of Ti implants in the oral and maxillofacial bones. The results showed that no wound infections, exposures or loss of the mesh were observed, which suggested that the long-term stability of the reconstructions was excellent. Researchers then conducted the study of the walls of paranasal sinus reconstruction and found that over time, complete re-pneumatization took place. Wind *et al.*<sup>5</sup> then used Ti to perform immediate cranioplasty on three patients and found that the patients had no further postoperative complications and no evidence of recurrent infection over a follow-up

**Table 1** Conventional nondegradable metallic biomaterials and their mechanical and biological performances in craniomaxillofacial reconstruction applications

Type	Alloy	Implant site	Results		Ref.
			Mechanical properties	Biocompatibility	
SS	SS	Cranial bone	<b><i>In vitro</i></b> -SS springs exerted about 1.7 times more force than Ti-Mo springs. <b><i>In vivo</i></b> - Force (SD) provided at variant sites: $F_{2\text{ mm}} = 2.42\text{ N}$ , $F_{9\text{ mm}} = 2.18\text{ N}$	<b><i>In vivo</i></b> -No infections and postoperative complications - There is osteogenic activity, but osteogenesis is slow.	11
			<b><i>In vitro</i></b> -SS springs exerted about 1.7 times more force than the memory metal (Ti-based). <b><i>In vivo</i></b> - Force (SD) provided at 2 mm arm spread: $F_{SS} = 2.3\text{ N}$ , $F_{Ti\text{-based}} = 1.3\text{ N}$ —	<b><i>In vivo</i></b> -No short- or long-term technical complications -Have bony ingrowth	12
Ti	Pure Ti	Cranial bone	<b>70% deformation</b> $YS = 750\text{ MPa}$ , $\epsilon = 8\%$	<b><i>In vivo</i></b> -New bone formation was found beside and above the miniplates and screws. -No connective tissue	14
	Pure Ti	Cranial bone	$Torgue_{Ti} = 0.53\text{ Nm} \approx Torgue_{xZCrNiMo} = 0.52\text{ Nm}$ —	<b><i>In vivo</i></b> -No further postoperative complications and recurrent infection (over a follow-up period of greater than 3 years)	5
	Pure Ti	Oral and Maxillofacial	—	<b><i>In vivo</i></b> -No wound infections and exposures -Excellent long-term stability of the reconstructions - Sinus walls were reconstructed.	15
	Ti-6Al-4V	Cranial bone	<b>-Construct E (MPa):</b> $146 \pm 9$ to E-PBF, $156 \pm 10$ to L-PBF <b>-Energy absorbed at peak load (J):</b> $3.1 \pm 0.1$ to E-PBF, $6.0 \pm 0.4$ to L-PBF	—	16 and 17
	Ti-6Al-4V	Mandibular bone	—	<b><i>In vitro</i></b> -Hydrophilic -Sample extracts from plasma polishing have serious cytotoxicity. -V significantly inhibited the proliferation of human gingival cells (human parotid gland epithelial cells and human prostatic growth factor).	18
	Ti-6Al-4V (porous)	Frontal skull	<b>Parallel:</b> $E = 12.9(\pm 0.9)\text{ GPa}$ , $CYS = 107.5(\pm 3.6)\text{ MPa}$ , $UCS = 148.4(\pm 3.5)\text{ MPa}$ <b>Perpendicular:</b> $E = 3.9(\pm 2.1)\text{ GPa}$ , $CYS = 49.6(\pm 20.6)\text{ MPa}$ , $UCS = 127.1(\pm 29.2)\text{ MPa}$	<b><i>In vivo</i></b> -The percentage of bone volume over tissue volume inside the pores steadily increased. - After 60 days the implants were filled completely with bony tissue. - The histological analysis revealed only scarce bone-implant contact.	19
	Ti-2Ag	Mandible	$\frac{1}{3}BS_{\text{hole Ti-Ag}} \approx \frac{2}{3}BS_{\text{hole Ti-Ag}} \approx BS_{CP\text{-Ti}}$	<b><i>In vitro</i></b> - Ag ion release: $\sim 200\text{ ppb}$ (4 h), $\sim 150\text{ ppb}$ (24 h) <sup>†</sup> - Lower cytotoxicity than CP-Ti	20
	Ti-Mo	Cranial bone	<b><i>In vitro</i></b> SS springs exerted about 1.7 times more force than the Ti-Mo springs. <b><i>In vivo</i></b> - Ti-Mo expander elements gave 1.39 (0.25) N at 2 mm arm spread and 1.09 (0.22) N at 9 mm arm spread.	<b><i>In vivo</i></b> - No infections and postoperative complications - There is osteogenic activity, but osteogenesis is slow.	11
Ti-Ni	Skull and mandible	<b><i>In vitro</i></b> - The force varied with the shape and the number of deformations cycling. - Have a permanent plastic strain of 4% <b><i>In vivo</i></b> -Implant movement	<b><i>In vivo</i></b> Clinical research confirmed the possibility of applying superelastic rings and springs in cranioplasty.	10 and 21	
Ti-Ni	Cranial bone	Microhardness values of nitinol averaged $350\text{ kg mm}^{-2}$	<b><i>In vivo</i></b> - No macrophage adhesion - New bone was similar in hardness to old bone. - Bone contact: $BC_{HA} > BC_{NiTi}$ - Bone ingrowth: $BI_{HA} < BI_{NiTi}$	22	
Ti-Ni	Cranial bone	—	<b><i>In vivo</i></b> New bone formation No apparent inflammatory response	23	

Table 1 (Contd.)

Type	Alloy	Implant site	Results		Ref.
			Mechanical properties	Biocompatibility	
Others	CP-Ta	Cranial bone	—	The rate of infection is alarmingly high in these large constructs	24
	CP-Ta (porous)	Cranial bone	UCS = 50–70 MPa UTS = 63 MPa, BS = 110 MPa	- There was a complete closure of the cranial defect by the new bone formation. - No inflammatory reaction - The connective tissue might form on the Ta surface in the early stage.	25 and 26
	CoCr	Mandibular bone	Prostheses were well tolerated by surrounding tissues and no evidence of wear of the metal implants causing osteolysis and formation of granulomatous.	- Normal organ function - New tissue and bone formation - Produced no evidence of tissue reaction	27
	Vitallium	Skull	—		28

$\epsilon$ : elongation; CP-Ta: commercially pure Ta; UCS: ultimate compression strength; CYS: compression yield strength; and BS: Bending strength. <sup>a</sup> The minimum inhibitory concentration of Ag ions for bacteria and yeast is 30 to 1000 ppb ( $\mu\text{g L}^{-1}$ ), depending on the type of bacterium and the initial concentration. The cytotoxic concentration of Ag ions to mouse fibroblasts (L929) and human fibroblasts during a short incubation was 500 and 2000 ppb, respectively.

period of greater than 3 years. However, the lack of mechanical properties and wear resistance of commercially pure titanium limited its development to some extent.

**2.2.2. Ti alloys.** Lewin *et al.*<sup>16,17</sup> investigated the effect of different pore shapes and different printing techniques on the mechanical properties of the additively manufactured Ti–6Al–4V meshes implanted in cranial bone. On the one hand, they compared the mechanical properties of two Ti meshes (D1 and D2) with different pore shapes and showed that the quasi-static applied pressure causes the D2 structure to fracture after 20 mm of deformation, while the D1 structure is only partially deformed. In the impact test, there was no significant difference between the two designs. In general, the deformation of D1 was better controlled and it was less dependent on the loading rate. Furthermore, Ti–6Al–4V samples were prepared by the hot isostatic laser powder bed melting technique (HIPed PBF-L) and the electron beam powder bed melting technique (PBF-EB), respectively, using the D2 model, and the mechanical testing showed that HIPed PBF-L had better mechanical properties.

The main reason for this discrepancy is the large geometric deviation of the samples prepared by PBF-EB. In addition, the large surface roughness of the samples prepared by the PBF-EB technique is also a critical issue, because the surface roughness concentrates stress and reduces the mechanical strength of the printed parts. In addition, larger roughness has a greater effect on cell activity. Bernhardt *et al.*<sup>18</sup> studied the influence of the surface roughness of additively manufactured Ti–6Al–4V implants on their material properties and *in vitro* biocompatibility. According to the rule that the effective roughness spectrum  $R_a = 0.2\text{--}2\ \mu\text{m}$  which is suitable for short-term oral and maxillofacial implants, the researchers subjected the samples prepared by laser melting deposition to different surface treatment procedures. The results showed that plasma electrolytic polishing would lead to excessive release of vanadium(v) ions, thus reducing cell activity. However, the biocompatibility of samples formed by vibration polishing and electropolishing basically met the requirements

of implantation. In addition, the combination of different polishing methods also affected cell activity.

Aluminum (Al) enrichment of the  $\alpha$ -phase in Ti was found to be detrimental to its passivity and corrosion resistance.<sup>39</sup> Moreover, Al may cause Alzheimer's disease,<sup>40,41</sup> while V is toxic to cells and can cause severe allergic reactions,<sup>42</sup> and hence, researchers focused their attention on Al-free and/or V-free Ti alloys, such as titanium–silver (Ti–Ag)<sup>20</sup> and nickel–titanium (NiTi).<sup>21–23</sup> As shown in Table 1, several papers about NiTi were published in recent decades focusing on cranial bone repair,<sup>21–23</sup> particularly for children,<sup>21</sup> because of its superelasticity. For instance, Lekston *et al.*<sup>10</sup> fabricated U- and  $\Omega$ -shaped springs using NiTi wires with different diameters and performed mechanical assessments both *in vivo* and *in vitro*, and the results indicated that the smaller the diameter of the ring, the larger the pre-deformation during formation and the higher the induced stress. Moreover, U-shaped springs increased the force and the increase in the deformation cycling decreased the force. Morawiec *et al.*<sup>21</sup> suggested that small formation favorably affected induced superelasticity. Simske and Sachdeva<sup>22</sup> proved that porous NiTi alloy had good biocompatibility and osteogenic properties through *in vivo* assessments. For osteogenic properties, they found that a relationship existed between the pore size and both bone apposition and ingrowth. Ayers *et al.*,<sup>23</sup> however, thought if implants were pressed into bone, bone growth might be independent of pore size. Meanwhile, above 150  $\mu\text{m}$  pore size, increased porosity was unnecessary to enhance early cartilaginous ingrowth.

However, the allergy, toxicity, and potential carcinogenicity associated with Ni ion release in NiTi alloys limit their use.<sup>43</sup> In order to overcome this problem, surface modifications such as oxidation treatment of NiTi to obtain a Ni-free surface<sup>44</sup> and several alternative Ni-free shape memory alloys, such as Nb-based, have been developed.<sup>45</sup>

There are also studies on the corrosion resistance of Ti in the craniomaxillofacial region, for example, in order to investigate the effect of the surface modification on the corrosion be-

behavior of Ti, long-term open circuit potential (OCP) measurement, cyclic voltammetry, and electrochemical impedance spectroscopy (EIS) were carried out by Suba *et al.*<sup>46</sup> Researchers have found that Ti-oxide layers produced by thermal and electrochemical processes prevent corrosion of pure Ti substrates, both *in vivo* and *in vitro*. However, the non-density of oxide coating and mechanical damage can make  $\text{Cl}^-$  in the body fluid directly in contact with the substrate, thus accelerating corrosion. However, the spontaneous formation of phosphate and hydrophosphate (apatite)<sup>47,48</sup> in the human body may close these pores, and thus realize the continuous re-passivation and repair of defects in the layer. González and Mirza-Rosca<sup>39</sup> and Marino and Mascaro<sup>49</sup> studied the corrosion of Ti in the oral environment. It was found that Mo, V and Fe improved the passivation of Ti and inhibited the active corrosion of the  $\beta$ -phase.<sup>39</sup> The EIS of grade 2 Ti was tested by immersion in artificial saliva, and the results showed that with the extension of immersion time, different structures of  $\text{TiO}_2$  (the anodically grown oxide is generally amorphous whereas thermal oxidation produced mainly the anatase phase) would be formed on the surface of Ti, which has long-term and stable passivation performance; however, the porous structure in grade 2 Ti reduced the corrosion resistance compared to solid pure Ti.<sup>49</sup>

### 2.3. Other nondegradable metallic biomaterials for craniomaxillofacial reconstruction

Ta implants in the cranio-maxillofacial area have a longer history than Ti because of its unique combination of properties such as high corrosion resistance<sup>25</sup> and biological inertia, as demonstrated experimentally by Ta cranioplasties in 11 cats.<sup>50</sup> One important finding in this animal study was the closure of the cranial defect by new bone formation.<sup>25</sup> The first Ta biomaterial implanted in a human skull can be traced back to 1941 in the US Naval Hospital in Washington, DC by Fulcher.<sup>51</sup> Afterwards, especially in World War II, more than 1000 cases of Ta cranioplasty were performed in the subsequent 5 years. However, the high thermal conductivity of Ta leading to cold or heat sensitivity was a shortcoming and there has been no definitive study on whether Ta implantation in human cranial and maxillofacial regions would have adverse effects on patients or the therapeutic effect on diseases, and Ta use in cranioplasty declined precipitously in the early 1950s.<sup>52</sup> However, with the advancement of manufacturing technology, the porosity and average pore size of Ta-based alloys can reach 75%~85% and 400–600  $\mu\text{m}$ , respectively. However, studies have shown that the average pore size of 400  $\mu\text{m}$  and the porosity of more than 70% can promote osteogenic differentiation and the formation of blood vessels and bone tissues.<sup>53–56</sup> In addition, Ta has a low bacterial adhesion rate but poor bactericidal effect.<sup>57</sup> Porous Ta also has excellent mechanical properties such as a large coefficient of friction (*i.e.*, high stability at the initial stage of implantation)<sup>58</sup> and elastic modulus (E) and compressive strength close to those of cortical bone (*i.e.*, load-bearing capacity).<sup>53,59</sup> These superior properties of Ta-based alloys enabled them to find widespread use in orthopedics and dentistry. However, the excellent bio-

logical properties of porous Ta are obtained at the expense of some excellent mechanical properties.<sup>60</sup> Therefore, balancing the biological and mechanical properties of porous Ta becomes a new challenge for Ta alloys.<sup>61</sup>

In addition, there have been a small number of cases of cranial and maxillofacial implants made of Co–Cr or vitallium alloys. For example, several Vitallium plates have been fabricated and tested,<sup>58</sup> which showed excellent *in vivo* biocompatibility with no adverse effects to the patients' skulls.<sup>62,63</sup> Kummoona *et al.*<sup>27</sup> concluded that Co–Cr implants could satisfy the patients' physiological needs for mandibular implants and muscle fiber attachments, based on long-term health tracking of patients with 2-part Co–Cr prosthesis.

## 3. Biodegradable metallic materials for craniomaxillofacial reconstruction

It is estimated that 10–12% of craniofacial implants are removed due to infection, exposure, pain and discomfort.<sup>64</sup> Therefore, the development of biodegradable materials continues to receive increasing attention. A variety of biodegradable metallic materials for biomedical use with potential for implantation into the craniomaxillofacial surface have been initially developed in recent years, such as splints and screws for fixation of maxillofacial fractures,<sup>65</sup> metallic membranes for guided craniomaxillofacial bone regeneration<sup>66,67</sup> and devices for anti-cranial osteolysis and/or antibacterial purposes.<sup>68,69</sup> However, most experiments remain in *in vitro* studies and are particularly underdeveloped for Fe and its alloys. In this section, we focus on the progress of research in craniomaxillofacial bone with three most typical biodegradable metals for biomedical use, namely Fe, Mg and Zn. The biodegradable metallic biomaterials and their main mechanical and biological performances in craniomaxillofacial reconstruction applications are summarized in Table 2.

### 3.1. Fe-based biodegradable metals for craniomaxillofacial reconstruction

Fe exists in almost all organisms and is essential for their development and survival.<sup>78</sup> It is a vital part of various enzymes involved in many biological processes, including DNA biosynthesis, oxygen transport, and cellular energy generation.<sup>79</sup>

Ferroalloy therefore became one of the earliest metals implanted in the human body. But in the field of metal implants, Fe is different from other metals in that it contains not only SS, a traditional implant with excellent corrosion resistance in the human body, but also new degradable metals that degrade less quickly than Zn and Mg (1–4 months), which researchers have begun to explore the way to increase their degradation rate. For instance, porous Fe-30Mn samples were prepared using 3D printing technology (Fe-30Mn 3D-P).<sup>70</sup> Mechanical testing showed that the YS of the porous samples was inferior to that of pure Fe and Fe-30Mn made by sintering and rolling, but it was the closest to that of natural bone, especially the Young's modulus ( $32.47 \pm 5.05$  GPa for Fe-30Mn



**Table 2** Biodegradable metallic biomaterials for craniomaxillofacial reconstruction and their main mechanical and biological performances in craniomaxillofacial reconstruction applications

Alloy	Implant site	Results			Ref.
		Mechanical properties	Degradability	Biological properties	
Fe-30Mn	Cranial bone	YS = 106.07 ± 8.13 MPa, UTS = 115.53 ± 1.05 MPa  $\epsilon = 0.73 \pm 0.15\%$	- Corrosion rate: $\nu = 2.81 \pm 0.88$ mm per year (nominal surface area); $\nu = 0.73 \pm 0.22$ mm per year (estimated true surface area)	<b>In vitro</b> - At the initial stage, it was toxic to MC3T3 cells. - At the later stage, it showed good cell compatibility.	70
Fe-35Mn	Cranial and maxillofacial bones	UTS = 228.1 MPa	- Corrosion unevenness	<b>In vitro</b>	71
Fe-34Mn-1Ca		YS = 189.7 MPa, $E = 39$ GPa, $\epsilon = 1.5\%$  UTS = 296.6 MPa, $E = 163$ GPa, $\epsilon = 0.1\%$ (Brittle fracture occurred)	- Corrosion rate: $\bar{\nu}_{\text{Fe-Mn}} = 0.03$ mm per year $\bar{\nu}_{\text{Fe-Mn-1Ca}} = 0.14$ mm per year	- Numerous cell-to-cell junctions and cytoplasmic expansion - Fe-Mn-Ca exhibited more live cells than Fe-Mn. - Good cell viability	
Pure Mg	Mandible	UTS = 86 MPa, YS = 20 MPa, $\epsilon = 13\%$	$V_{\text{cortical bone}} = 71.5\% > V_{\text{bone marrow}} = 9.6\%$ $\rightarrow \nu_{\text{cortical bone}} < \nu_{\text{bone marrow}}$	<b>In vivo</b> - Bone remodeling occurred in the area surrounding the screws. - New bone was seen growing up next to and in contact with the metal.	72
AZ31		The pull force of AZ31 was similar to SS	$V_{\text{cortical bone}} = 44.2\% > V_{\text{bone marrow}} = 20\%$ $\rightarrow \nu_{\text{cortical bone}} < \nu_{\text{bone marrow}}$  $V_{\text{screw head}} = 61.5\% < V_{\text{screw shaft}} \sim 39\%$	<b>In vivo</b> - New bone was seen growing up next to and in contact with the metal. - Significant bone overgrowth was also observed around the AZ31 screws.	
ZK60 + PLLA	Le Fort I	UTS = 372.94 ± 20.10 MPa YS = 251.16 ± 33.43 MPa  $\epsilon = 15.12 \pm 1.95\%$	$\rightarrow \nu_{\text{screw head}} > \nu_{\text{screw shaft}} \nu_{\text{Mg}} > \nu_{\text{ZK31}}$ - Gas formation - Uneven corrosion:  $\rightarrow \nu_{\text{cortical bone}} < \nu_{\text{bone marrow}}$	<b>In vivo</b> - The implant moved, the wound opened and became infected.	7
WE43	Frontal bone	—	<b>Volume change (<math>V_{12}-V_{24}</math>):</b> $\Delta V_{\text{WE43}} = 70 \pm 8.2\% - 43.0 \pm 20.9\%$ $> \Delta V_{\text{coated WE43}} = 80 \pm 6.3\% - 76.3 \pm 19.2\%$ <b>Production of gas:</b> $P_{\text{WE43}} = \frac{2}{3} > P_{\text{Coated WE43}} = \frac{1}{6}$ Gas occurrence decreased with time.	<b>In vivo</b> - Bone density (BV/TV): $\rho_{\text{Ti}} > \rho_{\text{coated WE43}} > \rho_{\text{WE43}}$ - Bone contact area: $\text{BIC}_{\text{Ti}} > \text{BIC}_{\text{coated WE43}} > \text{BIC}_{\text{WE43}}$ - No disturbances in wound healing or signs of foreign body reaction.	73
WE43	Le Fort I	- <b>Extruded WE43:</b> UTS = 303 MPa, YS = 195 MPa, $\epsilon = 6\%$  - <b>WE43:</b>  UTS = 260 MPa, YS = 160 MPa, $\epsilon = 6\%$	$V_{60 \text{ d}} = 60.07 \pm 15.40 = 60\% V_{\text{initial}}$ $V_{24 \text{ weeks}} = 19.37\% V_{\text{initial}}$	<b>In vivo</b> - Producing slight inflammation and gas in the initial stage (12 weeks). - New bone was well formed near the screw head and covered the screw body.	74
WE43	Cranial bone	- <b>As-cast:</b> YS = 157.5 ± 11.1 MPa  UTS = 384.9 ± 10.4 MPa - <b>As-cast + heat treatment</b>  ( <b>WE43-T5</b> ): YS = 244.1 ± 10.3 MPa	- Implant degradation was more pronounced for the WE43 group relative to the WE43-T5 group.	<b>In vivo</b> - Absence of increased inflammatory content at both bone and soft tissues interface; - New woven bone formation; - Lymph node Mg content quantification: No implant = 1126 ± 430 ppb WE43 = 2125 ± 411 ppb ≈ WE43-T5 ( $p > 0.38$ )	75
Mg-4Y-3RE	Cranium	UTS = 427.7 ± 17.1 MPa HV = 114, YS = 280 MPa, UTS = 316 MPa, $\epsilon = 5.3\%$ , CYS = 239 MPa, UCS = 402 MPa	<b>In vitro:</b> - $\nu_{\text{corrosion rate}} = 1.51$ mm per year - Local corrosion <b>In vivo:</b> - $\nu_{\text{corrosion rate}} = 0.90$ mm per year - Corrosion products were thick and irregular.	<b>In vivo</b> The biodegradable Mg-based implants have no adverse effects on the behavior or physical condition of rats	76

Table 2 (Contd.)

Alloy	Implant site	Results			Ref.
		Mechanical properties	Degradability	Biological properties	
Mg-6Zn-RE (Y, Gd, La and Ce)	Cranial bone	—	<b>In vivo</b> Rapid degradation, 2 weeks to fragmentation and cavitation	<b>In vivo</b> - Swelling at the wound area at the beginning of implantation, which subsides within 2 weeks - Excellent osteogenic properties - Milder inflammation - Rare earth elements accumulate in many organs and are metabolized through lymphatic vessels.	77
Pure Zn	Maxillofacial bone	YS = 92.16 ± 8.76 MPa, UTS = 108 ± 4.87 MPa, $\epsilon$ = 42.82 ± 2.69%	<b>In vivo</b> A relatively uniform corrosion mode	<b>In vivo</b> - Favorable biocompatibility <b>In vivo</b> - Well attached to the cranial defect area - New bones formed in the edge of implants. - No obstruction, histopathological changes and accumulation of degradation products	66
CP-Zn	Maxillofacial bone	A significant reduction in UTS	<b>In vitro</b> - Local corrosion (pore)  - The bigger the hole, the faster the corrosion. <b>In vivo</b> - The pure Zn membrane with 1000 $\mu$ m pores almost broke down at week 10.	<b>In vitro</b> - Favorable biocompatibility to 300 $\mu$ m - Noticeable toxicity to 1000 $\mu$ m <b>In vivo</b> - Well attached to the cranial defect area - No obstruction, histopathological changes and accumulation of degradation products - New bones formed in the edge of implants.	
Zn-xAg (x = 0.5, 1, 2%)	Cranial bone	- Increased tensile strength - Significant reduction in compressive strength - $\epsilon > 30\%$	<b>In vitro</b> - The addition of Ag accelerated the corrosion rate, with the highest corrosion rate for Zn-0.5Ag	<b>In vitro</b> - Antibacterial properties:  Zn < Zn-0.5Ag << Zn-1Ag < Zn-2Ag - Excellent cytocompatibility - Zn-2Ag significantly inhibited osteoblast differentiation. <b>In vivo (only Zn-2Ag)</b> - Good antibacterial properties and osseointegration in the femur. - Effective inhibition of titanium particle-induced cranial osteolysis.	69
Zn-xCu (x = 1, 2, 4 wt%)	Cranio-maxillofacial bone	<b>Zn-1Cu:</b> YS = 25.8 MPa, UTS = 32.5 MPa <b>Zn-2Cu:</b> YS = 50.1 MPa, UTS = 59.5 MPa <b>Zn-4Cu:</b> YS = 73.0 MPa, UTS = 105.4 MPa, $\epsilon$ = 42.2%	- $R_f^{Zn-Cu}$ was either maintained or increased over 40 days of immersion. - The rolling process appears not to enhance the corrosion resistance of the Zn-Cu alloys. - The degradation rate of the Zn-Cu alloy may be reduced, while Zn-4Cu has the highest corrosion resistance.	<b>In vitro</b> - Good cell-to-cell connections - No apparent cytotoxicity (reduction of cell viability <30%) - Promotes cell proliferation to as-rolled Zn-4Cu - Poor resistance to adherent mixed oral bacteria	68
Zn-0.5Cu-xFe (x = 0, 0.1, 0.2 and 0.4 wt%)	Cranial and maxillofacial bones	- <b>YS and UTS:</b> Zn < ZnCu < ZnCuFe  Max: ZnCu-0.4Fe - $\epsilon$ (%): 20.5 = ZnCu-0.4Fe < Zn Zn < ZnCu < 40 < ZnCuFe Max: ZnCu-0.1Fe	<b>In vitro</b> - Min: ZnCu, Max: ZnCu-0.4Fe  - Compared to $\alpha$ -MEM, ZnCuFe corrodes more rapidly in artificial saliva and increases significantly with increasing Fe content.	<b>In vitro</b> - ZnCu-0.2Fe initially meets the requirements of the GBR membrane for cell selection <sup>a</sup> . - ZnCu-0.2Fe has a good bacteriostatic and not bactericidal effect.	67

Table 2 (Contd.)

Alloy	Implant site	Results			Ref.
		Mechanical properties	Degradability	Biological properties	
Zn–2Mg	Cranial bone	HV = 97, YS = 235 MPa, UTS = 365 MPa, El. = 4.9%,  CYS = 231 MPa, UCS = 426 MPa	<b>In vitro:</b> - $v_{\text{corrosion rate}} = 0.09$ mm per year  <b>In vivo:</b> - $v_{\text{corrosion rate}} = 0.10$ mm per year - Corrosion was relatively slow and uniform.	<b>In vivo</b> - Favorable biocompatibility to rats	76
Zn–Mg–Fe	Mandibular bone	In a relatively long time (24 weeks), the mechanical properties of Zn–Mg–Fe did not decrease with the extension of time.	<b>In vivo:</b> $v_{4 \text{ weeks}} = 0.033 \pm 0.015$ mm per year, $v_{12 \text{ weeks}} = 0.079 \pm 0.009$ mm per year, $v_{24 \text{ weeks}} = 0.095 \pm 0.009$ mm per year	—	65

HV – Vickers hardness. <sup>a</sup> Blocks fibrous connective tissue cells and epithelial cells from surrounding soft tissue, but allows osteoblasts to enter the bone defect area.

3D-P compared to 14.1–17.3 GPa for natural cortical bone<sup>80</sup>). Corrosion testing and cytotoxicity assessment showed that Fe-30Mn 3D-P exhibited a significantly higher corrosion rate than pure Fe with localized corrosion and good biocompatibility.<sup>70</sup> Hong *et al.*<sup>71</sup> reported that the addition of Ca/Mg to Fe–Mn increased the corrosion rate of the alloy, especially the addition of Mg had a stronger effect on the corrosion rate (Fig. 1a and b); and the microporous Fe–Mn and Fe–Mn–1Ca metal sheets created by binder-jet 3D printing demonstrated good biocompatibility (Fig. 1c).

As Fe presents different valence states under physiological conditions, when it exists in the form of ferrous -Fe(II), it produces reactive oxygen species (ROS) through the Fenton reaction, which damage the functions of cells and organs. In addition, Fe-induced damage is aggravated by the fact that there is no physiological way of Fe excretion, apart from blood loss and (to a lesser extent) shedding of cells.<sup>81</sup> The precipitation form and content of Fe limit the implantation of Fe in cranial and maxillofacial areas.

### 3.2. Mg-based biodegradable metals for craniomaxillofacial reconstruction

The suggested daily intake of Mg lies in the range of 375–500 mg, and 0.75–1.15 mmol L<sup>-1</sup> within the blood

plasma.<sup>82,83</sup> Mg ions are involved in many reactions in the body, such as DNA synthesis, RNA transcription,<sup>84</sup> ion diffusion, and formation of cellular signals.<sup>85</sup> Excess Mg also collects in the kidneys and is excreted in the urine, which provides for Mg implantation in the body without disrupting the body Mg balance. In addition, the good osteogenic properties<sup>72,77</sup> and biomechanical stability of Mg in the craniomaxillofacial region have been demonstrated.<sup>86,87</sup> For example, three Mg screws on top plus two Mg screws at the bottom can maintain the mechanical stability of mandibular propulsion, while two Mg screws on top combined with one Mg screw at the bottom can meet the implantation requirements for sagittal split osteotomy of the mandibular branch.<sup>86</sup>

The excellent biological properties and good mechanical properties have led to a faster progress of research on Mg alloys in craniomaxillofacial surgery than Fe and Zn, and the Mg-based MAGNEZIX@CS screws developed by Syntellix AG are currently in clinical use, which are used to accomplish sturdy internal fixation of mandibular condylar fractures by incision and reduction.<sup>88</sup> In addition, sufficient mechanical properties and degradability have led to the use of Mg-based alloys for the preparation of metal membranes to guide cranial regeneration, such as Mg–Al–Zn alloy,<sup>89</sup> Mg–Zn–RE(Y, Gd, La and Ce),<sup>77</sup> WE43, and Mg3Gd alloys.<sup>90</sup>









Fig. 1 (a) CALPHAD simulation showing that addition of Ca or Mg to a binary Fe–Mn alloy can accelerate the corrosion of the alloy; (b) Tafel curves measured by potentiodynamic polarization of Fe–Mn, Fe–Mn–Ca, and Fe–Mn–Mg pressed pellets; (c) cell viability of MC3T3–E1 cells incubated with extracts of 3D printing Fe–Mn and Fe–Mn–1Ca for 3 d ( $n = 3$ )<sup>71</sup> (Copyright 2016 by Elsevier).



However, the corrosion behavior of Mg, such as excessive degradation and the production of hydrogen gas,<sup>91</sup> has led to particular interest in improving the composition and surface properties of Mg alloys for craniomaxillofacial bone applications. The corrosion reactions of Mg in a physiological environment can be expressed as follows:



Henderson *et al.*<sup>72</sup> studied the *in vivo* degradation behavior of Mg and AZ31 and showed that the degradation rate of Mg-based materials varied depending on the surrounding tissue, exhibiting a significantly lower degradation in the cortical bone than in the medullary cavity, besides, the degradation rate varied between different parts of the same screw.

Based on the fact that alloying has an important influence on the properties of Mg alloys, especially on corrosion behavior, Kubásek *et al.*<sup>76</sup> studied the *in vitro* and *in vitro* mechanical properties and corrosion behaviors of Mg-4Y-3RE and reported that rare earth element (REE) suppressed the formation of strong basal texture so that resulted in a decrease in the anisotropy of the mechanical properties of the Mg alloy, and its corrosion rate was still faster than that of the Zn-2Mg alloy (corrosion rate = ~0.1 mm per year) with local corrosion. In contrast, Zhao *et al.*<sup>77</sup> investigated the degradation behavior of Mg-6Zn-RE (REE = yttrium (Y), gadolinium (Gd), lanthanum (La) and cerium (Ce)) in critical-sized rat calvarial defects and showed that the large surface area accelerated the corrosion of the Mg alloy and that the corrosion channels formed inside the metal film decomposed the membrane into fragments within two weeks. However, these fragments continued to remain stable in the skull defect and formed gas bubbles due to the presence of surface deposits, and these fragments continued to decompose in the gas bubbles.<sup>77</sup> In addition, a rapid increase in the content of REEs was observed in the lymph nodes at 4 weeks of implantation due to the dissolution of deposits on the implant surface, which may have adverse effects on human health,<sup>92</sup> and it has also suggested that REEs may be phagocytosed by macrophages and enter the lymphatic system to complete metabolism. shows micro-CT images showing the new bone mass at the skull defect after implantation for 2 weeks, 5 weeks, 8 weeks, and a blank defect at 8 weeks.<sup>77</sup> It can be seen that an increase in bone mass was observed with increasing implantation time and the defect was completely covered at eight weeks.

Byun *et al.*<sup>7</sup> attempted coating the ZK60 Mg alloy with a poly(L-lactic)-acid (PLLA) polymer and investigated its degradation behavior in a LeFort I osteotomy canine model for max-

illofacial applications; the  $\mu$ -CT images indicated that all the PLLA-coated plates were completely dissolved in the five beagles and the wound dehiscence did not heal after 12 weeks of implantation. Rapid biodegradation of PLLA-coated ZK60 was observed due to the formation of microcracks during the bending process while fitting on the bone, which allowed the bodily fluid to penetrate into the coating and directly contact with ZK60. To improve the corrosion behavior and torsional strength of Mg alloys on the frontal bone, Schaller *et al.*<sup>73</sup> used a proprietary Magoxid electrolyte to form an electrocoating layer on WE43 and reported that this coating caused the degradation rate of WE43 to be slowed and stable. In addition, the presence of the coating also delayed the release of hydrogen gas, and the occurrence of soft tissue cavities caused by gas release decreased with increasing implantation time. Although the formation of a coating enhanced the bone-forming ability of WE43, it was still insufficient compared with Ti.

Byun *et al.*<sup>74</sup> and Torroni *et al.*<sup>75</sup> improved the properties of the WE43 Mg alloy by improving the preparation conditions to satisfy the conditions of cranial and maxillofacial implantation. In the former study,<sup>74</sup> the mechanical properties of WE43 were improved by tempering after casting, with 55% and 28% improvement in the YS and UTS for WE43-T5 specimens over the as-cast counterparts, respectively. Extrusion before heat treatment also improved the YS of WE43.<sup>75</sup> WE43 was biocompatible (*e.g.*, showing new bone formation and without long-term inflammation), except that the WE43-T5 produced hydrogen gas at the beginning of implantation.<sup>74</sup>


Overall, the Mg-based alloys in a physiological environment not only exhibit a rapid corrosion rate but also lead to the excessive evolution of hydrogen gas. Due to the special craniomaxillofacial anatomical characteristics, subcutaneous gas cavities may lead to wound healing disorders. These issues impede the widespread use of Mg-based alloys in craniomaxillofacial surgery.<sup>68</sup>

### 3.3. Zn-based biodegradable metals for cranio-maxillofacial reconstruction


Zn is the second most abundant micronutrient in living organisms<sup>93</sup> with normal plasma levels of 70–120 mg dL<sup>-1</sup>.<sup>94</sup> Humans consume an average of 4–15 mg day<sup>-1</sup> of Zn<sup>94</sup> to regulate various essential biological functions, including nucleic acid metabolism, signal transduction, apoptosis regulation, and gene expression in the human body,<sup>95</sup> especially for bone formation processes, such as activation of amino acid-tRNA synthetase in osteoblasts<sup>96</sup> and inhibition of differentiation of bone marrow cells to osteoclasts.<sup>97</sup> Furthermore, studies have shown that Zn has an important role in the prevention and treatment of skeletal and neurological disorders,<sup>98–100</sup> which provides strong support for the use of Zn alloys in craniomaxillofacial surgery. On the other hand, the moderate degradation behavior of Zn does not leave large amounts of corrosion products in the body that are difficult to be eliminated and its corrosion products,<sup>77</sup> such as Zn phosphate, may also improve the biocompatibility of Zn-based implants.<sup>101,102</sup>

**3.3.1. Pure Zn.** The critical role of Zn in the human body and its corrosive nature has promoted some research in craniomaxillofacial surgery. Guo *et al.*<sup>66</sup> prepared pure Zn membranes with different pore sizes (0, 300  $\mu\text{m}$ , and 1000  $\mu\text{m}$ ) for guided bone regeneration (GBR) using the degradability of Zn and its effect on bone marrow cell differentiation in order to investigate the feasibility of using pure Zn membranes for GBR in maxillofacial applications. This study demonstrated that the degradation rate increased with increasing pore size and the Zn membrane with 1000  $\mu\text{m}$  pore size corroded most rapidly, and even showed severe cracking after immersion for 28 d in Hanks' solution. The degradation rates of the Zn membranes with different pore sizes were consistent with the  $\text{Zn}^{2+}$  ion concentrations of the immersion solutions with  $C_{(0\ \mu\text{m})} = 14.4 \pm 1.5\ \mu\text{g ml}^{-1}$  for membranes without pores,  $C_{(300\ \mu\text{m})} = 19.1 \pm 0.9\ \mu\text{g ml}^{-1}$  for membranes with 300  $\mu\text{m}$  pore size, and  $C_{(1000\ \mu\text{m})} = 21.3 \pm 0.8\ \mu\text{g ml}^{-1}$  for membranes with 1000  $\mu\text{m}$  pore size. The 1000  $\mu\text{m}$  Zn membrane showed poor biocompatibility, insufficient mechanical strength, low bone regeneration ability, and even new bone collapse. In contrast, the 300  $\mu\text{m}$  Zn membrane showed sufficient mechanical properties (even stronger than pure Ti), excellent biocompatibility, and good osteogenic properties. Therefore, reducing the E by increasing the pore size is not a one-time-for-all solution, and the E, degradation rate and biocompatibility of the implant materials need to be balanced to meet the requirements of craniomaxillofacial implants.<sup>103</sup>

**3.3.2. Zn alloys.** Deficiencies in the properties of Zn and its alloys for craniomaxillofacial bone implants are usually improved by adding other elements in addition to increasing the porosity. The antimicrobial properties and osteolysis resistance of Zn alloys were improved by adding different concentrations of silver (0.5, 1, 2 wt%) and 2 wt% Ag addition was proved most effective.<sup>69</sup> The good antimicrobial properties were significant not only against coagulase-positive and negative staphylococci, the common types of bacteria tested, but also against drug-resistant strains (methicillin resistant *Staphylococcus aureus* and methicillin-resistant *Staphylococcus epidermidis*). The mechanical and degradation properties of these alloys are listed in Table 2.  $\text{Cu}^{2+}$  plays a role in improving the antimicrobial properties of Zn–Cu alloys, but previous research has shown that neither too high nor too low copper content can make the mechanical properties of Zn–Cu alloys meet the minimum requirements for biodegradable craniomaxillofacial bone implants (*i.e.*, YS > 230 MPa, UTS > 300 MPa, and  $\epsilon > 15\text{--}18\%$ ),<sup>104–106</sup> and that a lower copper content also makes Zn–Cu alloys insufficiently resistant to osteolysis.<sup>106</sup> Li *et al.*<sup>68</sup> reported the tensile properties of the as-cast and hot-rolled Zn–xCu ( $x = 0, 1, 2,$  and  $4\ \text{wt}\%$ ) (Fig. 2). It can be seen that for both the as-cast and as-rolled Zn–xCu alloys, the YS and UTS increased with increasing Cu addition, and the increase for the as-rolled Zn–Cu alloys was more significant, but  $\epsilon$  did not show a significant change with the Cu addition, and it was concluded that Zn–4Cu appeared to be




**Fig. 2** Tensile properties of as-cast and as-rolled Zn–xCu ( $x = 0, 1, 2,$  and  $4\ \text{wt}\%$ ) alloys: (a) stress–strain curves of as-cast alloys; (b) stress–strain curves of as-rolled alloys; (c) YS, UTS, and  $\epsilon$  of as-cast and as-rolled alloys<sup>68</sup> (Copyright 2019 by Elsevier).



**Fig. 3** Comparison of tensile properties of hot-extruded biodegradable Zn- $x$ Mg alloys ( $x = 0, 0.02, 0.08, 0.5, 0.8, 1, 2,$  and  $3$  wt%)<sup>107</sup> (Copyright 2021 by Elsevier).

the most suitable material for craniomaxillofacial bone implant applications based on its highest YS and UTS values among this series of Zn- $x$ Cu alloys. The corrosion of the Cu-containing alloys was more uniform due to the increase in the content or refinement of the second phase  $\text{CuZn}_5$ . Furthermore, Zn- $x$ Cu alloys showed good biocompatibility with a promotion effect on cell proliferation. However, it is not realistic to solely rely on the copper ions released by the metal matrix degradation to achieve a rapid and complete sterilization effect.

On the other hand, the high solubility of Cu in Zn makes the microstructure of the Zn-Cu alloys similar to that of pure Zn and it is difficult to form a second phase which would provide a reinforcing effect. Zhang *et al.*<sup>67</sup> added Fe to Zn to form an  $\text{FeZn}_{1.3}$  second phase and fabricated a non-porous Zn-0.5Cu- $x$ Fe ( $x = 0.1, 0.2, 0.4$  wt%) membranes with 1.5 mm thickness by hot extrusion for guided bone regeneration applications. It was demonstrated that the mechanical properties of the films were significantly improved, except for a significant decrease in the  $\epsilon$  of Zn-0.5Cu-0.4Fe. *In vitro* antimicrobial and cytocompatibility tests also showed good bacterial inhibition



**Fig. 4** Histological and micro-CT cross-sectional images of Zn-Mg-Fe alloy, PLLA, and Ti-6Al-4V bone plates and screws after implantation in beagles for 0, 4, and 12 weeks<sup>65</sup> (Copyright 2019 by Elsevier).






Fig. 5 Illustration of biomedical metals in the field of craniofacial repair and plastic surgery.

against *S. gordonii* media and mixed oral bacteria and appropriate cytocompatibility in Zn–0.5Cu–0.2Fe due to the release of Zn ions. Notably, the degradation rates of Zn–0.5Cu and Zn–0.5Cu–*x*Fe in  $\alpha$ -MEM and artificial saliva simulating the craniomaxillofacial environment differed considerably. For example, Zn–0.5Cu–0.2Fe did not degrade as fast as Zn–0.1Cu–0.1Fe in  $\alpha$ -MEM, but in artificial saliva, the former was significantly faster than the latter; overall, a significant increase in the degradation rate was observed with increasing the contents of Fe. This could be due to the organic components in  $\alpha$ -MEM decomposing with pH and adsorbing onto the Zn alloy, which in turn affects the corrosion behavior of the Zn alloys. Therefore, for biodegradable Zn-based alloys, the transient electrochemical corrosion results cannot directly translate into a prediction of their long-term degradation.

Fig. 3 shows a comparison of the tensile properties of hot-extruded biodegradable Zn–*x*Mg (*x* = 0, 0.02, 0.08, 0.5, 0.8, 1, 2, and 3 wt%) alloys. It can be seen that the addition of Mg significantly strengthened pure Zn,<sup>107</sup> while Kubásek *et al.*<sup>76</sup> showed that the Zn–2Mg alloy showed a more uniform and low degradation rate after implanted in rat cranium compared to WE43.

Based on this, Wang *et al.*<sup>65</sup> added Fe to Zn–Mg alloys (Zn  $\geq$  95, 0.001  $\leq$  Mg  $\leq$  2.5, and 0.01  $\leq$  Fe  $\leq$  2.5, wt%) and carried out a comprehensive study using a canine mandibular fracture model. The results showed that the Zn–Mg–Fe alloys corroded uniformly and slowly *in vivo* at a degradation rate of  $0.095 \pm 0.009$  mm per year with good biosafety and biocompatibility. At the same time, the Zn ions played a significant role in the bone formation process, which imparted the Zn–Mg–Fe alloy with superior osteogenic properties compared to Ti and PLLA. Fig. 4 shows histological and micro-CT cross-sectional images of the Zn–Mg–Fe alloy, PLLA, and Ti–6Al–4V bone plates and screws after implantation in beagles for 0, 4, and 12 weeks. It can be seen that at 4 weeks post-operation, more external callus appeared in the PLLA implant area than in the other two groups. At 12 weeks post-operation, large areas of wide and thick woven bone were formed around the Zn–Mg–Fe and Ti–6Al–4V alloys, whereas there were small and discontinuous areas of woven bone observed in the PLLA group.<sup>65</sup>

## 4. Summary and outlook

In summary, as illustrated in Fig. 5, biomedical metals have great research value in the field of craniofacial repair and plastic surgery. However, the research on large animal experiments and clinical trials is still limited. The design requirements for artificial implant materials include: (1) facilitating clinical manipulation, (2) being structurally compatible with the defect, (3) good biocompatibility (meaning that the implant can promote the development of tissues and cells in the environment of bone defects, including the proliferation and adhesion of osteogenic-related cells, the differentiation of bone progenitor cells, the integration of bone tissue with the implant material, and osteogenic-related angiogenesis<sup>108</sup>), (4) sufficient mechanical properties to provide a stable space for bone regeneration, and (5) minimizing complications.<sup>109,110</sup> These requirements make most of the currently developed biomedical metal materials still inadequate for implantation, including in the craniomaxillofacial area.<sup>111,112</sup> In addition, minimizing the impact of implant infection on the implantation is also an urgent issue that must be addressed in the field of craniofacial repair and plastic surgery. Therefore, appropriate materials and processing techniques must be continuously explored in the future. Some possible research directions on these aspects are proposed below.

### 4.1. Conventional nondegradable metallic biomaterials

**4.1.1. Improving antibacterial activity.** Infection with implant materials is detrimental to patient recovery and, once infected, it would lead to implant failure and is often associated with chronic and/or recurrent disease. On the other hand, infection is often difficult to treat because of antibiotic resistance, tolerance, and/or persistence.<sup>113–115</sup> Yavari *et al.*<sup>116</sup> applied a multilayer gelatin and chitosan-based coating containing bone morphogenetic protein (BMP)-2 and/or vancomycin to the surface of a porous Ti fabricated by selective laser melting. This structure with different types and concentrations of active substances in each layer allowed the continuous release of antibiotics to kill any bacteria that enter the body during the perioperative period and help host cells win the

“surface competition”.<sup>117</sup> The results also verified the bactericidal effect and bone formation promoting effect of the multi-layer structure. Of course, how the coating on the surface of the drug or material is firmly combined with the base metal is also a problem that cannot be ignored in design. In this regard, bionics is a strategy worth learning. For example, Wang *et al.*<sup>118</sup> learned from the molecular adhesion mechanism of the marine mussel foot protein, and used this protein to firmly “lock” metal ions ( $\text{Zn}^{2+}$ ) and bone induction growth factor (BMP-2 peptide) on the titanium group, which greatly improved the mechanical and osteogenic properties of the Ti group. In addition, the fluidity and buffering of body fluids significantly diminished the antimicrobial effect of the material *in vivo*<sup>67</sup> and therefore placed a greater demand on *in vitro* antimicrobial testing.

**4.1.2. Modulation of inflammatory response to promote osteogenic performance.** In the treatment of diseases by implantation of implants, it is inevitable to evoke a response of immune cells dominated by macrophages, which trigger inflammation, while inflammation is involved in tissue regeneration and the development of fibrosis,<sup>119</sup> in which fibrous tissue formed through early inflammation can provide a barrier to osseointegration, but earlier studies have shown that organisms cannot select by themselves the inflammatory response for the course of tissue regeneration and fibrosis, such as in amputated salamanders. Either an antifibrotic process or both a profibrotic and antifibrotic process mediated by immunity (this process is balanced) is generated,<sup>120</sup> and thus promoting wound healing with the aid of inflammatory response and achieving osseointegration need to be artificially regulated. Zhao *et al.*<sup>121</sup> developed a type I collagen (COL1) modified nano-porous network on Ti substrates *via* alkali treatment, polydopamine coating, and layer-by-layer (LBL) self-assembly, and this coating could regulate the differentiation of macrophages and inhibit the expression of inflammation and osteoclast related genes, thereby promoting subsequent vasculature/bone formation. Huang *et al.*<sup>122</sup> studied the effects of M1 type cells and M2 type cells, respectively, on human gingival fibroblasts attached to Ti substrates and similarly showed that in the M1 immune micro-environment, fibroblasts seem to aggravate inflammation in a positive cyclic manner, whereas in the M2 immune micro-environment, fibroblasts produce better soft tissue regeneration, adhesion to and contraction of metallic materials, and proposed the concept of an “inflammation-fibrous” complex. Yu *et al.*<sup>123</sup> promoted the repair of infectious (bacterial) fractures by directly regulating the release of exosomal mir-708-5p.

#### 4.2. Biodegradable metallic materials

Although it is difficult to obtain craniomaxillofacial bones at all ages of the human body, and the research on the mechanics and biology of craniomaxillofacial bones is very limited, there are still some data for reference. The data for 6 to 10 months are more representative as the samples are fresh, when they correspond to an  $E$  of 2.3 GPa to 6.4 GPa, an HV of 4.0 GPa to 7.2 GPa and a bending failure strain of 6.7%.<sup>124</sup> In addition, cranial tissue hardens with age, and previous studies

have shown that elastic modulus (determined by assuming a solid section) increases from approximately 0.5 GPa at birth<sup>125</sup> to as high as 10 GPa by the age of 6 years (at which point the elastic modulus of the cortical bone of the skull is  $9.8 \pm 71.24$  GPa and the elastic modulus of the sutures is  $1.10 \pm 0.53$  GPa). The triple-layered bone has an effective modulus of elasticity is  $3.69 \pm 0.92$  GPa).<sup>126</sup> By adulthood, the hardness of skull tissue will be three times that of children. The bending strength decreases correspondingly, and in adulthood, the bending strength is only one-fifth of that of young children.<sup>124,127</sup> Since maxillofacial fractures are the most common disease among craniomaxillofacial fractures, especially mandibular fractures, which can account for 70% of facial fractures,<sup>128</sup> researchers have done more research on the mechanical properties and stress distribution of maxillofacial bone, among which the elastic modulus is 1500 MPa.<sup>87</sup>

According to the above data of craniomaxillofacial bone, there is still a large gap between the mechanical properties of the degradable metal materials developed at present and those of natural bone, which may cause the movement of the degradable metal materials implanted into craniomaxillofacial bone, especially in the skull (in addition to the high hardness, there is also the presence of intracranial pressure). This unmet clinical need will continue to prompt the development of new alloys. Specifically, it is a great challenge to manufacture efficient tissue-engineered craniomaxillofacial bone grafts. It can reproduce the complex structure of craniomaxillofacial bone, which is the key factor in achieving natural and appropriate mechanical properties, so as to achieve the desired curative effect. Ceramic materials have high hardness, but their toughness is poor, while biomedical degradable materials have relatively high elongation. The combination of the two may further improve the strength of the metal matrix to meet the requirements of mechanical properties of implanted materials in the early stage of bone repair. For example, Bagherifard *et al.*<sup>129</sup> also sandblasted the surface of AZ331, and compared its mechanical properties with those of untreated Mg alloy. It is found that the microhardness of Mg alloy after sandblasting is 133% higher than that of untreated Mg alloy. Jamalian *et al.*<sup>130</sup> also sandblasted the surface of AZ31 and found that the surface grain of AZ31 is refined under high pressure. The microhardness of AZ31 after treatment was increased from 50 Hv to 92 Hv. In addition, ceramics can also protect metal substrates to a certain extent. For example, the protection of Ti substrates by Ca-P under impact force is described in section 2.

## Conflicts of interest

There are no conflicts to declare.

## Acknowledgements

The authors are thankful for the support by the National Natural Science Foundation of China (Grant No. 31700819),



the Young Elite Scientists Sponsorship Program by CAST (YESS, Grant No. 2018QNRC001), and the Fundamental Research Funds for the Central Universities and the Youth Teacher International Exchange & Growth Program (No. QNXM20220020).

## References

- 1 J. Rubio-Palau, A. Prieto-Gundin, A. Cazalla, M. Serrano, G. Fructuoso and F. Ferrandis, Bar, #243, A., Three-dimensional planning in craniomaxillofacial surgery, *Ann. Maxillofac. Surg.*, 2016, **6**(2), 281–286.
- 2 V. Martin and A. Bettencourt, Bone regeneration: Biomaterials as local delivery systems with improved osteoinductive properties, *Mater. Sci. Eng., C*, 2018, **82**, 363–371.
- 3 R. Mazzone, H. D. Netto and F. F. Nascimento, *Enxertos ósseos em implantodontia*, Napoleão, Nova Odessa, 2012.
- 4 A. Nouri and C. Wen, Chapter 3 - Stainless steels in orthopedics, in *Structural Biomaterials*, ed. C. Wen, Woodhead Publishing, 2021, pp. 67–101.
- 5 J. J. Wind, C. Ohaegbulam, F. M. Iwamoto, P. M. Black and J. K. Park, Immediate titanium mesh cranioplasty for treatment of postcraniotomy infections, *World Neurosurg.*, 2013, **79**(1), 207e11–207e13.
- 6 H. K. Lim, Y. J. Choi, W. C. Choi, I. S. Song and U. L. Lee, Reconstruction of maxillofacial bone defects using patient-specific long-lasting titanium implants, *Sci. Rep.*, 2022, **12**(1), 7538.
- 7 S.-H. Byun, H.-K. Lim, S.-M. Lee, H.-E. Kim, S.-M. Kim and J.-H. Lee, Biodegradable Magnesium Alloy (ZK60) with a Poly(l-lactic)-Acid Polymer Coating for Maxillofacial Surgery, *Metals*, 2020, **10**(6), 724.
- 8 J. Nagels, M. Stokdijk and P. M. Rozing, Stress shielding and bone resorption in shoulder arthroplasty, *J. Shoulder Elbow Surg.*, 2003, **12**(1), 35–39.
- 9 H. Pereira, I. Cengiz, F. Maia, F. Bartolomeu, J. M. Oliveira, R. Reis and F. S. Silva, Physicochemical properties and cytocompatibility assessment of non-degradable scaffolds for bone tissue engineering applications, *J. Mech. Behav. Biomed. Mater.*, 2020, **112**, 103997.
- 10 Z. Lekston, J. Drugacz and H. Morawiec, Application of superelastic NiTi wires for mandibular distraction, *Mater. Sci. Eng., A*, 2004, **378**(1–2), 537–541.
- 11 F. Gewalli, J. P. da Silva Guimarães-Ferreira, G. Maltese, U. Örtengren and C. Lauritzen, Expander Elements In Craniofacial Surgery: An Experimental Study In Rabbits, *J. Oral Surg.*, 2001, **35**(2), 149–156.
- 12 L. R. David, F. Gewalli, J. Guimarães-Ferreira, C. Sanger, S. Glazier and L. C. Argenta, Dynamic Spring-Mediated Cranioplasty in a Rabbit Model, *J. Craniofac. Surg.*, 2002, **13**(6), 794–801.
- 13 C. Lauritzen, Y. Sugawara, O. Kocabalkan and R. Olsson, Spring mediated dynamic craniofacial reshaping: Case report, *Scand. J. Plast. Reconstr. Surg. Hand Surg.*, 1998, **32**(3), 331–338.
- 14 E. S. J. Breme and G. Paulus, Commercially pure titanium Steinhäuser plate-screw system for maxillofacial surgery, *Biomaterials*, 1988, **9**(4), 310–313.
- 15 J. J. Kuttenberger and N. Hardt, Long-term results following reconstruction of craniofacial defects with titanium micro-mesh systems, *J. Cranio-Maxillofac. Surg.*, 2001, **29**(2), 75–81.
- 16 S. Lewin, J. Åberg, D. Neuhaus, H. Engqvist, S. J. Ferguson, C. Öhman-Mägi, B. Helgason and C. Persson, Mechanical behaviour of composite calcium phosphate–titanium cranial implants: Effects of loading rate and design, *J. Mech. Behav. Biomed. Mater.*, 2020, **104**, 103701.
- 17 S. Lewin, I. Fleps, J. Åberg, S. J. Ferguson, H. Engqvist, C. Öhman-Mägi, B. Helgason and C. Persson, Additively manufactured mesh-type titanium structures for cranial implants: E-PBF vs. L-PBF, *Mater. Des.*, 2021, **197**, 109207.
- 18 A. Bernhardt, J. Schneider, A. Schroeder, K. Papadopoulos, E. Lopez, F. Bruckner and U. Botzenhart, Surface conditioning of additively manufactured titanium implants and its influence on materials properties and in vitro biocompatibility, *Mater. Sci. Eng., C*, 2021, **119**, 111631.
- 19 S. Ponader, C. von Wilmowsky, M. Widenmayer, R. Lutz, P. Heintz, C. Korner, R. F. Singer, E. Nkenke, F. W. Neukam and K. A. Schlegel, In vivo performance of selective electron beam-melted Ti-6Al-4 V structures, *J. Biomed. Mater. Res., Part A*, 2010, **92**(1), 56–62.
- 20 J.-S. K. Jung-Hwan Lee, S.-K. Moon, S.-H. Uhm, B.-H. Choi, U.-H. Joo, K.-M. Kim and K.-N. Kim, Titanium-Silver Alloy Miniplates for Mandibular Fixation: In Vitro and In Vivo Study, *J. Oral Maxillofac. Surg.*, 2016, **74**(8), 1622e1–1622.e12.
- 21 H. Z. Morawiec, Z. H. Lekston, K. F. Kobus, M. C. Wegrzyn and J. T. Drugacz, Superelastic NiTi springs for corrective skull operations in children with craniosynostosis, *J. Mater. Sci. Mater. Med.*, 2007, **18**(9), 1791–1798.
- 22 S. Simske and R. Sachdeva, Cranial bone apposition and ingrowth in a porous nickel–titanium implant, *J. Biomed. Mater. Res.*, 1995, **29**(4), 527–533.
- 23 R. A. Ayers, S. J. Simske, T. Bateman, A. Petkus, R. Sachdeva and V. Gyunter, Effect of nitinol implant porosity on cranial bone ingrowth and apposition after 6 weeks, *J. Biomed. Mater. Res.*, 1999, **45**(1), 42–47.
- 24 S. R. Durham, J. G. McComb and M. L. Levy, Correction of large (> 25 cm<sup>2</sup>) cranial defects with “reinforced” hydroxyapatite cement: technique and complications, *Neurosurgery*, 2003, **52**(4), 842–845.
- 25 R. H. Pudenz, The repair of cranial defects with tantalum: an experimental study, *J. Am. Med. Assoc.*, 1943, **121**, 478–481.
- 26 L. D. Zardiackas, D. E. Parsell, L. D. Dillon, D. W. Mitchell, L. A. Nunnery and R. Poggie, Structure, metallurgy, and mechanical properties of a porous tantalum foam, *J. Biomed. Mater. Res.*, 2001, **58**(2), 180–187.
- 27 R. Kummoona, Temporomandibular joint reconstruction with a 2-part chrome-cobalt prosthesis, chondro-osseous

- graft, and silastic: clinical and experimental studies, *J. Craniofac. Surg.*, 2009, **20**(6), 2125–2135.
- 28 C. S. Beck, Repair of defects in skull by ready made vitalium plates, *J. Am. Med. Assoc.*, 1942, **118**(10), 798–799.
- 29 H. G. Luhr, Entwicklung der modernen Osteosynthese, *Mund-, Kiefer- und Gesichtschirurgie*, 2000, **4**(1), S084–S090.
- 30 B. Spiessl, G. Schargus and K. Schroll, Die stabile Osteosynthese bei Frakturen des unbezahnten Unterkiefers, *Schweiz Mschr Zahnheilk.*, 1971, **81**, 39.
- 31 R. Becker and E. Machtens, Druckplattenosteosynthese zur Frakturbehandlung und bei orthopädischchirurgischen Maßnahmen am Gesichtsschädel, *Osteo News*, 1973, **19**.
- 32 H. Niederdellmann and W. Schilli, Zur Plattenosteosynthese bei unterkieferfrakturen, *Dtsch zahnrztl Z*, 1973, **28**, 638.
- 33 F. Michelet and A. Moll, Traitements chirurgicaux des fractures du corps mandibulaire sans blocage, par plaques vissées miniatures insérées par voie endobuccale, *Rev. Odontostomatol.*, 1971, **29**(2), 87–105.
- 34 Q. Zhang, W. Wu, C. Qian, W. Xiao, H. Zhu, J. Guo, Z. Meng, J. Zhu, Z. Ge and W. Cui, Advanced biomaterials for repairing and reconstruction of mandibular defects, *Mater. Sci. Eng., C*, 2019, **103**, 109858.
- 35 S. W. On, S. W. Cho, S. H. Byun and B. E. Yang, Bioabsorbable Osteofixation Materials for Maxillofacial Bone Surgery: A Review on Polymers and Magnesium-Based Materials, *Biomedicines*, 2020, **8**(9), 300.
- 36 Q. Zhang, W. Wu, C. Qian, W. Xiao, H. Zhu, J. Guo, Z. Meng, J. Zhu, Z. Ge and W. Cui, Advanced biomaterials for repairing and reconstruction of mandibular defects, *Mater. Sci. Eng., C*, 2019, **103**, 109858.
- 37 K. Eichner, Messung der Krafte bei Kauvorgängen, *Dtsch*, 1963, **18**, 915–924.
- 38 H. Schwickerath, Kaukraft. Kaudruck. Belastbarkeit, *Dtsch*, 1976, **31**, 870–873.
- 39 J. E. G. González and J. C. Mirza-Rosca, Study of the corrosion behavior of titanium and some of its alloys for biomedical and dental implant applications, *J. Electroanal. Chem.*, 1999, **471**(2), 109–115.
- 40 M. Crapper and D. Boni, Aluminium in human brain disease: an overview, *Neurotoxicology*, 1980, **1**(4), 3–16.
- 41 D. P. Perl and A. R. Brody, Alzheimer's disease: X-ray spectrometric evidence of aluminum accumulation in neurofibrillary tangle-bearing neurons, *Science*, 1980, **208**(4441), 297–299.
- 42 P. G. Laing, A. B. Ferguson Jr. and E. S. Hodge, Tissue reaction in rabbit muscle exposed to metallic implants, *J. Biomed. Mater. Res.*, 1967, **1**(1), 135–149.
- 43 J. Black, Biologic performance of tantalum, *Clin. Mater.*, 1994, **16**(3), 167–173.
- 44 E. P. Strecker, D. Liermann, K. H. Barth, H. Wolf, N. Freudenberg, G. Berg, M. Westphal, P. Tsikuras, M. Savin and B. Schneider, Expandable tubular stents for treatment of arterial occlusive diseases: experimental and clinical results. Work in progress, *Radiology*, 1990, **175**(1), 97–102.
- 45 P. A. Ribeiro, R. Gallo, J. Antonius, L. Mimish, R. Sriram, S. Bianchi and C. G. Duran, A new expandable intracoronary tantalum (Strecker) stent: early experimental results and follow-up to twelve months, *Am. Heart J.*, 1993, **125**(2), 501–510.
- 46 C. Suba, M. Lakatos-Varsányi, A. Mikó, L. Kovács, N. Velich, B. Kádár and G. Szabó, Study of the electrochemical behaviour of Ti osteosynthesis plates used in maxillofacial surgery, *Mater. Sci. Eng., A*, 2007, **447**(1–2), 347–354.
- 47 D. Krupa, J. Baszkiewicz, B. Rajchel, A. Barcz, J. W. Sobczak and A. Biliński, Effect of sodium-ion implantation on the corrosion resistance and bioactivity of titanium, *Vacuum*, 2005, **78**(2), 161–166.
- 48 M. Pankuch, R. Bell and C. A. Melendres, Composition and structure of the anodic films on titanium in aqueous solutions, *Electrochim. Acta*, 1993, **38**(18), 2777–2779.
- 49 C. E. B. Marino and L. H. Mascaró, EIS characterization of a Ti-dental implant in artificial saliva media: dissolution process of the oxide barrier, *J. Electroanal. Chem.*, 2004, **568**, 115–120.
- 50 G. L. Burke, The corrosion of metals in tissues; and an introduction to tantalum, *Can. Med. Assoc. J.*, 1940, **43**(2), 125.
- 51 O. H. Fulcher, Tantalum as a metallic implant to repair cranial defects: a preliminary report, *J. Am. Med. Assoc.*, 1943, **121**(12), 931–933.
- 52 P. Flanigan, V. R. Kshetry and E. C. Benzel, World War II, tantalum, and the evolution of modern cranioplasty technique, *Neurosurg Focus*, 2014, **36**(4), E22.
- 53 S. Arabnejad, R. Burnett Johnston, J. A. Pura, B. Singh, M. Tanzer and D. Pasini, High-strength porous biomaterials for bone replacement: A strategy to assess the interplay between cell morphology, mechanical properties, bone ingrowth and manufacturing constraints, *Acta Biomater.*, 2016, **30**, 345–356.
- 54 F. S. L. Bobbert and A. A. Zadpoor, Effects of bone substitute architecture and surface properties on cell response, angiogenesis, and structure of new bone, *J. Mater. Chem. B*, 2017, **5**(31), 6175–6192.
- 55 S. Van Bael, Y. C. Chai, S. Truscillo, M. Moesen, G. Kerckhofs, H. Van Oosterwyck, J. P. Kruth and J. Schrooten, The effect of pore geometry on the in vitro biological behavior of human periosteum-derived cells seeded on selective laser-melted Ti6Al4 V bone scaffolds, *Acta Biomater.*, 2012, **8**(7), 2824–2834.
- 56 J. Van der Stok, O. P. Van der Jagt, S. Amin Yavari, M. F. P. De Haas, J. H. Waarsing, H. Jahr, E. M. M. Van Lieshout, P. Patka, J. A. N. Verhaar, A. A. Zadpoor and H. Weinans, Selective laser melting-produced porous titanium scaffolds regenerate bone in critical size cortical bone defects, *J. Orthop. Res.*, 2013, **31**(5), 792–799.
- 57 T. Miyazaki, H.-M. Kim, T. Kokubo, C. Ohtsuki, H. Kato and T. Nakamura, Mechanism of bonelike apatite formation on bioactive tantalum metal in a simulated body fluid, *Biomaterials*, 2002, **23**(3), 827–832.

- 58 Y. Zhang, P. B. Ahn, D. C. Fitzpatrick, A. D. Heiner, R. A. Poggie and T. D. Brown, Interfacial frictional behavior: cancellous bone, cortical bone, and a novel porous tantalum biomaterial, *J. Musculoskeletal Res.*, 1999, **03**(04), 245–251.
- 59 B. R. Levine, S. Sporer, R. A. Poggie, C. J. Della Valle and J. J. Jacobs, Experimental and clinical performance of porous tantalum in orthopedic surgery, *Biomaterials*, 2006, **27**(27), 4671–4681.
- 60 H. P. Tang, K. Yang, L. Jia, W. W. He, L. Yang and X. Z. Zhang, Tantalum Bone Implants Printed by Selective Electron Beam Manufacturing (SEBM) and Their Clinical Applications, *JOM*, 2020, **72**(3), 1016–1021.
- 61 G. Huang, S. T. Pan and J. X. Qiu, The Clinical Application of Porous Tantalum and Its New Development for Bone Tissue Engineering, *Materials*, 2021, **14**(10), 2647.
- 62 F. W. Geib, Vitallium skull plates, *J. Am. Med. Assoc.*, 1941, **117**(1), 8–12.
- 63 C. S. Venable, W. G. Stuck and A. Beach, The effects on bone of the presence of metals, based on electrolysis, *Ann. Surg.*, 1937, **103**, 917.
- 64 D. Meslemani and R. M. Kellman, Recent advances in fixation of the craniomaxillofacial skeleton, *Curr. Opin. Otolaryngol. Head Neck Surg.*, 2012, **20**(4), 304–309.
- 65 X. Wang, X. Shao, T. Dai, F. Xu, J. G. Zhou, G. Qu, L. Tian, B. Liu and Y. Liu, In vivo study of the efficacy, biosafety, and degradation of a zinc alloy osteosynthesis system, *Acta Biomater.*, 2019, **92**, 351–361.
- 66 H. Guo, D. Xia, Y. Zheng, Y. Zhu, Y. Liu and Y. Zhou, A pure zinc membrane with degradability and osteogenesis promotion for guided bone regeneration: In vitro and in vivo studies, *Acta Biomater.*, 2020, **106**, 396–409.
- 67 W. Zhang, P. Li, G. Shen, X. Mo, C. Zhou, D. Alexander, F. Rupp, J. Geis-Gerstorfer, H. Zhang and G. Wan, Appropriately adapted properties of hot-extruded Zn-0.5Cu-xFe alloys aimed for biodegradable guided bone regeneration membrane application, *Bioact. Mater.*, 2021, **6**(4), 975–989.
- 68 P. Li, W. Zhang, J. Dai, A. B. Xepapadeas, E. Schweizer, D. Alexander, L. Scheideler, C. Zhou, H. Zhang, G. Wan and J. Geis-Gerstorfer, Investigation of zinc-copper alloys as potential materials for craniomaxillofacial osteosynthesis implants, *Mater. Sci. Eng., C*, 2019, **103**, 109826.
- 69 X. Qu, H. Yang, B. Jia, M. Wang, B. Yue, Y. Zheng and K. Dai, Zinc alloy-based bone internal fixation screw with antibacterial and anti-osteolytic properties, *Bioact. Mater.*, 2021, **6**(12), 4607–4624.
- 70 D. T. Chou, D. Wells, D. Hong, B. Lee, H. Kuhn and P. N. Kumta, Novel processing of iron-manganese alloy-based biomaterials by inkjet 3-D printing, *Acta Biomater.*, 2013, **9**(10), 8593–8603.
- 71 D. Hong, D. T. Chou, O. I. Velikokhatnyi, A. Roy, B. Lee, I. Swink, I. Issaev, H. A. Kuhn and P. N. Kumta, Binder-jetting 3D printing and alloy development of new biodegradable Fe-Mn-Ca/Mg alloys, *Acta Biomater.*, 2016, **45**, 375–386.
- 72 S. E. Henderson, K. Verdelis, S. Maiti, S. Pal, W. L. Chung, D. T. Chou, P. N. Kumta and A. J. Almarza, Magnesium alloys as a biomaterial for degradable craniofacial screws, *Acta Biomater.*, 2014, **10**(5), 2323–2332.
- 73 B. Schaller, N. Saulacic, T. Imwinkelried, S. Beck, E. W. Liu, J. Gralla, K. Nakahara, W. Hofstetter and T. Iizuka, In vivo degradation of magnesium plate/screw osteosynthesis implant systems: Soft and hard tissue response in a calvarial model in miniature pigs, *J. Craniomaxillofac. Surg.*, 2016, **44**(3), 309–317.
- 74 S. H. Byun, H. K. Lim, K. H. Cheon, S. M. Lee, H. E. Kim and J. H. Lee, Biodegradable magnesium alloy (WE43) in bone-fixation plate and screw, *J. Biomed. Mater. Res., Part B*, 2020, **108**(6), 2505–2512.
- 75 A. Torroni, C. Xiang, L. Witek, E. D. Rodriguez, P. G. Coelho and N. Gupta, Biocompatibility and degradation properties of WE43 Mg alloys with and without heat treatment: In vivo evaluation and comparison in a cranial bone sheep model, *J. Craniomaxillofac. Surg.*, 2017, **45**(12), 2075–2083.
- 76 J. Kubásek, D. Dvorský, J. Šedý, Š. Msallamová, J. Levorová, R. Foltán and D. Vojtěch, The Fundamental Comparison of Zn-2Mg and Mg-4Y-3RE Alloys as a Perspective Biodegradable Materials, *Materials*, 2019, **12**(22), 3745.
- 77 M. Zhao, G. Liu, Y. Li, X. Yu, S. Yuan, Z. Nie, J. Wang, J. Han, C. Tan, C. Guo and D. Behavior, Transport Mechanism and Osteogenic Activity of Mg-Zn-RE Alloy Membranes in Critical-Sized Rat Calvarial Defects, *Coatings*, 2020, **10**(5), 496.
- 78 P. Aisen, C. Enns and M. Wessling-Resnick, Chemistry and biology of eukaryotic iron metabolism, *Int. J. Biochem. Cell Biol.*, 2001, **33**(10), 940–959.
- 79 B. J. Crielaard, T. Lammers and S. Rivella, Targeting iron metabolism in drug discovery and delivery, *Nat. Rev. Drug Discovery*, 2017, **16**(6), 400–423.
- 80 D. T. Reilly and A. H. Burstein, The Mechanical Properties of Cortical Bone, *JBJS J.*, 1974, **56**(5), 1001–1022.
- 81 K. Roemhild, F. von Maltzahn, R. Weiskirchen, R. Knuchel, S. von Stillfried and T. Lammers, Iron metabolism: pathophysiology and pharmacology, *Trends Pharmacol. Sci.*, 2021, **42**(8), 640–656.
- 82 F.-J. Kretz and J. Schäffer, *Anästhesie, Intensivmedizin, Notfallmedizin, Schmerztherapie*, Springer-Verlag, 2008.
- 83 M. J. Müller, J. Westenhöfer, A. Bosy-Westphal, C. Löser and O. Selberg, *Ernährungsmedizinische Untersuchungen, Ernährungsmedizinische Praxis*, Springer, 2007, pp. 1–196.
- 84 D. Fiorentini, C. Cappadone, G. Farruggia and C. Prata, Magnesium: Biochemistry, Nutrition, Detection, and Social Impact of Diseases Linked to Its Deficiency, *Nutrients*, 2021, **13**(4), 1136.
- 85 W. Qiao, K. H. M. Wong, J. Shen, W. Wang, J. Wu, J. Li, Z. Lin, Z. Chen, J. P. Matinlinna, Y. Zheng, S. Wu, X. Liu, K. P. Lai, Z. Chen, Y. W. Lam, K. M. C. Cheung and K. W. K. Yeung, TRPM7 kinase-mediated immunomodulation in macrophage plays a central role in magnesium

- ion-induced bone regeneration, *Nat. Commun.*, 2021, **12**(1), 2885.
- 86 J.-Y. Lee, J.-W. Lee, K.-M. Pang, H.-E. Kim, S.-M. Kim and J.-H. Lee, Biomechanical Evaluation of Magnesium-Based Resorbable Metallic Screw System in a Bilateral Sagittal Split Ramus Osteotomy Model Using Three-Dimensional Finite Element Analysis, *J. Oral Maxillofac. Surg.*, 2014, **72**(2), 402.
- 87 J. H. Lee, H. S. Han, Y. C. Kim, J. Y. Lee and B. K. Lee, Stability of biodegradable metal (Mg-Ca-Zn alloy) screws compared with absorbable polymer and titanium screws for sagittal split ramus osteotomy of the mandible using the finite element analysis model, *J. Craniomaxillofac. Surg.*, 2017, **45**(10), 1639–1646.
- 88 H. Leonhardt, A. Franke, N. M. H. McLeod, G. Lauer and A. Nowak, Fixation of fractures of the condylar head of the mandible with a new magnesium-alloy biodegradable cannulated headless bone screw, *Br. J. Oral Maxillofac. Surg.*, 2017, **55**(6), 623–625.
- 89 Y. Chen, S.-H. Ye, H. Sato, Y. Zhu, V. Shanov, T. Tiasha, A. D'Amore, S. Luketich, G. Wan and W. R. Wagner, Hybrid scaffolds of Mg alloy mesh reinforced polymer/extracellular matrix composite for critical-sized calvarial defect reconstruction, *J. Tissue Eng. Regener. Med.*, 2018, **12**(6), 1374–1388.
- 90 Y. Guo, W. Liu, S. Ma, J. Wang, J. Zou, Z. Liu, J. Zhao and Y. Zhou, A preliminary study for novel use of two Mg alloys (WE43 and Mg3Gd), *J. Mater. Sci.: Mater. Med.*, 2016, **27**(5), 82.
- 91 Z. Zhen, T.-f. Xi and Y.-f. Zheng, A review on in vitro corrosion performance test of biodegradable metallic materials, *Trans. Nonferrous Met. Soc. China*, 2013, **23**(8), 2283–2293.
- 92 J. Zhang, H. Li, W. Wang, H. Huang, J. Pei, H. Qu, G. Yuan and Y. Li, The degradation and transport mechanism of a Mg-Nd-Zn-Zr stent in rabbit common carotid artery: A 20-month study, *Acta Biomater.*, 2018, **69**, 372–384.
- 93 T. Kambe, T. Tsuji, A. Hashimoto and N. Itsumura, The Physiological, Biochemical, and Molecular Roles of Zinc Transporters in Zinc Homeostasis and Metabolism, *Physiol. Rev.*, 2015, **95**(3), 749–784.
- 94 P. J. Little, R. Bhattacharya, A. E. Moreyra and I. L. Korichneva, Zinc and cardiovascular disease, *Nutrition*, 2010, **26**(11), 1050–1057.
- 95 M. Dermience, G. Lognay, F. Mathieu and P. Goyens, Effects of thirty elements on bone metabolism, *J. Trace Elem. Med. Biol.*, 2015, **32**, 86–106.
- 96 H.-J. Seo, Y.-E. Cho, T. Kim, H.-I. Shin and I.-S. Kwun, Zinc may increase bone formation through stimulating cell proliferation, alkaline phosphatase activity and collagen synthesis in osteoblastic MC3T3-E1 cells, *Nutr. Res. Pract.*, 2010, **4**(5), 356–361.
- 97 C. J. Frederickson, J.-Y. Koh and A. I. Bush, The neurobiology of zinc in health and disease, *Nat. Rev. Neurosci.*, 2005, **6**(6), 449–462.
- 98 Y. Qiao, W. Zhang, P. Tian, F. Meng, H. Zhu, X. Jiang, X. Liu and P. K. Chu, Stimulation of bone growth following zinc incorporation into biomaterials, *Biomaterials*, 2014, **35**(25), 6882–6897.
- 99 Z. Xu, S. Kim and J. Huh, Zinc plays a critical role in the cardioprotective effect of postconditioning by enhancing the activation of the RISK pathway in rat hearts, *J. Mol. Cell. Cardiol.*, 2014, **66**, 12–17.
- 100 D. Zhu, Y. Su, Y. Zheng, B. Fu, L. Tang and Y. X. Qin, Zinc regulates vascular endothelial cell activity through zinc-sensing receptor ZnR/GPR39, *Am. J. Physiol.: Cell Physiol.*, 2018, **314**(4), C404–C414.
- 101 Y. Su, K. Wang, J. Gao, Y. Yang, Y.-X. Qin, Y. Zheng and D. Zhu, Enhanced cytocompatibility and antibacterial property of zinc phosphate coating on biodegradable zinc materials, *Acta Biomater.*, 2019, **98**, 174–185.
- 102 Y. Su, H. Yang, J. Gao, Y.-X. Qin, Y. Zheng and D. Zhu, Interfacial Zinc Phosphate is the Key to Controlling Biocompatibility of Metallic Zinc Implants, *Adv. Sci.*, 2019, **6**(14), 1900112.
- 103 H. Guo, D. Xia, Y. Zheng, Y. Zhu, Y. Liu and Y. Zhou, A pure zinc membrane with degradability and osteogenesis promotion for guided bone regeneration: In vitro and in vivo studies, *Acta Biomater.*, 2020, **106**, 396–409.
- 104 L. Heiden, More Predictive Cell-Based Assays, *Genet. Eng. Biotechnol. News*, 2015, **35**(2), 1, 10, 12–13.
- 105 J. Venezuela and M. S. Dargusch, The influence of alloying and fabrication techniques on the mechanical properties, biodegradability and biocompatibility of zinc: A comprehensive review, *Acta Biomater.*, 2019, **87**, 1–40.
- 106 H. Yang, B. Jia, Z. Zhang, X. Qu, G. Li, W. Lin, D. Zhu, K. Dai and Y. Zheng, Alloying design of biodegradable zinc as promising bone implants for load-bearing applications, *Nat. Commun.*, 2020, **11**(1), 401.
- 107 W. Pachla, S. Przybysz, A. Jarzebska, M. Bieda, K. Sztwiertnia, M. Kulczyk and J. Skiba, Structural and mechanical aspects of hypoeutectic Zn-Mg binary alloys for biodegradable vascular stent applications, *Bioact. Mater.*, 2021, **6**(1), 26–44.
- 108 X. Wei, W. Zhou, Z. Tang, H. Wu, Y. Liu, H. Dong, N. Wang, H. Huang, S. Bao, L. Shi, X. Li, Y. Zheng and Z. Guo, Magnesium surface-activated 3D printed porous PEEK scaffolds for in vivo osseointegration by promoting angiogenesis and osteogenesis, *Bioact. Mater.*, 2023, **20**, 16–28.
- 109 G. I. Benic and C. H. F. Hämmerle, Horizontal bone augmentation by means of guided bone regeneration, *Periodontology*, 2014, **66**(1), 13–40.
- 110 H. Saijo, U.-i. Chung, K. Igawa, Y. Mori, D. Chikazu, M. Iino and T. Takato, Clinical application of artificial bone in the maxillofacial region, *J. Artif. Organs*, 2008, **11**(4), 171.
- 111 S. Aydin, B. Kucukyuruk, B. Abuzayed, S. Aydin and G. Z. Sanus, Cranioplasty: Review of materials and techniques, *J. Neurosci. Rural Pract.*, 2011, **2**(2), 162–167.
- 112 D. P. Blake, The Use of Synthetics in Cranioplasty: A Clinical Review, *Mil. Med.*, 1994, **159**(6), 466–469.
- 113 C. R. Arciola, D. Campoccia, G. D. Ehrlich and L. Montanaro, Biofilm-Based Implant Infections in



- Orthopaedics, in *Biofilm-based Healthcare-associated Infections: Volume I*, ed. G. Donelli, Springer International Publishing, Cham, 2015, pp. 29–46.
- 114 C. R. Arciola, D. Campoccia and L. Montanaro, Implant infections: adhesion, biofilm formation and immune evasion, *Nat. Rev. Microbiol.*, 2018, **16**(7), 397–409.
- 115 L. Montanaro, P. Speziale, D. Campoccia, S. Ravaoli, I. Cangini, G. Pietrocola, S. Giannini and C. R. Arciola, Scenery of Staphylococcus implant infections in orthopaedics, *Future Microbiol.*, 2011, **6**(11), 1329–1349.
- 116 S. Amin Yavari, M. Croes, B. Akhavan, F. Jahanmard, C. C. Eigenhuis, S. Dadbakhsh, H. C. Vogely, M. M. Bilek, A. C. Fluit, C. H. E. Boel, B. C. H. van der Wal, T. Vermonden, H. Weinans and A. A. Zadpoor, Layer by layer coating for bio-functionalization of additively manufactured meta-biomaterials, *Addit. Manuf.*, 2020, **32**, 100991.
- 117 H. J. Busscher, H. C. v. d. Mei, G. Subbiahdoss, P. C. Jutte, J. J. A. M. v. d. Dungen, S. A. J. Zaat, M. J. Schultz and D. W. Grainger, Biomaterial-Associated Infection: Locating the Finish Line in the Race for the Surface, *Sci. Transl. Med.*, 2012, **4**(153), 153rv10–153rv10.
- 118 T. Wang, J. Bai, M. Lu, C. Huang, D. Geng, G. Chen, L. Wang, J. Qi, W. Cui and L. Deng, Engineering immunomodulatory and osteoinductive implant surfaces via mussel adhesion-mediated ion coordination and molecular clicking, *Nat. Commun.*, 2022, **13**(1), 160.
- 119 M. Mack, Inflammation and fibrosis, *Matrix Biol.*, 2018, **68–69**, 106–121.
- 120 J. W. Godwin, A. R. Pinto and N. A. Rosenthal, Macrophages are required for adult salamander limb regeneration, *Proc. Natl. Acad. Sci. U. S. A.*, 2013, **110**(23), 9415–9420.
- 121 Y. Zhao, L. Bai, Y. Zhang, R. Yao, Y. Sun, R. Hang, X. Chen, H. Wang, X. Yao, Y. Xiao and R. Hang, Type I collagen decorated nanoporous network on titanium implant surface promotes osseointegration through mediating immunomodulation, angiogenesis, and osteogenesis, *Biomaterials*, 2022, **288**, 121684.
- 122 P. Huang, J. Xu, L. Xie, G. Gao, S. Chen, Z. Gong, X. Lao, Z. Shan, J. Shi, Z. Zhou, Z. Chen, Y. Cao, Y. Wang and Z. Chen, Improving hard metal implant and soft tissue integration by modulating the “inflammatory-fibrous complex” response, *Bioact. Mater.*, 2023, **20**, 42–52.
- 123 C. Yu, L. Chen, W. Zhou, L. Hu, X. Xie, Z. Lin, A. C. Panayi, X. Zhan, R. Tao, B. Mi and G. Liu, Injectable Bacteria-Sensitive Hydrogel Promotes Repair of Infected Fractures via Sustained Release of miRNA Antagonist, *ACS Appl. Mater. Interfaces*, 2022, **14**, 34427–34442.
- 124 B. J. Igo, P. S. Cottler, J. S. Black and M. B. Panzer, The mechanical and microstructural properties of the pediatric skull, *J. Mech. Behav. Biomed. Mater.*, 2021, **120**, 104578.
- 125 B. C. a. S. S. Margulies, Material Properties of Human Infant Skull and Suture at High Rates, *J. Neurotrauma*, 2006, **23**(8), 1222–1232.
- 126 M. T. Davis, A. M. Loyd, H. Y. Shen, M. H. Mulroy, R. W. Nightingale, B. S. Myers and C. D. Bass, The mechanical and morphological properties of 6 years-old cranial bone, *J. Biomech.*, 2012, **45**(15), 2493–2498.
- 127 J. H. C. Lee, B. Ondruschka, L. Falland-Cheung, M. Scholze, N. Hammer, D. C. Tong and J. N. Waddell, An Investigation on the Correlation between the Mechanical Properties of Human Skull Bone, Its Geometry, Microarchitectural Properties, and Water Content, *J. Healthcare Eng.*, 2019, **2019**, 6515797.
- 128 S. Kanala, S. Gudipalli, P. Perumalla, K. Jagalanki, P. V. Polamarasetty, S. Guntaka, A. Gudala and R. P. Boyapati, Aetiology, prevalence, fracture site and management of maxillofacial trauma, *Ann. R. Coll. Surg. Engl.*, 2021, **103**(1), 18–22.
- 129 S. Bagherifard, D. J. Hickey, S. Fintová, F. Pastorek, I. Fernandez-Pariente, M. Bandini, T. J. Webster and M. Guagliano, Effects of nanofeatures induced by severe shot peening (SSP) on mechanical, corrosion and cytocompatibility properties of magnesium alloy AZ31, *Acta Biomater.*, 2018, **66**, 93–108.
- 130 M. Jamalain and D. P. Field, Effects of shot peening parameters on gradient microstructure and mechanical properties of TRC AZ31, *Mater. Charact.*, 2019, **148**, 9–16.