

Cite this: *Anal. Methods*, 2015, 7, 211Towards on-site testing of *Phytophthora* species†Lydia Schwenkbier,<sup>abc</sup> Sibyll Pollok,<sup>abc</sup> Stephan König,<sup>d</sup> Matthias Urban,<sup>a</sup>  
Sabine Werres,<sup>\*d</sup> Dana Cialla-May,<sup>abc</sup> Karina Weber<sup>\*abc</sup> and Jürgen Popp<sup>abc</sup>

Rapid detection and accurate identification of plant pathogens in the field is an ongoing challenge. In this study, we report for the first time on the development of a helicase-dependent isothermal amplification (HDA) in combination with on-chip hybridization for the detection of selected *Phytophthora* species. The HDA approach allows efficient amplification of the yeast GTP-binding protein (Ypt1) target gene region at one constant temperature in a miniaturized heating device. The assay's specificity was determined by on-chip DNA hybridization and subsequent silver nanoparticle deposition. The silver deposits serve as stable endpoint signals that enable the visual as well as the electrical readout. Our promising results point to the direction of a near future on-site application of the combined techniques for a reliable detection of *Phytophthora* species.

Received 26th September 2014  
Accepted 10th November 2014

DOI: 10.1039/c4ay02287d

www.rsc.org/methods

## Introduction

Fungal-like organisms of the genus *Phytophthora* belong to some of the world's most devastating plant pathogens.<sup>1</sup> *Phytophthora fragariae* (*P. fragariae*), *Phytophthora kernoviae* (*P. kernoviae*), *Phytophthora ramorum* (*P. ramorum*) and *Phytophthora rubi* (*P. rubi*) are on the EPPO A2 list of the European and Mediterranean Plant Protection Organization (EPPO) (<https://www.eppo.int/QUARANTINE/listA2.htm>). These species are sufficiently dangerous to recommend a regulation as quarantine organisms. The pathogen *P. ramorum*<sup>2</sup> for instance is the causal agent of Sudden Oak Death in the forests of the west coast of the United States (<http://www.suddenoakdeath.org>) and *Larix* decline in the United Kingdom.<sup>3</sup> A wide range of trees, shrubs and plants in natural and landscaped environments as well as in nursery industries can be affected by *Phytophthora* species. Therefore the prevention of the worldwide spread of these fungus-like plant pathogens due to an increasing trade between countries is of great importance. Thus, suitable detection methods are mandatory in order to facilitate effective screening to control and eradicate *Phytophthora*. In particular,

rapid and reliable approaches which are inexpensive plus field applicable are needed to significantly minimize the delay between sampling and diagnosis.

Common techniques for routine diagnosis of *Phytophthora* on the species level rely upon molecular biological, immunological or microbiological approaches.<sup>4–6</sup> The last two methods are time consuming, laborious, and require extensive knowledge of classical taxonomy. An accurate discrimination between various *Phytophthora* species was successfully realized by the polymerase chain reaction (PCR).<sup>5</sup> Nevertheless, their application in the field is hampered due to the need for thermal cycling instruments.<sup>7–12</sup>

An important step towards on-site detection of regulated *Phytophthora* species is provided by isothermal nucleic acid amplification techniques.<sup>7–12</sup> Recently, several articles highlighted the loop-mediated amplification (LAMP) for a DNA-based *Phytophthora* specification.<sup>13–15</sup> Here the Bst DNA polymerase amplifies the target gene region under conditions that omit the use of a thermal cycler.<sup>16–18</sup> Although LAMP allows a convenient usage and is highly sensitive, primer designing is arduous and requires dedicated software. Moreover, an initial heat denaturation of the double-stranded template DNA prior to the isothermal amplification is often mandatory. Thus, LAMP needs a two temperature profile and cannot be claimed as really isothermal.

A further improvement of isothermal amplification that mimics *in vivo* DNA replication is introduced by the helicase-dependent amplification (HDA).<sup>19–22</sup> Similar to the common PCR, the target gene region, which is enclosed by two primers, is selectively amplified. In more detail, a DNA helicase separates the double-stranded DNA and the resulting strands are immediately coated by single-stranded binding proteins (SSBs). Two sequence-specific primers bind to the template and get extended by the DNA polymerase. The newly synthesized DNA

<sup>a</sup>Leibniz Institute of Photonic Technology Jena (IPHT), Jenaer BioChip Initiative, Albert-Einstein-Straße 9, 07745 Jena, Germany. E-mail: karina.weber@ipht-jena.de; Fax: +49 3641 206399; Tel: +49 3641 206309/948390

<sup>b</sup>Friedrich Schiller University Jena, Institute of Physical Chemistry and Abbe Center of Photonics, Helmholtzweg 4, 07745 Jena, Germany

<sup>c</sup>InfectoGnostics Forschungscampus Jena, Zentrum für Angewandte Forschung, Philosophenweg 7, 07743 Jena, Germany

<sup>d</sup>Julius Kühn Institute-Federal Research Centre for Cultivated Plants (JKI), Institute for Plant Protection in Horticulture and Forests, Messeweg 11/12, 38104 Braunschweig, Germany. E-mail: sabine.werres@jki.bund.de; Fax: +49 531 2993009; Tel: +49 531 2994407

† Electronic supplementary information (ESI) available. See DOI: 10.1039/c4ay02287d

strands serve as matrices for a new amplification cycle which allows an exponential amplification.<sup>19</sup> This approach possesses several advantages compared to other isothermal amplification methods. Firstly, a helicase unwinds and separates the double-stranded DNA; a prior heat denaturation step and subsequent thermal cycling are unnecessary. Thus, HDA can be referred to as a real isothermal technique with performance at one constant temperature for the entire process. Secondly, only one specific primer pair has to be designed. Last but not least kits are commercially available that enable a more convenient usage.<sup>22,23</sup> These depicted properties offer promising potential towards the development of on-site detection systems for plant pathogens.<sup>19,24–26</sup> Optimized HDA protocols have already been adapted for the detection of several bacterial pathogens like *Clostridium difficile*,<sup>27,28</sup> *Staphylococcus aureus*,<sup>23,29,30</sup> *Neisseria gonorrhoeae*,<sup>20,23,31,32</sup> *Mycobacterium tuberculosis*,<sup>33,34</sup> as well as different viruses.<sup>35–39</sup>

Within this context, we adapted the HDA approach for *Phytophthora* pathogen detection for the first time. The DNA for this isothermal amplification was isolated from cultivated *Phytophthora* species or infected *rhododendron* leaves. A subsequent precise specification of the phylogenetically closely related *Phytophthora* species was realized by on-chip DNA hybridization.

## Experimental

### *Phytophthora* cultivation, plant samples and DNA extraction

Isolates of *P. ramorum* (BBA9/95), *P. fragariae* (BBA L1) and *P. kernoviae* (JKI 080-09-00-00-00-03) (culture collection of the Julius Kuehn Institute, Braunschweig, Germany) were cultivated on carrot piece agar<sup>2</sup> with the exception of *P. fragariae*, which was cultivated on red kidney bean agar.<sup>40</sup> 50 mg mycelia per isolate were harvested by scraping them from the agar surface. Furthermore, the DNA of *P. ramorum* BBA9/95 and *P. kernoviae* JKI-080-09-00-00-00-03 was extracted from 100 mg of artificially infected *rhododendron* leaves. The mycelium and infected plant material were frozen in liquid nitrogen and ground twice for 30 s at 70% speed in a mill (Retsch, Haan, Germany).

DNA extraction was performed by using the InviMag Plant DNA Mini Kit according to the recommendation of the manufacturer (Invitex, Berlin, Germany).

### Thermophilic helicase-dependent amplification

We decided to exploit an asymmetric thermophilic HDA (tHDA) approach that amplifies target DNA efficiently at 65 °C and requires less protein components than the ambient temperature platform.<sup>22,35</sup> This second-generation HDA approach led to higher specificity and sensitivity.<sup>20</sup> For the adaption of the *Phytophthora* target DNA amplification<sup>41</sup> to the HDA system we used the commercially available IsoAmp II tHDA kit from Bio-Helix. The IsoAmp® II Universal tHDA kit (BioHelix, Beverly, MA, USA) was utilized according to the manufacturer's recommendations. The reaction mix contained 1x annealing buffer II, 4 mM MgSO<sub>4</sub>, 40 mM NaCl, 1 μM BSA, 3.5 μl dNTPs, 75 nM biotin-labeled reverse primer, 25 nM forward primer and 3.5 μl

of IsoAmp enzyme mix in a final volume of 50 μl. The primers and capture probes (Table 1; Fig. 1) were designed within the yeast GTP-binding protein (*Ypt1*) target region,<sup>42,43</sup> using the program Sequencher 5.1. Primers fulfill the following criteria: (i) a length of 29–34 bp, (ii) an optimized melting temperature of 64 °C (±2 °C) and a G/C base content between 45 and 55%. Capture probes were constructed to achieve (i) the highest discrimination of target sequences in relation to non-corresponding sequences, (ii) a length of 30–35 bp and (iii) a melting temperature between 62 and 65 °C. A low tendency for sequence secondary structure formation is expressed in delta *G* values<sup>44</sup> between –1 and 1.5. The primer positions are close to the capture probes because the complete amplicon should not extend more than 110 bp to ensure amplification by the Bst polymerase. The tHDA reaction was conducted asymmetrically whereas the ratio between the forward and biotin-labeled reverse primer was 1 : 4.

The HDA reaction mixture was incubated for 90 min at 65 °C (thermophilic) in a miniaturized heating module allowing simple temperature management (Fig. 2). A Peltier heat pump element covered with a copper plate and a heat sink at the other side was used to create isothermal temperature conditions. A polycarbonate plate (thickness 4 mm) with drill-holes (diameter 4 mm, reaction volume 50 μl) was placed between these elements and sealed with a thin foil for incubating the HDA reaction mixture. The size of the heating element is 15 × 15 mm with an electric power of 8.5 Watt. The temperature was measured with a PT1000 platinum resistor thermometer pasted within the copper plate. An electronic controller used this temperature signal to generate a pulse-width modulation (PWM) signal to switch the Peltier element for holding a constant temperature of 65 °C. The operating points for this controller were set by USB connection from a PC.

### Agarose gel electrophoresis

Successful DNA amplification was verified on a 2% (w/v) agarose gel. For visualization the DNA was stained with GelRed (VWR International GmbH, Darmstadt, Germany) according to the recommendations of the manufacturer. The molecular weight marker 'GeneRuler 100 bp DNA Ladder' was purchased from Thermo Scientific.


### On-chip DNA hybridization

The preparation of the chips was performed as previously described.<sup>45,46</sup> The *Ypt1* region was chosen to design the species-specific capture probes<sup>41</sup> (Fig. 1). These capture probes (Eurofins MWG Operon, Ebersberg, Germany; Table 1) were dissolved in spotting buffer (160 mM Na<sub>2</sub>SO<sub>4</sub>, 130 mM Na<sub>2</sub>HPO<sub>4</sub>) to a final concentration of 20 μM and spotted within the electrode gaps of the chip platform (Nanoplotter 2.1 GeSim, Germany; spotting layout see ESI†). A biotin-labeled non-complementary probe was immobilized as a positive control to verify successful enzyme binding *via* biotin-streptavidin interaction and subsequent silver deposition. After UV-linking at 254 nm for 5 min the chips were washed with 0.1 × saline-sodium citrate (SSC)/0.5% sodium dodecyl sulphate (SDS).



Table 1 Primers and capture probes

DNA	Sequence 5'-3'	Modification
HDA_frag.F	GAC CAT TGG CGT CGA CTT TGT GAG TGC TA	
HDA_frag.R	GCA CGA TAA CGT CAG CAA TCG GAG AGC AAA TC	5'-Biotin
HDA_ram.F	CCA TCA AGC TCC AGA TTG TAC GTC TGC	
HDA_ram.R	GAG TAA AAT ATA GAT GTT AGC TGC ATG TCG TTG C	5'-Biotin
HDA_ker.F	GGC TGC ACG AGA TCG ATA GGT GAG TTC TAC	
HDA_ker.R	TCT CMC AGG CGT ATC TGA TTT AAC ACG TGT TCC	5'-Biotin
<i>P. kernoviae</i>	CAC CAC ATG AAT ACC TGC CAG GCG AGA TGC	5'-NH <sub>2</sub> -C6
<i>P. lateralis</i>	CGG GAG ATT TTT TCC CGC TTT CCT TGG GGT AAG	5'-NH <sub>2</sub> -C6
<i>P. ramorum</i>	CCC CCC ACT TTC CGT GGG TGA GTT TCC TTT	5'-NH <sub>2</sub> -C6
<i>P. pinifolia</i>	CCG CGG ACG AAA ACT AAC TCT CTT GTG TAG TG	5'-NH <sub>2</sub> -C6
<i>P. fragariae</i>	CTA GCC TTG CCA TTT CTA GGT CCA AAA AGG C	5'-NH <sub>2</sub> -C6
<i>P. rubi</i>	CTA GCC TTG CCA TTC CTA GGT CCA AAA AGG C	5'-NH <sub>2</sub> -C6
<i>P. austrocedrae</i>	CCT CCG TGG TTC ATG TAC AAA ACG TGC AGC	5'-NH <sub>2</sub> -C6
<i>P. cambivora</i>	GTC CAC CAT GGC TAA GTT TTG ACC TCC AGG	5'-NH <sub>2</sub> -C6
<i>P. cinnamomi</i>	CTG TCT GCC CCA TTC AAC AGA CGC TAA CGT C	5'-NH <sub>2</sub> -C6
Negative control I	GGA CAG GAG CGA TTC AGG ACY ATA ACA AGC AG	5'-NH <sub>2</sub> -C6
Negative control II	ATC GAG CTG GAC GGC AAG ACC ATC AAG CT	5'-NH <sub>2</sub> -C6
Positive control	AGA ATC AAG GAG CAG ATG CTG AAA AAA	5'-NH <sub>2</sub> , 3'-biotin

Fig. 1 This draft shows the positions of the various primer (upper part) and capture probes (lower part) within the *Ypt1* target gene sequence.

The specific detection of *Phytophthora* species was performed in a microfluidic device as previously described.<sup>41,46,47</sup> 20  $\mu$ l of the HDA products were dissolved in 50  $\mu$ l buffer (5  $\times$  SSC/0.1% SDS) and applied on the chips for 15 min at 58  $^{\circ}$ C using an interval flow and further processed.

### Optical and electrical signal readout

The amount of silver deposits was measured optically and electrically. The optical readout was realized by scanning with a reflecta ProScan 7200 slide scanner (reflecta GmbH, Rottenburg, Germany) with a 8 bit grey value and a resolution of 3600 dpi and the subsequent analysis of grey values was performed with ImageJ software (National Institutes of Health, USA). The

grey value is calculated by mean grey value calculation, subtracting the measured background value from the sample values and setting the positive control to 100%. The mean grey value of the internal hybridization control (negative control I) of all experiments was used to set the threshold which is three times the standard deviation (6.41%  $\pm$  6.46%).

For conductance measurement, the DC resistance is computed using an in-house developed portable chip-reader<sup>48</sup> and converted to electrical conductance.

## Results and discussion

### Optimization of helicase-dependent amplification for *Phytophthora* species

Within the present study we explored for the first time HDA as an attractive alternative amplification method for plant pathogens. The HDA reaction was established and optimized for selected *Phytophthora* species (Table 1) by using a robust, miniaturized heating module combined with an existent detection platform.

First, specific primers and capture probes were designed. The capture probes for *P. fragariae*, *P. ramorum* and *P. kernoviae* were recently published by Schwenkbier *et al.*<sup>41</sup> The positions of primers and capture probes were set to amplify a region within the yeast GTP-binding protein 1 (*Ypt1*) gene (Fig. 1). Established isothermal LAMP-based *Phytophthora* detection systems used




Fig. 2 HDA reaction module consisting of the polycarbonate reaction chamber and the electronically controllable heating device. The scale bars within the technical drawing are given in mm.



capture probes that hybridize with the internal transcribed spacer 1 region (ITS1). Due to the fact that by targeting the *Ypt1* instead of the ITS1 region a higher specificity is achievable concomitant with easier adaptability to other regulated *Phytophthora* species, we addressed this molecular target with our chip-based amplification and detection assay. Extensive studies with the *Ypt1* region from *Phytophthora* species showed that it is the best region to get species-specific base pairs within a length of 30–40 bp.<sup>49</sup>

Isolated genomic DNA from various *Phytophthora* cultures was used as a template for the amplification of specific *Ypt1* target gene regions via the tHDA approach (Fig. 1). The primer pair HDA\_frag. allowed the amplification of *P. fragariae* and *P. rubi* *Ypt1* DNA parts. These two species differ in only one single base. And as both *P. fragariae* and *P. rubi* are on the EPPO A2 list, it is not stringently required to discriminate between those plant pathogens. The amplification of *P. ramorum* was realizable with the primers HDA\_ram.F/R. Lastly, the primer set HDA\_ker. was used to amplify a fragment of *P. kernoviae* within the *Ypt1* gene region. A further HDA approach was conducted with genomic DNA isolated from *rhododendron* leaves infected with *P. ramorum/kernoviae*. Additionally, an asymmetric amplification strategy was chosen to generate ssDNA, which facilitates the subsequent hybridization.

The isothermal amplification was performed in a miniaturized HDA reaction module consisting of a polycarbonate plate providing reaction cavities of 50 µl and a heating device to ensure a constant temperature of 65 °C (Fig. 2). It offers several advantages including an accurate temperature control and a small size that ensure its portability. As there is a heater for both, bottom and top plus another cooling from the bottom, we achieve efficient heat conduction. Furthermore the system can be easily adapted to various chip formats and reaction volumes since the height is changeable.

The resulting HDA products were analyzed by agarose gel electrophoresis (Fig. 3). In general the asymmetric tHDA approach led to two distinct bands in the analytical gel. The faster migrating DNA emerged single-stranded and the higher molecular weight band represented double-stranded DNA. The asymmetric tHDA reaction was successfully realized with genomic DNA isolated from cultures (Fig. 3a) as well as from

infected *rhododendron* leaves (Fig. 3b). Thus, the newly designed HDA primer pairs allow for the successful amplification of *Ypt1* target gene regions of regulated *Phytophthora* species by asymmetric isothermal HDA.

### *Phytophthora* specification by on-chip hybridization

A subsequent on-chip hybridization step for proper discrimination of selected *Phytophthora* species concerning their *Ypt1* gene region was performed. The resulting hybridization signals were detectable by the naked eye due to the formed silver deposits in the case of matching the capture and target probe. We used HDA-generated DNA from *P. fragariae*, *P. kernoviae*, *P. ramorum* (template DNA isolated from cultures) and *P. ramorum* or *P. kernoviae*-infected plant samples (template DNA isolated from infected *rhododendron* leaves) to verify the functionality of on-chip hybridization. The results of five independent experiments for each of the *Phytophthora* isolates are displayed in Fig. 4 (left panel: diagrams with grey values for the spotted capture probes; right panel: chips with silver deposits). Specific signals were obtained for tHDA amplicons of *P. fragariae*, which indicated that a sufficient amount of ssDNA specifically bound to the matching capture probes (Fig. 4a). Besides the specific signals for *P. fragariae*, signals for *P. rubi* can also be detected. This can be explained by the fact that the amplified sequence of both *Phytophthora* species only differs in one single base. In the case of *P. kernoviae* the tHDA amplicons also yielded high and specific hybridization signals (Fig. 4b). In contrast, a slightly lower but still very distinct hybridization signal was detectable for *P. ramorum* (Fig. 4c). These experiments were repeated independently for at least five times per species to ensure the reproducibility. Due to the necessity of short fragments for the HDA, the hybridization efficiency is great and the signals are still specific. The target DNA binds selectively to the capture probes without showing any false-positive signals. Thus, the specificity of the capture probes is illustrated.

In addition to the grey value analysis an electrical detection was performed. A matching hybridization and subsequent enzyme binding result in the deposition of silver between the electrode gaps of the chip. The metallic silver enables the closure of the gap and the electric resistance on each individual spot can be measured. The conductivity signals of *P. fragariae*, *P.*




Fig. 3 Analytical agarose gel to monitor the successful amplification of the *Ypt1* target gene region of various cultivated *Phytophthora* species (left panel) and *Phytophthora*-infected *rhododendron* leaves (right panel).








Fig. 4 Grey values and electric conductance as endpoint signals for the successful on-chip hybridization of HDA products of culture samples of *P. fragariae* (a), *P. kernoviae* (b) and *P. ramorum* (c). In the diagram the signal intensities in % are depicted as normalized values related to the positive control. The dashed line indicates the calculated threshold.

*kernoviae* and *P. ramorum* reflect the results obtained by grey value analysis (Fig. 4).

The results for the infected plant samples are illustrated in Fig. 5 (left panel: diagrams with grey values for the spotted capture probes; right panel: chips with silver deposits). The

hybridization signals of amplified target DNA, which was isolated from *P. ramorum* or *P. kernoviae* infected *rhododendron* leaves appear significantly. Thus, also the combined technology of tHDA and on-chip hybridization allowed a discrimination of *Phytophthora* species in real plant samples.




Fig. 5 Grey values as endpoint signals for successful on-chip hybridization of HDA products from *rhododendron* leaves infected with *P. ramorum* (a) or *P. kernoviae* (b).

## Conclusions

Within this contribution, the identification of some EPPO-listed *Phytophthora* species by combining isothermal asymmetric tHDA with on-chip hybridization is introduced for the first time. The optical as well as electrical detection provided valid results for the analysis of the regulated plant pathogens.

Isothermal amplification techniques were developed as an alternative to PCR for target gene amplification omitting the use of a thermocycler. In particular, HDA enables truly isothermal amplification without the need for prior heat denaturation or elaborated primer design, which is mandatory for LAMP. To date, no report addresses the HDA technique for the amplification of plant pathogens, in particular *Phytophthora*. Our developed asymmetric tHDA approach was successfully applied to amplify isolated template DNA from *Phytophthora* cultures and infected plant material. For an effective discrimination of several regulated *Phytophthora* species the *Ypt1* region was chosen to design species-specific capture probes. These probes are located within a 450 bp region of the *Ypt1* gene. Current HDA protocols allow the amplification of DNA fragments with a maximum length of 120 bp, hence three different primer pairs had to be designed to cover all species investigated in this study. For the establishment of our HDA experiments we started with only one primer pair per reaction.

HDA combined with chip-based detection of regulated *Phytophthora* species offers great potential for on-site detection. Significant improvements can allow the use of portable testing devices directly in the field or at the location, where a suspicious

plant has to be investigated. This can concentrate sampling, detection as well as intervention and, thereby, reduce the delay between taking a sample of infected plants and obtaining a valid result. In order to realize a putative field application, isothermal nucleic acid amplification was optimized to substitute PCR, which requires a cost-intensive thermocycler. We demonstrated that the tHDA-based amplification as well as on-chip detection can be conducted in miniaturized and portable devices that enable on-site operating performance. The tHDA performance omits the need for thermal cycling and laborious technical requirements. Additionally, the development of disposable, low-cost chips can facilitate the near future availability of portable devices for chip-based DNA analytics. Further, the generated silver spots on the chips represent robust and long-lasting endpoint signals, which are already detectable by the naked eye. In contrast to a recently reported study based on HDA and fluorescence detection of bacterial pathogens, our colorimetric approach eliminated signal loss due to fading or expensive detection equipment. Also a conductance measurement is realizable *via* metallic silver, bridging the electrode gaps in the case of a matching DNA hybridization. The resulting decreased electrical resistance can be readout with our proprietary portable chip reader. Taken together, the presented results concerning an isothermal amplification and subsequent on-chip detection of *Phytophthora* pathogens in plant samples, realized in simple, modular, miniaturized devices, display great potential for upcoming on-site applications.

## Acknowledgements

The project “PhytoChip-Validierung” (28-1-54.081-10) is supported by the Federal Ministry of Food and Agriculture (BMEL) based on a decision of the Parliament of the Federal Republic of Germany *via* the Federal Office for Agriculture and Food (BLE) under the innovation support program. The heating system was provided by the LungCARD project. The project “JBCI 2.0” (InnoProfile-Transfer, Unternehmen Region), supported by the Federal Ministry of Education and Research (BMBF) Germany, is gratefully acknowledged.

## References

- 1 D. C. Erwin and O. K. Ribeiro, *Phytophthora Diseases Worldwide*, APS Press, St. Paul, Minn., 1996.
- 2 S. Werres, R. Marwitz, W. Veld, A. De Cock, P. J. M. Bonants, M. De Weerd, K. Themann, E. Ilieva and R. P. Baayen, *Mycol. Res.*, 2001, **105**, 1155–1165.
- 3 C. Brasier and J. Webber, *Nature*, 2010, **466**, 824–825.
- 4 O. Lazcka, F. J. Del Campo and F. X. Munoz, *Biosens. Bioelectron.*, 2007, **22**, 1205–1217.
- 5 P. A. O'Brien, N. Williams and G. E. S. Hardy, *Crit. Rev. Microbiol.*, 2009, **35**, 169–181.
- 6 F. N. Martin, Z. G. Abed, Y. Baldi and K. Ivors, *Plant Dis.*, 2012, **96**, 1080–1103.
- 7 P. Craw and W. Balachandran, *Lab Chip*, 2012, **12**, 2469–2486.



- 8 C.-C. Chang, C.-C. Chen, S.-C. Wei, H.-H. Lu, Y.-H. Liang and C.-W. Lin, *Sensors*, 2012, **12**, 8319–8337.
- 9 P. J. Asiello and A. J. Baeumner, *Lab Chip*, 2011, **11**, 1420–1430.
- 10 J. Kim and C. J. Easley, *Bioanalysis*, 2011, **3**, 227–239.
- 11 F. Sidoti, M. Bergallo, C. Costa and R. Cavallo, *Mol. Biotechnol.*, 2013, **53**, 352–362.
- 12 P. Gill and A. Ghaemi, *Nucleosides, Nucleotides Nucleic Acids*, 2008, **27**, 224–243.
- 13 J. A. Tomlinson, I. Barker and N. Boonham, *Appl. Environ. Microbiol.*, 2007, **73**, 4040–4047.
- 14 J. A. Tomlinson, M. J. Dickinson and N. Boonham, *Phytopathology*, 2010, **100**, 143–149.
- 15 T.-T. Dai, C.-C. Lu, J. Lu, S. Dong, W. Ye, Y. Wang and X. Zheng, *FEMS Microbiol. Lett.*, 2012, **334**, 27–34.
- 16 T. Notomi, H. Okayama, H. Masubuchi, T. Yonekawa, K. Watanabe, N. Amino and T. Hase, *Nucleic Acids Res.*, 2000, **28**, e63.
- 17 Y. Mori, K. Nagamine, N. Tomita and T. Notomi, *Biochem. Biophys. Res. Commun.*, 2001, **289**, 150–154.
- 18 Y. Mori, M. Kitao, N. Tomita and T. Notomi, *J. Biochem. Biophys. Methods*, 2004, **59**, 145–157.
- 19 M. Vincent, Y. Xu and H. M. Kong, *EMBO Rep.*, 2004, **5**, 795–800.
- 20 L. X. An, W. Tang, T. A. Ranalli, H. J. Kim, J. Wytiaz and H. M. Kong, *J. Biol. Chem.*, 2005, **280**, 28952–28958.
- 21 A. Motre, Y. Li and H. Kong, *Gene*, 2008, **420**, 17–22.
- 22 Y.-J. Jeong, K. Park and D.-E. Kim, *Cell. Mol. Life Sci.*, 2009, **66**, 3325–3336.
- 23 D. Andresen, M. von Nickisch-Rosenegk and F. F. Bier, *Clin. Chim. Acta*, 2009, **403**, 244–248.
- 24 D. Andresen, M. von Nickisch-Rosenegk and F. F. Bier, *Expert Rev. Mol. Diagn.*, 2009, **9**, 645–650.
- 25 M. Mahalanabis, J. do, H. Almuayad, J. Y. Zhang and C. M. Klapperich, *Biomed. Microdevices*, 2010, **12**, 353–359.
- 26 N. Ramalingam, T. C. San, T. J. Kai, M. Y. M. Mak and H.-Q. Gong, *Microfluid. Nanofluid.*, 2009, **7**, 325–336.
- 27 W. H. A. Chow, C. McCloskey, Y. Tong, L. Hu, Q. You, C. P. Kelly, H. Kong, Y.-W. Tang and W. Tang, *J. Mol. Diagn.*, 2008, **10**, 452–458.
- 28 S. Huang, J. Do, M. Mahalanabis, A. Fan, L. Zhao, L. Jepeal, S. K. Singh and C. M. Klapperich, *PLoS One*, 2013, **8**, e60059.
- 29 J. Goldmeyer, H. Li, M. McCormac, S. Cook, C. Stratton, B. Lemieux, F. Kong, W. Tang and Y.-W. Tang, *J. Clin. Microbiol.*, 2008, **46**, 1534–1536.
- 30 G. C. Frech, D. Munns, R. D. Jenison and B. J. Hicke, *BMC Res. Notes*, 2012, **5**, 430.
- 31 V. Doseeva, T. Forbes, J. Wolff, Y. Khripin, D. O'Neil, T. Rothmann and I. Nazarenko, *Diagn. Microbiol. Infect. Dis.*, 2011, **71**, 354–365.
- 32 Y. Tong, B. Lemieux and H. Kong, *BMC Biotechnol.*, 2011, **11**, 50.
- 33 E. Torres-Chavolla and E. C. Alocilja, *Biosens. Bioelectron.*, 2011, **26**, 4614–4618.
- 34 A. Motre, R. Kong and Y. Li, *J. Microbiol. Methods*, 2011, **84**, 343–345.
- 35 J. Goldmeyer, H. Kong and W. Tang, *J. Mol. Diagn.*, 2007, **9**, 639–644.
- 36 J. A. Jordan, C. O. Ibe, M. S. Moore, C. Host and G. L. Simon, *J. Clin. Virol.*, 2012, **54**, 11–14.
- 37 C. Domingo, P. Patel, J. Yillah, M. Weidmann, J. A. Mendez, E. R. Nakoune and M. Niedrig, *J. Clin. Microbiol.*, 2012, **50**, 4054–4060.
- 38 H.-J. Kim, Y. Tong, W. Tang, L. Quimson, V. A. Cope, X. Pan, A. Motre, R. Kong, J. Hong, D. Kohn, N. S. Miller, M. D. Poulter, H. Kong, Y.-W. Tang and B. Yen-Lieberman, *J. Clin. Virol.*, 2011, **50**, 26–30.
- 39 N. S. Miller, B. Yen-Lieberman, M. D. Poulter, Y.-W. Tang and P. A. Granato, *J. Clin. Virol.*, 2012, **54**, 355–358.
- 40 S. Werres and R. Casper, *J. Phytopathol.*, 1987, **118**, 367–369.
- 41 L. Schwenkbier, S. König, S. Wagner, S. Pollok, J. Weber, M. Hentschel, J. Popp, S. Werres and K. Weber, *Microchim. Acta*, 2013, **180**, 15–16.
- 42 L. Schena and D. E. L. Cooke, *J. Microbiol. Methods*, 2006, **67**, 70–85.
- 43 L. Schena, J. M. Duncan and D. E. L. Cooke, *Plant Pathol.*, 2008, **57**, 64–75.
- 44 M. Zuker, *Nucleic Acids Res.*, 2003, **31**, 3406–3415.
- 45 T. Schueler, R. Kretschmer, S. Jessing, M. Urban, W. Fritzsche, R. Moeller and J. Popp, *Biosens. Bioelectron.*, 2009, **25**, 15–21.
- 46 B. Seise, A. Brinker, R. Kretschmer, M. Schwarz, B. Rudolph, T. Kaulfuss, M. Urban, T. Henkel, J. Popp and R. Moeller, *Eng. Life Sci.*, 2011, **11**, 148–156.
- 47 S. Wuenscher, B. Seise, D. Pretzel, S. Pollok, J. Perelaer, K. Weber, J. Popp and U. S. Schubert, *Lab Chip*, 2014, **14**, 392–401.
- 48 M. Urban, R. Moller and W. Fritzsche, *Rev. Sci. Instrum.*, 2003, **74**, 1077–1081.
- 49 S. König, L. Schwenkbier, M. Riedel, S. Wagner, S. Pollok, J. Popp, K. Weber and S. Werres, *Plant Pathol.*, 2014, submitted.

