

CORRECTION

[View Article Online](#)  
[View Journal](#) | [View Issue](#)



Cite this: *Energy Environ. Sci.*, 2015, **8**, 1046

DOI: 10.1039/c5ee90006a

[www.rsc.org/ees](http://www.rsc.org/ees)

## Correction: Photoelectrochemical reduction of nitrates at the illuminated p-GaInP<sub>2</sub> photoelectrode

Heli Wang\* and John A. Turner

Correction for 'Photoelectrochemical reduction of nitrates at the illuminated p-GaInP<sub>2</sub> photoelectrode' by Heli Wang *et al.*, *Energy Environ. Sci.*, 2013, **6**, 1802–1805.

The previous IPCE spectrum (Fig. 4) was incorrect, due to our old instrument being out of calibration. With a new setup constructed, calibration was carried out by putting a reference cell inside the working glass cell to get rid of the effect of the glass window. A 375 nm filter was placed in front of the glass cell. IPCE was calculated based on the quantum efficiency of the reference cell.

Accordingly, Fig. 4 in our paper should be replaced with the following:




Fig. 4 IPCE for the p-GaInP<sub>2</sub> electrode at  $-1$  V (vs. Ag/AgCl) in pH 1 NH<sub>4</sub>NO<sub>3</sub> solution at room temperature. IPCE is over 60% until 620 nm; below 2% at 680 nm.

The Royal Society of Chemistry apologises for these errors and any consequent inconvenience to authors and readers.

