click for updates

Correction: Acoustic actuated fluorescence activated sorting of microparticles

Ola Jakobsson,^{*a} Carl Grenvall,^a Maria Nordin,^a Mikael Evander^a
and Thomas Laurell^{ab}

Cite this: *Lab Chip*, 2015, 15, 4625

DOI: 10.1039/c5lc90123e

www.rsc.org/loc

Correction for 'Acoustic actuated fluorescence activated sorting of microparticles' by Ola Jakobsson *et al.*, *Lab Chip*, 2014, 14, 1943–1950.

In the Materials and methods section, under the paragraph titled "Evaluating sort purity" there is an error in the formula:

$$\text{Recovery \%} = 100 - \frac{\text{purity waste outlet \%}}{\text{starting purity \%}}$$

The correct formula should read:

$$\text{Recovery \%} = 100 \times \left(1 - \frac{\text{purity waste outlet \%}}{\text{starting purity \%}} \right)$$

The correct version of the formula was used in all calculations and so the results are not affected by this error.

The Royal Society of Chemistry apologises for these errors and any consequent inconvenience to authors and readers.

^a Lund University – Dept. of Biomedical Engineering, Lund, Sweden. E-mail: ola.jakobsson@bme.lth.se

^b Dongguk University – Dept. Biomedical Engineering, Seoul, Republic of Korea