Cite this: *RSC Adv.*, 2015, 5, 89603

Centrifugo-thermopneumatic fluid control for valving and aliquoting applied to multiplex real-time PCR on off-the-shelf centrifugal thermocycler†

M. Keller,^{*ab} S. Wadle,^{ab} N. Paust,^{ab} L. Dreesen,^b C. Nuese,^b O. Strohmeier,^{ab}
R. Zengerle^{abc} and F. von Stetten^{ab}

We introduce microfluidic automation of geometrically multiplexed real-time PCR to off-the-shelf Rotor-Gene Q thermocyclers (RGQ, QIAGEN GmbH, Hilden, Germany). For centrifugal fluid control the RGQ provides low and constant rotation of 400 rpm, only. Compatibility to this very limited flexibility of centrifugal actuation is achieved by using thermal gas compression and expansion for valving and aliquoting. In contrast to existing thermo-pneumatic actuation, centrifugo-thermopneumatic (CTP) fluid control employs the induced change of partial vapor pressure by global temperature control as actuation parameter for two new unit operations: CTP siphon valving and CTP two-stage aliquoting. CTP siphon valving was demonstrated to reliably transfer sample liquid in all cases ($n = 35$) and CTP two-step aliquoting transfers metered aliquots of $18.2 \pm 1.2 \mu\text{l}$ (CV 6.7%, $n = 8$) into reaction cavities within 5 s ($n = 24$). Thermal characteristics of CTP two-stage aliquoting were found to be in good agreement with an introduced analytical model ($R^2 = 0.9876$, $n = 3$). A microfluidic disk segment comprising both new unit operations was used for automation of real-time PCR amplification of *Escherichia coli* DNA. Required primers and probes were pre-stored in the reaction cavities and a comparison with reference reactions in conventional PCR tubes yielded the same PCR efficiency, repeatability, and reproducibility.

Received 10th August 2015
Accepted 14th October 2015

DOI: 10.1039/c5ra16095b

www.rsc.org/advances

Introduction

In the field of nucleic acid amplification testing, real-time polymerase chain reaction (PCR) constitutes the method of choice.¹ Real-time PCR thermocyclers, featuring the required temperature cycling and fluorescence signal detection, have evolved to high-precision instruments.² In order to perform different reactions in parallel, multiple reagents must be manually pipetted (aliquoted) into different reaction tubes in a time-consuming and error-prone workflow. While this workflow may be automated by liquid handling robots, they are costly and especially not affordable at low- to medium-throughput laboratories.

To circumvent the problems described for manual PCR preparation several microfluidic solutions based on unit operations for the automation of aliquoting have been presented.^{3,4} In comparison to single-stage aliquoting, two-stage aliquoting structures⁵ first meter an introduced sample liquid into sub-volumes, before transferring the aliquots into reaction cavities.⁴ The approach was demonstrated to have a low risk of cross-contamination between PCR reagents inside different reaction cavities.⁶ Using two-stage aliquoting, different applications on the centrifugal microfluidic *LabDisk* platform operated in a modified Rotor-Gene thermocycler (QIAGEN GmbH, Hilden, Germany) were shown.^{6–9} The custom modification enabled the change in rotational frequency to make use of known valving concepts *e.g.* centrifugo-pneumatic valves⁵ in-between metering and transfer of metered aliquots into PCR reaction cavities. In contrast, microfluidic cartridges for off-the-shelf laboratory instruments, termed Microfluidic Apps,¹⁰ improve applicability of such microfluidic approaches. To realize a Microfluidic App for the Rotor-Gene and make use of its high-precision performance for real-time PCR without changing the instrument itself, novel valves for the introduction of a sample into an aliquoting structure and transfer of metered aliquots into reaction cavities are required. Such valves have to cope with the low and constant rotation of 400 rpm of the Rotor-

^aHahn-Schickard-Gesellschaft für angewandte Forschung e.V., Georges-Koehler-Allee 103, 79110 Freiburg, Germany. E-mail: mark.keller@hahn-schickard.de; Fax: +49 761 203-73299; Tel: +49 761 203-73223

^bLaboratory for MEMS Applications, IMTEK – Department of Microsystems Engineering, University of Freiburg, Georges-Koehler-Allee 103, 79110 Freiburg, Germany

^cBIOS – Centre for Biological Signalling Studies, University of Freiburg, Schanzlestr. 18, 79104 Freiburg, Germany

† Electronic supplementary information (ESI) available: Experimental details and a video of stationary stroboscopic images. See DOI: 10.1039/c5ra16095b

Gene. As the majority of passive valves require a change in centrifugal pressure and thus rotational speed, only time-dependent and active valves are considered suitable.⁴ Timing of valving by dissolving of membranes requires integration of such membranes and a paper strip to guide a propagating auxiliary liquid to them.¹¹ An alternative approach also requires an auxiliary liquid to open air vents inside a clepsidra structure in a timely manner.¹² Their space consumption on the cartridge, increased complexity of manufacturing and lack of both pre-storage and release mechanism of the auxiliary liquid are major drawbacks. For active valves, off-the-shelf Rotor-Gene thermocyclers offer global temperature actuation. Though only demonstrated for local temperature application, active valves have been demonstrated to utilize melting of paraffin-wax^{13,14} or thermo-pneumatic liquid actuation.^{15–17} Wax valves require additional fabrication steps and may cause union of processed reagents with the wax material. While thermo-pneumatic liquid actuation theoretically holds potential for operation on Rotor-Gene thermocyclers, it still lacks pumping efficiency considering the ratio of the space-consuming air volume and the liquid volume, which may be actuated by the thermally induced ideal gas expansion or contraction. Recently, thermo-pneumatic actuation was demonstrated for the deflection of elastic membranes to open and close fluidic paths, with the major drawback of having a hybrid cartridge assembly and required fabrication steps.¹⁸

In contrast, in this work we employ centrifugo-thermopneumatic (CTP) fluid control,^{19,20} which for the first time utilizes temperature-induced partial vapor pressure change inside an enclosed gas volume for valving and aliquoting. CTP fluid control thereby highly exceeds pumping efficiencies of known thermo-pneumatic actuation on centrifugal microfluidic platforms.^{15–17} We present an analytical model of CTP fluid control to introduce the functional principle of two novel unit operations: CTP siphon valving for initial sample transfer and CTP two-stage aliquoting for distribution of the transferred sample into multiple reaction cavities. CTP siphon valving for the first time in monolithic thermo-pneumatic liquid actuation features a distinct valving temperature, instead of gradually displacing liquid over the time of temperature change. CTP two-stage aliquoting is utilized for experimental verification of the analytical model and a quantification of its fluidic aliquoting performance is presented. The applicability of the centrifugal microfluidic disk segment “GeneSlice” comprising both CTP unit operations for automation of real-time PCR inside an off-the-shelf Rotor-Gene Q (RGQ) is demonstrated: geometric multiplexing of a DNA sample of the model organism *Escherichia coli* (*E. coli*) into eight reaction cavities with pre-stored primers and probes allows comparison with a manual reference (*i.e.* pipetting in conventional real-time PCR tubes) and determination of key PCR characteristics.

Theory

Workflow for geometrically multiplexed real-time PCR

The *GeneSlice* automates the depicted workflow for geometrically multiplexed real-time PCR (Fig. 1B) utilizing both known

and novel CTP unit operations (Fig. 1B, indicated in blue), which can be partly grouped into the process chain of main-amplification, as recently introduced by Strohmeier *et al.*⁴

After supply of a mixture of DNA sample and liquid PCR reagents (Fig. 1A, blue) by pipetting into a sample inlet, CTP siphon valving transfers the mixture into the CTP two-stage aliquoting structure. Here, the mixture is metered into eight sub-volumes and transfer of the metered aliquots into eight reaction cavities takes place. Each reaction cavity comprises a (different) pre-stored set of primers and probes (Fig. 1A, red), which is mixed with the aliquoted volume by rehydration (Fig. 1A, purple). Thermocycling and fluorescence detection are executed for all eight real-time PCRs in parallel.

Principle of centrifugo-thermopneumatic fluid control

We consider a chamber A that is partially filled with liquid and subsequently closed to enclose a gas volume V_{g,T_0} and a connection channel B to vented downstream structures as illustrated in Fig. 2. Upon global temperature change from T_0 to T , a gas volume expansion (or contraction) ΔV_g of the enclosed gas volume is induced. As a consequence, a portion ΔV_l of the liquid volume V_l in chamber A is displaced into the vented channel B.

At all times during temperature change an equilibrium of pressures in the depicted system of Fig. 2 is given, which can be described as follows,

$$p_g + \Delta p_{\text{cap},A} + \Delta p_{\text{cent},A} = \Delta p_{\text{cent},B} + \Delta p_{\text{cap},B} + p_0 \quad (1)$$

where $\Delta p_{\text{cap},A}$ and $\Delta p_{\text{cap},B}$ are the capillary pressures at the liquid–gas interfaces and $\Delta p_{\text{cent},A}$ and $\Delta p_{\text{cent},B}$ are the centrifugal pressures inside the chamber A and the channel B, respectively, p_0 is the atmospheric pressure and p_g is the total gas pressure inside the enclosed gas volume. The latter can be expressed as the sum of all partial gas pressures,

$$p_g = \sum_{i=1}^n p_i = p_{\text{air}} + p_{\text{vap}} \quad (2)$$

where p_{air} is the partial air pressure and p_{vap} is the partial pressure of the vapor in thermodynamic equilibrium with the liquid phase. The partial vapor pressure can be approximated with the Antoine equation at a given temperature T ,

$$p_{\text{vap}} = \frac{101\,325}{760} \times 10^{A - \frac{B}{C+T-273.15}} \quad (3)$$

where A , B , and C represent component-specific constants.

As the chemical species of each partial pressure takes up the entire enclosed gas volume and the amount n of air molecules, which is enclosed at the initial temperature T_0 remains constant throughout the temperature change to T , the volume expansion/contraction from V_{g,T_0} to $V_{g,T}$ can be described using the ideal gas law,

$$\frac{p_{\text{air},T_0} V_{g,T_0}}{R T_0} = n = \frac{p_{\text{air},T} V_{g,T}}{R T} \quad (4)$$

where R is the universal gas constant and p_{air,T_0} and $p_{\text{air},T}$ are the partial air pressures at temperature T_0 and T , respectively.

where $\Delta p_{\text{cent/cap},T_0/T}$ are the difference of the centrifugal/capillary pressures $\Delta p_{\text{cent/cap},A}$ in chamber A and $\Delta p_{\text{cent/cap},B}$ in channel B at temperature T_0/T and $p_{\text{vap},T_0/T}$ is the partial vapor pressure at temperature T_0/T described by eqn (3). From eqn (5), fundamental relations can be derived, which will be discussed for each unit operation in the following sections.

Centrifugo-thermopneumatic siphon valving

The diagram illustrates a two-chamber system. **Chamber A** (left) contains a gas phase (red dots) and a liquid phase (blue). The gas volume is $V_{g,T0}$ at pressure p_g . The liquid volume is $V_l - \Delta V_l$. The pressure difference between the gas and the center of the liquid is $\Delta p_{cent,A}$. The pressure difference between the gas and the capillary is $\Delta p_{cap,A}$. The volume change of the gas is ΔV_g . **Vented channel B** (right) is connected to Chamber A. The pressure at the top of the channel is p_0 . The pressure difference between the capillary and the center of the liquid is $\Delta p_{cap,B}$. The pressure difference between the center of the liquid and the vented channel is $\Delta p_{cent,B}$. The volume change of the liquid in the vented channel is ΔV_l . A curved arrow indicates a clockwise cycle.

RSC Adv., 2015, 5, 89603–89611 | 89605

Fig. 3 Functional principle of centrifugo-thermopneumatic (CTP) siphon valve. (A) CTP siphon valve comprising an inlet chamber with a branching volume buffer and a branching overflow siphon. The inlet chamber is filled with a sample volume V_1 using the sample inlet, which is then sealed gas-tight. (B) Liquid is propelled radially outwards by constant rotation at 400 rpm, which encloses a gas volume $V_{g,TRT}$. (C) Temperature increase from T_{RT} to T_2 results in gas volume expansion according to eqn (6), which displaces a portion ΔV_1 of the sample volume into both branching structures with identical radial positions of the menisci. (D) At a critical displaced volume portion, the overflow siphon crest is reached, and the entire sample volume is propelled radially outwards into a downstream CTP two-stage aliquoting structure by centrifugal forces. Liquid menisci are visualized with red arrows. Rotational frequency, temperature and gas volume expansion at each depicted stage (A–D) are plotted over time. Quantitative values in the plots are given for the microfluidic cartridge detailed in the “design and fabrication” section.

$$\Delta V_g = V_{g,T} - V_{g,T_0} = \frac{(p_0 - p_{vap,T_0}) V_{g,T_0} T}{T_0 (p_0 - p_{vap,T})} - V_{g,T_0} \quad (6)$$

In this case, the displaced portion ΔV_1 of sample volume equals the gas volume expansion ΔV_g at all times, and liquid is displaced into the volume buffer and overflow siphon channel (Fig. 3C). Although neglected as influence on the gas volume expansion, centrifugal pressures ensure identical radial positions of the menisci inside the volume buffer and overflow siphon channel. At a temperature T_2 the filling heights reach the siphon crest and downstream transfer of the entire sample volume by centrifugal forces is triggered. First, the volume buffer is emptied, as the enclosed gas volume inside the inlet chamber still holds back the liquid. Eventually, the enclosed gas volume is vented, which allows complete draining of the remaining liquid from the inlet chamber into the downstream CTP two-stage aliquoting structure.

Centrifugo-thermopneumatic two-stage aliquoting

CTP two-stage aliquoting (Fig. 4A) meters the sample into sub-volumes at elevated temperature T_{max} and subsequently transfers the metered aliquots into reaction cavities, both at constant rotational speed. For this, the sample from the upstream CTP siphon valve enters a feeding channel, which centrifugally routes the liquid into metering fingers. Each metering finger serves as an overflow vessel guiding excess liquid downstream

into the adjacent metering finger and eventually into a waste. As surface tension of the liquid decreases with temperature increase, metering accuracy is improved by the elevated temperature. At the radial outlet of each metering finger a dead-end reaction cavity is connected *via* a narrow $180 \times 400 \mu\text{m}$ connection channel. The filling of a metering finger leads to enclosure of a gas volume $V_{g,T_{max}}$ inside the connected reaction cavity. Liquid does not enter the reaction cavities during metering but flow comes to a stop at the equilibrium of pressures inside and outside the enclosed gas volume, which can be described analog to eqn (1) as follows,

$$p_g + \Delta p_{cap,A} = \Delta p_{cent,B} + \Delta p_{cap,B} + p_0 \quad (7)$$

where $\Delta p_{cent,B}$ is the centrifugal pressure and $\Delta p_{cap,A}$ and $\Delta p_{cap,B}$ is the capillary pressure at the advancing and the receding meniscus, respectively, of the liquid plug inside the metering finger. Analog to the CTP siphon valve, eqn (7) may be simplified to $p_g = p_0$ in this work due to the negligible influence of centrifugal and capillary pressures on the gas volume (*cf.* ESI S2†). Thus, the transfer of the metered aliquots into their corresponding reaction cavities may be modelled using eqn (6). The required thermodynamic equilibrium of the partial vapor pressure inside the reaction cavity and the liquid inside the metering finger is reached during a saturation phase at elevated temperature T_{max} with the aliquots still remaining in the metering fingers. Corresponding gas volume expansion pushes the liquid out of the

Fig. 4 Functional principle of centrifugo-thermopneumatic (CTP) two-stage aliquoting. (A) CTP two-stage aliquoting comprises a feeding channel with metering fingers, which are connected to dead-end reaction cavities via narrow channels. During metering gas volumes $V_{g,T_{\max}}$ inside each reaction cavity are enclosed. (B) Due to saturation of the gas volumes inside the reaction cavities with partial vapor pressure, bubbles are formed and escape through the metering fingers. (C) Temperature decrease from $T_{\max}$ to T_1 contracts the gas volume according to eqn (6), which aspirates liquid from the metering fingers into the reaction cavities. (D) Once the entire aliquot is transferred, the gas volume is vented again and is free to expand and contract during subsequent thermocycling for PCR. Liquid menisci are marked with red arrows. Temperature and gas volume contraction at each depicted stage (A–D) are plotted over time. The entire operation takes place at constant rotational speed of 400 rpm. Quantitative values in the plots are given for the microfluidic cartridge detailed in the “design and fabrication” section.

narrow channel and back to the metering finger eventually forming a gas bubble at the widening of the narrow channel into the metering finger. At a critical size, the gas bubble breaks off, rises through the metering finger and escapes (Fig. 4B). Here, the lowered surface tension of the liquid at elevated temperature supports gas bubble formation and its escape.

After completion of metering and saturation of the enclosed gas volumes, temperature is decreased from $T_{\max}$ to T_1 to contract the enclosed gas volume $V_{g,T}$ as described by eqn (6) ($T_0 = T_{\max}$). The volume contraction ΔV_g results in aspiration of an equal portion ΔV_l of the aliquot from the metering finger into the reaction cavity (Fig. 4C). Within one cool down step (given a sufficient ΔV_g for ΔV_l), the metering fingers are eventually emptied and the initially enclosed gas volumes inside the reaction cavities are vented again. Ventilation allows free gas volume expansion and contraction without liquid actuation during subsequent thermocycling for PCR (Fig. 4D).

Materials and methods

Microfluidic cartridge design and fabrication

A microfluidic cartridge comprising a CTP siphon valve for 180 μ l of sample volume to be transferred at up to $T_{\max} = 90$ °C and a CTP two-stage aliquoting structure for eight 20 μ l-aliquots with an excess waste volume was designed in SolidWorks 2011 (Dassault Systèmes SolidWorks, Waltham, MA, USA, Fig. 5A). The microfluidic cartridge, a disk segment “GeneSlice”

of our *LabDisk* platform, was manufactured in-house by the Hahn-Schickard Lab-on-a-Chip Design + Foundry Service (www.hahn-schickard.de/en/manufacturing/lab-on-a-chip-design-foundry-service) using a μ TSL replication process.²¹ The micro-thermoformed foil cartridges were sealed applying pressure-sensitive adhesive polyolefin foil (#900 320, HJ-Bioanalytik, Erkelenz, Germany), which included laser-cut (PLS3.60, Universal Laser Systems, Scottsdale, AZ, USA) air vents (Fig. 4A). For real-time PCR experiments, a mixture of primers (10 μ M) and probes (5 μ M, together 0.8 μ l) with trehalose as stabilizer²² (0.1 M, 1 μ l) was pre-stored by pipetting and air drying into all eight reaction cavities of each *GeneSlice* prior to sealing (cf. ESI Table S1† for sequences). Those *GeneSlices* were stored in aluminum barrier film flat bags (A 20T, Long Life for Art, Eichstetten, Germany) with silica gel desiccants (#610.002, ThoMar, Luetau, Germany) under nitrogen atmosphere until execution of experiments.

Experimental setup

GeneSlices were placed in a custom rotor holder (*GeneSlice 72 Rotor*, Hahn-Schickard, Freiburg, Germany, Fig. 5B) that can be inserted into an off-the-shelf RGQ similar to standard tube rotors (Fig. 5C), allowing a maximum of 72 reaction cavities by its spacing. The reaction cavities of the *GeneSlices* are aligned to the fluorescence detector inside the RGQ by two laser-cut alignment holes in each *GeneSlice* and corresponding alignment pins on the *GeneSlice 72 Rotor* (Fig. 5B). The *GeneSlice 72*

Fig. 5 Workflow of insertion of microfluidic disk segment "GeneSlice" into an off-the-shelf Rotor-Gene Q (RGQ, QIAGEN GmbH, Hilden, Germany). (A) Overview of *GeneSlice*. (B) Custom *GeneSlice* 72 Rotor, which holds up to four *GeneSlices*. Alignment pins are provided by the *GeneSlice* 72 Rotor for precise alignment of the *GeneSlice* reaction cavities to the fluorescence detector inside the RGQ. *GeneSlices* are held into place during processing by a locking ring. (C) Depiction of its insertion into an RGQ instead of the depicted standard tube rotor.

Rotor can hold up to four *GeneSlices* to be processed in parallel within the same run, each processing an individual sample.

For fluidic experiments, the sample volume of 180 μL consisted of a one-to-one mixture of PCR master mix (2x QuantiFast Probe (ROX), QIAGEN) and PCR-grade water (no-template control, NTC). In real-time PCR experiments for detection of *E. coli*, the PCR-grade water was partly substituted by 9 μL of genomic DNA (*E. coli*, O157:H7) dissolved in Nucleic Acid Dilution Buffer (both provided by QIAGEN). After sample insertion by pipetting, the sample inlet was sealed gas-tight with an adhesive tape (#95.1994, Sarstedt, Nuembrecht, Germany).⁶ The regular RGQ Series Software (version 2.1.0 build 9, QIAGEN) was used for programming all required temperature steps using the "hold at temperature" function: both temperatures (90 and 30 $^{\circ}\text{C}$) are set each to be hold for 3 min to allow complete metering of liquids, saturation of the gas volumes inside the reaction cavities and transfer of aliquots into the reaction cavities. Subsequent thermocycling for real-time PCR is seamlessly executed within the same thermal protocol (for real-time PCR experiments only).

For fluidic development and characterization as well as thermal characterization of CTP two-stage aliquoting, a test site, termed *strobe RGQ* (BioFluidix, Freiburg, Germany), was utilized (cf. ESI Video S1†). The *strobe RGQ* allows stroboscopic stationary image acquisition of rotating *GeneSlices* at arbitrary azimuthal positions inside an RGQ.²⁰ It was tested to have a similar heating and cooling behavior of an off-the-shelf RGQ. Acquired stroboscopic images were *post*-processed by measuring filling heights of liquids using ImageJ (National Institutes of Health, Bethesda, MD, USA). Measured values were fed into the CAD of SolidWorks to compute gas and liquid volumes.

Thermal and fluidic characterization of centrifugo-thermopneumatic two-stage aliquoting

To characterize the gas volume contraction inside the reaction cavities of the CTP two-stage aliquoting structure, three *GeneSlices* in individual runs were processed according to the experimental setup inside the *strobe RGQ*. Simultaneous to the image acquisition, the temperature of the gas phase inside reaction cavity #5 (Fig. 5A) of an additional *GeneSlice* was measured. For this, the additional *GeneSlice* was located on the opposite side of the *GeneSlice* 72 Rotor and thermocouples of a rotary temperature measurement system (TMS)²³ were inserted (cf. ESI Fig. S1† for positions of the thermocouples). The acquired temperature data was used to calculate a theoretical thermal volume contraction over time using eqn (6) ($p_0 = 98\,360\text{ Pa}$, $A = 8.07131$, $B = 1730.63$, $C = 233.426$, all for aqueous liquids, and T_{max} = measured temperature at the beginning of liquid transfer). The reaction cavity volume $V_{\text{g},T_{\text{max}}}$ was designed to 39.0 μL , based on a robustness analysis using the analytical model (cf. ESI S3†). For model verification, the gas volume during temperature decrease of reaction cavity #8 (Fig. 5A) of each *GeneSlice* was experimentally determined based on the liquid volume transferred and the total reaction cavity gas volume $V_{\text{g},T_{\text{max}}}$. The possibility for additional model verification using the gas volume expansion during CTP siphon valving is described and evaluated in ESI S4.† In addition, the transfer times of all eight aliquots per *GeneSlice* were determined using the time stamp of the stationary stroboscopic images.

The fluidic aliquoting performance of the *GeneSlice*, as an important characteristic for geometrically multiplexed real-time PCR, was quantified by processing one *GeneSlice* per run inside an off-the-shelf RGQ according to the experimental setup. Subsequent image acquisition of such processed *GeneSlices* was

performed at 400 rpm on a stroboscopic spin-stand, similar to,²⁴ and *post*-processed according to experimental setup. A total of eight runs were conducted resulting in 64 (8 × 8) quantified aliquots. Prior to the examination, the same method was used to find an appropriate fluidic resistance in-between the CTP siphon valve and CTP two-stage aliquoting structure (*cf.* ESI S5†).

Real-time PCR experiments

The applicability of CTP fluid control for automation of real-time PCR was examined by comparing real-time PCR results from *GeneSlices* and manual reference experiments in standard strip tubes (#981103, QIAGEN) with accordingly pre-stored primers and probes (*cf.* Microfluidic cartridge design and fabrication). Analog to the reaction cavities of *GeneSlices* in their *GeneSlice 72 Rotor*, tubes were arranged in groups of nine at same positions in a standard 72-well rotor (#9018903, QIAGEN), with the ninth tube simulating the thermal mass of the waste of a *GeneSlice*, both without primers and probes. Tubes for reference were manually filled with 20 µl from the real-time PCR reagents by manual pipetting. Two experimental series were conducted, (I) with a constant DNA concentration of 1000 copies in all *GeneSlice* reaction cavities/tubes and (II) with a dilution series with 0 (NTC), 10, 30, 100, 300, and 1000 copies in all *GeneSlice* reaction cavities/tubes. Each experimental series was carried out three times and four *GeneSlices* were included in each run. For the experimental series (II), one *GeneSlice* per run was reserved for the NTC reaction. The five DNA dilutions were allocated to the remaining six *GeneSlices* of two runs. The 1000 copies per reaction cavity (tube) dilution was analyzed in both runs and served as inter-run calibrator to combine results of both runs into a joint standard curve. The temperature protocol for real-time PCR, directly executed within the same run after the fluidic processing, consisted of 3

min at 95 °C for polymerase hot start followed by 40 cycles of 20 s at 95 °C and 25 s at 60 °C with subsequent fluorescence readout (FAM and ROX, the latter as passive reference dye). Real-time PCR data was *post*-processed according to ESI S6† to obtain quantification cycle (C_q) values. Raw fluorescence signals also served as failure alert mechanism to indicate reaction cavities, which may not have been filled with liquid.

Results and discussion

Thermal and fluidic characterization of centrifugo-thermopneumatic two-stage aliquoting

The determined gas volumes inside the reaction cavities during cool down are compared with calculations in Fig. 6. Each calculation is carried out with the exact starting temperature at the point of gas volume enclosure inside the reaction cavity. This yields individual gas contraction curves for each reaction cavity, as exemplarily depicted for one *GeneSlice* in Fig. 6A. A comparison of calculation and all three measurements is depicted in Fig. 6B and shows to be in good agreement ($R^2 = 0.9876$). The remaining deviation between calculation and measurement is likely to be caused by inaccuracies of volume determination and slight temporal shifts between image acquisition and temperature measurement.

The comparison of ideal air volume contraction with gas volume contraction of CTP fluid control, which includes a change in partial vapor pressure, shows that the change in partial vapor pressure is clearly dominating (Fig. 6A). Consequently the highly increased pumping efficiency of CTP fluid control allows smaller changes in temperature or smaller, space-efficient gas volumes, respectively, to actuate equal liquid volumes as by known thermo-pneumatic approaches.^{15–17} CTP fluid control thus renders compatible with global temperature

Fig. 6 Characterization of gas volume contraction in reaction cavities during cool down for centrifugo-thermopneumatic (CTP) two-stage aliquoting. (A) Comparison of calculated and measured volume contraction inside a single reaction cavity over temperature decrease. Additional depiction of three corresponding stroboscopic images (red arrows illustrate liquid menisci). The dashed line (blue) compares CTP fluid control with ideal air volume contraction to illustrate the dominance of partial vapor pressure. (B) Comparison of calculated and measured gas volume contraction for all measurements ($n = 3$).

actuation by ensuring temperatures to stay within safe ranges for common PCR reagents.

The transfer times of metered aliquots into their reaction cavities were measured to be 3.8 ± 0.7 s and below five seconds in all cases ($n = 24$). Thus, the hold time of 3 min at 30 °C for CTP two-stage aliquoting may be drastically reduced.

Quantification of aliquoted volumes resulted in a mean volume of 18.2 μ l with an inter-*GeneSlice* coefficient of variation (CV, standard deviation divided by mean) of 5.5% and mean intra-*GeneSlice* CV of $4.2 \pm 1.8\%$. The determined mean aliquoted volume remains below the nominal volume of 20 μ l, which is caused by minor liquid transfer from one metering finger to the downstream adjacent one and finally into the waste during the saturation phase (Fig. 4B). Due to the low CV values, it is assumed that the mean aliquoted volume can be increased to the nominal volume by increasing the volume of the metering fingers in the cartridge design if required by the application.

Real-time PCR experiments

The experimental series (I) was conducted to determine variations in the quantification cycle (C_q) values obtained from *GeneSlices* and reference reactions in conventional PCR tubes. Corresponding CV values are stated in Table 1 and are a measure of repeatability within one *GeneSlice* and run (intra-*GeneSlice* and intra-run CV, respectively) and of reproducibility of several runs (inter-run CV). In case of the reference reactions, “GeneSlice” was defined as the group of nine tubes, which were located at the same positions in the rotor. Both, repeatability and reproducibility of the *GeneSlice* are similar or superior to the reference.

In experimental series (II), all NTCs, both in reference and *GeneSlices*, stayed below threshold and were thus valid. Results from 10 copies per reaction from both systems were excluded from further *post*-processing due to C_q value variations. The PCR efficiency was determined to 76.2 ± 1.3 and $74.5 \pm 1.7\%$ for reference and *GeneSlice*, respectively, and the corresponding standard curves are depicted in Fig. 7.

Relative comparison of the PCR efficiencies obtained from experimental series (II) show that similar PCR performance may be expected from *GeneSlices* as from the reference in tubes, though the absolute efficiencies remain low for both systems, which may be caused by the assay design. Investigation of the raw fluorescence signals showed that CTP siphon valving and subsequent CTP two-stage aliquoting worked in all

Fig. 7 Mean quantification cycle (C_q) values obtained from real-time PCR with 30, 100, 300, and 1000 copies per reaction in *GeneSlices* and reference reactions in conventional PCR tubes with corresponding standard deviations.

cases ($n = 24$) and demonstrated its value as failure alert mechanism.

Conclusion and outlook

The functionality of CTP siphon valving and CTP two-stage aliquoting was successfully modeled by an analytical description of CTP fluid control. Experimental data of CTP two-stage aliquoting was demonstrated to be in good agreement with the analytical model, also demonstrating the advantageous dominance of partial vapor pressure change. Transfer times for aliquots into their reaction cavities were demonstrated to be below five seconds, which allows reduction in processing time in the future. Aliquoting results showed low CV values and proved sufficient for the demonstrated real-time PCR experiments. The applicability of CTP siphon valving and CTP two-stage aliquoting to real-time PCR was demonstrated by geometric multiplexing *E. coli* DNA into eight reaction cavities with pre-stored primers and probes. Here, similar performance parameters for *GeneSlices* and manual reference were achieved. CTP fluid control constitutes a space-efficient, well-modelled and easy to implement method for centrifugal microfluidic platforms to be operated at constant rotational frequency. Liquid is controlled by global temperatures, which stay within safe ranges for common PCR reagents at all times. The distinct valving temperature, similar to centrifugal burst pressure of passive valves, may allow to carry out biochemical reactions prior to valving: by using a crude lysis and subsequent direct DNA amplification, sample-to-answer nucleic acid testing in the presented *GeneSlice* may be realized. Alternatively, temperature-enhanced rehydration of lyophilized PCR reagents may take place prior to valving, which allows users to introduce pure DNA samples. Both approaches may increase the level of integration in the future. The demonstrated *GeneSlice* constitutes a powerful Microfluidic App for automation of geometrically

Table 1 Real-time PCR performance characteristics of 1000 copies per reaction in *GeneSlices* and reference reactions in conventional PCR tubes

Quantification cycle (C_q) value	<i>GeneSlice</i>	Reference
Mean [cycles]	26.9	26.9
Mean intra- <i>GeneSlice</i> CV [%]	0.5 ± 0.2	0.7 ± 1.5
Mean intra-run CV [%]	0.3 ± 0.2	0.5 ± 0.7
Inter-run CV [%]	0.5	0.5

multiplexed real-time PCR in an off-the-shelf Rotor-Gene Q. Since the RGQ may already be present at various laboratories, panel testing *e.g.* for infectious diseases, genotyping of KRAS point mutations⁸ or MRSA⁶ may now be implemented at such laboratories at no initial investments. In combination with the *GeneSlice* for nested PCR,²⁰ allowing sensitive DNA detection in the fields of forensic science and food analysis, these Microfluidic Apps may bring automation of a variety of applications to a broad group of RGQ users.

Acknowledgements

This work was financially supported from the German Federal Ministry of Education and Research (BMBF) as part of the initiative “MicroMountains Applications – Applikationszentren der Mikrosystemtechnik” (16SV3593). We thank Dr Ralf Himmelreich (formerly QIAGEN, now Fraunhofer ICT-IMM), Markus Jeziorski (formerly QIAGEN, now F. Hoffmann-La Roche), Dr Andy Wende (QIAGEN), Dr Maximilian Focke (formerly IMTEK, now QIAGEN) and Dr Günter Roth (University Freiburg) for helpful discussion, scientific advice and providing PCR reagents.

References

- 1 M. G. Roper, C. J. Easley and J. P. Landers, *Anal. Chem.*, 2005, **77**, 3887–3894.
- 2 H. D. de Paz, P. Brotons and C. Muñoz-Almagro, *Expert Rev. Mol. Diagn.*, 2014, **14**, 827–843.
- 3 R. Gorkin, J. Park, J. Siegrist, M. Amasia, B. S. Lee, J.-M. Park, J. Kim, H. Kim, M. Madou and Y.-K. Cho, *Lab Chip*, 2010, **10**, 1758–1773.
- 4 O. Strohmeier, M. Keller, F. Schwemmer, S. Zehnle, D. Mark, F. von Stetten, R. Zengerle and N. Paust, *Chem. Soc. Rev.*, 2015, **44**, 6187–6229.
- 5 D. Mark, T. Metz, S. Haeberle, S. Lutz, J. Ducreé, R. Zengerle and F. von Stetten, *Lab Chip*, 2009, **9**, 3599–3603.
- 6 M. Focke, F. Stumpf, B. Faltin, P. Reith, D. Bamarni, S. Wadle, C. Müller, H. Reinecke, J. Schrenzel, P. Francois, D. Mark, G. Roth, R. Zengerle and F. von Stetten, *Lab Chip*, 2010, **10**, 2519–2526.
- 7 M. Focke, F. Stumpf, G. Roth, R. Zengerle and F. von Stetten, *Lab Chip*, 2010, **10**, 3210–3212.
- 8 O. Strohmeier, S. Laßmann, B. Riedel, D. Mark, G. Roth, M. Werner, R. Zengerle and F. von Stetten, *Microchim. Acta*, 2014, **181**, 1681–1688.
- 9 O. Strohmeier, N. Marquart, D. Mark, G. Roth, R. Zengerle and F. von Stetten, *Anal. Methods*, 2014, **6**, 2038–2046.
- 10 D. Mark, F. von Stetten and R. Zengerle, *Lab Chip*, 2012, **12**, 2464–2468.
- 11 D. J. Kinahan, S. M. Kearney, O. P. Faneuil, M. T. Glynn, N. Dimov and J. Ducreé, *RSC Adv.*, 2014, **5**, 1818–1826.
- 12 Y. Ukita, Y. Takamura and Y. Utsumi, *Sens. Actuators, B*, 2015, **220**, 180–183.
- 13 K. Abi-Samra, R. Hanson, M. Madou and R. A. Gorkin III, *Lab Chip*, 2011, **11**, 723–726.
- 14 W. Al-Faqheri, F. Ibrahim, T. H. G. Thio, J. Moebius, K. Joseph, H. Arof, M. Madou and A. Han, *PLoS One*, 2013, **8**, e58523.
- 15 K. Abi-Samra, L. Clime, L. Kong, R. Gorkin, T.-H. Kim, Y.-K. Cho and M. Madou, *Microfluid. Nanofluid.*, 2011, **11**, 643–652.
- 16 T. H. G. Thio, F. Ibrahim, W. Al-Faqheri, J. Moebius, N. S. Khalid, N. Soin, M. K. Kahar and M. Madou, *Lab Chip*, 2013, **13**, 3199–3209.
- 17 T. H. G. Thio, F. Ibrahim, W. Al-Faqheri, N. Soin, M. Kahar Bador and M. Madou, *PLoS One*, 2015, **10**, e0121836.
- 18 M. M. Aeinehvand, F. Ibrahim, S. W. Harun, A. Kazemzadeh, H. A. Rothan, R. Yusof and M. Madou, *Lab Chip*, 2015, **15**, 3358–3369.
- 19 M. Focke, O. Strohmeier, P. Reith, G. Roth, D. Mark, R. Zengerle and F. von Stetten, in *Proc. of MicroTAS 2011*, 2011, pp. 659–661.
- 20 M. Keller, J. Naue, R. Zengerle, F. von Stetten and U. Schmidt, *PLoS One*, 2015, **10**, e0131845.
- 21 M. Focke, D. Kosse, D. Al-Bamerni, S. Lutz, C. Müller, H. Reinecke, R. Zengerle and F. von Stetten, *J. Micromech. Microeng.*, 2011, **21**, 115002.
- 22 M. Rombach, D. Kosse, B. Faltin, S. Wadle, G. Roth, R. Zengerle and F. von Stetten, *BioTechniques*, 2014, **57**, 151–155.
- 23 J. Burger, M. Focke, A. Gross, T. Jäger, A. Lastochkin, D. Mark and G. Roth, WO2012046178 A1, 2012.
- 24 M. Grumann, T. Brenner, C. Beer, R. Zengerle and J. Ducreé, *Rev. Sci. Instrum.*, 2005, **76**, 025101.