



Cite this: *Phys. Chem. Chem. Phys.*,
2016, **18**, 31338

Correction: Achieving high resolution and optimizing sensitivity in spatial frequency encoding NMR spectroscopy: from theory to practice

Bertrand Plainchont, Daisy Pitoux, Ghanem Hamdoun, Jean-Michel Ouvrard, Denis Merlet, Jonathan Farjon and Nicolas Giraud*

DOI: 10.1039/c6cp90272c
www.rsc.org/pccp

Correction for 'Achieving high resolution and optimizing sensitivity in spatial frequency encoding NMR spectroscopy: from theory to practice' by Bertrand Plainchont *et al.*, *Phys. Chem. Chem. Phys.*, 2016, **18**, 22827–22839.

The authors would like to correct an error in the published results: the E-Snob pulse whose amplitude profile is presented in Fig. 8 in the main text, and Fig. S4 in ESI, was calibrated for 90° (instead of 270°).

The authors note that this calibration to 90° does not correspond to the optimal setting for an E-Snob shape. E-Snob pulses calibrated for 270° are actually as suitable as E-Burp2 and Gaussian pulses for performing a Spatial Frequency Encoding excitation step (see for instance *eMagRes*, 2016, 5, 1377–1382, DOI: 10.1002/9780470034590.emrstm1528). The authors apologize for this error.

The Royal Society of Chemistry apologises for these errors and any consequent inconvenience to authors and readers.

