



NJC

New Journal of Chemistry
www.rsc.org/njc

A journal for new directions in chemistry



ISSN 1144-0546



PAPER

Sudeshna Chandra *et al.*
Fabrication of a label-free electrochemical immunosensor using a redox active ferrocenyl dendrimer





Cite this: *New J. Chem.*, 2016, 40, 9046

Fabrication of a label-free electrochemical immunosensor using a redox active ferrocenyl dendrimer

Sudeshna Chandra,^{*a} Christian Gäbler,^b Christian Schliebe,^b Heinrich Lang^b and Dharendra Bahadur^c

We report an IgG (=immunoglobulin) electrochemical immunosensor using a newly synthesized redox-active ferrocenyl dendrimer of generation 2 (**G₂Fc**) as a voltammetric transducer. The ferrocenyl dendrimer N(CH₂CH₂C(O)NHCH₂CH₂NHC(O)Fe(η⁵-C₅H₄)(η⁵-C₅H₅))(CH₂CH₂N(CH₂CH₂C(O)NHCH₂CH₂NHC(O)Fe(η⁵-C₅H₄)(η⁵-C₅H₅))₂)₂ (**G₂Fc**) was used as a functional moiety to immobilize the antibody on the surface of the electrode. A sandwich immunosensor of the type IgG/Bovine serum albumin (BSA)/anti-IgG/**G₂Fc**/glassy carbon electrode (GCE) was fabricated. The electrochemical properties of **G₂Fc** were thoroughly studied in aqueous and non-aqueous electrolytes with varying scan rates. The incubation time was optimized for better analytical performance of the immunosensor. It is found that the developed amperometric immunosensor is sensitive to a concentration of IgG as low as 2 ng mL⁻¹.

Received (in Montpellier, France)
15th March 2016,
Accepted 30th August 2016

DOI: 10.1039/c6nj00830e

www.rsc.org/njc

Introduction

Electro-active dendrimers containing multiple redox units have generated great interest for their use as electrode modifiers for the development of biosensors.¹ A wide variety of dendrimers have been fabricated with centrally and peripherally located redox-active units. For example, dendrimers functionalized with ferrocenyl end-grafted building blocks and their derivatives have shown promising results.² Ferrocenes are capable of transferring multiple numbers of electrons under the same potential, and therefore, exhibit stable reversible redox properties. In ferrocenyl functionalized dendrimers, the redox centers are usually separated from one another, and therefore, show negligible electrostatic effects.³ Recently, it was shown that even for a 14 000 ferrocenyl bearing dendrimer, no electrostatic effects were observed.⁴ However, in small rigid dendrimers, there can be some electrostatic effects seen between the ferrocenyl end-grafted moieties due to a short inter-site distance between the terminal groups.^{5,6} In such cases, the cyclic voltammogram wave is fragmented into smaller ones indicating multiple step electron transfer processes. The electron transfer between redox-active dendrimers and a variety of other biomolecules plays a key role

in promising applications in, for example, the area of biosensors.⁷

Cancer biomarkers are molecules, generally proteins, that are over-expressed in cancer cells and whose measurement or identification plays an important role in the diagnosis of cancer.⁸ These biomarkers of cancer encompass physical symptoms, mutated DNAs and RNAs, secreted proteins and serum levels. Therefore, the detection of cancer biomarkers at an early stage can provide crucial information for foundational research of life sciences, clinical diagnosis and prevention of disease. Elevated concentration of biomarkers in body fluid is an early indication of some type of cancerous disease and among all the biomarkers, IgG are the most common and extensively used clinical cancer biomarkers for several carcinomas like human cervical, breast, liver, colon and lung cancer.⁹ All these carcinomas produce IgG in both cytoplasmic and secreted forms. The most important goal of a cancer biomarker is the detection of the smallest number of tumor cells in human fluid.^{8,10} Thus, to realize the potentiality of biomarkers in diagnosis, novel bioanalytical techniques must be developed with improved accuracy and sensitivity.

A sandwich immunosensor is a major analytical technique that can be used for the sensitive and selective detection of cancer biomarkers. However, there are only a handful of studies available on this topic.¹¹ Electrochemical immunosensors are fabricated by immobilizing the antigen-antibody on the surface of the electrode transducer and the conventional methods are physical adsorption, covalent binding and polymer entrapment.¹² Qiao *et al.*¹³ reported a change in the electrochemical signal *i.e.*,

^a Department of Chemistry, Sunandan Divatia School of Science, NMIMS University, Vile Parle (West) Mumbai-400056, India. E-mail: sudeshna.chandra@nmims.edu; Tel: +91 22-42199956

^b Technische Universität Chemnitz, Faculty of Natural Sciences, Institute of Chemistry, Inorganic Chemistry, 09107 Chemnitz, Germany

^c Department of Metallurgical Engineering and Material Science, Indian Institute of Technology Bombay, Mumbai-400076, India



a decrease in the redox peak intensity upon the adsorption of an antibody on the electrodes due to its non-electrochemical activity which partially blocked the electron transfer between the electrode and $[\text{Fe}(\text{CN})_6]^{3-/4-}$. After the antigen–antibody reaction, a further decrease of the redox peaks was observed. Thus, there is a change in the electrochemical signal. The analyte is measured through the detection of electrochemical signals produced due to antigen–antibody reactions.¹³ Electrochemical immunosensors are known for fast and sensitive analysis, are inexpensive and involve no prior sample pre-treatment. Electrochemical amperometric immunosensors are even more promising because they can achieve a very low detection limit with high sensitivity. Amperometry is gaining attention due to its role in the development of immunosensors and therefore a novel immobilization technique with enhanced sensitivity and selectivity with less response time is of considerable interest.¹⁴

A number of electroactive materials have been investigated as signaling species like metal nanoparticles and quantum dots.^{15–18} Various electroactive materials, transducer techniques and their response to different cancer biomarkers are summarized in Table 1. An amperometric immune-sensor based on gold nanoparticles was reported by Liang and Mu in which the nanoparticles were used to modify the electrodes for the determination of paraoxon, a carcinoembryonic antigen and interleukin-6.¹⁹ Wu *et al.*²⁰ reported on the use of electrochemical biosensors based on gold nanoparticles/poly(lactide) nanocomposites for sensitive recognition of leukemia K562 cells.

Among various signaling species,²³ dendrimers are recognized as promising candidates for bioconjugation for the construction of electrochemical biosensors.²⁴ Highly branched dendrimers provide structural homogeneity, controlled composition, comparable size to biomolecules, internal porosity and multiple functional groups for conjugation reactions.²⁵ Further, the end-grafted groups can be synthetically manipulated for obtaining a desired structure according to the functional requirement. To this end, the surface of polyamidoamine (PAMAM)-dendrimers can be modified with redox-active ferrocenyl building blocks to obtain well-structured end-grafted ferrocenyl dendrimers that can be used as signaling species for fabricating electrochemical biosensors.²⁴ The ferrocenyl units are multiple and electrochemically equivalent redox units and can act as an electron “pool”, and both Fe(II) and Fe(III) forms are stable on the electrochemical time scale. The ferrocenyl-terminated PAMAM

(Fc) dendrimer is capable of transferring multiple numbers of electrons under the same applied potential. In this way, redox-active dendrimers can be used for dual purposes: one for building a film over the electrode and the other for immobilizing biomolecules for sensing.²⁵

Herein, a novel electrochemical immunosensor for the detection of cancer biomarkers is proposed, which is based on a redox-active ferrocenyl dendrimer on a glassy carbon electrode (GCE). The electrochemical response by cyclic voltammetry (CV) has been used to detect over-expressed immunoglobulin (IgG) in the presence of a specific probe. The CV measurements were used to characterize each step of surface modification resulting due to the binding of IgG to anti-IgG. The change in voltammetric response, due to the formation of an immune complex produced by the antibody–antigen reaction was used for the sensitive detection of cancer biomarkers by amperometry.¹⁴ The reported detection technique is simple, convenient, reproducible and sensitive. The redox active **G₂Fc** showed promising results in the recognition of cancer biomarkers at a very low concentration and at an early stage (Scheme 1).

Experimental


Materials and instruments

All reactions were carried out under a nitrogen or argon atmosphere using standard Schlenk techniques. Dichloromethane was used from a MBRAUN (MB-SPS 800) solvent drying and purification system. As starting materials, the H₂N end-grafted dendrimer (**1**)²⁶ and ferrocene carboxylic acid fluoride (**2**)²⁷ were prepared according to the published procedures.²⁸ All other reagents were used as received from commercial suppliers. The NMR spectrum was recorded using a Bruker Avance III 500 spectrometer. The ¹H NMR spectrum was recorded at 500.3 MHz and the ¹³C{¹H} one at 125.7 MHz, respectively. Chemical shifts are reported in δ units (parts per million) downfield from tetramethylsilane with the solvent as a reference signal (¹H NMR: standard internal CDCl₃, δ 7.26; ¹³C{¹H} NMR: standard internal CDCl₃, δ 77.16). High resolution mass spectra were recorded using a Bruker Daltonik micrOTOF-QII spectrometer (ESI-TOF). Elemental analysis was carried out using a Thermo Flash AE 1112 series instrument.

Table 1 Detection of cancer biomarkers and their reported sensitivity by various nanomaterials

Sl. no.	Sensing materials	Techniques	Cells/cancer markers	Sensitivity/response	Ref.
1	TiO ₂ /CNT	CV	Leukemia	Not mentioned	16
2	Fe ₃ O ₄ /PB	DPV	IgG	5 nM	15
3	Nano-Au/NiHCF Nanoparticles	CV	CEA	0.1 ng mL ⁻¹	14
4	Au/poly(lactide) nanocomposites	EIS	Leukemia	800 cells per mL	20
5	Graphene/Au NPs	EIS	α -Fetoprotein	0.1 ng mL ⁻¹	34
6	Chitosan/Au NPs	Amperometry	CEA	0.08 ng mL ⁻¹	46
7	CdFe ₂ O ₄ -SiO ₂	Amperometry	IgG	0.18 μ g mL ⁻¹	12
8	PAMAM dendrimer	QCM	Human IgG	7 nM	21
9	AuNPs/PAMAM	DPV	PSA	0.5 ng mL ⁻¹	22





Scheme 1 Representation of the immunosensor based on ferrocenyl dendrimer (G_2Fc).

Electrochemistry

Measurements in an aqueous electrolyte. Cyclic voltammetry (CV) and amperometric experiments were performed on a CH Instrument Work Station 660D at ambient temperature. A 3-electrode system was used for the measurements with a bare and working glassy carbon electrode (GCE), a platinum wire as a counter electrode and an Ag/AgCl reference electrode. The electrolytes were purged with N_2 for at least 15 min for complete removal of O_2 . Purging with N_2 was continued during measurements. The electrochemical response of the immunosensor towards IgG is based on the current response of the redox-active material (concentration: 1 mg mL of G_2Fc) after the formation of an antibody–antigen complex.²⁹ Amperometry was carried out in an electrochemical cell containing 0.1 M PBS and 1 mM H_2O_2 at +0.3 V under constant stirring. The current–time profile was recorded after a steady state current was achieved.

Measurements in non-aqueous electrolyte. Electrochemical measurements of G_2Fc (1.0 mmol L^{-1} using $[N^rBu_4][B(C_6F_5)_4]$) in dichloromethane were performed in a dried, argon purged cell at 25°C using a Radiometer Voltalab PGZ 100 electrochemical workstation interfaced with a personal computer. As supporting electrolyte 0.1 mol L^{-1} solutions of $[N^rBu_4][B(C_6F_5)_4]$ were used. A three-electrode cell containing a Pt auxiliary electrode, a glassy carbon working electrode (surface area 0.031 cm^2) and an Ag/Ag⁺ (0.01 mmol L^{-1} $[AgNO_3]$) reference electrode fixed on a Luggin capillary were applied. The working electrode was pre-treated by polishing on a Buehler microcloth first with 1 mm and then with a $\frac{1}{4}$ mm diamond paste. The reference electrode was constructed from a silver wire inserted into a 0.01 mmol L^{-1} $[AgNO_3]$ and 0.1 mol L^{-1} $[N^rBu_4][B(C_6F_5)_4]$ acetonitrile solution in a Luggin capillary with a vycor tip. This Luggin capillary was inserted in a second Luggin capillary containing a 0.1 mol L^{-1} $[N^rBu_4][B(C_6F_5)_4]$ dichloromethane solution and a vycor tip. Under the same experimental conditions all reduction and oxidation potentials were reproducible within 5 mV. Experimental potentials

were referenced against an Ag/Ag⁺ reference electrode but the presented results are referenced against ferrocene as an internal standard as required by IUPAC.³⁰ To achieve this, each experiment was repeated in the presence of 1 mmol L^{-1} decamethylferrocene (Fc^*). Data were processed on a Microsoft Excel worksheet to set the formal reduction potentials of the FcH/FcH^+ couple to 0.0 V. Under our conditions the Fc^*/Fc^{*+} couple was at $-619 \text{ mV vs. } FcH/FcH^+$, $\Delta E_p = 60 \text{ mV}$, while the FcH/FcH^+ couple itself was at $220 \text{ mV vs. } Ag/Ag^+$, $\Delta E_p = 61 \text{ mV}$.³¹

Anti-goat Immunoglobulin (anti-IgG), Immunoglobulin (IgG) and bovine serum albumin (BSA) were obtained from Sigma Aldrich and stored at -20°C . Phosphate buffer saline (PBS, 0.1 M , pH 7.2) was used as electrolyte for the electrochemical measurements. Solvents and reagents were purchased, were of reagent grade and were used as such.

Synthesis of the ferrocenyl end-grafted dendrimer (G_2Fc)

The polyamidoamine (PAMAM)-based dendrimer **1** was modified^{32,33} and functionalized peripherally by redox-active ferrocenyl units to form a new electro-active functional material, ferrocenyl end-grafted dendrimer (G_2Fc).

Synthesis. 100 mg (0.15 mmol) of the H_2N end-grafted dendrimer $N(CH_2CH_2C(O)NHCH_2CH_2NH_2)(CH_2CH_2N(CH_2CH_2C(O)NHCH_2CH_2NH_2)_2)$ (**1**) (Fig. 1) and 206.6 mg (0.9 mmol) of ferrocene carboxylic acid fluoride were dissolved in 100 mL of anhydrous dichloromethane and stirred for 5 d at 30°C . The reaction mixture was cooled to ambient temperature, the solvent was removed and the crude product was washed five times with acetonitrile (each 20 mL) at 35°C . After drying under vacuum, the title compound was obtained as an orange solid in 46% (118 mg , 0.068 mmol) yield.

¹H NMR ($CDCl_3$). $\delta = 2.25\text{--}2.70$ (m, 28 H, $NCH_2CH_2C(O)$), (NCH_2CH_2N), 3.48 (m, 20 H, $HNCH_2CH_2NH$), 4.20 (s, 25 H, H/C_5H_5), 4.33 (m, 10 H, $H/C_5H_4C(O)$), 4.82 (m, 10 H, $H/C_5H_4C(O)$), 7.30–7.55 (m, 5 H, NH), 7.80–8.30 (m, 5 H, NH) ppm. ¹³C{¹H} NMR ($CDCl_3$): $\delta = 34.1$ ($CH_2C(O)$), 39.5 ($HNCH_2CH_2NH$),






Fig. 1 Synthesis of ferrocenyl end-grafted dendrimers G_2Fc .

40.5 (HNCH₂CH₂NH), 50.5, (NCH₂CH₂C(O)), 51.5 (NCH₂CH₂N), 68.6 (C_α/C₅H₄C(O)), 70.0 (C₅H₅), 70.7 (C_β/C₅H₄C(O)), 75.9 (C_i/C₅H₄C(O)), 171.8 (C(O)), 173.9 (C(O)). HRMS: m/z : calcd for C₈₄H₁₀₃Fe₅N₁₃O₁₀: 867.7430, found 867.7487 [M + H]²⁺; elemental analysis calcd (%) for C₈₄H₁₀₃Fe₅N₁₃O₁₀: C 58.18, H 5.99, N 10.50; found C 57.71, H 6.27 N 10.06.

Fabrication of the immunosensor

Ferrocenyl-functionalized dendrimers G_2Fc (~10 μg) were drop-casted on the GCE and kept for drying under an ambient atmosphere for a few hours. Then, 10 μL of 20 μg mL⁻¹ anti-goat IgG (0.1 M PBS, pH 7.2) was immobilized onto the G_2Fc /GCE. The electrode was incubated at 37 °C for 1 h and the unbound anti-goat IgG was washed away with PBS. The immobilized electrode was then incubated with 10 mg mL⁻¹ of BSA for 60 min at 37 °C to avoid non-specific adsorption by blocking the active sites.³⁴ The modified electrode was stored at 4 °C when not in use.

Results and discussion

G_2Fc dendrimers have been used as a redox responsive receptor system for immobilization of cancer biomarkers. The dendrimers can easily adsorb on the surfaces and hence the modification of electrodes is easier. The interaction of the active recognition units with the ferrocenyl end-grafted dendrimers enhanced the sensing ability of the system not only in solution but also when confined on the electrode surface.

Electrochemical properties of ferrocenyl end-grafted dendrimers (G_2Fc)

The redox properties of G_2Fc have been monitored in aqueous and non-aqueous electrolytes.

In an aqueous electrolyte. The electrochemical behavior of a G_2Fc -modified GCE (=glassy carbon electrode) was investigated in an unstirred buffer solution (pH 7.0, 5 mL) containing 0.1 M KCl at room temperature under an inert atmosphere by cyclic voltammetry (CV). The respective G_2Fc (ferrocenyl = Fc) showed

the oxidation peak at 0.72 V and the reduction peak at 0.33 V in an aqueous electrolyte at a scan rate of 50 mV s⁻¹. The ratio of the oxidation peak current-to-reduction peak current (I_{pa}/I_{pc}) was calculated to be ~1. The peak current increased significantly with a shift in the anodic and cathodic peak towards lower potential, showing the promoting electro-active effect of the ferrocenyl end-grafted dendrimer electrode. The peak-to-peak separation (ΔE_p) of the Fc-dendrimer-coated GCE was calculated to be 390 mV. The single electrochemical process at 0.35 V (Ag/AgCl) is attributed to the ferrocenyl moieties on the dendrimers.²⁸

Subsequently, the influence of the potential scan rate on the cyclic voltammograms of the G_2Fc /GCE electrodes was investigated (Fig. 2a). It was found that there is a shift in the peak potential along with an increase in the peak current as the scan rates increased from 10 to 200 mV s⁻¹. The ΔE_p values are about 400–500 mV and remain fairly constant at lower scan rates, but increase with higher scan rates. This indicates that at a higher scan rate, the charge transfer is not sufficiently fast and the electrochemical reaction is basically charge-transfer controlled and not diffusion controlled (please note that we did not get linearity, when the peak current was plotted against the square root of the scan rate, Fig. 2b). This is indicative of a quasi-reversible system in which the current is controlled by both charge-transfer and mass-transfer. However, the system approaches reversibility as the scan rate increases.

Multiple successive scans were carried out to check the stability of the (G_2Fc)-modified GCE and it was found that the electro-activity is retained throughout the study. Moreover, the redox response was unchanged even after several days.

In a non-aqueous electrolyte. The cyclic voltammograms of G_2Fc , measured in the presence of [ⁿBu₄N][B(C₆F₅)₄] as electrolyte³⁵ at room temperature under an inert atmosphere, are shown in Fig. 3a. The application of [ⁿBu₄N][B(C₆F₅)₄] as supporting electrolyte within electrochemical measurements has been demonstrated.³⁶ The CV measurements were carried out at 20 °C and all potentials were referenced to the FcH/FcH⁺ redox couple.³⁷ The ferrocenyl groups were oxidized at the same






Fig. 2 (a) Cyclic voltammograms of **G₂Fc** in an aqueous electrolyte at varying scan rates, (b) current vs. square root of the scan rate depicting a non-linear response. CVs were measured at room temperature in PBS buffer (pH 7.0, 5 mL) containing 0.1 M KCl.

potential showing one reversible redox event, with differences between the cathodic and the anodic peak potential ΔE_p falling between 200 and 300 mV depending on the scan rate (Fig. 3b). Peak currents are proportional to $\nu^{1/2}$, confirming the essentially reversible character of the ferrocenyl-based oxidation events. The electrostatic effect is sufficiently low because the terminal redox centers are located apart from each other and therefore a single CV wave is observed for all the five Fc redox centers.³⁸

The potential shift (ΔE_p) is higher in an aqueous electrolyte due to diffused electrostatic interactions and more H-bond interactions with the amido groups of the PAMAM dendrimers, respectively.³⁹ Furthermore, the electrochemical (E) and chemical (C) reactions influence the shape of the voltammograms. In the present study, both electron-transfer kinetics and chemical reactions play an important role and based on the above results, it can be concluded that the oxidation of **G₂Fc** proceeds *via* a CE mechanism.⁴⁰ In both aqueous and non-aqueous electrolytes, the occurrence of a single redox process with simultaneous multi-electron transfer at the same potential indicates that the

iron centers of the ferrocenyl units are non-interacting in the mixed-valent dendrimer.⁴¹

Cyclic voltammetric characterization of the immunosensor

Cyclic voltammetry was used to monitor the stepwise assembly of the electrochemical immunosensor for its potential application in biosensing. Cyclic voltammetric studies of the **G₂Fc**-modified glassy carbon electrode in a 0.1 M PBS solution of pH 7.2 showed a pair of well-defined redox peaks. The peak current decreased significantly with the immobilization of anti-goat IgG. After the immunosensor was blocked with BSA, a further decrease in the peak current was observed (Fig. 4). This may be due to the attachment of the protein BSA to the immunosensor.²⁹ It was also found that the pH of the PBS solution affects the electrochemical immunosensor, due to denaturation of the BSA proteins. The current responses of the sensor were investigated at various pH values and the best current response was obtained at pH 7.2. Hence, for all further studies, PBS of pH 7.2 was used as optimum buffer solution.




Fig. 3 (a) Cyclic voltammograms of **G₂Fc** in the non-aqueous electrolyte at varying scan rates, (b) plot of current density versus square root of the scan rate depicting the non-linear response of the system; CVs were measured in (1.0 mmol L⁻¹ using [ⁿBu₄N][B(C₆F₅)₄]) in dichloromethane at room temperature.






Fig. 4 Cyclic voltammograms of the **G₂Fc** (black) recorded in 0.1 M solution of PBS on immobilization with anti-IgG (red), BSA (blue) and IgG (green) at a scan rate of 50 mV s⁻¹.

In addition, the effect of incubation time on the current response of the immunosensor was examined. It was found that the current response decreases with incubation time and levels within 60 min. Therefore, 60 min was set as the optimum incubation time for the immobilization of BSA. Serum albumin is a weak polyacid and can be adsorbed on both positively and negatively charged surfaces depending on the pH. At pH 7.2, BSA (pI = 4.8) is negatively charged, while **G₂Fc** remains positively charged and this leads to the binding of BSA through electrostatic attractive strength.⁴²

The concentration of the antibody plays a vital role in influencing the sensitivity of an immunoassay. Therefore, the effect of anti-IgG antibody concentration was studied with a series of concentrations ranging from 0.5 μg mL⁻¹ to 20 μg mL⁻¹ as a function of the current inhibition ratio, *i.e.*, the relative change in the current response. It was found that with the increase in the concentration of the antibody up to 5 μg mL⁻¹, there was an increase in the current peak which then rapidly decreased as the antibody concentration was increased further. This may be ascribed to the molecular bonding between the antibody and the dendrimers which saturated at 5 μg mL⁻¹. This led to the formation of a film that hindered the electron transmission towards the electrode surface thereby reducing the peak current. The maximum inhibition ratio was attained at a concentration of 5 μg mL⁻¹, which was therefore taken as the optimized concentration for further studies.

The incubation time of an antibody and an antigen is important for achieving a satisfactory sensitivity of an immunoassay. Therefore, antigen-antibody interaction was studied as a function of absorbance in a time scale from 10 to 30 min. It was found that the intensity of the antibody-antigen reaction increased slightly with longer incubation time, and no significant difference was observed between the tests for 20 min and 30 min. Hence, an incubation time of 20 min was taken for all the studies thereafter.

Performances of the immunosensor

The current signal of the fabricated immunosensor BSA/anti-IgG/**G₂Fc**/GCE was further investigated with immobilizing IgG




Fig. 5 Effect of the concentration of IgG on BSA/anti-IgG/**G₂Fc**/GCE.

(Fig. 5). A significant decrease in the current was observed, which may be due to the formation of immune-conjugates that block the tunneling of mass and electron transfer.⁴³ The current signal was found to be directly related to the amount of IgG captured on the electrode surface. With the increase in the concentration of IgG, there is an increase in the amount of immune-conjugates formed on the immunosensor, decreasing the peak current. A good linear relationship between the current response and the IgG concentration in the range 100 to 1000 ng mL⁻¹ with a correlation coefficient of 0.98664 was observed. This is associated with the electron flux involved in the redox reaction at the solution-immunosensor interface. The efficient and fast recognition of IgG may also be due to the electrostatic attraction forces between the carbonyl group of the ferrocenyl dendrimers and the amine groups of IgG.⁴⁴ The hydrophilicity of the dendrimer enhances the adsorption of IgG, prevents denaturation and also exposes the active sites for binding with the complementary IgG.⁴⁵

A control experiment was carried out to determine the degree of non-specific binding of anti-IgG. The control electrode did not produce any considerable signal amperometrically (0.05–0.12 μA), which showed that there was no significant non-specific adsorption of the antibody on the electrode surface. Further, the specificity of the immunosensor was also tested by incubating the BSA/anti-IgG/**G₂Fc**/GCE probe in solutions containing 50 μg mL⁻¹ of BSA and α-feto protein (AFP); however, no change in the peak current was observed. Therefore, it may be concluded that the current response of the BSA/anti-IgG/**G₂Fc**/GCE probe to the target antibody (IgG) is due to a highly specific antigen-antibody interaction.

Amperometric response to IgG

The amperometric current response of BSA/anti-IgG/**G₂Fc**/GCE was recorded in 0.1 M PBS at room temperature. The current response was monitored at +0.3 V vs. Ag/AgCl. Under the optimum experimental conditions, the relationship between the current and the concentration of IgG was measured in a wide concentration range (2–50 ng mL⁻¹). The data indicate






Fig. 6 Amperometric current–time response profile of IgG on BSA/anti-IgG/G₂Fc/GCE. Inset shows the linearity of current response with increasing concentration of IgG.

that the immunosensors are capable of estimating different concentrations of IgG and the amperometric response increased with the increase in the IgG concentration. As can be seen from Fig. 6, there is an immediate increase in the steady state current, which is due to the binding of IgG with the anti-IgG immobilized on G₂Fc at the electrode surface. The steady state currents (inset of Fig. 6) exhibited linearity in the range of 5–50 ng mL^{−1} IgG and the linear dependencies of IgG yielded an equation of I (μA) = $(-0.2046 \pm 0.00003) + (-0.02016 \pm 0.000001)[\text{IgG}]$ (ng mL^{−1}) with a correlation coefficient of 0.9815. The sensitivity of the developed immune-sensor was 0.020 ± 0.00003 μA mL ng^{−1} and the limit of detection was ~ 2 ng mL^{−1}, based on three measurements for the standard deviation (95% confidence level, signal-to-noise ratio $k = 3$, number of repeated measurement = 3). The results also show that the electrode retains its ability to exhibit linear amperometric responses over a period of time without denaturing the antibody.

The stability of the immunosensor was evaluated over a period of one month. A group of immunosensors was kept at 4 °C under identical conditions and was periodically checked for their amperometric response. No apparent change in the current response was observed.

As far as the working of an electrochemical immunosensor is concerned, immobilization of the sensing molecules on the electrode surface remains the crucial step. The immobilization process should be simple, fast and effective and efforts should be made to optimize the loading of the receptor molecules. Dendrimers on one hand provide porosity to the system and on the other hand provide functionalities to entrap the biomolecule. Tailoring of redox-active dendrimers provides enhanced electro-activity to the system and enlarges the sensor surface for binding the antibodies. This would result in a higher biosensor signal.⁴⁶ It may be assumed that both electron transfer and diffusion contribute to the signal transformation between the electroactive dendrimers and the antibody.

Conclusion

Within this study a redox-active ferrocenyl dendrimer-modified glassy carbon electrode for the detection of cancer biomarkers is described. The electrochemical behavior of redox-active G₂Fc depends on the pH value and the type of electrolyte. The attachment of the biomarker and the binding events were studied by amperometry and cyclic voltammetry, respectively, with the results showing that the binding events result in a decrease in the exposed surface area giving rise to an increase in resistance and a lowering of the current intensity. The immunosensor requires very small amounts of the sample and exhibits shorter response times with good sensitivity towards IgG cancer biomarkers. The developed redox-active dendrimer-based immunosensor has the potential for further development into practical cancer diagnosis systems, which can be used in point-of-care quantitative tests for cancer biomarkers.

Acknowledgements

The authors are grateful to the Deutsche Forschungsgemeinschaft (DFG), Germany and the Department of Science and Technology (DST), India vide Reference No. SR/WOS-A/CS-45/2010 for financial support.

References

- 1 S. Sengupta, *Tetrahedron Lett.*, 2003, **44**, 7281–7284.
- 2 S. M. Grayson and J. M. J. Fréchet, *Chem. Rev.*, 2001, **101**, 3819–3868.
- 3 C. Ornelas, J. Ruiz and D. Astruc, *Organometallics*, 2009, **28**, 4431–4437.
- 4 C. Ornelas, J. Ruiz, C. Belin and D. Astruc, *J. Am. Chem. Soc.*, 2009, **131**, 590–601.
- 5 A. K. Diallo, J. C. Daran, F. Varret, J. Ruiz and D. Astruc, *Angew. Chem., Int. Ed.*, 2009, **48**, 3141–3145.
- 6 A. K. Diallo, C. Absalon, J. Ruiz and D. Astruc, *J. Am. Chem. Soc.*, 2011, **133**, 629–641.
- 7 D. Astruc, *Nat. Chem.*, 2012, **4**, 255–267.
- 8 J. F. Rusling, C. V. Kumar, J. S. Gutkinde and V. Patel, *Analyst*, 2010, **135**, 2496–2511.
- 9 X. Qiu, X. Zhu, L. Zhang and Y. Mao, *Cancer Res.*, 2003, **63**, 6488–6495.
- 10 S. Chandra, N. Barola and D. Bahadur, *Chem. Commun.*, 2011, **47**, 11258–11260.
- 11 M. Yang, A. Javadi and S. Gong, *Sens. Actuators, B*, 2011, **155**, 357–360.
- 12 Z. M. Liu, H. F. Yang, Y. F. Li, Y. L. Liu, G. L. Shen and R. Q. Yu, *Sens. Actuators, B*, 2006, **113**, 956–962.
- 13 L. Qiao, X. Wang and X. Sun, *Int. J. Electrochem. Sci.*, 2014, **9**, 1399–1414.
- 14 Y. R. Yuan, R. Yuan, Y. Q. Chai, Y. Zhuo and X. M. Miao, *J. Electroanal. Chem.*, 2009, **626**, 6–13.
- 15 D. Tang, J. Tang, B. Su, H. Chen, J. Huang and G. Chen, *Microchim. Acta*, 2010, **171**, 457–464.



- 16 Q. Shen, S. K. You, S. G. Park, H. Jiang, D. Guo, B. Chen and X. Wang, *Electroanalysis*, 2008, **20**, 2526–2530.
- 17 Y. Zhu, W. C. Alvin Koh and Y. B. Shim, *Electroanalysis*, 2010, **22**, 2908–2914.
- 18 G. Liu, J. Wang, J. Kim, M. Jan and G. Collins, *Anal. Chem.*, 2004, **76**, 7126–7130.
- 19 R. Liang and W. J. Mu, *Anal. Chim. Acta*, 2006, **580**, 128–135.
- 20 X. Wu, H. Jiang, J. Zheng, X. Wang, Z. Gu and C. Chen, *J. Electroanal. Chem.*, 2011, **656**, 174–178.
- 21 L. Svobodová, M. Šnejdárková, V. Polohová, I. Grman, P. Rybrá and T. Hianik, *Electroanalysis*, 2006, **18**, 1943–1949.
- 22 B. Kavosia, A. Salimia, R. Hallaja and K. Amaniam, *Biosens. Bioelectron.*, 2014, **52**, 20–28.
- 23 (a) W. Chen, D. H. Xu, L. Q. Liu, C. F. Peng, Y. Y. Zhu, W. Ma, A. Bian, Z. Li and L. B. Wang, *Anal. Chem.*, 2009, **81**, 9194–9198; (b) S. Gupta, S. Huda, P. K. Kilpatrick and O. D. Velez, *Anal. Chem.*, 2007, **79**, 3810–3820; (c) G. S. Lai, F. Yan and H. X. Ju, *Anal. Chem.*, 2009, **81**, 9730–9736.
- 24 S. J. Kwon, E. Kim, H. Yang and J. Kwak, *Analyst*, 2006, **131**, 402–406.
- 25 H. C. Yoon, M. Y. Hong and H. S. Kim, *Anal. Chem.*, 2000, **72**, 4420–4427.
- 26 S. Dietrich, A. Nicolai and H. Lang, *J. Organomet. Chem.*, 2011, **696**, 739–747.
- 27 T. H. Galow, J. Rodrigo, K. Cleary, G. Cooke and V. M. Rotello, *J. Org. Chem.*, 1999, **64**, 3745–3746.
- 28 E. R. de Jong, E. Manoury, J. C. Daran, C. O. Turrin, J. Chiffre, A. Sournia-Saquet, W. Knoll, J. P. Majoral and A. M. Caminade, *J. Organomet. Chem.*, 2012, **718**, 22–30.
- 29 A. Ray and S. Ghorai, *Tetrahedron Lett.*, 2011, **52**, 2980–2983.
- 30 G. Gritzner and J. Kuta, *Pure Appl. Chem.*, 1984, **56**, 461–466.
- 31 A. Nafady and W. E. Geiger, *Organometallics*, 2008, **27**, 5624–5631.
- 32 S. Dietrich, S. Chandra, C. Georgi, S. Thomas, D. Makarov, S. Schulze, M. Hietschold, M. Albrecht, D. Bahadur and H. Lang, *Mater. Chem. Phys.*, 2012, **132**, 292–299.
- 33 S. Chandra, M. D. Patel, H. Lang and D. Bahadur, *J. Power Sources*, 2015, **280**, 217–226.
- 34 K. J. Huang, D. J. Niu, J. Y. Sun and J. J. Zhu, *J. Electroanal. Chem.*, 2011, **656**, 72–77.
- 35 (a) W. E. Geiger and F. Barrière, *Acc. Chem. Res.*, 2010, **43**, 1030–1039; (b) F. Barrière and W. E. Geiger, *J. Am. Chem. Soc.*, 2006, **128**, 3980–3989; (c) F. Barrière, N. Camire, W. E. Geiger, U. T. Mueller-Westerhoff and R. Sanders, *J. Am. Chem. Soc.*, 2002, **124**, 7262–7263.
- 36 (a) H. J. Gericke, N. I. Barnard, E. Erasmus, J. C. Swarts, M. J. Cook and M. A. S. Aquino, *Inorg. Chim. Acta*, 2010, **363**, 2222–2232; (b) E. Fourie, J. C. Swarts, D. Lorcy and N. Bellec, *Inorg. Chem.*, 2010, **49**, 952–959; (c) J. C. Swarts, A. Nafady, J. H. Roudebush, S. Trupia and W. E. Geiger, *Inorg. Chem.*, 2009, **48**, 2156–2165; (d) D. Chong, J. Slote and W. E. Geiger, *J. Electroanal. Chem.*, 2009, **630**, 28–34; (e) V. N. Nemykin, G. T. Rohde, C. D. Barrett, R. G. Hadt, J. R. Sabin, G. Reina, P. Galloni and B. Floris, *Inorg. Chem.*, 2010, **49**, 7497–7509; (f) V. N. Nemykin, G. T. Rohde, C. D. Barrett, R. G. Hadt, C. Bizzarri, P. Galloni, B. Floris, I. Nowik, R. H. Herber, A. G. Marrani, R. Zanoni and N. M. Loim, *J. Am. Chem. Soc.*, 2009, **131**, 14969–14978; (g) A. Hildebrandt, S. W. Lechrich, D. Schaarschmidt, R. Jaeschke, K. Schreiter, S. Spange and H. Lang, *Eur. J. Inorg. Chem.*, 2012, 1114–1121; (h) A. Hildebrandt, D. Schaarschmidt, R. Claus and H. Lang, *Inorg. Chem.*, 2011, **50**, 10623–10632; (i) A. Hildebrandt and H. Lang, *Organometallics*, 2013, **32**, 5640–5653; (j) U. Pfaff, A. Hildebrandt, D. Schaarschmidt, T. Rüffer, P. J. Low and H. Lang, *Organometallics*, 2013, **32**, 6106–6117; (k) A. Hildebrandt and H. Lang, *Dalton Trans.*, 2011, **40**, 11831–11837.
- 37 D. Miesel, A. Hildebrandt, M. Korb, P. J. Low and H. Lang, *Organometallics*, 2013, **32**, 2993–3002.
- 38 D. Astruc, C. Ornelas and J. Ruiz, *Acc. Chem. Res.*, 2008, **41**, 841–856.
- 39 O. Reynes, T. Gulon, J. C. Moutet, G. Royal and E. Saint-Aman, *J. Organomet. Chem.*, 2002, **656**, 116–119.
- 40 C. A. Nijhuis, B. A. Boukamp, B. Jan Ravoo, J. Huskens and D. N. Reinhoudt, *J. Phys. Chem. C*, 2007, **111**, 9799–9810.
- 41 M. Zamora, S. Herrero, J. Losada, I. Cuadrado, C. M. Casado and B. Alonso, *Organometallics*, 2007, **26**, 2688–2693.
- 42 L. J. Feng, X. H. Zhang, D. M. Zhao and S. F. Wang, *Sens. Actuators, B*, 2011, **152**, 88–93.
- 43 Y. Kong, J. Li, S. Wu, W. Cheng, R. K. Rana and J. J. Zhu, *Sens. Actuators, B*, 2013, **183**, 187–193.
- 44 P. Thevenot, W. Hu and L. Tang, *Curr. Top. Med. Chem.*, 2008, **8**, 270–280.
- 45 D. Kim and A. E. Herr, *Biomicrofluidics*, 2013, **7**, 041501.
- 46 Y. Liu, R. Yuan, Y. Chai, C. Hong, K. Liu and S. Guan, *Microchim. Acta*, 2009, **167**, 217–224.

