



Cite this: *Analyst*, 2017, **142**, 3666

## Development of lithium attachment mass spectrometry – knudsen effusion and chemical ionisation mass spectrometry (KEMS, CIMS)†

A. Murray Booth,<sup>a</sup> Thomas J. Bannan, \*<sup>a</sup> Med Benyazzar,<sup>b</sup> Asan Bacak,<sup>a</sup> M. Rami Alfara,<sup>a,c</sup> David Topping<sup>a</sup> and Carl J. Percival‡<sup>a</sup>

Lithium ion attachment mass spectrometry provides a non-specific, non-fragmenting, sensitive and robust method for the detection of volatile species in the gas phase. The design, manufacture and results of lithium based ion attachment ionisation sources for two different mass spectrometry systems are presented. In this study trace gas analysis is investigated using a modified Chemical Ionization Mass Spectrometer (CIMS) and vapour pressure measurements are made using a modified Knudsen Effusion Mass Spectrometer (KEMS). In the Li<sup>+</sup> CIMS, where the Li<sup>+</sup> ionization acts a soft and unselective ionization source, limits of detection of 0.2 ppt for formic acid, 15 ppt for nitric acid and 120 ppt for ammonia were achieved, allowing for ambient measurements of such species at atmospherically relevant concentrations. In the first application of Lithium ion attachment in ultra-high vacuum (UHV), vapor pressures of various atmospherically relevant species were measured with the adapted KEMS, giving measured values equivalent to previous results from electron impact KEMS. In the Li<sup>+</sup> KEMS vapour pressures <10<sup>-3</sup> mbar can be measured without any fragmentation, as is seen with the initial electron impact (EI) set up, allowing the vapor pressure of individual components within mixtures to be determined.

Received 11th July 2017,  
Accepted 1st September 2017

DOI: 10.1039/c7an01161j

rsc.li/analyst

## Introduction

Mass spectrometry has provided considerable insight into atmospheric science, notably in the fields of gas kinetics, trace gas analysis and aerosol composition elucidation. Mass spectrometry has been used in aircraft measurements, field campaigns<sup>1</sup> and laboratory studies<sup>2</sup> and typically consists of an ionizer, an analyser and a detector. The choice of ion source for the different applications each have advantages and drawbacks. Chemical ionization (CI),<sup>3</sup> for example, has proved popular for trace gas analysis and gas kinetics, providing a source of ions at high pressures (including up to atmospheric pressure) which is highly specific to the target analyte. However, it requires suitable ionization reactions for the analyte to detect and is thus not universal. Electron impact

(EI) has proved a consistently popular as an ionization method, through the provision of high sensitivity and non-selectivity, albeit with the twin drawbacks of including the requirements for high vacuum (10<sup>-6</sup> torr) and the inherent fragmentation of gas analytes that results in complicated mass spectra.

In order to avoid some of the inherent drawbacks of EI and CI ionization, an alternative method (ion attachment) has previously been developed by Fujii *et al.*<sup>4</sup> It offers a non-specific, non-fragmenting ionisation method resulting in mass spectra consisting of adduct ions formed from the gas-phase attachment of a primary ion and neutral gas species. Thermionic emission of a primary ion, such as lithium ions, which exhibit strong affinity to any neutral species in the gas phase produces adducts ions which consist of analytes which have been “captured” by lithium ions at low kinetic energy. This yields unfragmented adducts which are then easily detected. Adduct ion formation is obtained when a lithium ion Li<sup>+</sup> attaches to a sample neutral/molecule (M), producing a single stable molecular ion adduct as follows:



Under high pressure regimes (10<sup>-1</sup> to 10<sup>-2</sup> mbar), a neutral bath gas (g) such as nitrogen, argon, or methane is used to

<sup>a</sup>School of Earth, Atmospheric and Environment Science, University of Manchester, UK. E-mail: Thomas.bannan@manchester.ac.uk

<sup>b</sup>Photon Science Institute, University of Manchester, UK


<sup>c</sup>National Centre for Atmosphere Science, UK

†Electronic supplementary information (ESI) available. See DOI: 10.1039/c7an01161j

‡Current address: Jet Propulsion Laboratory, 4800 Oak Grove Drive, Pasadena, CA 91109.



carry ions towards the detection region whilst acting as a third body during the ion-sample attachment process to help remove some of the excess heat or collision energy. This leads to a more stable adduct which can be more readily detected and improves sensitivity:



Fujii and Ohta<sup>5</sup> investigated several lithium bearing materials ( $\text{Li}_2\text{O}$ ,  $\text{Li}_2\text{SiO}_3$ ,  $\text{Li}_3\text{PO}_4$ , zeolite A and  $\beta$ -eucryptite) as potential sources for lithium ion production.  $\beta$ -eucryptite was shown to provide the highest emission current densities ( $2 \times 10^{-6} \text{ A cm}^{-2}$ ) and is the material used by Fujii *et al.*,<sup>4</sup> Selvin and Fujii,<sup>6</sup> Nakamura *et al.*,<sup>7</sup> and Edwards *et al.*<sup>8</sup>

Presented here are initial results from the testing and construction of lithium-ion attachment sources for two different mass spectrometers. First, a Chemical Ionization Mass Spectrometer (CIMS) used for making trace gas phase measurements will be modified. Use of the  $\text{Li}^+$  ionization source in this application will allow for a non-specific, non-fragmenting ionization method for trace gas detection. Secondly, in the first application of Lithium ion attachment in ultra-high vacuum (UHV), the Knudsen Effusion Mass Spectrometer (KEMS) will be used for measuring vapour pressures. The primary advantage of the  $\text{Li}^+$  KEMS over the standard EI ionization KEMS is that individual peaks represent the vapour pressure of individual compounds, in comparison to EI where the sum of the total peaks is used to extract data and no single component data from mixtures is available.  $\text{Li}^+$  therefore will allow the vapor pressure of individual components within mixtures to be probed.

## Materials and methods

### Filament production

Fujii *et al.*,<sup>4</sup> Selvin and Fujii,<sup>6</sup> Nakamura *et al.*,<sup>7</sup> and Edwards *et al.*,<sup>8</sup> have made  $\beta$ -eucryptite from heating a crushed mix of  $\text{Al}_2\text{O}_3:\text{SiO}_2:\text{Li}_2\text{O}$  in a 1:1:1 molar ratio. In this work filaments were using  $\text{Al}_2\text{O}_3:\text{SiO}_2:\text{Li}_2\text{O}$  and  $\text{Al}_2\text{O}_3:\text{SiO}_2:\text{Li}_2\text{CO}_3$ , both with 1:1:1 molar ratios. The basic filament consists of the  $\beta$ -eucryptite material, which has been fused over a metallic conductor or gauze and heating this structure using electrical currents results in the thermionic emission of  $\text{Li}^+$  ions. The heat source that used in this study is produced by resistively heating the material through use of DC currents at low voltage. The specific design for the filaments for the CIMS and KEMS will be discussed in their respective sections.

### Chemical ionisation mass spectrometer

The CIMS system was constructed and designed by the Georgia Institute of Technology and has been described in detail by Nowak *et al.*<sup>9</sup> It is designed to sample gases from high pressure (up to atmospheric) using a differentially pumped arrangement and this specific instrument has been used for both ground and airborne studies.<sup>10,11</sup> The differential pumping system has three main regions; an ion-molecule

region where sample ionisation takes place with a pressure between 10 and 0.1 mbar. This is linked *via* a biased pinhole to the collision dissociation chamber (CDC) where there is a DC electric field to pull apart weakly bound cluster ions, the pressure in this region is around  $10^{-2}$  mbar. This is then linked *via* another biased pinhole and an octopole ion guide to the final region of the instrument which houses the quadrupole mass analyser. The quadrupole filters ions onto a channeltron detector maintained at a pressure of  $10^{-5}$  mbar.

For the CIMS filament, a source was constructed using off the shelf KF40 components. The whole assembly was fitted in the ion-molecule region of the CIMS instrument (Fig. 1). The filament wire, the details of which are discussed later, was mounted using oxygen free high purity copper clamps. The clamps were connected to a linear motion feedthrough *via* insulated alumina standoffs allowing the source to be moved relative to a biased pinhole. This helps ions to be transmitted to the next section of the instrument. Standard high current electrical KF feedthroughs are used to pass low voltage DC current to the filament. Currents of up to 20 Amperes can be produced using a regulated power supply (Manson EP 907). A neutral gas flow of nitrogen is maintained between the emitter or ion source region and the sample area. This allows for thermalisation and easier transmission of Lithium ions and facilitates the attachment of such ions onto the sample neutral gas species.

For the CIMS filament production, in the first instance, the filament wire was high purity 0.25 mm diameter tungsten wire from Advent materials formed into a spiral. Tungsten was chosen by virtue of its extremely high melting point (3422 °C). However, in the presence of oxygen it rapidly oxides at high temperatures meaning it can only be used under vacuum or an inert atmosphere. Filaments were also produced using platinum (melting point 1768 °C) and iridium (melting point 2466 °C) wire, both far more prone to burning out in comparison to the tungsten. Platinum in particular is quite brittle and the act of forming a spiral results in sufficient defects in the metal to significantly reduce the lifetime of the filament and thus an inappropriate metal for a filament. All the wire filaments were produced by making a slurry of the  $\beta$ -eucryptite pre mix using some liquid (usually water) and painting this onto the wire with a paintbrush, a small current was then passed through the wire (~2 amps) to drive off the solvent and dry the  $\beta$ -eucryptite paste onto the wire. Once under vacuum




Fig. 1 Schematic for CIMS filament holder.



the level of heating required for thermionic emission ( $\sim 4\text{--}6$  amps) is sufficient to melt the  $\beta$ -eucryptite pre mix paste onto the wire. An improved filament production method was also used consisting of a mesh, rather than a wire to provide a much larger area for lithium emission and hence a large number of emitted lithium ions. A platinum–iridium (Pt90:Ir10) mesh ( $>99.9\%$  purity 150 mesh, woven from 0.043 mm wire sourced from Alfa-Aesar) was trialed to avoid the mechanical brittleness associated with pure platinum. Dry  $\beta$ -eucryptite pre mix was placed on the mesh and was then heated sufficiently to produce a ‘glassy mix’. The  $\text{Al}_2\text{O}_3:\text{SiO}_2:\text{Li}_2\text{CO}_3$  mix releases excess  $\text{CO}_2$  upon heating and yields a transparent white/greyish glass structure.<sup>12</sup> To avoid filament ‘burnout’ the temperature of the wires and meshes were measured using an optical pyrometer (Land Instruments Cyclops 100) to obtain the surface temperature at various currents and pressures. This yielded a safer method for undertaking measurements and prevented heating any part of the holding metallic wire above its melting point.

The CIMS ion optics were altered and optimised from normal chemical ionization mode to allow for the transmission of primary ions from the source (Table SM1†). The settings for optimisation of the primary ion counts are slightly different from those for adduct counts. The optimum voltages for the first pinhole/aperture (1 mm diameter) towards the exit of the ion-molecule region are strongly dependent on the operating pressure (Table SM2 and Fig. SM1†). A higher pressure requires a higher accelerating voltage to the first pinhole due to the increased mean free path in the bath gas which is balanced against an increased adduct stability from thermalisation with the bath gas ( $\text{N}_2$  removing the excess energy from the lithium ion – sample molecule collision).

### Knudsen effusion mass spectrometer

The KEMS has been described in detail in Booth *et al.*,<sup>13,14</sup> and used recently in Bannan *et al.*,<sup>15</sup> with a schematic of the instrument being shown in Fig. 2. The core of the system is a Balzers-Pfiefer QMA 410 quad analyser with a cross-beam electron impact ion source for sampling molecular beams of organics. The detector is a secondary electron multiplier (SEM) and the




Fig. 2 Schematic of the Knudsen effusion mass spectrometer.

base pressure of the system is  $10^{-7}$  mbar. This is the same analyzer as those on aerodyne’s quad-aerosol mass spectrometer (AMS)<sup>16</sup> so the source could, in principle, be fitted onto an AMS. By swapping the ionization source in the KEMS from EI to  $\text{Li}^+$ , specific and identifiable individual peaks in the mass spectrum will represent the vapour pressure of individual compounds, in comparison to EI where the sum of the total peaks is used to extract data. The  $\text{Li}^+$  KEMS will therefore allow the vapor pressure of individual components within mixtures to be probed. Previous discussions on potential for sources of uncertainty in the KEMS using the EI source include varying sample phase state, statistical noise, ionisation cross sections or lack of appropriate calibration standard in multiple volatility ranges, all described in detail by Booth *et al.* (2009).<sup>13</sup>

For the KEMS filament a standard Balzers-Pfiefer cross-beam electron impact source was adapted.  $\beta$ -Eucryptite coated tungsten (0.25 mm diameter,  $>99.9\%$  purity, Sigma Aldrich) and iridium (0.25 mm diameter,  $>99.9\%$  purity, Advent Metals) wires were spot welded between the two electrodes of the Balzers-Pfiefer crossbeam electron impact source with a Wehnelt repeller around the wire. The assembly is mounted on an alumina holder previously used to hold an electron impact filament. As with the CIMS, the ion optics were first optimized for  $\text{Li}^+$  primary ions, as detailed in Table SM3.†

## Results and discussion

### CIMS

Of the filaments tested in the CIMS, the best results were achieved by using a Pt:Ir mesh and the  $\text{LiCO}_3$  based  $\beta$ -eucryptite synthesis method, giving a factor 500 improvement in primary ion current over the tungsten wire method. The emission current as a function of filament temperature was investigated. Emission current was monitored as a function of peak temperature for a 0.25 mm Tungsten spiral filament and a platinum iridium mesh style filament, shown in the Fig. SM2.† The emission current peaked at around  $0.002\ \mu\text{A}$  for the spiral filament at starts to level off as the filament gets past  $1350\ \text{C}$ . Above  $1350\ \text{C}$  the  $\beta$ -eucryptite in the center of the spiral starts to ablate away whilst material on the sides gets hot enough to emit, leading to a plateau in the emission current. The emission for the mesh increases exponentially to a peak of  $0.6\ \mu\text{A}$  before the heat caused uncoated sections of the filament to melt.

Fig. 3 illustrates a background mass scan then a mass scan during an 8 ppb acetone calibration, showing a strong signal at  $m/z$  65. The acetone dimer was also detected at  $m/z$  123 and its relative intensity was controlled by varying the CDC voltages (Table SM2†). Initially while the source was still emitting other alkali metals  $\text{CH}_3\text{COCH}_3\cdot\text{Na}$  and  $\text{CH}_3\text{COCH}_3\cdot\text{K}^{39/41}$  adducts were observed, as was the  $\text{H}_2\text{O}\cdot\text{Li}^+$  adduct from residual water vapour in the system. The heavier alkali metals are easier to eject from the filament but constitute a very small fraction (0.01% of the Li) of the material. They rapidly deplete during initial operation (usually within 1 hour).






Fig. 3 Background and calibration spectra for acetone permeation source.

Formic acid concentrations, at above expected ambient concentrations (up to 0.5% v/v) using the  $\text{Li}^+$  source were also obtained. This resulted in the production of  $\text{HCOOH} \cdot (\text{H}_2\text{O})_x \cdot \text{Li}^+$  clusters (Fig. SM3†), the intensity of which was highly dependent on the CDC voltage, as was the case with the acetone dimer. The production of formic acid clusters is also observed when measuring formic acid using CIMS with an iodide ionization gas.

The  $3\sigma$  limit of detections (LODs) using the improved source are 4 ppt for acetone, 0.2 ppt for formic acid, 15 ppt for nitric acid, 0.2 ppt for formic acid, and 120 ppt for ammonia as calculated from permeation tube calibrations. It should be noted that the simultaneous detection of both an acid and a base is not possible for iodide CIMS for example, highlighting the versatility of lithium attachment approach.

A serendipitous reaction of a volatile organic compound stored in a cylinder provided a further test of the  $\text{Li}^+$  CIMS development, allowing observation of a wide range of reaction products (Table 1). A 1L gas cylinder was filled with 1% v/v of acetone, 1% v/v  $\text{NO}_2$ , and 1% v/v  $\text{SF}_6$  topped up with  $\text{N}_2$  and left at room temp for approximately 24 hours before sampling, allowing time for reactions to occur within the mixture. The cylinder was then measured using  $\text{Li}^+$  CIMS. A number of additional  $m/z$  peaks were observed and the conjectured components are shown in Table 3, showing the effectiveness of this ionization method, despite some as yet unidentified peaks. Future tests can be implemented using an instrument with a higher resolving power to improve on species identification.

Emission current was monitored at constant temperature as a function of time (Fig. SM4†). The longest running filament emission current displayed an exponential decay in emission with a half-life of between 50 to 60 hours, with a filament lifetime of up to 150 hours, however there was a considerable variation in lifetime depending on the quality of manufacture. The filament lifetime in this study, despite a large variation, is a significant improvement over a stated lifetime of between 1–10 hours for Fujii *et al.*<sup>17</sup> In spite of such improvements made to the manufacturing and operation of the filaments and the high sensitivities and low limits of detection achieved, the problem of burnout (from uneven heating and subsequent melting of the wire or mesh) remained persistent, in addition

Table 1 Species present in reaction mix (excluding shoulder peaks from 6Li adducts)

$m/z$	Peak height	Assignment
6	$4.2 \times 10^5$	$\text{Li}^+$ primary (minor shoulder)
7	$1.7 \times 10^6$	$\text{Li}^+$ primary
23	$3 \times 10^3$	$\text{Na}^+$ primary
25	$1.4 \times 10^5$	$\text{H}_2\text{O} \cdot \text{Li}^+$
27	$1 \times 10^5$	Unknown
30	$1 \times 10^3$	Unknown
35	$4 \times 10^2$	$\text{N}_2 \cdot \text{Li}^+$
37	$3 \times 10^3$	$\text{NO} \cdot \text{Li}^+$
39	$3 \times 10^2$	$\text{O}_2 \cdot \text{Li}^+$
46	$1 \times 10^3$	Unknown
51	$1.6 \times 10^3$	$\text{CH}_3\text{NO} \cdot \text{Li}^+$
53	$1.5 \times 10^3$	$\text{NO}_2 \cdot \text{Li}^+$
65	$3.2 \times 10^4$	$\text{CH}_3\text{COCH}_3$ (acetone)
67	$9 \times 10^3$	$(\text{NO})_2 \cdot \text{Li}^+$
79	$6 \times 10^2$	Unknown
83	$6 \times 10^3$	$\text{CH}_3\text{CH}_3\text{NO}_2$
85	$2 \times 10^3$	Unknown
87	$2 \times 10^2$	Unknown
89	$4 \times 10^2$	Unknown
97	$3 \times 10^2$	Unknown
99	$3 \times 10^2$	$\text{N}_2\text{O}_4$
101	$3 \times 10^2$	Unknown
105	$4 \times 10^2$	$\text{CH}_3\text{CONO}_2\text{CH}_3$ (nitro acetone)
123	$3 \times 10^2$	$(\text{CH}_3\text{COCH}_3)_2$ (acetone dimer)
125	$2 \times 10^2$	Unknown
127	$1.5 \times 10^2$	Unknown
153	$1 \times 10^3$	$\text{SF}_6$

the filament coating was found to be mechanically very brittle and easily shatters resulting in a loss of  $\text{Li}^+$  emission. Due to the variation in lifetime of the filaments it is unlikely that our  $\text{Li}^+$ -CIMS would be a reliable ionization option for field and aircraft work at this time, however this achieved increase in lifetime makes the  $\text{Li}^+$  technology a viable option for applications such as laboratory studies and selected measurements where the advantages of such ionisation method could be exploited.

### Knudsen effusion mass spectrometry

In comparison to the CIMS method, the attachment probability in the KEMS is much lower and the primary ion signal is orders of magnitude greater than the secondary. This is because under UHV the molecules are travelling in straight lines and only have a single chance to collide. Furthermore the collisions also produce less stable adducts as a result of no additional gas molecules to carry away excess energy from the reaction. With optimised ion optics that allowed adducts above the limit of detection to be detected, primary emission of trace alkali metals such as rubidium, caesium and francium were also observable. The small size of the opening on the crossbeam ion source meant there was no improvement from using a mesh over a wire. Even a comparatively small amount of lithium (~2 mm diameter) bearing material on a wire, in comparison to the CIMS, was enough so saturate the entrance to the ionization area. In UHV burnout was less of a problem as Tungsten wire, with its much higher melting point (3422 °C) could be used to make the filament. The lower






Fig. 4 Comparison of EI vs.  $\text{Li}^+$  KEMS spectra for benzophenone.

ionization efficiency of  $\text{Li}^+$  compare to EI meant that the Knudsen cell in the system had to be moved much closer to the ion source (1 cm rather than 60 cm as was previously the case).

Fig. 4 shows a comparison of an electron impact and lithium ion attachment spectra of benzophenone (two benzene rings bridged by a carbonyl), a compound frequently used for calibrations of vapour pressures. The 70 eV electron impact spectra shows the molecular fragment at  $m/z$  182, and major fragments at 105 (benzene ring and CO), 77 (benzene ring) and 55. The  $\text{Li}^+$  (bottom) only produces a single molecular adduct peak at  $m/z$  189 (e.g.  $182 + 7$ ) with a small shoulder for contributions from the minor  $6\text{Li}^+$  primary, also visible are the primary emission of  $\text{Fr}^{85/87}$ .

#### Validation of single component and mixture vapor pressure measurements

Work following the Bilde *et al.*,<sup>18</sup> study culminated in the recent inter-comparison study by Krieger *et al.*<sup>19</sup> The Krieger *et al.*,<sup>19</sup> study defined the homologous series of polyethylene glycols (PEG) ( $(\text{H}-(\text{O}-\text{CH}_2-\text{CH}_2)_n-\text{OH})$  for  $n = 3$  to  $n = 8$ ) as the first series of compounds with very good agreement over a wide range of VPs as measured by a range of vapor pressure measurement techniques. PEG samples therefore provide an ideal base mark for characterising the  $\text{Li}^+$  KEMS. Reported values from the Krieger *et al.* (2017)<sup>19</sup> study show that the VP at 298 K of PEG 3 is  $6.68(+1.10/-0.95) \times 10^{-2}$  (Pa) and PEG 4 is  $1.69(+0.1/-0.10) \times 10^{-2}$  (Pa). PEG 3 and PEG 4, were both measured with the  $\text{Li}^+$  KEMS both individually and when mixed. For PEG 3 the  $\text{Li}^+$  measurement gave a vapour pressure at 298 K of  $2.49 \times 10^{-2}$  (Pa) and for PEG 4  $1.08 \times 10^{-2}$  (Pa). When PEG 3 and 4 were mixed inside of the

Knudsen cell, PEG 3 had a measured VP at 298 K of  $2.29 \times 10^{-2}$  (Pa) and for PEG 4  $1.04 \times 10^{-2}$  (Pa). This general good agreement between the reported values of the Krieger *et al.* (2017)<sup>19</sup> study and the measurements of the  $\text{Li}^+$  KEMS show the validity of using this method from this simple test. Benzophenone was used as the internal reference for these measurements.

In CI, a much more specific method than the  $\text{Li}^+$  ionization, sensitivities of measured species to the reagent ion varies considerably. For example in Bannan *et al.*,<sup>20</sup> sensitivities to different species ranged by over an order of magnitude. The sensitivity of the  $\text{Li}^+$  KEMS to different reference compounds was therefore probed by calculating vapour pressures of oxalic acid, malonic acid, *p*-ansic acid, benzophenone and *cis*-pinonic acid using each of these compounds for calibration for one another. These results are compared to previous published EI measurements with the KEMS, as detailed in Table 2. Fig. 5 shows the measured vapour pressure of malonic acid, with repeats, using different calibrations. Using an average value from all calibrations, extracted values for malonic, *P* ansic and *cis* pinonic acid from the  $\text{Li}^+$  KEMS are over reported by factors of 1.3, 1.4 and 3.3 respectively and oxalic acid is under measured by a factor of 0.75. Oxalic acid as a reference performs most poorly at recreating the measurements with the measured  $\text{Li}^+$  VP being 42% of that reported by EI, however on average with all calibrations and measurements used, the EI measurements are represented well by the  $\text{Li}^+$  attachment.

#### Secondary organic aerosol analysis

$\text{Li}^+$  KEMS was also utilised to investigate the vapour pressure of secondary organic aerosol produced from a system of atmospheric relevance, *i.e.* testing the use of the  $\text{Li}^+$  KEMS for a mixture of unknown composition. Secondary organic aerosols were generated in the Manchester Aerosol Photochemical Chamber. Briefly, the chamber consists of an 18 m<sup>3</sup> FEP, cube shaped, Teflon bag illuminated by a bank of halogen lamps and two 6 kW Xenon arc lamps simulating the solar spectrum (further details can be found in Alfara *et al.*<sup>21,22</sup>). The air charge in the bag was dried and filtered for gaseous impurities and particles, prior to humidification with high purity de-ionised water. The biogenic SOA precursor  $\alpha$ -pinene was injected into the chamber with an initial mixing ratio of 250 ppb.  $\text{NO}_x$  and  $\text{O}_3$  were added with initial mixing ratios of 98 and 40 ppb respectively. The relative humidity was 55% and

Table 2 KEMS determined solid state vapour pressures at 298 K of malonic, oxalic, *P*-ansic acid *cis* pinonic acid using EI ionization from Booth *et al.* (2009)<sup>13</sup> and  $\text{Li}^+$  using different calibration compounds

	EI measurement VP (Pa)	$\text{Li}^+$ malonic cal VP (Pa)	$\text{Li}^+$ oxalic cal VP (Pa)	$\text{Li}^+$ <i>P</i> ansic cal VP (Pa)	$\text{Li}^+$ <i>cis</i> pinonic cal VP (Pa)	$\text{Li}^+$ benzo cal VP (Pa)
Malonic acid	$5.25 \times 10^{-4}$		$1.24 \times 10^{-3}$	$8.11 \times 10^{-4}$	$2.29 \times 10^{-4}$	$3.65 \times 10^{-4}$
Oxalic acid	$1.40 \times 10^{-2}$	$1.30 \times 10^{-2}$		$1.21 \times 10^{-2}$	$5.65 \times 10^{-3}$	$1.14 \times 10^{-2}$
<i>P</i> ansic acid	$5.33 \times 10^{-4}$	$5.70 \times 10^{-4}$	$1.34 \times 10^{-3}$		$2.48 \times 10^{-4}$	
<i>cis</i> pinonic acid	$1.51 \times 10^{-4}$	$3.47 \times 10^{-4}$	$8.18 \times 10^{-4}$	$3.24 \times 10^{-4}$		






Fig. 5 Temperature dependent measurements of malonic acid (VP at 298 K is  $5.73 \times 10^{-4}$ ) (Booth *et al.* 2009)<sup>13</sup> using different calibrants with the  $\text{Li}^+$  attachment and EI KEMS.

the temperature was 25 °C. The duration was 97 minutes from lights on to lights off. The total aerosol content of the chamber was then collected on to a filter. The estimated mass of sample on the filter was approximately 1 mg. A background filter was also produced using the same procedure but excluding the reactants ( $\alpha$ -pinene,  $\text{O}_3$  and  $\text{NO}_x$ ).

In order to provide an internal calibration, the filter samples were loaded into a knudsen cell together with a sample of known vapour pressure (benzophenone). This allowed us to both calibrate in terms of what ion currents corresponded to what vapour pressures, and to normalise the total signal to account for any variations in the lithium emission current between experiment runs. Fig. 6 shows a com-




Fig. 6 EI and Li of aerosol generated in the Manchester Photochemical Aerosol Chamber.

parison of  $\text{Li}^+$  and EI of  $\alpha$ -pinene-SOA aerosol from a filter once the background has been subtracted. The  $\text{Li}^+$  spectra show a much higher contribution from high  $m/z$  signals demonstrating conclusively that the EI ionization is fragmenting almost all of the molecules in the sample. Table 3 shows the vapour pressure contribution from the different peaks in the  $\text{Li}^+$  sample as well as the total vapour pressure for comparison with EI. The vapour pressures of the individual components are within the range expected for compounds capable of partitioning between the particle and gas phase of an aerosol (roughly  $0.1\text{--}10^{-5}$  Pa). The total vapour pressure is quite high for particle phase compounds at  $1.6 \times 10^{-1}$  Pa but this can be explained by the relatively high concentrations in the experimental conditions driving more material into the particle phase than would normally be the case in the atmosphere. The total vapour pressure derived from  $\text{Li}^+$  measurements is approximately double that derived from EI ( $1.6 \times 10^{-1}$  vs.  $5.3 \times 10^{-2}$  Pa). This may arise from the necessary assumptions about the ionization cross sections of the molecules in the sample. It is assumed that the ion cross section is identical for all molecules in the sample as well as for the reference compounds. Measurement of the total ion current for the  $\text{Li}^+$  and EI signals yields an estimate of the relative sensitivity. This shows that EI is  $\sim 75$  times more sensitive than  $\text{Li}^+$ .

## Conclusions

A lithium ion attachment source for a Chemical Ionisation Mass Spectrometer (CIMS) has been constructed and demonstrated. Its potential uses for a variety of analytes including acids and bases and acetone has been shown. Improvements on this approach have shown up to a 500 fold increase in yield when a gauze is used to hold the lithium ion emitting material (mesh style filament). The source was successfully miniaturized and mounted in a Knudsen Effusion Mass Spectrometer (KEMS). Effused species adducts were detected under UHV conditions with no fragmentation. Spectral signatures were compared between EI and Lithium attachment approaches. High  $m/z$  spectral peaks were easily observed using the Lithium attachment method whereas these were absent in the case of the EI method. A direct measurement of  $\alpha$ -pinene secondary aerosol volatility has been made and compared with results from EI methods. This showed that in the KEMS the  $\text{Li}^+$  is less sensitive than EI by a factor of  $\sim 75$ .

Table 3 Vapour pressures for  $\alpha$ -pinene SOA

$m/z$	44	137	155	172	291	Total (Li)	Total (EI)
Vapour pressure (Pa)	$1.0 \times 10^{-2}$	$5.0 \times 10^{-2}$	$1.2 \times 10^{-2}$	$2.1 \times 10^{-2}$	$7.0 \times 10^{-2}$	$1.6 \times 10^{-1}$	$5.3 \times 10^{-2}$



## Conflicts of interest

There are no conflicts to declare.

## Acknowledgements

The authors wish to thank NERC for funding on grant NE/H003061/1.

## References

- 1 K. A. Pratt and K. A. Prather, Mass spectrometry of atmospheric aerosols—Recent developments and applications. Part I: Off-line mass spectrometry techniques, *Mass Spectrom. Rev.*, 2012a, **31**(1), 1–16.
- 2 K. A. Pratt and K. A. Prather, Mass spectrometry of atmospheric aerosols—Recent developments and applications. Part II: On-line mass spectrometry techniques, *Mass Spectrom. Rev.*, 2012b, **31**(1), 17–48.
- 3 M. S. Munson and F. H. Field, Chemical ionization mass spectrometry. I. General introduction, *J. Am. Chem. Soc.*, 1966, **88**(12), 2621–2630.
- 4 T. Fujii, M. Ogura and H. Jimba, Chemical ionization Mass Spectrometry with lithium ion attachment to the molecule, *Anal. Chem.*, 1989, **61**(9), 1026–1029.
- 5 T. Fujii and M. Ohta, Filament thermionic sources of Li<sup>+</sup> ions in low heating power, *J. Phys. D: Appl. Phys.*, 1995, **28**(6), 1268.
- 6 P. C. Selvin and T. Fujii, Lithium ion attachment mass spectrometry: Instrumentation and features, *Rev. Sci. Instrum.*, 2001, **72**(5), 2248–2252.
- 7 M. Nakamura, Y. Shiokawa, T. Fujii, M. Takayanagi and M. Nakata, Detection of quasimolecular ion of Cu (hfac) (tmvs) by ion attachment mass spectrometry, *J. Vac. Sci. Technol., A*, 2004, **22**(6), 2347–2350.
- 8 S. J. Edwards, C. G. Freeman, M. J. McEwan and P. F. Wilson, A selected ion flow tube investigation of the gas phase chemistry of Li<sup>+</sup>, *Int. J. Mass Spectrom.*, 2006, **255**, 164–169.
- 9 J. B. Nowak, J. A. Neuman, K. Kozai, L. G. Huey, D. J. Tanner, J. S. Holloway and F. C. Fehsenfeld, A chemical ionization mass spectrometry technique for airborne measurements of ammonia, *J. Geophys. Res.: Atmos.*, 2007, **112**(D10).
- 10 T. J. Bannan, A. Bacak, J. Muller, A. M. Booth, B. Jones, M. Le Breton, K. E. Leather, P. Xiao, D. E. Shallcross and C. J. Percival, Importance of direct anthropogenic emissions of formic acid measured by a chemical ionisation mass spectrometer (CIMS) during the Winter ClearLo Campaign in London, January 2012”, *Atmos. Environ.*, 2014, **83**, 301–310.
- 11 M. Le Breton, A. Bacak, J. B. Muller, T. J. Bannan, O. Kennedy, B. Ouyang and M. J. Daniels, The first airborne comparison of N<sub>2</sub>O<sub>5</sub> measurements over the UK using a CIMS and BBCEAS during the RONOCO campaign, *Anal. Methods*, 2014, **6**(24), 9731–9743.
- 12 M. Benyazzar and R. S. Mason, A New Technique for the Preparation of Long Life Lithium Ion Emitters and Application in a Lithium Attachment Mass Spectrometry System, 25th Int. Kinetics Symposium, Manchester, UK, 2008.
- 13 A. M. Booth, T. Markus, G. McFiggans, C. J. Percival, M. R. McGillen and D. O. Topping, Design and construction of a simple Knudsen Effusion Mass Spectrometer (KEMS) system for vapour pressure measurements of low volatility organics, *Atmos. Meas. Tech.*, 2009, **2**(2), 355–361.
- 14 A. M. Booth, T. Bannan, M. R. McGillen, M. H. Barley, D. O. Topping, G. McFiggans and C. J. Percival, The role of ortho, meta, para isomerism in measured solid state and derived sub-cooled liquid vapour pressures of substituted benzoic acids, *RSC Adv.*, 2012, **2**(10), 4430–4443.
- 15 T. J. Bannan, A. M. Booth, B. T. Jones, S. O'Meara, M. H. Barley, I. Riipinen and D. Topping, Measured Saturation Vapor Pressures of Phenolic and Nitro-aromatic Compounds, *Environ. Sci. Technol.*, 2017, **51**(7), 3922–3928.
- 16 J. T. Jayne, D. C. Leard, X. Zhang, P. Davidovits, K. A. Smith, C. E. Kolb and D. R. Worsnop, Development of an aerosol mass spectrometer for size and composition analysis of submicron particles, *Aerosol Sci. Technol.*, 2000, **33**(1–2), 49–70.
- 17 T. Fujii, P. C. Selvin, M. Sablier and K. Iwase, Lithium ion attachment mass spectrometry for on-line analysis of trace components in air: direct introduction, *Int. J. Mass Spectrom.*, 2001, **209**(1), 39–45.
- 18 M. Bilde, K. Barsanti, M. Booth, C. D. Cappa, N. M. Donahue, E. U. Emanuelsson and P. Ziemann, Saturation vapor pressures and transition enthalpies of low-volatility organic molecules of atmospheric relevance: from dicarboxylic acids to complex mixtures, *Chem. Rev.*, 2015, **115**(10), 4115–4156.
- 19 U. K. Krieger, F. Siegrist, C. Marcolli, E. U. Emanuelsson, F. M. Gøbel, M. Bilde, A. Marsh, J. P. Reid, A. J. Huisman, I. Riipinen, N. Hyttinen, N. Myllys, T. Kurtén, T. Bannan and D. Topping, A reference data set for validating vapor pressure measurement techniques: Homologous series of polyethylene glycols, *Atmos. Meas. Tech. Discuss*, 2017, DOI: 10.5194/amt-2017-224.
- 20 T. J. Bannan, A. M. Booth, A. Bacak, J. Muller, K. E. Leather, M. Le Breton and S. Visser, The first UK measurements of nitryl chloride using a chemical ionization mass spectrometer in central London in the summer of 2012, and an investigation of the role of Cl atom oxidation, *J. Geophys. Res.: Atmos.*, 2015, **120**(11), 5638–5657.
- 21 M. R. Alfarra, J. F. Hamilton, K. P. Wyche, N. Good, M. W. Ward, T. Carr and G. B. McFiggans, The effect of photochemical ageing and initial precursor concentration on the composition and hygroscopic properties of



$\beta$ -caryophyllene secondary organic aerosol, *Atmos. Chem. Phys.*, 2012, **12**(14), 6417–6436.

22 M. R. Alfarra, N. Good, K. P. Wyche, J. F. Hamilton, P. S. Monks, A. C. Lewis and G. McFiggans,

Water uptake is independent of the inferred composition of secondary aerosols derived from multiple biogenic VOCs, *Atmos. Chem. Phys.*, 2013, **13**(23), 11769.

