# **RSC Advances**



View Article Online

View Journal | View Issue

# PAPER

Check for updates

Cite this: RSC Adv., 2017, 7, 24789

Received 15th February 2017 Accepted 29th April 2017 DOI: 10.1039/c7ra01815k

rsc.li/rsc-advances

## 1. Introduction

Bacteremia is one of the main causes of high morbidity and mortality worldwide, especially for neonate and children. Bacteremia cases in United States and California occurred at an average of 1.7 million hospitalizations per year.1 Morbidity and mortality rates of bacteremia are higher in China because of relatively weak domestic basic sanitation, lack of standardized antibiotic use, and low specimen submission.2-4 Staphylococci are extremely important pathogens of bacteremia. Coagulasenegative Staphylococci (CNS) are normal inhabitants of the human skin and mucous membranes. Proper acquisition of blood specimens seriously affects the accuracy of laboratory tests. Recently, CNS was recognized as etiologic agents of a wide variety of infections and was found to be the most frequent organisms isolated from blood cultures.4,5 Staphylococcus aureus (S. aureus), a major human pathogen, and resistant strains, especially meticillin-resistant S. aureus (MRSA), play important roles in nosocomial infection. Rapid detection of S. aureus and CNS in humans is a serious challenge in clinical research.<sup>5</sup>

The diagnostic "gold standard" of infectious disease is identification of bacterial pathogens. Note that traditional methods of bacterial identification are still mainly applied by

## Fast detection of volatile organic compounds from *Staphylococcal* blood cultures by CDI-MS

Yaping Hang, <sup>(D)</sup><sup>a</sup> Konstantin Chingin,<sup>b</sup> Juchao Liang,<sup>b</sup> Xiaozhong Wang<sup>a</sup> and Longhua Hu<sup>\*a</sup>

Rapid recognition of *Staphylococcal* bacteremia in humans is a serious challenge in clinical research. Here we report on direct analysis of molecular volatile organic compounds (VOCs) of blood cultures emitted by 35 patients with a suspicion of *Staphylococcal* bacteremia and 31 patients positively diagnosed with *Staphylococcal* bacteremia were fingerprinted by atmospheric corona discharge ionization mass spectrometry (CDI-MS) for the presence of four common *Staphylococci* (*Staphylococcus aureus, S. hominis, S. haemolyticus,* and *S. epidermidis*) which statistically account for *ca.* 90% of cases of *Staphylococcal* bacteremia in humans. All the infected blood cultures revealed characteristic and clearly distinct MS patterns specific to the presence of one of the four *Staphylococci* after an incubation time of 3–16 h. Technical replicates were incubated over 2–3 days for reference diagnosis using current conventional bacteremia detection. The results of our 6 month hospital study confirm peaks at *m/z* 87 (isobutyric acid) and *m/z* 101 (isovaleric acid) signals were specifically present in *Staphylococcal* and indicate that high suitability of CDI-MS for rapid, reliable and cost-efficient diagnosis of *Staphylococcal* bacteremia in humans.

majority of the clinical microbiology laboratories. The drawbacks of the current conventional identification methods highlight several unmet needs in *Staphylococci* diagnosis, such as necessity for accuracy and rapidity. As a result of these drawbacks, the processes of proliferation, smear microscopy, isolation, biochemical identification, and treatment are prolonged. Hence, rapid and accurate diagnosis could reduce the incidence of antibiotic abuses and unnecessary treatments.

Volatile organic compounds (VOCs) can be regarded as the "chemical information molecule" of bacteria. VOCs have great potential for development and wide application of as biomarkers for bacterial diagnosis.<sup>6,7</sup> In this study, homemade and new high-sensitivity mass spectrometry (CDI-MS), followed by headspace analysis, was applied in the identification of common *Staphylococci* for early clinical diagnosis and treatment.

Other headspace-based chemical analytical methods can be used for the identification of volatile metabolites. However, these methods have drawbacks; chromatography demands higher operation and relatively complicated pre-treatment with poor practicability;<sup>8</sup> the electronic nose analysis is not highly sensitive and is not capable of analyzing chemical compositions;<sup>9</sup> the rising matrix-assisted laser desorption ionizationtime of flight-MS is inadequate for CNS identification, which is based on the difference of bacterial protein spectra, still requires bacteria detection after the isolation and culture;<sup>10</sup> and secondary electrospray ionization-MS also needs simple pretreatment with formic acid, with water used as an electrospray

<sup>&</sup>lt;sup>a</sup>Department of Clinical Laboratory, The Second Affiliated Hospital of Nanchang University, 1 Minde Road, Nanchang 330006, Jiangxi Province, P. R. China. E-mail: bingbinghang@hotmail.com; Tel: +86-791-86300410

<sup>&</sup>lt;sup>b</sup>Jiangxi Key Laboratory for Mass Spectrometry and Instrumentation, East China Institute of Technology, Nanchang 330013, P. R. China

solution.<sup>11</sup> Various methods have significant shortcomings.<sup>12</sup> Previously *Staphylococcal* VOCs observed by mass spectrometry showed in Table 1,<sup>13-15</sup> bacteremia screening by ambient CDI-MS, however, has obvious potential advantages:<sup>16-19</sup> the speed of diagnosis (<1 day) is higher than that by culture methods (total 2–3 days), greatly reducing the turnaround time for clinical testing; Another advantage of VOC analysis is its high throughput (*ca.* 10–20 s per sample) because absolutely no sample pretreatment is necessary. CDI water vapor in the air with high ionization efficiency and without the use of high-pressure gas, methanol, and other toxic reagents. Spectrum is relatively simple and easy to miniaturize. This work is an attempt to address the problems of time-consuming bacterial identification methods.

### 2. Materials and methods

#### 2.1 Sample ethical statement

All experiments were performed in compliance with the guideline "Biomedical research ethics review method involving people" (China), and approved by the medical research ethics committee at the Second Affiliated Hospital of Nanchang University. Informed consents were obtained from human participants of this study.

# 2.2 Conventional identification and preservation of pathogens

Bacteria samples were isolated from patients in the Second Affiliated Hospital of NanChang University. Samples have previously been screened for the presence of bacteria. BacT/ALERT 3D automated blood culture instrument (BD, Shanghai, China) was used to enrich the bacteria through phenotype screening of gram stain, oxidase, and  $H_2O_2$  enzymes. The identification of clinical samples was carried out with VITEK2-compact automated microbial analyzer (BioMerieux, Shanghai, China). When one of the four pathogens of interest (*S. aureus* and CNS) was detected, the remainder of the sample was stored in frozen milk at -80 °C (approximately 2 mL) for further use.

#### 2.3 Preparation of the bacterial suspension

Preserved strains or clinical isolates were recovered by culturing in blood agar plate at 35  $^{\circ}$ C overnight. The fresh single colony was inoculated in 0.45% sterile NaCl solution followed by dilution of bacterial suspension at different concentrations. Turbidity was determined with a turbidity meter (Vitek 2 DensiCHEK Plus, BioMerieux, Shanghai, China).

#### 2.4 Sample preparation

2.4.1 Simulated blood culture. Negative blood cultures for reference measurements were selected based on the lack of bacteremia symptoms. Each negative blood culture was collected without splitting (48 mL) in glass bottles (Hemoline, BioMerieux, Shanghai, China), and 25 mL of liquid was removed and equally split into five fractions. Four out of the five fractions were respectively inoculated with 1 mL of  $3 \times 10^5$  CFU per mL S. aureus, S. hominis, S. haemolyticus, and S. epidermidis suspension; the remaining fifth fraction was not inoculated. For the blank background reference, 1 mL of 0.45% sterile NaCl solution was used. All the fractions were incubated in 10 mL centrifuge tubes (Solarbio, Beijing, China). Patients with suspected bacteremia: 11 mL was removed from 48 mL of hemoline and was split into two fractions. The first 5 mL fraction was injected with 1 mL of 0.45% sterile NaCl solution, A second 6 mL fraction was incubated at the same condition for 16 h; 5 mL of each fraction of simulated and suspected blood cultures was removed for MS analysis; the remaining 1 mL was used for biochemical identification. After incubation at 35 °C at different times, the time growth curves of VOCs were observed. 5 mL of confirmed bacteremia blood culture was directly detected by MS analysis.

#### 2.5 MS analysis of VOCs emitted by blood cultures

**2.5.1 Detecting condition.** The indoor temperature was 20 °C; humidity was 30%; and nitrogen was used as the nebulizing gas at a flow rate of  $3.2 \text{ L s}^{-1}$  (0.1 MPa). Briefly, discharge needle was normally maintained at a high voltage (+4 kV in positive and -3.5 kV in negative ion detection mode) at ambient pressure. The angle between the needle and the outlet tubing was 30°. The distance from the tip of the needle to the end of the outlet tubing was 2 mm. The distance from the tip of the needle to the inlet of the LTQ capillary was 6 mm as shown in Fig. 1.

**2.5.2 Preparation of sample interface.** For CDI-MS analysis, open centrifuge tube with a microbial culture was connected to the mounted cap. Changing the sample simply required disconnecting the tube and connecting a tube with the next sample to the same cap. This operational workflow allowed the high throughput and reproducibility of analysis. The headspace

Table 1 St	taphylococcal	VOCs	previously	y observed I	oy mass s	pectrometry
------------	---------------	------	------------	--------------	-----------	-------------

Biomarker	Method	Specificity	Time (h)
Isovaleric acid	GC-MS	Staphylococci	25
1-Pyrroline-5-carboxylate	LC-ESI MS/MS	S. aureus/S. epidermidis	40-64
Enolase	LC-ESI MS/MS	S. aureus/epidermidis	40-64
Catalase	LC-ESI MS/MS	S. aureus	40-64
Aconitate hydratase	LC-ESI MS/MS	S. epidermidis	40-64
<i>m</i> / <i>z</i> 6230	MALDI-TOF MS	S. saprophyticus class A	25
<i>m</i> / <i>z</i> 3875	MALDI-TOF MS	S. aureus CC5	25
<i>m</i> / <i>z</i> 6592	MALDI-TOF MS	S. aureus CC8	25



VOCs of blood cultures were continuously transferred into the ionization region with nitrogen gas via plastic tubing. The assembly was firmly fixed in front of the instrument.

2.5.3 Detecting by CDI-MS. MS experiments were carried out on a commercial linear ion trap mass spectrometer (LTQ-XL, ThermoFischer, San Jose, CA, USA) interfaced with a homemade corona discharge-assisted extractive ionization source (SDAPCI), fully combining the advantages of various MS methods<sup>11,20,21</sup> (schematically shown in Fig. 1). LTQ-Tune automatically optimized ion lens, detection systems, and other parameters. Analyses of the blank background reference and samples were done alternately. CDI-MS analysis of each replicate was obtained for 30–60 s in the m/z range of 15–200. The entire analysis sequence was independently repeated three times. Each fingerprint is averaged from the 120 scans obtained over the initial sampling period of 10 s, when the headspace VOCs concentration is at a maximum. Chemicals identities of the sample were based on collision-induced dissociation (CID) experiments, CID analysis, up to MS3, with a collision energy (CE) of 10-30% and a mass-to-charge window width of 1.5 units,

were performed on the characteristic ions for the sake of molecular structure identification.

#### 2.6 Data processing

PCA was used to recognize difference/similarity between the microbial VOCs fingerprints based on the MS raw data by virtue of Matlab software (version 7.0, Mathworks, Inc., Natick, MA), respectively. Single, full-scan MS data was exported to Microsoft Excel as a one-to-one function. The file used the m/z values as independent variables and the relative intensities of the full-scan MS as dependent variables. Hence, matrix was assembled of a total of MS and the total number of m/z values collected in the MS. This matrix was readily loaded from the Excel file into the Matlab software for PCA analysis, respectively. The principle components for output were automatically determined by the Matlab software with the "princomp" function. The scores and loadings of the first two principle components (PCs) were 2D plotted in Matlab, and integral time of PCA was approximately several minutes.

## 3. Results and discussion

Previously, our team has analyzed the bacterial VOCs *in vitro* by ambient MS.<sup>18,19,22</sup> In the present study, we extended the scope of the investigation and focused on clinical diagnosis of *Staphylococcal* bacteremia in humans. Over a period of six months, we detected blood cultures of 35 suspects and 31 positively diagnosed patients, including 9 *S. aureus* isolates, 9 *S. hominis* isolates, 6 *S. haemolyticus* isolates, 7 *S. epidermidis* isolates, and 19 simulated blood cultures. A total of 14 cases among 35 suspects were identified as *S. aureus*, and further research was needed for the other remaining pathogens. The full agreement between the results by the ambient MS analysis and the results by the conventional biochemical analysis demonstrated the potential of ambient MS for clinical applications.

First of all, the main VOC peaks of *S. aureus* were screened by positively diagnosed specimens. The baseline level in ion



Fig. 2 Single ion chromatograms for *m*/*z* 101 signals in the CDI-screen of *S. aureus*. 8 strains was analyzed in succession. The samplings of ambient air are indicated with asterisk; bk: blank background reference; SA: *S. aureus*.

#### **RSC** Advances

chromatograms corresponds to the sampling of an empty centrifuge tube. Compared with the blank background reference, the signal intensity for m/z 101 observed was dramatically increased in the headspace of *S. aureus* blood cultures after culturing at 35 °C for 72 h. The intensity increased up to the 10<sup>6</sup> level (Fig. 2). The spike in the signal intensity at the beginning of each sampling in Fig. 2 can be attributed to the volatility of the ionized chemicals. Bacterial specific biomarker VOCs are critical for rapid pathogen identification based on  $MS^7$ , but individual VOC is limited to the identification of a particular bacteria. Joint detection of multiple metabolites must be hastened to increase the sensitivity and specificity of VOC detection methods.<sup>23</sup> Fig. 3 displays typical CDI-VOC

fingerprints in the main m/z 60–200 range of simulated blood cultures inoculated with one of the four bacterial species (*S. aureus, S. haemolyticus, S. hominis*, and *S. epidermidis*) at a concentration of 10<sup>4</sup> CFU per mL and grown aerobically at 35 °C for 16 h. Peaks observed in the background-subtracted spectra do not necessarily belong to bacterial VOCs. Therefore, each mass spectrum was shown after the subtraction of signals from blank background reference from the same individual and incubated under the same conditions, all the signal to noise ratios of bacterial VOC fingerprints we detected were much higher than 3 after he subtraction of signals from blank background reference (*e.g.*, signal to noise ratio of m/z 101 mainly were 195, 168, 206, 225, 251, 201), it indicated that the



Fig. 3 Positive (top) and negative (bottom) ion mode VOC fingerprints of simulated blood cultures inoculated with one of the four *Staphylococcal* species.



Fig. 4 CDI-MS spectrum of volatiles emitted by S. aureus culture after 16 h incubation, and MS/MS experiments of m/z 101.

#### Paper

instrument was stable and sensitive to our requirements. A number of blood metabolic-related signals, *e.g.*, m/z 59, m/z 90, and m/z 92, are shared between different bacteria. Thus, the fingerprints can be readily distinguished by visual inspection. Separation of the bacterial metabolism associated with signals from the metabolism of blood was necessary to obtain the specific bacterial VOC fingerprints.<sup>24</sup> The signals of *S. aureus* were relatively higher intensity. The signal intensities for m/z 47, m/z 89, m/z 95, and m/z 108 were dramatically increased, with the intensity reaching up to the 10<sup>5</sup> level. Generally low signal peaks of *S. haemolyticus* and *S. hominis* were observed. By contrast, the signal peaks of *S. epidermidis* were higher than the other two CNS besides the common signal. Intensities of m/z 73, m/z 106, and m/z 108 were even higher than *S. aureus*, whereas the remaining signal peaks were low.

Based on the earlier published reports, the relative molecular mass of compounds corresponding to m/z,<sup>25</sup> and the high specific signals were confirmed by MS/MS experiments (Fig. 4).



**Fig. 5** PCA score plot of VOCs fingerprints from four simulated *Staphylococci* blood cultures in CDI-mode. Strains marked in different shapes or colors and the distance between each graph indicate the respective degree of difference; degree of dispersion between the same kinds of graphs indicates reproducibility.

We temporarily assigned some interesting signals in bacterial fingerprints to ethanol (m/z 47,  $CDI^+$ ), nitric acid (m/z 64,  $CDI^+$ ), phenol (m/z 95,  $CDI^+$ ), benzaldehyde (m/z 108,  $CDI^+$ ), acetic acid (m/z 59,  $CDI^-$ ), ammonium propionate (m/z 90,  $CDI^-$ ), and aniline (m/z 92,  $CDI^-$ ). The biomarker signals were isovalerate (m/z 101,  $CDI^-$ ) and isobutyric acid (m/z 89,  $CDI^+$ ; m/z 87,  $CDI^-$ ). Notably, m/z 101 signal specifically existed in all four *Staphylococci* at different intensities, thus helping to differentiate *Staphylococci* from the other bacterial species. Moreover, the presence of m/z 132 and m/z 134 suggested that isovalerates are secondary metabolites of these two substances. All the features of *Staphylococcal* spectrum introduced above can be used for the identification of pathogens.

Handling multi-variable MS data by PCA has significant advantages.<sup>26,27</sup> The replicates of *S. aureus* are clustered at the largest distance from the rest of the bacterial strains; *S. haemolyticus* are distant from *S. aureus* with highest distinction; *S. hominis* and *S. epidermidis* are close to each other but the boundaries are clear (detailed in Fig. 5). Overall, the discrimination of VOC fingerprints from the four *Staphylococci* is satisfactory using PCA, especially for *S. aureus* and *S. haemolyticus*, which are indistinguishable by traditional methods because of the presence of hemolysis. Early exclusion of contamination with CNS will greatly reduce the abuse of antibiotics and reduce health care costs for patients.

Moreover, the incubation time of 16 h used in the above simulated experiment is substantially shorter than in traditional clinical routines (48–72 h), thereby allowing space for further improvement. Therefore, we also investigated the release time of VOCs. Results revealed that the pathogen fingerprint identification saved time for at least 48 h. The relative signal intensities of the different bacterial VOCs grew at different rates, whereas the intensities of some signals remained nearly constant. The peaks at m/z 101 and m/z 87 were paid considerable attention because of their possibility as biomarker of *Staphylococci*; steady increase in the signal intensity in the beginning of incubation (<12 h) was followed by a rapid increase (Fig. 6a). However, no significant relationship with the initial concentration of bacteria was



Fig. 6 The intensity of MS signals at m/z 101, m/z 87 (a) and m/z 64 (b) in the VOC mass spectrum of simulated S. aureus blood cultures as a function of culture time or initial S. aureus concentration.

established. The characteristic releases of isobutyric acid (m/z 87 in negative ion mode) and isovaleric acid  $(m/z \ 101$  in negative ion mode) by SA cultures are respectively linked to the metabolism of valine and leucine precursor.13,14 Therefore, the signal response of particular biomarkers can be selectively enhanced for targeted analysis by increasing the nutrient substance in culturing medium. Furthermore, signal intensity of nitrate  $(m/z \ 64$  in positive ion mode) was proportional to the initial concentration of bacteria, whereas there was nothing to do with the culture time and remained constant. This appearance may be related to the ability of bacterial metabolism. There also remains space for further analytical improvement (e.g., adjustment of the medium, change of culture environment or optimization of mass spectrometry conditions). We can optimize the conditions (including m/z range, peak intensity and peak area%) for each m/z peak of all the VOCs in the standard strains blood culture, then establish the fingerprint database. We need to do internal quality control with the standard strains conventionally, the establishing of cutoff value for qualitative experiment and preparation of calibrators for quantitative testing are the goal of our next phase of experimentation. The integration of this approach in clinical practice will be further promoted by the growing availability of atmospheric corona discharge ionization mass spectrometers in bioanalytical laboratories and core facilities.

## 4. Conclusion

VOCs are good biomarkers of pathogens, *Staphylococci* were readily distinguished based on characteristic fingerprints. The obvious advantages of CDI-MS are particularly suitable for the detection of VOCs. However, the potential of using ambient ionization approaches in real clinical practice remains largely unexplored. The results of our 6 month hospital study strongly confirmed the potential of CDI-MS for clinical applications. Therefore, *Staphylococcal* blood culture processed by the rapid and noninvasive CDI-MS analysis can be further interrogated by more expensive and time-consuming biochemical assays for complementary analysis.

## Acknowledgements

This work is jointly supported by National Natural Science Foundation of China (NNSFC) (no. 81460327, 21105010) and major Project Science and Technology Agency of Jiangxi Province (no. 20152ACG70014).

## References

- 1 S. K. Gohil, C. Cao, M. Phelan, T. Tjoa, C. Rhee, R. Platt and S. S. Huang, C. Centers for Disease and P. Prevention Epicenters, *Clin. Infect. Dis.*, 2016, **62**, 695–703.
- 2 Q. Liu, W. Li, X. Du, W. Li, T. Zhong, Y. Tang, Y. Feng, C. Tao and Y. Xie, *PLoS One*, 2015, **10**, e0130701.
- 3 J. N. Wu, T. E. Gan, Y. X. Zhu, J. M. Cao, C. H. Ji, Y. H. Wu and B. Lv, *J. Zhejiang Univ.*, *Sci.*, *B*, 2015, **16**, 70–77.

- 4 Y. Jiang, L. Kuang, H. Wang, L. Li, W. Zhou and M. Li, *Intern. Med.*, 2016, **55**, 597–603.
- 5 D. J. Anderson, R. W. Moehring, R. Sloane, K. E. Schmader, D. J. Weber, V. G. Fowler Jr, E. Smathers and D. J. Sexton, *PLoS One*, 2014, **9**, e91713.
- 6 B. Audrain, M. A. Farag, C. M. Ryu and J. M. Ghigo, *FEMS Microbiol. Rev.*, 2015, **39**, 222–233.
- 7 V. Shestivska, K. Dryahina, J. Nunvar, K. Sovova, D. Elhottova, A. Nemec, D. Smith and P. Spanel, *J. Breath Res.*, 2015, **9**, 027104.
- 8 Y. Wang, M. Wang, J. Li, S. Yao, J. Xue, W. Zou and C. Hu, *J. Antibiot.*, 2014, **67**, 511–518.
- 9 R. M. Schnabel, M. L. Boumans, A. Smolinska,
  E. E. Stobberingh, R. Kaufmann, P. M. Roekaerts and
  D. C. Bergmans, *Respir. Med.*, 2015, **109**, 1454–1459.
- 10 C. D. Doern and S. M. Butler-Wu, J. Mol. Diagn., 2016, 18, 789–802.
- 11 S. P. Pasilis and G. J. Van Berkel, *Encyclopedia of Spectroscopy and Spectrometry*, 2017, 3rd edn, pp. 819–829.
- 12 S. Sethi, R. Nanda and T. Chakraborty, *Clin. Microbiol. Rev.*, 2013, **26**, 462–475.
- 13 A. G. James, D. Cox and K. Worrall, *Flavour Fragrance J.*, 2013, 28, 231–237.
- 14 A. G. James, C. J. Austin, D. S. Cox, D. Taylor and R. Calvert, *FEMS Microbiol. Ecol.*, 2013, **83**, 527–540.
- 15 K. Cheng, H. Chui, L. Domish, D. Hernandez and G. Wang, Proteomics: Clin. Appl., 2016, **10**, 346–357.
- K. Sekimoto, M. Sakakura, T. Kawamukai, H. Hike, T. Shiota,
  F. Usui, Y. Bando and M. Takayama, *Analyst*, 2016, 141, 4879–4892.
- 17 Z. Aladaghlo, A. R. Fakhari and K. S. Hasheminasab, Microchem. J., 2016, **129**, 41-48.
- 18 J. Liang, Y. Hang, K. Chingin, L. Hu and H. Chen, *RSC Adv.*, 2014, 4, 25326.
- 19 K. Chingin, J. Liang, Y. Hang, L. Hu and H. Chen, *RSC Adv.*, 2015, 5, 13952–13957.
- 20 X.-X. Li, L.-F. Chen, Y.-Z. Ouyang, F. Feng and H.-W. Chen, *Chin. J. Anal. Chem.*, 2016, 44, 25–31.
- 21 H.-W. Chen, J.-H. Lai, Y.-F. Zhou, Y.-F. Huan, J.-Q. Li, Z. Xie, Z.-C. Wang and M.-B. Luo, *Chin. J. Anal. Chem.*, 2007, 35, 1233–1240.
- 22 L. Hu, J. Liang, K. Chingin, Y. Hang, X. Wu and H. Chen, *RSC Adv.*, 2016, **6**, 8449–8455.
- 23 S. Sethi, R. Nanda and T. Chakraborty, *Clin. Microbiol. Rev.*, 2013, **26**, 462–475.
- 24 K. Kim, C. Mall, S. L. Taylor, S. Hitchcock, C. Zhang, H. I. Wettersten, A. D. Jones, A. Chapman and R. H. Weiss, *PLoS One*, 2014, **9**, e86223.
- 25 L. D. Bos, P. J. Sterk and M. J. Schultz, *PLoS Pathog.*, 2013, 9, e1003311.
- 26 C. Sales, M. I. Cervera, R. Gil, T. Portolés, E. Pitarch and J. Beltran, *Food Chem.*, 2017, **216**, 365–373.
- 27 J. Ding, S. Yang, D. Liang, H. Chen, Z. Wu, L. Zhang and Y. Ren, *Analyst*, 2009, **134**, 2040–2050.