



Cite this: *RSC Adv.*, 2017, 7, 25665

Fluorescent detection of point mutation *via* ligase reaction assisted by quantum dots and magnetic nanoparticle-based probes

S. Heidari Sharafdarkolaei,^a M. Motovali-Bashi ^a and P. Gill ^{*bc}

A nanodiagnostic genotyping method was presented for point mutation detection directly in human genomic DNA based on ligase reaction coupled with quantum dots and magnetic nanoparticle-based probes. For this purpose, allele-specific probes, including a biotin-labeled common probe and two biotin-labeled allele-specific probes were designed for mutant and wild alleles of human beta globin gene (IVS-II-I G → A point mutation). When genomic DNA carried the mutation site, the common probe and allele-specific probe were ligated to form exponential amplified biotin-labeled fluorescence ligation products. These ligated products were captured by streptavidin-coated magnetic nanoparticles at one end and then attached to a QD 605-streptavidin conjugate at the other end to be detected fluorescently. Thereafter, the genotypes were identified conveniently according to the fluorescence color of quantum dots using a rotor-gene 6000Q real-time rotary analyzer. The results demonstrated the sensitivity and specificity percentages of this nanomolecular mutation detection method were 85.45% and 95.77% respectively. In addition, this method could be a high throughput and high sensitivity detection system that represents suitable non-PCR based nanodiagnostics for detection of other point mutations.

Received 1st April 2017
 Accepted 3rd May 2017

DOI: 10.1039/c7ra03767h
rsc.li/rsc-advances


Introduction

Gene point mutations present important biomarkers for genetic diseases. However, existing point mutation detection methods suffer from low sensitivity, specificity and tedious assay processes.¹ A major goal of mutation diagnostic methods is the ability to detect low abundant mutations with enough sensitivity to detect the mutant DNA without generating false positives or false negatives. Nucleic acid amplification methods fall into three categories: target, probe and signal amplification systems.² The majority of diagnostic methods rely on PCR amplification of target sequences.³ The main disadvantages of PCR based methods although highly sensitive are amplification errors due to mis-priming, limited accuracy of discriminating single nucleotide variations, and limited multiplexing capability.^{1,4-6} Although a number of alternative PCR-free methods have been introduced, ligation-based techniques are the most widely used for diagnostic methods due to their exceptional specificity on base discrimination using ligase enzyme and robust multiplexing capabilities.^{1,7}

The ligase detection reaction (LDR) is a probe amplification method that uses two adjacent probes and the thermostable ligase.⁸ LDR employs two-allele specific probes known as discriminating probes, which differ in the base at their 3'-end and one common probe. Only if the discriminating primer fully complement to the target sequence, the ligase joins this primer to common one to form ligation products. Common or discriminating primer or both of them can be labelled, which is used for detection step.⁹⁻¹⁴ LDR has certain advantages over PCR based methods. LDR amplifies the probe molecule instead of producing fragments of DNA through polymerization of nucleotides so; few steps are involved in the amplification cycle, thereby reducing the time needed to complete the entire process. In addition, PCR requires high-quality DNA because it involves the amplification of large chains of nucleotides, whereas LDR does not, because the target sequence is usually small (36 to 60 nucleotides). Ligase reactions have exceptional specificity on base discrimination because of high fidelity thermostable ligase enzyme.^{1,7} The use of a thermostable ligase minimizes target independent ligation since the ligation reaction can be performed at or near the melting temperature (T_m) of the oligonucleotide probes.

Colloidal semiconductor nanocrystals, which are also called quantum dots (QDs) as a popular kind of fluorescent material that have several unique photophysical properties. Quantum dots have been found to be superior to traditional organic dyes as fluorescent markers with numerous advantageous features,

^aDepartment of Genetics, Faculty of Sciences, University of Isfahan, Isfahan, Iran
^bNanomedicine Group, Immunogenetics Research Center, Mazandaran University of Medical Sciences, Sari, Iran
^cThalassemia Research Center, Mazandaran University of Medical Sciences, Sari, Iran.
 E-mail: pooriagill@yahoo.com; p.gill@mazums.ac.ir



involving people" (I.R. Iran), and approved by the ethical committee of Mazandaran University of Medical Science. Informed consents were obtained from human participants of this study. Fifty persons who their IVSII-1 (G → A) beta globin gene mutation investigated using ARMS-PCR method established in Mazandaran Thalassemia Research Center. Genomic DNA was isolated from leucocytes in peripheral blood of normal and patients subjects using phenol-chloroform extraction protocol. In addition, the genomic DNA concentration was assessed at 260 nm wavelength using UV-vis spectrophotometer.

LDR amplification

LDR was performed using a Bio-Rad MJ mini thermal cycler. Extracted genomic DNA was used as LDR template. A set of 3 probes, including one common probe and two variant-specific probes were designed and ordered. All probes were labelled with biotin and attach to streptavidin coated magnetic nanoparticles and streptavidin coated quantum dots. Each LDR reaction contained 10 μ L reaction volumes: 1 μ L of $10\times$ Taq ligase buffer; 4 units TaqDNA ligase; 20 μ M each probes; template (20 ng genomic DNA). The conditions of LDR were the following: initial denaturation at 94 $^{\circ}$ C for 4 min, followed by 30 cycles of denaturation at 94 $^{\circ}$ C for 30 s, then 65 $^{\circ}$ C for 5 min.

Coupling LDR ligated products

The proposed method employed MNPs to capture LDR ligated products through biotin-streptavidin interaction. The capturing process involved two major steps: (1) capturing ligated products using streptavidin-coated MNPs; (2) removal of non-specific absorption components on the surface of MNPs using washing buffer. Hence, 10 μ L of LDR products with 2 μ L of MNPs (1 mg ml⁻¹) were incubated for 30 min at 37 $^{\circ}$ C. Then MNPs were absorbed using a magnetic plate and supernatant was discharged. Next, the MNPs were re-dispersed in incubation buffer and this washing procedure was repeated three times. Subsequently, 5 μ L QD (0.2 μ M) was added to the MNP-LDR mixture for 30 min at 37 $^{\circ}$ C for shaping MNP-LDR-QD nanocomplexes. Then, the nanocomplexes transferred to a magnetic plate, supernatant was discharged, and washing procedure was repeated three times again. Finally, deposited nanocomplexes were resuspended in 10 μ L incubation buffer for fluorescent measurement.

Fluorescent assay

Fluorescence ligation products on the surface of MNPs were detected by the rotor-gene 6000 real-time rotary analyzer and the corresponding signal were obtained which refer to QD 605 ITK Streptavidin in orange area.

TEM characterization

A few drops of each QDs, MNPs and ligation products sample were dispersed onto a 3 mm copper grid covered with a continuous carbon film and dried at room temperature. The TEM image of each sample was obtained on a CM30

transmission electron microscope (Philips EO, Netherlands) operating at 150 kV.

Clinical performance of LDR-QD system

Fifty DNA samples were analyzed by LDR-QD system. True positive (TP), true negative (TN), false positive (FP) and false negative (FN) were determined by comparing the results of routine test (ARMS-PCR) and LDR-QD method. In this study, TP means routine detection test reported samples with IVS II-I (G-A) mutation and LDR-QD test confirmed it; TN means routine detection test reported samples without IVS II-I (G-A) mutation and LDR-QD test confirmed it; FP means detection test didn't report samples with IVS II-I (G-A) mutation but LDR-QD test reported samples with IVS II-I (G-A) mutation and FN means routine detection test reported samples with IVS II-I (G-A) mutation but LDR-QD test reported samples without IVS II-I (G-A) mutation. The sensitivity and specificity percentages were determined as follows: % sensitivity = TP/TP + FN, and % specificity = TN/TN + FP. The sensitivity of LDR-QD test were referred to the ability of the test for identifying those samples with IVS II-I (G-A) mutation, correctly. The specificity of LDR-QD test was referred to the ability of the test for identifying those samples without IVS II-I (G-A) mutation, correctly.^{35,36}

Results

TEM

To investigate the formation of the MNP-QD conjugate, TEM images were obtained from MNPs and QDs before and after ligation detection reaction (LDR). TEM micrographs showed that MNPs surrounded by QDs, which indicated the formation of the MNP-QD conjugates (Fig. 2).

QD-LDR signals

In the presence of a mismatch between two adjacent probes and template, ligation was not occurred. During LDR process, 3' and 5'-biotin-LPA labelled ligation product was amplified for mutated template and 3' and 5'-LPG-labeled ligation product was amplified for wild type template. These ligation products were attached to the QD streptavidin conjugate through biotin-streptavidin interaction and captured by streptavidin coated MNPs. The signal of QD was determined by detection of the fluorescence signal using rotor-gene 6000 real-time rotary analyzer in wavelength of 605 nm (orange area) (Fig. 3).

QD-LDR genotyping

Genotypes were determined by detecting the fluorescence signals of QDs attached to ligation products (Fig. 3). Two LDR reactions were performed in separated tubes which was used wild allele specific probe (LPG) at one tube and mutant allele specific (LPA) at the other one. Both tubes exhibited fluorescence signal in heterozygote mutant sample, the tube that included of LPG exhibited signal in normal sample and the tube that consist of LPA had signal in homozygote mutant sample (Fig. 4).






Fig. 2 TEM schematics (up) and micrographs (down) of streptavidin coated MNPs and QD605 ITK streptavidin conjugate. Left, streptavidin coated MNPs; middle, QD605 ITK streptavidin conjugates; right, MNPs–QD conjugate formations in the presence of ligase enzyme and connection of two probes together.




Fig. 3 Standard fluorescence signals of wild alleles, IVSII-1 (G–A) homozygote alleles, and heterozygote alleles gained by rotor-gene real-time rotary analyzer. Two LDR reactions were performed in separated tubes which is used wild allele specific probe (LPG) in one tube and mutant allele specific (LPA) in another one. Genotypes were determined by detecting the fluorescence signal of QDs attached to ligation products.

QD–LDR clinical performance

The clinical performance of the test was determined by comparing the results of LDR–QD test and ARMS. Table 2 summarized reliability values of LDR–QD test for diagnosis of β -thalassemia. In this study, among 50 persons with and without IVSII-1 (G–A) mutation, the false negative and positive results were 2 and 8 alleles respectively; however, the true positive and negative results were 47 and 43 alleles respectively. Moreover, the comparison between the results of the routine tests with the LDR–QD results indicated the sensitivity and specificity of this assay were 85.45% and 95.77% respectively (Table 2). Our results suggested that LDR–QD system could be suitable for detection of IVSII-1 (G–A) mutation.

High sensitivity and specificity of this assay for point mutation detection was attributed to the high-efficiency ligase based LDR and the signal amplification ability of QD-labelled probe.




Fig. 4 QD–LDR genotyping assay wild, homozygote, and heterozygote DNAs. The QD signal of 6 samples were analyzed with optical detector. According to fluorescence intensity of LPG (wild specific probe) and LPA (mutant specific probe), samples 1 and 5 were homozygotes, samples 2 and 6 were normals, samples 3 and 4 were heterozygotes.

Discussion

Beta thalassemia is considered the most common autosomal single gene disorder worldwide that can be found in more than 60 countries with a carrier population of up to 150–200 million. So far, close to 200 different mutations in the β -globin gene have been reported that associated with the onset of the disease^{37,38} the frequency and geographic distribution of beta thalassemia mutations is useful to establish a program for carrier screening, genetic counselling, prenatal diagnosis and for physicians to

Table 2 Clinical performances of LDR–QD test in comparison to ARMS method

LDR–QD test	ARMS test		Clinical performances	
	Mutant allele	Wilde Allele	% sensitivity	% specificity
Mutant allele	47	2	85.45	95.77
Wilde allele	8	43		





Table 3 Comparisons of conventional assay with the present research for point mutation detection

Detection method	Mutation(s)	Mechanism	Number of probes	Fluorescent molecule(s)	Time (hours)	Disadvantages	Ref. no.
Allele-specific real-time PCR	Codon 13 (G → A) of KRAS	Microfluidic cartridge containing preloaded primers and probes which enables the parallel detection of point mutation by allele-specific real-time PCR	4	FAM	4	Fabrication of the microfluidic system is expensive, a T-T, G-T, C-T, or C-A 3'-terminal mismatch error prone of allele specific PCR amplification	42
High resolution melt analysis	Fr 8/9 (+G), IVS1-1 (G > A), IVS1-5 (G > C), IVS1-110 (G > A), and CD44 (-C)	Is the quantitative analysis of the melt curve of a DNA fragment following amplification by PCR	2	EvaGreen	4	Depends strongly on good PCR, instruments and dyes. Heterozygote detection does not depend on the type of single base substitution or the variant position within the PCR product. Small insertions and deletions may be somewhat more difficult to detect than substitutions	43
Ligase chain reaction	CD17 (A → T)	Two pairs of probe hybridization, ligation and signal detection	6	FAM and ROX	3	Prone to false-positive amplification	44
Gapped-ligase chain reaction	Codon12 and 13 (G → A)	PCR products use as template and DNA polymerase fills in a gap between annealed probes which are subsequently joined by DNA ligase	4	FAM, ROX and JOE	More than 4	Need of polymerase, nucleotides and extension steps	45
Ligase detection reaction	IVS1-1 (G → A), IVS1-6 (T → C), IVS1-110 (G → A), codon 39 (C → T), IVS1-1 (G → A)	PCR amplification, probe hybridization and ligation to PCR products, signal detection	5	Cy5	More than 4	Combined PCR-LDR process suffers from the complications of PCR	46
Nano based ligase detection reaction	IVSII-1 (G → A)	Probe hybridization and ligation, separation by MNPs and quantum dot signal detection	3	QD	3	Providing non-PCR-based fluorescent analyzer	This study

establish specific therapeutic approaches for patients with beta thalassaemia.

In this report, we described a nanodiagnostic method for beta-thalassaemia point mutation using quantum dot-LDR system that could detect directly point mutations in human genomic DNA. Combining ligase detection reaction (LDR) and magnetic nanoparticle technique used for increased sensitivity, and employing fluorescence ligation products allowed detection analysis. The method has been demonstrated as convenient point mutation detection in human genome with the genotyping of the IVSII-1 (G → A) point mutation of β -globin gene associated with thalassaemia disease. Although LDR was highly specific in base recognition, but it had very limited sensitivity. Consequently, LDR was usually combined with PCR that exponentially amplified the ligation products to a detectable level. The combined PCR-LDR process significantly improves the assay sensitivity but suffers from the complications of PCR.^{1,39} To our knowledge, this is the first report to demonstrate a practical biological application of LDR-quantum dot PCR-free system.

Streptavidin-biotin interaction was specifically chosen for this assay. This binding interaction was quick, reliable and strong.^{17,40} The success of the system depend on three components: (1) accuracy of the allele discrimination reaction by thermal DNA ligase, which has been well-established; (2) specificity biotin-streptavidin interaction between magnetic nanoparticles and quantum dots with ligation products; (3) the sensitive detection resulting from LDR product coupled with quantum dot and enrichment of streptavidin-coated magnetic nanoparticle.

The assay was able to differentiate between the homogeneous mutant, the heterozygous mutant and the wild type by analyzing the sample using QD-labelled mutant and wild type-specific probes. The method is simple, cost-effective, and paves a new way for point mutation detection as well as molecular diagnosis.⁴¹ While there have been studies such as those described by Meng and Battistella, which have reported the use of ligase detection system for point mutation detection, this study is unique in its integration of quantum dot and magnetic nanoparticles with LDR assay (Table 3).

While there have been studies which have reported the use of ligation reaction for point mutation detection, this study is unique in its integration of both quantum dots and magnetic nanoparticles in LDR-PCR free system.⁴⁷⁻⁴⁹

Meng and Battistella described a convenient genotyping method capable of detecting point mutations directly in human genomic DNA based on the combination of ligase chain reaction (LCR) and microbead-enrichment technique. LCR probes, including a biotin-labelled common probe and two fluorescence-labelled allele-specific probes, were designed for two alleles of a mutated site. When genomic DNA carried the mutated site, the common probe and allele-specific probe were ligated to form exponential amplified biotin-labelled fluorescence ligation products. Streptavidin-coated microbeads enriched these ligated products, and genotypes were identified conveniently according to the fluorescence colour of microbeads using fluorescent microscopy. The method had been demonstrated as convenient point mutation detection in

human genome with the genotyping of the CD17 (A → T) point mutation of β -globin gene associated with thalassaemia disease.⁴⁴

In the LCR, thermostable ligase lacking blunt-end ligation activity should be used to prevent the generation of background signals produced by the blunt-end ligation of the complementary probe pair duplexes.⁵⁰ In a manner analogous to PCR, the ligated products could be exponentially amplified because the ligated probes serve as a template for the ligation reaction of the complementary probe pair in the following cycle. However, despite the improved efficiency, the LCR was prone to false-positive amplification. Because non-specifically ligated probes could also be used as a template so, the LCR was not as popular.⁵¹

Battistella *et al.* developed a genotyping assay based on LDR system that followed by a universal PCR (U-PCR) of genomic DNA-templated LDR product. The assay was designed to detect 7 prevalent mutations in the beta globin gene.⁴⁶ Although this strategy had sufficiently high sensitivity and specificity to differentiate multiple targets but the combined PCR-LDR process suffer from the complications of PCR.

Several point mutation detection methods have been developed such as restriction fragment length polymorphisms (RFLP),⁵² temporal temperature gradient gel electrophoresis (TGGE),⁵³ denaturing gradient gel electrophoresis (DGGE),^{54,55} reverse dot blot hybridization, amplification refractory mutation system (ARMS),⁵⁶ high resolution melting (HRM).³⁵ However, these technologies have some shortcomings. The main disadvantages of PCR based methods although highly sensitive are amplification errors due to mispriming, limited accuracy of discriminating single nucleotide variations, post-PCR processing steps such as gel electrophoresis and limited multiplexing capability.

To overcome these problems we presented a novel SNP assay that for the first time exploited MNPs and QDs coupled with LDR technique, which is removed amplification and post-PCR processing steps. Unique optical properties of QDs increase multiplexing capability of the assay. High throughput magnetic separation and high precision QD signal detection improve the sensitivity and selectivity of this assay.

Conclusion

This nanodiagnostic ligation assay was developed for accurately SNP/point mutation detection directly in human genomic DNA. This method did not require amplification step and use very fast and highly specific biotin-streptavidin interaction. Separating and labelling steps do in one tube using this quick and easily biotin-streptavidin interaction. Quantum dots have a broader excitation spectra and a narrow more sharply defined emission peak, which was not presented in organic fluorophores. Due to these properties, a single light source can be used to excite multicolour quantum dots simultaneously without signal overlap and allows the use of many probes to track several targets simultaneously, too. The large Stokes shift (difference between peak absorption and peak emission wavelengths) reduces autofluorescence, which increases sensitivity. In future



using two different types of quantum dots signal can be detected in one tube, which reduces sample volume, processing times. Finally, this method matches the golden rules for the development of diagnostic applications. These include (1) low costs, (2) multiplexing capabilities, (3) short processing times, (4) low sample consumption, (5) easy adaptability to new formats and targets, (6) easily standardized and programmed.

Acknowledgements

This study was supported partially by University of Isfahan and Vice Chancellor of Research and Technology, Mazandaran University of Medical Sciences. The authors thank also Central Research Laboratory, Mazandaran university of Medical Sciences. Also, the authors thank from Iran Nanotechnology Initiative (INI).

Notes and references

- 1 Y. Song, Y. Zhang and T. H. Wang, *Small*, 2013, **9**, 1096–1105.
- 2 P. Gill, M. Forouzandeh, N. Eshraghi, M. Ghalami, M. Safa and M.-R. Noori-Dalooi, *Mol. Cell. Probes*, 2008, **22**, 103–109.
- 3 H. Sun, Y. Yang, L. Yang, B. Su, G. Jiang, K. Fei and D. Lu, *BioMed Res. Int.*, 2014, **2014**, 407537.
- 4 J. Ragoussis, *Annu. Rev. Genomics Hum. Genet.*, 2009, **10**, 117–133.
- 5 P. Y. Kwok and X. Chen, *Curr. Issues Mol. Biol.*, 2003, **5**, 43–60.
- 6 Z. Tsuchihashi and N. C. Dracopoli, *Pharmacogenomics J.*, 2002, **2**, 103–110.
- 7 W. Cao, *Trends Biotechnol.*, 2004, **22**, 38–44.
- 8 F. Barany, *Proc. Natl. Acad. Sci. U. S. A.*, 1991, **88**, 189–193.
- 9 S. Watanabe, K. Hagihara, K. Tsukagoshi and M. Hashimoto, *Anal. Chem.*, 2014, **86**, 900–906.
- 10 X. L. Wang, S. G. Xie, L. Zhang, W. X. Yang, X. Wang and H. Z. Jin, *World J. Gastroenterol.*, 2008, **14**, 120–124.
- 11 R. Sinville, J. Coyne, R. J. Meagher, Y. W. Cheng, F. Barany, A. Barron and S. A. Soper, *Electrophoresis*, 2008, **29**, 4751–4760.
- 12 A. E. Dent, C. T. Yohn, P. A. Zimmerman, J. Vulule, J. W. Kazura and A. M. Moormann, *Am. J. Trop. Med. Hyg.*, 2007, **77**, 250–255.
- 13 P. Yi, W. Lu, J. Guo, Q. Liu, Z. Chen, J. Han and L. Li, *Cell Biochem. Biophys.*, 2011, **61**, 629–636.
- 14 X. T. Gregg, R. Kralovics and J. T. Prchal, *Am. J. Hematol.*, 2000, **63**, 184–191.
- 15 H. Zhou, J. Liu, J. J. Xu and H. Y. Chen, *Anal. Chem.*, 2011, **83**, 8320–8328.
- 16 H. Zhang, X. Fu, L. Liu, Z. Zhu and K. Yang, *Anal. Biochem.*, 2012, **426**, 30–39.
- 17 H. C. Yeh, Y. P. Ho, M. Shih Ie and T. H. Wang, *Nucleic Acids Res.*, 2006, **34**, e35.
- 18 T. Jamieson, R. Bakhshi, D. Petrova, R. Pocock, M. Imani and A. M. Seifalian, *Biomaterials*, 2007, **28**, 4717–4732.
- 19 W. R. Algar, A. J. Tavares and U. J. Krull, *Anal. Chim. Acta*, 2010, **673**, 1–25.
- 20 E. Abbasi, T. Kafshdooz, M. Bakhtiary, N. Nikzamir, M. Nikzamir, M. Mohammadian and A. Akbarzadeh, *Artif. Cells, Nanomed., Biotechnol.*, 2016, **44**, 885–891.
- 21 E. Petryayeva, W. R. Algar and I. L. Medintz, *Appl. Spectrosc.*, 2013, **67**, 215–252.
- 22 H. Xu, M. Y. Sha, E. Y. Wong, J. Uphoff, Y. Xu, J. A. Treadway, A. Truong, E. O'Brien, S. Asquith, M. Stubbins, N. K. Spurr, E. H. Lai and W. Mahoney, *Nucleic Acids Res.*, 2003, **31**, e43.
- 23 X. He and N. Ma, *Anal. Chem.*, 2014, **86**, 3676–3681.
- 24 S. Su, J. Fan, B. Xue, L. Yuwen, X. Liu, D. Pan, C. Fan and L. Wang, *ACS Appl. Mater. Interfaces*, 2014, **6**, 1152–1157.
- 25 C. Y. Zhang, H. C. Yeh, M. T. Kuroki and T. H. Wang, *Nat. Mater.*, 2005, **4**, 826–831.
- 26 D. Diaz-Diestra, B. Thapa, J. Beltran-Huarac, B. R. Weiner and G. Morell, *Biosens. Bioelectron.*, 2017, **87**, 693–700.
- 27 Y. Tan, Y. Wang, M. Li, X. Ye, T. Wu and C. Li, *Biosens. Bioelectron.*, 2017, **91**, 741–746.
- 28 X. Gao, L. Yang, J. A. Petros, F. F. Marshall, J. W. Simons and S. Nie, *Curr. Opin. Biotechnol.*, 2005, **16**, 63–72.
- 29 A. P. Alivisatos, W. Gu and C. Larabell, *Annu. Rev. Biomed. Eng.*, 2005, **7**, 55–76.
- 30 W. C. Chan, D. J. Maxwell, X. Gao, R. E. Bailey, M. Han and S. Nie, *Curr. Opin. Biotechnol.*, 2002, **13**, 40–46.
- 31 A. Fu, W. Gu, C. Larabell and A. P. Alivisatos, *Curr. Opin. Neurobiol.*, 2005, **15**, 568–575.
- 32 E. Palecek and M. Fojta, *Talanta*, 2007, **74**, 276–290.
- 33 X. Zhang, D. B. Reeves, I. M. Perreard, W. C. Kett, K. E. Griswold, B. Gimi and J. B. Weaver, *Biosens. Bioelectron.*, 2013, **50**, 441–446.
- 34 M. K. Pal, M. Rashid and M. Bisht, *Biosens. Bioelectron.*, 2015, **73**, 146–152.
- 35 M. Zafari, P. Gill, M. Kowsaryan, A. Alipour and A. Banihashemi, *J. Matern.-Fetal Neonat. Med.*, 2015, 1–6.
- 36 A. Rafati and P. Gill, *Microchim. Acta*, 2015, **182**, 523–530.
- 37 M. R. Mahdavi, H. Karami, M. T. Akbari, H. Jalali and P. Roshan, *Case Reports in Hematology*, 2013, **2013**, 906292.
- 38 S. Hajihoseini, M. Motovali-Bashi, M. A. Honardoost and N. Alerasool, *Iran. J. Public Health*, 2015, **44**, 380–387.
- 39 M. Hashimoto, F. Barany, F. Xu and S. A. Soper, *Analyst*, 2007, **132**, 913–921.
- 40 K. Sato, K. Hosokawa and M. Maeda, *J. Am. Chem. Soc.*, 2003, **125**, 8102–8103.
- 41 Z. Yuan, Y. Zhou, S. Gao, Y. Cheng and Z. Li, *ACS Appl. Mater. Interfaces*, 2014, **6**, 6181–6185.
- 42 O. Strohmeier, S. Laßmann, B. Riedel, D. Mark, G. Roth, M. Werner, R. Zengerle and F. von Stetten, *Microchim. Acta*, 2014, **181**, 1681–1688.
- 43 M. Ramezanzadeh, M. Salehi and R. Salehi, *Adv. Biomed. Res.*, 2016, **5**, 71.
- 44 X. Meng, X. Yang, K. Wang, Q. Guo, Y. Tan, Q. Mo and X. Xu, *Talanta*, 2010, **80**, 1725–1729.
- 45 S. Jenner and D. Techel, *Tumour Biol.*, 2015, **36**, 6361–6368.
- 46 S. Battistella, F. Damin, M. Chiari, K. Delgrosso, S. Surrey, P. Fortina, M. Ferrari and L. Cremonesi, *Clin. Chem.*, 2008, **54**, 1657–1663.
- 47 Y. Sun, X. Lu, F. Su, L. Wang, C. Liu, X. Duan and Z. Li, *Biosens. Bioelectron.*, 2015, **74**, 705–710.



- 48 M. B. Wabuye, H. Farquar, W. Stryjewski, R. P. Hammer, S. A. Soper, Y. W. Cheng and F. Barany, *J. Am. Chem. Soc.*, 2003, **125**, 6937–6945.
- 49 Y. Cheng, Q. Du, L. Wang, H. Jia and Z. Li, *Anal. Chem.*, 2012, **84**, 3739–3744.
- 50 G. W. Shin, B. Chung and G. Y. Jung, *Electrophoresis*, 2014, **35**, 1004–1016.
- 51 J. P. Schouten, C. J. McElgunn, R. Waaijer, D. Zwijnenburg, F. Diepvens and G. Pals, *Nucleic Acids Res.*, 2002, **30**, e57.
- 52 J. Li, *Zhongguo Yixue Kexueyuan Xuebao*, 1992, **14**, 42–45.
- 53 R. V. Shaji, E. S. Edison, B. Poonkuzhali, A. Srivastava and M. Chandy, *Clin. Chem.*, 2003, **49**, 777–781.
- 54 M. Losekoot, R. Fodde, C. L. Hartevelde, H. van Heeren, P. C. Giordano and L. F. Bernini, *Br. J. Haematol.*, 1990, **76**, 269–274.
- 55 A. M. Dozy and Y. W. Kan, *Clin. Genet.*, 1994, **45**, 221–227.
- 56 J. A. Tan, J. S. Tay, L. I. Lin, S. K. Kham, J. N. Chia, T. M. Chin, N. B. Aziz and H. B. Wong, *Prenatal Diagn.*, 1994, **14**, 1077–1082.

