


 Cite this: *RSC Adv.*, 2017, 7, 28419

# Multimode microsensors based on Ag–TiO<sub>2</sub>–graphene materials used for the molecular recognition of carcinoembryonic antigen in whole blood samples

 Livia Alexandra Gugoasa,<sup>a</sup> Ahmed Jassim Muklive AlOgaidi,<sup>b</sup> Raluca-Ioana Stefan-van Staden,<sup>id</sup>\*<sup>ab</sup> Ahed El-Khatib,<sup>c</sup> Marcela-Corina Rosu<sup>d</sup> and Stela Pruneanu<sup>d</sup>

Ag–TiO<sub>2</sub>–graphene pastes modified with inulin, and L-alanine *tert*-butyl ester nitrate (L-Ala-C<sub>4</sub>-L-Lac), respectively, were proposed for the molecular recognition of carcinoembryonic antigen (CEA) from whole blood samples. High levels of CEA are associated with different types of cancer, one of them being colorectal cancer. Both sensors gave high reliability for the measurements, and also high sensitivity and selectivity were recorded. The modes used were stochastic mode for qualitative and quantitative analysis of CEA, and differential pulse voltammetry for the quantitative measurement of CEA. The lowest limit of determination (20.5 fg mL<sup>−1</sup>) was recorded when the multimode sensor based on Ag–TiO<sub>2</sub>–graphene paste was modified with the inulin IQ.

 Received 4th April 2017  
Accepted 22nd May 2017

DOI: 10.1039/c7ra03842a

rsc.li/rsc-advances

## Introduction

Carcinoembryonic antigen (CEA) is a glycoprotein encoded by a gene from the superfamily of immunoglobulins<sup>1</sup> and it can be found in the chromosome 19q.<sup>2</sup> CEA is the most common tumor marker used in the diagnosis of several types of cancer. After its discovery in 1964 (ref. 3 and 4) over expression of CEA levels were mainly used in gastrointestinal cancer with focus on colorectal tumors.

Determination of this biomarker in bodily fluids plays a crucial role in biomedical research. To date, CEA was determined using various analytical methods (Table 1) from serum samples. A fast screening method is needed for both diagnosis of patients as well as checking the efficiency of cancer treatment.

The group of Stefan-van Staden proposed multimode sensors as new reliable tools for the analysis of biomarkers from biological samples.<sup>16,17</sup> This new type of sensors can perform in more than one mode, *e.g.*, differential pulse voltammetry (DPV), stochastic mode, potentiometric mode, measurements of one or more analytes from the same sample. Multimode sensors

offer a qualitative and a quantitative analysis by their response obtained in stochastic mode and a second quantitative analysis by their response in either amperometric or potentiometric modes. Because of this, stochastic mode is used first for the qualitative analysis, and the first quantification of the analyte(s). For this paper the second mode was amperometric mode with DPV measurements.

In this paper, we proposed two multimode microsensors based on Ag–TiO<sub>2</sub>–graphene (Ag–TiO<sub>2</sub>/rGO) materials modified with inulin IQ and L-alanine *tert*-butyl ester L-lactate (L-Ala-C<sub>4</sub>-Lac) for the screening tests of whole blood samples for CEA. The selection of the modifiers was done based on their properties: possibility to form channels for stochastic sensing, and also good electrocatalytic activity needed on DPV mode. The modification of graphene material with Ag–TiO<sub>2</sub> nanoparticles facilitated the increasing the surface of the sensors, as well as the producing of channels, and increasing of the conductivity of graphene. The modes used were stochastic and differential pulse voltammetry (DPV).

## Experimental

### Materials and reagents

Carcinoembryonic antigen (CEA), inulin, L-alanine *tert*-butyl ester L-lactate (L-Ala-C<sub>4</sub>-Lac), sodium azide, monosodium and disodium phosphate were purchased from Sigma Aldrich (Milwaukee, USA) and paraffin oil (d<sub>4</sub>,<sup>20</sup> 0.86 g cm<sup>−1</sup>) from Fluka (Buchs, Switzerland). TiO<sub>2</sub>-P25 powder was purchased from Degussa (Germany). Triton X-100 and acetylacetone, used as organic additives, were bought from Fluka (Switzerland) and

<sup>a</sup>Laboratory of Electrochemistry and PATLAB, National Institute of Research for Electrochemistry and Condensed Matter, 202 Splaiul Independentei Str., Bucharest-6, 060021, Romania. E-mail: ralucaivanstaden@gmail.com; Fax: +40 213163113; Tel: +40 751507779

<sup>b</sup>Faculty of Applied Chemistry and Material Science, Politehnica University of Bucharest, Bucharest, Romania

<sup>c</sup>University of Medicine and Pharmacy "Carol Davila", Bucharest, Romania

<sup>d</sup>National Institute for Research and Development of Isotopic and Molecular Technologies, Donat Street, No. 67-103, Cluj-Napoca, RO-400293, Romania



**Table 1** Analytical methods for the determination of carcinoembryonic antigen

Method	Biological samples	Detection limit	Reference
Electrochemiluminescence (ECL)	Serum	1.52 fg mL <sup>-1</sup>	5
Electrochemistry-immunosensor		8 pg mL <sup>-1</sup>	6
		90 pg mL <sup>-1</sup>	7
Electrochemistry-biosensor		0.2 fg mL <sup>-1</sup>	8
		8 pg mL <sup>-1</sup>	9
		30 pg mL <sup>-1</sup>	10
Surface Plasmon Resonance (SPR)		1 ng mL <sup>-1</sup>	11
Fluorescence		3 pg mL <sup>-1</sup>	12
		0.3 ng mL <sup>-1</sup>	13
Lateral-flow immunoassay (LFIA)		0.25 ng mL <sup>-1</sup>	14
Fluorescence-spectroscopy		49 pg mL <sup>-1</sup>	15

Merck (Germany), respectively. Silver nitrate (AgNO<sub>3</sub>) was obtained from Alfa-Aesar (Germany). Ascorbic acid (C<sub>6</sub>H<sub>8</sub>O<sub>6</sub>) and ethanol were acquired from Fluka (Germany). Graphene oxide (GO) was prepared from natural graphite (Sigma Aldrich, Germany) using a modified Hummer's method.<sup>18</sup>

Monosodium and disodium phosphates were used for the preparation of 0.1 mol L<sup>-1</sup> phosphate buffer solution (PBS) of pH = 7.4. Deionized water obtained from a Millipore Direct-Q 3 System (Molsheim, France) was used for the preparation of all solutions. All standard solutions were prepared in buffer solution pH = 7.4, with NaN<sub>3</sub> 0.1% in a ratio water : buffer solution of 1 : 1 (v/v). Serial dilution technique was used for the preparation of solutions of different concentrations (2.05 × 10<sup>-14</sup> g mL<sup>-1</sup> to 5.00 × 10<sup>-6</sup> g mL<sup>-1</sup>). CEA solutions were stored in the fridge at 2–8 °C. All chemicals were of analytical grade.

### Preparation of Ag–TiO<sub>2</sub> graphene (Ag–TiO<sub>2</sub>/rGO) nanocomposite


In the first step, Ag–TiO<sub>2</sub> nanoparticles were obtained by a combined chemical/thermal approach. A suspension of TiO<sub>2</sub> nanoparticles in double-distilled water (0.15 g mL<sup>-1</sup>) was added to a mixture of Triton X-100 and acetylacetone (3 : 2 v/v) and magnetically stirred at room temperature. Then, silver nitrate (AgNO<sub>3</sub>) and ascorbic acid (C<sub>6</sub>H<sub>8</sub>O<sub>6</sub>) (1 : 2 mass ratio) were added into the TiO<sub>2</sub> homogenous suspension. AgNO<sub>3</sub> acted as Ag nanoparticles precursor and C<sub>6</sub>H<sub>8</sub>O<sub>6</sub> as reducing agent for

Ag<sup>+</sup> ions. The final Ti : Ag molar ratio was 50 : 1. The mixture was magnetically stirred at 80 °C until the color changed from white to dark gray, indicating the chemical reduction of Ag<sup>+</sup> to Ag<sup>0</sup>. Finally, the dried powder was thermally treated at 450 °C for 2 h, in the air. The resulted gray material, denoted as Ag–TiO<sub>2</sub>, was naturally cooled to room temperature.

In the second step, Ag–TiO<sub>2</sub> nanoparticles were mixed with graphene oxide (GO) and then thermally treated, in order to obtain Ag–TiO<sub>2</sub>/reduced graphene oxide (Ag–TiO<sub>2</sub>/rGO). Briefly, GO and Ag–TiO<sub>2</sub> nanoparticles (20 : 1 mass ratio) were dispersed in 20% ethanol (Fluka, Germany) solution and the suspension was ultrasonicated for 30 minutes. Then, the suspension was magnetically stirred at 50 °C, until complete drying. Following, the material was subjected to 1 hour thermal treatment at 550 °C, in argon/hydrogen controlled atmosphere. The thermal treatment is necessary for the reduction of GO which is an insulating material. The reduced graphene oxide (rGO) obtained after the thermal treatment has a low content of oxygen-containing groups and high conductivity, being beneficial for charge transfer.<sup>19</sup>

### Design of multimode microsensors based on carbon matrices

Paraffin oil was added to the Ag–TiO<sub>2</sub>/rGO powder until a homogenous paste was formed. To each 50 mg paste were added 50 µL of modifier (10<sup>-3</sup> mol L<sup>-1</sup> solution of either inulin, or ionic liquid: L-alanine *tert*-butyl ester L-lactate (L-Ala-C<sub>4</sub>-Lac)) to give the

**Fig. 1** Construction diagram of the multimode sensor.




Fig. 2 Examples of stochastic diagrams for the assay of CEA in whole blood samples using (a) inulin IQ/Ag-TiO<sub>2</sub>/rGO, (b) L-AlaC<sub>4</sub>-Lac/Ag-TiO<sub>2</sub>/rGO.

modified paste. Each modified paste was inserted in a plastic tip with a diameter of the active site of 300  $\mu\text{m}$  (Fig. 1). The electric contact was done using an Ag wire. When not in use, the multi-mode sensors were stored at room temperature, in a dry place.

### Apparatus and methods

All measurements were performed with AUTOLAB/PGSTAT 12 (Methrom) connected to a computer. A three electrode system electrochemical cell was employed (Fig. 1). Ag/AgCl (0.1 mol L<sup>-1</sup> KCl) electrode serves as a reference electrode in the cell and a Pt wire as auxiliary electrode in the cell, respectively.

A stochastic mode was used for the measurements of the qualitative parameter  $t_{\text{off}}$  and the quantitative parameter  $t_{\text{on}}$ , at a constant potential (125 mV vs. Ag/AgCl) and was used for both qualitative and quantitative analysis of CEA (Fig. 2). The known value of  $t_{\text{off}}$  was followed in each stochastic diagram to identify CEA and further for its quantification using the recorded  $t_{\text{on}}$  value ( $1/t_{\text{on}} = f(\text{conc.})$ ).

Differential pulse voltammetry (DPV) was used for the quantitative analysis of CEA in whole blood samples. All parameters used for DPV scans are summarized in Table 2. The unknown concentrations of CEA from whole blood samples were determined from the calibration graphs  $H = f(\text{conc.})$ , where  $H$  is the value of peak height (Fig. 3).

Table 2 Working parameters used in DPV mode for the assay of CEA using multimode microsensors

Parameter	Microsensors based on	
	Inulin-IQ/ Ag-TiO <sub>2</sub> /rGO	L-Ala-C <sub>4</sub> -L-Lac/ Ag-TiO <sub>2</sub> /rGO
Potential range (mV)	400 to >1500	1000 to >2200
Scan rate (mV s <sup>-1</sup> )	50	50
Modulation amplitude (mV)	10	10
Step potential (mV)	10	10
Peak position (mV)	1650	1570




Fig. 3 Examples of voltammograms for the assay of CEA in whole blood samples using (a) inulin IQ/Ag-TiO<sub>2</sub>/rGO, (b) L-AlaC<sub>4</sub>-Lac/Ag-TiO<sub>2</sub>/rGO.

TEM images were obtained with an H-7650 120 kV Automatic Microscope-Hitachi while STEM-EDX images were obtained with an SU-8230 STEM system, Hitachi (Japan). Before investigations, the samples were diluted in ethanol then dropped on



copper grid (200 mesh) and dried at room temperature for several minutes.

Phase formation and crystalline properties of the composite sample were revealed by X-ray powder diffraction, using a Bruker X-ray diffractometer (Bruker, Germany) with  $\text{CuK}_\alpha$  radiation ( $\lambda = 1.54056 \text{ \AA}$ ).

### Samples

Eleven blood samples were collected from patients confirmed with colon cancer from the University Hospital in Bucharest (Ethics committee approval nr 75/2015). Informed consent was obtained from all patients. The samples were screened using the multimode sensors, without any treatment sampling procedure.

## Results and discussions

### TEM and XRD characterization of Ag-TiO<sub>2</sub>/rGO composite

Fig. 4a and b show representative TEM images of TiO<sub>2</sub> and Ag-TiO<sub>2</sub> nanoparticles, respectively. One can see that there are no

major morphological differences between TiO<sub>2</sub> and Ag-TiO<sub>2</sub> nanoparticles. The majority of nanoparticles have round shape, with the size below 50 nm. Due to the low concentration of AgNO<sub>3</sub> used in the synthesis of the composite material, only a small number of silver nanoparticles were identified by EDX mapping (Fig. 4c). The elemental maps of Ti, O and Ag in the Ag-TiO<sub>2</sub> sample can be seen in Fig. 4d–f, respectively.

As known, the Ag<sup>+</sup> has a larger ionic radius (1.29 Å) than that of Ti<sup>4+</sup> (0.745 Å)<sup>20</sup> therefore silver ions are not readily incorporated at Ti sites in the TiO<sub>2</sub> lattice, but mostly deposited on TiO<sub>2</sub> surface. In addition, the substitution of Ti<sup>4+</sup> by Ag<sup>+</sup> would be accompanied by an important lattice expansion, due to the large difference between them. This was not observed in our XRD measurement.

The Ag-TiO<sub>2</sub> nanoparticles were mixed with graphene oxide then the material was thermally reduced at 550 °C, resulting the final Ag-TiO<sub>2</sub>/rGO composite (Fig. 5a and b). The interaction between Ag-TiO<sub>2</sub> nanoparticles and the sp<sup>2</sup> carbon network of rGO is mainly physical. As a result of the oxygen-containing




Fig. 4 TEM image of: TiO<sub>2</sub> nanoparticles (a) and Ag-TiO<sub>2</sub> nanoparticles (b); elemental maps of Ti, O and Ag in Ag-TiO<sub>2</sub> sample (c–f); scale bar 100 nm (a–f).








Fig. 5 TEM images of Ag-TiO<sub>2</sub>/rGO composite; scale bar 100 nm (a); 600 nm (b).

group removal from the GO structure, the graphene sheets are closely associated and stacked *via* van der Waals' interactions. The two representative TEM images of composite indicate that the Ag-TiO<sub>2</sub> nanoparticles are uniformly dispersed on the graphene surface. The high transparency of the synthesized material suggests that graphene sheets are very thin being composed of 2–3 layers. Also, the wrinkles and folding of the layers, specific to graphene, can be clearly seen. TiO<sub>2</sub> and Ag attached to graphene have beneficial effects, by increasing the surface area of the composite material (Ag-TiO<sub>2</sub>/rGO).


Phase formation and crystalline properties of the composite sample were next studied by X-ray powder diffraction. The X-ray pattern of Ag-TiO<sub>2</sub> powder illustrates the presence of both anatase (PDF card no. 21-1272) and rutile (PDF card no. 21-1276) crystalline phases of TiO<sub>2</sub>, denoted as TiO<sub>2</sub>-A and TiO<sub>2</sub>-R, respectively (Fig. 6). By comparison, Ag-TiO<sub>2</sub>/rGO composite exhibits only the diffraction peaks of TiO<sub>2</sub> anatase crystalline phase. No peaks related to the presence of silver nanoparticles were identified, due to the low concentration of Ag in the composites (Ag-TiO<sub>2</sub> and Ag-TiO<sub>2</sub>/rGO).

The broad peak at  $2\theta = 21.25^\circ$  can be attributed to reduced graphene oxide (rGO). Its position correlates with the distance

between the graphene layers (*d*-spacing) which was found to be 0.43 nm (slightly higher than the distance between graphene layers in graphite – 0.335 nm). The graphene crystallite size was estimated from the FWHM of the rGO diffraction peak, using Scherrer equation<sup>21</sup> and was determined to be 0.8 nm. By taking into account the *d*-spacing value (0.43 nm) we determined that the graphene crystallites are composed of two layers. Such finding was in excellent agreement with TEM results which clearly revealed the transparent morphology of graphene sheets (few-layer graphene).

#### Response characteristics of multimode microsensors used in stochastic mode

Stochastic response relies on the conductivity of channels when the analyte is going through it: under an applied potential, the molecule of the analyte is extracted from the solution to the membrane–solution interface. Molecular recognition of the analyte takes place in 2 stages. In the first stage the molecule of the analyte blocks the entrance in the pore (the current is dropping to zero) for a certain period of time which is called  $t_{\text{off}}$  and represents the qualitative parameter of the analyte. In the second stage the molecule is binding to the pore wall and the following equation of equilibrium takes place:



where Ch is channel, and *i* is the interface, followed by redox processes. The binding time is called  $t_{\text{on}}$  and represents the quantitative parameter.

In Table 3 are presented the response characteristics of the multimode microsensors in stochastic mode. The best multimode microsensor for stochastic mode was the one based on inulin IQ/Ag-TiO<sub>2</sub>/rGO, because it had the highest sensitivity and the lowest determination limit (20.5 fg mL<sup>-1</sup>), making it a good choice for early detection of CEA.

#### Response characteristics of multimode microsensors used in DPV mode

Response characteristics of the multimode microsensors in DPV mode are given in Table 4. For the assay of CEA (Table 4), the microsensor based on inulin IQ/Ag-TiO<sub>2</sub>/rGO showed the




Fig. 6 X-ray diffraction patterns of Ag-TiO<sub>2</sub> (magenta) and Ag-TiO<sub>2</sub>/rGO (blue) composites, showing the characteristic peaks of TiO<sub>2</sub> (anatase and rutile) and of reduced graphene oxide (rGO).



Table 3 Response characteristics of multimode microsensors in stochastic mode<sup>a</sup>

Microsensors based on	Calibration equation and correlation coefficient ( <i>r</i> )	Linear concentration range ( $\mu\text{g mL}^{-1}$ )	$t_{\text{off}}$	Sensitivity ( $\text{s}^{-1}/\text{mol L}^{-1}$ )	Limit of determination ( $\mu\text{g mL}^{-1}$ )
L-Ala-C <sub>4</sub> -L-Lac/Ag-TiO <sub>2</sub> /rGO	$1/t_{\text{on}} = 0.01 + 2.91 \times 10^{-1} \times C, r = 0.9991$	$1.60 \times 10^{-3}$ to $2.00 \times 10^{-1}$	2.5	$2.91 \times 10^{-1}$	$1.6 \times 10^{-3}$
Inulin-IQ/Ag-TiO <sub>2</sub> /rGO	$1/t_{\text{on}} = 0.018 + 9.40 \times 10^4 \times C, r = 0.9999$	$2.05 \times 10^{-8}$ to $5.12 \times 10^{-7}$	2	$9.40 \times 10^4$	$2.05 \times 10^{-8}$

<sup>a</sup>  $\langle 1/t_{\text{on}} \rangle = \text{s}^{-1}$ ;  $\langle C \rangle = \mu\text{g mL}^{-1}$ .

Table 4 Response characteristics of multimode microsensors for DPV mode

Sensor based on	Equation of calibration <sup>a</sup> and correlation coefficient ( <i>r</i> )	<i>E</i> (mV)	Limit of detection ( $\mu\text{g mL}^{-1}$ )	Limit of determination ( $\mu\text{g mL}^{-1}$ )	Linear conc. range ( $\mu\text{g mL}^{-1}$ )
L-Ala-C <sub>4</sub> -L-Lac/Ag-TiO <sub>2</sub> /rGO	$H = 6.1 \times 10^{-9} + 4.0 \times 10^{-9} \times C, r = 0.9996$	1570	$4.77 \times 10^{-5}$	$1.60 \times 10^{-3}$	$1.60 \times 10^{-3}$ to $2.00 \times 10^{-1}$
Inulin-IQ/Ag-TiO <sub>2</sub> /rGO	$H = 1.2 \times 10^{-8} + 6.5 \times 10^{-6} \times C, r = 0.9981$	1650	$4.78 \times 10^{-7}$	$1.02 \times 10^{-7}$	$1.02 \times 10^{-7}$ to $1.60 \times 10^{-3}$

<sup>a</sup>  $\langle H \rangle = \text{A}$ ;  $\langle C \rangle = \text{mg mL}^{-1}$ .

lowest determination limit ( $1.02 \text{ fg mL}^{-1}$ ). The same microsensor had covered the widest linear concentration range ( $1.02 \text{ fg mL}^{-1}$  to  $0.20 \text{ } \mu\text{g mL}^{-1}$ ) for the assay of CEA.

Both types of modifiers – ionic liquids and inulins are nanostructured materials, which can give the necessary pores needed for stochastic sensing, but the inulins had also a ring in their structures which can function as channel, and therefore they are better materials for the design of stochastic sensors; this property was also reflected on the response characteristics of the proposed sensors.

Based on the experimental data, the multimode microsensor of choice, which can be used reliable for both modes is the one based on Ag-TiO<sub>2</sub>/rGO modified with inulin IQ.

The reliability of the design was checked by construction and evaluation of 5 different sensors of each category. For each category, the sensitivity was evaluated initially, and every day, when the % RSD values recorded were lower than 1.0% for 3 months.

Interferences were checked *versus* CA19-9, p53, and HER-1. For stochastic mode, different  $t_{\text{off}}$  values were recorded with the proposed sensors for the compounds proposed as interferences and CEA, and accordingly the proposed sensors and selective in the stochastic mode. Also, measurements performed in mixed solution methods for the DPV mode shown that the amperometric selectivity coefficients recorded were lower than  $10^{-4}$ , proving that the sensors are selective also in this mode, and accurate measurements can be performed.

Although the limits of determination are a bit higher than those reported to date in the literature,<sup>5,8</sup> the advantages of the proposed sensors and method are: simplicity, the biological fluid can be used as taken from the patients, there is a first step of qualitative analysis based on identification of the signature of the analyte which indicates reliable the presence of CEA, also the quantitative analysis used for the assay of concentration is performed using two techniques: stochastic mode and DPV mode – and this is increasing the precision and accuracy of the determination.

## Interferences

The multimode sensors' selectivity was checked using both modes: stochastic and DPV, *versus* other cancer biomarkers: p53, KRAS, HER-1, NSE, and CYFRA-21. For the stochastic mode, the signatures ( $t_{\text{off}}$  values) obtained for the specified possible interferences were different from the one obtained for CEA using the proposed sensors; accordingly, the sensors are selective in the stochastic mode.

Mixed solution method was selected for the determination of the selectivity amperometric coefficients of the proposed sensors in the DPV mode. The results presented in Table 5 for the amperometric selectivity coefficients shown that the other cancer biomarkers did not interfere in the assay of CEA using the DPV mode.

Table 5 Amperometric selectivity coefficients obtained in the DPV mode for the proposed multimode sensors

Sensor based on	$K_{\text{sel}}^{\text{amp}}$				
	P53	KRAS	HER-1	NSE	CYFRA-21
L-Ala-C <sub>4</sub> -L-Lac/Ag-TiO <sub>2</sub> /rGO	$5.16 \times 10^{-3}$	$6.70 \times 10^{-3}$	$6.77 \times 10^{-3}$	$6.96 \times 10^{-3}$	$6.52 \times 10^{-3}$
Inulin-IQ/Ag-TiO <sub>2</sub> /rGO	$5.22 \times 10^{-3}$	$6.68 \times 10^{-3}$	$9.52 \times 10^{-5}$	$5.87 \times 10^{-3}$	$5.90 \times 10^{-3}$



Table 6 The assay of CEA in whole blood samples using multimode microsensors in stochastic and DPV modes<sup>a</sup>

Nr	Inulin IQ/Ag-TiO <sub>2</sub> /rGO, ng mL <sup>-1</sup>		L-AlaC <sub>4</sub> -Lac/Ag-TiO <sub>2</sub> /rGO, ng mL <sup>-1</sup>		ELISA, ng mL <sup>-1</sup>	t-Test
	Stochastic mode	DPV mode	Stochastic mode	DPV mode		
1	96.40 ± 0.21	98.40 ± 0.32	96.50 ± 0.22	99.00 ± 0.32	95.95 ± 0.54	1.49
2	116.00 ± 0.17	114.96 ± 0.35	113.90 ± 0.20	111.16 ± 0.30	110.02 ± 0.55	1.30
3	69.80 ± 0.18	67.55 ± 0.27	68.49 ± 0.13	68.09 ± 0.23	67.00 ± 0.52	1.78
4	29.06 ± 0.20	28.90 ± 0.20	27.10 ± 0.16	27.90 ± 0.20	27.22 ± 0.60	1.27
5	87.50 ± 0.23	91.80 ± 0.43	93.30 ± 0.20	89.20 ± 0.29	87.50 ± 0.46	1.92
6	63.10 ± 0.18	66.00 ± 0.37	63.73 ± 0.22	66.56 ± 0.42	62.59 ± 0.51	1.68
7	17.50 ± 0.09	15.23 ± 0.27	15.03 ± 0.12	17.18 ± 0.11	15.29 ± 0.50	1.03
8	68.96 ± 0.21	68.30 ± 0.22	68.25 ± 0.21	66.67 ± 0.25	66.50 ± 0.47	1.60
9	13.40 ± 0.08	13.36 ± 0.19	13.56 ± 0.14	13.80 ± 0.27	13.23 ± 0.40	0.98
10	96.40 ± 0.21	101.92 ± 0.32	100.00 ± 0.22	102.96 ± 0.30	99.87 ± 0.42	1.27
11	0.30 ± 0.05	0.31 ± 0.08	0.32 ± 0.07	0.32 ± 0.09	0.27 ± 0.12	1.02

<sup>a</sup> All values are average of three determinations.

### Analytical applications

Normal levels of CEA in a healthy patient are less than 2.5 ng mL<sup>-1</sup>. High levels of CEA in blood predict colon cancer tumors, therefore a fast, simple and cheap method to analyse the concentration of CEA in hospitalized patients can help physicians to ask the patients to do further investigations for a fast diagnostic. Results obtained from the screening of whole blood samples, using multimode sensors, in stochastic and DPV mode are shown in Table 6. Statistical analysis based on pair *t*-test was performed for a confidence level of 99.00%. The results have shown that there is no significant difference between the values obtained using the multimode microsensors used in either stochastic or DPV modes, and ELISA (standard method) at 99.00% confidence level ( $t_{\text{theoretical}} = 4.032$ ). Accordingly, the proposed sensors are a good alternative for ELISA, for screening tests and quantitative assay of CEA in whole blood samples.

### Conclusions

In this paper we proposed two multimode microsensors for the molecular recognition of carcinoembryonic antigen (CEA). For the design of the proposed multimode sensors an Ag-TiO<sub>2</sub>/rGO paste was modified with inulin IQ or two types of ionic liquids. The best multimode sensor which reached very low limits of determination in both stochastic and DPV modes (LOQs in the magnitude orders of fg mL<sup>-1</sup>) was the one based on Ag-TiO<sub>2</sub>/rGO modified with inulin IQ. The proposed method is more simple, fast, and accurate compared with those proposed previously and can be used for the screening of the whole blood samples, as taken from the patients.

### Acknowledgements

This work was supported by UEFISCDI, Partnership 22/2014 and by a grant of CNCS/CCCDI – UEFISCDI, project number PN-III-P2-2.1-PED-2016-0392, within PNCDI III. The authors want to thank Alin Sebastian Porav for performing the TEM investigation.

### References

- 1 D. M. P. Thomson, J. Krupey, S. O. Freedman and P. Gold, The radioimmunoassay of circulating carcinoembryonic antigen of the human digestive system, *Proc. Natl. Acad. Sci. U. S. A.*, 1969, **64**, 161–167.
- 2 J. A. Thompson, F. Grunert and W. Zimmermann, Carcinoembryonic antigen gene family: molecular biology and clinical perspective, *J. Clin. Lab. Anal.*, 1991, **5**, 344–366.
- 3 P. Gold and S. O. Freedman, Demonstration of tumor-specific antigens in human colonic carcinomata by immunological tolerance and absorption techniques, *J. Exp. Med.*, 1965, **121**, 439–462.
- 4 P. Gold and S. O. Freedman, Specific carcinoembryonic antigens of the human digestive system, *J. Exp. Med.*, 1965, **122**, 467–481.
- 5 D. Wang, Y. Li, Z. Lin, B. Qiu and L. Guo, Surface-Enhanced Electrochemiluminescence of Ru@SiO<sub>2</sub> for Ultrasensitive Detection of Carcinoembryonic Antigen, *Anal. Chem.*, 2015, **87**, 5966–5972.
- 6 D. Feng, L. Li, X. Fang, X. Han and Y. Zhang, Dual signal amplification of horseradish peroxidase functionalized nanocomposite as trace label for the electrochemical detection of carcinoembryonic antigen, *Electrochim. Acta*, 2014, **127**, 334–341.
- 7 J. Liu, J. Wang, T. Wang, D. Li, F. Xi, J. Wang and E. Wang, Three-dimensional electrochemical immunosensor for sensitive detection of carcinoembryonic antigen based on monolithic and macroporous graphene foam, *Biosens. Bioelectron.*, 2015, **65**, 281–286.
- 8 S. Samanman, A. Numnuam, W. Limbut, P. Kanatharana and P. Thavarungkul, Highly-sensitive label-free electrochemical carcinoembryonic antigen immunosensor based on a novel Au nanoparticles-graphene-chitosan nanocomposite cryogel electrode, *Anal. Chim. Acta*, 2015, **853**, 521–532.
- 9 J. Huang, J. Tian, Y. Zhao and S. Zhao, Ag/Au nanoparticles coated graphene electrochemical sensor for ultrasensitive analysis of carcinoembryonic antigen in clinical immunoassay, *Sens. Actuators, B*, 2015, **206**, 570–576.



- 10 G. Li, Q. Xue, J. Feng and W. Sui, Electrochemical biosensor based on nanocomposites film of thiol graphene–thiol chitosan/nano gold for the detection of carcinoembryonic antigen, *Electroanalysis*, 2015, **27**, 1245–1252.
- 11 R. Li, F. Feng, Z. Chen, Y. Bai, F. F. Guo, F. I. Wu and G. Zhau, Sensitive detection of carcinoembryonic antigen using surface plasmon resonance biosensor with gold nanoparticles signal amplification, *Talanta*, 2015, **140**, 143–149.
- 12 X. Yang, Y. Zhuo, S. Zhu, Y. Luo, Y. Feng and Y. Xu, Selectively assaying CEA based on a creative strategy of gold nanoparticles enhancing silver nanoclusters' fluorescence, *Biosens. Bioelectron.*, 2015, **64**, 345–351.
- 13 H. Miao, L. Wang, Y. Zhuo, Z. Zhou and X. Yang, Label-free fluorimetric detection of CEA using carbon dots derived from tomato juice, *Biosens. Bioelectron.*, 2016, **86**, 83–89.
- 14 F. Liu, H. Zhang, Z. Wu, H. Dong, L. Zhou, D. Yang, Y. Ge, C. Jia, H. Liu, Q. Jin, J. Zhao, Q. Zhang and H. Mao, Highly sensitive and selective lateral flow immunoassay based on magnetic nanoparticles for quantitative detection of carcinoembryonic antigen, *Talanta*, 2016, **161**, 205–210.
- 15 W. Qin, K. Wang, K. Xiao, Y. Hou, W. Lu, H. Xu, Y. Wo, S. Feng and D. Cui, Carcinoembryonic Antigen Detection with “Handing”-Controlled Fluorescence Spectroscopy Using a Color Matrix for Point-of-care Applications, *Biosens. Bioelectron.*, 2016, **90**, 508–515.
- 16 R. I. Stefan-van Staden, L. A. Gugoasa, B. Calenic, J. F. van Staden and J. Legler, Screening of children saliva samples for bisphenol A using stochastic, amperometric and multimode microsensors, *Anal. Chem. Res.*, 2014, **1**, 1–7.
- 17 L. A. Gugoasa, R. I. Stefan-van Staden, B. Calenic and J. Legler, Multimode sensors as new tools for assessing the levels of testosterone, dihydrotestosterone and estradiol in children's saliva, *J. Mol. Recognit.*, 2015, **28**, 10–19.
- 18 F. Pogacean, C. Socaci, S. Pruneanu, A. R. Biris, M. Coros, L. Magerusan, G. Katona, R. Turcu and G. Borodi, Graphene based nanomaterials as chemical sensors for hydrogen peroxide – A Comparison study of their intrinsic peroxidase catalytic behavior, *Sens. Actuators, B*, 2015, **213**, 474–483.
- 19 S. Pei and H.-M. Cheng, The reduction of graphene oxide, *Carbon*, 2012, **50**, 3210–3228.
- 20 F. Bensouici, T. Souier, A. A. Dakhel, A. Iratni, R. Tala-Ighil and M. Bououdina, Synthesis, characterization and photocatalytic behavior of Ag doped TiO<sub>2</sub> thin film, *Superlattices Microstruct.*, 2015, **85**, 255–265.
- 21 H. P. Klug and L. E. Alexander, *X-ray diffraction procedures for polycrystalline and amorphous materials*, John Wiley & Sons, New York, 1974.

