

Cite this: *RSC Adv.*, 2018, 8, 39561

# A disposable microcapsule array chip fabricated by ice printing combined with isothermal amplification for Salmonella DNA detection†

Enqi He,<sup>†</sup> Ting Cao,<sup>†a</sup> Liangyuan Cai,<sup>a</sup> Dan Guo,<sup>b</sup> Yinglin Zhou,<sup>†\*a</sup>  
Xinxiang Zhang<sup>\*a</sup> and Zhihong Li<sup>\*cd</sup>

In this work, a novel microcapsule array chip was fabricated for the detection of Salmonella DNA by integrating an ice-printing technique with DNA isothermal amplification. Reaction solutions were previously sealed in the microcapsule array chip via ice printing. To protect the relatively fragile DNA isothermal amplification system, an extra polystyrene (PS) film was introduced to isolate the reaction solution from photopolymer precursor, which was proved to be a vital step for providing a clean and stable environment for DNA amplification reaction. Detection operation can be done by simply injecting sample DNA into the microcapsule by an easily accessible syringe, and the result can be directly obtained through color change within 90 minutes. This method shows good sensitivity, specificity and stability.

Received 22nd August 2018  
Accepted 18th November 2018

DOI: 10.1039/c8ra07045h

rsc.li/rsc-advances

## 1. Introduction

Portable, disposable, and cost-effective devices that can be easily operated by inexperienced personnel with only a few steps are called for in point-of-care assays.<sup>1,2</sup> Motivated by this need, test papers are developed and commercialized with diverse designs,<sup>3–5</sup> among which the HCG Test Strip for early pregnancy testing<sup>6</sup> is the most successful biochemical trial. However, most biochemical assays, especially those concerning DNA-related reactions, require strict experimental conditions that a test paper cannot provide. Firstly, a sealed environment is necessary for system storage, because fragile reagents like bio-enzymes may be deactivated in open conditions due to possible contamination or oxidation. Secondly, because most DNA-related reactions need incubation at a certain temperature, closed small vessels are required to contain reaction solutions, which should not change or volatilize during incubation. Additionally, it is advantageous if the device would endure relatively high temperature for hours.

Thirdly, the device should suit signal readout methods. For example, a transparent device is preferred for colorimetric assays and non-fluorescence substrate material is preferred for fluorescence-based assays. Fourthly, if DNA amplification reaction is required, it is extremely important that the reaction device must be disposable and should not be opened when discarded to avoid the contamination of DNA amplification products. The reason is that the released product DNA may interfere with the next assays because of the high sensitivity of the amplification method.

Most of the significant isothermal amplification devices are fabricated using a soft lithography method instead of conventional top-down lithography process with the advantages of low toxicity, low cost, wide substrate ranges especially non-flat surfaces, and no need for a photo-reactive surface.<sup>7</sup> Generally, a liquid precursor of polymers is poured over the mold or the sacrificial layer and then liquid precursors are cured into the solid by incubation at some certain conditions for a period of time. For example, polydimethylsiloxane (PDMS), which has been widely used in recent isothermal amplification devices<sup>8–10</sup> showing easy and flexible fabrication process and good biocompatibility, is cured to solid at 70 °C for almost 1 h. For most previous cases, devices are completely made before integrated with biological systems, and thus can hardly achieve pre-packaged bio-systems, for bio-reagents may suffer from deactivation under long-time incubation at high temperature, and the liquid solution cannot stay in a certain shape for the liquid precursors to be stereotyped.

Our group has developed a microcapsule chip based on 3D ice-printing technique<sup>11</sup> that is promising address this issue. In general, a small volume of reaction solution is printed on

<sup>a</sup>Beijing National Laboratory for Molecular Sciences (BNLMS), Key Laboratory of Bioorganic Chemistry and Molecular Engineering of Ministry of Education, College of Chemistry, Peking University, Beijing 100871, China. E-mail: zhouyl@pku.edu.cn; zxx@pku.edu.cn; Fax: +86-10-62754112; Tel: +86-10-62754112

<sup>b</sup>State Key Laboratory of Tribology, Department of Mechanical Engineering, Tsinghua University, Beijing 100084, China

<sup>c</sup>National Key Laboratory of Science and Technology on Micro/Nano Fabrication, Institute of Microelectronics, Peking University, Beijing 100871, China. E-mail: zhhl@pku.edu.cn; Fax: +86-10-62751789; Tel: +86-10-62766581

<sup>d</sup>Center for Nano and Micro Mechanics, Tsinghua University, Beijing 100084, China

† Electronic supplementary information (ESI) available. See DOI: 10.1039/c8ra07045h

‡ These two authors contribute equal to this work.

a cooled membrane substrate and is frozen into ice. Then transparent photopolymer is poured onto the ice spot and polymerized under LED to seal the reaction solution. The fabricated microcapsules can be stored below 0 °C for long-term preservation. When in use, a small volume of sample solution (1–2 µL) is injected directly into the microcapsule for simple visual detection. This technique is potential to fit bio-reagents containing detecting devices in two ways: firstly, the low temperature of ice keeps the bio-reagents in a relatively stable physical and chemical state, making it compatible with the entire device fabrication process. Secondly, using this bottom-up one-step molding design, bio-system can be pre-packed in the device in the device fabrication process, which is significant for easy and rapid detection. Most existing isothermal DNA amplification devices cannot achieve the pre-pack function, for they have to be completely made before integrated with biological systems.

On-site point-of-care detection of pathogens has become increasingly important in the fields of food safety testing, environmental monitoring and human health care.<sup>12–16</sup> Since the abundance of pathogens is usually very low, DNA amplification is essential in the detection process. Loop-mediated isothermal amplification (LAMP)<sup>17–20</sup> is an isothermal method for nucleic acids amplification with high specificity and sensitivity using a set of four or six specially designed primers. Compared with the most widely applied DNA amplification method called polymerase chain reaction (PCR),<sup>21,22</sup> which needs specific temperature cycling instrument, LAMP can be carried out in easily accessible thermal controllers such as ovens at around 65 °C. Moreover, the LAMP result can be simply determined through turbidity<sup>23,24</sup> or color change,<sup>25,26</sup> which makes it remarkably advantageous in point-of-care assays at home or in the field.<sup>27,28</sup>

In this work, we first integrated ice-printing technique with LAMP to fabricate a novel microcapsule array chip for Salmonella DNA detection. However, our previous ice-printing technique is not adequate for DNA isothermal amplification-based biochemical assay, mainly because raw materials of photopolymer (unpolymerized monomers and polymerization initiators) may exude during high-temperature incubation, and thus interfere with DNA isothermal amplification system. Additionally, the porous structure formed by photo polymerization might adsorb reagents during incubation process. To address this issue, we herein improved 3D ice-printing technique to fit relatively fragile DNA isothermal amplification system. A polystyrene (PS) film is fabricated encapsulating icy reaction solution before photopolymer sealing by solvent volatilization of PS solution. The PS film isolates and protects the reaction solution from sealing photopolymer, providing a clean and stable environment for bio-enzyme based reaction. Thus, this design well suits LAMP reaction and is successfully applied into Salmonella DNA detection.

## 2. Experimental section

### 2.1. Materials and reagents

DNA oligonucleotides were synthesized by Sangon Biotech. Co., Ltd. (Shanghai, China), and have the following

sequences: 5'-GGCGATATTGGTGTATTATGGGG-3', F3; 5'-AACGATAAACTGGACCACGG-3', B3; 5'-GACGACTGGTACTGATC-GATAGTTTTTC AACGTTTCCTGCGG-3', FIP; 5'-CCGGTGAAATTATCGC CACACAAAACCCACCGCCAGG-3', BIP; 5'-GACGAAAGAGCGTGGTAATTAAC-3', LF; 5'-GGGCAATTCGTTATTGGCGATAG-3', LB. Bst 2.0 WarmStart polymerase, 10× Isothermal Amplification Buffer (200 mM Tris-HCl, 500 mM KCl, 100 mM (NH<sub>4</sub>)<sub>2</sub>SO<sub>4</sub>, 20 mM MgCl<sub>2</sub>, 1% Triton X-100, pH 8.8 at 25 °C), and MgSO<sub>4</sub> (100 mM) were obtained from New England Biolabs (Beverly, MA). Deoxy-nucleotide solution mixture (dNTPs) was provided by Takara Biotechnology Co. Ltd (Dalian, China). Calcein and MnCl<sub>2</sub> were purchased from Sigma-Aldrich (St. Louis, MO, USA). Salmonella DNA, *Escherichia coli* DNA, and Shigella DNA were extracted right from the corresponding bacteria using bacteria genomic DNA extraction kit from Tiangen Biotech Co. Ltd (Beijing, China), and are diluted into certain concentrations. Commercial milk purchased from local supermarkets was mixed with overnight cultured Salmonella bacteria to obtain contaminated milk, which was then treated by the same bacteria genomic DNA extraction kit to get the injection sample. Other reagents were at least of analytical reagent grade. All solutions were prepared with ultrapure water purified with a Milli-Q purification system (Branstead, MA, U.S.A.).

### 2.2. Preparation of LAMP reaction solution

LAMP primers are previously dissolved into 100 µM separately in pure water, and afterwards 25× LAMP primer mix is prepared as the following concentration: *c* (FIP) = 40 µM, *c* (BIP) = 40 µM, *c* (F3) = 5 µM, *c* (B3) = 5 µM, *c* (LF) = 10 µM, *c* (LB) = 10 µM. Calcein is dissolved in 1× Isothermal Amplification Buffer into 1 mM. MnCl<sub>2</sub> is dissolved in pure water into 10 mM. Using the above stock solutions, LAMP reaction solution is freshly prepared right before the fabrication of microcapsule array chip. The final concentrations of all reagents are listed below: Isothermal Amplification Buffer 1× (contains 2 mM MgSO<sub>4</sub>), extra MgSO<sub>4</sub> (6 mM, 8 mM in total), LAMP primer mix 1×, *c* (dNTP) = 2.8 mM each, *c* (calcein) = 100 µM, *c* (MnCl<sub>2</sub>) = 1 mM, *c* (Bst 2.0 WarmStart polymerase) = 480 U mL<sup>-1</sup>.

### 2.3. Chip fabrication based on ice-printing technique

The microcapsule array chip is fabricated using our homemade ice-printing instrument consisted of a set of X-Y-Z-axis platform to move the printing nozzle to a certain spot, a printing pump to pump reaction solution, a printing nozzle to drop solution and a Peltier cooler driven by condensate water to maintain cool surface (below -20 °C) in the fabrication process. The whole system is placed inside a glovebox filled with pure nitrogen to maintain dry atmosphere, avoiding frosting on the cold surface.

The main frame of the chip is made by simply sticking polyethylene terephthalate (PET) film to plastic cofferdam by commercial glue. The inner side of the PET film is cleaned by ethanol and pure water before use. 30% polystyrene solution is dissolved in chloroform (different concentrations ranging from



15–40% and different solvents are prepared in optimization experiments).

The detailed fabrication procedure is illustrated as follows. Firstly, the printing nozzle is moved to a certain spot and pumps out 10  $\mu\text{L}$  LAMP reaction solution onto the PET film. The solution is frozen into ice as soon as it reaches the cool surface. The radius of curvature of the ice droplet is measured to be 4.6 mm, and the contact angel is  $68^\circ$  (Fig. S1†). Multiple droplets can be pumped successively to form an array. Secondly, 8  $\mu\text{L}$  the 30% PS solution is dropped onto the icy solution to form a contact film after solvent volatilization. It is important to make sure that the PS solution covers the ice droplet well. It takes about 20 minutes before the film gets dried throughout. Thirdly, photopolymer precursor is added slowly onto the chip to avoid bubbles, and it polymerizes under 420 nm LED for 5 minutes to seal the microcapsule. With good light transmittance property, PS film and polymer would not hinder visual colorimetric detection (Fig. S2†). In the whole process, the system is supposed to be cooled to below  $-20^\circ\text{C}$  to guarantee the reaction solution frozen. The chip is stored at  $-20^\circ\text{C}$  before use.

#### 2.4. Direct visual detection of Salmonella DNA

The chip is taken out from refrigerator and placed in room temperature for 5 minutes to let the icy solution melt right before use. 1  $\mu\text{L}$  of Salmonella DNA (or *Escherichia coli* DNA, and Shigella DNA) is injected into the microcapsule by a syringe, and the injecting hole is directly sealed by transparent tape. Then, the chip is incubated at  $65^\circ\text{C}$  for 90 minutes. Finally, the colorimetric detection results can be simply observed by naked eyes. Photos are taken by cell phone.

### 3. Results and discussion

#### 3.1. Fabrication of the chip and its principle for Salmonella DNA detection

Fig. 1(A) shows the self-made instrument used in this work, consisted of X-Y-Z-axis platforms, a printing pump, a printing nozzle and a Peltier cooler. The cooler is driven by condensate water system, and can cool the surface to below  $-20^\circ\text{C}$ . The general procedure of fabricating ice-printing microcapsule array is explained in Fig. 1(B). First, 10  $\mu\text{L}$  of reaction solution for LAMP containing Salmonella LAMP primers, DNA polymerase, colorimetric reagents and corresponding buffer is printed on the cold PET substrate and is frozen into ice (a). Then, previous dissolved PS solution is dropped on the ice droplet and a PS film is formed after solvent volatilization (b). Next, photopolymer precursor is added and polymerized under 420 nm LED to seal the microcapsule (c). Multiple microcapsules array can be fabricated at one time on a chip for multiple assays, and the whole fabrication process takes no more than 30 minutes. The obtained chips are stored at  $-20^\circ\text{C}$ .

When in use, sample Salmonella DNA is injected directly into the microcapsule using a syringe after thawing the frozen chip, and the hole is simply sealed with a transparent tape (Fig. S3 and S4†). LAMP reaction is carried out in every





Fig. 1 (A) Ice-printing instrument used in this work. (B) Fabrication process of microcapsule array chip.

microcapsule with 6 primers,<sup>29</sup> consisted of two outer primers (forward primer F3 and backward primer B3), two inner primers (forward inner primer FIP and backward inner primer BIP) and two loop primers (forward loop LF and backward loop LB). The outer primers are used in initial steps to produce starting structures, while the inner primers lead the subsequent



Scheme 1 Principle of LAMP colorimetric detection using metal indicator.





explosive cycling strand displacement reaction, generating broccoli-like final products with different lengths. Calcein (green color), a metal indicator, and  $\text{MnCl}_2$  are previously added to the reaction solution. Before amplification, the solution appears orange, which is the color for  $\text{Mn}^{2+}$  binding with calcein. In the intense amplification reaction process, a large amount of pyrophosphate ions are generated from dNTP, which forms insoluble salt ( $\text{Mn}_2\text{P}_2\text{O}_7$ ) with  $\text{Mn}^{2+}$ . The released calcein binds with free  $\text{Mg}^{2+}$  in the solution, and appears green (Scheme 1). By simply telling the color (orange or green), the amplification conditions can be determined, which is suitable for point-of-care assays.

### 3.2. Function of the PS film and optimization

Notably, the PS film formation step is very important in the microcapsule fabrication process. Raw materials and polymerization residues may exude during high-temperature incubation, and thus influences the DNA isothermal amplification system, resulting in inefficient DNA amplification. Additionally, the porous structure formed by photo polymerization might adsorb reagents during incubation process, which influences the homogeneity of the reaction solution. Moreover, the metal indicator-based colorimetric system is also influenced. As Fig. 2 shows, reaction system would change color from orange to green in the absence of the target (no LAMP reaction) without fabrication of PS film, which severely interferes the colorimetric assay.

Since the PS film is formed by solvent volatilization of the PS solution, the nature of the solvent determines the quality of the film as well as the film formation time. Four common PS solvents, namely dichloromethane ( $\text{CH}_2\text{Cl}_2$ ), chloroform ( $\text{CHCl}_3$ ), tetrahydrofuran (THF), and toluene, were tried and the results are explained as follows respectively. THF may absorb water from the ice droplet due to its water-miscible property, which leads to uneven dispersion of PS, resulting in multiple small pores on the film. Because the high volatility of dichloromethane, the outer side of dissolved PS gets dried immediately, which hinders throughout volatilization of the inner side of the film. On the contrary, toluene volatilizes very slowly, and the volatilization process takes more than 1.5 h, which wastes too much time. Compared with the above three solvents, chloroform is the most appropriate PS solvent with moderate volatility, which helps to form an even and compact PS film within 20 minutes. PS concentration, and the volume of PS solution are also optimized to 30% and 8  $\mu\text{L}$  to obtain best film quality.




Fig. 2 Comparison of the microcapsule solution conditions after incubation in the absence of target with and without PS film.




Fig. 3 Sensitivity performance of the microcapsule array chip for visual Salmonella DNA detection. Salmonella DNA is previously extracted from Salmonella and diluted to different concentrations (6 copies to  $6 \times 10^6$  copies). DNA-free water is used as negative control. (A) Photos of the sensitivity assay. (B) The normalized integrated optical densities (IOD) of the green color from the photos under different concentrations. Error bars represent results from three independent experiments.

### 3.3. Performance of Salmonella DNA detection

The sensitivity performance of this ice-printed microcapsule array chip for Salmonella DNA detection is shown in Fig. 3(A). 6, 60,  $6 \times 10^2$ ,  $6 \times 10^3$ ,  $6 \times 10^4$ ,  $6 \times 10^5$ , and  $6 \times 10^6$  copies of Salmonella DNA and DNA-free water (negative control) are injected into microcapsules, respectively, and incubated in an oven or any heater at hand at  $65^\circ\text{C}$  for 90 minutes. Simple choices of “orange or green” can be immediately and clearly made by naked eyes just on the spot owing to the colorimetric principle of LAMP and the transparent property of the whole chip. Notably, down to 600 copies of Salmonella DNA can be detected, which is comparable to the limit of detection of other visual detection methods based on LAMP and better than conventional PCR<sup>29–31</sup> (Table S1†). To better illustrate this test, we measure the Integrated optical densities (IOD) of the green color (parameter: hue 48–70) using software Image J. The difference of green color IOD between experimental group and negative control is calculated and normalized by dividing by the IOD of negative control, *i.e.*,

$$\text{Normalized IOD} = \frac{\text{Measured IOD} - \text{IOD of negative control}}{\text{IOD of negative control}} \times 100\%$$

The processed normalized IODs are showed in Fig. 3(B), which indicates no gradual color change among different concentrations, suggesting that our device a “yes or no” sensor





Fig. 4 Specificity tests of the microcapsule array chip for visual Salmonella DNA detection. The concentrations of Salmonella DNA, *Escherichia coli* DNA and Shigella DNA are all  $10^5$  copies. DNA-free water is used as negative control.




Fig. 5 Visual detection of Salmonella contaminated milk. DNA-free water is used as negative control.

like most LAMP based colorimetric sensors<sup>25,26</sup> rather than a quantitative sensor.

Fig. 4 shows the specificity test of Salmonella DNA assay against DNA-free water, *Escherichia coli* DNA, and Shigella DNA. After LAMP reaction, the microcapsule with Salmonella DNA injection turns green, however, other microcapsules stay orange. This visual result indicates good DNA discrimination ability of our detection. The stability performance, which is tested after storing the microcapsule chip in freeze for 0–15 days, is shown in Fig. S5† Distinguished color change of the Salmonella DNA assay is observed compared with the negative control, which demonstrates good stability. Furthermore, the stability of the materials used is also tested (Fig. S6†).

To further verify its ability for Salmonella detection in real samples, we treat commercial milk using our device. Salmonella bacteria is cultured overnight and mixed with commercial milk. The mixture is treated following DNA extraction process to obtain the injection sample as usual. In Fig. 5, the left microcapsule injected the contaminated milk sample gives a positive green color.

## 4. Conclusions

We proposed a microcapsule array chip for visual Salmonella DNA detection based on improved ice-printing technique, which meets the needs of point-of-care systems with fragile bio-agents on all aspects including storage, incubation needed reaction and sealed discard without contamination. The chip fabrication process is simple and fast without traditional complicated lithography techniques. Only 10  $\mu$ L reaction solution is contained in each microcapsule, which gives rise to reagents saving. Notably, with the isothermal amplification method, no annealing and temperature cycling steps are needed in the whole DNA amplification process, which stays away from instrumental constraints and can be easily carried out

in ordinary heating conditions like water baths or ovens. LAMP also enables colorimetric detection based on orange to green color change, leading to easy, rapid, and on-spot result determination. With excellent sensitivity, specificity and stability, this system is promising in real point-of-care detection of Salmonella DNA as well as other pathogen DNA.

Hopefully, automated production can be realized to save labor and further improve efficiency owing to the simple and programmable chip fabrication process. Moreover, by fabricating multiple microcapsules and using LAMP solutions targeting different analytes on one chip, high-throughput and multiple-target assays can be further performed. This novel device also provides an alternative method on other bio-assays with fragile reagents, for instance, nucleic acids and proteins. Our future work will focus on integrating DNA extraction into our device by designing microfluidics systems.

## Conflicts of interest

There are no conflicts to declare.

## Acknowledgements

This work was supported by the National Key Research and Development Program of China (No. 2016YFA0200802) and the National Natural Science Foundation of China (No. 21675004 & No. 21575005).

## Notes and references

- 1 E. Petryayeva and W. R. Algar, *RSC Adv.*, 2015, 5, 22256–22282.
- 2 Z. Fang, W. Wu, X. Lu and L. Zeng, *Biosens. Bioelectron.*, 2016, 56, 192–197.
- 3 N. K. Thom, G. G. Lewis, K. Yeung and S. T. Phillips, *RSC Adv.*, 2014, 4, 1334–1340.
- 4 Q. Ju, M. O. Noor and U. J. Krull, *Analyst*, 2016, 141, 2838–2860.
- 5 J. R. Choi, J. Hu, R. Tang, Y. Gong, S. Feng, H. Ren, T. Wen, X. Li, W. A. B. W. Abas, B. Pingguan-Murphy and F. Xu, *Lab Chip*, 2016, 16, 611–621.
- 6 C. Gnoth and S. Johnson, *Geburtshilfe Frauenheilkd.*, 2014, 74, 661–669.
- 7 Y. Xia and G. M. Whitesides, *Annu. Rev. Mater. Sci.*, 1998, 28, 153–184.
- 8 K. Hsieh, A. S. Patterson, B. S. Ferguson, K. W. Plaxco and H. T. Soh, *Angew. Chem., Int. Ed.*, 2012, 124, 4980–4984.
- 9 Y. Liu, Y. Zhao, Y. Qin, X. Du, Q. Wang and J. Lyu, *RSC Adv.*, 2016, 6, 13399–13406.
- 10 X. Fang, Y. Liu, J. Kong and X. Jiang, *Anal. Chem.*, 2010, 82, 3002–3006.
- 11 H. Z. Zhang, F. T. Zhang, X. H. Zhang, D. Huang, Y. L. Zhou, Z. H. Li and X. X. Zhang, *Anal. Chem.*, 2015, 87, 6397–6402.
- 12 L. Schwenkbier, S. Pollok, A. Rudloff, S. Sailer, D. Cialla-May, K. Weber and J. Poppa, *Analyst*, 2015, 140, 6610–6618.
- 13 S. J. Oh, B. H. Park, G. Choi, J. H. Seo, J. H. Jung, J. S. Choi, D. H. Kim and T. S. Seo, *Lab Chip*, 2016, 16, 1917–1926.



- 14 Z. Fang, W. Wu, X. Lu and L. Zeng, *Biosens. Bioelectron.*, 2014, **56**, 192–197.
- 15 Y. Xia, Z. H. Liu, S. Q. Yan, F. Yin, X. J. Feng and B. F. Liu, *Sens. Actuators, B*, 2016, **228**, 491–499.
- 16 M. Pacheco, B. Jurado-Sanchez and A. Escarpa, *Anal. Chem.*, 2016, **90**(4), 2912–2917.
- 17 T. Notomi, H. Okayama, H. Masubuchi, T. Yonekawa, K. Watanabe, N. Amino and T. Hase, *Nucleic Acids Res.*, 2000, **28**, e63.
- 18 C. D. Guevara, V. V. Swaminathan, B. Reddy, J. Huang, Y. Liu and R. Bashir, *RSC Adv.*, 2016, **6**, 103872–103887.
- 19 M. S. Draz and X. Lu, *Theranostics*, 2016, **6**, 522–523.
- 20 L. Zhu, Y. Xu, N. Cheng, P. Xie, X. Shao, K. Huang, Y. Luo and W. Xu, *Sens. Actuators, B*, 2017, **242**, 880–888.
- 21 R. Saiki, D. Gelfand, S. Stoffel, S. Scharf, R. Higuchi, G. Horn, K. Mullis and H. Erlich, *Science*, 1988, **239**, 487–491.
- 22 H. A. Erlich, D. Gelfand and J. J. Sninsky, *Science*, 1991, **252**, 1643–1651.
- 23 Y. Mori, K. Nagamine, N. Tomita and T. Notomi, *Biochem. Biophys. Res. Commun.*, 2001, **289**, 150–154.
- 24 Y. Mori, M. Kitao, N. Tomita and T. J. Notomi, *J. Biochem. Biophys. Methods*, 2004, **59**, 145–157.
- 25 N. Tomita, Y. Mori, H. Kanda and T. Notomi, *Nat. Protoc.*, 2008, **3**, 877–882.
- 26 H. Nikbakht, P. Gill, A. Tabarraei and A. Niazi, *RSC Adv.*, 2014, **4**, 13575–13580.
- 27 M. Safavieh, M. U. Ahmed, E. Sokullu, A. Ng, L. Braescu and M. Zourob, *Analyst*, 2014, **139**, 482–487.
- 28 X. Fang, H. Chen, S. Yu, X. Jiang and J. Kong, *Anal. Chem.*, 2010, **83**, 690–695.
- 29 C. Techathuvanan, F. A. Draughon and D. H. D'Souza, *J. Food Sci.*, 2010, **75**, M165–M172.
- 30 J. L. Yang, G. P. Ma, R. Yang, S. Q. Yang, L. Z. Fu, A. C. Cheng, M. S. Wang, S. H. Zhang, K. F. Shen and R. Y. Jia, *J. Appl. Microbiol.*, 2010, **109**, 1715–1723.
- 31 J. Li, L. Zhai, X. Bie, Z. Lu, X. Kong, Q. Yu, F. Lu, C. Zhang and H. Zhao, *Food Control*, 2016, **60**, 230–236.

