


 Cite this: *J. Anal. At. Spectrom.*, 2019, **34**, 2340

## Correction: Preparation and characterization of primary magnesium mixtures for the *ab initio* calibration of absolute magnesium isotope ratio measurements

 Björn Brandt,<sup>a</sup> Jochen Vogl,<sup>ID</sup>\*<sup>a</sup> Janine Noordmann,<sup>b</sup> Angela Kaltenbach<sup>b</sup> and Olaf Rienitz<sup>b</sup>

DOI: 10.1039/c9ja90057h

[www.rsc.org/jaas](http://www.rsc.org/jaas)

 Correction for 'Preparation and characterization of primary magnesium mixtures for the *ab initio* calibration of absolute magnesium isotope ratio measurements' by Björn Brandt *et al.*, *J. Anal. At. Spectrom.*, 2016, **31**, 179–196.

Eqn (3) on p. 184 describes how the density of air  $\rho_{\text{air}}$  was calculated from the ambient conditions air pressure ( $p$ ), relative humidity ( $\varphi$ ) of the air, and air temperature  $\vartheta$ . In the denominator of the original there was a typo:  $\varphi$  needs to be replaced with  $\vartheta$ . Eqn (3) reads correctly:

$$\rho_{\text{air}} = \frac{0.348444 \text{ kg m}^{-3} \text{ hPa}^{-1} \times p - \varphi(0.252 \text{ kg m}^{-3} \text{ }^{\circ}\text{C}^{-1} \times \vartheta - 2.0582 \text{ kg m}^{-3})}{273.15 + \vartheta \text{ }^{\circ}\text{C}^{-1}}$$

The Royal Society of Chemistry apologises for these errors and any consequent inconvenience to authors and readers.

<sup>a</sup>BAM Federal Institute for Materials Research and Testing, Unter den Eichen 87, 12205 Berlin, Germany. E-mail: [jochen.vogl@bam.de](mailto:jochen.vogl@bam.de)

<sup>b</sup>Physikalisch-Technische Bundesanstalt (PTB), Bundesallee 100, 38116 Braunschweig, Germany. E-mail: [olaf.rienitz@ptb.de](mailto:olaf.rienitz@ptb.de)

