



Cite this: *Lab Chip*, 2019, **19**, 919

Rotating magnetic particles for lab-on-chip applications – a comprehensive review

C. P. Moerland, L. J. van IJzendoorn and M. W. J. Prins *

Magnetic particles are widely used in lab-on-chip and biosensing applications, because they have a high surface-to-volume ratio, they can be actuated with magnetic fields and many biofunctionalization options are available. The most well-known actuation method is to apply a magnetic field gradient which generates a translational force on the particles and allows separation of the particles from a suspension. A more recently developed magnetic actuation method is to exert torque on magnetic particles by a rotating magnetic field. Rotational actuation can be achieved with a field that is uniform in space and it allows for a precise control of torque, orientation, and angular velocity of magnetic particles in lab-on-chip devices. A wide range of studies have been performed with rotating MPs, demonstrating fluid mixing, concentration determination of biological molecules in solution, and characterization of structure and function of biomolecules at the single-molecule level. In this paper we give a comprehensive review of the historical development of MP rotation studies, including configurations for field generation, physical model descriptions, and biological applications. We conclude by sketching the scientific and technological developments that can be expected in the future in the field of rotating magnetic particles for lab-on-chip applications.

Received 3rd December 2018,
Accepted 1st February 2019

DOI: 10.1039/c8lc01323c

rsc.li/loc

Introduction

Magnetic nano- and micro-particles (MPs) can be actuated in fluids using applied magnetic fields. Their large surface-to-

volume ratio and the availability of a wide range of particle types and methods for surface functionalization have enabled the application of MPs as carriers (e.g. for biomolecule and cell extraction), as labels (e.g. for biomolecule detection) and as probes in biophysical studies (e.g. to quantify mechanical properties of molecules and cells). Several reviews have discussed principles and applications of magnetic actuation, see e.g. Gijs *et al.* 2010 (MPs in microfluidics),¹ Van Reenen *et al.* 2014 (lab-

Department of Applied Physics, Department of Biomedical Engineering, Institute for Complex Molecular Systems, Eindhoven University of Technology, Eindhoven, The Netherlands. E-mail: m.w.j.prins@tue.nl



Christian Moerland

Christian Moerland received his BSc in 2012 and MSc in 2014 in Applied Physics at Eindhoven University of Technology. He worked as a PhD student in the Nano-Mag project (European 7th framework project on standardization of magnetic nanoparticles) under the supervision of Leo van IJzendoorn and Menno Prins. His PhD research focused on fundamental physics and applications of torque generation in micrometer-sized multicore magnetic particles.



Leo van IJzendoorn

Leo van IJzendoorn is Associate Professor in the group Molecular Biosensing for Medical Diagnostics at Eindhoven University of Technology and he is a core member of the Institute for Complex Molecular Systems at TU/e. He obtained his MSc in Physics (1981) and PhD (1985) at Leiden University on molecular spectroscopy of interstellar ices in the group of Prof. J. M. Greenberg. From 1985–2004 he worked at Philips Research and Eindhoven University of Technology in the field of ion beam analysis. Since 2004 he has investigated particle-based detection of biomolecules in complex environments for point of care applications, including the actuation of magnetic particles and their applications in sensing technologies.



on-chip biosensing),² Schrittwieser *et al.* 2016 (homogeneous bioassays)³ and Munaz *et al.* 2018 (magnetophoresis).⁴

The most well-known method to actuate MPs is by applying a magnetic field gradient that generates a translational force on the MPs. A characteristic property of gradient-based actuation is a non-uniform force on the magnetic particles. A more recently developed magnetic actuation method is to rotate particles by applying a magnetic field of which the direction and/or amplitude varies over time. Rotational actuation can be achieved with a field that is uniform in space and it allows for a precise control of torque, orientation, and angular velocity of MPs in lab-on-chip devices. In the past years a wide range of studies have been performed with rotating MPs, demonstrating applications such as fluid mixing, concentration determination of biological molecules in solution and torsional characterization of biomolecules at the single-molecule level.

This review starts with a brief description of the basic physical principles of rotational actuation of MPs and categorizes the experimental configurations. Then a detailed description is given of the internal structure of MPs, how this underlies their magnetic properties and the torque generated by an applied field, and an overview is given of field-generating setups reported in the literature. Finally we discuss a number of lab-on-chip studies with rotating MPs that exemplify the field and we give an outlook into future scientific and technological developments.

Physical principles and basic configurations

In this section the physical origin of magnetic particle rotation will be addressed and the main experimental configurations found in literature will be categorized.

The rotational force on an object, the torque, is represented with symbol $\vec{\tau}$ and is expressed in the unit N m (New-

ton times meter). A magnetic object that is placed in a magnetic field, experiences a torque of magnetic origin given by:

$$\vec{\tau} = \vec{m} \times \vec{B}. \quad (1)$$

In eqn (1) \vec{m} is the magnetic moment of the object and \vec{B} the magnetic flux density caused by the applied external field.

The torque tends to rotate the object so that the magnetic moment and the magnetic field become aligned; and when they are aligned, the torque reduces to zero. Note that this equation does not contain the gradient of the magnetic field, so a magnetic field that is spatially completely uniform can generate a torque. This is in contrast to the magnetic force, $\vec{F} = \nabla(\vec{m} \cdot \vec{B})$, which is nonzero only in a spatially non-uniform magnetic field, *i.e.* the application of force requires the presence of a gradient in the magnetic field.

Many different MPs are used in research and technology. They can be nanometer sized and consist of a single-domain magnetic crystal or consist of a few fused or agglomerated magnetic crystals. In this paper we will refer to such nanocrystals as “cores”. Larger magnetic particles ranging in size from approximately 200 nm to several micrometers can contain up to tens of thousands of cores embedded in a non-magnetic polymer or glassy matrix. The MPs have a total magnetic moment \vec{m} with a magnitude and orientation that depends on the internal magnetic structure of the MP and on the applied magnetic field.

Every individual core has a magnetic moment of which the orientation can be perturbed by thermal energy ($k_B T$), by an applied magnetic field ($U_m = \vec{m}_{\text{core}} \cdot \vec{B}$), and by core–core interactions. Re-magnetization of each individual core occurs when the perturbation energy is larger than the internal energy barrier for re-magnetization. The magnitude of the barrier depends on the magnetic anisotropy of the core, which has several contributions like its volume and shape. These mechanisms lead to the appearance of magnetic torque in the particle, a topic that is further detailed in section ‘Magnetic cores and magnetic particles’.

The torque and rotation of magnetic particles in a lab-on-chip device can be controlled by applying a magnetic field of which the direction and/or amplitude varies over time. We refer to “uniaxial field actuation” when the amplitude of the field changes over time but not the direction of the field; in literature this is also referred to as AC fields (alternating current). We refer to “rotating field actuation” when the direction of the field changes over time. See Dieckhoff *et al.*⁵ for a discussion on the effects of the types of driving field (uniaxial or rotating).

Fig. 1 sketches five experimental configurations in which rotational actuation of MPs is exploited for different functionalities, namely for mixing (Fig. 1A), for determining the concentration of biological molecules (Fig. 1B and C) or for biophysical characterization of cells or molecules (Fig. 1D and E). In the following paragraphs we will summarize the main application areas based on these experimental



Menno Prins

Menno Prins received his MSc and PhD in physics and worked for nearly 20 years in Philips Research on microfluidics and particle-based sensing technologies for medical applications. Since 2014 he has been full professor at Eindhoven University of Technology, in the Departments of Biomedical Engineering and Applied Physics, where he chairs the Molecular Biosensing cluster. He investigates particle-based technologies for the detection

and continuous monitoring of biomolecules. To stimulate education and innovation in the field of biosensing, he founded and organizes SensUs, the annual international student competition on biosensors for health (www.sensus.org).



configurations, addressing the physical measurement principles, the magnetic properties of the involved particles, and the magnetic actuation devices used in these experiments. A more extensive description of the application areas can be found in the section “Applications of rotating MPs”.

Mixing of fluid can be effectuated by applying a rotating field to a chain of MPs^{2,6–15} (Fig. 1A). The chains are formed due to attractive magnetic dipole–dipole forces between the particles. The resulting chains of magnetic particles have a strong shape anisotropy that causes them to align along the applied magnetic field, so that a rotating field causes rotation of the chains and locally mixes the fluid.

A second functionality of rotating MPs is to determine the concentration of target molecules in a solution (Fig. 1B and C). One technique is to detect a change in hydrodynamic size when target molecules bind to the MPs^{16–29} (Fig. 1B). The binding is made target-specific by coupling capture molecules to the particles, *e.g.* antibodies or aptamers. Preferably small particles are used (40–400 nm diameter) so that the relative volume change induced by the targets is significant. The hydrodynamic size of a MP can be measured *via* the rotational hydrodynamic drag that scales with the volume of the particle. A change of the hydrodynamic size of MPs can be detected by optical or magnetic signals. Optical detection of rotating particles requires an optical asymmetry. Alternatively, one can magnetically measure the rotational response of particles, which requires a magnetic moment that is physically coupled to the orientation of the particle, *i.e.* when the magnetic moment is a remnant moment. An early publication by Chung *et al.*¹⁷ in 2004 describes a measurement of hydrodynamic size by magnetic detection under uniaxial field actuation (AC fields), where a slower response speed was observed when target proteins were present in solution.

Another way to measure the concentration of target molecules is by a cluster assay. Here MPs are provided with capture molecules and form clusters in the presence of target molecules^{30–46} (Fig. 1C). The clusters can be detected by opti-

cal or by magnetic signals. Optical detection exploits the modulation of scattering due to the elongated shape of clusters compared to single particles, by measuring the amplitude or the frequency of the oscillations in the scattered light. An early publication by Baudry⁸ describes the dependence of the amplitude of the oscillations on the formation of particle clusters.

A third functionality of rotating MPs is to probe mechanical properties of single cells (Fig. 1D) or single molecules (Fig. 1E). In cell studies, the cells are immobilized on a substrate and MPs are attached to the cell membrane. The mechanical properties of the membrane and underlying cell structure are probed by applying a uniaxial or a rotating magnetic field, also called magnetic twisting cytometry.^{47–60} The deformation of the cell can be determined by measuring the magnetic response of ensembles of MPs, or by tracking the movement and rotation of single particles in an optical microscope. The anisotropy of cellular rigidity⁶⁰ as well as influence of the binding mechanism of the MP to the cell⁵¹ was investigated.

Studies on the mechanical properties of single molecules have focused on macromolecules like DNA and proteins^{61–89} (Fig. 1E). The molecules are sandwiched between a magnetic particle and the surface of a fluid cell. A mechanical torque is applied by magnetic fields and the rotation of individual particles is tracked using an optical microscope. Experiments have shown that the macromolecules behave as torsional springs, that DNA has supercoiling behavior, and that the influence of proteins and enzymes binding to DNA can be studied.

In the following sections we will discuss how torque is generated in MPs, the magnetic field actuation systems that have been used, and subsequently we will review lab-on-chip studies based on application and control of torque and rotation.

Magnetic cores and magnetic particles

The magnetic content of a magnetic particle typically consists of ferro- or ferri-magnetic nano-crystals. These nano-crystals




Fig. 1 Schematic overview of experimental configurations that have been studied with rotating magnetic MPs. Related literature is indicated in Table 2. Panel A: Chains of MPs are rotated and used for mixing of fluids or to increase the molecular capturing rate for biosensor applications. Panel B: MPs capture target molecules from solution, which changes the hydrodynamic radius of the MPs. The hydrodynamic radius is determined by rotational actuation, detecting the magnetic or optical scattering response in a time varying magnetic field. Panel C: The concentration of target molecules is determined *via* target-induced formation of clusters of MPs. The clusters are measured by rotational actuation, detecting the magnetic or optical scattering response in a time varying magnetic field. Panel D: Magnetic particles are attached to cells and are rotated by an applied field. The MPs probe the mechanical compliance of the cell and its membrane. Panel E: Mechanical properties of single biological molecules (DNA/protein/enzyme) are probed by sandwiching the molecule between a substrate and a rotating magnetic particle.



consist of a single-crystalline domain with strong internal magnetic coupling, so that they represent a single magnetic domain that is magnetized along an easy axis of magnetic anisotropy. Such nano-crystals are referred to as cores. When an external magnetic field is applied, the moment of the core tends to align with the field. The reorientation time of a single magnetic moment depends on the size, shape and atomic structure of the core. In 1949 Louis Néel developed a theory that relates the thermal magnetic relaxation time (in the absence of an applied field) of a single core to its volume:

$$\tau_N = \tau_0 \exp\left(\frac{KV}{k_B T}\right), \quad (2)$$

where τ_N is the Néel relaxation time, V the volume of the core and K its effective magnetic anisotropy energy density along the easy axis, which is influenced by the lattice-, shape- and surface anisotropy of the core. Typically values of K are reported to be in order of 10^4 J m^{-3} and depend on the shape and crystal structure of the cores,^{90,91} $k_B T$ is the thermal energy and $1/\tau_0$ is an attempt frequency⁹² that is typically taken as a constant. For iron oxide cores the attempt frequency is between 10^8 and 10^{10} per second. The resulting relaxation times range over many orders of magnitude; typically from a few nanoseconds to microseconds for cores with a diameter of a few nanometers, and several orders of magnitude longer for cores of 20 nm or larger.

Core–core interactions such as dipolar and/or exchange interactions are not taken into account in the Néel model, but these can also have a strong influence on the magnetic relaxation time.⁹³ The core–core interactions are defined by the mesoscopic structure of the MP. Three types of mesoscopic structures can be distinguished. The simplest case is a single nanocrystal that behaves as a single magnetic domain (Fig. 2A). The second type consists of multiple fused single crystals that behave as a single magnetic entity due to the strong magnetic (exchange) coupling (Fig. 2B). The third type consists of an agglomerate of single nanocrystals that do not have exchange coupling but do have strong dipolar interactions due to their close proximity (Fig. 2C). Such nanocrystal-

based particles are used in many applications, including hyperthermia and magnetic particle imaging.⁹⁴ Reviews on the synthesis methods of magnetic nanoparticles have been written by Willard *et al.* in 2004 (ref. 95) and more recently by Wu *et al.* in 2015.⁹⁶ Classifications of magnetic nanoparticles, synthesis methods and analysis methods have been reviewed and summarized by Bogren *et al.* in 2015 (ref. 97) and the standardization has been described by Wells *et al.* in 2017.⁹⁸

Fig. 2D shows the architecture of a multi-core MP that is typically used in lab-on-chip applications. The magnetic cores (brown) are fixed within a non-magnetic material (gray) and the outer surface of the MP is bio-functionalized (green).⁹⁹

In 1976 John Ugelstad *et al.*⁹⁹ described an emulsion polymerization process to produce porous monodisperse micrometer-sized polystyrene particles in which magnetic cores can be fixed. The cores are distributed quite homogeneously throughout the particle.¹⁰⁰ The creation of these MPs was a major breakthrough in the field. The particles show strong magnetic properties due to the large number of nanometer sized cores inside the particle.

A MP that is placed in a magnetic field can respond by internally changing its magnetization direction, called Néel relaxation, or by changing its physical orientation, called Brownian relaxation.¹⁰¹ Both processes lower the magnetic energy by orienting the magnetic moments of the cores toward the external field. There are two limiting cases of Néel relaxation, namely fast relaxation for superparamagnetic particles and slow relaxation for magnetically blocked particles (particles with a barrier for re-magnetization that exceeds the thermal energy). In the superparamagnetic limit the remagnetization barrier of the cores is smaller than the thermal energy so the magnetic relaxation time is short, *i.e.* shorter than the timescale of the lab-on-chip experiment. The magnetization of an ensemble of superparamagnetic cores in thermal equilibrium can be modeled like atomic paramagnets, only with a significantly larger magnetic moment. This implies that the magnetic moment of the MP is zero in absence of an external field. For the small field regime (B typically $< 50 \text{ mT}$) the magnetization is approximately linear with the external field and occurs on a timescale that is short




Fig. 2 Magnetic cores and magnetic particle. (A) TEM image of a single-nanocrystal core (CSIC11 produced by Consejo Superior de Investigaciones Científicas, Spain). (B) TEM image of a core consisting of fused crystals (FeraSpin-L produced by nanoPET, Germany). (C) TEM image of a multi-nanocrystal agglomerate (MM06 produced by MicroMod, Germany). (D) Example of an architecture of a bio-functional MP.⁹⁸ (1) magnetic cores in brown, (2) core aggregate, (3) non-magnetic shell in gray, and (4) biofunctional layer in green. The particles shown in panels A–C were synthesized in the NanoMag FP7 project (<http://nanomag-project.eu/>).



compared to the timescale of the experiment. One of the first to use the term “superparamagnetism” were Bean and Livingston in 1959.¹⁰² In the same year they also reported on the possibility to apply a torque to a powder of aligned magnetic nanoparticles.¹⁰³ Due to the magnetic anisotropy of the cores, the magnetic susceptibility, χ , is anisotropic and torque is generated when the field is not aligned with the easy axis.¹⁰⁴ For blocked particles, the re-magnetization barrier of the cores is larger than the thermal energy and the MP has a remnant magnetic moment, with a magnetic relaxation time that is much longer than the experimental timescale. To apply a torque on ferromagnetic MPs they are typically magnetized by applying a short (<1 s) strong (>100 mT) magnetic pulse.^{53,105} The pulse creates a remnant moment that is constant over time during the measurement. When the resulting remnant moment is placed in a field that is small (much smaller than the original pulse, *i.e.* <100 mT), and misaligned with the moment, then a torque is exerted.

For an ensemble of MPs, the total remnant moment can be measured using a commercially available magnetometer, for example a superconducting quantum interference device (SQUID) or a vibrating sample magnetometer (VSM). For an individual MP, the remnant moment can be quantified *via* the magnetic torque by tracking the rotation of a single particle in an optical microscope.¹⁰⁶ For a particle that is not attached and that can freely rotate in solution at the same frequency as the applied magnetic field, the torque on the particle is approximately equal to its rotational hydrodynamic drag. The hydrodynamic drag of a smooth sphere with radius R in a fluid with dynamic viscosity η , is given by the Stokes equation, which together with eqn (1) gives the following equation for the angular motion of the particle:

$$\vec{m} \times \vec{B} - 8\pi\eta R^3 \frac{d\vec{\theta}}{dt} = 0, \quad (3)$$

where $\vec{\theta}$ is the angular orientation of the particle.¹⁰⁷ Inertial forces have been ignored since these can be neglected for small particles (between 10 nm and 10 μm).¹⁰⁶ When a particle is bound to a substrate *via* a cell or a molecular complex (*cf.* Fig. 1D and E), the rotational motion is restricted by the elastic properties of the biological matter. A general analytical expression is not available for the torsional elasticity of biological matter due to its complex structure. In case of small deformations of a biological macromolecule, the response can be approximated by a linear torsional spring, $\tau = -k\theta$, where typically $\tau < 1 \text{ pN } \mu\text{m}$ (*i.e.* 1 pN using a 1 μm arm).⁷⁰

Superparamagnetic micrometer-sized multicore MPs, like the ones invented by John Ugelstad *et al.*,⁹⁹ typically contain tens of thousands of non-remnant nanometer-sized cores fixed in a non-magnetic matrix. If these cores are randomly oriented, one would expect an isotropic magnetic susceptibility and thus it would be impossible to generate a magnetic torque. However experiments have shown that a torque can be exerted on a single multicore superparamagnetic parti-

cle.^{106,108} The torque of single superparamagnetic particles with a diameter of 2.8 μm has been calibrated in small fields (<5 mT on Dynal M-280 particles¹⁰⁶) as well as in larger fields (10–100 mT on Dynal M-270 particles¹⁰⁸). For the small fields, the magnetic torque was calibrated based on the known hydrodynamic drag of a smooth sphere, which revealed a linear dependence of the torque on the applied field. This linear dependence of the torque was attributed to a small remnant moment in the superparamagnetic particles.¹⁰⁶ For larger magnetic fields (10–100 mT), the angular thermal motion of surface-coupled particles was recorded to calibrate the magnetic torque. A quadratic scaling of the magnetic torque with field was found,¹⁰⁸ indicating that there is a net anisotropy in the MP,¹⁰³ or in other words, an easy-axis and a hard-axis.¹⁰⁴

A quadratic scaling of torque with field was also found for two-particle clusters of superparamagnetic particles. Clusters of particles have a non-spherical shape and thus a magnetic shape anisotropy. For such clusters a quadratic scaling of torque with field was observed, also for small fields,⁴¹ caused by the shape anisotropy of the clusters. The torque of clusters of particles depends only on the magnetic moment and diameter of the particles. In contrast, the magnetic torque of single particles depends on the size distribution and arrangement of nanocrystals within the particle, which can vary strongly between particles.^{109,110} The torque varies by 30–60% between particles of the same type and even more, up to a factor 100, between particles of different types.²⁷

Systems for magnetic torque actuation

In this section we discuss various systems to generate magnetic fields for rotational actuation of magnetic particles in lab-on-chip applications. Two main categories can be distinguished: permanent-magnet actuators and electromagnetic actuators. The geometry and type of magnetic actuator depends on the required field strength, field frequency, and sometimes the need for an additional field gradient. An instrument that can control the field direction and magnitude in two or three dimensions is typically referred to as magnetic torque tweezers (MTT),¹¹¹ a term that was introduced around the year 2000. A fundamental difference between magnetic tweezers and for instance optical- or mechanical tweezers, is that magnetic fields tend to spread out in space and therefore typically many particles are actuated at the same time. The torque that can be generated in a MTT system is comparatively large. Values for the torque generated in a MTT system using micrometer-sized MPs are between 1 and 100 pN μm .^{84,106,108} In optical torque tweezers the torque is between 0.1 and 1 pN μm ;^{84,112,113} higher torques could in principle be achieved using stronger optical fields, but limiting factors are the optical power, the generated heat inside the particles and heat that biological systems can withstand.

Fig. 3 sketches three main classes of MTT systems. Panels A and B sketch MTTs based on permanent magnets (pMTT)



and panel C a MTT system with electromagnets (eMTT). The system of panel A uses two parallel linear magnets with opposite magnetic orientations. Generally a space is left open between the two magnets to allow for optical imaging through the gap.¹¹⁴ The field gradient creates a large force oriented toward the magnet. The field strength and force are varied by changing the distance between magnets and substrate. A torque is applied by mechanically rotating the magnet. The pMTT system with circular magnet in panel B creates a field at the position of the particle with a field component perpendicular to the substrate. A torque is applied by adding a small linear magnet at one side of the circular magnet. The main advantage of the circular-magnet pMTT with respect to the linear-magnet pMTT is that the MP is pulled to the side, so that its rotation can be optically tracked by recording the *xy*-position of the particle rather than the angle of the particle itself.⁶⁴ This removes the need to add a label to the particle for angular tracking. Typical in-plane field strengths are between 5 and 500 mT (ref. 108) with an out-of-plane gradient between 10 and 1000 T m⁻¹.¹¹⁴ The magnitude of the field strength and the gradient are coupled by the design of the magnet. The maximum speed of field rotation is rather limited (typically \sim 0.5 Hz) because the magnet needs to be physically rotated. pMTTs^{65,115} are typically used for single-

molecule experiments on DNA (Fig. 1E) since these do not require high frequency actuation. Furthermore, in these experiments the coupling between the out-of-plane gradient and the in-plane field magnitude is not an issue as the DNA molecule typically needs to be stretched.

Electromagnetic MTT systems (eMTT) generate fields using current-carrying coils in combination with optional soft-magnetic materials. eMTTs typically have smaller fields than pMTTs, but eMTTs are much more flexible for modulating the magnitude-, orientation-, and frequency of the field. eMTT systems were already used three decades ago in magnetic twisting cytometry experiments (Fig. 1D). Uniaxial fields (AC fields) were used to apply mechanical stress on cell membranes.^{54,55} Later, eMTTs were also used to study single biological molecules.⁷⁰ Typically a single coil or a set of two coils is used to control the magnetic field amplitude along a single axis. By adding pairs of coils, field control is extended to two or three spatial dimensions. The systems are typically designed to have a homogenous, gradient free, magnetic field in the center of the system. Fig. 3C sketches an example with a single pair of coils. The coils can be equipped either with or without a soft-magnetic material inside the coil. Rotating fields are created by applying phase-shifted currents through orthogonal sets of coils in the system. No physical movement




Fig. 3 Sketches of three main classes of MTT systems, for actuation of particles in solution or particles coupled to a substrate (not to scale). North and south poles of permanent magnets are sketched in red and blue respectively. Electromagnets are sketched with coils. Dotted lines represent lines of the magnetic field. The field-induced magnetic moments (\vec{m}) of the particles and gradient-induced forces (\vec{F}) are indicated. Panel A sketches a pMTT system (p = permanent magnet) with two linear magnets.⁸⁰ Panel B sketches a cross-section of a pMTT system with a circular magnet.⁸⁰ Both pMTT systems generate a large gradient-induced force on the particle. Panel C sketches an eMTT system (e = electromagnet) with cores of soft magnetic material.^{116–119} The eMTT system has an in-plane field orientation and no field gradient in the middle of the magnetic system.

is needed to create field modulation or field rotation, therefore high frequencies can be achieved (>100 Hz without a soft magnetic material inside the coils). For creating high frequency fields, a limiting factor is the currents and heating induced by time varying fields, due to Faraday's law of induction and eddy currents.

For creating high amplitude fields the limiting factors are the Joule heating of the coils and/or the electric current that the power supply can deliver. The maximum field strengths for coils without a soft-magnetic material are in the order of a few milli Tesla at the sample location. If larger magnetic fields are required active cooling of the coils is needed. Another approach to enhance the field strength is to add a soft-magnetic material inside the coils.^{116,119} This enhances the field strength significantly (approximate 50 times) due to the high magnetic susceptibility of the soft-magnetic material. A disadvantage is that soft-magnetic materials have a remnant moment, typically 0.5–5% of the saturation magnetization, limiting the accuracy and the maximum frequency at which the field can be changed. Depending on the pole tip geometry, fields of 10 to 100 mT can be reached, with almost no field gradient. Table 1 summarizes the typical magnetic fields and rotation frequencies that can be achieved with the different magnetic systems. Typical values for the torque generated by an MTT and micrometer sized MPs are between 0 and 100 pN μ m.^{84,106,108}

Applications of rotating MPs

In this section we highlight early works and recent publications describing applications of rotating MPs as sketched in Fig. 1. We will discuss (1) the physical measurement principles, (2) the used detection methods (magnetic or optical), (3) the used types of particles (ferro- or superparamagnetic, size), (4) the used magnetic systems (eMTT or pMTT), and (5) we give examples of the physical systems that have been studied. An overview of this section is given in Table 2.

Mixing

Fluid mixing is important in lab-on-chip devices because a homogeneous composition of fluid ensures efficient reactions with low variability. However, mixing on a micrometer scale is difficult due to the low Reynolds numbers. Magnetic particles can be used to induce mixing in a lab-on-chip system (*cf.* Fig. 1A) by applying rotating gradient-free magnetic fields to particles in solution. Chains of particles are formed

due to dipole–dipole interactions between the MPs. These chains align themselves along the field lines due to the large shape anisotropy of the self-assembled chains.

When the magnetic field is rotated the chains act as rotors in the fluid. The average size of the rotors can be controlled by the magnitude of the field. Due to the absence of a gradient the rotors remain in the same position and induce local mixing. Studies have shown that rotational chain mixing can enhance the capturing rate of molecular targets by about a factor 4.^{10,13,15} An early publication describing the use of self-assembled chains for mixing was written by Vuppu *et al.* in 2003,⁶ based on superparamagnetic MPs in combination with a motorized permanent magnet.

One year later the same group published a paper showing that these rotating magnetic particle chains can be used to capture biological molecules.⁹ This principle has been extensively studied in the years thereafter¹¹ and was used in many bio-sensing applications^{35,37,38,42,43} (Fig. 1B and C) to significantly shorten assay times, typically from 60–120 min to 15–30 min. In later studies by Gao *et al.* an octopole magnetic tweezer was developed (Fig. 4A).¹⁵ The system can create 3D rotating magnetic fields with and without field gradient, which can for example be used to create particle chains, pull these to a substrate (Fig. 4B), and subsequently redisperse the chains to achieve distributed single particles¹⁴ (Fig. 4C). A more recent trend is to integrate magnetic micro-mixers on a chip, which gives compact systems for fluid mixing in very small volumes.⁷

Biomarker detection and concentration determination

An important application field of MPs in lab-on-chip applications is the detection and concentration determination of biomarkers for point-of-care medical diagnostic purposes. Typically the biomarkers to be measured are captured onto MPs and thereafter a secondary process is applied to detect the captured substance. Commonly chemiluminescent labels and reagents are used for detection² but a disadvantage of this method is that additional reagents and fluidic processing steps are needed.

To simplify the assay and allow direct detection in solution,³ measurement concepts are being studied based on particle rotation. One method is to measure a change of the hydrodynamic size of magnetic particles, as in Fig. 1B. The surface of the particles is functionalized with specific capture molecules, *e.g.* antibodies or aptamers that target the antigen

Table 1 Characteristics of magnetic torque tweezer systems as sketched in Fig. 3. pMTT refers to MTT with permanent magnet. eMTT refers to MTT with electromagnet. The eMTT can be built either with or without a soft magnetic material core inside the coil. The reported values are typical values rather than hard limitations. Note that for an eMTT system, a higher field typically implies a lower frequency

	Typical field (mT)	Typical frequency (Hz)	Field gradient
pMTT	100	1	Large, out of plane
eMTT with core	10	10	Small, in plane
eMTT without core	1	1000	Small, in plane



or bacteria of interest. There are two ways to detect a change of the hydrodynamic size of magnetic particles; the first is

via the magnetic response and the second by optical scattering.

Table 2 Overview of experimental studies with rotating magnetic particles (MPs). Applications are indicated (mixing, concentration detection, and biophysical characterization; see Fig. 1), the use of clusters of MPs or single MPs, the type of particle used (SP = superparamagnetic particle, FP = ferromagnetic particle), the tweezer system (eMTT = electromagnetic torque tweezers, pMTT = permanent magnetic torque tweezers), the detection method (magnetic or optical), the studied biological system, and the measurement principle. An example of an early publication and also more recent works are listed

Application	Cluster or single particle	Magnetic Particle (type & size range)	Magnetic tweezer	Detection method	Studied system	Measurement principle	Early publication	More recent works
Mixing (Fig. 1A)	Cluster	SP μm	eMTT	n.a.	n.a.	n.a.	2003 Vuppu ⁶	2009 Bruls ¹¹ actuation of clusters increases mixing at a surface 2014 Gao ¹⁵ actuation of clusters enhances mixing and capture in solution 2014 Van Reenen ² review paper
Concentration detection (Fig. 1B)	Single	FP nm	eMTT	Magnetic	Protein	Out-of-phase field component	2004 Astalan ¹⁶	2008 Fornara ²⁴ antibody concentration in undiluted serum
		SP/FP nm and μm	eMTT	Optical scattering or transmission	Cell/protein	Oscillation frequency or phase lag	2007 McNaughton ²²	2017 Schrittweiser ²¹ protein detection in diluted serum and saliva
Concentration detection (Fig. 1C)	Cluster	SP μm	eMTT	Optical scattering or transmission	Protein	Oscillation frequency or amplitude	2006 Baudry ⁸	2012 Ranzoni ⁴³ PSA in undiluted blood plasma 2016 Uddin ³⁸ Lab on a disc 2018 Van Vliembergen ¹²⁴ inter-particle distance
		FP nm	eMTT	Magnetic	Protein	In-phase component	2007 Hong ³¹	2010 Chieh ⁴⁰ VEGF in buffer, measured using a SQUID
Biophysical characterization (Fig. 1D)	Single	FP μm	pMTT and eMTT	Magnetic	Cell	Oscillation frequency	1987 Valberg ⁴⁷	2001 Puig-De-Morales ⁵⁷ frequency demodulation to measure mechanical cell properties
		SP/FP μm	eMTT	Optical microscopy	Cell	Spatial position and/or angular position	1993 Berg ⁴⁸	2008 del Alamo ⁵⁰ anisotropy in mechanical properties of cells 2012 Irmischer ⁵¹ torsional and translational mechanical properties of cell membranes
Biophysical characterization (Fig. 1E)	Single	SP μm	pMTT	Optical microscopy	DNA	Distance to surface or angular position	1996 Strick ⁶¹	2000 Strick ⁸³ DNA-protein interactions 2005 Koster ⁸⁶ enzyme-induced uncoiling of DNA 2012 Bryant ⁶⁵ review 2016 Berghuis ⁷⁹ high-throughput probing of DNA-protein interactions
		SP μm	eMTT	Optical microscopy	Protein	Angular position	2013 Van Reenen ⁷⁰	2016 Gutiérrez-Mejía ⁸¹ conformation changes in cardiac troponin






Fig. 4 Example of a magnetic system for mixing of fluids using MPs. (A) Octopole electromagnet developed by Gao *et al.*¹⁵ (B) Particle chains formed by a rotating field. (C) Redispersion of the particles, achieving single particles distributed over the substrate.^{120,121}

In 2003 two groups^{16,17} showed how the AC magnetic susceptibility (magnetization in a field with alternating amplitude) as a function of the driving field frequency can be used to measure the concentration of target proteins *via* the hydrodynamic size of the particles, see Fig. 5A. The theoretical concept of this measurement method was proposed by Connolly and St Pierre in 2001.¹²² Driving fields were used in the range between kHz to MHz, generated using eMTTs without magnetic cores. The field causes the MPs to align their magnetic moments with the field, by magnetic relaxation, called Néel relaxation, or by physical reorientation, called Brownian relaxation.¹⁰¹ Small (20–100 nm) ferromagnetic MPs were used, with $\tau_N \gg 1$ second. For these ferromagnetic MPs the dominant relaxation mechanism is Brownian, *i.e.* physical reorientation, which depends on the hydrodynamic volume. The method has been refined by optimizing the magnetic particles, see Fornara *et al.* in 2007.²⁴ They were able to measure a dose-response curve of a target antibody in undiluted serum.

In 2014 Dieckhoff *et al.* measured the phase lag of the magnetic moment using a rotating field instead of an AC field.²⁰ By tuning the driving frequency they were able to measure the concentration of antibodies in buffer using only one field frequency. They found a limit of detection on the order of 0.1 nM. Schrittweiser *et al.* proposed to measure the change in hydrodynamic size of hybrid plasmonic-ferromagnetic rods by optically measuring the phase lag between the

oscillations in the transmitted light and the rotating driving field, as shown in Fig. 5B.¹⁹ In a recent paper they reported experiments with non-plasmonic nanorods and were also able to measure this phase lag.²¹

An alternative rotation-based detection technique makes use of the scattering signal of optically anisotropic MPs in an AC- or rotating magnetic field. One of the first to use this optical detection technique was McNaughton *et al.* in 2007.²² They showed how the growth of a cell attached to a micrometer sized MP could be monitored *via* the scattered light. In these experiments a rotating field was used with a frequency such that the maximum magnetic torque was smaller than the hydrodynamic drag, *i.e.* the field rotates faster than the particle. This can be achieved by using small fields (1–5 mT) and relatively fast rotation speeds (10–100 Hz) in combination with micrometer sized superparamagnetic particles. This technique is referred to as asynchronous magnetic bead rotation (AMBR). If the hydrodynamic drag increases due to the growth of the cell, the average particle rotation frequency decreases.²² By tracking the rotation frequency over time the effect of anti-bacterial drugs could be tested by looking at the change in rotation frequency.^{25,28}

The abovementioned techniques involve a change of hydrodynamic radius. Another approach for measuring the concentration of biological molecules makes use of the analyte-induced formation of clusters of particles, as in Fig. 1C. In




Fig. 5 Methods to detect biological molecules *via* a change of hydrodynamic size of MPs, *cf.* the measurement principle shown in Fig. 1B. (A) Schematic representation of the experiment by A. Fornara *et al.*²⁴ Ferromagnetic nanoparticles were provided with specific capture molecules. The functionalized nanoparticles were mixed with a biological sample and the magnetic relaxation of the mixture was measured. (B) Schematic illustration of the measurement principle designed by Schrittweiser *et al.*¹⁹ Nanoprobe follow the rotating external magnetic field at a certain phase lag ϕ , which depends on their hydrodynamic volume and increases upon analyte binding. The angle ϕ follows from the measured transmission, which depends on the orientation of the nanoprobe's long axis with respect to the polarization direction of the incoming light (*i.e.*, the angle α).

these assays the number of clusters or the cluster size depends on the concentration of target. The clusters are detected *via* a change in magnetic response or optical scattering.

In 2006 two publications reported the detection of biological targets *via* the magnetic response of clusters of MPs. Hong *et al.*³⁰ made use of the change of AC magnetic susceptibility caused by the formation of clusters of particles. They applied a mixed frequency driving field (see Fig. 6A) and measured the AC magnetic susceptibility at a sum frequency to suppress the magnetic noise from the driving field. Using a SQUID to measure the magnetic signal instead of a pickup coil, Chieh *et al.* achieved a limit of detection in the pg ml^{-1} range.⁴⁰ A SQUID is more sensitive than coils, but requires cryogenic temperatures, which complicates system miniaturization.

In 2006 Baudry *et al.* showed the detection of a biological target by optically measuring the change of the rotational Brownian relaxation speed of clusters of MPs.⁸ A magnetic field was applied to form chains of MPs and the redispersion time of particles was measured upon removal of the magnetic field. The redispersion time is larger when target-induced clusters are present than when such biochemically-formed clusters are absent, due to the slower Brownian motion of clusters compared to single particles. The cluster assay was accelerated by the magnetic chain formation, to a time on the order of 5 min. The formation of chains can also be measured by recording the modulation of scattered light in a rotating magnetic field, as demonstrated around 2010 by Park *et al.*¹²³ and Donolato *et al.*^{36,38} who monitored the transmission of light, and by Ranzoni *et al.*^{41–43} who monitored the scattered light under an angle (see Fig. 6B). In these experiments spherical 100–1000 nm superparamagnetic MPs were used. Only clusters or chains of particles cause the scattered light to fluctuate in a rotating magnetic or AC field, because single particles possess no optical anisotropy. The magnitude of the oscillations of the scattered or transmitted light is proportional to the number of clusters, which in turn depends on the target concentration. The transmission-based measurement principle in combination with AC actuation along the optical axis, as demonstrated by Donolato *et al.*, led in

2013 to the creation of BluSense Diagnostics, a spin-out company from the Technical University of Denmark.

BluSense combines magnetic actuation with existing blu-ray technology, where the rotational motion of a disk-shaped cartridge gives microfluidic actuation and optical detection is used to measure cluster formation.³⁸ With a detection angle of the scattered light at 90 degrees, Ranzoni *et al.*⁴³ demonstrated that the concentration of PSA in undiluted blood plasma could be measured in the low picomolar range, by recording the amplitude of the $2f$ signal (at twice the driving frequency). Vliembergen *et al.*^{124,125} have extended the optical detection method by analyzing the rotation-angle dependent scattering signals. Sharp peaks in the spectrum allowed them to measure nanometer-scale inter-particle distances within the clusters. The results show that particle roughness is an important factor in the data. In the future this approach can lead to novel ways to distinguish between specific and non-specific binding in lab on chip cluster assays.

Biophysical characterization

The study of single proteins and single cells is of importance for understanding the basic properties of biological systems and for the development of bioanalytical tools. One subfield in this research area is force spectroscopy, where single molecules and cells are studied under different force conditions. The forces can be applied to the biological systems in a variety of ways, including the use of sharp tips (*e.g.* atomic force microscopy) and particles (*e.g.* optical tweezers and magnetic tweezers). A review comparing the different methods has been written by Neuman *et al.* in 2008.¹¹¹ Here we focus on the use of rotating MPs and gradient-free magnetic fields.

One of the first applications of rotating MPs was for studying the mechanical properties of cell membranes (Fig. 1D), by so-called magnetic twisting cytometry (MTC).⁴⁷ An experiment done by Valberg *et al.* in 1987 is shown in Fig. 7. Typically ferromagnetic MPs were used in combination with an AC field generated by an eMTT. The magnetic relaxation time of these ferromagnetic MPs is chosen to be significantly




Fig. 6 Magnetic and optical detection of clusters of MPs to detect concentrations of biological molecules in fluid. (A) Sketch of the setup used by Hong *et al.*^{30,31} The sample is placed in three concentric coils, where the two outer coils generate the driving field and the inner coil picks up the magnetic response of the sample. (B) Sketch of the setup used by Ranzoni *et al.*^{42,43} The sample is placed in a quadrupole electromagnet that generates a rotating magnetic field. The sample is illuminated using a laser and the scattered light from the particles is recorded using a photodiode.






Fig. 7 Study of mechanical properties of cells using rotating MPs. Shown are two lung macrophages, from experiments by Valberg *et al.* in 1987.⁴⁷ The dark spots are aggregates of magnetic nanoparticles. The direction of the applied field is indicated with the arrows. Small arrow heads indicate the reorientation of the magnetic aggregates.

longer than the mechanical relaxation times of the cells. The experiments start by attaching magnetic particles to the cell surface, for instance using antibodies. Next a strong magnetic pulse (typically 100 mT) is applied to magnetize all MPs in a selected direction. Thereafter a weak magnetic field (weak enough to not change the particle magnetization) is applied along another direction, which applies a magnetic torque to the cell membrane. Using a flux gate the field is measured giving information about the orientation of the magnetic particles.⁵⁶ By rotating the magnetic field with respect to the sample the storage ($G'(\omega)$) and loss modules ($G''(\omega)$) are resolved for different frequencies.^{57,59} All magnetic measurements are ensemble averages and do not provide information about cell-to-cell variations or spatial anisotropy in the cell mechanical response.

In later work, researchers switched from detecting the magnetic signal to optically tracking the spatial position and/or angular orientation of individual particles. Both ferromagnetic and superparamagnetic particles, typically larger than 1 μm , have been used in these experiments. An example of early work using optical detection was a study on the torque generated by a flagellar motor of *Escherichia coli* by Berg *et al.* in 1993.⁴⁸ They tracked the rotation speed of the particle and compared it for different magnetic field strengths.

Another example was the tracking of the spatial xy -position of a MP relative to a cell, giving the local deformation of the cell.⁶⁰ In another study the three-dimensional Euler angles as well as the spatial position of a superparamagnetic MP bound to cell surface were tracked, while applying a rotating field using an eMTT. This allowed to determine the storage and loss modules as a function of the spatial and angular orientations. Strong differences were observed in the storage modules between the different angular orientations.

Rotating MPs are also well-suited for studying the torsional properties of biological systems at the single-molecule level^{69,80} (Fig. 1E). Typically, a molecular system is captured between a planar substrate and a MP, and the rotation of the MP is recorded using optical microscopy. The first molecule to be studied was double-stranded DNA (dsDNA), where the

molecule was stretched by force as well as torsionally actuated by the MP.^{61,62,73} Stretching requires relatively high forces, so a pMTT system and micrometer-sized superparamagnetic particles were used. The torsional properties of the dsDNA were studied by measuring the height change of the MP (for instance *via* the optical diffraction pattern) as a function of the number of rotations of the particle.^{79,82,88} In these experiments the angular orientation of the MP was not tracked directly. Due to the large magnetic field and torque (10^{-17} Nm), it could be assumed that the number of rotations of the magnet corresponded to the number of rotations of the MP, and thus the number of windings of the dsDNA. The technique has allowed studies on the structure of over-wound and supercoiled dsDNA,⁷³ the effect of DNA-binding proteins,⁸⁹ enzymatic uncoiling of supercoiled dsDNA,⁸³ and the speed of enzymatic uncoiling compared to cleaving.⁸⁶

Later studies involved direct tracking of the angular orientation of the MP. Angular tracking can be achieved by attaching an optical marker that breaks rotational symmetry, *e.g.* a small non-magnetic particle. One solution is to attach a small nonmagnetic bead as marker to the DNA molecule itself. This labelling technique has allowed for studying the local structure, sequence specific effects and dynamical effects in supercoiled DNA.⁷¹ Another technique to track MP rotation is by using a circular magnet, as in Fig. 3B. The magnet creates a displacement of the center of the MP with respect to the anchoring point of the DNA in the fluid cell. As a consequence, rotation of the MP becomes related to the off-center spatial position of the MP. This technique has been used to monitor changes in the twist of nucleic acids at minimal torque.⁶⁴ If required the DNA can be torsionally actuated by adding a set of coils or a small linear magnet.^{63,66}

More recently the torque properties of proteins have been studied using MPs. The mechanical properties of proteins are very different from DNA: proteins are globular molecules (rather than filaments) with a size of around 10 nm. This gives a number of important differences: (i) there is no need to apply a stretching force when studying proteins, (ii) proteins are small and stiff, so angular and spatial displacements of the MP are small, and (iii) in the experiment the MP is close to the substrate, so artefacts due to particle-surface interactions need to be carefully ruled out. Janssen *et al.* developed an eMTT system with minimal field gradients.^{131,132} The system was used to study the torsional properties of proteins as a function of the angle of rotation⁷⁰ and as a function of the concentration of surfactants.⁷⁷ Fluorescent nanoparticles were attached to the MP for angular tracking, see Fig. 8. Panel C shows a microscopy image of an MP with a single fluorescent marker. By applying a continuously rotating magnetic field and recording the maximum angular deformation, the torsional rigidity of protein G-IgG and IgG-IgG complexes could be quantified.⁷⁰ This technique has also been used to study the torsional rigidity of a heterotrimeric protein complex, namely the cardiac troponin complex that regulates muscle contraction of the heart. Using antibody targeting of




Fig. 8 Study of single protein molecules using rotating MPs in an eMTT, by A. van Reenen *et al.*⁷⁰ (a) MPs are bound to a glass substrate by *via* a molecular complex. (b) Inside a fluid cell, the particles are actuated by a rotating magnetic field generated by a quadrupole electromagnet. (c) The particle orientation is visualized by means of attached small green fluorescent spheres. The bright spot in the center of the particle is caused by the excitation light.

specific protein domains, the calcium-induced conformational change of the protein complex could be recorded *via* a change of torsional rigidity.⁸¹ The work paves the way for nanomechanical mapping of native proteins and their conformational changes, using available libraries of monoclonal antibodies.

Conclusions and outlook

Nanometer and micrometer sized magnetic particles combined with gradient-free magnetic fields allow for the generation of well-defined mechanical torques and rotations in lab on a chip devices. Foundations of the field were laid in the previous century, when superparamagnetic behavior of magnetic particles was observed and monodisperse micrometer-sized superparamagnetic particles were developed. More recently, lab on chip functionalities have been studied with rotating magnetic particles, ranging from fluid mixing, detection of biomolecules, to biophysical studies of molecules and cells. These applications could be developed due to the large and well-controlled torques and rotation, the inherent spatial uniformity, and the bio-orthogonal nature of magnetic fields.

The mixing functionality (Fig. 1A) has been well-developed in terms of experiments as well as theory. Many basic studies have been performed and fluid-mechanical models exist to describe the development of flow by rotating particles. Biomarker quantification (Fig. 1B and C) is being studied using rotation of single particles as well as clusters of particles. These applications involve biofunctionalization of the particles, capture of target molecules by the particles, and readout *via* magnetic or optical signals. An advantageous feature is that the detection is performed in solution, without a fluidic washing process or other separation steps, which strongly simplifies assay integration into lab on chip devices. Further developments can be expected, particularly in the field of cluster assays (Fig. 1C), to achieve high analytical performance in a wide range of assays, by designing new particles, biofunctionalizations and assay implementations, with the prospect to enable affinity-based biosensing that is fast, reliable, integrated, cost-effective, and broadly applicable.

Another approach that has seen strong development is the biophysical probing of biomolecules and cells using rotating MPs (Fig. 1D and E). Methodologies were first developed for studying cells, but in recent years the focus has shifted completely to studying single molecules. Single-molecule studies (Fig. 1E) have given unique insights into the properties and function of macromolecules such as DNA, DNA–protein interactions, and also native protein complexes.

Single-molecule techniques are important for answering scientific questions but the methodologies also penetrate into bioanalytical sciences. Commercial analytical instruments are now available based on earlier academic studies on single-molecule methodologies for the quantitation of proteins^{126,127} and for DNA sequencing.^{128,129} The same trend can be expected for sensing based on magnetic rotation and torque, when the single-molecule methods will become suited to test unknown biological samples. As an example, the rotation-based methodologies based on target capturing by affinity molecules such as aptamers or antibodies⁸¹ will enable the detection and characterization of native proteins in biological samples. Due to the fact that unmodified native proteins are measured, the methods can have impact beyond research only. Proteins and their functional response can be characterized straight from industrial processes (for biotechnological process monitoring) and from patient samples (for diagnostics and patient monitoring). We foresee that this paves the way for biomarker response fingerprinting at the single-molecule level, with the prospect to lead to lab-on-chip systems with high specificity and high sensitivity, *e.g.* particle-based point-of-care testing devices² and devices for continuous biomarker monitoring.¹³⁰

In summary, the controlled application of torque and rotation to magnetic particles has developed strongly and steadily over the past decades, delivering a range of unique actuation and probing functionalities of fluids, molecules and cells in lab on chip systems. The field is set to further develop innovative approaches for characterizing biomolecular and biological systems as well as for integrating lab on chip functionalities into advanced bioanalytical tools.

Conflicts of interest

There are no conflicts to declare.

References

- M. A. M. Gijs, F. Lacharme and U. Lehmann, *Chem. Rev.*, 2010, **110**, 1518–1563.
- A. van Reenen, A. M. de Jong, J. M. J. den Toonder and M. W. J. Prins, *Lab Chip*, 2014, **14**, 1966–1986.
- S. Schrittawieser, B. Pelaz, W. Parak, S. Lentijo-Mozo, K. Soulantica, J. Dieckhoff, F. Ludwig, A. Guenther, A. Tschöpe and J. Schotter, *Sensors*, 2016, **16**, 828.
- A. Munaz, M. J. A. Shiddiky and N.-T. Nguyen, *Biomicrofluidics*, 2018, **12**, 031501.
- J. H. Dieckhoff, T. Yoshida, K. Enpuku, M. Schilling and F. Ludwig, *IEEE Trans. Magn.*, 2012, **48**, 3792–3795.
- A. K. Vuppu, A. A. Garcia and M. A. Hayes, *Langmuir*, 2003, **19**, 8646–8653.
- Q. Xiong, C. Y. Lim, J. Ren, J. Zhou, K. Pu, M. B. Chan-Park, H. Mao, Y. C. Lam and H. Duan, *Nat. Commun.*, 2018, **9**, 1743.
- J. Baudry, C. Rouzeau, C. Goubault, C. Robic, L. Cohen-Tannoudji, A. Koenig, E. Bertrand and J. Bibette, *Proc. Natl. Acad. Sci. U. S. A.*, 2006, **103**, 16076–16078.
- A. K. Vuppu, A. A. Garcia, M. A. Hayes, K. Booksh, P. E. Phelan, R. Calhoun and S. K. Saha, *J. Appl. Phys.*, 2004, **96**, 6831–6838.
- K. Tanaka and H. Imagawa, *Talanta*, 2005, **68**, 437–441.
- D. M. Bruls, T. H. Evers, J. A. H. Kahlman, P. J. W. van Lankveld, M. Ovsyanko, E. G. M. Pelssers, J. J. H. B. Schleipen, F. K. de Theije, C. A. Verschuren, T. van der Wijk, J. B. A. van Zon, W. U. Dittmer, A. H. J. Immink, J. H. Nieuwenhuis and M. W. J. Prins, *Lab Chip*, 2009, **9**, 3504.
- Y. Gao, M. Hulsen and J. M. J. den Toonder, in *14th International Conference on Miniaturized Systems for Chemistry and Life Science*, 2010, pp. 1055–1057.
- J.-T. Lee, L. Sudheendra and I. M. Kennedy, *Anal. Chem.*, 2012, **84**, 8317–8322.
- Y. Gao, A. van Reenen, M. A. Hulsen, A. M. de Jong, M. W. J. Prins and J. M. J. den Toonder, *Lab Chip*, 2013, **13**, 1394.
- Y. Gao, A. van Reenen, M. A. Hulsen, A. M. de Jong, M. W. J. Prins and J. M. J. den Toonder, *Microfluid. Nanofluid.*, 2014, **16**, 265–274.
- A. P. Astalan, F. Ahrentorp, C. Johansson, K. Larsson and A. Krozer, *Biosens. Bioelectron.*, 2004, **19**, 945–951.
- S. H. Chung, A. Hoffmann, S. D. Bader, C. Liu, B. Kay, L. Makowski and L. Chen, *Appl. Phys. Lett.*, 2004, **85**, 2971–2973.
- J. Dieckhoff, M. Schilling and F. Ludwig, *Appl. Phys. Lett.*, 2011, **99**, 112501.
- S. Schrittawieser, F. Ludwig, J. Dieckhoff, K. Soulantica, G. Viau, L.-M. Lacroix, S. M. Lentijo, R. Boubekri, J. Maynadié, A. Huetten, H. Brueckl and J. Schotter, *ACS Nano*, 2012, **6**, 791–801.
- J. Dieckhoff, A. Lak, M. Schilling and F. Ludwig, *J. Appl. Phys.*, 2014, **115**, 024701.
- S. Schrittawieser, B. Pelaz, W. J. Parak, S. Lentijo-Mozo, K. Soulantica, J. Dieckhoff, F. Ludwig and J. Schotter, *Sci. Rep.*, 2017, **7**, 4752.
- B. H. McNaughton, R. R. Agayan, R. Clarke, R. G. Smith and R. Kopelman, *Appl. Phys. Lett.*, 2007, **91**, 224105.
- K. Ērglis, Q. Wen, V. Ose, A. Zeltins, A. Sharipo, P. A. Janmey and A. Cēbers, *Biophys. J.*, 2007, **93**, 1402–1412.
- A. Fornara, P. Johansson, K. Petersson, S. Gustafsson, J. Qin, E. Olsson, D. Ilver, A. Krozer, M. Muhammed and C. Johansson, *Nano Lett.*, 2008, **8**, 3423–3428.
- B. H. McNaughton, P. Kinnunen, R. G. Smith, S. N. Pei, R. Torres-Isea, R. Kopelman and R. Clarke, *J. Magn. Magn. Mater.*, 2009, **321**, 1648–1652.
- Q. Wei, H.-M. Song, A. P. Leonov, J. A. Hale, D. Oh, Q. K. Ong, K. Ritchie and A. Wei, *J. Am. Chem. Soc.*, 2009, **131**, 9728–9734.
- P. Kinnunen, I. Sinn, B. H. McNaughton and R. Kopelman, *Appl. Phys. Lett.*, 2010, **97**, 223701.
- I. Sinn, P. Kinnunen, T. Albertson, B. H. McNaughton, D. W. Newton, M. A. Burns and R. Kopelman, *Lab Chip*, 2011, **11**, 2604.
- P. Kinnunen, I. Sinn, B. H. McNaughton, D. W. Newton, M. A. Burns and R. Kopelman, *Biosens. Bioelectron.*, 2011, **26**, 2751–2755.
- C.-Y. Hong, C. C. Wu, Y. C. Chiu, S. Y. Yang, H. E. Horng and H. C. Yang, *Appl. Phys. Lett.*, 2006, **88**, 212512.
- C.-Y. Hong, W. S. Chen, Z. F. Jian, S. Y. Yang, H. E. Horng, L. C. Yang and H. C. Yang, *Appl. Phys. Lett.*, 2007, **90**, 074105.
- A. Hecht, P. Commiskey, N. Shah and R. Kopelman, *Biosens. Bioelectron.*, 2013, **48**, 26–32.
- M. Donolato, P. Antunes, T. Z. G. de la Torre, E.-T. Hwu, C.-H. Chen, R. Burger, G. Rizzi, F. G. Bosco, M. Strømme, A. Boisen and M. F. Hansen, *Biosens. Bioelectron.*, 2015, **67**, 649–655.
- P. Antunes, D. Watterson, M. Parmvi, R. Burger, A. Boisen, P. Young, M. A. Cooper, M. F. Hansen, A. Ranzoni and M. Donolato, *Sci. Rep.*, 2015, **5**, 16145.
- D. Ramiandrisoa, E. Brient-Litzler, A. Daynes, E. Compain, J. Bibette and J. Baudry, *New Biotechnol.*, 2015, **32**, 467–472.
- J. Yang, M. Donolato, A. Pinto, F. G. Bosco, E.-T. Hwu, C.-H. Chen, T. S. Alström, G.-H. Lee, T. Schäfer, P. Vavassori, A. Boisen, Q. Lin and M. F. Hansen, *Biosens. Bioelectron.*, 2016, **75**, 396–403.
- B. Tian, J. Ma, T. Zardán Gómez de la Torre, Á. Bálint, M. Donolato, M. F. Hansen, P. Svedlindh and M. Strømberg, *ACS Sens.*, 2016, **1**, 1228–1234.
- R. Uddin, R. Burger, M. Donolato, J. Fock, M. Creagh, M. F. Hansen and A. Boisen, *Biosens. Bioelectron.*, 2016, **85**, 351–357.
- W. U. Dittmer, T. H. Evers, W. M. Hardeman, W. Huijnen, R. Kamps, P. de Kievit, J. H. M. Neijzen, J. H. Nieuwenhuis, M. J. J. Sijbers, D. W. C. Dekkers, M. H. Hefti and M. F. W. C. Martens, *Clin. Chim. Acta*, 2010, **411**, 868–873.



40 J. J. Chieh, S.-Y. Yang, H.-E. Horng, C. Y. Yu, C. L. Lee, H. L. Wu, C.-Y. Hong and H.-C. Yang, *J. Appl. Phys.*, 2010, **107**, 074903.

41 A. Ranzoni, X. J. A. Janssen, M. Ovsyanko, L. J. van IJzendoorn and M. W. J. Prins, *Lab Chip*, 2010, **10**, 179–188.

42 A. Ranzoni, J. J. H. B. Schleipen, L. J. van IJzendoorn and M. W. J. Prins, *Nano Lett.*, 2011, **11**, 2017–2022.

43 A. Ranzoni, G. Sabatte, L. J. van IJzendoorn and M. W. J. Prins, *ACS Nano*, 2012, **6**, 3134–3141.

44 P. Kinnunen, B. H. McNaughton, T. Albertson, I. Sinn, S. Mofakham, R. Elbez, D. W. Newton, A. Hunt and R. Kopelman, *Small*, 2012, **8**, 2477–2482.

45 C. H. Chang, Z. X. Lai, H. L. Lin, C. C. Yang, H. H. Chen, S. Y. Yang, H. E. Horng, C. Y. Hong, H. C. Yang and H. C. Lin, *Int. J. Nanomed.*, 2012, **7**, 4335–4340.

46 S. H. Liao, H. C. Yang, H. E. Horng, J. J. Chieh, K. L. Chen, H. H. Chen, J. Y. Chen, C. I. Liu, C. W. Liu and L. M. Wang, *Appl. Phys. Lett.*, 2013, **103**, 243703.

47 P. A. Valberg and H. A. Feldman, *Biophys. J.*, 1987, **52**, 551–561.

48 H. C. Berg and L. Turner, *Biophys. J.*, 1993, **65**, 2201–2216.

49 A. Cēbers and M. Ozols, *Phys. Rev. E: Stat., Nonlinear, Soft Matter Phys.*, 2006, **73**, 021505.

50 J. C. del Alamo, G. N. Norwich, Y.-s. J. Li, J. C. Lasheras and S. Chien, *Proc. Natl. Acad. Sci. U. S. A.*, 2008, **105**, 15411–15416.

51 M. Irmscher, A. M. de Jong, H. Kress and M. W. J. Prins, *Biophys. J.*, 2012, **102**, 698–708.

52 M. L. Rodriguez, P. J. McGarry and N. J. Sniadecki, *Appl. Mech. Rev.*, 2013, **65**, 060801.

53 N. Wang and D. E. Ingber, *Biophys. J.*, 1994, **66**, 2181–2189.

54 N. Wang, E. Planus, M. Pouchelet, J. J. Fredberg and G. Barlovatz-Meimon, *Am. J. Physiol.*, 1995, **268**, C1062–C1066.

55 N. Wang and D. E. Ingber, *Biochem. Cell Biol.*, 1995, **73**, 327–335.

56 B. Fabry, G. N. Maksym, R. D. Hubmayr, J. P. Butler and J. J. Fredberg, *J. Magn. Magn. Mater.*, 1999, **194**, 120–125.

57 M. Puig-De-Morales, M. Grabulosa, J. Alcaraz, J. Mullol, G. N. Maksym, J. J. Fredberg and D. Navajas, *J. Appl. Physiol.*, 2001, **91**, 1152–1159.

58 H. Huang, C. Y. Dong, H.-S. Kwon, J. D. Sutin, R. D. Kamm and P. T. C. So, *Biophys. J.*, 2002, **82**, 2211–2223.

59 V. M. Laurent, S. Hénon, E. Planus, R. Fodil, M. Balland, D. Isabey and F. Gallet, *J. Biomech. Eng.*, 2002, **124**, 408.

60 S. Hu, L. Eberhard, J. Chen, J. C. Love, J. P. Butler, J. J. Fredberg, G. M. Whitesides and N. Wang, *Am. J. Physiol.*, 2004, **287**, C1184–C1191.

61 T. R. Strick, J.-F. Allemand, D. Bensimon, A. Bensimon and V. Croquette, *Science*, 1996, **271**, 1835–1837.

62 J. D. Moroz and P. Nelson, *Proc. Natl. Acad. Sci. U. S. A.*, 1997, **94**, 14418–14422.

63 J. Lipfert, J. W. J. Kerssemakers, T. Jager and N. H. Dekker, *Nat. Methods*, 2010, **7**, 977–980.

64 J. Lipfert, M. Wiggin, J. W. J. Kerssemakers, F. Pedaci and N. H. Dekker, *Nat. Commun.*, 2011, **2**, 439.

65 Z. Bryant, F. C. Oberstrass and A. Basu, *Curr. Opin. Struct. Biol.*, 2012, **22**, 304–312.

66 X. J. A. Janssen, J. Lipfert, T. Jager, R. Daudey, J. Beekman and N. H. Dekker, *Nano Lett.*, 2012, **12**, 3634–3639.

67 F. C. Oberstrass, L. E. Fernandes and Z. Bryant, *Proc. Natl. Acad. Sci. U. S. A.*, 2012, **109**, 6106–6111.

68 C. Brème and F. Heslot, *PLoS One*, 2013, **8**, e55154.

69 S. Forth, M. Y. Sheinin, J. Inman and M. D. Wang, *Annu. Rev. Biophys.*, 2013, **42**, 583–604.

70 A. van Reenen, F. Gutiérrez-Mejía, L. J. van IJzendoorn and M. W. J. Prins, *Biophys. J.*, 2013, **104**, 1073–1080.

71 F. C. Oberstrass, L. E. Fernandes, P. Lebel and Z. Bryant, *Phys. Rev. Lett.*, 2013, **110**, 178103.

72 J. Lipfert, G. M. Skinner, J. M. Keegstra, T. Hensgens, T. Jager, D. Dulin, M. Kober, Z. Yu, S. P. Donkers, F.-C. Chou, R. Das and N. H. Dekker, *Proc. Natl. Acad. Sci. U. S. A.*, 2014, **111**, 15408–15413.

73 J. F. Allemand, D. Bensimon, R. Lavery and V. Croquette, *Proc. Natl. Acad. Sci. U. S. A.*, 1998, **95**, 14152–14157.

74 J. C. Cordova, D. K. Das, H. W. Manning and M. J. Lang, *Curr. Opin. Struct. Biol.*, 2014, **28**, 142–148.

75 E. J. Banigan and J. F. Marko, *Phys. Rev. E: Stat., Nonlinear, Soft Matter Phys.*, 2014, **89**, 062706.

76 R. S. Bejhed, T. Z. G. de la Torre, M. Donolato, M. F. Hansen, P. Svedlindh and M. Strömberg, *Biosens. Bioelectron.*, 2015, **66**, 405–411.

77 F. Gutiérrez-Mejía, L. J. van IJzendoorn and M. W. J. Prins, *New Biotechnol.*, 2015, **32**, 441–449.

78 D. Dulin, T. J. Cui, J. Cnossen, M. W. Docter, J. Lipfert and N. H. Dekker, *Biophys. J.*, 2015, **109**, 2113–2125.

79 B. A. Berghuis, M. Köber, T. van Laar and N. H. Dekker, *Methods*, 2016, **105**, 90–98.

80 J. Lipfert, M. Lee, O. Ordu, J. W. J. Kerssemakers and N. H. Dekker, *J. Visualized Exp.*, 2014, **19**, 1–18.

81 F. A. Gutiérrez-Mejía, *PhD thesis*, Eindhoven University of Technology, 2016.

82 F. Mosconi, J. F. Allemand, D. Bensimon and V. Croquette, *Phys. Rev. Lett.*, 2009, **102**, 078301.

83 T. R. Strick, V. Croquette and D. Bensimon, *Nature*, 2000, **404**, 901–904.

84 G. Romano, L. Sacconi, M. Capitanio and F. Pavone, *Opt. Commun.*, 2003, **215**, 323–331.

85 Y. Harada, *Sci. Technol. Adv. Mater.*, 2004, **5**, 709–713.

86 D. A. Koster, V. Croquette, C. Dekker, S. Shuman and N. H. Dekker, *Nature*, 2005, **434**, 671–674.

87 A. Celedon, I. M. Nodelman, B. Wildt, R. Dewan, P. Searson, D. Wirtz, G. D. Bowman and S. X. Sun, *Nano Lett.*, 2009, **9**, 1720–1725.

88 D. Klaue and R. Seidel, *Phys. Rev. Lett.*, 2009, **102**, 028302.

89 J. Lipfert, S. Klijnghout and N. H. Dekker, *Nucleic Acids Res.*, 2010, **38**, 7122–7132.

90 R. Řezníček, V. Chlan, H. Štěpánková, P. Novák and M. Maryško, *J. Phys.: Condens. Matter*, 2012, **24**, 055501.

91 J. Carrey, B. Mehdaoui and M. Respaud, *J. Appl. Phys.*, 2011, **109**, 083921.

92 A. Aharoni, *Phys. Rev. B: Solid State*, 1973, **7**, 1103–1107.

93 S. Mørup, M. F. Hansen and C. Frandsen, *Beilstein J. Nanotechnol.*, 2010, **1**, 182–190.





94 Q. A. Pankhurst, J. Connolly, S. K. Jones and J. Dobson, *J. Phys. D: Appl. Phys.*, 2003, **36**, R167–R181.

95 M. A. Willard, L. K. Kurihara, E. E. Carpenter, S. Calvin and V. G. Harris, *Int. Mater. Rev.*, 2004, **49**, 125–170.

96 W. Wu, Z. Wu, T. Yu, C. Jiang and W.-S. Kim, *Sci. Technol. Adv. Mater.*, 2015, **16**, 023501.

97 S. Bogren, A. Fornara, F. Ludwig, M. del Puerto Morales, U. Steinhoff, M. Hansen, O. Kazakova and C. Johansson, *Int. J. Mol. Sci.*, 2015, **16**, 20308–20325.

98 J. Wells, O. Kazakova, O. Posth, U. Steinhoff, S. Petronis, L. K. Bogart, P. Southern, Q. Pankhurst and C. Johansson, *J. Phys. D: Appl. Phys.*, 2017, **50**, 383003.

99 J. Ugelstad and F. K. Hansen, *Rubber Chem. Technol.*, 1976, **49**, 536–609.

100 G. Fonnum, C. Johansson, A. Molteberg, S. Mørup and E. Aksnes, *J. Magn. Magn. Mater.*, 2005, **293**, 41–47.

101 J. Dieckhoff, D. Eberbeck, M. Schilling and F. Ludwig, *J. Appl. Phys.*, 2016, **119**, 043903.

102 C. P. Bean and J. D. Livingston, *J. Appl. Phys.*, 1959, **30**, S120–S129.

103 J. D. Livingston and C. P. Bean, *J. Appl. Phys.*, 1959, **30**, S318–S319.

104 K. Kurosawa, M. Miura and S. Saito, *J. Phys. C: Solid State Phys.*, 1980, **13**, 1521–1527.

105 D. Cohen and I. Nemoto, *IEEE Trans. Biomed. Eng.*, 1984, **31**, 261–273.

106 X. J. A. Janssen, A. J. Schellekens, K. van Ommering, L. J. van IJzendoorn and M. W. J. Prins, *Biosens. Bioelectron.*, 2009, **24**, 1937–1941.

107 R. I. Ovseenko and Y. G. Ovseenko, *Fluid Dyn.*, 1971, **3**, 78–82.

108 M. M. van Oene, L. E. Dickinson, F. Pedaci, M. Köber, D. Dulin, J. Lipfert and N. H. Dekker, *Phys. Rev. Lett.*, 2015, **114**, 218301.

109 D. R. Baselt, G. U. Lee, M. Natesan, S. W. Metzger, P. E. Sheehan and R. J. Colton, *Biosens. Bioelectron.*, 1998, **13**, 731–739.

110 J. Xu, K. Mahajan, W. Xue, J. O. Winter, M. Zborowski and J. J. Chalmers, *J. Magn. Magn. Mater.*, 2012, **324**, 4189–4199.

111 K. C. Neuman and A. Nagy, *Nat. Methods*, 2008, **5**, 491–505.

112 A. La Porta and M. D. Wang, *Phys. Rev. Lett.*, 2004, **92**, 190801.

113 S. Bayoudh, T. A. Nieminen, N. R. Heckenberg and H. Rubinsztein-dunlop, *J. Mod. Opt.*, 2003, **50**, 1581–1590.

114 Z. Yu, D. Dulin, J. Cnossen, M. Köber, M. M. van Oene, O. Ordu, B. A. Berghuis, T. Hensgens, J. Lipfert and N. H. Dekker, *Rev. Sci. Instrum.*, 2014, **85**, 123114.

115 D. Kiline and G. U. Lee, *Integr. Biol.*, 2014, **6**, 27–34.

116 C. Jiang, T. A. Lionberger, D. M. Wiener and E. Meyhofer, *Rev. Sci. Instrum.*, 2016, **87**, 084304.

117 F. Mosconi, J. F. Allemand and V. Croquette, *Rev. Sci. Instrum.*, 2011, **82**, 034302.

118 L. Sacconi, G. Romano, R. Ballerini, M. Capitanio, M. De Pas, M. Giuntini, D. Dunlap, L. Finzi and F. S. Pavone, *Opt. Lett.*, 2001, **26**, 1359.

119 S. Schuerle, S. Erni, M. Flink, B. E. Kratochvil and B. J. Nelson, *IEEE Trans. Magn.*, 2013, **49**, 321–330.

120 A. van Reenen, A. M. de Jong and M. W. J. Prins, *Lab Chip*, 2015, **15**, 2864–2871.

121 A. van Reenen, *PhD thesis*, Eindhoven University of Technology, 2014.

122 J. Connolly and T. G. St Pierre, *J. Magn. Magn. Mater.*, 2001, **225**, 156–160.

123 S. Y. Park, H. Handa and A. Sandhu, *Nano Lett.*, 2010, **10**, 446–451.

124 R. W. L. van Vliembergen, L. J. van IJzendoorn and M. W. J. Prins, *Langmuir*, 2018, **34**, 179–186.

125 R. W. L. van Vliembergen, L. J. van IJzendoorn and M. W. J. Prins, *Opt. Express*, 2016, **24**, A123.

126 J. Todd, B. Freese, A. Lu, D. Held, J. Morey, R. Livingston and P. Goix, *Clin. Chem.*, 2007, **53**, 1990–1995.

127 D. M. Rissin, C. W. Kan, T. G. Campbell, S. C. Howes, D. R. Fournier, L. Song, T. Piech, P. P. Patel, L. Chang, A. J. Rivnak, E. P. Ferrell, J. D. Randall, G. K. Provuncher, D. R. Walt and D. C. Duffy, *Nat. Biotechnol.*, 2010, **28**, 595–599.

128 D. Branton, D. W. Deamer, A. Marziali, H. Bayley, S. A. Benner, T. Butler, M. Di Ventra, S. Garaj, A. Hibbs, X. Huang, S. B. Jovanovich, P. S. Krstic, S. Lindsay, X. S. Ling, C. H. Mastrangelo, A. Meller, J. S. Oliver, Y. V. Pershin, J. M. Ramsey, R. Riehn, G. V. Soni, V. Tabard-Cossa, M. Wanunu, M. Wiggin and J. A. Schloss, *Nat. Biotechnol.*, 2008, **26**, 1146–1153.

129 J. Eid, A. Fehr, J. Gray, K. Luong, J. Lyle, G. Otto, P. Peluso, D. Rank, P. Baybayan, B. Bettman, A. Bibillo, K. Bjornson, B. Chaudhuri, F. Christians, R. Cicero, S. Clark, R. Dalal, A. DeWinter, J. Dixon, M. Foquet, A. Gaertner, P. Hardenbol, C. Heiner, K. Hester, D. Holden, G. Kearns, X. Kong, R. Kuse, Y. Lacroix, S. Lin, P. Lundquist, C. Ma, P. Marks, M. Maxham, D. Murphy, I. Park, T. Pham, M. Phillips, J. Roy, R. Sebra, G. Shen, J. Sorenson, A. Tomaney, K. Travers, M. Trulson, J. Vieceli, J. Wegener, D. Wu, A. Yang, D. Zaccarin, P. Zhao, F. Zhong, J. Korlach and S. Turner, *Science*, 2009, **323**, 133–138.

130 E. W. A. Visser, J. Yan, L. J. van IJzendoorn and M. W. J. Prins, *Nat. Commun.*, 2018, **9**, 2541.

131 X. J. A. Janssen, A. van Reenen, L. J. van IJzendoorn, A. M. de Jong and M. W. J. Prins, *Colloids Surf. A*, 2011, **373**, 88–93.

132 X. J. A. Janssen, J. M. van Noorloos, A. Jacob, L. J. van IJzendoorn, A. M. de Jong and M. W. J. Prins, *Biophys. J.*, 2011, **100**, 2262–2267.