RSC Advances



PAPER

View Article Online
View Journal | View Issue



Cite this: RSC Adv., 2019, 9, 41970

Sheath-less high throughput inertial separation of small microparticles in spiral microchannels with trapezoidal cross-section†

Ala'aldeen Al-Halhouli, *\omega* \text{abc} Ahmed Albagdady *\omega* and Andreas Dietzel *\omega* b

Various mechanisms of different designs have emerged for the purpose of microparticle separation and cell sorting. The main goals behind such designs are to create high throughput and high purity sample isolation. In this study, high efficiency, high throughput and precise separation of microparticles under inertial lift and drag forces induced by trapezoidal curvilinear channels are reported. This work is the first to focus and recover 2 from 5 μ m and 2 from 10 μ m particles in spiral channels in a sheath-less flow device, which reduces the overall complexity of the system and allows for higher throughput. The new microfluidic chip design is fabricated in glass using femtosecond laser ablation. In addition, mathematical force calculations were conducted during the design phase of the microfluidic channels and compared with experiments. The results show a close prediction of the equilibrium position of the tested microparticles.

Received 30th July 2019 Accepted 28th November 2019

DOI: 10.1039/c9ra05916d

rsc.li/rsc-advances

Introduction

Size-based microparticle separation has drawn a lot of interest due to its applications in biomedical and biological research.1-6 Conventional separation techniques such as centrifugation, filtration, and sedimentation are associated with wearing and clogging, which affect their efficiency and reliability.7 Processing biological samples through microfluidic devices like focusing,8 on-chip labelling,9 separation,10,11 mixing12,13 and diagnostics10 is an area of high relevance to the life science community. Several methods have been developed to address this matter, which can be classified as active or passive.14 The active methods rely on external forces to manipulate the microparticle's position, such as dielectrophoretic (DEP)15 and magnetic forces.16 On the contrary, passive methods do not require any kind of external forces and separate microparticles based on physical properties such as channel geometry, particle size, and the characteristics of the fluid medium. Pinched flow fractionation, 17 deterministic lateral displacement¹⁸ and inertial migration¹⁹ are well-known passive techniques, they depend on inertial forces inside the fluid that can typically be neglected in microfluidics when operating at low Reynolds numbers in the straight microfluidic Stokes regime.

focus in different positions to be selectively isolated through

microparticles are driven away from the walls due to the

asymmetric wake that forms when a particle is close to the

wall,20 while the parabolic velocity profile in the Poiseuille flow induces an opposing shear gradient lift force by which the particles migrate towards the channel walls. To achieve high purity microparticle separation, the minimum number of force equilibrium positions for each particle size has to be obtained. In square microchannels (h/w = 1), four streamlines of particles form close to the channel's center where the wall and shear gradient lift forces are balanced, whereas in rectangular channels with (h/w < 1), two symmetric focusing positions result around the channel's center. Complex secondary flows are induced by using curved channel arrangements as in spiral microfluidics,21 or by forcing the fluid to follow a curved path as in contraction-expansion arrays (CEAs).22 In either method, two counter-rotating vortices (Dean vortices) develop at the upper and lower halves of the channel's cross-section due to the shift of the maximum velocity point towards the outer wall,23 which enhances particle focusing and adds an additional force that induces particle migration in a lateral direction across the channel's width.24 Spiral microchannels consist of a multi-turn loop; the side walls of the channels are referred to as the inner and outer walls, depending on which is closer to the spiral center. Given certain channel dimensions, larger particles experience higher lift forces and, therefore, have the tendency to migrate towards the inner wall. Moreover, drag forces are dominant on smaller particles, which re-circulate them to eventually settle in their equilibrium position²⁵ where its location depends on the Reynolds number and the ratio of inertial lift forces to Dean drag force.26 Microparticle separation requires particles to

[&]quot;NanoLab, School of Applied Technical Sciences, German Jordanian University (GJU), Amman, Jordan. E-mail: alaaldeen.alhalhouli@gju.edu.jo

^bInstitut für Mikrotechnik, Technische Universität Braunschweig, Braunschweig, Germany

^{&#}x27;Faculty of Engineering, Middle East University, Amman 11831, Jordan

[†] Electronic supplementary information (ESI) available. See DOI: 10.1039/c9ra05916d

microfluidics community.

Paper

separate outlets; however, in a Dean flow fractionation device with a rectangular cross-section, the distance between the particles' focusing position is minimal because the centers of the Dean vortices are located at the center of the channel's cross-section. Therefore, spiral channels with trapezoidal cross-sections27,28 are used to shift the centers of vortices towards the deepest wall, creating a wider distance between focused particles of different sizes. Recently, multiple inertial mechanisms besides spiral microfluidics have been developed to separate smaller microparticles. Wang et al.29 demonstrated an asymmetric serpentine microfluidic channel of 20 μ m \times 10 μm (width \times height) to focus 2 μm particles and 10 $\mu m \times 5$ μm for 0.92 μm. Since the channels are small, a high-pressure syringe pump was needed to allow for high flow rates up to 1400 µL min⁻¹ to increase the channel's throughput. Also, Cruz et al.30 presented a method to focus 0.5 and 0.7 μm particles and separate 1 and 2 µm particles with two curves connected with a small turn; the channel was 20 µm in width and 10 μ m in depth to match F_D and F_L within the range of available flow rates provided by the pump. Cruz et al.30 turned to a high-performance liquid chromatography (HPLC) pump to achieve high pressure values, which is uncommon to the

Several attempts have been made to focus and separate 2 μm in spiral channels, however, most designs used a sheath fluid and worked at low flow rates due to the high pressure induced in the small channel geometry as reported by Bhagat $\it et~al.$ and Lee $\it et~al.$ 31,32 In a previous work, 33 2 μm particles were focused by implementing microchambers along the last turn of a spiral channel; that design produced thin focusing streams but could not separate microparticles since all sizes focused near the inner wall. In this paper, we study the trapezoidal microfluidic channel's ability to separate small microparticles down to 2 μm in high throughput without utilizing a high-pressure pump and without adding a sheath fluid that complicates the testing setup and might limit its use in low setup areas, achieving a separation between 2 and 5 μm , and 2 and 10 μm fluorescent microparticles.

Implementing a sheath-less inertial focusing device reduces the setup complexity and most importantly, the number of expensive syringe pumps,³⁴ because a pump would be needed for every buffer inlet. The total flow rate (sample flow rate + the buffer solution flow rate) that is often used to describe the focusing flow rate, can be higher than its sheath-less counterpart; however, unlike sheath-less focusing, the output fluid of a sheath-based device often has similar or lower concentration in comparison with the sample fluid, due to the addition of a buffer solution that dilutes the sample inside the focusing channels. Therefore, sheath-based focusing would not be the best choice for use in a particle concentration device.

Materials and methods

Microfluidic chip design and fabrication

Trapezoidal microfluidic chips with different dimensions were fabricated and tested for the purpose of high-efficiency separation. The microchannels consisted of an 8-turn spiral with radius increasing from 4.9 to 11.9 mm and an asymmetric branch at the end of the last turn to isolate the inner and outer wall fluids. The width of the channel cross-section was 220 μ m and the heights were optimized after a few iterations to be 60 and 40 μ m at the outer and inner wall, respectively.

Mechanism of operation

Microparticles experience inertial forces when flowing inside a microfluidic channel, which affect their lateral and axial positions. In a straight channel, only the wall-induced lift force (F_{LW}) and shear-gradient lift force (F_{LS}) are large enough to induce particle migration. The resultant of these forces is often referred to as the inertial lift force and can be represented as follows:³⁵

$$F_{\rm L} = \frac{2\rho U_{\rm f}^2 a_{\rm c}^4}{D_{\rm h}} \tag{1}$$

where $a_{\rm c}$ is the particle diameter, $U_{\rm f}$ is the average fluid velocity, ρ is the fluid density and $D_{\rm h}$ is the channel's hydraulic diameter. Particle size is the dominant factor; therefore, aiming to focus on small particles is a challenge when counting only on inertial forces. On the contrary, the Dean drag force resulting from a high pressure gradient generated by the sharp velocity profile^{36,37} would be larger than the inertial force when calculated for smaller particles. The strength of the Dean vortices can be represented by the dimensionless Dean number $D_{\rm e}$ as follows:^{23,38}

$$D_{\rm e} = {\rm Re}\sqrt{\frac{D_{\rm h}}{2R}} \tag{2}$$

where R is the radius of curvature and Re is the Reynolds number. The velocity that the particles experience during lateral migration is called the Dean velocity and can be calculated as $U_{\rm Dean} = 1.8 \times 10^{-4} \, D_{\rm e}^{1.63}$. The resulting Dean force exerted on particles passing in a curved channel can be calculated using Stokes' Law:

$$F_{\rm D} = 3\pi\mu U_{\rm Dean} a_{\rm c} \tag{3}$$

where μ is the dynamic viscosity of the fluid. The Dean drag force drives the particles in a direction opposite to the inertial lift forces and since it is less sensitive (linear dependence) to particle diameter, smaller particles are more affected and would be focused near the center of the Dean vortices (outer wall). Fig. 1(A) shows the used trapezoidal microfluidic chip with one inlet and two outlets, while Fig. 1(B) illustrates the direction of inertial and drag forces acting on particles inside the rectangular and trapezoidal microchannels as well as the skewed velocity profile due to the channel's curvature. The ratio of inertial lift and drag forces $R_f = F_L/F_D$. can be calculated with the help of eqn (1) and (3). Small values of R_f indicate the dominance of the Dean drag forces over inertial forces and particles would more likely be focused near the outer wall. On the other hand, as $R_{\rm f}$ increases, inertial forces become more dominant and particles would be focused closer to the inner wall.

Fig. 1 Trapezoidal microfluidic chip representation. (A) Fabricated spiral microfluidic chip with one inlet and two outlets, the microchannels are filled with red dye for better contrast. (B) Illustration of the spiral channel principle of operation showing a comparison between Dean vortices in rectangular and trapezoidal channel cross-sections.

Chip fabrication

A laser workstation (microSTRUCT-C, 3D-Micromac, Chemnitz, Germany) equipped with a Yb:KGW solid-state femtosecond laser (Pharos, Light Conversion, Vilnius, Lithuania) was used to engrave the microchannels on a glass wafer. The machine's software offered two options for importing the drawings one method was to convert DXF files created on AutoCAD (AutoCAD 2018, Autodesk, San Rafael, CA, USA) into a machine-readable graphics file with certain filling strategies that define the lines through which the laser will pass when ablation is initiated. This method is versatile and simple but it produces large files leading to long ablation times. An alternative method is given by writing a Visual Basic for Applications (VBA) script that draws the required laser path (see Fig. 2(B)), where tighter lines produce a larger depth. Both methods were used to fabricate the spiral chip; the VBA script, described in previous work, ²¹ to draw




Fig. 2 The fabricated trapezoidal channel investigated by 3D microscopy. (A) Three-dimensional reconstruction of the trapezoidal microchannel showing the inner and outer walls. (B) Top view of the branch at the end of the spiral shape where the red lines indicate the paths generated by the VB script to produce trapezoidal channels. (C) Channel cross-section profile as referenced in (B), showing the depth at each wall.




Fig. 3 Fluorescent green light intensity line profile (across the outlet channel) for 5 μ m particles with 0.0125% (w/v) and 0.0025% (w/v) suspension concentration at 0.7 mL min⁻¹. Higher concentration values may increase the number of equilibrium positions.

the spiral shape with multi-pitch strategy, and the filling strategy for the branch and holes. After laser ablation, the engraved glass wafer (BOROFLOAT®, Schott AG, Mainz, Germany) was dipped in glass etching solution (phosphoric acid, hydrofluoric acid and water, 20:6:9)39,40 for 90 s then rinsed with DI water to smoothen the ablated surface and remove any residual glass particles resulting from the engraving process. A wafer-cleaning machine (Fairchild Convac, Neuenstadt, Germany) was used to clean the wafer before it was bonded to a blank wafer by placing it in a muffle furnace at 620 °C for 6 hours. During the laser process development, a 3D microscope (Keyence VHX 5000, Japan-imported through Keyence Germany) was used to evaluate the ablated surface's shape as shown in Fig. 2(A), and to accurately measure the dimensions to finally produce a 220, 60 μ m \times 40 μ m (w, $h_1 \times h_2$) trapezoidal channel (see Fig. 2(C)).

Particle solution preparation

Fluorescent polystyrene microparticles of 2 μm and 10 μm (exc./ emm.: 468/508 nm, 1% solids concentration, Fluoro-Max, Thermo Fisher Scientific, Massachusetts, USA) and 5 μm (exc./ emm.: 502/518 nm, 2.5% solids concentration, microParticles GmbH, Berlin, Germany) were used in this work. For the 2 μm and 10 μm particles, 50 μL of particles solution were diluted in 10 mL DI water resulting in a 0.005% (w/v) suspension, while 10 μL of particles solution were diluted in 10 mL DI water for 5 μm particles, resulting in a 0.0025% (w/v) suspension concentration, which will be further discussed in a subsequent section.

Results and discussion

A set of experiments was designed to test the fabricated microfluidic chip with 2, 5, 10 μ m fluorescent microparticles. The particle solution was filled in a glass syringe (SETonic GmbH, Ilmenau, Germany) and loaded into the syringe pump (NEMESYS low pressure module 14:1, Cetoni GmbH,

Table 1 Dean and lift forces calculations for different flow rates and particle sizes

Particle diameter (μm)	Flow rate (mL min ⁻¹)	Dean number $D_{\rm e}$	$F_{ m L}$ (pN)	$F_{\mathrm{D}}\left(\mathrm{pN}\right)$	$R_{ m f}$
2	0.1	1.1094	0.11102	4.0265	0.0276
	0.5	5.5469	2.7755	55.494	0.0500
	1	11.0939	11.102	171.76	0.0646
5	0.1	1.1094	4.3367	10.066	0.4308
	0.5	5.5469	108.42	138.74	0.7815
	1	11.0939	433.67	429.40	1.0099
10	0.1	1.1094	69.388	20.132	3.4466
	0.5	5.5469	1734.7	277.47	6.2518
	1	11.0939	6938.8	858.81	8.0795

Korbussen, Germany). Live monitoring of the experiments was performed with a microscope 0065 camera (XC30) fixed on an inverted microscope (CKX53, both Olympus, Shinjuku, Tokyo, Japan) with blue light excitation (460-495 nm) from a mercury short-arc lamp (USH-103D, Ushio, Marunouchi, Chiyoda-ku, Tokyo) for particle fluorescence imaging. Force calculations on various chip designs were conducted during the design phase. The ratio $R_{\rm f}$ of inertial lift to drag forces has to be $R_{\rm f} \ll 1$ for the particles to focus near the outer wall, indicating the dominance of the drag forces. On the contrary, larger values of $R_{\rm f} \gg 1$ indicate the dominance of inertial forces. Table 1 shows detailed force calculations for 2, 5 and 10 µm particles for the final design at various flow rates. The lift forces exceeded the Dean forces for the 10 µm particles, which were, therefore, expected to focus near the inner wall. The Dean forces were greater than the lift forces for 2 µm particles, which were expected to focus near the outer wall. The magnitude of forces is

also of great importance because even if the ratio between forces is significantly smaller or larger than 1, particles will not focus in a thin line unless the channel can provide enough force to migrate them. In the case of the 5 μ m particles, it can be seen in Table 1 that the ratio $R_{\rm f}$ is very close to unity; this indicates that particles will experience two equilibrium positions on both ends of the channel, which will be further discussed in a subsequent section.

The effect of particle concentration

When designing a microfluidic chip for microparticle/cell separation, one of the aims is to maximize throughput, which directly relates to the suspension concentration. Some interactions and forces are typically neglected in mathematical models for particle focusing such as particle–particle interaction, which is often avoided in experiments by choosing a low particle fraction.⁴¹ Higher particle concentration affects the focusing




Fig. 4 The 2, 5 and 10 μ m fluorescent microparticle trajectories at flow rates from 0.1–1.0 mL min⁻¹. The dashed lines represent the channel's outline.

RSC Advances

Table 2 A review of focusing/separation of 2 um particles from the literature in spiral microchannels

Group	Turns	Channel geometry	Mechanism	$Width/height(\mu m)$	Particle sizes (μm)	Total flow rate
Bhagat et al. 2008 ³¹	5	Rectangular	Spiral, sheath flow	100/50	7.32, 1.9	20 μL min ⁻¹ 1 : 1 sample to sheath ratio
Lee et al. 2019 ³²	1.5	Rectangular	Spiral, sheath flow	50/20	0.81, 2.29, 4.7	78 μL min ⁻¹ 1 : 5
Al-Halhouli <i>et al.</i> 2019 ³³	8	Rectangular	Spiral, sheath-less, microchambers	200/50	2, 5, 10 focusing only	sample to sheath ratio $350 \ \mu L \ min^{-1}$
Current work	8	Trapezoidal	Spiral, sheath-less	220/60, 40	2, 5, 10	600 μL min ⁻¹ (2, 5 μm), 1000 μL min ⁻¹ (2, 10 μm)

efficiency and, in some cases, results in multiple focusing positions for the same particle size. 42 Fig. 3 shows two fluorescence intensity plots for the 5 µm particle trajectory at 0.7 mL min⁻¹. Once the flow rate has stabilized in the channel and reached a steady-state, another secondary focusing position appears at the outer wall, which limits the chip's operation.

However, reducing the particle suspension concentration from 0.0125% to 0.0025% (w/v) was able to solve this problem within the tested range of operation. A critical flow rate at which the equilibrium positions of particles split into two (0.7 mL min⁻¹ in the case of 0.0125% (w/v)) might just shift to higher flow rate values. It is noticeable in Fig. 3 that the concentration reduction also shifted the equilibrium position towards the inner wall. Normally, a small concentration variation does not notably affect the particle equilibrium position; however, in this case, the particles were not fully stable in the first place because the lift and Dean forces were almost identical at the tested flow rates. Flow rates below 0.7 mL min⁻¹ returned one equilibrium position even for 0.0125% (w/v) concentration, but when the 2 µm particles were added to test the separation, a second equilibrium position started to emerge at lower flow rates.

Microparticle separation

Before injecting the chip with the particle solution, the chip was primed with DI water for a few minutes to remove air bubbles and any residual particles. Solutions with 2, 5 and 10 µm particles were pumped individually at flow rates varying from $0.1-1 \text{ mL min}^{-1}$.

The 2 µm particles were recovered from the outer wall at all tested flow rates, while the 5 µm (0.0025% (w/v) suspension) particles were focused near the inner wall at flow rates 0.5-1.0 mL min $^{-1}$. The 10 μ m particles also focused on the inner wall for flow rates 1.0-1.5 mL min⁻¹. Below 0.3 mL min⁻¹, the focusing position was not stable for all tested particles due to the low magnitude of inertial forces. The experiments as shown in Fig. 4 agree with the calculations as summarized in Table 2, thereby confirming that the 2 µm particles experienced higher Dean forces at all flow rates and consequently focused at the outer wall, whereas the 5 µm (after suspension concentration optimization) and 10 µm particles experienced higher lift forces and focused at the inner wall. To verify the separation capability, mixtures of 2, 5 µm and 2, 10 µm particles were prepared at the same concentrations tested before with monodisperse suspension (as shown in Fig. 4). The acquired results confirmed that mixed particle populations allow the successful separation of 2, 5 and 10 µm particles at different flow rates. As shown in Fig. 5, the 2 µm and 5 µm particles were separated and recovered at 0.6 mL min⁻¹, while the 2 µm and 10 µm particles were separated and recovered at 1.0 mL min⁻¹. Furthermore, it is noticeable that the gap between focusing the streams of 81 µm for the 2 and 5 μm particle separation and 72 μm for the 2 and 10 μm particle separation is large enough to provide high purity separation. These results show the versatility of the fabricated chip to separate three particle sizes and possibly more if tested for 15 or 20 µm particles.

Table 2 shows a comparison between earlier designs that could focus or separate 2 µm particles in spiral channels, with the results obtained in this work. Unlike the current results, it is evident that all previous designs reported in the literature separated particles at relatively low flow rates. Since smaller particles require higher lift force values, smaller channel dimensions are required to achieve focusing; however, the size reduction comes at a great cost of high pressure values, even at low flow rates. Therefore, in addition to the complex setup needed for sheath-based focusing, expensive syringe pumps with high linear force are required. Furthermore, particle separation in a sheath-less device is harder because the operator only has control over the sample flow rate and there is no way to actively change the particle position by varying another input parameter, but once the device is well optimized, it becomes easier to repeat and use.




Fig. 5 The separation of 2 μ m particles from 10 μ m (a) at the flow rate of 1.0 mL min⁻¹, and 2 μ m particles from 5 μ m (b) at the flow rate of 0.6 $mL min^{-1}$

Conclusion

Paper

It has been shown that trapezoidal spiral channels are able to focus particles as small as 2 µm due to a secondary flow without the implementation of a sheath flow. The fluorescent microparticle experiments confirmed predictions based on the calculations of inertial forces, where the expected focusing position depended on the ratio between the inertial lift and drag forces and the magnitude of the dominant force. Reducing the suspension concentration from 0.0125% (w/v) to 0.0025% (w/v) was essential for 5 µm particles since the ratio of lift to drag forces was close to unity; therefore, secondary focusing was expected at high flow rates. As a result, the successful separation of 2 and 5 µm particles, and 2, 10 µm particles was achieved with a spatial separation of 81 µm for 2 and 5 µm particles and 72 µm for 2 and 10 µm particles. Recently, 30,32 spiral channels that were able to focus \sim 2 μm particles exhibited small hydraulic diameters and a sheath flow; however, in our design, the sheath-less bigger channel reduced the complexity of the microsystem and the required syringe pumps and increased the throughput. Separation operated at a flow rate of 0.6-1.0 mL min⁻¹, while in the literature, ^{30–32} the flow rate was less than 200 μL min⁻¹ even in other passive separation mechanisms besides spiral microfluidic channels.

Conflicts of interest

There are no conflicts to declare.

Acknowledgements

This research is funded by the German Academic Exchange Services (DAAD) under the project entitled "Inertial focusing for continuous nanoparticles separation in femtosecond laser 3D micromachined curved channels". Special thanks to Zaid Doofesh for his notable contribution in chips fabrication and early experiments. The first author would like to thank GJU for the sabbatical year that allowed significant progress in finalizing the paper.

References

- T. Yang, P. Paiè, G. Nava, F. Bragheri, R. M. Vazquez,
 P. Minzioni, M. Veglione, M. Di Tano, C. Mondello,
 R. Osellame and I. Cristiani, *Lab Chip*, 2015, 15, 1262–1266.
- M. Terashima, E. S. Freeman, R. E. Jinkerson and M. C. Jonikas, *Plant J.*, 2015, 81, 147–159.
- 3 D. T. Pederick, K. L. Richards, S. G. Piltz, R. Kumar, S. Mincheva-Tasheva, S. A. Mandelstam, R. C. Dale, I. E. Scheffer, J. Gecz, S. Petrou, J. N. Hughes and P. Q. Thomas, *Neuron*, 2018, **97**, 59–66.e5.
- 4 P. Li, Z. S. Stratton, M. Dao, J. Ritz and T. J. Huang, *Lab Chip*, 2013, 13, 602–609.
- 5 R. Moloudi, S. Oh, C. Yang, M. Ebrahimi Warkiani and M. W. Naing, *Microfluid. Nanofluidics*, 2018, **22**, 1–14.
- 6 E. J. Lim, T. J. Ober, J. F. Edd, G. H. McKinley and M. Toner, *Lab Chip*, 2012, **12**, 2199.

- 7 M. E. Warkiani, C. P. Lou, H. B. Liu and H. Q. Gong, *Biomed. Microdevices*, 2012, 14, 669–677.
- 8 P. Paiè, F. Bragheri, D. Di Carlo and R. Osellame, *Microsyst. Nanoeng.*, 2017, 3, 17027.
- 9 G. Velve-Casquillas, M. Le Berre, M. Piel and P. T. Tran, *Nano Today*, 2010, 5, 28–47.
- 10 M. E. Warkiani, A. K. P. Tay, B. L. Khoo, X. Xiaofeng, J. Han and C. T. Lim, *Lab Chip*, 2015, 15, 1101–1109.
- 11 M. E. Warkiani, B. L. Khoo, L. Wu, A. K. P. Tay, A. A. sga, S. Bhagat, J. Han and C. T. Lim, *Nat. Protoc.*, 2016, 11, 134–148.
- 12 C. Y. Lee, C. L. Chang, Y. N. Wang and L. M. Fu, *Int. J. Mol. Sci.*, 2011, **12**, 3263–3287.
- 13 A. Al-Halhouli, A. Alshare, M. Mohsen, M. Matar, A. Dietzel and S. Büttgenbach, *Micromachines*, 2015, **6**, 953–968.
- 14 W. Al-Faqheri, T. H. G. Thio, M. A. Qasaimeh, A. Dietzel, M. Madou and A. Al-Halhouli, *Microfluid. Nanofluidics*, 2017, 21, 1–23.
- 15 P. R. C. Gascoyne, X.-B. Wang, Y. Huang and F. F. Becker, *IEEE Trans. Ind. Appl.*, 1997, 33, 670–678.
- 16 B. Verbruggen, T. Tóth, M. Cornaglia, R. Puers, M. A. M. Gijs and J. Lammertyn, *Microfluid. Nanofluidics*, 2014, **18**, 91–102.
- 17 A. V. Larsen, L. Poulsen, H. Birgens, M. Dufva and A. Kristensen, *Lab Chip*, 2008, **8**, 818–821.
- 18 L. R. Huang, E. C. Cox, R. H. Austin and J. C. Sturm, *Science*, 2004, **304**, 987–990.
- 19 Z. Wu, Y. Chen, M. Wang and A. J. Chung, *Lab Chip*, 2016, **16**, 532–542.
- 20 L. Zeng, S. Balachandar and P. Fischer, *J. Fluid Mech.*, 2005, **536**, 1–25.
- 21 A. Al-Halhouli, W. Al-Faqheri, B. Alhamarneh, L. Hecht and A. Dietzel, *Micromachines*, 2018, 9, 171.
- 22 M. G. Lee, J. H. Shin, C. Y. Bae, S. Choi and J. K. Park, *Anal. Chem.*, 2013, 85, 6213–6218.
- 23 N. Nivedita, P. Ligrani and I. Papautsky, *Sci. Rep.*, 2017, 7, 1–10.
- 24 J. M. Martel and M. Toner, Phys. Fluids, 2012, 24, 032001.
- 25 P. Sajeesh and A. K. Sen, *Microfluid. Nanofluidics*, 2014, 17, 1–52.
- 26 A. Russom, A. K. Gupta, S. Nagrath, D. Di Carlo, J. F. Edd and M. Toner, New J. Phys., 2009, 11, 075025.
- 27 G. Guan, L. Wu, A. A. Bhagat, Z. Li, P. C. Y. Chen, S. Chao,C. J. Ong and J. Han, *Sci. Rep.*, 2013, 3, 1–9.
- 28 M. E. Warkiani, G. Guan, K. B. Luan, W. C. Lee, A. A. S. Bhagat, P. Kant Chaudhuri, D. S. W. Tan, W. T. Lim, S. C. Lee, P. C. Y. Chen, C. T. Lim and J. Han, *Lab Chip*, 2014, 14, 128–137.
- 29 L. Wang and D. S. Dandy, Adv. Sci., 2017, 4, 1700153.
- 30 J. Cruz, T. Graells, M. Walldén and K. Hjort, *Lab Chip*, 2019, 19, 1257–1266.
- 31 A. A. S. Bhagat, S. S. Kuntaegowdanahalli and I. Papautsky, *Lab Chip*, 2008, **8**, 1906–1914.
- 32 J. H. Lee, S. K. Lee, J. H. Kim and J. H. Park, *Sens. Actuators A Phys.*, 2019, **286**, 211–219.
- 33 A. Al-Halhouli, A. Albagdady, W. Al-Faqheri, J. Kottmeier, S. Meinen, L. J. Frey, R. Krull and A. Dietzel, RSC Adv., 2019, 9, 19197–19204.

This article is licensed under a Creative Commons Attribution-NonCommercial 3.0 Unported Licence.

Open Access Article. Published on 18 December 2019. Downloaded on 11/27/2025 11:08:47 PM.

- 34 Y. Zhou, Z. Ma and Y. Ai, Microsyst. Nanoeng., 2018, 4, 5.
- 35 A. A. S. Bhagat, S. S. Kuntaegowdanahalli, N. Kaval, C. J. Seliskar and I. Papautsky, *Biomed. Microdevices*, 2010, 12, 187–195.
- 36 K. Cheng and M. Akiyama, *Int. J. Heat Mass Transf.*, 1970, 13, 471–490.
- 37 T. T. Chandratilleke and A. Nursubyakto, *Int. J. Therm. Sci.*, 2003, **42**, 187–198.
- 38 W. R. Dean, *Proc. R. Soc. A Math. Phys. Eng. Sci.*, 1928, **121**, 402–420.
- 39 P. Erfle, J. Riewe, H. Bunjes and A. Dietzel, *Microfluid. Nanofluidics*, 2017, 1–12.
- 40 P. Erfle, J. Riewe, H. Bunjes and A. Dietzel, *Micromachines*, 2019, 10, 220.
- 41 A. Özbey, M. Karimzadehkhouei, S. Akgönül, D. Gozuacik and A. Koşar, *Sci. Rep.*, 2016, **6**, 1–11.
- 42 X. Casadevall i Solvas and A. DeMello, Matters, 2016, 1-7.