



Cite this: *RSC Adv.*, 2019, 9, 29078

## Rotary manifold for automating a paper-based *Salmonella* immunoassay†

Cody S. Carrell,<sup>a</sup> Rachel M. Wydallis,<sup>a</sup> Mridula Bontha,<sup>c</sup> Katherine E. Boehle,<sup>a</sup> J. Ross Beveridge,<sup>c</sup> Brian J. Geiss <sup>b</sup> and Charles S. Henry <sup>\*,a</sup>

Foodborne pathogens are responsible for hundreds of thousands of deaths around the world each year. Rapid screening of agricultural products for these pathogens is essential to reduce and/or prevent outbreaks and pinpoint contamination sources. Unfortunately, current detection methods are laborious, expensive, time-consuming and require a central laboratory. Therefore, a rapid, sensitive, and field-deployable pathogen-detection assay is needed. We previously developed a colorimetric sandwich immunoassay utilizing immuno-magnetic separation (IMS) and chlorophenol red- $\beta$ -D-galactopyranoside for *Salmonella* detection on a paper-based analytical device ( $\mu$ PAD); however, the assay required many sample preparation steps prior to the  $\mu$ PAD as well as laboratory equipment, which decreased user-friendliness for future end-users. As a step towards overcoming these limitations in resource-limited settings, we demonstrate a reusable 3D-printed rotational manifold that couples with disposable  $\mu$ PAD layers for semi-automated reagent delivery, washing, and detection in 65 minutes. After IMS to clean the sample, the manifold performs pipette-free reagent delivery and washing steps in a sequential order with controlled volumes, followed by enzymatic amplification and colorimetric detection using automated image processing to quantify color change. *Salmonella* was used as the target pathogen in this project and was detected with the manifold in growth media and milk with detection limits of  $4.4 \times 10^2$  and  $6.4 \times 10^2$  CFU mL<sup>-1</sup> respectively. The manifold increases user friendliness and simplifies immunoassays resulting in a practical product for in-field use and commercialization.

Received 4th September 2019  
Accepted 6th September 2019

DOI: 10.1039/c9ra07106g

rsc.li/rsc-advances

### 1. Introduction

Foodborne pathogens pose a major health risk around the world. Each year they are responsible for 600 million infections and over 400 000 deaths around the world (9.4 million and 1300 in the US).<sup>1,2</sup> Screening for specific pathogens could prevent large and deadly outbreaks; however, traditional pathogen detection assays are inadequate for testing large volumes of samples in a timely manner.<sup>3</sup> The most effective screening tests should be inexpensive, portable, rapid, and easy to use such that the assay can be performed in the field by an untrained end user.<sup>4</sup>

Foodborne pathogen detection is typically accomplished using culture-based methods.<sup>5</sup> Culture-based methods are well established, selective, sensitive, and benefit from decades of data to cross reference. Despite these advantages, culturing can take days to complete, requires trained laboratory personnel,

transportation to a central laboratory for analysis, and is prone to false negatives due to viable but non-culturable pathogens.<sup>3,4,6</sup> Therefore, culturing is an ill-suited tool to screen for foodborne pathogens, especially in resource-limited settings.

To decrease testing times, numerous rapid assays have been developed, including lateral flow immunoassays (LFAs), polymerase chain reaction (PCR), loop mediated isothermal amplification (LAMP), and enzyme linked immunosorbent assays (ELISAs). Other rapid assays have used nanomaterials and specific recognition elements like aptamers, phages, or antibodies for capture and detection.<sup>5,7</sup> While faster than culture-based methods, PCR and ELISA are both expensive, require complex instrumentation, and are unsuitable for in-field testing.<sup>8</sup> LAMP is a promising tool for resource limited settings, but it still requires heating to 65 °C, four sets of primers, and trained technicians for most detection techniques.<sup>9</sup> Finally, LFAs have been used for decades to detect pathogens, antibodies, and small biomarkers.<sup>10,11</sup> Although they are well established diagnostic tools, LFAs consistently suffer from poor detection limits without pre-enrichment steps such as culturing or PCR amplification.<sup>11,12</sup> Additionally, complex food sample matrices are difficult to use in an LFA format without significant sample pre-treatment or cleaning.

<sup>a</sup>Department of Chemistry, Colorado State University, USA. E-mail: chuck.henry@colostate.edu<sup>b</sup>Department of Microbiology, Immunology & Pathology, Colorado State University, USA<sup>c</sup>Department of Computer Science, Colorado State University, USA

† Electronic supplementary information (ESI) available. See DOI: 10.1039/c9ra07106g



Microfluidic paper-based analytical devices ( $\mu$ PADs) have become a common platform for point-of-care and field-based assays.<sup>13–17</sup>  $\mu$ PADs are small, portable, inexpensive, easy to dispose of, frequently require no external instrumentation, and can effectively store reagents.<sup>14,16,18,19</sup> Recently,  $\mu$ PADs have been used to detect multiple foodborne pathogens with the goal of creating a simple screening test.<sup>20–23</sup> In these works, traditional antibody-based immunoassays, enzymatic detection, and electrochemical methods were adapted for use in  $\mu$ PADs. Although they are a promising technology,  $\mu$ PADs frequently suffer from inadequate sensitivity and high detection limits compared to traditional methods like PCR and ELISA.<sup>14</sup> To improve detection limits and sensitivity in traditional immunoassays, washing steps and signal enhancement reagents are used.<sup>3,24,25</sup> However, these steps require timed and sequential delivery of reagents and/or washing agents, which  $\mu$ PADs are not typically designed to do without significant manual intervention from the end user.

Several groups have successfully demonstrated sequential reagent delivery in  $\mu$ PADs. Fu *et al.* accomplished this in 2010 with a 2D  $\mu$ PAD containing multiple inlets,<sup>26</sup> and Govindarajan *et al.* introduced folding, or origami,  $\mu$ PADs in 2012, where reagents are delivered in each folding step.<sup>27</sup> Since then, many other devices have been proposed including key shaped devices, sliding  $\mu$ PADs, rotating devices, dissolving sugar bridges, and complex channel geometries.<sup>26,28–32</sup> Although they accomplish the goal of sequential reagent delivery, many of these devices suffer from inconsistent results, as the end user is ultimately performing the folding or sliding, and others require multiple pipetting steps to continually add buffer.<sup>33</sup> Additionally, the more complex channels necessary for sequential delivery in these systems can result in lost analyte to the cellulose fibers and decreased sensitivity.<sup>24</sup> Instead of relying on paper alone for reagent delivery, several groups have developed paper–plastic hybrid devices for pathogen detection.<sup>34–36</sup> In this work we propose a  $\mu$ PAD housed in a re-usable 3D-printed rotational manifold capable of reproducible sequential reagent delivery for the detection of *Salmonella* via an immuno-magnetic separation (IMS) sandwich immunoassay.

Among foodborne pathogens, *Salmonella* is the most harmful, causing more worldwide infections (>90 million) and deaths (~155 000) than any other foodborne pathogen.<sup>37</sup> Therefore, *Salmonella* was chosen as the target in this study and the manifold was used to detect *Salmonella* in growth media and milk. Milk was used as a real-world sample as milk-borne infections are common in unpasteurized, raw milk.<sup>38</sup> Specificity against *E. coli* and other bacteria naturally found in milk was also demonstrated. To simplify the assay, we employed colorimetric detection using an enzymatic reaction that cleaves chlorophenol red- $\beta$ -D-galactopyranoside (CPRG) to chlorophenol red (CPR), which is indicated by a yellow to red color change. The color change can be qualitatively observed by the naked eye or quantitatively analyzed through image analysis. We used a flood fill algorithm and analysis system that lowered the limit of detection by an order of magnitude over previous color analysis techniques. Our target limit of detection (LOD) was  $\leq 10^2$  colony forming units (CFU) per mL, which has been

reported as the infectious dose of *Salmonella* in food samples.<sup>39,40</sup> This LOD was realized in milk and media samples and is also the lowest detection limit that has been realized for direct colorimetric detection of *Salmonella* on a paper-based sensor with no pre-enrichment.<sup>41,42</sup> By combining a 3D-printed manifold with paper layers we have increased the capabilities of a  $\mu$ PAD without sacrificing user-friendliness. We believe that many of the pitfalls of traditional  $\mu$ PADs may be solved by integrating inexpensive plastic platforms or manifolds like the one described here.

## 2. Materials and methods

### 2.1. Rotational manifold

The 3D-printed manifold was printed using a FormLabs Form2 printer with their Clear V2 resin. All 3D-printed parts were designed in OnShape, a cloud-based CAD software. The paper channels, waste-pads, and sample layers are made of Fusion 5 paper, a conjugate-release membrane from GE healthcare (Waukesha, WI). The reagent layer and sample layers were made from Fusion 5 pieces, 3M transparency sheets, and 10 mil Fellowes® lamination sheets. The two layers are sealed using a TruLam laminator at 360 °F. The surfaces of the reagent and sample layers are coated in Never-Wet®, a hydrophobic spray-on product from Rust-Oleum®. All designs for the two layers and paper pieces were created in CorelDraw (Fig. S1†) and were cut with a 30 W CO<sub>2</sub> laser cutter (Epilog Model 10000).

### 2.2. IMS assay reagents

The reagents used in the *Salmonella* immunoassay include: Dynabeads™ M-280 Tosylactivated (Invitrogen™, purchased from Thermo Fisher Scientific, Milwaukee, WI); anti-*Salmonella typhimurium* 0–4 monoclonal mouse antibody (Abcam, Cambridge, MA ab8274); *Salmonella* antibody–biotin conjugate (Invitrogen™, purchased from Thermo Fisher Scientific); streptavidin– $\beta$ -galactosidase conjugate (Invitrogen™, purchased from Thermo Fisher Scientific), and chlorophenol red- $\beta$ -D-galactopyranoside (purchased from Sigma Aldrich, St. Louis, MS). The bacterial strains used in this work were *Salmonella enterica* serovar *typhimurium* and *Escherichia coli* DH5 $\alpha$ . The cells were grown in a Difco (*Salmonella*) and LB (*E. coli*) media for 12 h to a final concentration  $\sim 10^8$  CFU mL<sup>–1</sup>. After each growth the cell concentration was quantified by serial dilution and plating. The cells were spiked into and diluted in growth media or milk at the desired concentrations before running the assay with the manifold. The milk used was DairyGold ultra-high pasteurized whole milk. The buffer used in the manifold was a 1 $\times$  solution of phosphate buffered saline (PBS) at pH 7.4 diluted in Milli-Q water. 2.5–4.5  $\mu$ m pink fluorescent magnetic beads from Spherotech (Product #FP-4058-2) were used to perform the bead retention studies with a Dino-Lite EDGE fluorescent digital microscope at an excitation wavelength of 575 nm.

### 2.3. Assay steps

One mL of liquid sample is incubated with 5  $\mu$ L of 5 mg mL<sup>–1</sup> magnetic bead–*Salmonella* antibody conjugate for 15 min at



room temperature in a microcentrifuge tube. After a washing step using an external DynaMag-2 (Thermo Fisher Scientific) magnet for IMS, the magnetic bead-antibody-*Salmonella* conjugate is reconstituted in 60  $\mu\text{L}$  of PBS buffer. A 15  $\mu\text{L}$  aliquot of the concentrated solution was placed on the sample layer and the remainder of the assay (enzyme labeling and detection) is performed in the rotational manifold.

## 2.4. Device operation

The manifold consists of four 3D printed pieces: the manifold top, center, and bottom, and the sample layer insert (Fig. 1). The top, center, and bottom pieces are all held together by bolts and springs. The springs allow the center piece to rotate while the top and bottom pieces are held in place. The center piece contains a slot for the reagent card. Within the reagent card are eight reagent channels and waste pads. The reagent card is made of five layers: a transparency sheet with a hydrophobic coating, the paper reagent channels, a 3 mil lamination sheet, the paper waste pads, and a 10 mil lamination sheet. The waste pads and sample layers are stacked on top of each other but are separated by the 3 mil lamination sheet. Four reagent channels store reagents and four are empty channels used for washing. Two reagent and two washing channels are needed to run the assay in the current format, so eight total channels allow two tests to be run with one reagent card. In the four channels needed, the first channel contains 0.4  $\mu\text{g}$  of biotin-labeled *Salmonella* antibody that is deposited and dried prior to the assay; the second is empty for washing; the third contains 0.2  $\mu\text{g}$  of dried streptavidin  $\beta$ -galactosidase; and the fourth is empty. 5  $\mu\text{L}$  aliquots of 40  $\mu\text{g mL}^{-1}$  and 80  $\mu\text{g mL}^{-1}$  streptavidin  $\beta$ -galactosidase and biotin-ab are added to the reagent channels to obtain 0.2  $\mu\text{g}$  and 0.4  $\mu\text{g}$  of dried reagent respectively. These reagents are all added and dried before the assay is run so the end-user does not need to perform the reagent addition steps.

When the manifold is assembled, the reagent layer rotates with the center piece and the sample layer insert remains stationary. As the center piece rotates, the top and bottom pieces move vertically while keeping pressure on the center

piece. After a 45° rotation the center piece clicks into a slot and the top and bottom pieces collapse down with it. A video in the ESI† shows the device rotation, clicking mechanism, and flow through the layers.

With each rotation, an opening in the reagent layer is aligned with a wick in the bottom piece of the manifold and the next reagent layer and waste pad are aligned with the sample layer. The wick sits in a buffer reservoir and, buffer flows through the reagent channel, through sample layer, and into a waste pad. The flow through the system either delivers a reagent to the sample layer or simply washes excess reagent away from the sample layer. Magnetic beads, and anything conjugated to them, will not wash through the sample layer because of a magnet that's fitted underneath the sample layer in the sample layer insert. The end-user knows to rotate the device after the waste pad is saturated ( $\sim 2$  min).

## 2.5. Colorimetric detection

The final step of the assay is to add 25  $\mu\text{L}$  of a colorimetric substrate (CPRG) to the sample layer with a pipette. In the presence of  $\beta$ -galactosidase, and therefore *Salmonella*, CPRG (Yellow) will be turned over to CPR (Red). The [CPR] can be monitored *via* image capture and analysis. The image of the colored sample layer is taken inside a light box. The light box consists of two shells. The inner shell houses the sample layer and is made of clear acrylic. The acrylic was sanded until cloudy to diffuse light and prevent glare on the sample. The outer shell covers the first and is made of black acrylic so no ambient light can reach the sample layer. The outer shell contains 35 white and warm white LEDs built into the walls and powered by a 9 V battery. The inner and outer shells contain an opening for a smartphone to capture an image (Fig. S2†).

In this work we used a Samsung Galaxy S7 smartphone to capture static images of the sample layer at different time points. In the past, our group has frequently used manual processing with NIH ImageJ to quantify color changes.<sup>42–45</sup> To decrease the manual labor associated with ImageJ we used a flood-fill algorithm available in the OpenCV computer vision




Fig. 1 CAD rendering (A) and images (B) of the rotational manifold. In the CAD renderings all gray portions are 3D printed, dark blue are lamination or transparency sheets, and white are exposed Fusion 5 paper, and light blue are Fusion 5 paper covered by lamination or transparency sheets.



library.<sup>46</sup> Flood-fill identifies the color pixels in our sample region by employing a recursive region growing process. The regions grow out from a seed known to be in the sample. Flood-fill adds pixels relatively similar in color until no more similar pixels are adjacent to the region. In our case this means it stops at the edge of the oval paper that bounds our sample. The average RGB color values in the resulting set of pixels is then analyzed for each sample.<sup>47</sup> When compared to manual analysis using NIH ImageJ, the automated algorithm returned nearly identical results (Fig. S3†).

We define the signal in each trial as the difference in color between the blank and the sample. To quantify this difference the vector length in 3D RGB space between our blank RGB coordinates and the sample RGB coordinates ( $\Delta\text{RGB}$ ) was calculated using eqn (1) as previously described.<sup>48,49</sup>

$$\Delta\text{RGB} = \sqrt{(R_s - R_0)^2 + (G_s - G_0)^2 + (B_s - B_0)^2} \quad (1)$$

where  $R_s$ ,  $G_s$ , and  $B_s$  are the RGB values for the sample being tested, and  $R_0$ ,  $G_0$ , and  $B_0$  are the RGB values for each blank. The magnitude of the signal was larger for  $\Delta\text{RGB}$  versus a single-color channel, thus increasing sensitivity. Additionally, eqn (1) yields 9 data points with only 3 repeats:  $\Delta S_1\text{Bl}_1$ ,  $\Delta S_1\text{Bl}_2$ ,  $\Delta S_1\text{Bl}_3$ ,  $\Delta S_2\text{Bl}_1$ ,  $\Delta S_2\text{Bl}_2$ ,  $\Delta S_2\text{Bl}_3$ ,  $\Delta S_3\text{Bl}_1$ ,  $\Delta S_3\text{Bl}_2$ ,  $\Delta S_3\text{Bl}_3$ , where S and Bl are the sample and the blank respectively and the number denotes the trial. The larger signal and number of data points resulted in an order of magnitude improvement in detection limit when compared to the use of a single channel (Fig. S4†).

### 3. Results and discussion

#### 3.1. Assay steps and design

In a previous publication we presented an immuno-magnetic separation sandwich immunoassay for *Salmonella* detection.<sup>42</sup> The chemistry of the immunoassay remains the same in this work and is discussed in depth in our previous publication. Briefly, *Salmonella* is isolated from a sample using *Salmonella*-antibody labeled magnetic beads. After a short incubation with the sample the beads are removed from solution with an external magnet, washed, and reconstituted in a small volume of buffer to concentrate the sample. A secondary *Salmonella*-antibody labeled with biotin is then introduced to the *Salmonella*-magnetic bead complex solution. After an incubation and washing step, streptavidin-labeled  $\beta$ -galactosidase is added to the *Salmonella*-magnetic bead-biotin conjugate solution. Finally, after another incubation and washing step a small volume of the *Salmonella*-magnetic bead-biotin- $\beta$ -galactosidase conjugate solution reacts with the substrate CPRG. After a set reaction time, a color change from yellow to red indicates the presence of  $\beta$ -galactosidase, and therefore *Salmonella*. Previously, all steps of the assay were performed in solution in a microcentrifuge tube until the final step, where a small aliquot of magnetic bead-*Salmonella*-enzyme conjugate was added to a paper spot with dried substrate. The assay worked but required thirteen pipetting steps. To improve ease-of-use, we designed a rotational manifold to minimize pipetting steps by washing and delivering reagents to the test zone (Fig. 2).

Here, the initial step of sample incubation with antibody-labeled magnetic beads remains in-solution to ensure a sufficient number of bacteria are available for conjugation to the magnetic beads. After the bacteria is isolated from the sample, a 15  $\mu\text{L}$  aliquot of the magnetic bead-bacteria complex is added to a small paper sample-layer and the remaining steps are completed in the device.

#### 3.2. Device design and operation

Schematic drawings and a photograph of the manifold are shown in Fig. 1. An important feature of the device is that the buffer volume delivered in each step can be controlled by changing the size of the waste pad. Fusion 5 paper holds  $0.422 \pm 0.006 \mu\text{L mm}^{-2}$  (Fig. S5†), so by changing the surface area of the waste pad the volume of buffer used in each step can be customized. Volume control of each reagent delivery and washing step is critical when using the device for different assays that may need more thorough washing. Flow through each layer of the device is illustrated in Fig. 3. The sample layer sits on the 3D-printed sample layer insert, which fits into the bottom manifold piece with lego-style fittings. A magnet screws into the bottom of the insert and is positioned directly underneath the sample layer. The magnet ensures that the magnetic beads are not washed away during reagent delivery and washing steps (Fig. S6†). Using magnetic beads and magnets to create a test zone allowed us to use membranes other than nitrocellulose, which is expensive and has a short shelf life.<sup>50</sup> Additionally, the sample in the manifold is stationary on the sample layer and therefore will not be lost to the membrane during flow through the device as is the case in traditional LFAs. Sample loss to paper channels is a problem in paper-based devices,<sup>24,51</sup> and a stationary sample can improve detection limits.


After the final washing step, the insert and sample layer are removed from the device. 25  $\mu\text{L}$  of 2.5 mM CPRG is pipetted onto the sample layer containing the conjugated system and the substrate reacts with any enzyme present for 40 min. A picture is then taken of the sample layer inside the light box. The image is analyzed as described in the materials and methods to determine the *Salmonella* concentration. Using the rotational manifold, the number of pipetting steps needed to complete the IMS sandwich immunoassay was decreased from thirteen to four, which significantly reduces labor and complexity for the end user. Each sample addition step requires  $\sim 1$  second of user input (rotating the device) compared to traditional pipetting and IMS that can take multiple minutes per step. Additionally, the waste generated during each step of a traditional in-solution IMS assay exceeds 1 mL, while our assay uses roughly 0.1 mL per step, reducing cost and storage needs.

#### 3.3. Assay optimization and parameters

**3.3.1 Paper type.** Fusion 5 was chosen as the paper type for all portions of the manifold: reagent layer, waste pad, and sample layer. Fusion 5 was developed as a substrate for all portions of an LFA.<sup>52</sup> In our device, Fusion 5 demonstrated faster flow and lower noise from non-specific enzyme adsorption than other papers tested (Whatman I and IV). The faster








**Fig. 2** Schematic demonstrating the concept of sequential reagent delivery using a rotating reagent storage card and a stationary sample layer (A) add sample containing magnetic beads conjugated to your target analyte to the sample layer. (B) Add the sample layer to the device. Biotinylated antibodies will be introduced. (C) Rotate the device to a washing step used to remove excess biotinylated antibodies. (D) Continue rotating until streptavidin  $\beta$ -galactosidase has been introduced and washed. (E) Remove sample layer, add substrate CPRG, and observe color change.


flow rates increased washing efficiency and decreased assay time. Additionally, Fusion 5 contains plastic binder meant to stabilize its mechanical properties, allowing us to laminate the reagent layer under high pressure and heat without decreasing flow rates.

**3.3.2 Sample layer.** A major challenge in developing the rotational manifold and corresponding assay was minimizing the noise in the blank signal. Although using Fusion 5 helped improve washing,  $\beta$ -galactosidase would consistently get trapped on the edges or ends of the sample layer. To overcome this challenge, we blocked the surface of the sample layer with 10  $\mu$ L of 5 mg mL<sup>-1</sup> bovine serum albumin (BSA) and updated the shape to an oval. The benefit of an oval sample layer is that there are no corners where the enzyme can be trapped. Once these changes were implemented, the washing efficiency of the device was improved, and blank samples remained yellow after all washing steps were performed (Fig. 4).

**3.3.3 Reaction times.** The initial sample incubation with the magnetic beads was 15 min. In more complex samples this incubation may need to be lengthened; however, in our liquid sample matrices 15 min was sufficient. The reaction times for each reagent delivery step were roughly 2 min, or the time it took for the waste pad to completely saturate. Once saturated, flow through the waste pad stopped and no new reagents reached the immobilized magnetic bead complex, eliminating any benefit of longer incubation. The shape of the waste pad was fan-like to provide continuous capillary pressure and more constant flow rates.<sup>53</sup> Finally, the substrate and enzyme reacted for 40 min. In optimization studies of CPRG and  $\beta$ -galactosidase the longer the wait time the higher the signal, so long as the sample layer did not dry out. Above 25  $\mu$ L of substrate, the sample layer began to leak, and at 25  $\mu$ L the sample began to dry out after 60 min (Fig. S7†). The 40 min reaction time was a compromise



**Fig. 3** Flow through the device begins in the buffer reservoir where PBS is wicked through a paper wick into the reagent channel. Buffer flows through the reagent channel and delivers reagents to the conjugated system on the sample layer before washing away excess reagents to the waste pad.



**Fig. 4** Effect of sample layer shape and blocking with BSA. The blank sample was an image of the sample layer with CPRG without any  $\beta$ -galactosidase ever introduced. All other samples had  $\beta$ -galactosidase washed through the sample layer into the waste pads.



between assay speed and sensitivity. If total assay time was not a concern, one could lower detection limits by increasing the reaction time and minimizing evaporation with a small air-tight enclosure. With the 40 min reaction, the entire assay can be performed in roughly 65 min. In samples with higher bacteria concentrations, a qualitative color change could be observed in as little as 5 min after adding the substrate. To reach a lower limit of detection, however, 40 min of reaction time was needed to differentiate between the blank and the lowest bacteria concentration.

**3.3.4 Image capture.** An image captured by a smartphone was used to quantify *Salmonella*. Smartphones are an excellent tool for in-field measurements as most end-users will already own a smartphone, cutting down on the cost of the assay and increasing user-friendliness. Additionally, numerous papers on colorimetric  $\mu$ PADs use a smartphone for detection.<sup>54</sup> We correct for differences in ambient lighting with a custom light box that contains an opening for a smartphone to take a picture. The light-box improved consistency between images (Table S1†). The % RSD of three identical blank images of the sample layer taken in three different ambient light settings decreased from 19.9% without the box to 3.3% in the box.

### 3.4. Colorimetric *Salmonella* detection

A *Salmonella* detection assay was used to test the functionality of the manifold. Here, *Salmonella* cultured in Difco nutrient broth was diluted to concentrations spanning from  $10^2$  to  $10^7$  CFU mL<sup>-1</sup>. The results are shown in Fig. 5 ( $n = 3$ ). The data was fit to a 4-parameter logistic curve (Equation ESI†) with a  $\chi^2 = 13.02$  ( $\alpha = 0.05$ ,  $\chi_{\text{critical}} = 15.51$ ).<sup>55</sup> The logistic fit was chosen because the antibody binding kinetics of the sandwich immunoassay are the limiting steps. The LOD was calculated by finding  $3 \times \text{SD} + \text{mean}$  of the blank and plugging it into the fit equation. An LOD of  $4.40 \times 10^2$  CFU mL<sup>-1</sup> was realized using the manifold. In our previous entirely in-solution assay, we found a LOD on the same order of magnitude, which demonstrates the capabilities of our device to perform a multi-step immunoassay on a  $\mu$ PAD.<sup>42</sup>




Fig. 5 Dose response curve ( $n = 3$ ) for *Salmonella* in media detected using the rotational manifold. The curve was fit to a 4-parameter logistic model and a LOD of  $2.9 \times 10^5$  was achieved.

### 3.5. Specificity and real-world samples

To confirm the specificity of our *Salmonella* assay,  $10^7$  CFU mL<sup>-1</sup> DH5 $\alpha$  *E. coli* in media was tested with the device using *Salmonella* specific antibodies (Fig. 6A). The low signal for *E. coli* samples demonstrates the specificity of the assay. A low signal also indicates efficient washing throughout the system as *E. coli* produces  $\beta$ -galactosidase naturally, and any excess enzyme would increase the signal.

Finally, the assay was run using milk spiked with *Salmonella* (Fig. 6B). In milk, we found a detection limit of  $6.36 \times 10^2$  CFU mL<sup>-1</sup>. The data was once again fit to a 4-parameter logistic curve ( $\chi^2 = 4.13$ ,  $\alpha = 0.05$ ,  $\chi_{\text{critical}} = 14.07$ ) and the LOD found using the fit equations and  $3 \times \text{SD} + \text{the mean}$  of the blank. Although the detection limit is slightly higher, the assay performed nearly as well in milk *versus* growth media. The small difference in LOD could be attributed to decreased antibody binding efficiency in the first step of the assay due to non-specific adsorption of other biomolecules in milk onto the magnetic beads. The international microbiological criteria for dairy products states that *Salmonella* should have a concentration of  $<1$  cell per mL.<sup>56</sup> The only way to meet the required LOD with our device, and any other inexpensive point-of-care system, is through some form of pre-enrichment like culturing. Although we believe that the rotary manifold is an innovative way to perform




Fig. 6 (A) Specificity study using *E. coli* at  $10^7$  CFU mL<sup>-1</sup> compared to blank samples in milk, growth media, and a positive *Salmonella* sample. (B) Dose-response curve ( $n = 3$ ) of *Salmonella* in milk detected using the rotational manifold. The curve was fit to a 4-parameter logistic model and a LOD of  $6.4 \times 10^2$  CFU mL<sup>-1</sup> was achieved.



IMS as a rapid screening tool, further technologies will need to be developed to meet the strict requirements of the food safety industry.

## 4. Conclusions and future directions

The field of microfluidic paper-based devices has grown tremendously over the last 10 years in academic settings. There are many reasons for the disconnect between published works and commercial products, but the most commonly cited argument is poor sensitivity and specificity in  $\mu$ PADs.<sup>14,19</sup> Attempts to improve these criteria are often successful but result in assays that are far too complicated for their intended use at the point-of-care with untrained end-users. We believe that coupling  $\mu$ PADs with other inexpensive materials could enhance sensitivity and specificity without sacrificing usability. In this work we designed a 3D-printed rotational manifold to perform multiple reagent delivery and washing steps in a sequential order. The washings steps increase sensitivity over traditional LFAs and the manifold ensures that the additional steps do not infringe on usability. The manifold was used to detect *Salmonella*, however with a simple change of reagents could be used to detect a large host of other pathogens, proteins, and/or biomarkers. In the future we will continue working with this device to detect additional biomolecules in a multiplexed format. We also recognize that although we simplified testing and reduced manual input, the four pipetting steps still required are not ideal. Some liquid transfer will always be necessary in an IMS assay, however, future work will be dedicated to decreasing user intervention further. For example, work has begun to implement a motorized rotational mechanism and further automate image capture and analysis with a web-based interface. With devices like the one described here, we anticipate an increase in new  $\mu$ PADs on the market.

## Conflicts of interest

There are no conflicts to declare.

## Acknowledgements

Funding for this project was provided by the United States Department of Agriculture, Grant Number #16-74000589 through the National Wildlife Research Center. The authors would also like to thank the members of the Henry and Geiss laboratories and Dr Alan Franklin for their guidance throughout the project.

## References

- 1 E. Scallan, R. M. Hoekstra, F. J. Angulo, R. V. Tauxe, M. A. Widdowson, S. L. Roy, J. L. Jones and P. M. Griffin, Foodborne Illness Acquired in the United States-Major Pathogens, *Emerging Infect. Dis.*, 2011, **17**(1), 7–15.
- 2 Group, F. D. B. E. R., *WHO Estimates of the Global Burden of Foodborne Diseases*, Organization, W. H., 2015.
- 3 J. W. F. Law, N. S. Ab Mutalib, K. G. Chan and L. H. Lee, Rapid methods for the detection of foodborne bacterial pathogens: principles, applications, advantages and limitations, *Front. Microbiol.*, 2015, **5**, 19.
- 4 V. Velusamy, K. Arshak, O. Korostynska, K. Oliwa and C. Adley, An overview of foodborne pathogen detection: In the perspective of biosensors, *Biotechnol. Adv.*, 2010, **28**(2), 232–254.
- 5 V. M. Bohaychuk, G. E. Gensler, R. K. King, J. T. Wu and L. M. McMullen, Evaluation of detection methods for screening meat and poultry products for the presence of foodborne pathogens, *J. Food Prot.*, 2005, **68**(12), 2637–2647.
- 6 P. K. Mandal, A. K. Biswas, K. Choi and U. K. Pal, Methods for Rapid Detection of Foodborne Pathogens: An Overview, *Am. J. Food Technol.*, 2011, **6**(2), 15.
- 7 J. Chen, S. M. Andler, J. M. Goddard, S. R. Nugen and V. M. Rotello, Integrating recognition elements with nanomaterials for bacteria sensing, *Chem. Soc. Rev.*, 2017, **46**(5), 1272–1283.
- 8 X. Zhao, C. W. Lin, J. Wang and D. H. Oh, Advances in Rapid Detection Methods for Foodborne Pathogens, *J. Microbiol. Biotechnol.*, 2014, **24**(3), 297–312.
- 9 Y. Mori and T. Notomi, Loop-mediated isothermal amplification (LAMP): a rapid, accurate, and cost-effective diagnostic method for infectious diseases, *J. Infect. Chemother.*, 2009, **15**(2), 62–69.
- 10 Y. Zhao, H. R. Wang, P. P. Zhang, C. Y. Sun, X. C. Wang, X. R. Wang, R. F. Yang, C. B. Wang and L. Zhou, Rapid multiplex detection of 10 foodborne pathogens with an up-converting phosphor technology-based 10-channel lateral flow assay, *Sci. Rep.*, 2016, **6**, 21342.
- 11 G. A. Posthuma-Trumpie, J. Korf and A. van Amerongen, Lateral flow (immuno) assay: its strengths, weaknesses, opportunities and threats. A literature survey, *Anal. Bioanal. Chem.*, 2009, **393**(2), 569–582.
- 12 J. Hu, S. Q. Wang, L. Wang, F. Li, B. Pingguan-Murphy, T. J. Lu and F. Xu, Advances in paper-based point-of-care diagnostics, *Biosens. Bioelectron.*, 2014, **54**, 585–597.
- 13 A. W. Martinez, S. T. Phillips, M. J. Butte and G. M. Whitesides, Patterned paper as a platform for inexpensive, low-volume, portable bioassays, *Angew. Chem., Int. Ed.*, 2007, **46**(8), 1318–1320.
- 14 K. Yamada, H. Shibata, K. Suzuki and D. Citterio, Toward practical application of paper-based microfluidics for medical diagnostics: state-of-the-art and challenges, *Lab Chip*, 2017, **17**(7), 1206–1249.
- 15 A. W. Martinez, S. T. Phillips, G. M. Whitesides and E. Carrilho, Diagnostics for the Developing World: Microfluidic Paper-Based Analytical Devices, *Anal. Chem.*, 2010, **82**(1), 3–10.
- 16 Y. Y. Yang, E. Noviana, M. P. Nguyen, B. J. Geiss, D. S. Dandy and C. S. Henry, Paper-Based Microfluidic Devices: Emerging Themes and Applications, *Anal. Chem.*, 2017, **89**(1), 71–91.
- 17 Y. Y. Xia, J. Si and Z. Y. Li, Fabrication techniques for microfluidic paper-based analytical devices and their



- applications for biological testing: A review, *Biosens. Bioelectron.*, 2016, **77**, 774–789.
- 18 A. M. Lopez-Marzo and A. Merkoci, Paper-based sensors and assays: a success of the engineering design and the convergence of knowledge areas, *Lab Chip*, 2016, **16**(17), 3150–3176.
  - 19 C. Carrell, A. Kava, M. Nguyen, R. Menger, Z. Munshi, Z. Call, M. Nussbaum and C. Henry, Beyond the lateral flow assay: A review of paper-based microfluidics, *Microelectron. Eng.*, 2019, **206**, 45–54.
  - 20 K. Kant, M.-A. Shahbazi, V. P. Dave, T. A. Ngo, V. A. Chidambara, L. Q. Than, D. D. Bang and A. Wolff, Microfluidic devices for sample preparation and rapid detection of foodborne pathogens, *Biotechnol. Adv.*, 2018, **36**, 1003–1024.
  - 21 J. C. Jokerst, J. A. Adkins, B. Bisha, M. M. Mentele, L. D. Goodridge and C. S. Henry, Development of a Paper-Based Analytical Device for Colorimetric Detection of Select Foodborne Pathogens, *Anal. Chem.*, 2012, **84**(6), 2900–2907.
  - 22 J. Bhardwaj, S. Devarakonda, S. Kumar and J. Jang, Development of a paper-based electrochemical immunosensor using an antibody-single walled carbon nanotubes bio-conjugate modified electrode for label-free detection of foodborne pathogens, *Sens. Actuators, B*, 2017, **253**, 115–123.
  - 23 J. Park, J. H. Shin and J. K. Park, Pressed Paper-Based Dipstick for Detection of Foodborne Pathogens with Multistep Reactions, *Anal. Chem.*, 2016, **88**(7), 3781–3788.
  - 24 M. P. Nguyen, N. A. Meredith, S. P. Kelly and C. S. Henry, Design considerations for reducing sample loss in microfluidic paper-based analytical devices, *Anal. Chim. Acta*, 2018, **1017**, 20–25.
  - 25 E. Fu, P. Kauffman, B. Lutz and P. Yager, Chemical signal amplification in two-dimensional paper networks, *Sens. Actuators, B*, 2010, **149**(1), 325–328.
  - 26 E. Fu, B. Lutz, P. Kauffman and P. Yager, Controlled reagent transport in disposable 2D paper networks, *Lab Chip*, 2010, **10**(7), 918–920.
  - 27 A. V. Govindarajan, S. Ramachandran, G. D. Vigil, P. Yager and K. F. Bohringer, A low cost point-of-care viscous sample preparation device for molecular diagnosis in the developing world; an example of microfluidic origami, *Lab Chip*, 2012, **12**(1), 174–181.
  - 28 R. B. Channon, M. P. Nguyen, A. G. Scorzelli, E. M. Henry, J. Volckens, D. S. Dandy and C. S. Henry, Rapid flow in multilayer microfluidic paper-based analytical devices, *Lab Chip*, 2018, **18**(5), 793–802.
  - 29 A. Apilux, Y. Ukita, M. Chikae, O. Chailapakul and Y. Takamura, Development of automated paper-based devices for sequential multistep sandwich enzyme-linked immunosorbent assays using inkjet printing, *Lab Chip*, 2013, **13**(1), 126–135.
  - 30 H. Liu, X. Li and R. M. Crooks, Paper-Based SlipPAD for High-Throughput Chemical Sensing, *Anal. Chem.*, 2013, **85**(9), 4263–4267.
  - 31 B. Lutz, T. Liang, E. Fu, S. Ramachandran, P. Kauffman and P. Yager, Dissolvable fluidic time delays for programming multi-step assays in instrument-free paper diagnostics, *Lab Chip*, 2013, **13**(14), 2840–2847.
  - 32 X. E. Sun, B. W. Li, C. Y. Tian, F. B. Yu, N. Zhou, Y. H. Zhan and L. X. Chen, Rotational paper-based electrochemiluminescence immunodevices for sensitive and multiplexed detection of cancer biomarkers, *Anal. Chim. Acta*, 2018, **1007**, 33–39.
  - 33 E. Fu and C. Downs, Progress in the development and integration of fluid flow control tools in paper microfluidics, *Lab Chip*, 2017, **17**(4), 614–628.
  - 34 L. Lafleur, D. Stevens, K. McKenzie, S. Ramachandran, P. Spicar-Mihalic, M. Singhal, A. Arjyal, J. Osborn, P. Kauffman, P. Yager and B. Lutz, Progress toward multiplexed sample-to-result detection in low resource settings using microfluidic immunoassay cards, *Lab Chip*, 2012, **12**(6), 1119–1127.
  - 35 R. H. Tang, H. Yang, J. R. Choi, Y. Gong, J. Hu, T. Wen, X. J. Li, B. Xu, Q. B. Mei and F. Xu, Paper-based device with on-chip reagent storage for rapid extraction of DNA from biological samples, *Microchim. Acta*, 2017, **184**(7), 2141–2150.
  - 36 U. M. Jalal, G. J. Jin and J. S. Shim, Paper-Plastic Hybrid Microfluidic Device for Smartphone-Based Colorimetric Analysis of Urine, *Anal. Chem.*, 2017, **89**(24), 13160–13166.
  - 37 S. E. Majowicz, J. Musto, E. Scallan, F. J. Angulo, M. Kirk, S. J. O'Brien, T. F. Jones, A. Fazil, R. M. Hoekstra and Int. Collaboration Enteric Dis., B., The Global Burden of Nontyphoidal *Salmonella* Gastroenteritis, *Clin. Infect. Dis.*, 2010, **50**(6), 882–889.
  - 38 S. P. Oliver, K. J. Boor, S. C. Murphy and S. E. Murinda, Food Safety Hazards Associated with Consumption of Raw Milk, *Foodborne Pathog. Dis.*, 2009, **6**(7), 793–806.
  - 39 M. J. Blaser and L. S. Newman, A Review of Human Salmonellosis. 1. Infective Dose, *Rev. Infect. Dis.*, 1982, **4**(6), 1096–1106.
  - 40 M. H. Kothary and U. S. Babu, Infective dose of foodborne pathogens in volunteers: A review, *J. Food Saf.*, 2001, **21**(1), 49–73.
  - 41 C. Z. Li, K. Vandenberg, S. Prabhulkar, X. N. Zhu, L. Schneper, K. Methee, C. J. Rosser and E. Almeida, Paper based point-of-care testing disc for multiplex whole cell bacteria analysis, *Biosens. Bioelectron.*, 2011, **26**(11), 4342–4348.
  - 42 M. Srisa-Art, K. E. Boehle, B. J. Geiss and C. S. Henry, Highly Sensitive Detection of *Salmonella typhimurium* Using a Colorimetric Paper-Based Analytical Device Coupled with Immunomagnetic Separation, *Anal. Chem.*, 2018, **90**(1), 1035–1043.
  - 43 K. E. Boehle, J. Gilliland, C. R. Wheeldon, A. Holder, J. A. Adkins, B. J. Geiss, E. P. Ryan and C. S. Henry, Utilizing Paper-Based Devices for Antimicrobial-Resistant Bacteria Detection, *Angew. Chem., Int. Ed.*, 2017, **56**(24), 6886–6890.
  - 44 K. E. Boehle, C. S. Carrell, J. Caraway and C. S. Henry, Paper-Based Enzyme Competition Assay for Detecting Falsified beta-Lactam Antibiotics, *ACS Sens.*, 2018, **3**(7), 1299–1307.
  - 45 J. A. Adkins, K. Boehle, C. Friend, B. Chamberlain, B. Bisha and C. S. Henry, Colorimetric and Electrochemical Bacteria





- Detection Using Printed Paper- and Transparency-Based Analytic Devices, *Anal. Chem.*, 2017, **89**(6), 3613–3621.
- 46 G. Bradski, The OpenCV library, *Dr. Dobbs's Journal*, 2000, **25**(11), 120–125.
- 47 K. E. Boehle, E. Doan, S. Henry, J. R. Beveridge, S. L. Pallickara and C. S. Henry, Single board computing system for automated colorimetric analysis on low-cost analytical devices, *Anal. Methods*, 2018, **10**(44), 5282–5290.
- 48 R. C. Murdock, L. Shen, D. K. Griffin, N. Kelley-Loughnane, I. Papautsky and J. A. Hagen, Optimization of a Paper-Based ELISA for a Human Performance Biomarker, *Anal. Chem.*, 2013, **85**(23), 11634–11642.
- 49 L. Shen, J. A. Hagen and I. Papautsky, Point-of-care colorimetric detection with a smartphone, *Lab Chip*, 2012, **12**(21), 4240–4243.
- 50 M. Sajid, A. N. Kawde and M. Daud, Designs, formats and applications of lateral flow assay: A literature review, *J. Saudi Chem. Soc.*, 2015, **19**(6), 689–705.
- 51 X. Li, D. R. Ballerini and W. Shen, A perspective on paper-based microfluidics: Current status and future trends, *Biomicrofluidics*, 2012, **6**(1), 011301.
- 52 K. Jones, FUSION 5: A New Platform for Lateral Flow Immunoassay Tests, *Lateral Flow Immunoassay*, 2009, pp. 115–129.
- 53 J. A. Adkins, E. Noviana and C. S. Henry, Development of a Quasi-Steady Flow Electrochemical Paper-Based Analytical Device, *Anal. Chem.*, 2016, **88**(21), 10639–10647.
- 54 D. M. Zhang and Q. J. Liu, Biosensors and bioelectronics on smartphone for portable biochemical detection, *Biosens. Bioelectron.*, 2016, **75**, 273–284.
- 55 T. K. Christopoulos and E. P. Diamandis, Theory of immunoassays, *Immunoassay*, 1996, 25–50.
- 56 *Scientific Criteria to Ensure Safe Food*, National Academies Press, Washington, DC, 2003, <https://www.ncbi.nlm.nih.gov/books/NBK221563/>.

