RSC Advances



REVIEW

View Article Online
View Journal | View Issue



Cite this: RSC Adv., 2019, 9, 40577

Iron oxide nanoparticles as multimodal imaging tools

Edouard Alphandéry Pabe

In medicine, obtaining simply a resolute and accurate image of an organ of interest is a real challenge. To achieve this, it has recently been proposed to use combined methods in which standard imaging (MRI, PAI, CT, PET/SPEC, USI, OI) is carried out in the presence of iron oxide nanoparticles, thus making it possible to image certain tissues/cells through the specific targeting of these nanoparticles, hence resulting in improved imaging contrast and resolution. Here, the advantages and drawbacks of these combined methods are presented as well as some of their recent medical applications.

Received 21st October 2019 Accepted 25th October 2019

DOI: 10.1039/c9ra08612a

rsc.li/rsc-advances

Introduction

Iron oxide nanoparticles (IONP) can be used to image organs/ cells, which capture or accumulate them, such as the liver, 1 spleen,2 lymph nodes,3,4 bone marrow, and those of the mononuclear phagocytic system, whether these organs/cells are tumorigenic or not.5,6 Other examples of IONP imaging applications include the detection of apoptosis,7 inflammation,8 angiography,9 ruptured atherosclerotic plaque,10 multiple sclerosis,5 integrity of the blood-brain barrier,11 and vasculature, e.g. coronary arteries12 (Fig. 1). The development of IONP for these applications stems from their advantages compared with non-nanoparticle based systems such as: (i) their longer residence/circulation time,13 (ii) their faculty to act as a contrast agent for several imaging methods simultaneously,14 (iii) their ability to specifically target an organ/tissue of interest via passive, active or magnetic targeting with an efficacy that varies depending on studies and leads to a percentage of IONPs that target the tumor relative to the quantity of injected IONP that is between $4 \times 10^{-4}\%$ and 7%, ¹⁵ and (iv) their use as 'theranostic' probes, where IONP therapeutic activity comes from localized ROS or heat production, exposure of IONP to various excitation sources, or conjugation of drugs to IONP.16 Furthermore, since IONP are already used on humans for therapeutic applications either to treat iron anemia disease or to carry out magnetic hyperthermia treatment of cancer, 15 one could easily foresee their clinical use for imaging applications.

Here, I review the use of IONP as imaging tools in various imaging methods, *i.e.* magnetic resonance imaging (MRI),

magnetic particle imaging (MPI), photo-acoustic imaging (PAI), computing tomography (CT), positron emission tomography (PET)/single photon emission computed tomography (SPECT), ultrasound imaging (USI), and optical imaging (OI). IONP appeal comes from the fact that they enable:

- Adjustment of IONP T_1/T_2 contrasting strength by tuning the properties of these nanoparticles (size, charge, assembly, surface) for MRI;¹⁷
- Use of a device generating a magnetic field that can both image and heat IONP through magnetic hyperthermia for MPI;¹⁸
- Improved imaging resolution using common diagnostic devices for most imaging methods;¹⁹
 - Localized functional imaging for PET/SPECT;20
- A broad spectrum of different types of detections, *e.g.* drug release and intracellular imaging, for optical imaging.²¹

The review presented here is broader in scope than previous ones since it covers more imaging techniques and is not restricted to IONP synthesized by a specific method.^{22–25}

INOP advantages/drawbacks and parameters influencing IONP their imaging power in these various imaging methods are summarized in Table 1 and Fig. 2.

Magnetic resonance imaging (MRI)

Since human body is mainly composed of water, specific water properties could be measured to image all parts of the organism. MRI was developed to measure proton relaxation times of water molecules following a two-steps excitation process in which a static magnetic field (B_0) first produces longitudinal magnetization, *i.e.* alignment of proton nuclear spins parallel to B_0 , and then a radiofrequency pulse yields transverse magnetization, *i.e.* alignment of proton nuclear spins perpendicular to B_0 . Upon removal of the radiofrequency pulse, the proton nuclear spins relax longitudinally and transversely with relaxation times of T_1 and T_2 , respectively. IONP can be used as contrast agents to improve the quality of MRI images by

^aParis Sorbonne Université, Muséum National d'Histoire Naturelle, UMR CNRS7590, IRD, Institut de Minéralogie, de Physique des Matériaux et deCosmochimie, IMPMC, 75005 Paris, France

^bNanobacterie SARL, 36 Boulevard Flandrin, 75116, Paris, France

^{*}Institute of Anatomy, UZH University of Zurich, Institute of Anatomy, Winterthurerstrasse 190, CH-8057, Zurich, Switzerland. E-mail: edouardalphandery@hotmail.com; Tel: +4133632697020

RSC Advances Review




Fig. 1 Various applications of IONP as imaging tools.

decreasing the values of T_1 and/or T_2 . IONP contrast agents (CA) can be divided between T_1 and T_2 CA that lead to bright and dark MRI images associated with the recovery of the longitudinal magnetization (positive contrast, T_1) or loss of transverse magnetization (negative contrast, T_2). The relaxivity of an assembly of IONP and water molecules (R_i) can be expressed as a function of the relaxivities of isolated water protons (R_{water}) and IONP magnetic moment (R_{IONP}), using the relation: $R_i = 1/2$ $T_i = R_{\text{water}} + C_{\text{IONP}} R_{\text{IONP}}$ (i = 1 or 2). This relation indicates how R_i can be maximized by using large values of R_{IONP} and C_{IONP} . Furthermore, on the one hand, the Solomon-Bloemberger-Morgan theory indicates that R_1 can be maximized by increasing IONP proton molecular thumbling time and decreasing proton residence lifetime. On the other hand, the outer-sphere diffusion model suggests that the effect of IONP on T_1 or T_2 relaxivities essentially depends on the stability or strength of IONP magnetic moment. According to this model, IONP with less stable magnetic moments and smaller sizes essentially increase T_1 relaxivities while IONP with more stable magnetic moments and larger sizes increase T_2 relaxivities. However, these theories are simplified. They don't take into

account the whole stet of parameters that can influence the values of T_1 and T_2 relaxivities, which include: 26,27

- Size of IONP, with small IONP acting as a positive contrast agent, *e.g.* r_1 increases from 3–30 mM⁻¹ s⁻¹ for IONP of ~20–65 nm to 66 mM⁻¹ s⁻¹ for IONP of ~5 nm,²⁸ and large IONP behaving as negative contrast agents, *e.g.* r_2 increases from 35–130 mM⁻¹ s⁻¹ for IONP of 4–5 nm to 218–385 mM⁻¹ s⁻¹ for IONP of 12–14 nm,²⁸
- Shape of IONP, which affects the stability of IONP magnetic moment and magnetization, leading "theoretically" to well-distributed magnetization for ellipsoids and to localization of magnetization in specific locations of IONP for other geometries, *e.g.* in the corners of IONP cubes, a parameter that seems to influence IONP contrasting power if water molecules are unable to reside where IONP magnetization is located.
- Crystallinity of IONP, which determines whether IONP affects T_1 and/or T_2 through: (i) IONP crystal phase, *e.g.* spinel/inverse spinel crystal phases of Fe₂O₃/Fe₃O₄ with ferrimagnetic behaviors enhance T_2 contrast while IONP with antiferromagnetic behaviors, low magnetic moments, and small sizes such as α -FeOOH mainly impact T_1 contrast, (ii) the presence of doping

Review **RSC Advances**

Table 1 Advantages and drawbacks of the various imaging techniques (PA, OI, CT, USI, PET/SPEC, MRI, MPI) used in combination with IONP

Imaging methods	Advantages	Drawbacks
MRI	Widely used (available clinically)	Negative contrast
MKI	High penetration depth	Artifacts preventing detailed anatomical imaging
	Rapid signal acquisition	No direct measurement of concentration
	Functional information using fMRI	Not so good resolution (25–100 μM)
	Large area covered with one scan (>2500 cm ²)	1100 30 ξουά Γεσοιατίοι (25 100 μινή)
MPI	No penetration depth limit	Slow signal acquisition
1411 1	Direct measurement of concentration	No widely used (not available clinically)
	No radiation	No functional information
	Good resolution (~1 mm)	140 failetional information
	Can combine imaging/therapeutic activity (MPH)	
	Large area covered with one scan (can be adjusted	
	by varying coil size)	
PAI	No radiation	Low penetration depth (∼5 cm)
IAI	Rapid signal acquisition	Not widely used (only prototypes available in clinic)
	Good resolution (~0.5 nm)	No direct measurement of IONP concentration
	Large area covered with one scan (\sim 15 cm ²)	No uncer measurement of form concentration
	Functional information	
СТ	Large area covered with one scan (>2500 cm ² ,	
O1	similar to MRI)	
	Rapid signal acquisition	Not so good resolution (5–200 μm)
	No penetration depth limit	Radiation
		No functional information
	Widely used (available clinically) Large area covered with one scan?	No direct measurement of low IONP concentration
	Large area covered with one scan:	Limited soft tissue resolution
PET/SPECT	Widely used (available clinically)	Radiation
PE1/SPEC1	Widely used (available clinically) Simultaneous anatomical/functional imaging	
	Can combine imaging/therapeutic activity	Not so good good resolution (2-10 mm) Not straightforward to bind a radiotreacer to IONF
	can combine imaging/therapeutic activity	and maintain the complex IONP-radiotrace active in
		vivo without its destruction
	No penetration depth limit	Not straightforward to acquire signal
		Not straightforward to acquire signal
	Can cover the whole body No radiation	
Sonography	Direct measurement of concentration in some cases	No direct detection (use of magnete metica
Sonography	Widely used (available clinically)	No direct detection (use of magneto-motive
		mechanism or additional contrast agent such as
	No radiation	micro-bubble) Not so good resolution (50–500 μM)
	Rapid signal acquisition	Limited penetration depth (\sim 5–20 cm) No functional information
	Large area covered with one scan (\sim 200 cm ²)	
Ontical mathed	No radiation	No direct measurement of IONP concentration Limited penetration depth (\sim 1–5 cm)
Optical method	Area covered depends on diameter/number of	
		Not so good resolution (1–5 mm)
	optical fibers used	Not widely used alinically
	Can combine therapeutic/imaging activity Functional information in some cases	Not widely used clinically Not straightforward to bind a fluorophore to IONP
	Functional information in some cases	and maintain the complex IONP-fluorophore
		luminescent in vivo
	Measurement of IOND correction in several	
All toobniques	Measurement of IONP concentration in some cases Can be combined with eachother	Not straightforward to acquire signal
All techniques	Can be combined with eachother	One technique does not gather all advantages
		IONP can be eliminated in blood (capture by macrophages/opsonization)

material in the crystal, e.g. doping XFe_2O_4 with $X = Mn^{2+}$, Fe^{2+} , Co^{2+} , or Ni^{2+} leads to the largest magnetic moment for $X = Mn^{2+}$ and a very significant effect on T_2 with $r_2 = 358 \text{ mM}^{-1} \text{ s}^{-1}$ (1.5 T), and most interestingly (iii) the formation of mixed crystalline structures such as $Gd_xFe_yO_z$, possibly enabling dual-mode T_1 – T_2 contrast agents.

• Surface of IONP, which can lead to an enhanced IONP contrasting ability when: (i) surface defects are minimized, (ii) the number of metallic atoms at IONP surface that are close to water molecules is maximized, (iii) a specific binding material covers the surface, (iv) oxygen atoms occupy surface vacancies in a well-adjusted manner, (iv) coating material and coating RSC Advances Review

Parameters needing to be optimized for using IONP as imaging tools




Fig. 2 Parameters that need to be optimized for using IONP as imaging tools in PA, OI, CT, USI, PET/SPEC, MRI, MPI.

thickness are chosen to optimize the interactions between IONP magnetic moments and protons of water molecules.

• Assembly of IONP, which can change IONP contrasting power by modifying the diffusion properties of either IONP or water molecules surrounding IONP or both, e.g. a decrease in diffusion of water molecules surrounding IONP assemblies was associated with an increase in T_2 .

While some of these parameters favor one type of contrast to the detriment of the other one, *e.g.* a large nanoparticle size could enhance negative contrast and decrease positive contrast, other ones such as a mixed nanoparticle structure containing both metallic (Iron) and Gd atoms can promote simultaneously T_1 and T_2 contrasting ability.

In addition, since IONP contrasting ability depends on IONP interaction with the excitation sources, *i.e.* a combination of magnetic field and radiofrequency, the parameters of the excitation sources, in particular the strength or homogeneity/inhomogeneity of the magnetic field applied or the temperature of measurement, are other important parameters that influence the values of T_1 and T_2 .

IONP, commercialized as negative or positive contrast agents, are listed in Table 2. They have been used for imaging: (i) liver/spleen using Ferumoxides/AMI-24/AMI-25/Endorem/Feridex or Ferucarbotran/SHU555A/Resovist/Cliavist, ²⁹ (ii) lymph node/bone marrow imaging using Ferumoxtran-10/AMI-227/Sinerem/Combidex or Ferucarbotran/SHU555C/Supravist, ³⁰ (iii) angiography using Feruglose/NC100250/Clariscan or Ferumoxytol/AMI-7228/Feraheme (231), (iv) delineation of the bowel from adjacent organs and tissues, using AMI-121/Lumi-rem/Gastromark, ³²

The commercialization of most IONP mentioned in Table 2 has stopped for the following reasons. First, IONP are often eliminated in blood where they are captured by macrophages, making the targeting of specific organs by IONP a difficult task. Second, IONP can result in the formation of artifacts/black holes, which can prevent the realization of a detailed image of anatomical/tissular structures. Third, a quantity/concentration of IONP can't be deduced from an MRI image since IONP only indirectly affect an MRI image. Fourth, most IONP are T_2 contrast agents, which are not favored by clinicians due to their darkening contrasting effect on an MRI image.

MPI (magnetic particle imaging)

Magnetic particle imaging (MPI)^{34,35} is a new appealing imaging technique, which relies on the application of two types magnetic fields. A first magnetic field gradient results in the formation of a first region where the strength of the magnetic field is sufficiently high to saturate IONP magnetic moments and align them parallel to the applied magnetic field, and a second region, also called field free point (FFP), where the magnetic field strength is zero, and IONP magnetic moments are randomly oriented. When IONP located in the FFP are exposed to a second oscillating magnetic field of strength 10-100 mT and frequency 10-100 KHz,36 their magnetization oscillates with time, leading to an electromagnetic signal, which is detected and further converted into an MPI image. As for MRI, the quality of MPI imaging depends on IONP properties, such as IONP size, leading to an increase in spatial resolution by a factor 4 using IONP of 20 nm compared with Resovist of 45-60 nm. Several applications of IONP in MPI have been suggested such as:

Open Access Article. Published on 06 December 2019. Downloaded on 12/5/2025 12:58:55 PM.

This article is licensed under a Creative Commons Attribution 3.0 Unported Licence.

Table 2 Properties of IONP as contrast agents in MRI, such as IONP blood half-life, values of IONP relaxivities r_1 and r_2 , recommended IONP clinical dose expressed in quantity of iron in IONP properties for imaging various parts of the organism with MRI

Properties as MRI con	Properties as MRI contrast agents of commercialized iron oxide nanoparticles	ed iron oxide nanopa	rticles				
Name/reference	Composition	Blood half life in patients	$\mathop{(\mathrm{mmol}^{-1}}_{2} \mathrm{s}^{-1})$	$r_2 \pmod{1}$	Type of contrast	Clinical dose	Application/ administration
Ferumoxides (AMI-24 AMI-25), endorem. feridex	SPION (120–180 nm) + dextran coating (ref. 66)	10 min (ref. 65)	24 (ref. 68 and 73)	100–160 (ref. 72 and 73)	Negative	30 µmol Fe kg ⁻¹ (ref. 66)	Liver/spleen I (ref. 69)
Ferucarbotran SHU555A, Resovist/ Cliavist	SPION (60–80 nm) + carboxy-dextran coating (ref. 66)	12 min (ref. 65)	25 (ref. 68 and 73)	164–177 (ref. 72 and 73)	Negative	8–12 μmol Fe kg ^{–1} (ref. 66)	Liver/spleen IV (ref. 68 and 69)
Ferumoxtran-10 AMI-227, Sinerem Combidex	USPIO (30 nm) + dextran coating (ref. 66)	24-30 h (ref. 65)	22 (ref. 68 and 73)	44–85 (ref. 72 and 73)	Positive or negative	45 μ mol Fe kg^{-1} (ref. 66)	Lymph node bone marrow IV (ref. 68 and 69)
Ferucarbotran SHU555C, Supravist	USPIO (20–25 nm) + carboxy-dextran coating (ref. 66)	6–8 h (ref. 65)	7 (ref. 68)	57 (ref. 68)	Positive	40 μ mol Fe kg ⁻¹ (ref. 66)	Perfusion lymph node bone marrow IV (ref. 68 and 69)
Feruglose NC100250, Clariscan	USPIO (10–20 nm) + carbohydrate-polyethylene glycol (ref. 66)	2 h (ref. 65)	20 (ref. 68)	35 (ref. 68)	Positive	36 µmol Fe kg ⁻¹ (ref. 66)	Perfusion angiography IV (ref. 68 and 69)
Ferumoxytol AMI-7228, Feraheme	USPIO (3 nm) + carboxy- methyldextran (ref. 67)	14 h (ref. 65)	15 (ref. 71)	89 (ref. 71)	Positive	50–400 mg per patient (ref. 71)	Angiography IV (ref. 68 and 70)
AMI-121 Lumirem and Gastromark	SPION (300 nm) + silica coating (ref. 74)	NA	3 (ref. 68)	72 (ref. 68)	Negative	105 mg per patient (ref. 74)	GI oral (ref. 74, 69 and 31)

RSC Advances Review

- Cell concentration measurement, i.e. it is possible to obtain a relation between the MPI signal and the cellular concentration,37 with a high resolution, down to ~250 cells and 7.8 ng of Fe. It however necessitates knowing on the one hand the quantity of IONP internalized in each cell and on the other hand how IONP internalization affects MPI signal. Whereas the quantity of IONP nanoparticles internalized in each cell can easily be determined in vitro, it seems more difficult to evaluate it in vivo without extracting a tissue sample from the organism for analysis, restricting the use of this method. In addition, whereas the MPI signal of some IONP is relatively independent of their environment and of the fact that they are internalized or not, due to a minor Brownian contribution to their MPI signal, other IONP display a MPI signal that strongly depends on IONP environment,38 making the judicious choice of IONP crucial for accurate MPI measurement.
- Imaging of several different types of IONP, since IONP with different sizes yield different MPI signals. This can enable the detection of various biological entities if each IONP of different size targets a different entity.³⁹
- Measurement of blood flux, which can on the one hand enable evaluating the displacement speed of IONP in blood by MPI and on the other hand certain specific blood trajectory such as that resulting from bleeding, *i.e.* IONP leakage outside of blood vessels.⁴⁰
- Imaging of some organs or tissues (tumors), which are either targeted by IONP following different targeting methods, *i.e.* passive through the EPR effect, magnetic *via* the application of a magnetic field gradient, or active using an antibody bound to IONP that targets a specific cell receptor, or directly injected with IONP, yielding imaging of organ, *e.g.* lung⁴¹ or brain,⁴² or vasculature,³⁷ imaging without facing the same hurdle as in MRI in which IONP can leads to a hypo-intense signal that is similar to that of air inside lung.
- MPI combined with MHT, which can be foreseen since the properties of the applied oscillating magnetic field (strength and frequency) are similar in MPI and MHT (magnetic hyperthermia), possibly enabling combining imaging and therapeutic methods with IONP.¹⁸

MPI present a series of advantages such as: (i) a high penetration depth due to a theoretical unlimited propagation of the magnetic field, (ii) a direct measurement of IONP concentration, (iii) a positive contrast, (iv) an operating mechanism that does not rely on the use of toxic radiations, (v) a signal that should not be strongly affected by background tissue, (vi) a signal that can be acquired quickly in a spatially selective manner, (vii) a high resolution (sub-mm). However, MPI also suffers from a series of drawbacks such as: (i) the absence of anatomic imaging (MPI only images IONP), (ii) possible formation of eddy currents that can yield overheating if the strength/frequency of the oscillating magnetic field is too high, (iii) the necessity to use IONP whose magnetic moment couples with the external oscillating magnetic field without depending on IONP environment, (iv) the absence of a MPI equipment that can be used in the clinic.

Photo-acoustic imaging (PAI)

The mechanism of photo-acoustic imaging relies on the excitation of a tissue with a nanosecond pulsed laser of 680-970 nm emission wavelength, 10 Hz frequency, and 10-100 mJ cm⁻² power, resulting in a slight temperature increase and thermoelastic expansion of the tissue, which further generates an acoustic wave that provides an image of tissue following detection and processing of the signal by an echograph of typical ultrasound frequency 5-50 MHz. 43-45 Chromophores such as melanin and hemoglobin that are strong optical absorbers can act as intrinsic PAI contrast agents. However, in some cases they are not available, e.g. in some cancer cells, then requiring the use of other extrinsic PAI contrast agents such as IONP. IONP can behave as PAI contrast agents without the presence of another substance. In this case, the coupling mechanism between the incident laser light and IONP can involve surface plasmon waves.46 IONP can also be conjugated to a fluorophore such as Indocyanine Green,21 which enhance laser light absorption. An important number of applications of PAI has been described, 45 such as the imaging of certain cancers (breast, prostate, bladder, melanoma, ovarian), lymph nodes possibly containing metastasis, circulating tumor cells, as well as neonatal brain, gastrointestinal, thyroid, or intra-operative imaging. PAI displays a number of advantages such as: (i) an operating mechanism that does not use radiation, (ii) fast imaging acquisition, i.e. typically less than 1 minute,47 (iii) a relatively good resolution, e.g. \sim 0.5 mm was reached for breast imaging,48 which can be increased by increasing the ultrasound frequency, (iv) the relatively large area that can be covered with one scan, e.g. a surface of 10-50 cm² of a breast tumor can be imaged,48 (v) the possibility to obtain functional information derived from a variation in hemoglobin or melanin concentration, (vi) the ability to combine PAI with other imaging techniques such as MRI or CT. However, this technique also suffers from a relatively low penetration depth (typically of the order of a few cm), which is due to the use of laser light and to the ultrasound frequency when it is too high, and to a too small number of PAI equipment that can be used in the clinic. As an example, the Twente Photoacoustic Mammoscope (PAM) was built at the University of Twente (Enschede, the Netherlands) to image breast tumor cells in the clinic, but this type of equipment remains a prototype.49 It needs to be more widely developed and made available in different hospitals to spread out the use of the PAI imaging technique.

Scanner, computing tomography (CT)

X-ray scanners, also designated as computing tomography (CT), expose patients to an X-ray beam, resulting in a non-uniform absorption of X-rays by the patients, *i.e.* the absorption is different parts for various parts of the organism depending on their consistency (*Z*-number). Then the transmitted (non-absorbed) X-rays are detected, providing an image of these different parts with various contrasts. In fact, the interaction between X-rays and the organism is due to the photo-electric effect, whose strength is mainly proportional to Z^3 (*Z*: atomic number), and Compton scattering, which is enhanced at high electron and

Review **RSC Advances**

mass density.50 The unit of attenuation measurement obtained in CT is Hounsfield unit (HU), leading to low CT number (HU value) for weakly absorbing material (-1000 HU for air and 0 HU for water), to intermediate HU values for slightly absorbing material (40 to 80 HU for tissue), and to larger HU values for stronger absorbing material (400 to 1000 HU for bone).51 Substances with high HU values (strongly attenuating X-ray) appear white or light gray, while those with low HU values (weakly attenuating X-ray) appear dark gray or black. The photoelectric and Compton effects suggest that material with high Z and large mass density would be the best contrast agent. Although the Z value of iron, i.e. 26, and the density of maghemite comprised in most IONP for medical application, i.e. 5 g cm⁻³, are not the largest values that can be reached with nanomaterials, and are larger than those of water (Z = 1 for hydrogen, and Z = 8 for oxygen, r = 1 g cm⁻³ for water) that constitutes the majority of organic living material, they seem sufficient to provide. They seem are sufficient to provide a contrast with CT. Indeed, IONP were shown to act as CT contrast agent, leading to CT number that increases with increasing IONP concentration, i.e. from 27 HU at 2.5 mg mL⁻¹ of IONP to 113 HU at 25 mg mL⁻¹ of IONP.⁵² However, compared with iodine, IONP were shown to absorb \sim 5-6 times less X-ray. The advantages of CT lie in: (i) the acquisition of large scans (typically 10-20 cm) quickly (under 5 seconds) with image reconstruction within less than one minute, (ii) its non-invasiveness, (iii) high contrast resolution, (iv) no depth penetration limit, (v) a relatively low cost, (vi) quantitative information on contrast agents can be obtained, (vii) its wide availability in hospitals. Disadvantages of CT come from: (i) its low sensitivity compared with other imaging techniques, *i.e.* CT detection limit is $\sim 10^{-3}$ M, 51 compared with 10^{-5} M for Gd chelates in MRI and 10^{-9} M for nuclear techniques about, 51 (ii) the exposure of patient to relatively high dose of ionizing radiation, (iii) limited soft tissue resolution.

Positron emission tomography (PET)/single photon emission computed tomography (SPECT)

The mechanism of PET and SPECT relies on the use of a radiotracer, which targets a specific part of the organism such as a tumor, then emits a signal made of positrons (PET) or gamma rays (SPECT) that is measured, hence enabling the detection of this part of the organism. PET and SPECT are functional imaging methods mimicking a substance of interest such as glucose for 18F-FDG, which gets fixed on a specific tissue, e.g. a tumor in case of ¹⁸F-FDG that consumes more glucose than a healthy tissue.⁵³ Using IONP, it was possible to bring several improvement to the standard PET/SPECT imaging method by: (i) binding different radio-tracers to IONP [18F]fluorodeoxyglucose (FDG), copper-61/64, gallium-66/68, zirconium-89, and iodine-124 for PET,54 and 99mTc, 125I, 111I, 125I and 131I for SPECT, 54-56 that increase radio-tracer lifetime and targeting efficacy,55 (ii) enabling simultaneous anatomical and functional imaging by combining PET/SPEC with MRI, taking advantage of the MRI contrasting ability of IONP, (iii) enlarging the SPECT/ PET imaging capacity to a therapeutic activity through the use of a theranostic IONP probe that can trigger drugs delivery, immunotherapy, hyperthermia, or photodynamic therapy. 19,57

In addition, PET/SPECT lead to high detection sensitivity, e.g. PET was shown to be 200 times more sensitive than MRI using nanoparticles, the possibility to image the whole organism without limitation in depth of penetration, and the acquisition of quantitative information since in optimal operating conditions PET/SPECT signal should be proportional to the number/ concentration of radiotracer in the targeted region.54 For these reasons, the development of PET/SPECT can be foreseen in various medical fields such as disease detection or the assessment of the efficacy of a medical treatment.

Ultrasound imaging/sonography

It was reported that IONP could be used in ultrasound (sonography) imaging. On the one hand, IONP could be inserted within a material that is already an ultrasound contrast agent such as a microbubble or liposomes, then enabling a combination between USI and another imaging modality such as MRI that is made possible by the presence of IONP.58 On the other hand, although a direct detection of IONP by ultrasound was not reported to the author knowledge, IONP presence could be detected indirectly by measuring tissue resistance resulting from IONP motion/vibration under the application of a high intensity pulsed magnetic field through a technique cold magneto-motive ultrasound whose efficacy essentially depends on IONP susceptibility and could be combined with ultrasound imaging to provide both anatomical tissue information and estimate the quantity of IONP or photoacoustic imaging to improve the resolution of PAI imaging. Magneto-motive approaches could be used to carry out dynamic imaging, i.e. imaging of moving biological systems, such as circulating tumor cells, metastases diffusing in/out of lymph nodes or stem cells, to determine viscoelastic property of soft tissues, and could easily be combined with other imaging modalities.59

Optical imaging methods

Optical imaging can be carried out with IONP-fluorophore complexes, essentially using NIR fluorophores such as organic fluorochromes, e.g. cyanine dyes such as Cy5.5. Such complexes are either made of IONP surrounded by various coatings such as silica shell, spacer, lipid bilayer, polymer, which contain or are labeled with fluorophores or of IONP directly coated with a fluorescent semiconducting material. Such methods present the advantages of being relatively easy to implement, of yielding under optimized conditions a high spatial/temporal resolution, and of being complementary to other imaging techniques such as MRI. In order for this method to be efficient, absorption by the organism/tissue/water/hemoglobin/melanin/proteins that occurs between 200 and 650 nm shall be avoided by choosing fluorescent probes with absorption/emission within the range of 650-1450 nm. Furthermore, fluorophores should not be prone to photo-bleaching, not be quenched by iron oxide, be stable in biological media, i.e. have a long life-time in such environment, and display well-separated emission and absorption spectra to be to distinguishable them. The use of this method is facilitated by the existence of numerous systems that can excite/detect fluorescence in different conditions, i.e. in real time, in vitro, in vivo, such as: optical scanners, e.g.

RSC Advances Review

fluorescence mediated tomography, fluorescence reflectance tomography, optical coherence tomography, a large number of fluorescence microscopies, flow cytometry, spectrophotometry, intra-vital microscopy, intravascular non-invasive near-infrared (NIRF) imaging, clinical endoscopy, and equipment for fluorescence detection during surgery. Such method can be used for monitoring magnetofection efficacy,60 for multi-modal imaging with MPI, MRI and PAI,61,62 for detecting various biological entities such as tumors, 16,56 apoptotic cells,7 and sentinel lymph nodes,4 and for delineating infiltrating tumors such as glioblastoma.63

Multimodal imaging

Since IONP can be used with various imaging methods, they could be considered for multimodal imaging. The latter presents the advantage of enabling the combination of information coming from different imaging methods.⁶⁴ For example, in PET/ MRI, MRI is used for anatomical imaging while PET offers molecular information.

Conclusion/perspectives

From the author's point of view, the challenges ahead lie in:

- Improving both IONP (better targeting, size control, fabrication process) and imaging methods (better resolution, miniaturization, lower cost, more limited use of hospital infrastructure);
- Yielding a sufficiently large percentage of IONP in the organ of interest so that IONP can act as contrast agents;
- Developing a single imaging method that brings together the maximum benefit of the different imaging techniques with the minimum of their drawbacks;
- Building a device/platform that gathers all the different imaging techniques in one unit;
- Identifying the specific medical need upstream and then providing the imaging method best suited to this need, an approach that seems appropriate when medical needs remain globally unchanged, which may be the case within a limited time period in specialized hospital units.
- IONP present the advantage of being compatible and even improving a large series of different imaging techniques, making them appealing for multimodal imaging.64

Abbreviations

OI

PAI

BBB Blood brain barrier CAContrast agent CTComputing tomography **IONP** Iron oxide nanoparticles MRI Magnetic resonance imaging MPI Magnetic particle imaging MPH/MHT Magnetic particle hyperthermia MPS Mononuclear phagocytic system MS Multiple sclerosis

Optical imaging

Photo-acoustic imaging

ROS Radical oxygen species

US Ultrasounds

Ultrasound imaging (sonography) USI

Conflicts of interest

EA has been working in the company Nanobacterie.

Acknowledgements

I would like to thank the BPI ('banque publique d'investissement, France'), the region of Paris ('Parris Région Entreprise, France'), the French Research Tax Credit program ('crédit d'impôt recherche'), the incubator Paris Biotech Santé, the ANRT (CIFRE 2014/0359, CIFRE 2016/0747, CIFRE 2013/0364, CIFRE 2015/976), the Eurostars programs (Nanoneck-2 E9309 and Nanoglioma E11778), the AIR program ('aide à l'innovation responsable') from the region of Paris (A1401025Q), the ANR ('Agence Nationale de la Recherche') Méfisto, as well as the Universities Paris 6 and Paris 11. We also would like to thank the Nomis Foundation and Markus Reinhard for their support.

References

- 1 J. Lu, J. Sun, F. Li, J. Wang, D. Kim, C. Fan, T. Hyeon and D. Ling, Highly Sensitive Diagnosis of Small Hepatocellular Carcinoma Using pH-Responsive Iron Oxide Nanocluster Assemblies, J. Am. Chem. Soc., 2018, 140, 10071-10074.
- 2 K. El-Boubbou, Magnetic iron oxide nanoparticles as drug carriers: preparation conjugation and delivery, Nanomedicine, 2018, 13, 929-952.
- 3 R. M. Clauson, M. Chen, L. M. Scheetz, B. Berg and B. Chertok, Size-Controlled Iron Oxide Nanoplatforms with Lipidoid-Stabilized Shells for Efficient Magnetic Resonance Imaging-Trackable Lymph Node Targeting and High-Capacity Biomolecule Display, ACS Appl. Mater. Interfaces, 2018, 10, 20281-20295.
- 4 S. Mehralivand, H. Van der Poel, A. Winter, P. L. Choyke, P. A. Pinto and B. Turkbey, Sentinel lymph node imaging in urologic oncology, Transl. Androl. Urol., 2018, 7, 887-902.
- 5 S. M. Dadfar, K. Roemhild, N. I. Drude, S. Von Stillfried, R. Knüchel, F. Kiessling and T. Lammers, Iron oxide nanoparticles: Diagnostic, therapeutic and theranostic applications, Adv. Drug Delivery Rev., 2019, 138, 302-325.
- 6 Y. Hu, S. Mignani, J.-P. Majoral, M. Shen and W. Shi, Construction of iron oxide nanoparticle-based hybrid platforms for tumor imaging and therapy, Chem. Soc. Rev., 2018, 47, 1874-1900.
- 7 M. M. Mekawya, A. Saito, A. Sumiyoshic, J. J. Riera, H. Shimizud, R. Kawashima and T. Tominagaa, Hybrid magneto-fluorescent nano-probe for live apoptotic cells monitoring at brain cerebral ischemia, Mater. Sci. Eng., C, 2019, 100, 485-492.
- 8 S. Hedgire, C. Krebill, G. R. Wojtkiewicz, I. Oliveira, B. B. Ghoshhajra, U. Hoffmann and M. G. Harisinghani, Ultrasmall superparamagnetic iron oxide nanoparticle

Review

uptake as noninvasive marker of aortic wall inflammation on MRI: proof of concept study, *Br. J. Radiol.*, 2018, **1092**, 20180461.

- 9 Y. Cai, Y. Wang, H. Xu, C. Cao, R. Zhu, X. Tang, T. Zhang and Y. Pan, Positive magnetic resonance angiography using ultrafine ferritin-based iron oxide nanoparticles, *Nanoscale*, 2019, 11, 2644–2654.
- 10 J. M. S. Chan, C. Monaco, M. Wylezinska-Arridge, J. L. Tremoleda, J. E. Cole, M. Goddard, M. S. H. Cheung, K. K. Bhakoo and R. G. J. Gibbs, Imaging vulnerable plaques by targeting inflammation in atherosclerosis using fluorescent-labeled dual-ligand microparticles of iron oxide and magnetic resonance imaging, *J. Vasc. Surg.*, 2018, 67, 1571–1548e3.
- 11 A. Ivask, E. H. Pilkington, T. Blin, A. Käkinen, H. Vija, M. Visnapuu, J. F. Quinn, M. R. Whittaker, R. Qiao, T. P. Davis, P. C. Ke and N. H. Voelcker, Uptake and transcytosis of functionalized superparamagnetic iron oxide nanoparticles in an *in vitro* blood brain barrier model, *Biomater. Sci.*, 2018, 6, 314–323.
- 12 E. D. Lehrman, A. N. Plotnik, T. Hope and D. Saloner, Ferumoxytol-enhanced MRI in the peripheral vasculature, *Clin. Radiol.*, 2019, 74, 37–50.
- 13 H. Arami, A. P. Khandhar, A. Tomitaka, E. Yu, P. W. Goodwill, S. M. Conolly and K. M. Krishnan, In vivo multimodal magnetic particle imaging (MPI) with tailored magneto/optical contrast agents, *Biomaterials*, 2015, 52, 251–261.
- 14 T. H. Shin, Y. Choi, S. Kim and J. C. Shin, Recent advances in magnetic nanoparticle-based multi-modal imaging, *Chem. Soc. Rev.*, 2015, 44, 4501–4516.
- 15 E. Alphandéry, Biodistribution and targeting properties of iron oxide nanoparticles for treatments of cancer and iron anemia disease, *Nanotoxicology*, 2019, **13**, 573–596.
- 16 C. Song, W. Sun, Y. Xiao and X. Shi, Ultrasmall iron oxide nanoparticles: synthesis, surface modification, assembly, and biomedical applications, *Drug Discovery Today*, 2019, **24**, 835–844.
- 17 D. Bonvin, D. T. L. Alexander, A. Millán, R. Piñol, B. Sanz, G. F. Goya, A. Martínez, J. A. M. Bastiaansen, M. Stuber, K. J. Schenk, H. Hofmann and M. M. Ebersold, Tuning Properties of Iron Oxide Nanoparticles in Aqueous Synthesis without Ligands to Improve MRI Relaxivity and SAR, Nanomaterials, 2017, 7, 225.
- 18 Z. W. Tay, P. Chandrasekharan, A. Chiu-Lam, D. W. Hensley, R. Dhavalikar, X. Y. Zhou, E. Y. Yu, P. W. Goodwill, B. Zheng, C. Rinaldi and S. M. Conolly, Magnetic Particle Imaging-Guided Heating *in vivo* Using Gradient Fields for Arbitrary Localization of Magnetic Hyperthermia Therapy, *ACS Nano*, 2018, 12, 3699–3713.
- 19 N. V. S. Vallabani and S. Singh, Recent advances and future prospects of iron oxide nanoparticles in biomedicine and diagnostics, *3 Biotech*, 2018, **8**, 279.
- 20 H. Yuan, M. Q. Wilks, M. D. Normandin, G. E. Fakhri, C. Kaittanis and L. Josephson, Heat-induced radiolabeling and fluorescence labeling of Feraheme nanoparticles for

- PET/SPECT imaging and flow cytometry, *Nat. Protoc.*, 2018, 13, 392-412.
- 21 H. Wang, X. Li, B. W. C. Tse, Y. Yang, C. A. Thorling, Y. Liu, M. Touraud, J. B. Chouane, X. Liu, M. S. Roberts and X. Liang, Indocyanine green-incorporating nanoparticles for cancer theranostics, *Theranostics*, 2018, 8, 1227–1242.
- 22 Z. Shen, A. Wu and X. Chen, Iron Oxide Nanoparticle Based Contrast Agents for Magnetic Resonance Imaging, *Mol. Pharmaceutics*, 2017, 14, 1352–1364.
- 23 I. Fernández-Barahona, M. Muñoz-Hernando and F. Herranz, Microwave-Driven Synthesis of Iron-Oxide Nanoparticles for Molecular Imaging, *Molecules*, 2019, 24, 1224, DOI: 10.3390/molecules24071224.
- 24 J. Dulinska-Litewka, A. Lazarczyk, P. Hałubiec, O. Szafranski, K. Karnas and A. Karewicz, Superparamagnetic Iron Oxide Nanoparticles—Current and Prospective Medical Applications, *Materials*, 2019, 12, 617, DOI: 10.3390/ma12040617.
- 25 C. Song, W. Sun, Y. Xiao and X. Shi, Ultrasmall iron oxide nanoparticles: synthesis, surface modification, assembly, and biomedical applications, *Drug Discovery Today*, 2019, **24**, 835–844.
- 26 W. Zhang, L. Liu, H. Chen, K. Hu, I. Delahunty, S. Gao and J. Xie, Surface impact on nanoparticle-based magnetic resonance imaging contrast agents, *Theranostics*, 2018, 8, 2521–2548.
- 27 Z. Zhou, L. Yang, J. Gao and X. Chen, Structure-Relaxivity Relationships of Magnetic Nanoparticles for Magnetic Resonance Imaging, Adv. Mater., 2019, 31, 1804567.
- 28 C. Blanco-Andujar, A. Walter, G. Cotin, C. Bordeianu, D. Mertz, D. Felder-Flesch and S. Begin-Colin, Design of iron oxide-based nanoparticles for MRI and magnetic hyperthermia, *Nanomedicine*, 2019, 11, 1889–1910.
- 29 Z. R. Stephen, F. M. Kievit and M. Zhang, Magnetite Nanoparticles for Medical MR Imaging, *Mater. Today*, 2011, 14, 330–338.
- 30 R. Dinniwell, P. Chan, G. Czarnota, *et al.*, Pelvic lymph node topography for radiotherapy treatment planning from ferumoxtran-10 contrast-enhanced magnetic resonance imaging, *Int. J. Radiat. Oncol., Biol., Phys.*, 2009, 74, 844–851.
- 31 S. Stoumpos, M. Hennessy, A. T. Vesey, A. Radjenovic, R. Kasthuri, D. B. Kingsmore, P. B. Mark and G. Roditi, Ferumoxytol magnetic resonance angiography: a dose-finding study in patients with chronic kidney disease, *Eur. J. Radiol.*, 2019, **29**, 3543–3552.
- 32 Y. Xiáng, J. Wáng and J.-M. Idée, A comprehensive literatures update of clinical researches of superparamagnetic resonance iron oxide nanoparticles for magnetic resonance imaging, *Quant. Imaging Med. Surg.*, 2017, 7, 88–122.
- 33 S. Tong, H. Zhu and G. Bao, Magnetic iron oxide nanoparticles for disease detection and therapy, *Mater. Today*, 2019, DOI: 10.1016/j.mattod.2019.06.003.
- 34 J. W. M. Bulte, Superparamagnetic iron oxides as MPI tracers: A primer and review of early applications, *Adv. Drug Delivery Rev.*, 2019, **138**, 293–301.
- 35 N. Panagiotopoulos, R. L. Duschka, M. Ahlborg, G. Bringout, C. Debbeler, M. Graeser, M. Kaethner, K. Lüdtke-Buzug,

RSC Advances

H. Medimagh, J. Stelzner, T. M. Buzug, J. Barkhausen, F. M. Vogt and J. Haegele, Magnetic particle imaging: current developments and future directions, *Int. J. Nanomed.*, 2015, **10**, 3097–3114.

- 36 A. Meola, J. Rao, N. Chaudhary, G. Song, X. Zheng and S. D. Chang, Magnetic Particle Imaging in Neurosurgery, 2019, 125, 261–270.
- 37 X. Y. Zhou, Z. W. Tay, P. Chandrasekharan, E. Y. Yu, D. W. Hensley, R. Orendorff, K. E. Jeffris, D. Mai, B. Zheng, P. W. Goodwill and S. M. Conolly, Magnetic particle imaging for radiation-free, sensitive and high-contrast vascular imaging and cell tracking, *Curr. Opin. Chem. Biol.*, 2018, 45, 131–138.
- 38 H. Arami, R. M. Ferguson, A. P. Khandhar and K. M. Krishnan, Size-dependent ferrohydrodynamic relaxometry of magnetic particle imaging tracers in different environments, *Med. Phys.*, 2013, **40**, 071904.
- 39 H. Kratz, M. Taupitz, A. A. de Schellenberger, O. Kosch, D. Eberbeck, S. Wagner, L. Trahms, B. Hamm and J. Schnorr, Novel magnetic multicore nanoparticles designed for MPI and other biomedical applications: from synthesis to first *in vivo* studies, *PLoS One*, 2018, 13, e0190214.
- 40 M. G. Kaul, J. Salamon, T. Knopp, H. Ittrich, G. Adam, H. Weller and C. J. Wegner, Magnetic particle imaging for in vivo blood flow velocity measurements in mice, *Phys. Med. Biol.*, 2018, 63, 064001.
- 41 F. Wegner, T. M. Buzug and J. Barkhausen, Take a Deep Breath–Monitoring of Inhaled Nanoparticles with Magnetic Particle Imaging, *Theranostics*, 2018, **8**, 3691–3692.
- 42 L. C. Wu, Y. Zhang, G. Steinberg, H. Qu, S. Huang, M. Cheng, T. Bliss, F. Du, J. Rao, G. Song, L. Pisani, T. Doyle, S. Conolly, K. Krishnan, G. Grant and M. Wintermark, A Review of Magnetic Particle Imaging and Perspectives on Neuroimaging, Am. J. Neuroradiol., 2019, 40, 206–212.
- 43 P. Armanetti, A. Flori, C. Avigoa, L. Conti, B. Valtancoli, D. Petroni, S. Doumett, L. Cappiello, C. Ravagli, G. Baldi, A. Bencini and L. Menichetti, Spectroscopic and photoacoustic characterization of encapsulated iron oxide super-paramagnetic nanoparticles as a new multiplatform contrast agent, Spectrochim. Acta, Part A, 2018, 199, 248–253.
- 44 J. P. Thawani, A. Amirshaghaghi, L. Yan, J. M. Stein, J. Liu and A. Tsourkas, Photoacoustic-Guided Surgery with Indocyanine Green-Coated Superparamagnetic Iron Oxide Nanoparticle Clusters, Small, 2017, 13, 1701300.
- 45 S. Zackrisson, S. M. W. Y. van de Ven and S. S. Gambhir, Light In and Sound Out: Emerging Translational Strategies for Photoacoustic Imaging, *Cancer Res.*, 2014, 74, 979–1004.
- 46 H. Chen, Z. Yuan and C. Wu, Nanoparticle Probes for Structural and Functional Photoacoustic Molecular Tomography, *BioMed Res. Int.*, 2015, 2015, 757101.
- 47 P. Beard, Biomedical photoacoustic imaging, *Interface Focus*, 2011, **1**, 602–631.
- 48 S. Manohar and M. Dantuma, Current and Future Trends in Photoacoustic Breast Imaging, *Photoacoustics*, 2019, DOI: 10.1016/j.pacs.2019.04.004.

- 49 K. S. Valluru, K. E. Wilson and J. K. Willmann, Photoacoustic Imaging in Oncology: Translational Preclinical and Early Clinical Experience, *Radiology*, 2016, 280, 332–349.
- 50 J. Kim, P. Chhour, J. Hsu, H. I. Litt, V. A. Ferrari, R. Popovtzer and D. P. Cormode, Use of Nanoparticle Contrast Agents for Cell Tracking with Computed Tomography, *Bioconjugate Chem.*, 2017, 28, 1581–1597.
- 51 D. P. Cormode, P. C. Naha and Z. A. Fayad, Nanoparticle Contrast Agents for Computed Tomography: A Focus on Micelles, *Contrast Media Mol. Imaging*, 2014, **9**, 37–52.
- 52 S. Goel, C. G. England, F. Chen and W. Cai, Positron Emission Tomography and Nanotechnology: A Dynamic Duo for Cancer Theranostics, *Adv. Drug Delivery Rev.*, 2017, 113, 157–176.
- 53 A. L. Dias, R. Kunzel, R. S. Levenhagen and E. Okuno, Application of computed tomography images in the evaluation of magnetic nanoparticles biodistribution, *J. Magn. Magn. Mater.*, 2010, 322, 2405–2407.
- 54 F. Ai, A. Ferreira, F. Chen and W. Cai, Engineering of Radiolabeled Iron Oxide Nanoparticles for Dual-Modality Imaging, *Wiley Interdiscip. Rev.: Nanomed. Nanobiotechnol.*, 2016, **8**, 619–630.
- 55 J. Pellico, J. Llop, I. Fernández-Barahona, R. Bhavesh, J. Ruiz-Cabello and F. Herranz, Iron Oxide Nanoradiomaterials: Combining Nanoscale Properties with Radioisotopes for Enhanced Molecular Imaging, Contrast Media Mol. Imaging, 2017, 2017, 1549580.
- 56 M. De Simone, D. Panetta, E. Bramanti, C. Giordano, M. C. Salvatici, L. Gherardinia, A. Menciassie, S. Burchiellif, C. Cintia and P. A. Salvadoria, Magnetically driven nanoparticles: 18FDGradiolabelling and positron emission tomography biodistribution study, *Contrast Media Mol. Imaging*, 2016, 11, 561–571.
- 57 R. A. Revia and M. Zhang, Magnetite nanoparticles for cancer diagnosis, treatment, and treatment monitoring: recent advances, *Mater. Today*, 2016, **19**, 157–158.
- 58 X. Qian, X. Han and Y. Chen, Insights into the unique functionality of inorganic micro/nanoparticles for versatile ultrasound theranostics, *Biomaterials*, 2017, **142**, 13–30.
- 59 M. Evertsson, P. Kjellman, M. Cinthio, R. Andersson, T. A. Tran, R. Zandt, G. Grafström, H. Toftevall, S. Fredriksson, C. Ingvar, S. E. Strand and T. Jansson, Combined Magnetomotive ultrasound, PET/CT, and MR imaging of 68Ga-labelled superparamagnetic iron oxide nanoparticles in rat sentinel lymph nodes in vivo, Sci. Rep., 2017, 7, 4824.
- 60 M. Zuvin, E. Kuruoglu, V. O. Kaya, O. Unal, O. Kutlu, H. Y. Acar, D. Gozuacik and A. Koşar, Magnetofection of Green Fluorescent Protein Encoding DNA Bearing Polyethyleneimine-Coated Superparamagnetic Iron Oxide Nanoparticles to Human Breast Cancer Cells, ACS Omega, 2019, 4, 12366–12374.
- 61 L. Meng, X. Ma, S. Jiang, G. Ji, W. Han, B. Xu, J. Tian and W. Tian, High-efficiency fluorescent and magnetic multimodal probe for long-term monitoring and deep penetration imaging of tumors, *J. Mater. Chem. B*, 2019, 7, 5345–5351.

Review **RSC Advances**

- 62 C. Song, W. Sun, Y. Xiao and X. Shi, Ultrasmall iron oxide nanoparticles: synthesis, surface modification, assembly, and biomedical applications, Drug Discovery Today, 2019, 24, 835-844.
- 63 C. Lee, G. R. Kim, J. Yoon, S. E. Kim, J. S. Yoo and Y. Piao, In vivo delineation of glioblastoma by targeting tumorassociated macrophages with near-infrared fluorescent silica coated iron oxide nanoparticles in orthotopic xenografts for surgical guidance, Sci. Rep., 2018, 8, 11122.
- 64 J. Lemaster, F. Chen, T. Kim, A. Hariri and J. V. Jokerst, Development of a Trimodal Contrast Agent for Acoustic and Magnetic Particle Imaging of Stem Cells, ACS Appl. Nano Mater., 2018, 1, 1321-1331.
- 65 H. E. Daldrup-Link, Ten Things You Might Not Know about Iron Oxide Nanoparticles, Radiology, 2017, 284, 616-629.
- 66 J. S. Weinstein, C. G. Varallyay, E. Dosa, S. Gahramanov, В. Hamilton, W. D. Rooney, L. L. Muldoon and Ε. A. Neuwelt, Superparamagnetic iron oxide nanoparticles: diagnostic magnetic resonance imaging and potential therapeutic applications in neurooncology and central nervous system inflammatory pathologies, a review, J. Cereb. Blood Flow Metab., 2010, 30, 15-35.
- 67 J. P. Bullivant, S. Zhao, B. J. Willenberg, B. Kozissnik, C. D. Batich and J. Dobson, Materials Characterization of Feraheme/Ferumoxytol and Preliminary Evaluation of Its Potential for Magnetic Fluid Hyperthermia, Int. J. Mol. Sci., 2013, 14, 17501-17510.
- 68 Y. X. J. Wang, S. H. Hussain and G. P. Krestin, Superparamagnetic iron oxide contrast agents: physicochemical characteristics and applications in MR imaging, Eur. Radiol., 2001, 11, 2319-2331.
- 69 J. Lodhia, G. Mandarano, N. J. Ferris and S. F. Cowell, Development and use of iron oxide nanoparticles (Part 1): Synthesis of iron oxide nanoparticles for MRI, Biomed. Imaging Intervention J., 2010, 6(2), e12.
- 70 G. B. Toth, C. G. Varallyay, A. Horvath, et al., Current and Potential Imaging Applications of Ferumoxytol for Magnetic Resonance Imaging, Kidney Int., 2017, 92(1), 47-66.

- 71 L. H. Deddens, G. A. F. Van Tilborg, W. J. M. Mulder, H. E. De Vries and R. M. Dijkhuizen, Imaging Neuroinflammation after Stroke: Current Status of Cellular and Molecular MRI Strategies, Cerebrovasc. Dis., 2012, 33, 392-402.
- 72 Notice of Lumirem, Guerbet laboratory.
- 73 Y. X. Wang, Superparamagnetic iron oxide based MRI contrast agents: Current status of clinical application, Quant. Imaging Med. Surg., 2011, 1, 35-40.
- 74 F. Bertorelle, C. Wilhelm, J. Roger, F. Gazeau, C. Ménager and V. Cabuil, Fluorescence-Modified Superparamagnetic Nanoparticles: Intracellular Uptake and Use in Cellular Imaging, Langmuir, 2006, 22, 5385-5391.
- 75 C. G. Varallyay, E. Nesbit, R. Fu, S. Gahramanov, B. Moloney, E. Earl, L. L. Muldoon, X. Li, W. D. Rooney and E. A. Neuwelt, High-resolution steady-state cerebral blood volume maps in patients with central nervous system neoplasms using ferumoxytol, a superparamagnetic iron oxide nanoparticle, J. Cereb. Blood Flow Metab., 2013, 33, 780-786.
- 76 E. A. Vermeij, M. I. Koenders, M. B. Bennink, et al., The In-Vivo Use of Superparamagnetic Iron Oxide Nanoparticles to Detect Inflammation Elicits a Cytokine Response but Does Not Aggravate Experimental Arthritis, PLoS One, 2015, 10(5), e0126687.
- 77 J. Chatterjee, Y. Haik and C.-H. Chen, Size dependent magnetic properties of iron oxide nanoparticles, J. Magn. Magn. Mater., 2003, 257, 113-118.
- 78 Q. Feng, Y. Liu, J. Huang, K. Chen, J. Huang and K. Xiao, Uptake, distribution, clearance, and toxicity of iron oxide nanoparticles with different sizes and coatings, Sci. Rep., 2018, 8, 2082.
- 79 F. Li, Z. Liang, J. Liu, J. Sun, X. Hu, et al., Dynamically Reversible Iron Oxide Nanoparticle Assemblies for Targeted Amplification of T1-Weighted Resonance Imaging of Tumors, Nano Lett., 2019, 19, 4213-4220.
- 80 E. Tysiak, P. Asbach, O. Aktas, et al., Beyond blood brain barrier breakdown - in vivo detection of occult neuroinflammatory foci by magnetic nanoparticles in high field MRI, J. Neuroinflammation, 2009, 6, 20.