

Showcasing research from Dr. Saggiomo, Laboratory of BioNanoTechnology (*BioNT*), Wageningen University, The Netherlands, in collaboration with WETSUS and Leiden University Medical Centre.

COvalent monolayer patterns in Microfluidics by PLasma etching Open Technology – COMPLIT

Plasma microcontact patterning and replica molding are combined to make PDMS/glass microfluidic devices with stripe patterns of cyclodextrins covalently attached on the glass surface inside microchannels. Using the COMPLIT strategy, two-legged guest molecules, selectively, come together on the grooves, Adamantane Binding Beta-cyclodextrin Efficiently Yielding Red Orthogonally Allocated Dyes.

As featured in:



See Vittorio Saggiomo *et al.*, *Analyst*, 2020, **145**, 1629.



Cite this: *Analyst*, 2020, **145**, 1629

Received 29th November 2019,

Accepted 12th January 2020

DOI: 10.1039/c9an02407g

rsc.li/analyst

COvalent monolayer patterns in Microfluidics by PLasma etching Open Technology – COMLOT†

Stan B. J. Willems,^a Jaccoline Zegers,^a Anton Bunschoten,^{a,c}
 R. Martijn Wagterveld,^b Fijs W. B. van Leeuwen,^{a,c} Aldrik H. Velders^{a,c} and
 Vittorio Saggiomo^{a*}

Plasma microcontact patterning (PμCP) and replica molding were combined to make PDMS/glass microfluidic devices with β-cyclodextrin (β-CD) patterns attached covalently on the glass surface inside microchannels. The supramolecular reactivity, re-usability and association constant of β-CD with Cy5–Ad₂ was tested by analyzing signal-to-noise ratios of patterns vs. spacing with fluorescence microscopy.

Introduction

(Bio)sensors using chemically modified surfaces have proven instrumental for the sensitive detection of (bio)analytes.^{1–6} Therefore, the incorporation of modified surfaces within a microfluidic device is becoming an attractive field for creating lab-on-a-chip platforms, *e.g.* the detection of enzymes on chitosan functionalized surfaces.⁷ Within microfluidic devices, the fabrication of polydimethylsiloxane (PDMS)/glass hybrid systems are highly popular because of their versatility and low-cost.^{8,9} These microfluidic devices allow for facile fluid handling, can analyze small volumes and support direct analyte analysis using microscopy.¹⁰ Subsequently, analyte quantification defines the final utility of the microfluidic sensor. In light of this, analyte binding to patterns of molecules within the microfluidic device can be compared with non-functionalized (control) spacing in one field of view in order to quickly establish signal to background ratio's.¹¹

The fabrication of PDMS/glass microfluidic devices, also known as 'replica molding',¹² works by activation of the PDMS

replica containing micro-sized grooves with a glass surface in the plasma oven, and subsequently applying conformal contact between PDMS and glass surface. The grooves in the PDMS replica form a microchannel with desired size features on the glass surface and the microfluidic device is sealed. When the two surfaces are in contact the silanols react to form silyl ethers and covalently 'seal' PDMS to glass, forming a mixed PDMS/glass microfluidic device.¹³ One major drawback of this methodology is that creating patterns of molecules on the glass surface before plasma treatment is not feasible, as the plasma exposure for sealing the microfluidic device will remove the molecular patterns.

We have recently shown that plasma etching, in combination with plasma microcontact patterning (PμCP), is a reproducible method for creating patterns of covalently bound molecules on glass surfaces.¹⁴ In brief, a PDMS stamp, of the kind normally used for microcontact printing (μCP), is placed in contact with a covalently modified glass surface and the construct is subjected to plasma etching. Exposed molecules which are not protected by the PDMS stamp are decomposed by plasma, resulting in molecular patterns. It should be noted that the PDMS 'stamp' is now actually functioning as a 'mask': instead of adding molecular patterns to the surface through 'stamping' as with μCP, patterns are now generated by removing molecular functionality through plasma etching from those areas where the PDMS stamp is not in contact with the surface. Another important advantage with PμCP is that the plasma exposed regions between the stamp patterns and outside the stamp area now contain reactive silanols which can be used for covalent bonding with PDMS surfaces. Based on this utility, we speculated that in order to create a microfluidic device with embedded patterns of covalently functionalized monolayer patterns, PμCP could be combined with PDMS replica molding. Previously, this concept has only been carried out to create surfaces that contain patterns of physisorbed proteins within microfluidic channels.^{15,16} More labor intensive methods, without patterns, involve the manufacturing of glass–glass microfluidic

^aLaboratory of BioNanoTechnology, Wageningen University and Research, Bornse Weiland 9, 6708 WG Wageningen, the Netherlands.

E-mail: vittorio.saggiomo@wur.nl

^bWetsus, European Centre of Excellence for Sustainable Water Technology, Oostergoweg 9, 8911 MA Leeuwarden, the Netherlands

^cInterventional Molecular Imaging, Department of Radiology, Leiden University and Medical Centre, 2300 RC Leiden, the Netherlands

†Electronic supplementary information (ESI) available. See DOI: 10.1039/c9an02407g



systems using vacuum UV light or fluorosilanes in combination with piranha etching.^{17,18}

We are advocating Open Technologies such as ESCARGOT,¹⁹ for cheap microfluidics and HARICOT,²⁰ for responsive materials: here we present the COMPLIT concept: COvalent monolayer patterns in Microfluidics by PLasma etching Open Technology. Therefore, introducing covalent surface modification, as presented here, provides the user with a reliable and cost-friendly patterning methodology that can simply be incorporated in microfluidic channels for simple qualitative as well as more advanced and quantitative applications.

As proof of concept for combined PμCP/replica molding with a covalent modified surface, we investigated the fabrication of β-cyclodextrin (β-CD) patterns within a microfluidic device (Fig. 1). Cyclodextrins are well known for their applications in food technology, pharmaceuticals and drug delivery, but have also emerged as versatile platforms for material science through their interactions with metal ions and inorganic nanoparticles.²¹ Through supramolecular host-guest chemistry, the inside of β-CD's hydrophobic cavity (host) can be used for immobilization of guest functionalized molecules, such as Cyanine 5-diadamantane (Cy5-Ad₂, structure in Fig. S1, ESI†),²² or for capturing adamantane modified micro-particles on flow through host-guest chemistry.²³ Here, we validated the functionalization of β-CD patterns on the glass surface within the fabricated microfluidic device through addition of Cy5-Ad₂, which allowed us to use fluorescence microscopy to study the immobilization and release of Cy5-Ad₂ under flow. In addition, the binding kinetics of Cy5-Ad₂ were also determined in a multichannel COMPLIT device *via* analysis of different Cy5-Ad₂ concentrations on one microfluidic slide.

Results and discussion

Fig. 1a illustrates the steps required for microfluidic device fabrication. A fully functionalized β-CD surface was first obtained following a literature protocol from Onclin *et al.*²⁴ (Fig. 1a(i)): the glass functionalization with heptakis amino β-CD was carried out using 3-aminopropyltriethoxysilane (APTES) for obtaining an amine functionalized surface, 1,4-phenylenediisothiocyanate (PDITC) for creating an amine reactive surface, and finally incubation of heptakis amino β-CD. Surface patterning of β-CD was realized *via* PμCP with a PDMS stamp molded from a silicon wafer that was patterned through UV-lithography (150 μm broad line features and 50 μm broad spacing with 50 μm height, cut to about 0.75 cm²; four plasma cycles of 1 min).^{14,25} PμCP also generates a reactive surface for subsequent microchannel fabrication (Fig. 1a(ii)). The size of the PDMS stamp, along with the length and width of the spacing between features on the PDMS stamp, were chosen to achieve sufficient penetration of plasma gas molecules through the PDMS stamp spacing for etching the modified glass surface. Furthermore, the resolution of features can also

be limited by UV-lithography fabrication of silicon substrates, which is >10 μm without a clean-room. Previously, we reached sub-10 μm resolution by rotating the PDMS stamp during PμCP of a functionalized surface, which removes more of the surface functionalization.¹⁴

In parallel, PDMS was poured into a 3D printed mold containing a ridge of size 20 × 2 × 0.5 mm (length × width × height), forming the PDMS replica after curing in the oven. This will form a microchannel over the glass surface with a volume of 20 μL. The PDMS replica was activated in the plasma oven along with the stamp protected patterns for the final plasma cycle of 1 min, which ensures high quality and irreversible plasma bonding between the PDMS replica and glass surface (Fig. 1a(ii)).²⁶ Following the plasma treatment, the stamp was removed from the glass surface and the activated PDMS replica was then placed on the patterned β-CD glass surface to chemically attach the PDMS to the glass surface, creating a microfluidic channel over the patterned area (Fig. 1a(iii)). In order to promote homogenous sealing of the microfluidic device, we ensured conformal contact between the PDMS replica and glass surface, then attached clamps to keep the microchannel firmly pressed against the glass surface which was followed by further bonding in the oven (Fig. S2, ESI†). The resulting microfluidic device is depicted in Fig. 1b.

Confirmation of β-CD pattern formation was realized by flowing a 0.23 μM solution of Cy5-Ad₂ through the microfluidic device for 15 min at 50 μL min⁻¹. Excess Cy5-Ad₂ was washed away with PBS and DI water, resulting in clear 150 μm broad line patterns when analyzed with a fluorescence microscope (Fig. 1b(ii)). Importantly, no binding of Cy5-Ad₂ was noticed outside the β-CD patterned area, which serves as an additional control for the analyte detection. A more detailed sequence of events is depicted in Fig. S3 (ESI†) showing the microscope images of surface before Cy5-Ad₂ addition, during addition and after washing away excess Cy5-Ad₂ with PBS and DI water. After washing away excess Cy5-Ad₂, the fluorescence intensity of remaining Cy5-Ad₂ on β-CD is generally *ca.* 4 times higher than the background signal between patterns, which is a result of the high affinity divalent host-guest interactions between Cy5-Ad₂ and β-CD (Fig. S3b(ii), ESI†). We also illustrated the Cy5-Ad₂ addition and washing steps by carrying out a time lapse experiment coupled with a live profile plot that follows the fluorescence intensity during these events (Video S1†). Cy5-Ad₂ addition is carried out within 1 min at 50 μL min⁻¹ flow rate and shows a bright signal over the whole surface. Consecutively washing the surface with PBS at 200 μL min⁻¹ decreases the fluorescence intensity to show clear patterns of Cy5-Ad₂ adhered to β-CD already within 2 min, which correlates with the situations depicted in Fig. S3 (ESI†).

An important characteristic of (bio)sensors incorporated within simple microfluidic devices is re-usability of the sensor in order to run different analytical experiments over the same sensor. In our case, supramolecular host-guest chemistry of β-CD with adamantane guest molecules allows for non-






Fig. 1 (a) Scheme showing the fabrication of the patterned microfluidic device: a fully functionalized β -CD surface is obtained after silanization with 3-aminopropyltriethoxysilane (APTES) and 1,4-phenylenediisothiocyanate (PDITC) cross linking of heptakis amino β -CD (i), then a PDMS stamp with line features is placed on top of modified surface for PuCP and plasma etched together with the PDMS replica (ii), and finally the PDMS replica is chemically bonded to the patterned β -CD surface via plasma bonding to create a β -CD patterned microfluidic device (iii). (b) Photograph of a single channel patterned microfluidic device (i) and microscope images of channel using 5 \times objective and 11 \times 2 tile scan after addition of Cy5-Ad2 using the Brightfield (lower left) and Cy5-fluorescence (upper right) settings (ii) (scale bar is 2 mm).



covalent, reversible interactions.²⁷ Hence, stripping the surface of immobilized Cy5-Ad₂ analytes was also studied using a known procedure based on removal of diadamantane functionalized guest molecules with 10 mM β -CD,²⁸ which was carried out using a flow rate of 800 $\mu\text{L min}^{-1}$. In Fig. 2a, the 1st removal of Cy5-Ad₂ and subsequent repeat addition and removal are shown. From the profile plots it should be noted that the intensity values were normalized for the minimal values within the graph (normalizing the spacing intensity to 1). Fig. 2b illustrates that the printed surface was tested for a total of 4 additions of Cy5-Ad₂ and that the loading-rate diminishes by approximately 10% in the consecutive recycling steps, but the pattern intensity remains preserved at *ca.* 3.5 as shown for the 3rd and 4th uses. The washing step with β -CD does not completely remove immobilized Cy5-Ad₂ (Fig. 2a(ii) and (iv)). However, the average pattern intensity after removal remains approximately between 1–1.5 for all removal steps which indicates reversibility is reproducible (Fig. 2b). The sensor technology presented holds potential for monitoring extraction and release of analytes through supramolecular interactions for guest molecules which have similar binding affinities as Cy5-Ad₂ (approximately between 10^7 – 10^{10} M^{-1} in aqueous solutions).^{29,30}

Within lab-on-a-chip technology, the fabrication of (bio) arrays for carrying out multiple analyses using one chip is quite popular. This is also particularly useful for analyzing binding kinetics of analytes through varying analyte concentration. In this case, using one surface increases the accuracy of the quantitative analysis with techniques such as fluorescence microscopy. The inclusion of analyte binding patterns on the sensor also further improves the quantitative analysis by providing the user with a signal-to-noise ratio. To demonstrate the utility of our microfluidic device in such a test, we exchanged the single channel PDMS device for a multiple channel one containing 5 channels with the same groove dimensions as used before. The P μ CP step on β -CD

functionalized surfaces was slightly modified to incorporate β -CD patterns within all 5 channels: before plasma etching, P μ CP stamps were cut in 5 pieces, with each piece having approximately the same width of one microchannel, and then placed on the β -CD glass surface (Fig. 3a). At this stage it is important that sufficient space (approximately 2 mm) is left between the P μ CP stamps and that the stamps are placed within the area where the microchannels will be formed when bonding the PDMS replica. Otherwise, resulting patterns of β -CD will overlap neighboring channels, which hinders proper sealing of the PDMS replica between channels and causes leaching of solution from one channel to another.

Next, 5 \times serial dilution of a Cy5-Ad₂ solution was made and flowed over the β -CD patterned microfluidic device with multiple channels (Fig. 3b), starting from the highest concentration (top channel) to the lowest concentration (bottom channel) (Fig. 3c). A sharp increase in fluorescence intensity was observed close to the channel walls *i.e.* a so-called ‘edge’ effect. This could be attributed to reflection of the PDMS channel wall when using fluorescence microscopy and was therefore not included in the quantitative analysis. The average pattern intensity in Fig. 3d was determined by capturing images of each microchannel using the 10 \times objective and normalizing profile plots for the minimal values, *i.e.* spacing between β -CD patterns. Moreover, a control experiment with 0.23 μM of Cy5-Ad₂ was carried out to show that using multiple strips of PDMS stamps during the P μ CP step does not significantly alter the density of β -CD units over the different patterned areas on one surface (Fig. S4, ESI†). As expected, the fluorescence increases when higher concentrations of Cy5-Ad₂ are used, and the trend in the graph resembles a typical binding affinity profile when fitted to the Langmuir adsorption model (Fig. 3d).[‡] Although Cy5-Ad₂ binds in a divalent fashion with β -CD units on the surface, it should be noted that by using the Langmuir equation the overall binding affinity is




Fig. 2 (a) Re-usability of microfluidic devices showing β -CD patterned surface after addition and washing away excess Cy5-Ad₂ (i), after removal of immobilized Cy5-Ad₂ by washing with β -CD solution (ii), after 2nd addition and washing away excess Cy5-Ad₂ (iii), and after 2nd removal of immobilized Cy5-Ad₂ by washing with β -CD solution (iv). Images were taken at the same location to allow for fair comparison and were captured with 10 \times objective. Scale bars are 100 μm . The profile plots below the images show the normalized intensity of patterns and spacing vs. the distance. (b) Graph showing the average pattern intensity for 4 uses/additions of Cy5-Ad₂ and 3 removals with β -CD. Error bars represent standard deviation in average pattern intensity determined from the profile plot of each image.






Fig. 3 (a) Scheme explaining fabrication of multichannel microfluidic device with β -CD patterns (b) Photograph of patterned microfluidic device with 5 channels. (c) Microscope image of area depicted in (b) after addition of different concentrations of Cy5-Ad₂. The concentration used is written above the designated microchannel. The image was captured using a 5x objective and 8×4 tilescan. Scale bar is 2 mm. (d) Graph showing the average pattern intensity obtained with the different concentrations of Cy5-Ad₂ used. The binding (K_a) and dissociation (K_d) constants are also shown which are based on fitting results to the Langmuir equation. Error bars represent standard deviation in average pattern intensity determined from profile plots of microscope images from each channel.

determined. When analyzing more complex samples (*e.g.* proteins that contain several binding sites), determining the intrinsic binding affinity and effective molarity of a single interaction pair can give more details about relevant binding kinetics.^{30–32} Fitting the data allowed for determining a Cy5-Ad₂ dissociation constant (K_d , the analyte concentration at half of the maximum fluorescence) of $0.3 \times 10^{-7} \text{ M}$ for β -CD patterns which yields an overall binding constant (K_a) of $3.3 \times 10^7 \text{ M}^{-1}$. The acquired binding constant is in line with expectations for divalent host-guest interactions for the β -CD cavity and adamantane.^{29,30} These results show that these patterned microfluidic devices can also be used for carrying out analysis of different concentrations of analytes and suggest that they could also be used for comparing binding affinity of different analytes on one surface.

Conclusions

In this work, we introduce COMPLIT technology: covalent monolayer patterns in Microfluidics by Plasma etching Open Technology. Microfluidic devices were fabricated containing covalently functionalized molecular patterns at the inside of the channel wall by combining plasma microcontact patterning (PμCP) and replica molding. Incorporation of the patterned surfaces in the microfluidic device allows for simple fluid handling and approaching sensor applications within aqueous flow streams. As a proof of concept for sensing applications, we studied the reversible host-guest interactions between patterns of β -CD, included within the microfluidic device, and Cy5-Ad₂. The immobilization and removal of Cy5-Ad₂ was easily analyzed using a fluorescence microscope and it



was possible to re-use the β -CD patterned microfluidic device several times, which is beneficial for extraction applications and avoids laborious repetition of the surface functionalization. It is also possible to study binding of several analytes on one surface through creating a microfluidic device with multiple channels, all containing β -CD patterns. For this, we looked into determining the binding constant of Cy5-Ad₂ for β -CD layers, which correlated with expectations for diadaman-tane functionalized molecules. β -CD modified surfaces could also be used as selective “turn-off” colorimetric sensors for quantitative detection of heavy metal ions in aqueous streams.^{33,34} Furthermore, the versatility of the adamantane- β -CD couple is similar to the (strept)avidin-biotin couple, and therefore allows for incorporating (bio)molecules functionalized with adamantane on the surface. Alternatively, COMPLIT can be used for incorporating other molecules for the surface functionalization besides β -CD. In principle, any molecule containing a primary amine can be patterned in the microfluidic device using the same methodology presented here, e.g. biotin-NH₂, amino sugars for lectin detection or antibodies for antigen detection, provided the (bio)sensor's functionality is preserved after attachment to the surface. The microfluidic device can be used for various applications with aqueous or light organic solvents such as ethanol and methanol, however, use of highly hydrophobic solvents which swell PDMS, such as toluene or chloroform should be avoided. Through this enabling technology for fabrication of patterned PDMS/glass hybrid microfluidic devices, a way is opened for researchers less experienced in surface functionalization and working on different fields ranging from (bio) sensing applications, such as immobilization of micron-sized entities,²³ to analytical chemistry and medical applications like lab- or organs-on-a-chip.^{35,36}

Conflicts of interest

There are no conflicts to declare.

Acknowledgements

This work was performed in the cooperation framework of Wetsus, European Centre of Excellence for Sustainable Water Technology (<http://www.wetsus.eu>). Wetsus is co-funded by the Dutch Ministry of Economic Affairs and Ministry of Infrastructure and Environment, the Province of Fryslân, and the Northern Netherlands Provinces. The authors like to thank the members of the research theme Source Separated Sanitation for the shared knowledge and financial support. We acknowledge Mark Rood for preparing the Cy5-Ad₂ dye.

Notes and references

† In our case, the Langmuir adsorption model is represented by the equation: $y = K_d x / (y_{\max} + x) + 1$, where y is the average fluorescent pattern intensity, y_{\max} is the

maximum amount of fluorescent pattern intensity subtracted by 1, and x is the concentration of Cy5-Ad₂ in solution. The y values in the equation are corrected by adding 1 because the experimental pattern ratio starts at 1.

- 1 A. Kumar, H. A. Biebuyck and G. M. Whitesides, *Langmuir*, 1994, **10**, 1498–1511.
- 2 M. Mrksich and G. M. Whitesides, *Annu. Rev. Biophys. Biomol. Struct.*, 1996, **25**, 55–78.
- 3 S.-H. Hsu, M. D. Yilmaz, C. Blum, V. Subramaniam, D. N. Reinhoudt, A. H. Velders and J. Huskens, *J. Am. Chem. Soc.*, 2009, **131**, 12567–12569.
- 4 D. Dorokhin, S.-H. Hsu, N. Tomczak, D. N. Reinhoudt, J. Huskens, A. H. Velders and G. J. Vancso, *ACS Nano*, 2010, **4**, 137–142.
- 5 A. Gonzalez-Campo, S. H. Hsu, L. Puig, J. Huskens, D. N. Reinhoudt and A. H. Velders, *J. Am. Chem. Soc.*, 2010, **132**, 11434–11436.
- 6 R. de laRica, R. M. Fratila, A. Szarpak, J. Huskens and A. H. Velders, *Angew. Chem., Int. Ed.*, 2011, **50**, 5704–5707.
- 7 Z. Wang, L. He, J. Lv and M. Kimura, *Mater. Res. Express*, 2019, **6**, 045403.
- 8 J. C. McDonald and G. M. Whitesides, *Acc. Chem. Res.*, 2002, **35**, 491–499.
- 9 K. Ren, J. Zhou and H. Wu, *Acc. Chem. Res.*, 2013, **46**, 2396–2406.
- 10 J. P. Lafleur, A. Jönsson, S. Senkbeil and J. P. Kutter, *Biosens. Bioelectron.*, 2016, **76**, 213–233.
- 11 R. S. Kane, S. Takayama, E. Ostuni, D. E. Ingber and G. M. Whitesides, in *The Biomaterials: Silver Jubilee Compendium*, ed. D. F. Williams, Elsevier Science, Oxford, 1999, pp. 161–174, DOI: 10.1016/B978-008045154-1.50020-4.
- 12 Y. Xia, J. J. McClelland, R. Gupta, D. Qin, X.-M. Zhao, L. L. Sohn, R. J. Celotta and G. M. Whitesides, *Adv. Mater.*, 1997, **9**, 147–149.
- 13 D. C. Duffy, J. C. McDonald, O. J. A. Schueller and G. M. Whitesides, *Anal. Chem.*, 1998, **70**, 4974–4984.
- 14 S. B. J. Willems, L. M. I. Schijven, A. Bunschoten, F. W. B. van Leeuwen, A. H. Velders and V. Saggiomo, *Chem. Commun.*, 2019, **55**, 7667–7670.
- 15 S. W. Rhee, A. M. Taylor, C. H. Tu, D. H. Cribbs, C. W. Cotman and N. L. Jeon, *Lab Chip*, 2005, **5**, 102–107.
- 16 M. Rosso, V. van Steijn, L. C. P. M. de Smet, E. J. R. Sudhölter, C. R. Kleijn and M. T. Kreutzer, *Appl. Phys. Lett.*, 2011, **98**, 174102.
- 17 K. Shirai, K. Mawatari and T. Kitamori, *Small*, 2014, **10**, 1514–1522.
- 18 S.-I. Funano, N. Ota, A. Sato and Y. Tanaka, *Chem. Commun.*, 2017, **53**, 11193–11196.
- 19 V. Saggiomo and A. H. Velders, *Adv. Sci.*, 2015, **2**, 1500125.
- 20 A. H. Velders, J. A. Dijkman and V. Saggiomo, *Appl. Mater. Today*, 2017, **9**, 271–275.
- 21 D. Prochowicz, A. Kornowicz and J. Lewiński, *Chem. Rev.*, 2017, **117**, 13461–13501.
- 22 M. T. M. Rood, S. J. Spa, M. M. Welling, J. B. ten Hove, D. M. van Willigen, T. Buckle, A. H. Velders and F. W. B. van Leeuwen, *Sci. Rep.*, 2017, **7**, 39908.



- 23 S. B. J. Willems, A. Bunschoten, R. M. Wagterveld, F. W. B. van Leeuwen and A. H. Velders, *ACS Appl. Mater. Interfaces*, 2019, **11**, 36221–36231.
- 24 S. Onclin, A. Mulder, J. Huskens, B. J. Ravoo and D. N. Reinhoudt, *Langmuir*, 2004, **20**, 5460–5466.
- 25 R. Picone, B. Baum and R. McKendry, in *Methods in Cell Biology*, ed. M. Piel and M. Théry, Academic Press, 2014, vol. 119, pp. 73–90.
- 26 B. Millare, M. Thomas, A. Ferreira, H. Xu, M. Holesinger and V. I. Vullev, *Langmuir*, 2008, **24**, 13218–13224.
- 27 G. Chen and M. Jiang, *Chem. Soc. Rev.*, 2011, **40**, 2254–2266.
- 28 A. Mulder, S. Onclin, M. Peter, J. P. Hoogenboom, H. Beijleveld, J. ter Maat, M. F. Garcia-Parajo, B. J. Ravoo, J. Huskens, N. F. van Hulst and D. N. Reinhoudt, *Small*, 2005, **1**, 242–253.
- 29 A. Mulder, T. Auletta, A. Sartori, S. Del Ciotto, A. Casnati, R. Ungaro, J. Huskens and D. N. Reinhoudt, *J. Am. Chem. Soc.*, 2004, **126**, 6627–6636.
- 30 J. Huskens, A. Mulder, T. Auletta, C. A. Nijhuis, M. J. W. Ludden and D. N. Reinhoudt, *J. Am. Chem. Soc.*, 2004, **126**, 6784–6797.
- 31 A. Perl, A. Gomez-Casado, D. Thompson, H. H. Dam, P. Jonkheijm, D. N. Reinhoudt and J. Huskens, *Nat. Chem.*, 2011, **3**, 317–322.
- 32 D. Di Iorio, M. L. Verheijden, E. van der Vries, P. Jonkheijm and J. Huskens, *ACS Nano*, 2019, **13**, 3413–3423.
- 33 P. Suresh, I. A. Azath and K. Pitchumani, *Sens. Actuators, B*, 2010, **146**, 273–277.
- 34 K. Kanagaraj, K. Bavanidevi, T. J. Chow and K. Pitchumani, *RSC Adv.*, 2014, **4**, 11714–11722.
- 35 A. Polini, L. Prodanov, N. S. Bhise, V. Manoharan, M. R. Dokmeci and A. Khademhosseini, *Expert Opin. Drug Discovery*, 2014, **9**, 335–352.
- 36 P. S. Dittrich and A. Manz, *Nat. Rev. Drug Discovery*, 2006, **5**, 210–218.

