CORRECTION

Check for updates

Cite this: Analyst, 2020, 145, 1950

Correction: Review: a comprehensive summary of a decade development of the recombinase polymerase amplification

Jia Li,^a Joanne Macdonald ()*^{b,c} and Felix von Stetten ()*^{a,d}

DOI: 10.1039/c9an90127b

rsc.li/analvst

Correction for 'Review: a comprehensive summary of a decade development of the recombinase polymerase amplification' by Jia Li *et al., Analyst*, 2019, **144**, 31–67.

The authors regret that the original version of the review contained some incorrect data. Corrections to the original article are listed as follows.

Some of the values in Table 5 were incorrect. The corrected version of Table 5 is presented below.

In section 2.3, data from ref. 41 were not cited correctly and should be removed. The respective passage should read as follows: "However, shorter amplicons (79 nucleotides;³⁷ 94 nucleotides^{38–40}) and longer amplicons up to 1500 nucleotides⁶ have also been reported."

In the Fig. 2 caption, "(*Bsu* or *Sau*)" should be removed after "recombinase" and inserted after "polymerase". The corrected passage should read as follows: "The recombinase disassembles from the nucleoprotein filament once the strand exchange is performed, and will be available for the next pair of primers. Next, the DNA polymerase (*Bsu* or *Sau*) extends from the 3' end of primers."

In section 2.5, ref. 76 is not relevant as it is the same as ref. 77, and should be disregarded. The respective passage should read as follows: "However, several research groups have studied RPA reaction temperatures that lie outside of the recommended range.^{38,44,45,60,62–75,77,78} The largest temperature range was tested between 15 °C and 50 °C;^{62,64,69,70,77} and results indicated the marginal reaction temperature to produce a positive result should be greater than 30 °C.^{62–64,66,67,69,71,74,77},"

In section 2.5, ref. 63 was not interpreted correctly. The corrected text should read: "Moreover, Lillis *et al.*⁶³ showed that the ambient temperature also had an effect on RPA reaction: the RPA reaction was unstable if the ambient temperature was below 30 °C, even at extended reaction time."

In section 2.8, ref. 107 is not required and should be deleted. The corrected passage should read: "For the TwistAmp® nfo kit, however, two types of amplicons are generated... (note that only the dual-labelled product will generate a positive signal in the test zone of a lateral flow strip detection based on a sandwich assay).^{105,106,108},"

In section 2.8, the problems reported in ref. 117 were not sufficiently reflected in the original version of the review. Ref. 117 should be reported separately and the passage should read as follows: "As with lateral flow strip detection, direct usage of RPA amplicons is possible, but it is recommended to dilute the amplicons with the running buffer (*e.g.* 1/100 dilution) before running on the strip to (1) improve its wicking performance¹¹⁴ and (2) avoid "faint ghost band" effects.^{45,54,115,116} However, the dilution of the amplicon does not always prevent the appearance of a faint band, which can lead to specificity problems in the assay.¹¹⁷"

In section 3.2, ref. 166 was not cited correctly. The corrected version should read: "Results suggest that electrochemical detection could be up to 10-fold more sensitive than optical detection (by enzyme linked oligonucleotide assay).¹⁶⁶"

In section 3.2, a reference was not provided for the sensitivity of the GeneXpert MTB/RIF assay. A reference should be added to the end of the following passage: "This ruthenium compound-based electrochemical detection achieved 11 CFU mL^{-1} of

View Article Online

^aLaboratory for MEMS Applications, IMTEK – Department of Microsystems Engineering, University of Freiburg, Georges-Köhler-Allee 103, 79110 Freiburg, Germany. E-mail: Felix.von.Stetten@Hahn-Schickard.de; Tel: +49 761 203-73243

^bInflammation and Healing Research Cluster, Genecology Research Centre, School of Science and Engineering, University of the Sunshine Coast, Qld, Australia. E-mail: jmacdon1@usc.edu.au; Tel: +61 7 5456 5944

^cDivision of Experimental Therapeutics, Columbia University, New York, NY, USA. E-mail: jm2236@columbia.edu

^dHahn-Schickard, Georges-Köhler-Allee 103, 79110 Freiburg, Germany

Mycobacterium tuberculosis analytical sensitivity, which is even more sensitive than the GeneXpert MTB/RIF (Cepheid Inc.) detection (a World Health Organisation recommended tuberculosis diagnostic system that employs PCR real-time fluorescent detection; 131 CFU mL⁻¹)." The added reference is shown below as ref. 1.

In section 3.4, ref. 113 was not cited correctly. Instead of *Mycobacterium bovis*, *Mycobacterium tuberculosis* was used for demonstration. The corrected sentence should read: "Liu *et al.*¹¹³ demonstrated a duplex detection of *IS6110* and *IS1081* insertion sequences of *Mycobacterium tuberculosis* using RPA-SMR assay, and achieved 3.2 and 12 genomic DNA copies per reaction analytical sensitivity respectively."

In section 4.1, ref. 182 and 183 were not cited precisely. The corrected version of the text should read: "The "microcliff" structured microchip demonstrated by Yeh *et al.* encased 200 to 1500 wells (30–100 nL per well),¹⁸² and 224 wells (100 nL per well),¹⁸³ which allowed detection of 10^3-10^5 and $10-10^5$ copies per μ L of MRSA DNA, respectively."

In section 4.2, ref. 194 should be deleted after the following sentence: "For the latter, one demonstration is on the digital video disk (DVD) by Maquieira research group, and the resulting signals can be detected by a DVD player (Fig. 12B).^{188,191}"

Ref. 101 in the original article was incorrect and should be replaced with the correct reference, shown below as ref. 2.

Ref. 107 was not cited in the original article and should be disregarded.

cess Article. Published on 23 January 2020. Downloaded on 8/24/2025 5:37:52 AM.	This article is licensed under a Creative Commons Attribution 3.0 Unported Licence.
Open Acc	(cc)) BY

Table 5 RPA literature describing clinical/field trials

nicai ceificity of nchmark thod nbared	RPA Ref.	e same 125	e same 92	e same 126	e same 127	e same 117	e same	gher 128	e same	gher 129
Clunical Clu sensitivity of spe benchmark ber method me compared cor	to RPA to 1	The same Th	Higher Th	Higher Th	Lower Th	Higher Th	Higher Th	Higher Hig	Higher Th	Higher Hig
Limit of detection of	benchmark method	1000 RNA copies	I	1	I	8 genomic copies/ reaction in YFV RNA extracts		I		10 ^{2.5} cysts per mL of stool
Benchmark	method	Real-time RT ^a -PCR	Roche Cobas Amplicor CT assay	Real-time PCR	Real-time PCR	Real-time RT ^a -PCR		Culture		Real-time PCR
Clinical	specificity	100%	100%	100%	100%	100%	100%	95.4%	100%	96%
Clinical	sensitivity	100%	83%	96%	100%	80%	71.4%	87.5%	91.4%	73%
	Clinical/field sample(s)	16 fecal and 14 nasal swab specimens collected from cattle showing intestinal and/or respiratory manifestations	70 self-collected first void morning urine samples from young adults (19 males and 51 females)	50 vaginal/anal samples collected from women	A total of 10 human stool samples clinically verified to contain cryptosporidium by a reference laboratory and 11 stool samples from healthy volunteers presumed to be uninfected	34 samples of monospecific pools of wild-caught mosquitoes collected from Kedougou, southern Senegal	27 RNA samples of mosquito pools	121 specimens including induced and expectorated sputum $(n = 119)$ and	respiratory washes (bronchial and tracheal, n = 2) collected from a total of 101 tuberculosis suspect cases (no more than 3 specimens/individual were tested)	104 clínical stool samples
	Limit of detection	10 to 100 RNA copies (19 RNA copies by probit analysis)	5–12 pathogens/ reaction	98 genome copies	100 oocysts per mL stool	44 genomic copies/ reaction in YFV RNA extracts; 21 genomic copies/reaction of	YFV-spiked human plasma samples	6.25 fg	20 fg	10^{3} - $10^{3.5}$ cysts per mL of stool
Detection	method	Real-time fluorescent detection	Lateral flow strip detection	Real-time fluorescent detection	Lateral flow strip detection	Real-time fluorescent detection on the tube scanner	Real-time fluorescent detection on the microfluidic platform	Real-time fluorescent detection		Lateral flow strip detection
	Analyte(s)	Nucleocapsid (N) gene of bovine coronavirus	Chlamydia trachomatis CDS2 gene	cAMP factor (<i>cfb</i>) gene of group B streptococci	DNA target sequence specific to <i>Cryptosporidium</i> spp.	5'-Untranslated region of Yellow fever virus (YFV)		IS6110 gene of Mycobacterium tuberculoss (MTB)	IS1081 gene of Mycobacterium tuberculoss	<i>Giardia</i> beta giardin gene

Analyte(s)	Detection method	Limit of detection	Clinical/field sample(s)	Clinical sensitivity	Clinical specificity	Benchmark method	Limit of detection of benchmark method	Clinical sensitivity of benchmark method compared to RPA	Clinical specificity of benchmark method compared to RPA	Ref.
IS6110 gene of Mycobacterium tuberculoss	Real-time photonic detection	10 ⁻⁶ -Fold diluted MTB sample	42 clinical samples including 13 smear and culture positive samples and 22 smear and culture neositive samples	86%	95%	Real-time PCR	I	Higher	Higher	130^{b}
A highly conserved 3'untranslated region that cover DENV 1-4	Real-time fluorescent detection	DENV serotype 1: 237 RNA copies; DENV serotype 2: 618 RNA copies; DENV	Inactivated DENV 1-4 spiked plasma and 31 DENV positive samples in Kedougou region in Senegal	98%	I	Real-time RT ^a -PCR	I	Higher	I	131
		serotype 3: 363 RNA copies; DENV serotype 4: 383 RNA copies	RNA of 90 plasma samples extracted and tested between 2012–2013 by RT ^a -PCR in Bangkok (Thailand)	72%				Higher	I	
47 kDa gene sequence from the Karp strain of <i>Orientia</i> <i>tsutsugamushi</i> (47-RPA) and the 17 kDa gene sequence from the Willmington strain	Lateral flow strip detection	47 kDa gene: 53 DNA copies/reaction 17 kDa gene: 20 copies/reaction	10 positive and 10 negative human samples 	80%	100%	PCR PCR	47 kDa gene: 10 DNA copies/reaction 17 kDa gene: 6 DNA copies/reaction	Higher 	Higher —	95
of Rickettsta typhi Ribosomal 18S DNA of Entamoeba histolytica	Lateral flow strip detection	2.5 fg from serial dilutions of pure DNA extracted from parasites; 40 parasites from spiked stool sample	32 samples of DNA extracted from clinical samples	100%	100%	Real-time PCR	2.5 fg from serial dilutions of pure DNA extracted from parasites	The same	The same	132
A sequence designed based on ITS sequences of the <i>Madurella</i> <i>mycetomatis</i> type strain CBS 109801	Gel electrophoresis detection	0.23 ng of DNA	12 patient biopsy specimens	100%	100%	Conventional PCR	I	The same	The same	133
Ebola virus (EBOV) nucleocapsid sequence	Real-time fluorescent detection	5 genomic copies/ reaction of a molecular RNA standard; 15 genomic copies/ reaction in EBOV- spiked human plasma samples	928 post-mortem swab samples	100%	100%	Real-time RT ^a -PCR	1	The same	The same	134

Open Access Article. Published on 23 January 2020. Downloaded on 8/24/2025 5:37:52 AM.

(cc) BY

Table 5 (Contd.)

View Article Online

Correction

	Detection			Clinical	Clinical	Benchmark	Limit of detection of	Clinical sensitivity of benchmark method compared	Clinical specificity of benchmark method compared	د ډ
Antaryce(s) Orf virus (ORFV) DNA polymerase gene segments	Real-time fluorescent detection	100 DNA copies	22 samples collected from suspected cases of Orf, 8 nasal swabs collected from	86%	100%	Real-time PCR		Higher	The same	135 135
Leader peptidase A (<i>LepA</i>) gene of <i>Streptococcus</i> <i>pneumoniae</i>	Real-time fluorescent detection	4.1 genome equivalents/reaction	experimentary inteceed scheep and 5 samples obtained from healthy goats 15 blood samples including 11 confirmed culture positive and 4 confirmed culture negative for	100%	100%	Real-time PCR	5.1 genome equivalents/reaction	The same	The same	67
Orf virus (ORFV) DNA polymerase gene segments	Lateral flow strip detection	80 copies/reaction of DNA plasmid	<i>Streptococcus pneumoniae</i> 24 ORFV-spiked tissues samples, 53 samples collected from goats with suspected ORFV infection, 8 nasal swabs samules and 5 tissues	100%	100%	Real-time PCR	Ι	The same	The same	64
<i>Leishmania donovani</i> (LD) kinetoplast minicircle DNA	Real-time fluorescent detection	100 DNA copies applying the LD DNA linearised plasmid; 1 genomic DNA copy	samples from healthy goats 96 buffy coats and skin biopsies collected from visceral leishmaniasis, asymptomatic and post-kala-	100%	100%	Real-time PCR	Ι	The same	The same	121
Highly pathogenic porcine reproductive and respiratory syndrome virus	Real-time fluorescent detection	70 RNA copies/ reaction	azar dermal leishmaniasis 68 tissue samples and 10 serum samples collected from suspected pigs of HP-PRRSV, 35 serum samples and 12 tissue	97.6%	100%	Real-time RT ^a -PCR	I	Higher	The same	136
Nor2 gene 100% conserved sequence of a major capsid protein gene of all cyprinid	Gel electrophoresis detection	10 copies of genomic DNA	samples conceted nom healthy pigs 12 confirmed latently infected fish and 1 confirmed uninfected fish	100%	100%	Real-time PCR	1	Lower	The same	99
herpesvirus 3 strains cAMP factor (cfb) gene of group B strentorocci	Real-time fluorescent detection	6.25–12.5 genome equivalents	124 clinical samples	100%	100%	Real-time PCR	3.1–6.25 genome equivalents	The same	The same	137
Non-structure protein 1 (nsP1) of Chikungunya virus	Real-time fluorescent detection	80 genome copies of extracted RNA from CHIKV isolate LR	58 suspect Chikungunya fever cases	100%	100%	Real-time RT ^a PCR	80 genome copies of extracted RNA from CHIKV isolate LR	The same	The same	87
(CLILIAN) A sequence designed in NS2A region conserved among all Zika virus lineages	Real-time fluorescent detection	suan 21 RNA copies	25 positive and 9 negative urine samples collected during the Zika virus epidemic in Tuparetama, Brazil	92%	100%	Real-time RT ^a -PCR		Higher	The same	138

Correction

Open Access Article. Published on 23 January 2020. Downloaded on 8/24/2025 5:37:52 AM.

(cc) BY

Table 5 (Contd.)

1954 | Analyst, 2020, 145, 1950-1960

	Detection			Clinical	Clinical	Benchmark	Limit of detection of	Clinical sensitivity of benchmark method compared	Clinical specificity of benchmark method compared	e F
lyte(s)	method	Limit of detection	Clinical/neid sample(s)	sensitivity	specificity	method	benchmark method	to KPA	to kPA	Keī.
otein-coupled nokine receptor CR) gene of py skin disease s (LSDV)	Real-time fluorescent detection	100 DNA copies (179 DNA copies by probit analysis)	12 negative skin samples and 22 skin nodules of suspected LSDV-infected cattle collected during the summer of 2012 in Dakahlia Governorate. Foort	100%	100%	Real-time PCR	37 DNA copies	The same	The same	139
0 gene of obacterium avium sp. P)	Real-time fluorescent detection	16 plasmid copies per μL; 500 fg genomic DNA/ reaction	Archived DNA of MAP positive blood ($n = 14$), sperm ($n = 18$), faccal ($n = 12$) and tissue ($n = 4$) samples and 20 MAP- negative faecal samples	89.5%	I	Real-time PCR	1 plasmid copies per μL; 50 fg genomic DNA/ reaction	Higher	I	140
4 gene of tate cancer	Real-time fluorescent detection	1000 RNA copies	9 urine samples obtained from prostate cancer and 2 urine samples from healthy individuals	%06	100%	Real-time RT ^a -PCR	I	The same	The same	141
gene of porcine ovirus (PPV)	Real-time fluorescent detection	300 DNA copies	101 clinical tissue samples (serum, liver, kidney, lymph node, spleen and duodenum) collected from pig farms with suspected cases of PPV in Gansu province, China, and 27 clinical samples (serum, kidney and duodenum) collected from healthv pigs	94.4%	100%	Real-time PCR	I	Higher	The same	5
leocapsid gene pe 2 porcine oductive and iratory syndrome a (PRRSV)	Real-time fluorescent detection	100 RNA copies (690 RNA copies by probit analysis)	60 clinical samples (lýmph node, lung, spleen and liver) collected from diseased pigs suspected of having PRRS from 5 pigs farms in Hebei province, China from 2015-2016	I	I	Real-time RT ^a -PCR	100 RNA copies		I	142
chrome b gene heileria annulata	Lateral flow strip detection	2 pg genomic DNA	17 anticoagulated blood samples collected from tropical theileriosis endemic areas in Gansu province, China	1	I	Real-time PCR	I		1	67
like gene of io owensii	Real-time filuorescent detection	2 plasmid copies (2.84 plasmid copies by probit analysis)	138 clinical shrimp obtained from immersion bioassay, including 70 shrimp acute hepatopancreatic necrosis disease (AHPND) infected shrimp and 68 non-AHPND infected shrimp	100%	100%	Real-time PCR	I	Lower	Lower	143

Open Access Article. Published on 23 January 2020. Downloaded on 8/24/2025 5:37:52 AM.

(cc)) BY

Table 5 (Contd.)

Correction

								Clinical sensitivity of benchmark method	Clinical specificity of benchmark method	
Analyte(s)	Detection method	Limit of detection	Clinical/field sample(s)	Clinical sensitivity	Clinical specificity	Benchmark method	Limit of detection of benchmark method	compared to RPA	compared to RPA	Ref.
rRNA gene of Fasciola henatica	Gel electrophoresis detection	1.6 pg μL ⁻¹ DNA copies	102 human stool samples selected from banked	87.8%	100%	Real-time PCR	1.6 pg μL ⁻¹ DNA conies	Lower	The same	144
	Lateral flow strip detection	1.0 pg µL ⁻¹ DNA copies	specimens	95.2%	90.4%			Lower	Higher	
N gene of pest des petits ruminants	Real-time fluorescent	100 plasmid copies	32 clinical samples collected from suspected cases of	%06	100%	Real-time RT ^a -PCR	10 plasmid copies	Higher	The same	145
VITUS (PPKV)	detection Lateral flow strip detection	150 plasmid copies	PPKV IN Gansu province, China and 5 samples obtained from healthy sheen	%06	100%			Higher	The same	
ITS2 gene of Phytophthora	Real-time fluorescent	50 fg μL ⁻¹ of genomic DNA	24 potato leaf samples collected from fields with	33.3%	100%	LAMP	50 fg μL ⁻¹ of genomic DNA	Higher	Lower	146
unjestaris	detection		symptoms of late blight infections in New Brunswick and Quebec provinces, Canada memorinaly							
ORF2 gene of porcine circovirus tyne 2 (PCV2)	Real-time fluorescent detection	100 plasmid copies	canada, respectively 65 clinical samples (spleen, inguinal lymph node, tonsil, hung and serum) collected	100%	100%	Real-time PCR	80 plasmid copies	The same	The same	69
	Lateral flow strip detection	100 plasmid copies	from suspected PCV2 infection pigs from 8 pig farms in Shandong province, China; 37 clinical samples (inguinal	100%	100%			The same	The same	
			lymph node, tonsil, lung and serum) collected from Gansu Province, China, and 10 PCV1 positive samples conserved in the							
gD gene of pseudorabies virus	Real-time fluorescent	100 DNA copies	laboratory 76 clinical samples (tonsil, heart, spleen, lymph nodes,	93.3%	100%	Real-time PCR	I	Higher	The same	70
	detection Lateral flow strip detection	160 DNA copies	lung and serum) collected from pig farms in Shandong province, China, and 26 clinical samples (lymph nodes, tonsil and serum) collected from	93.3%	100%			Higher	The same	
B1 gene of Toxoplasma gondii	Lateral flow strip detection	0.1 oocysts/reaction	healthy pigs 35 soil samples and 15 water samples collected from parks, residential areas, schools and gutterways in Lanzhou city, Gansu rovince, China, during August 2016	100%	100%	Nested PCR	1 oocyst/reaction	The same	The same	71

Correction

Open Access Article. Published on 23 January 2020. Downloaded on 8/24/2025 5:37:52 AM.

(cc) BY

Table 5 (Contd.)

Analvte(s)	Detection method	Limit of detection	Clinical/field samnle(s)	Clinical sensitivity	Clinical snecificity	Benchmark merhod	Limit of detection of benchmark method	Clinical sensitivity of benchmark method compared to RPA	Clinical specificity of benchmark method compared fo RPA	Ref.
RNA transcript of TMPRSS2:ERG (a fusion gene for prostate cancer)	RPA fluocculation assay	10 ⁵ RNA copies	Clinical urine specimens from 10 metastatic castration-resistant promising prostate cancer protects and 5 healthy	70%	100%	Conventional RT ^a -PCR		The same	The same	101
VP2 gene of porcine parvovirus	Real-time fluorescent detection	100 DNA copies (103 DNA copies by probit analysis)	control patterns node, lung, spleen, kidney and duodenum collected from pigs with reproductive disorders, diarrhea or respiratory disease in Hebei province, China from 2014 to 2016	100%	100%	PCR	100 DNA copies	The same	The same	147
G-protein-coupled chemokine receptor (GPCR) gene of	Real-time fluorescent detection	300 plasmid copies	107 clinical samples (liver, lung, kidney, spleen, skin and blood) collected from	%26	100%	Real-time PCR		Higher	The same	148
Capripoxpirus	Lateral flow strip detection	300 plasmid copies	14 suspected sheep and 6 suspected goats in Gansu province which were characterised by pyrexia, excessive salivation and generalised pock lesions in the skin during the period of October 2014 to August 2015	%26	100%			Higher	The same	
Nucleocapsid protein gene of canine distemper virus	Real-time fluorescent detection	9.4 RNA copies (31.8 RNA copies by probit analysis)	32 nasal/oropharyngeal swabs collected from 20 dogs of both sexes (various breeds and ages) from the animal hospital of Agricultural University of Hebei and 12 raccoon dogs from the farms in Hebei Province, China from 2014 to 2016	100%	100%	Real-time RT ^a -PCR	94 RNA copies	The same	The same	149
<i>imp</i> gene of <i>Candidatus</i> Phytoplasma oryzae	Real-time fluorescent detection	1–10 plasmid copies	66 Napier grass samples from various geographical locations in western Kenya	100%	57.1%	Real-time PCR	I	Lower	The same	79
	Lateral flow strip detection	10–100 plasmid copies		I						

Open Access Article. Published on 23 January 2020. Downloaded on 8/24/2025 5:37:52 AM.

(cc) BY

Table 5 (Contd.)

Table 5 (Contd.)										
Analyte(s)	Detection method	Limit of detection	Clinical/field sample(s)	Clinical sensitivity	Clinical specificity	Benchmark method	Limit of detection of benchmark method	Clinical sensitivity of benchmark method compared to RPA	Clinical specificity of benchmark method compared to RPA	Ref.
<i>imp</i> gene of <i>Candidatus</i> Phytoplasma mali	Real-time fluorescent detection	10 copies of cloned plasmid	38 roots of field samples from apple (Malus domestica) trees collected in autumn	100%	100%	Real-time PCR	1	The same	The same	72
	Lateral flow strip detection	10 copies of cloned plasmid	2014, in spring 2015 and in June 2016 in private orchards or in the experimental field of the Institute for fruit growing in Samochvalovichi, Belarus	100%	100%			The same	The same	
N gene of rabies	Real-time fluorescent detection	1000 RNA copies per μL of strains SAD B19, Bobcat USA and Kelev	A panel of RNA from 33 field samples	%26		Real-time PCR	1 RNA copies per µL of strains SAD B19, Bobcat USA and Kelev	Higher	I	150
<i>KRAS</i> oncogenic mutation gene G12D on Exon 12	Real-time silicon photonic microring-based	1% to 100% of the mutant cells	70 frozen tissues samples from colorectal cancer patients in Bio-Resource	100%	100%	Conventional PCR	30% to 100% of the mutant cells	Lower	The same	151
<i>KRAS</i> oncogenic mutation genes G13D on Exon 13	detection		Center of Asian Medical Center, including 24 samples with the G12D mutation (34.3%), 26 samples with G13D mutation (37.1%) and 20 samples with no mutation (28.6%)	100%	100%			Lower	The same	
A consensus region that covers all 7 S-segment clades of Crimean-Congo Hemorrhagic fever virus (CCHFV)	Real-time fluorescent detection	500 RNA copies (251 RNA copies by probit analysis)	21 extracted patient sera samples obtained in relation to outbreaks of CCHFV in 2013–2015 in Tajikistan	88%	100%	Real-time PCR	I	Higher	The same	152
Canine parvovirus 2 (CPV-2) nucleocapsid protein gene	Real-time fluorescent detection	10 copies of recombinant plasmid	91 fecal swab samples collected from the dogs from 2012 to 2016	100%	100%	Real-time PCR	10 copies of recombinant plasmid	The same	The same	153
Ğ gene of bovine ephemeral fever virus (BEFV)	Lateral flow strip detection	8 plasmid copies/ reaction (corresponding to 24 RNA copies)	104 clinical blood specimens and 24 tissue samples including 16 lung tissue specimens, 8 lymph gland specimens collected from suspected dairy cattle cases of BEFV infections in eastern China	97.89%	90.91%	PCR PCR	I	Higher	Higher	74

Correction

This article is licensed under a Creative Commons Attribution 3.0 Unported Licence.

(cc) BY

Open Access Article. Published on 23 January 2020. Downloaded on 8/24/2025 5:37:52 AM.

Analyst

Detection	- - - - -			Clinical	Clinical	Benchmark	Limit of detection of	Clinical sensitivity of benchmark method compared	Clinical specificity of benchmark method compared	e 6
method Limit of detection Clinical/field san	LIMIT OF DETECTION CUNICAL/Held San	Clinical/field san	nple(s)	sensitivity	specificity	method	benchmark methoa	to KPA	to KPA	ke
Lateral flow strip 8 plasmid copies/ 320 individual fe detection reaction collected between 2016 and Septem 2016 and Septem from 10 different from 10 different farms located in geographic region Shandong provin Shandong provin	8 plasmid copies/ 320 individual fe reaction collected between 2016 and Septem from 10 different farms located in geographic region Shandong provin	320 individual fe collected between 2016 and Septem from 10 different farms located in geographic regioi Shandong provin	cal samples a September aber 2017 dairy 10 distinct ns of ce, China	100%	97.63%	Real-time PCR	8 plasmid copies/ reaction	The same	Higher	77
Real-time10 plasmid copiesSamples of spltfluorescent $(15 \text{ plasmid copies})$ head kidney $(n$ detectionby probit analysis)water $(n = 5)$	10 plasmid copies Samples of sple $(15 \text{ plasmid copies})$ head kidney $(n$ by probit analysis) water $(n = 5)$	Samples of splt head kidney $(n$ water $(n = 5)$	en (n = 78), $= 78) and$	100%	84.89%	PCR PCR	10 plasmid copies (11 plasmid copies by probit analysis)	The same	Higher	154
Real-time 3.767 log 10 genomic Stool samples (fluorescent copies (LGC) collected in 20 detection shenzhen Cent control and Pr	3.767 log 10 genomic Stool samples (copies (LGC) collected in 20 Shenzhen Cent Control and Pr	Stool samples (collected in 20 Shenzhen Cent Control and Pr	(<i>n</i> = 44) 17 by cer for Disease evention	100%	100%	Real-time PCR	2.026 log 10 genomic copies (LGC)	The same	The same	155
Stool samples (collected from suspected hanc disease at the p department of Hospital (South University, Gua	Stool samples (collected from suspected hanc disease at the p department of Hospital (South University, Gua	Stool samples (collected from - suspected hand disease at the p department of - Hospital (South University, Gua	n = 134) patients with l-foot-mouth ediatrics Zhujiang nern Medical ngzhou,	89.5%	100%			Lower	The same	
Lateral flow strip 10 copies 62 animal (includence) detection (recombinant Apodemus agrar plasmid); 12 copies nonvegicus, Micr of genomic DNA and Neomys fod samples include and multiple and Neomys fod samples include and multiple 2 infected in the v trapped in the v	10 copies 62 animal (incluction of combinant (recombinant Apodemus agrar plasmid); 12 copies nonvegicus, Micropations, Micro	62 animal (incl <i>Apodemus agrar</i> <i>norvegicus, Micr</i> and <i>Neomys fod</i> samples includ animals trapped animals trapped animf 55 unimfect trapped in the v	uding <i>ius, Rattus</i> <i>otus fortis</i> <i>otus fortis</i> <i>iug</i> 5 infected ing 5 infected ing 5 infected ing 1 in the wild, e laboratory ted animals wild	100%	100%	PCR	12 copies of genomic DNA	The same	The same	156
Lateral flow strip 10 copies DNA of spleer detection (recombinant 5-week old C5 plasmid); 7 copies of Coxiella burne genomic DNA mice and 9 co infected mice	10 copiesDNA of spleer(recombinant5-week old C5plasmid); 7 copies ofCoxiella burnegenomic DNAmice and 9 coinfected mice	DNA of spleer 5-week old C5 <i>Coxiella burne</i> mice and 9 co infected mice	us from 7BL/6 female <i>tii</i> -infected introl PBS-	100%	100%	Real-time PCR	7 copies of genomic DNA	The same	The same	157

This article is licensed under a Creative Commons Attribution 3.0 Unported Licence.

(cc)) BY

Table 5 (Contd.)

Open Access Article. Published on 23 January 2020. Downloaded on 8/24/2025 5:37:52 AM.

The Royal Society of Chemistry apologises for these errors and any consequent inconvenience to authors and readers.

Acknowledgements

The authors thank Valentina Paz Gacitúa Miranda, Hahn-Schickard, for thoroughly studying all of the referenced literature and for spotting gaps and mistakes.

References

World Health Organization, Xpert MTB/RIF Implementation Manual: Technical and Operational 'How-To'; Practical Considerations. No. WHO/HTM/TB/2014.1. World Health Organization, 2014, Geneva, Switzerland, ISBN: 978 92 4 150670 0.
 K. M. Koo, E. J. Wee, P. N. Mainwaring and M. Trau, *Sci. Rep.*, 2016, 6, 30722.