# Lab on a Chip



**View Article Online** 

COMMUNICATION

Check for updates

Cite this: Lab Chip, 2020, 20, 4349

Received 6th August 2020, Accepted 4th November 2020

DOI: 10.1039/d0lc00793e

rsc.li/loc

Established digital bioassay formats, digital PCR and digital ELISA, show extreme limits of detection, absolute quantification and high multiplexing capabilities. However, they often require complex instrumentation, and extensive off-chip sample preparation. In this study, we present a dipstick-format digital biosensor (digital dipstick) that detects bacteria directly from the sample liquid with a minimal number of steps: dip, culture, and count. We demonstrate the quantitative detection of *Escherichia coli* (*E. coli*) in urine in the clinically relevant range of  $10^2-10^5$  CFU ml<sup>-1</sup> for urinary tract infections. Our format shows 89% sensitivity to detect *E. coli* in clinical urine samples (n = 28) when it is compared to plate culturing (gold standard). The significance and uniqueness of this diagnostic test format is that it allows a non-trained operator to detect urinary tract infections in the clinically relevant range in the home setting.

### Introduction

In digital bioassays, a reaction solution is first partitioned into a large number of isolated microreactors such that most reactors contain 0 or 1 target analyte particles, whereafter an assay in each of the microreactors reveals the presence of the analyte quantitatively and discretely. Digital bioassays are of interest for their potentially extremely low limit of detection (single molecule detection), absolute quantification, and multiplexing of the bioassays, in an uncomplicated assay format.<sup>1</sup> Digital bioassays include digital polymerase chain  $(dPCR)^2$ digital recombinase polymerase reactions amplification (dRPA),<sup>3,4</sup> and digital loop-mediated isothermal amplification  $(dLAMP)^5$  for nucleic acid detection; digital enzyme-linked immunosorbent assays (digital ELISA)6,7 for protein detection, and; digital colony-forming-unit (CFU)

# Digital dipstick: miniaturized bacteria detection and digital quantification for the point-of-care<sup>+</sup>

Emre Iseri, <sup>(1)</sup>\*<sup>a</sup> Michael Biggel,<sup>b</sup> Herman Goossens,<sup>b</sup> Pieter Moons<sup>b</sup> and Wouter van der Wijngaart<sup>\*a</sup>

assays<sup>8,9</sup> for bacterial detection. Platforms for digital bioassays, such as droplet microfluidic systems,<sup>10</sup> pump-valve systems,<sup>11</sup> planar emulsion arrays,<sup>12</sup> SlipChip,<sup>13</sup> lab-on-a-disc,<sup>14</sup> and SIMPLE chip,<sup>3</sup> either require sophisticated instrumentation<sup>4–9,11,12,14,15</sup> or extensive sample preparation.<sup>3–5,8,10,12–14,16</sup> The resulting cost of testing and the need for trained personnel have limited the translation of digital assays to the point-of-care (PoC) setting for common infectious diseases like urinary tract infections (UTIs).<sup>17,18</sup>

Today's gold standard diagnostic test for UTI detection is culture-based phenotypic testing,<sup>19</sup> which takes 1–3 days and requires trained personnel and laboratory facilities. Other diagnostic formats for the detection of UTIs include culturebased assays (Flexicult,<sup>17,20</sup> dip-slides<sup>21,22</sup>), nitride sticks,<sup>23,24</sup> nucleic acid based tests (PCR,<sup>25</sup> LAMP<sup>5</sup>), immunoassays,<sup>26–29</sup> and single cell imaging based tests.<sup>9</sup> These tests are limited by their need for operation by trained personnel, lab facilities or readout instruments that add cost and complexity, or suffer from high rates of false negative results (Table 1).

Here, we provide a digital assay in dipstick format (digital dipstick) with extremely simple handling that measures *Escherichia coli* (*E. coli*) concentrations in clinical urine samples without sample preparation. Digital dipsticks have the potential to be used as a self-diagnostic test for UTIs by virtue of their three key characteristics:

1. Clinically relevant results at a potentially low cost.

2. No complicated external equipment or technical skills required during testing.

3. Easy manual readout by the naked eye or potentially automated using a cellphone camera.

### Results and discussion

#### **Digital dipsticks**

Digital dipsticks are coffee spoon-like plastic sticks made out of 1 mm thick polymethyl methacrylate (PMMA) plates (Fig. 1A and S1A†). The dipsticks contain an array of 180 small hourglass-shaped through-holes (from hereon "wells")

<sup>&</sup>lt;sup>a</sup> KTH Royal Institute of Technology, Stockholm, Sweden. E-mail: iseri@kth.se, wouter@kth.se

<sup>&</sup>lt;sup>b</sup> University of Antwerp, Antwerp, Belgium

 $<sup>\</sup>dagger$  Electronic supplementary information (ESI) available. See DOI: 10.1039/ d0lc00793e

#### Communication

Method

Table 1 Overview of common dia

gnostic tests for urinary tract infections. NA means not assessed					
Validity	Time to result	Required operator skills	Complexity/cost		
Gold standard	24-72 h	Trained	High		
>90%	24 h	Trained	High		

View Article Online

Lab on a Chip

Urine culture <sup>19</sup>	Gold standard	24-72 h	Trained	High
Flexicult (culture based) <sup>17,20</sup>	>90%	24 h	Trained	High
Nitride dipsticks <sup>23,24</sup>	40-70%	2-5 min	Untrained	Low
Dip-slide (culture based) <sup>21,22</sup>	73%	18–24 h	Trained	Low
Nucleic acid based <sup>5,25</sup>	NA	1–5 h	Trained	High
Immunoassays <sup>26–29</sup>	NA	20-120 min	Trained	High
Single cell imaging based <sup>9</sup>	NA	20-60 min	Trained	High
Digital dipstick (this work)	NA	8–12 h	Untrained	Low

#### A. FABRICATION



Fig. 1 Digital dipstick fabrication and operation. A. Schematic of digital dipstick preparation: dipsticks with 180 holes are laser cut and drilled out of a 1 mm thick PMMA plate, placed on an aluminium mold (grey), filled with chromogenic agar (light yellow) at 50 °C, cooled to 20 °C and demolded. B. Schematic of digital dipstick operation: first, place the dipstick in sample liquid (yellow) containing bacteria (purple); second, insert the dipstick into the incubation box (orange), which removes excess urine on the dipstick surface and keeps the dipstick humid during incubation at 37 °C (>10 hours); third, count the holes that are colored (blue), which indicates the occurrence of bacterial growth.

(Fig. S1B<sup>†</sup>). Each well lodges chromogenic agar (from hereon "gel"), which contains reagents needed for bacteria detection. Each agar gel has a central cylindrical through-hole (from hereon "hole") with a designed 200 µm diameter and inner volume of 31.4 nL.

#### The dipstick operation

The dipstick operation is simple (Fig. 1B and ESI<sup>†</sup> Video S1). First, the dipstick is placed in the liquid sample, e.g., raw urine, for a few seconds to capture sample in the wells. Excess sample on the dipstick surface is swiped off by inserting the dipstick into an incubation box, thereby isolating the sample in the separate wells (ESI<sup>†</sup> Video S2). The box keeps the environment humid, protecting gels from desiccation during incubation at 37 °C. In positive wells (wells containing CFU), enzymatic reactions in the growing bacteria induce a color change in the reagents in the gel.<sup>30</sup> Bacterial growth increases the amount of enzyme and thus amplifies the color signal. Negative wells (wells without CFU)

remain uncolored. The color *versus* no color difference of the wells forms a digital signal, and simple counting of colored wells by the naked eye or cellphone camera allows quantifying the bacterial concentration *via* the most probable number method.<sup>31</sup>

# We developed a colorimetric bioassay for miniaturized samples to detect *E. coli*

An investigation of the commercially available chromogenic agar (ChromAgar orientation) revealed that miniaturizing the volumes of gel during bacterial culture results in a decrease of bacterial growth and color signal (Fig. S2†). We speculate, this results from the decreased volume-per-surface ratio, resulting in a reduced availability of nutrient and reaction reagent. After testing gels with specific mixtures of signal-generating substance (ChromAgar orientation and RedGal) and nutrition broths (Rapid Coliform ChromoSelect Broth) (Fig. S3†), we selected the mixture resulting in the most vivid color change during culturing *E. coli* for further experiments.

#### Sensor response

We monitored dipsticks during 24 h of incubation at 37 °C (Fig. 2A). As a sample, we used phosphate-buffered saline (PBS) spiked with *E. coli* in a ten-fold dilution of concentrations from  $5 \times 10^7$  to  $5 \times 10^1$  (n = 3). We imaged each dipstick at the start of incubation and at different time points after. Images were processed with a home-built algorithm using ImageJ and MATLAB. The time-to-color-change scales with the bacterial culture concentration. The highest tested concentration,  $5 \times 10^7$  CFU ml<sup>-1</sup>, resulted in a detectable color change after 4 h, whereas concentrations

below 10<sup>3</sup> CFU ml<sup>-1</sup> needed at least 10 h of incubation. Fig. 2B shows the positive hole fraction (the number of positive holes divided by the total number of holes) of dipsticks after 12 h of incubation, which is enough to quantify samples at the clinically relevant range for UTI diagnosis. We used the most probable number method<sup>31</sup> for curve fitting, and a best-fit analysis indicates that an average of 92 nl urine was captured in each well (between 22 nl and 155 nl with 95% CI). We defined the limit of detection (LoD) to be 253 CFU ml<sup>-1</sup>, which is the highest concentration in the sensor response curve, with 95% CI band, for which the device has the highest probability of having 1 positive well. We defined the upper limit of quantification (ULoQ) to be 33 500 CFU ml<sup>-1</sup>, which is the lowest concentration in the sensor response curve, with 95% CI band, for which the device has the highest probability of having 1 negative well. We also did colony counting of dilution series on agar plates to measure standard deviation in prepared samples (Fig. S4<sup>†</sup>).

#### Validation with spiked urine samples (Fig. 3)

We evaluated the digital dipstick using urine from a healthy donor spiked with *E. coli* type strain ATCC 11775. Digital dipsticks were dipped into urine containing a ten-fold dilution series of *E. coli* resulting in *E. coli* concentrations ranging from  $10^7$  to  $10^1$  (n = 3). Overnight incubation at 37 °C (>12 h) resulted in a clearly visible color change (blue) in CFU-positive wells and no color change in negative controls. The results indicate it is possible to differentiate bacteriuria from presumptive contamination in the clinically relevant concentration range for urinary tract infections (UTIs), where concentrations above  $10^3$  CFU ml<sup>-1</sup> in patients presenting with



Fig. 2 Characterization of digital dipsticks. A. Positive hole fraction vs. incubation time. B. Sensor response curve: positive hole fraction vs. *E. coli* concentration in PBS after 12 h incubation. Samples measured with the digital dipstick assay consisted of *E. coli* in phosphate-buffered saline (PBS) with a concentration between  $5 \times 10^7$  and  $5 \times 10^1$  CFU ml<sup>-1</sup>. Solid round dots indicate individual measurements; diamonds indicate average values (n = 3); vertical error bars indicate standard deviation; horizontal error bars indicate the standard deviation of the *E. coli* concentration in PBS samples; colored dashed lines connecting averages are plotted for eye guidance only. The curved solid blue line is the best-fit digital assay response curve using the most probable number method,<sup>31</sup> indicating that an average of 92 nl urine was captured in each well. The blue band indicates the 95% confidence interval ( $2\sigma$ ) around the sensor response curve.



**Fig. 3** *E. coli* detection in urine. A. Photographs of digital dipsticks after their use to quantify *E. coli* spiked in urine of a healthy donor with concentrations of  $10^2$ ,  $10^3$ ,  $10^4$ ,  $10^5$  and  $10^6$  CFU ml<sup>-1</sup>. Width of each dipstick is 18 mm. B. Quantification with digital dipstick readouts: positive hole fraction vs. the concentration of *E. coli* spiked in urine of a healthy donor (n = 3). Solid round dots indicate individual measurements; diamonds indicate average values (n = 3); vertical lines indicate standard deviation; vertical dashed lines indicate the clinically relevant concentration thresholds for UTIs.

symptoms are indicative of infection (symptomatic bacteriuria) and where concentrations above  $10^5$  CFU ml<sup>-1</sup> in the absence of symptoms are indicative of asymptomatic bacteriuria.<sup>32,33</sup>

#### Validation with clinical urine samples

We compared digital dipstick assays to standard urine culture on agar plates (gold standard) for 28 samples of patients with confirmed *E. coli* infection (Fig. 4). 25 out of 28 results were correlated, whereas 3 out of 28 dipstick measurements gave false-negative results, meaning 89% sensitivity to detect *E. coli* in clinical samples. The digital dipsticks confirmed that one sample had a concentration below the threshold for symptomatic bacteriuria, three samples had concentrations between the thresholds for symptomatic and asymptomatic bacteriuria, and 21 samples had concentrations above the ULoQ. Two samples with concentration between  $10^5$  CFU ml<sup>-1</sup> and one sample with concentration between  $10^4$  and  $10^5$  CFU ml<sup>-1</sup> resulted in a false negative in which two readouts were below LoD.

We also tested digital dipsticks with PBS samples spiked with species of *Klebsiella pneumoniae* and *Proteus mirabilis*. Although the assay is currently only optimized for *E. coli* detection, the dipsticks detect these species, showing the potential multiplexing capability of the test for different bacterial species (Fig. S5†). We foresee that our device can further be developed and optimised for bacteria identification or antibiotic susceptibility testing in future studies.

# Experimental

#### Fabrication of digital dipstick frames

Digital dipstick frames are fabricated out of 1 mm-thick polymethyl methacrylate (PMMA) plates using a  $CO_2$  laser

(VLS 2.30, Universal Laser Systems, Vienna, Austria). To get hourglass-shaped arrays, we first engraved, with 80% laser power and 50% laser speed one side of the plate with a 10 by 18 array of circles with 1.2 mm diameter and 1.5 mm center



**Fig. 4** Digital dipstick validation with clinical urine samples. The *E. coli* concentration measured using digital dipsticks *vs.* the concentration determined using plate culturing (gold standard). Small solid round dots indicate individual measurements; the large round dot indicates the result of 21 measurements; vertical error bars for measured values between LoD and ULoQ indicate the 95% CI; vertical error bars for measured values below LoD or above ULoQ indicate the interpretation outside the quantitation range; horizontal error bars indicate the clinical interpretation of plate culture test; diagonal solid lines indicate the boundaries for false positives and false negatives; dashed lines indicate the LoD and ULoQ; dotted lines indicate the threshold for symptomatic and asymptomatic bacteriuria.

to center spacing. After that, the other side of the plate was engraved using the same laser settings after manual alignment using premade alignment markers in the laser platform. After engraving, circular holes of 0.5 mm diameter were drilled in each engraved well, and the frame was cut into a dipstick format using 90% laser power and 10% laser speed. A dipstick frame with engraved/drilled holes is shown in Fig. S1.<sup>†</sup>

# Development of the colorimetric bioassay for miniaturized samples

When we used the commercially available chromogenic agar (CHROMagar<sup>TM</sup> orientation, Paris, France), we observed that the color signal generated by bacterial was insufficient in miniaturized samples. We, therefore, investigated the effect of miniaturization. Chromogenic agar was prepared according to the user manual (32 g L<sup>-1</sup> powder in DI water) and cast on glass slides in layers of various thicknesses (1 mm, 600 µm, 300 µm and 150 µm), defined with the help of spacers. *Escherichia coli* (*E. coli*) type strain ATCC 11775 was grown on the cast agar inside a humid incubator at 37 °C for 24 hours. The red colour indicating growth decreased with decreasing thickness of the agar (Fig. S2†).

To improve signal generation, we modified the chromogenic agar by adding more nutrients or/and chromogenic substrates (Fig. S3<sup>†</sup>). Different combinations of chromogenic agar with added Miller's Luria broth (LB broth (Miller), Sigma-Aldrich, MO, USA) in concentrations of 20 g  $L^{-1}$  and 40 g  $L^{-1}$  and with additional Reg-Gal (6-chloro-3indolyl-β-D-galactopyranoside, Sigma-Aldrich, MO, USA) with concentrations of 60  $\mu$ g ml<sup>-1</sup> and 120  $\mu$ g ml<sup>-1</sup>, which generate red color in the presence of E. coli-specific enzyme β-galactosidase, were tested for detecting growth of E. coli. Additionally, chromogenic agar modified with E. coli specific chromogenic broth (Rapid Coliform ChromoSelect Broth, Sigma-Aldrich, MO, USA) with concentrations of 16 g  $L^{-1}$  and 32 g  $L^{-1}$ , which generate blue colour in the presence of E. coli-specific enzyme β-galactosidase, were tested for E. coli growth. After 24 hours humid incubation at 37 °C, samples in small glass vials (Fig. S3A†) and 300 µm thick agar pieces (Fig. S3B<sup>†</sup>) show that the most vivid color change is provided by chromogenic agar prepared with chromogenic broth concentration of 32 g L<sup>-1</sup>. We used this modified chromogenic agar gel for the rest of our experiments in this study.

#### Fabrication of aluminum mold

An aluminum mold was machined by computer numerical control (CNC) milling in the shape of an array of cylindrical pins of 200  $\mu$ m diameter, 1.1 mm height and 1.5 mm spacing, *i.e.*, to fit into the laser-cut holes of dipsticks. After milling, the mold was coated with Teflon (Teflon<sup>TM</sup> AF 1600, Teflon, DE, USA) to prevent adhesion of gel during molding.

#### Fabrication of digital dipsticks

Chromogenic agar was prepared as described above and kept at elevated temperature (>50 °C) to keep it in liquid form. The aluminum mold was placed on top of a hotplate (80 °C), and few drops of warm chromogenic agar were poured inside the mold. Then, the prefabricated digital dipstick frames were placed on the aluminium mold. Liquid agar spontaneously filled the wells of the frame. To proper filling in every well, a few drops of chromogenic agar were added on top of the dipsticks. After that, the aluminum mold was placed on a cooling block at 0 °C, resulting in the gelation of the chromogenic agar. After gelation, the dipsticks were demolded and the top and the bottom part of the dipstick were cleaned from excess agar by scalpel. Fabricated dipsticks were stored at 4 °C to prevent desiccation and used within the same day of fabrication. Devices with similar content of agarose have a reported shelf life of 4 months at 4 °C.<sup>34</sup> Other studies show that the shelf-life of agar plates can be extended to 24 months at the room temperature by vacuum-sealed packaging under controlled atmosphere.35 Thus, we believe that long-term storage (>6 months) of digital dipsticks is possible with modified atmosphere packaging, a method frequently used in food packaging.<sup>36</sup>

#### Fabrication of incubation boxes

Incubation boxes were fabricated in a FormLabs Form 3 printer using the standard clear transparent resin.

#### Characterization of digital dipsticks for E. coli detection

Ten-fold dilution series of *E. coli* type strain ATCC 11775 in phosphate-buffered saline (PBS) (Gibco<sup>TM</sup> Dulbecco's phosphate-buffered saline, Thermo Fisher Scientific, Göteborg, Sweden) were prepared with concentrations between 5 × 10<sup>7</sup> CFU ml<sup>-1</sup> and 5 × 10<sup>1</sup> CFU ml<sup>-1</sup>. Plate-counting experiments were performed on diluted samples to evaluate the variation in colony count at the different dilutions.

Plate counting experiments. For the concentrations of  $5 \times$  $10^5$  CFU ml<sup>-1</sup> (dilution #3),  $5 \times 10^4$  CFU ml<sup>-1</sup> (dilution #4), and 5 × 10<sup>3</sup> CFU ml<sup>-1</sup> (dilution #5), 1  $\mu$ l of solution (n = 3) was pipetted on an agar plate, as shown in Fig. S4<sup>†</sup> (top agar plate with 90 mm diameter, segmented by red marker). For 5 ×  $10^2$  CFU ml<sup>-1</sup> (dilution #6), 10 µl of solution (n = 3) was pipetted on an agar plate, as shown in Fig. S4<sup>†</sup> (bottom left agar plate with 55 mm diameter, segmented by the red marker). For 5  $\times$  10<sup>1</sup> CFU ml<sup>-1</sup> (dilution #7), 100 µl of solution (n = 1) was pipetted on an agar plate, as shown in Fig. S4<sup>†</sup> (bottom right agar plate with 55 mm diameter). Plates were incubated at 37 °C for 24 hours. The whole set of experiments was repeated five times with different dilution series. Results of plate counting experiments are shown in Table S1.<sup>†</sup> For the concentration of  $5 \times 10^5$  CFU ml<sup>-1</sup>, we were not able to count individual colonies. The average and the standard deviation data of these experiments were used as the x-axis error bars in Fig. 2B.

Digital dipstick characterization experiments. Triplicate digital dipstick assays were performed to measure the concentration of each sample of ten-fold diluted E. coli in PBS. Clean PBS was used as a negative control. Operation of the digital dipsticks is simple, as demonstrated in Fig. 1 and ESI† Video S1. First, dipsticks were placed in sample liquids for 1-2 seconds. Subsequently, they were individually placed in an incubation box which swipes off the excess sample liquid on the dipstick surfaces and isolates the sample in the separate wells (ESI<sup>†</sup> Video S2). The incubation box was placed in an incubation chamber at 37 °C for 24 hours. A photograph of each dipstick was taken at hours 0, 1, 2, 3, 4, 5, 6, 7, 8, 9, 10, 12, and 24. Each photograph was analysed for colour change in each well of the dipsticks by using ImageI and Matlab. A detailed description of the image processing can be found below, and the Matlab code is provided in ESI<sup>†</sup> File S1. Positive hole fraction data from image processing is provided in Table S2<sup>†</sup> and plotted in Fig. 2. We did not observe any colour change in negative control samples.

**Data analysis.** We assume that: one CFU in a single well of the dipstick is sufficient to change the color of the well; each well captures the same amount of liquid, and; having a positive well does not change the probability of having another positive well, *i.e.*, the probability of capturing CFU in wells is independent, and there is no cross-contamination between wells. Under these assumptions, a Poisson distribution can be used to characterize the device statistically. The most probable positive hole fraction, MPF, for a given sample concentration, *c* and sample volume captured in individual wells, *v*, is

$$MPF = 1 - e^{-c\nu}.$$
 (1)

To find the volume of sample liquid captured in individual wells,  $\nu$ , we compared the results at hour 12 with the theoretical predictions (eqn (1)) for sampling volumes values  $\nu = 20$  nl, 21 nl, ..., 300 nl and used the least square error method. The best fit indicates a well sampling volume,  $\nu = 92$  nl. The dipstick response for  $\nu = 92$  nl well volumes is plotted in Fig. 2B, and this value was used for further analysis.

We used eqn (1) to calculate the concentration, c, from a measured positive hole fraction, m, as

$$c = -\ln(m-1)/\nu. \tag{2}$$

The average,  $\mu$ , and standard deviation,  $\sigma$ , were calculated for each triplicate measurement (n = 3)

$$\mu = \sum_{i=1}^{n} c_i / n, \qquad (3)$$

$$\sigma = \frac{1}{n} \sqrt{\sum_{i=1}^{n} \left(c_i - \mu\right)^2}.$$
(4)

We further calculated the average positive hole fraction, M, the standard deviation *y*-axis error bar values  $\pm \sigma$ , and the 95% confidence interval (CI) band  $M \pm 2\sigma$ , using eqn (1):

$$M = 1 - e^{-\nu.\mu},$$
 (5)

$$M \pm \sigma = 1 - e^{-\nu(\mu \pm \sigma)}, \tag{6}$$

$$M \pm 2\sigma = 1 - e^{-\nu . (\mu \pm 2\sigma)}$$
. (7)

*M*, and  $M \pm \sigma$  values are calculated and plotted in Fig. 2B for each concentration.  $M + 2\sigma$  and,  $M - 2\sigma$ , values are used to determine the 95% CI band by using the previously used algorithm for finding the best fit value of 92 nl. For  $M + 2\sigma$ , and  $M - 2\sigma$ , 155 nl and 22 nl of sampling volume in each well, respectively, give the best fitting curve for the 95% CI, which is plotted in Fig. 2B.

#### Automated positive hole fraction counting

The following automated image processing protocol was developed to determine the positive hole fraction in a dipstick automatically. Each photograph had three dipsticks measuring the same concentration next to each other. Images were first opened in ImageJ, and center coordinates of top left, top right and bottom left wells of each dipstick were manually extracted to a text file for further analysis. Photographs and text files of coordinates were imported to Matlab, and the following steps were performed to detect the blue color change and count the number of positive wells in each dipstick for every time point:

1. The diameter of wells and the position of each well in pixel units for each dipstick was calculated based on the coordinate data taken from ImageJ.

2. Rectangular areas on the top and the bottom of the dipsticks were selected as a region of interest (ROI) and white balancing of the photo was processed based on these ROIs.

3. Based on the calculated position and diameters of wells, 180 circular ROI were selected for the dipstick.

4. For detecting blue color, the blue channel value of each pixel in each ROI was compared to the red channel of each pixel. Five different thresholds were set for five different brightness levels such that the difference between blue and red channel was higher for pixels to be defined as "blue".

5. To define the well as "positive", the number of "blue" pixels should be larger than the defined threshold.

6. Finally, the number of positive wells is counted and exported for further data analysis (Table S2<sup>†</sup>).

The Matlab script is provided in the ESI<sup>†</sup> File S1.

#### Validation with spiked urine samples

Ten-fold dilution series of *E. coli* type strain ATCC 11775 concentrations in between  $10^7$  CFU ml<sup>-1</sup> and  $10^1$  CFU ml<sup>-1</sup> were prepared with a urine sample of a healthy donor. Triplicate digital dipstick assays were used to measure the concentration of each sample, including a negative control, which was the untreated urine of the healthy donor. Positive hole fractions are plotted in Fig. 3B, and the data is available in Table S3.† Gel pieces in some wells were lost during the

fabrication and operation process, and these wells were excluded during the hole fraction calculations.

#### Trials to detect different species with digital dipsticks

Although our test was optimised only for *E. coli* detection, the chromogenic agar we used contains different substances to detect different species. We performed tests on five other common bacteria species that cause urinary tract infections (*Klebsiella pneumoniae*, *Proteus mirabilis*, *Enterococcus faecium*, *Staphylococcus aureus*, *Acinetobacter baumannii*) to investigate the potential multiplexing capability of digital dipsticks. We used concentrations of these bacteria  $>5 \times 10^4$  CFU ml<sup>-1</sup> in PBS and ran duplicate assays. We observed color change for *Proteus mirabilis* (dark yellow) and *Klebsiella pneumoniae* (light blue/turquoise) (Fig. S5†). We did not discern colour changes for the other three species. We did not further investigate these results since our study focuses on the detection of *E. coli*, which cause more than 85% of uncomplicated UTIs.

#### Validation with clinical samples

Urine samples were daily collected from the Antwerp University Hospital (UZA), and stored for more than 24 hours at 4 °C. All samples collected were verified for E. coli infection, and only E. coli positive samples were withheld. Also, samples confirmed to have multi-bacterial infection were excluded from this study. Each sample was placed on a blood agar plate and a cystine-lactose-electrolyte deficient (CLED) agar plate using the EddyJet spiral plating tool (50 µl log mode). 300 µl of samples were pipetted to EddyJet cups and were manually fed to the machine. Plates were incubated at 37 °C and colonies were counted according to the EddyJet manual instructions. Counting was performed if the growth was in the detection range of spiral plating, which is between  $2 \times 10^2$  and  $4 \times 10^5$  CFU ml<sup>-1</sup>. Since digital dipsticks have similar detection range, we did not plate dilution series from samples for a wider range of quantification.

Simultaneously, tests with digital dipstick assays were performed on the patient samples. Due to the restricted volume of the samples, for each sample, 0.5 ml was pipetted on a disposable petri dish and the dipstick was swiped over the urine sample from both sides to ensure urine was captured in the holes of the dipstick. Dipsticks were placed in a humid chamber at 37 °C overnight (12–15 h). The number of positive holes was counted to obtain the concentration and results are plotted in Fig. 4.

# Conclusions

To conclude, we presented the digital dipstick, which miniaturizes and digitizes culture-based diagnosis. The test simplicity and rapidity indicate the potential for a lower test cost and shorter turn-around time than today's gold standard, plate culturing. Sampling and compartmentalization neither require sophisticated external equipment nor technical skills. Moreover, the digital bioassay format provides straightforward quantification, for which readout can be performed *via* cell phone, enabling automation and easy communication of results between user and healthcare settings. We demonstrated the clinically relevant detection range for PoC diagnosis of UTI. The significance of uniquely combining microbial culture and digital assays in a simple dipstick format lays in its potential for self-diagnosing uncomplicated UTI at home or for systematically testing for asymptomatic bacteriuria in elderly patients.<sup>37</sup>

# Ethics statement

Ethical approval for the study was received from the ethics committee of the Antwerp University Hospital (UZA) (approval no. 19/22/281).

# Author contributions

EI and WvdW designed the device. EI fabricated prototypes, performed experiments for proof of concept, device characterization and did image processing. For the clinical trials, EI, MB, PM, HG and WvdW designed the experiments, EI fabricated the samples, EI and MB performed the experiments and EI analyzed the data. EI and WvdW wrote the manuscript with the input from all authors.

# Conflicts of interest

Authors declare no competing interests.

# Acknowledgements

Emre Iseri and Michael Biggel were funded through the European Union's Horizon 2020 research and innovation programme under the Marie Sklodowska-Curie grant agreement no. 675412. We thank Herland Lab at KTH for their donation of the Escherichia Coli strain ATCC 11775. We thank University Hospital Antwerp for the donation of patient urine samples.

# Notes and references

- 1 Y. Zhang and H. Noji, Anal. Chem., 2017, 89, 92-101.
- 2 B. Vogelstein and K. W. Kinzler, *Proc. Natl. Acad. Sci. U. S. A.*, 1999, **96**, 9236–9241.
- 3 E. C. Yeh, C. C. Fu, L. Hu, R. Thakur, J. Feng and L. P. Lee, *Sci. Adv.*, 2017, 3, e1501645.
- 4 F. Schuler, F. Schwemmer, M. Trotter, S. Wadle, R. Zengerle,
  F. Von Stetten and N. Paust, *Lab Chip*, 2015, 15, 2759–2766.
- 5 N. G. Schoepp, T. S. Schlappi, M. S. Curtis, S. S. Butkovich, S. Miller, R. M. Humphries and R. F. Ismagilov, *Sci. Transl. Med.*, 2017, 9(410), DOI: 10.1126/scitranslmed.aal3693.
- 6 D. M. Rissin, C. W. Kan, T. G. Campbell, S. C. Howes, D. R. Fournier, L. Song, T. Piech, P. P. Patel, L. Chang, A. J. Rivnak, E. P. Ferrell, J. D. Randall, G. K. Provuncher, D. R. Walt and D. C. Duffy, *Nat. Biotechnol.*, 2010, 28, 595–599.

#### Communication

- 7 S. H. Kim, S. Iwai, S. Araki, S. Sakakihara, R. Iino and H. Noji, *Lab Chip*, 2012, **12**, 4986–4991.
- 8 O. Scheler, N. Pacocha, P. R. Debski, A. Ruszczak, T. S. Kaminski and P. Garstecki, *Lab Chip*, 2017, **17**, 1980–1987.
- 9 Ö. Baltekin, A. Boucharin, E. Tano, D. I. Andersson and J. Elf, *Proc. Natl. Acad. Sci. U. S. A.*, 2017, **114**, 9170–9175.
- C. M. Hindson, J. R. Chevillet, H. A. Briggs, E. N. Gallichotte, I. K. Ruf, B. J. Hindson, R. L. Vessella and M. Tewari, *Nat. Methods*, 2013, **10**, 1003–1005.
- 11 E. A. Ottesen, W. H. Jong, S. R. Quake and J. R. Leadbetter, *Science*, 2006, **314**, 1464–1467.
- 12 K. A. Heyries, C. Tropini, M. Vaninsberghe, C. Doolin, O. I. Petriv, A. Singhal, K. Leung, C. B. Hughesman and C. L. Hansen, *Nat. Methods*, 2011, **8**, 649–651.
- 13 F. Shen, W. Du, J. E. Kreutz, A. Fok and R. F. Ismagilov, *Lab Chip*, 2010, **10**, 2666–2672.
- 14 F. Schuler, M. Trotter, M. Geltman, F. Schwemmer, S. Wadle, E. Domínguez-Garrido, M. López, C. Cervera-Acedo, P. Santibáñez, F. Von Stetten, R. Zengerle and N. Paust, *Lab Chip*, 2016, **16**, 208–216.
- 15 K. Leirs, P. Tewari Kumar, D. Decrop, E. Pérez-Ruiz, P. Leblebici, B. Van Kelst, G. Compernolle, H. Meeuws, L. Van Wesenbeeck, O. Lagatie, L. Stuyver, A. Gils, J. Lammertyn and D. Spasic, *Anal. Chem.*, 2016, **88**, 8450–8458.
- 16 Q. Zhu, Y. Gao, B. Yu, H. Ren, L. Qiu, S. Han, W. Jin, Q. Jin and Y. Mu, *Lab Chip*, 2012, **12**, 4755–4763.
- 17 C. C. Butler, N. A. Francis, E. Thomas-Jones, M. Longo, M. Wootton, C. Llor, P. Little, M. Moore, J. Bates, T. Pickles, N. Kirby, D. Gillespie, K. Rumsby, C. Brugman, M. Gal, K. Hood and T. Verheij, *Br. J. Gen. Pract.*, 2018, **68**, e268–e278.
- 18 L. Brookes-Howell, E. Thomas-Jones, J. Bates, M.-J. Bekkers, C. Brugman, E. Coulman, N. Francis, K. Hashmi, K. Hood, N. Kirby, C. Llor, P. Little, M. Moore, A. Moragas, K. Rumsby, T. Verheij and C. Butler, *BJGP Open*, 2019, 3(2), DOI: 10.3399/bjgpopen18X101630.
- 19 G. Schmiemann, E. Kniehl, K. Gebhardt, M. M. Matejczyk and E. Hummers-Pradier, *Dtsch. Arztebl. Int.*, 2010, **107**, 361–367.

- 20 M. Blom, T. L. Sørensen, F. Espersen and N. Frimodt-Møller, Scand. J. Infect. Dis., 2002, 34, 430–435.
- 21 D. Guttmann and G. R. Naylor, BMJ, 1967, 3, 343-345.
- 22 R. Winkens, H. Nelissen-Arets and E. Stobberingh, *Fam. Pract.*, 2003, 20(4), 410–412.
- 23 M. S. Lachs, Ann. Intern. Med., 1992, 117, 135.
- 24 M. Juthani-Mehta, M. Tinetti, E. Perrelli, V. Towle and V. Quagliarello, *Infect. Control Hosp. Epidemiol.*, 2007, 28, 889–891.
- 25 L. E. Lehmann, S. Hauser, T. Malinka, S. Klaschik, S. U. Weber, J. C. Schewe, F. Stüber and M. Book, *PLoS One*, 2011, 6(2), DOI: 10.1371/journal.pone.0017146.
- 26 C. M. Shih, C. L. Chang, M. Y. Hsu, J. Y. Lin, C. M. Kuan, H. K. Wang, C. Te Huang, M. C. Chung, K. C. Huang, C. E. Hsu, C. Y. Wang, Y. C. Shen and C. M. Cheng, *Talanta*, 2015, **145**, 2–5.
- 27 I. P. Alves and N. M. Reis, *Biosens. Bioelectron.*, 2019, 145, 111624.
- 28 G. Kokkinis, B. Plochberger, S. Cardoso, F. Keplinger and I. Giouroudi, *Lab Chip*, 2016, 16, 1261–1271.
- 29 E. Eltzov and R. S. Marks, Anal. Chem., 2016, 88, 6441-6449.
- 30 Z. Samra, M. Heifetz, J. Talmor, E. Bain and J. Bahar, J. Clin. Microbiol., 1998, 36, 990–994.
- 31 W. G. Cochran, *Biometrics*, 1950, 6, 105.
- 32 G. Bonkat, R. Pickard, R. Bartoletti, T. Cai, F. Bruyere, S. E. Geerlings, B. Köves, F. Wagenlehner, A. Pilatz, B. Pradere and R. Veeratterapillay, *EAU Guidelines on Urological Infections*, 2018.
- 33 M. L. Wilson and L. Gaido, *Clin. Infect. Dis.*, 2004, 38, 1150-1158.
- 34 P. Yagupsky, M. Rider and N. Peled, Eur. J. Clin. Microbiol. Infect. Dis., 2000, 19, 694–698.
- 35 J. V. Rogers, J. Microbiol. Exp., 2015, 2(7), DOI: 10.15406/ jmen.2015.02.00075.
- 36 J. M. Farber, J. Food Prot., 1991, 54, 58-70.
- 37 M. Biggel, S. Heytens, K. Latour, R. Bruyndonckx, H. Goossens and P. Moons, *BMC Geriatr.*, 2019, **19**, 170.