



Cite this: *Phys. Chem. Chem. Phys.*,
2022, 24, 19481

Received 14th April 2022,
Accepted 30th July 2022

DOI: 10.1039/d2cp01731h

rsc.li/pccp

Sensitive multi-species photoacoustic gas detection based on mid-infrared supercontinuum source and miniature multipass cell

Tommi Mikkonen,^a Tuomas Hieta,^b Goëry Genty^a and Juha Toivonen^{*a}

We report multipass broadband photoacoustic spectroscopy of trace gases in the mid-infrared. The measurement principle of the sensor relies on supercontinuum-based Fourier transform photoacoustic spectroscopy (FT-PAS), in which a scanning interferometer modulates the intensity of a mid-infrared supercontinuum light source and a cantilever microphone is employed for sensitive photoacoustic detection. With a custom-built external Herriott cell, the supercontinuum beam propagates ten times through a miniature and acoustically non-resonant gas cell. The performance of the FT-PAS system is demonstrated by measuring the fundamental C–H stretch bands of various hydrocarbons. A noise equivalent detection limit of 11 ppb is obtained for methane (40 s averaging time, 15 $\mu\text{W cm}^{-1}$ incident power spectral density, 4 cm^{-1} resolution), which is an improvement by a factor of 12 compared to the best previous FT-PAS systems. High linearity and good stability of the sensor provide reliable identification of individual species from a gas mixture with strong spectral overlap, laying the foundation for sensitive and selective multi-species detection in small sample volumes.

1 Introduction

Photoacoustic spectroscopy (PAS) is an extensively used technique for trace gas sensing due its high sensitivity and low gas consumption.^{1,2} These properties result from a unique operation principle converting the detection from optical into acoustic domain. Modulated incident radiation induces periodic absorption in a closed gas cell, which leads to periodic pressure variations *via* molecular thermal relaxation. The pressure waves proportional to the gas concentration are typically detected with a capacitive microphone. More recently, detection schemes utilizing pressure transducers with enhanced sensitivity such as piezoelectric quartz tuning forks and optically read cantilevers have been demonstrated.^{3–5} The detection of multiple species with high selectivity requires a spectrally broadband system that can be implemented using a Fourier transform photoacoustic spectroscopy (FT-PAS) approach.^{6,7} In FT-PAS, a scanning interferometer modulates each wavenumber of a broadband light source at a distinct acoustic frequency, after which all the pressure waves at these frequencies are simultaneously detected in an acoustically non-resonant gas cell. Broadband thermal emitters are typically used as the light source in FT-PAS, but recently supercontinuum (SC) sources

and frequency combs have been shown to yield significantly better performance.^{8–12}

The spatial coherence of these light sources enables efficient coupling to an optical cavity, which amplifies optical power inside the gas cell resulting in a proportional enhancement in the photoacoustic (PA) signal and sensitivity. Although non-resonant multipass cavities provide less amplification than resonant cavities,^{13,14} they are often preferred due to simpler operation, better mid-infrared (mid-IR) availability and wider spectral bandwidth, which is especially crucial for broadband systems. In single-frequency PAS, various types of multipass cavities have been reported, including simple zigzag configurations,^{12,15–17} circular cells,¹⁸ retroreflectors¹⁹ and Herriott cells^{20–24} that are especially attractive for conventional cylindrical gas cells. The reflectors in most miniature multipass systems are internal,^{15–18,24} *i.e.* in contact with the gas medium, to eliminate optical power losses from windows, and the background signal from these mirrors is minimized by ways of wavelength modulation^{17,18,24} or acoustic resonances,^{15,16} which however are not applicable to FT-PAS.

Here, we report the first broadband (spectral bandwidth over 100 cm^{-1}) multipass PAS system in the mid-IR by employing an external Herriott-type mirror configuration. Radiation from a mid-IR SC source is guided ten times through a miniature gas cell equipped with a cantilever microphone that has been shown to perform exceptionally well in broadband PA detection due to its sensitivity over a wide frequency range.²⁵ The multipass system provides a sixfold signal enhancement in a

^a Photonics Laboratory, Physics Unit, Tampere University, FI-33014 Tampere, Finland. E-mail: juha.toivonen@tuni.fi

^b Gasera Ltd., Lemminkäisenkatu 59, FI-20520 Turku, Finland



bandwidth of 400 cm^{-1} around 3000 cm^{-1} , which results in an order of magnitude improvement in sensitivity for FT-PAS systems and detection limits for hydrocarbons in the low parts per billion (ppb) level. We also study the stability of the system and demonstrate its high selectivity showing great promise for multi-species trace gas detection.

2 Experimental

The experimental setup is illustrated in Fig. 1. A home-built SC source is generated by injecting sub-nanosecond pulses with 70 kHz repetition rate into a combination of optical fibers, yielding a spectrum that extends from 1547 nm (6464 cm^{-1}) to 3700 nm (2700 cm^{-1}).^{8,26} The SC light is collimated into a beam of 2 mm diameter using a reflective fiber collimator, and then spectrally filtered. A long-pass filter with a cut-off at 2989 nm (3346 cm^{-1} , Northumbria Optical Coatings SLWP-2989) removes optical power from spectral regions not under interest and thus suppresses excess absorption from water (H_2O) molecules in the gas cell. As we have shown previously,⁹ eliminating unutilized absorption decreases the noise level significantly, as the intensity noise of the SC is coupled to the PA signal *via* light absorption by gas molecules. The filtered SC light with 21 mW of optical power is directed into a Fourier transform spectrometer (FTS, Bruker IRCube), whose maximum spectral resolution is 1 cm^{-1} . The optical path difference scanning velocity of the FTS is about 1 mm s^{-1} , which leads to modulation frequencies of 270–340 Hz at the wavenumber range of the filtered SC. At such low modulation frequencies the relaxation processes do not limit the sensitivity of the system and the response is similar for different gas mixtures.

About 30% of the incoming light is transmitted through the FTS and focused towards a commercial PA analyser (Gasera PA201) using a parabolic mirror ($f = 20\text{ cm}$). Within the PA analyser, a miniature non-resonant gas cell is 95 mm long,

4 mm in diameter, about 7 ml by total volume and sealed with anti-reflection coated BaF_2 windows (Thorlabs WG00530-E). Four mass flow controllers (Brooks Instruments 0154 and three pieces of Bronkhorst F-201CV) and a gas exchange system incorporated into the PA analyser control the gas flow into the cell with pressure and temperature set to 1 bar and 23°C , respectively.

The multipass configuration for this gas cell was designed through careful modelling of the optical system using Matlab. Specifically, we simulated the propagation of a Gaussian beam²⁷ in different Herriott configurations^{28,29} by varying the properties of the mirrors (radius of curvature and separation) and the input beam (radius of curvature and beam waist). The optimum number of passes was found to be ten using mirrors with 200 mm radius of curvature and separated by 138 mm. Such gold-plated spherical mirrors (LBP Optics) with diameters of 12.7 mm were placed outside the gas cell, where the SC radiation is guided off-axially through a small aperture (1 mm in diameter, displaced 1.1 mm from the mirror center) in the front mirror. A simulated intensity pattern at the front mirror is shown at the bottom-right corner in Fig. 1. A theoretical signal enhancement (dashed black line in Fig. 2(b)) for this multipass system is slightly wavenumber dependent around 6.2, considering the transmittance curve of the windows (96% on average) reported by the manufacturer and a constant 98% mirror reflectance.

An integrated microphone placed on the side of the gas cell is a thin (a few μm) silicon slab with an area of a few mm^2




Fig. 1 A schematic of the experimental setup: SC – supercontinuum; LPF – long-pass filter; FTS – Fourier transform spectrometer; BS – beam splitter; PM – parabolic mirror; PA – photoacoustic. The simulation on the bottom-right corner shows the theoretical intensity pattern at the front mirror. The smaller white ring represents the input aperture and the larger ring illustrates the gas cell.




Fig. 2 (a) Measured FT-PAS spectra of 36 ppm CH_4 in N_2 using the multipass enhancement (black line) and allowing only one beam pass through the gas cell (blue line). Both spectra were averaged over 40 s (eight scans) at a spectral resolution of 4 cm^{-1} . We note that the maximum resolution of the FTS (1 cm^{-1}) was not utilized here. In addition to the optical domain on the bottom, the corresponding modulation frequencies in the acoustic domain are shown on the top. (b) Signal enhancement using the multipass arrangement, evaluated from the measured spectra (solid blue line) and estimated from the window transmittance and mirror reflectance (dashed black line).



attached to its frame only from one side.³⁰ Pressure variations displace the free end of the cantilever whose position is monitored with a spatial-type interferometer.^{31,32} The microphone with a bandwidth of about 700 Hz records an interferogram containing information on all the excited acoustic waves whose frequencies are directly proportional to specific absorbed wavenumbers. A three-term Blackman–Harris apodization function is applied for the resampled (reference HeNe laser) interferogram, and the absorption spectrum is obtained by Fourier transform.

3 Results and discussion

Multipass enhancement and system performance

We measured the rovibrational absorption spectra of the fundamental C–H stretch band of methane (CH₄) around 3000 cm^{−1} to demonstrate multipass enhancement and to compare the performance of the system with previous studies. Fig. 2 shows the absorption spectrum of 36 parts per million (ppm, volume mixing ratio) of CH₄ in nitrogen (N₂), measured both with a single and ten beam-pass through the gas cell. Single pass was realized by blocking the beam before the back spherical mirror. Each spectrum was averaged over eight scans at a spectral resolution of 4 cm^{−1} resulting in a total averaging time of 40 s. A small H₂O residual was subtracted from both spectra using a methane-free spectral region with water absorption around 3200 cm^{−1} and a separate pure H₂O measurement. We calculated the detection enhancement arising from the multipass configuration by comparing the raw spectra with and without the multipass enhancement. This division shown in Fig. 2(b) (solid blue line) is in excellent agreement with the theoretical prediction of a signal enhancement of 5–8 at 2800–3200 cm^{−1}, especially in a region with a high signal-to-noise ratio.

For the multipass enhanced system, we estimated the limit of detection as $LOD = 3\sigma c/S$, where c is the gas concentration (36 ppm), S is the strongest spectral peak evaluated at 3017 cm^{−1} and σ is the noise level calculated from a separate 40 seconds long recording with only N₂ in the gas cell, as this represents the noise level at low CH₄ concentrations.⁹ Due to a small H₂O residual in the N₂ measurement, the noise level was calculated as the standard deviation of the baseline at 2200–2800 cm^{−1}. As will be shown later, this choice is justified by the flat noise level across the investigated wavenumber region. Based on this calculation, a detection limit of 32 ppb in 40 s for CH₄ was obtained.

The performance of our system is compared to that of previous FT-PAS studies in Table 1. Note that the power spectral density is reported without any multipass enhancement (a double-pass arrangement was used in ref. 8, 10 and 11). Significantly, our system shows a factor of 12 improvement in the time-normalized detection limit compared to the best previous FT-PAS system⁹ due to the multipass enhancement and a more sensitive PA analyser. In general, the differences between the LODs in Table 1 also result from the power spectral densities, spectral resolutions and from the procedures for estimating the noise level at the detection limit. The LOD

Table 1 A summary of recent FT-PAS experiments using supercontinuum (SC) or optical frequency comb (OFC) light sources in terms of CH₄ limit of detection (LOD, 3σ , 100 s), spectral resolution and incident power spectral density (PSD)

Source	LOD (ppb)	Res. (cm ^{−1})	PSD (μW cm ^{−1})	Ref.
SC	990	4	31	8
OFC	3300	0.033	21	10
OFC	660	0.02	240	11
SC	240	13	19	9
SC	20	4	18	This work

obtained here is comparable with other relatively simple broadband mid-IR techniques such as SC-based direct absorption spectroscopy and dual-comb spectroscopy with achieved CH₄ detection limits of 47 ppb and 2.4 ppm (for ~100 times weaker absorption cross sections), respectively.^{33,34} However, both of these techniques require two orders of magnitude larger sample volumes compared to our approach.

The performance of the system was further characterized by measuring the C–H stretch bands of 30 ppm ethane (C₂H₆) and 100 ppm ethene (C₂H₄) separately at 1 cm^{−1} spectral resolution. The raw spectra were averaged over 200 s (ten scans), corrected for H₂O absorption using the procedure described for CH₄ and normalized by the spectral envelope of the SC measured with a monochromator (Spectral Products DK480 1/2) and numerically corrected by the transmittance of the filter and the windows. The raw C₂H₄ spectrum and the H₂O residual are shown in Fig. 3(a). The corrected spectra in Fig. 3(b) are in excellent agreement with simulations based on the HITRAN database³⁵ and the instrument lineshape function. The increase in the baseline at the low wavenumber part of the C₂H₆ spectrum is due to low power spectral density of the SC in this region, which amplifies the background noise in the normalization process.

Calibration and noise characterization

We studied the effect of the gas concentration on the amplitude of the measured photoacoustic signal and the noise level. The concentration–amplitude dependence is used for calibration, which is always required in PAS, and the noise relation gives insight into the unique noise mechanism of FT-PAS studied in more detail elsewhere.⁹ Samples of C₂H₄ with varying concentration between 6 and 900 ppm were prepared by diluting C₂H₄ from a 1% gas bottle to N₂. The measured spectra for the different concentrations with 1 cm^{−1} resolution and averaged over 100 s (5 scans) are shown in Fig. 4(a) on a logarithmic scale. One can see how both the signal amplitude and noise level increase with the concentration. We also observe that within a single measurement, the noise level remains constant across the characterized wavenumber range, which justifies the use of the noise value in a specific region (red area in Fig. 4(a)) to assess the detection limit. For each spectrum, the amplitude was calculated from the strongest spectral peak of C₂H₄ after H₂O subtraction, and the noise level was estimated as the standard deviation of the raw PA amplitude in an absorption-free region at 2600–2800 cm^{−1}.






Fig. 3 (a) A raw FT-PAS spectrum of 30 ppm C_2H_6 and residual H_2O . (b) Absorption spectra of 30 ppm C_2H_6 and 100 ppm C_2H_4 at 1 cm^{-1} spectral resolution, obtained experimentally (left axis) by subtracting the H_2O residual from the raw spectra and normalizing by the envelope of the SC spectrum, and modelled (right axis, inverted) using the HITRAN database and the instrument lineshape function.

Fig. 4(b) and (c) illustrate quantitatively the dependence of the signal amplitude and noise as a function of C_2H_4 concentration up to 200 ppm. Uncertainties in the C_2H_4 concentrations result from dilution errors caused by the gas bottle uncertainty (2%) and the errors in the mass flow controllers (1–17% depending on the concentration and corresponding to absolute errors of 1–5 ppm). Uncertainties in the peak amplitudes and the noise levels are mainly caused by the fluctuation in the SC power, which was estimated from a long-term stability measurement. As can be expected, the amplitude of the acoustic signal scales linearly with the concentration ($R^2 = 0.9997$). At C_2H_4 concentrations beyond 200 ppm, the amplitude begins to saturate in accordance with the Beer–Lambert law. This observed amplitude nonlinearity at high concentrations was compared to a theoretical nonlinearity of the C_2H_4 amplitude to retrieve the effective optical path length. This simulation, based on a reference spectrum from the HITRAN database,³⁶ the Beer–Lambert law and the instrument lineshape function, yields an optical path length of $58 \pm 10\text{ cm}$ corresponding to a path length enhancement of 6.1 ± 1.1 and matching well to our achieved signal enhancement in the CH_4 demonstration.

The noise level exhibits similar dependence with the C_2H_4 concentration as the amplitude, because the intensity noise of the SC couples to the PA signal *via* light absorption by gas molecules. Higher concentration yields more absorption and therefore both higher amplitude and noise. This concentration dependent noise component is shown as a dotted green line in Fig. 4(c), extracted from the fit. However, the noise has another contribution from the thermal fluctuations of gas molecules




Fig. 4 (a) Measured absorption spectra of C_2H_4 at three concentrations and a background spectrum (no light entering the cell), all averaged over 100 s (5 scans) at a spectral resolution of 1 cm^{-1} . The spectra of C_2H_4 include H_2O residual, which is especially visible at the lowest concentration. Red area shows the noise region used in estimating the detection limits, which is justified by the flat noise level across the wavenumber region. (b) Peak amplitude and (c) noise level with respect to the applied C_2H_4 concentration. Noise has two contributions: Brownian noise (σ_B) from the thermal fluctuations of gas molecules corresponding to the background spectrum in (a), and a concentration dependent component from the light source (σ_{SC}).

(Brownian noise), which begins to dominate below 50 ppm, causing the noise level to approach the fundamental limit. This background noise level (dashed grey line in Fig. 4(c)) was calculated from the measurement with no light entering the cell (black line in Fig. 4(a)) and is consistent with the fitting result. As the noise at low concentrations equals the fundamental noise level, light absorption by the cell windows is expected to be small, which supports our choice for the external multipass system.

Long-term stability

To evaluate the long-term stability of the SC-based FT-PAS system, we measured 100 ppm of C_2H_4 every 40 s for 8320 s. Due to the limitations in the FTS, each 20 s scan followed by a 20 s break, cutting the averaging time in half. Furthermore, distorted measurements (30%) due to external vibrations were removed using Hampel's outlier test for average amplitudes in the frequency range of 710–870 Hz ($7030\text{--}8610\text{ cm}^{-1}$) resulting in an effective averaging time of 2880 s. Water was again subtracted from each spectrum by spectrally fitting pure H_2O and C_2H_4 spectra to each raw spectrum with an algorithm (linsolve function in Matlab) that solves a linear least squares






Fig. 5 (a) Separated average amplitudes of H_2O and C_2H_4 during the stability measurement, in which a 20 s scan was recorded every 40 s for 5760 s. Due to this dead time, in reality the drifts and variations in the amplitudes occur twice as slow as shown in the plot. (b) Calculated limit of detection (LOD) of C_2H_4 for different effective integration times. Water absorption was subtracted from each spectrum to account for the increase of H_2O concentration during the measurement.

problem using QR factorization with column pivoting. The result of this fitting routine is shown in Fig. 5(a), which displays the separated water and ethene traces during the measurement. A significant and slow increase in the H_2O amplitude is due to desorption from the cell walls. The C_2H_4 amplitude, on the other hand, stays constant on average but exhibits variations most likely due to SC instability. Corresponding variations are also visible in the H_2O trace. The relative precision of the system (4.8%) was estimated from the standard deviation of the ethene trace relative to its mean. The minimum precision of the system (at 1 cm^{-1} spectral resolution and for a single scan) is 200 ppb, set by the Brownian noise.

Using a water corrected ethene amplitude at 2980 cm^{-1} and noise from the raw spectrum (estimated as described previously), we calculated C_2H_4 LODs for various integration times. The result of this averaging process shown in Fig. 5(b) displays high stability of the system and a detection limit of 130 ppb in 2880 s for C_2H_4 . However, as previously discussed, the correct noise level for the detection limit calculation should be from the N_2 measurement, and thus a practical C_2H_4 detection limit is expected to be about two times smaller. The curve in Fig. 5(b) follows the inverse square root law well on average, as the main noise contributions (pulse-to-pulse SC fluctuations and Brownian noise) exhibit white noise characteristics in this spectral region. Deviations from the theoretical line are caused by the long-term variations of the SC power spectral density and the increase of H_2O inside the cell, which raises the noise level during the measurement.

Multi-species detection

We demonstrated the advantage of the high system linearity and spectral resolution with the wide spectral bandwidth in retrieving the concentrations of three species from a complex gas mixture. We first measured the reference spectra of ethane (30 ppm), ethene (100 ppm) and propane (20 ppm) by diluting each from a 1% gas bottle to N_2 . We then prepared a gas mixture containing 20 ppm ethane, 69 ppm ethene and 23 ppm propane, injected the mix to the gas cell and measured the spectrum. Water residual was removed from all spectra. The same least squares spectral fitting routine was used to match the weighted reference spectra to the spectrum of the gas mixture. The resulting scaled spectra of the individual species and the measured gas mixture spectrum are shown in Fig. 6(a) and the residual of the spectral fit in Fig. 6(b).

The concentration of each species in the gas mixture was calculated by multiplying the concentration of the reference measurement with the optimum weight from the fitting routine. Fig. 6(c) illustrates the correspondence between the applied and the retrieved concentrations and shows that the linear spectral fitting is able to retrieve concentrations within the measurement uncertainty. Uncertainties in the applied concentrations result from the dilutions errors caused by the four mass flow controllers, which produce uncertainty also in the retrieved concentrations (*via* reference measurements) together with the system precision estimated previously.




Fig. 6 (a) Measured FT-PAS spectra of a gas mixture (black) and individual species (blue, red and green), whose magnitudes were retrieved from a simple least squares fitting routine. (b) Residual of the spectral fitting. (c) Concentrations retrieved from the gas mixture (black dots), in comparison with the applied concentrations (red dashed lines). The spectral resolution and the total gas pressure of the measurement were 1 cm^{-1} and 1 bar, respectively.



4 Conclusions

A broadband mid-infrared optical gas sensing technique, Fourier transform photoacoustic spectroscopy (FT-PAS) with an external and compact Herriott-type multipass configuration has been demonstrated. Our FT-PAS system incorporates a home-built supercontinuum (SC) source with a high power spectral density and a cantilever enhanced photoacoustic analyser with a small sample volume. We characterized the fundamental C–H stretch bands of four hydrocarbons, methane, ethane, ethene and propane in nitrogen, at spectral resolutions of 1 cm^{-1} and 4 cm^{-1} with an excellent agreement with simulated absorption spectra. The Herriott cell for ten beam passes provides a sixfold broadband signal enhancement, resulting in a detection limit of 32 ppb in 40 s for methane. This is a factor of 12 lower compared to the best previous FT-PAS demonstrations and on the same scale with other broadband optical sensors with significantly higher gas consumption,^{33,34} when normalized by the averaging time. Moreover, we demonstrated high linearity and good stability of the system, which was shown to provide a simple and accurate spectral analysis of complex gas mixtures.

The detection limit of SC-based FT-PAS can be improved with more transparent windows and further optimization of the multipass arrangement. Ideally, the PA analyser should be redesigned with the multipass enhancement in mind, as increasing the diameter of the gas cell allows more beam passes but decreases the sensitivity of PA detection. The detection limit can be further lowered with commercial SC sources, which currently provide three times higher power spectral density around 3000 cm^{-1} compared to our device. Such sources would also enhance the precision of the system, which can be further improved by recording the optical power of the exiting radiation for each scan. As this light has propagated through a scanning Fourier transform spectrometer (FTS), an entire SC spectrum could also be continuously acquired for better spectral correction. The selectivity of the current system is limited by the maximum spectral resolution of the FTS. However, complex samples could benefit from higher spectral resolution provided by FTS with a longer optical path difference, which have been reported to reach isotopic selectivity with the cost of significantly increased footprint and lower sensitivity.¹¹ Dual-comb approach could decrease the size of a high-resolution instrument, but the only system reported so far has limited sensitivity and spectral bandwidth.¹² Lower gas pressures would further improve the selectivity of a high-resolution system.¹⁰

Conflicts of interest

There are no conflicts of interest to declare.

Acknowledgements

This work was supported by the Graduate School of Tampere University and the Academy of Finland Flagship Programme PREIN (320165).

References

- 1 F. J. Harren and S. M. Cristescu, *Encyclopedia of Analytical Chemistry: Applications, Theory and Instrumentation*, 2006, pp.1–29.
- 2 Z. Bozók, A. Pogány and G. Szabó, *Appl. Spectrosc. Rev.*, 2011, **46**, 1–37.
- 3 P. Patimisco, G. Scamarcio, F. K. Tittel and V. Spagnolo, *Sensors*, 2014, **14**, 6165–6206.
- 4 K. Wilcken and J. Kauppinen, *Appl. Spectrosc.*, 2003, **57**, 1087–1092.
- 5 T. Tomberg, M. Vainio, T. Hieta and L. Halonen, *Sci. Rep.*, 2018, **8**, 1–7.
- 6 G. Busse and B. Bullemer, *Infrared Phys.*, 1978, **18**, 631–634.
- 7 M. M. Farrow, R. K. Burnham and E. M. Eyring, *Appl. Phys. Lett.*, 1978, **33**, 735–737.
- 8 T. Mikkonen, C. Amiot, A. Aalto, K. Patokoski, G. Genty and J. Toivonen, *Opt. Lett.*, 2018, **43**, 5094–5097.
- 9 T. Mikkonen, Z. Eslami, G. Genty and J. Toivonen, *Opt. Lett.*, 2022, **47**, 1713–1716.
- 10 I. Sadiq, T. Mikkonen, M. Vainio, J. Toivonen and A. Foltynowicz, *Phys. Chem. Chem. Phys.*, 2018, **20**, 27849–27855.
- 11 J. Karhu, T. Tomberg, F. S. Vieira, G. Genoud, V. Hänninen, M. Vainio, M. Metsälä, T. Hieta, S. Bell and L. Halonen, *Opt. Lett.*, 2019, **44**, 1142–1145.
- 12 T. Wildi, T. Voumard, V. Brasch, G. Yilmaz and T. Herr, *Nat. Commun.*, 2020, **11**, 1–6.
- 13 S. Borri, P. Patimisco, I. Galli, D. Mazzotti, G. Giusfredi, N. Akikusa, M. Yamanishi, G. Scamarcio, P. De Natale and V. Spagnolo, *Appl. Phys. Lett.*, 2014, **104**, 091114.
- 14 T. Tomberg, T. Hieta, M. Vainio and L. Halonen, *Analyst*, 2019, **144**, 2291–2296.
- 15 A. Miklós, S.-C. Pei and A. Kung, *Appl. Opt.*, 2006, **45**, 2529–2534.
- 16 J. Saarela, J. Sand, T. Sorvajärvi, A. Manninen and J. Toivonen, *Sensors*, 2010, **10**, 5294–5307.
- 17 K. Chen, B. Zhang, S. Liu, F. Jin, M. Guo, Y. Chen and Q. Yu, *Sens. Actuators, A*, 2019, **290**, 119–124.
- 18 A. Manninen, B. Tuzson, H. Looser, Y. Bonetti and L. Emmenegger, *Appl. Phys. B: Lasers Opt.*, 2012, **109**, 461–466.
- 19 Y. Ma, S. Qiao, Y. He, Y. Li, Z. Zhang, X. Yu and F. K. Tittel, *Opt. Express*, 2019, **27**, 14163–14172.
- 20 C. Hornberger, M. König, S. Rai and W. Demtröder, *Chem. Phys.*, 1995, **190**, 171–177.
- 21 M. Nägele and M. Sigris, *Appl. Phys. B: Lasers Opt.*, 2000, **70**, 895–901.
- 22 M. Nägele, D. Hofstetter, J. Faist and M. W. Sigris, *Analytical Sciences/Supplements Proceedings of 11th International Conference of Photoacoustic and Photothermal Phenomena*, 2002, pp. s497–s499.
- 23 K. Krzempek, A. Hudzikowski, A. Gluszek, G. Dudzik, K. Abramski, G. Wysocki and M. Nikodem, *Appl. Phys. B: Lasers Opt.*, 2018, **124**, 74.
- 24 B. Zhang, K. Chen, Y. Chen, B. Yang, M. Guo, H. Deng, F. Ma, F. Zhu, Z. Gong, W. Peng and Q. Yu, *Opt. Express*, 2020, **28**, 6618–6630.
- 25 C. B. Hirschmann, J. Uotila, S. Ojala, J. Tenhunen and R. L. Keiski, *Appl. Spectrosc.*, 2010, **64**, 293–297.



- 26 C. Amiot, A. Aalto, P. Ryczkowski, J. Toivonen and G. Genty, *Appl. Phys. Lett.*, 2017, **111**, 061103.
- 27 A. A. Tovar and L. W. Casperson, *J. Opt. Soc. Am. A*, 1995, **12**, 1522–1533.
- 28 C. G. Tarsitano and C. R. Webster, *Appl. Opt.*, 2007, **46**, 6923–6935.
- 29 J. Altmann, R. Baumgart and C. Weitkamp, *Appl. Opt.*, 1981, **20**, 995–999.
- 30 P. Sievilä, N. Chekurov, J. Raittila and I. Tittonen, *Sens. Actuators, A*, 2013, **190**, 90–95.
- 31 J. Kauppinen, K. Wilcken, I. Kauppinen and V. Koskinen, *Microchem. J.*, 2004, **76**, 151–159.
- 32 T. Kuusela and J. Kauppinen, *Appl. Spectrosc. Rev.*, 2007, **42**, 443–474.
- 33 K. E. Jahromi, M. Nematollahi, Q. Pan, M. A. Abbas, S. M. Cristescu, F. J. Harren and A. Khodabakhsh, *Opt. Express*, 2020, **28**, 26091–26101.
- 34 L. A. Sterczewski, J. Westberg, M. Bagheri, C. Frez, I. Vurgaftman, C. L. Canedy, W. W. Bewley, C. D. Merritt, C. S. Kim, M. Kim, J. R. Meyer and G. Wysocki, *Opt. Lett.*, 2019, **44**, 2113–2116.
- 35 C. Hill, I. E. Gordon, R. V. Kochanov, L. Barrett, J. S. Wilzewski and L. S. Rothman, *J. Quant. Spectrosc. Radiat. Transfer*, 2016, **177**, 4–14.
- 36 L. Rothman, I. Gordon, Y. Babikov, A. Barbe, D. C. Benner, P. Bernath, M. Birk, L. Bizzocchi, V. Boudon, L. Brown, A. Campargue, K. Chance, E. Cohen, L. Coudert, V. Devi, B. Drouin, A. Fayt, J.-M. Flaud, R. Gamache, J. Harrison, J.-M. Hartmann, C. Hill, J. Hodges, D. Jacquemart, A. Jolly, J. Lamouroux, R. L. Roy, G. Li, D. Long, O. Lyulin, C. Mackie, S. Massie, S. Mikhailenko, H. Müller, O. Naumenko, A. Nikitin, J. Orphal, V. Perevalov, A. Perrin, E. Polovtseva, C. Richard, M. Smith, E. Starikova, K. Sung, S. Tashkun, J. Tennyson, G. Toon, V. Tyuterev and G. Wagner, *J. Quant. Spectrosc. Radiat. Transfer*, 2013, **130**, 4–50.

