


 Cite this: *RSC Adv.*, 2022, **12**, 6918

Antimicrobial peptide GL13K immobilized onto SLA-treated titanium by silanization: antibacterial effect against methicillin-resistant *Staphylococcus aureus* (MRSA)†

 Yusang Li, ^a Ruiying Chen, ^d Fushi Wang, ^{ac} Xinjie Cai ^{*ab} and Yining Wang ^{*ab}

Infection is the main reason for implant failure, and the incidence of drug-resistant bacterial infection has increased in recent years. Further, methicillin-resistant *Staphylococcus aureus* (MRSA)-related implant infection has become a serious worldwide threat. New strategies, other than antibiotics, to tackle drug-resistance, are of high clinical significance. Antimicrobial peptides show clear superiority over conventional antibiotics in inhibiting drug-resistant bacteria. In the present study, we combined the antimicrobial peptide, GL13K, with sandblasting and acid-etching (SLA)-treated titanium using a silane coupling agent. Field emission scanning electron microscopy images showed the morphology of the coating. Attenuated total reflectance Fourier transform infrared spectroscopy and X-ray photoelectron spectroscopy results confirmed loading of GL13K, and the hydrophilicity of the SLA-GL13K coating was evaluated by water contact angle analysis. The releasing study of samples showed that the coating has a sustained releasing profile. SLA-GL13K coating exhibited strong contact- and release-killing abilities against MRSA, *E. coli*, and *S. aureus*. Meanwhile, Cell Counting Kit 8 analysis and examination of cell morphology demonstrated that the SLA-GL13K coating had good cytocompatibility at antibacterial concentrations. Overall, all these results suggest that SLA-GL13K coating can be successfully fabricated using silanization, and is a promising candidate for controlling MRSA-induced implant-related infection.

Received 28th June 2021
 Accepted 22nd February 2022
 DOI: 10.1039/d1ra04974g
rsc.li/rsc-advances

Introduction

The superior biological and mechanical properties of titanium for long-term rehabilitation mean that it is among the most popular metallic materials used for biomedical applications.¹ Titanium implants have been widely-used in various fields, including oral implantology, hip or knee arthroplasty, and bone fracture fixation,² however, implant infection has become one of the most common and severe complications of biomaterials. The infection rate of orthopedic surgery is 2–5%,³ rising to approximately 30% after internal fixation of an open fracture.⁴ Implant-related infections can strongly impede osteogenesis

and osseointegration, leading to the need for complex revision procedures, implant failure, or even osteomyelitis.^{5,6}

In addition, drug-resistant bacteria have emerged as a severe global healthcare problem.⁷ Methicillin-resistant *Staphylococcus aureus* (MRSA) is one of the most important and common drug-resistant bacteria in the clinic. Approximately 25–32% of implant infections after fracture fixation are related to MRSA.⁷ MRSA can synthesize extracellular polymeric substances to form biofilms, which strongly impede penetration of systemically administered antibiotics and inhibition of bacterial growth.⁸ MRSA can also produce β-lactamase enzymes, which can hydrolyze antibiotic molecules, and penicillin-binding proteins to reduce the binding affinities of antibiotics.⁹ Relative to normal infections, treatment of patients with MRSA-related infections is more difficult, and the prognosis is significantly worse, due to high recurrence rates and increased sequelae.^{10,11}

Recent studies have focused on local delivery of antibacterial agents, which can maximize *in situ* antibacterial effects while avoiding potential systemic toxicity and drug resistance.¹² Antimicrobial peptides (AMPs) have been introduced into titanium surface modification as effective agents against colonization of microorganisms these years. AMPs are positively-charged bioactive molecules with a few hydrophobic

^aThe State Key Laboratory Breeding Base of Basic Science of Stomatology (Hubei-MOST) and Key Laboratory of Oral Biomedicine Ministry of Education, School and Hospital of Stomatology, Wuhan University, Wuhan, China. E-mail: xinjie.cai@whu.edu.cn; wang.yn@whu.edu.cn

^bHospital of Stomatology Wuhan University, Department of Prosthodontics, Wuhan, China

^cHospital of Stomatology Wuhan University, Department of Cariology and Endodontics, Wuhan, China

^dShanghai Ninth People's Hospital, Shanghai Jiao Tong University School of Medicine, College of Stomatology, Shanghai Jiao Tong University, Department of Implant Dentistry, Shanghai, China

† Electronic supplementary information (ESI) available. See DOI: [10.1039/d1ra04974g](https://doi.org/10.1039/d1ra04974g)



residues.¹³ As a constituent of the biological innate immune system with broad-spectrum activity, AMPs inhibit bacteria, fungi, and viruses with low levels of host cytotoxicity and bacterial resistance.¹⁴ AMPs can induce electrostatic interactions with anionic phospholipids in bacterial cell membranes, resulting in membrane rupture and cell death;¹⁵ this unique mechanism decreases the likelihood of bacterial resistance.¹⁵ Among various AMPs, the antimicrobial peptide, GL13K, is a 13-amino-acid peptide modified from the human parotid secretory protein (PSP; BPIFA2) with good biocompatibility.¹³ Similar to other AMPs, GL13K could interact with bacterial membranes in different ways and then leading to its rupture,^{16,17} and could also inhibit bacteria from forming biofilm.¹⁸ Various findings support GL13K as an effective and safe antibacterial agent for titanium surface modification;^{13,19} however, there has been little research investigating the antibacterial ability of GL13K against drug-resistant bacteria.

Sandblasting and acid-etching (SLA) treatment has been applied for titanium surface modification in recent years. The SLA process creates a hydrophilic titanium surface with micron-scale roughness, which can increase osteoblast differentiation and bone formation.²⁰ SLA-treated implants have higher removal torque values and bone-to-implant contact (BIC) than untreated implants, and the roughened surface can also provide a greater contact area for binding of the antimicrobial peptides.^{21,22} However, the rough surface with a greater contact area also favors bacterial adhesion and biofilm formation.²³ For this reason, various antibacterial agents have been incorporated into SLA-treated titanium surfaces, while no research has combined GL13K and SLA in titanium surface modification to our knowledge. Therefore, we hypothesized that immobilization of GL13K onto SLA-treated titanium surfaces could simultaneously prohibit bacterial growth and favor osseointegration.

In the present study, we combined GL13K antimicrobial coating with SLA technology for titanium implant surface modification to prevent infection, particularly with drug-resistant bacteria (MRSA), in orthopedic and dentistry applications. Therefore, we covalently immobilized GL13K onto SLA-treated titanium surfaces using the silane coupling agent, 3-aminopropyltriethoxysilane (KH-550). The surface morphology, physicochemical properties, releasing profile, cellular responses, and antimicrobial activities of the coating were evaluated.

Materials and methods

Materials

The antimicrobial peptide, GL13K, was supplied by China Peptides Biotechnology Co., Ltd (Shanghai, China). Commercially pure titanium (grade 2, 99.6% purity) was supplied by Baoji Titanium Industry Co., Ltd (Shanxi, China). The MC3T3-E1 cell line (American Type Culture Collection [ATCC] catalog CRL-2594), *Staphylococcus aureus* (ATCC catalog 25923), *Escherichia coli* (ATCC catalog 25922), and MRSA (ATCC catalog 43300) were obtained from the American Type Culture Collection. All other chemical reagents were local products of analytical grade.

Sample preparation

Titanium substrate preparation. Titanium pieces were cut into disks (10 mm diameter, 1 mm thick). To obtain the SLA surface, working surfaces were coarse grit blasted with 0.25–0.50 mm corundum grit for 1 min, and then acid etched in hydrochloric acid/sulfuric acid (1 : 1) at 60 °C for 30 min. Next, disks were rinsed with deionized water, and ultrasonically cleaned twice with acetone, ethanol, and deionized water for 20 min each procedure.^{24,25}

Silanization. For activation, all samples were immersed in 5 M NaOH and were water bathed at 60 °C for 1 h.²⁶ Then, they were rinsed with deionized water and dried. The silane coupling agent, KH-550, was hydrolyzed using ethanol and water. Activated samples were immersed in KH-550 solution in a water bath at 37 °C for 40 min. After that, samples were dried at room temperature.

Immobilization of GL13K onto silanized SLA surfaces. Scheme 1 shows the preparation process of the coatings. GL13K was dissolved in distilled water at 125 and 500 µg ml⁻¹. Silanized samples were immersed into each peptide solution and then freeze-dried at -80 °C under vacuum. Samples used for antibacterial and cell culture analyses were prepared under aseptic condition. Untreated smooth titanium samples were used as controls, and silanized SLA treated samples are referred to as the SLA + SIL (SS) group. Samples with GL13K immobilized onto silanized SLA surfaces are referred to as SLA + SIL + 125 µg ml⁻¹ GL13K (SSG125) and SLA + SIL + 500 µg ml⁻¹ GL13K (SSG500) groups, according to the different concentrations of peptide.


Physicochemical characterization of coatings

Field emission scanning electron microscopy (Zeiss, Oberkochen, Germany) was used to analyze the surface structure of coatings, which were sputter-coated with gold before microscopic examination. Further, coating surface chemistry was investigated using attenuated total reflectance Fourier transform infrared spectroscopy (ATR-FTIR; Thermo Nicolet 5700, USA). Spectra were obtained at 45 °C at a spectral resolution of 4 cm⁻¹, with 32 scans every 90 s interval, using air as the reference spectrum. The chemical composition of the coatings was analyzed by X-ray photoelectron spectroscopy (XPS; ESCALAB 250Xi, Thermo Fisher Scientific, Britain) with an Al-K α X-ray source (1486.6 eV). The selected resolution for the general spectra was 100.0 eV of pass energy and 1.0 eV per step, while the resolution for narrow scans was 30.0 eV of pass energy and 0.1 eV per step. Binding energies were based on C 1s at 285.0 eV for calibration.

Water contact angles (WCA) of all samples were tested using the drop shape analysis system, CASTs3.0 (SL200KS, KINO Industry Co., Ltd, USA). Milli-Q water (1 µl) was used to conduct tests at room temperature. The contact angle (θ) was defined as the angle between the solid substrate and the liquid phase. The data were tracked and measured for 60 s and the test was repeated five times for each group.

GL13K release study

For the GL13K release study, each sample of SSG125 and SSG500 was immersed into 0.5 ml PBS solution and was incubated at



Scheme 1 Schematic of the sample preparation.

37 °C. At pre-determined time intervals, 0.5 ml of the solution was removed and the GL13K concentration was measured by Nano-drop 2000 spectrophotometer (Thermo Fisher Scientific, US) through absorbance at 205 nm (A205 method). After the removal, 0.5 ml fresh PBS was added into the system. An absorbance/GL13K concentration calibration curve was portrayed using a series of concentrations of GL13K peptide dissolved in PBS.

In vitro antibacterial study

Preparation of bacteria. MRSA was used to study antibacterial activity against drug-resistant bacteria. *Escherichia coli* and *Staphylococcus aureus* were used to study the activity of SLA-GL13K coating against common Gram-positive and Gram-negative bacteria. *E. coli* were cultured in Luria-Bertani broth, while *S. aureus* and MRSA were cultured in tryptic soy broth, at 37 °C. When cultures reached logarithmic phase, the concentration of each bacterial suspension was adjusted to 1×10^4 cells per ml.

Antibacterial activity assay by the spread plate method. To determine the antibacterial effect of GL13K coating, we performed contact and release antibacterial assays. Sterilized samples were placed into 24-well plates, with the working surface upward. Bacterial suspension (1 ml) was added into each well, and then cultured at 37 °C for 16 h.

To measure the release antibacterial capacity of samples, the OD value of the bacteria suspension after culturing was measured, and serial 10-fold dilutions were performed. Then, viable counts were estimated following the spread plate method. For contact antibacterial capacity, samples cultured with bacteria were removed with sterile forceps, gently washed with PBS three times, and ultrasonicated in 2 ml PBS for 5 min. The obtained suspension was then diluted and re-cultured on solid medium and viable counts were estimated.

Observation of bacterial biofilm by confocal laser scanning microscopy. After culture with *E. coli*, *S. aureus*, or MRSA for 24 h, samples were gently rinsed three times with PBS and transferred into fresh 12-well plates. Samples were stained by incubation in reagent containing 2 μ M calcein-AM and 1 μ M propidium iodide in PBS for 15 min. Stained samples were subsequently viewed by confocal laser scanning microscopy (CLSM; Carl Zeiss, Jena, Germany).

In vitro cell analysis

Cell culture and seeding. MC3T3-E1 cells were cultured in Alpha Minimal Essential Medium (α -MEM, HyClone, 22 Logan,

UT, USA) supplemented with 10% fetal bovine serum (Tianhang Biotechnology Co., Ltd, Zhejiang, China) and 1% penicillin-streptomycin solution (HyClone) in an incubator at 37 °C with a 5% CO₂ humidified atmosphere. Culture medium was refreshed every 3 days. When cells reached 80–90% confluence, they were trypsinized and counted using a Beckman–Coulter automated cell counter (Beckman–Coulter, CA, USA) before use in subsequent experiments.

Analysis of cell morphology and skeleton. To observe cell morphology, MC3T3-E1 cells were seeded at 1×10^5 cm⁻² and cultured on the surface of samples in 24-well plates. After culturing for 24 h, samples were gently washed three times with PBS and fixed with 4% paraformaldehyde for 15 min. Then, samples were washed with PBS, permeabilized with 0.5% Triton X-100 solution for 5 min, and incubated in 1% bovine serum albumin for 30 min, to block non-specific binding sites. Next, filamentous actins were stained with rhodamine-phalloidin (R-415 kit, Molecular Probes, Invitrogen, OR, USA) for 30 min and cell nuclei stained with 2-(4-amidinophenyl)-6-indolecarbamidine dihydrochloride (0.1 μ g ml⁻¹ in PBS) (Invitrogen, Basel, Switzerland) for 10 min in dark. After rinsing with PBS, samples were visualized using a fluorescence microscope equipped with a CCD camera.

Cell proliferation. Cell proliferation was quantified using a cell-counting kit 8 assay (CCK8; Dojindo Laboratories, Japan). MC3T3-E1 cells were seeded at 1×10^4 cells per cm² in 24-well plates, and cultured with extract liquid of samples for 1, 3, and 7 days. At each time point, culture plates were gently washed with PBS three times. Then, 400 μ l of α -MEM containing 10% CCK-8 solution was added into wells. After incubation at 37 °C for 1 h, supernatants were transferred to a 96-well plate. Absorbance was measured at 450 nm using an ELX808 Ultra Microplate Reader (Bio-Tek Instruments, Inc., USA). Optical density values were determined in at least triplicate and reflected the viable cell counts in each well.

Statistical analysis

Quantitative data are expressed as mean \pm standard deviation. Statistical analysis was carried out by one-way analysis of variance (ANOVA) and *post hoc* testing. The level of significance was set at $P < 0.05$.

Results and discussion

Physicochemical characterization of titanium disk coatings

Surface topography description. Compared with the surface of the untreated control group, SEM images of SLA-treated



samples showed a coarse surface, with micron-scale concave features produced by sandblasting and acid-etching (Fig. 1A and B).²⁷ The nano-roughness of an SLA surface can influence coating immobilization, cell adhesion, and surface properties.²⁸ After treatment with NaOH and silanization, needle-shaped submicron scales crystallites were visible, due to the formation of hydroxide (Fig. 1C). After immobilization of GL13K, crystallites in the SSG125 group were connected

by the organic component and some holes were also filled (Fig. 1D and I). In the SSG500 group, the surface morphology was typical of an organic layer. The organic coating wrapped the needle-shaped inorganic structures while revealing their shapes beneath (Fig. 1E and J). In SSG500 group the organic layer was clearly thicker and denser than that in SSG125 group, indicating increased peptide binding.




Fig. 1 SEM micrographs of titanium sample surfaces at 1k and 50k magnification. Smooth (A and F), SLA (B and G), SS (C and H), SSG125 (D and I), and SSG500 (E and J).



ATR-FTIR analysis. ATR-FTIR can measure the chemical bonding interactions between structures, hence provide information about the existence and absence of certain functional groups and the chemical structure of samples. The infrared light is passed through the ATR crystal and interact with the sample surface, then the total internal reflection forms an evanescent wave which extends into the sample. Being partially absorbed by the sample, the evanescent wave slightly attenuated, and the specific infrared spectrum was generated according to it. ATR-FTIR spectra of the samples are shown in Fig. 2A–E. In the SS group, the silane coating generated a strong absorption peak at 1000 cm^{-1} (Fig. 2C), which is the characteristic peak caused by the asymmetric stretching vibration of Si–O–Si, confirming the presence of silane coating on the SLA-treated titanium disks.²⁹ SSG125 spectra included characteristic peptide peaks of amide I–CONH stretching (1536 cm^{-1}) and amide II–NH₂ stretching (1625 cm^{-1}), indicating the presence of peptide on the titanium surface (Fig. 2D);³⁰ in the SSG500 group

these two peaks were clearly stronger than those in the SSG125 group, indicating that SSG500 group samples bound relatively more peptide than the SSG125 group (Fig. 2E). Spectra from both the smooth and SLA samples had no characteristic peaks in the range examined (Fig. 2A and B).

XPS. XPS is a surface-sensitive technique based on the photoelectric effect. By irradiating a sample with a beam of X-rays and measuring the kinetic energy of the emitted electrons. It can detect the elements of sample surface and obtain the spectra of it. XPS spectra are shown in Fig. 3. XPS analysis showed that the surfaces of the smooth and SLA-treated samples were primarily composed of Ti 2p and O 1s, while a small C 1s peak was visible, due to inevitable surface contamination with carbon (Fig. 3A and B). Increased intensity of the Si 2p peak confirmed the KH-550 coating of the silanized sample (Fig. 3C). Fig. 3D and E show increased N 1s and C 1s signals following treatment with GL13K, which demonstrates that the peptide was immobilized on the silanized titanium.




Fig. 2 FTIR spectra of sample surfaces. (A) Smooth, (B) SLA, (C) SS, (D) SSG125, (E) SSG500.




Fig. 3 XPS survey scans (0–700 eV) of sample surfaces. (A) Smooth, (B) SLA, (C) SS, (D) SSG125, (E) SSG500.



Further, the N 1s and C 1s signals were stronger in the SSG500 samples, indicating an increase in bound peptide. Notably, the Si 2p peak was also present in the SSG125 spectrum, while it was barely visible in the SSG500 spectrum. We speculate that the thicker and denser peptide coating in SSG500 samples shielded the KH-550 coating, while the thin peptide coating in SSG125 samples failed to completely cover the coating beneath. The results of XPS analysis were in accordance with the detected ATR-FTIR spectra.

WCA analysis. The results of WCA analysis are shown in Fig. 4. The WCA of the smooth titanium surface was $48.51^\circ \pm 1.221^\circ$ and decreased to $9.865^\circ \pm 0.4692^\circ$ following SLA treatment. As shown in SEM images, SLA treatment altered the titanium surface topography and visibly increased surface roughness, and the increased roughness was previously reported to result in a lower WCA in SLA-treated titanium.³¹ In addition to surface topography, surface chemistry is another vital element that influences surface wettability.³² In SLA group samples, the extensive hydroxylation/hydration of the oxide layer also contributed to increased surface hydrophilicity.³³ The hydroxylated oxide surface of SLA implants can enhance surface reactivity with surrounding ions, amino acids, and proteins, thus promoting implant osseointegration.³²

After silanization, the WCA increased to $75.4^\circ \pm 1.583^\circ$. KH-550 coating is proven to increase the hydrophobicity of materials because it reacts with hydroxyl groups on inorganic

surfaces, lowering the surface energy.³⁴ WCA values in the SLAS125 and SLAS500 groups were $91.64^\circ \pm 2.843^\circ$ and $109.7^\circ \pm 2.205^\circ$, respectively. In general, a solid surface is considered hydrophobic if the WCA is $>90^\circ$. Like most AMPs, GL13K peptide composition is rich in cationic and hydrophobic amino acids, including Ala, Ile, and Leu.³⁵ From SEM images, the superficial areas of SSG125 and SSG500 were visibly smaller than those in SLA samples; hence, both the amino acid sequence and the organic surface characteristics may contribute to the hydrophobicity of GL13K.

GL13K release study

The release profiles of SSG125 and SSG500 are shown in Fig. 5. Fig. 5A illustrates the standard curve of absorbance of GL13K in PBS at 205 nm ($y = 0.01554x - 0.2935$, $R^2 = 0.9972$). For SSG125, most of the peptide was released at the first 2 hours of incubation. Then a small amount of peptide (approximately 18%) was constantly released from the second hour to the twelfth hour. After that, the amount of released GL13K became negligible. For SSG500, approximately 68% of the peptide was rapidly released at the first 4 hours, then the speed of releasing gradually slowed down. A sustained release was observed for about 36 hours, followed by a negligible amount of releasing afterward.

It was speculated that during the sample preparation, a certain amount of GL13K peptide was covalently immobilized




Fig. 4 (A) The water contact angle (WCA) of the different sample surfaces. The SLA samples have hydrophilic surfaces, while the surfaces of SSG125 and SSG500 samples are hydrophobic. (B) Statistical analysis of the WCA results. (** $p < 0.01$, *** $p < 0.0001$).




Fig. 5 (A) Standard curve of GL13K in PBS ($y = 0.01554x - 0.2935$, $R^2 = 0.9972$). (B) Cumulative release profile of GL13K from SSG125 and SSG500 groups.

onto titanium surface through silane coupling agent, while another amount of peptide was deposited and adsorbed on the surface physically. As a result, there was a rapid release initially, and followed by a sustained release for a longer time. And the sustained release is more obvious in the SSG500 group than the SSG125 group, which might indicate that a higher amount of peptide was covalently bonded to SSG500 samples. The rapid initial release could ensure a functional high concentration locally at an early stage, and then the sustained releasing would help to maintain the biological function of the coating.³⁶

In vitro cellular study

Proliferation of MC3T3-E1 cells in the different experimental groups, determined using the CCK8 assay at 1, 4, and 7 days, are shown in Fig. 6. Cell counts in the SLA group were greater than those of the untreated smooth group at all three time points, with significant differences on days 4 and 7. On day 1, the SSG500 group had a significantly lower cell count than the control group. However, on days 4 and 7, the cell counts in the

SSG500 group were significantly higher than those of the control group. To conclude, the incorporation of GL13K did not affect the viability or proliferation of MC3T3-E1 cells.

The immunofluorescence images presented in Fig. 7 show the morphology and actin skeleton of cells cultured on different surfaces. Cells grown on the SLA surface showed cellular spreading, with bright and well-pronounced elongated lamellipodia and filopodia protruding from the cell edges (Fig. 7B). By comparison, cells on the smooth titanium surface exhibited irregular morphology with few visible filopodia, and the majority of cells were round-shaped and aggregated together (Fig. 7A). In silanized samples, cells also exhibited clear spreading and interactions, with filopodia extending outward (Fig. 7C). Cells cultured on SSG125 and SSG500 also showed a typical polygonal osteoblastic shape, with numerous membrane channels between adjacent cells (Fig. 7D and E). These tunneling nanotubes can mediate intercellular transfer of organelles, plasma membrane components, and cytoplasmic molecules.^{37,38} Fig. 7F shows the statistical analysis of total cell area of MC3T3-E1 cells cultured on different samples. The cell




Fig. 6 Proliferation of MC3T3-E1 cells culturing with the samples of Smooth, SLA, SS, SSG125, and SSG500 groups. (* $p < 0.05$, ** $p < 0.01$, *** $p < 0.005$, **** $p < 0.0001$).




Fig. 7 Fluorescence images of MC3T3-E1 cells cultured on smooth (A), SLA (B), SS (C), SSG125 (D), and SSG500 (E) samples. (F) Statistical graph of total cell area within fluorescence images. (* $p < 0.05$, ** $p < 0.01$).



area of SLA, SSG125, and SSG500 groups is significantly higher than that of the control group and no significant difference exists between any other two groups. This result is basically in consistent with the result of the CCK8 test. Combining the statistical cell area and proliferation results with the cell morphology in Fig. 7, it is indicated that SSG125 and SSG500 groups have no cytotoxic effect on MC3T3-E1 cells.

The nano-roughness surface of SLA-treated titanium samples has been proven to have desirable effects on osteoblasts.³⁹ In our study, the proliferation and morphology of MC3T3-E1 cells were also significantly better when grown on SLA samples relative to smooth samples, indicating a higher likelihood of stable osseointegration. Further, cells cultured on samples with SSG125 and SSG500 coatings showed favorable proliferation and viability characteristics, with clear cell stretching and interaction. These results indicate that SSG125 and SSG500 have good cytocompatibility. The main factor that has contributed to that is the large amounts of cholesterol and other sterols in animal cell membranes, which could help to inhibit antimicrobial peptide activity and stabilize membranes.^{40,41} Also, the membranes of animals are not heavily populated with negatively charged molecules such as phosphatidylglycerol, which also reduces the risk of targeting of positively charged AMPs.⁴²

In vitro antibacterial study

The results of contact and release antibacterial capacity experiments are shown in Fig. 8 and 9. Notably, in results of both the contact and release antibacterial assays, the number of bacteria in the SLA group was higher than that in the smooth group, with significant differences for *E. coli* and MRSA. By releasing GL13K to bacterial suspensions, bacteria numbers in SSG125 samples

reduced by 82.82%, 80.14%, and 43.64% for the *E. coli*, *S. aureus*, and MRSA groups, while in the SSG500 samples bacteria numbers were decreased by 100.00%, 100.00%, and 79.09% in the *E. coli*, *S. aureus*, and MRSA groups, respectively, compared with SLA samples. Regarding contact-killing capacity, the SSG125 sample significantly reduced the bacteria number by 98.60%, 98.91%, and 73.26% in the *E. coli*, *S. aureus*, and MRSA groups, while those in the SSG500 sample were decreased by 100.00%, 100.00%, and 88.37%, respectively, compared with SLA samples. These data demonstrate that GL13K-coated samples exhibit both contact and release antibacterial activity against these three bacteria significantly. Antibacterial and cytotoxicity results of samples with higher concentrations (SSG1000, SSG2000) are shown in Fig. S2 and S3 in the ESL.[†]

The live and dead staining could help to identify the living and dead bacteria based on the color changing. As shown in Fig. 10, by fluorescence microscopy, live bacteria with intact membranes appeared green, while dead bacteria with damaged membranes were stained red. The fluorescence intensity of bacterial membranes was high in smooth, SLA, and SS samples, indicating that a large number of bacteria had adhered to the surface. Further, imaging revealed almost no living *E. coli* and *S. aureus* on the surface of SSG125 and SSG500 samples. However, some living MRSA were present on the surface in SLAS125 samples and green spots were also visible on the surface of SLAS500 samples, consistent with the CFU data.

MRSA is one of the most important drug-resistance bacteria associated with implant infection, while *E. coli* and *S. aureus* are also common pathogenic bacteria in implant-related infection. Hence, we chose these three bacteria as representative to examine the antibacterial ability of SLA-GL13K coating against drug-resistant bacteria, Gram-positive bacteria, and Gram-




Fig. 8 The percentage reduction of *E. coli*, *S. aureus*, and MRSA in release and contact antibacterial experiment of different samples. (* $p < 0.05$, ** $p < 0.01$, *** $p < 0.005$, **** $p < 0.0001$).






Fig. 9 (A) Optical images of bacterial colonies of *E. coli*, *S. aureus*, and MRSA from the release antibacterial experiment. (B) Optical images of bacterial colonies of *E. coli*, *S. aureus*, and MRSA from the contact antibacterial experiment.

negative bacteria. To our knowledge, this study is the first investigation of the antibacterial properties of GL13K against MRSA.

In addition to promoting the adhesion of osteoblasts, SLA-treated samples can also promote bacterial adhesion;²³ however, after coating with GL13K, samples presented significant antibacterial activity and good cytocompatibility. In the high concentration (SSG500) group, *E. coli* and *S. aureus* were completely killed and no cytotoxicity to MC3T3-E1 cells was detected. Hence, for common pathogenic bacteria such as *E. coli* and *S. aureus*, SLA-GL13K coating showed a significant antibacterial effect. Notably, as for MRSA, GL13K peptides also

have significant antibacterial activity, although the effect was weaker compared to the effect against *E. coli* and *S. aureus*.

Although AMPs are a diverse group of bioactive molecules, most of them share common properties that related to the antibacterial activity: positive charge, hydrophobicity, and amphiphilicity.⁴³ As a result, the antibacterial mechanism of most AMPs is basically identical and GL13K is not an exception.⁴⁴ Like most of the AMPs, GL13K is believed to kill bacteria by interacting with the cell membranes.⁴⁵ As a small molecule, GL13K could easily pass through the cell wall of bacteria as the peptidoglycan mesh is relatively porous.⁴⁶ As we know, bacterial membranes are negatively charged.²³ In contrast, most of the






Fig. 10 Fluorescent images of biofilm formation on sample surfaces obtained by CLSM after live and dead staining. Scale bar = 50 μ m.

AMPs are cationic peptides,⁴⁷ so the electrostatic attraction between AMPs and bacterial membranes would ensure the attachment of AMPs, therefore providing the chance for AMPs to interact with the membranes and cause the death of the bacteria.⁴⁸ Based on that, several models were proposed. For example, barrel-stave model, toroidal-pore model, carpet model and aggregate model are the most studied and widely accepted ones.^{49,50} Although having minor differences, these models all share the same key point: membrane disruption.

This mechanism is different from that of antibacterial compounds such as β -lactam antibiotics, which usually target a specific site on the cell wall, cell membrane, nucleic acid, or proteins.⁵¹ Aiming at it, some bacteria have developed antibiotic resistance by genetic mutation or the acquisition of exogenous resistance genes from other bacteria, which highly decreased the antibacterial efficacy of many antibiotics.⁵² However, the biochemically simple and efficient mechanism of AMPs like GL13K is predicted to prevent bacteria from developing drug resistance, while no research of the antibacterial properties of GL13K against MRSA was reported before.⁵³

Our study has proved the significant antibacterial effect of GL13K against MRSA, while the effect is relatively weaker compared to that against *E. coli* and *S. aureus*. The mutation of MRSA is complex. For example, by blocking murein hydrolysis in cell wall substrates, cell wall turnover and autolysis of MRSA are inhibited, which increases the thickness and density of the cell wall.⁵⁴ This mutation could partially prevent the AMPs from passing through the cell wall and interacting with the cell membranes, and thus protect MRSA from being disrupted by GL13K. This may explain the reason why the antibacterial effects of GL13K against MRSA were relatively inferior to those against *E. coli* and *S. aureus* in our study. Nevertheless, accumulating evidence shows that different AMPs may have synergistic potentiation,⁵⁵ and some AMPs can also synergize with

conventional antibiotics or metal particles, such as ceftriaxone, amoxicillin-clavulanate, linezolid, imipenem, silver nanoparticles (AgNPs), etc.⁵⁶ The synergy is proved to improve the antibacterial effects, while further reduce the likelihood of resistance.¹⁵ So, the combination application of SLA-GL13K coating and some other antibacterial agents might improve its efficacy and could be further studied. Also, as a 13-amino-acid peptide, GL13K has a short sequence. Evidence has revealed that short AMPs have better pharmacological profiles, better metabolic stability, and lower toxicity than long AMPs.⁵⁴ Therefore, the SLA-GL13K coating might be a promising agent for the control of MRSA-related infection. In addition, for samples as SSG125 and SSG500, GL13K was coated on the titanium surface through silanization and hence gained the property of sustained releasing after a rapid releasing stage. It could ensure an efficient high concentration locally in the early stage and might maintain the biological function through the sustained release.

Conclusions

In summary, the antimicrobial peptide, GL13K, was covalently conjugated onto SLA-treated titanium surfaces using the silane coupling agent, KH-550. The results of SEM, ATR-FTIR, XPS, and WCA confirmed successful immobilization of the peptide, and the release measurement of GL13K indicated a sustained-release property of the samples. SLA-GL13K coating exhibited antibacterial properties against *E. coli* and *S. aureus*, as well as significant antibacterial effects against drug-resistant MRSA. Further, the coating had a positive effect on osteoblast proliferation and adhesion *in vitro*. Based on these findings, and within the limitations of this study, we conclude that SLA-GL13K coating is a promising candidate for early-stage prevention of infections at implant sites; however, further *in*



vivo antibacterial and biocompatible evaluation will be necessary to examine the full potential of SLA-GL13K coatings for biomedical applications.

Conflicts of interest

There are no conflicts of interest to declare.

Acknowledgements

This work was financially supported by the National Natural Science Foundation of China (No. 81901063) and the Natural Science Foundation of Hubei Province (2017CFB183 and 2018CFB419).

Notes and references

- 1 C. N. Elias, J. H. C. Lima, R. Valiev and M. A. Meyers, *JOM*, 2008, **60**, 46–49.
- 2 M. Geetha, A. K. Singh, R. Asokamani and A. K. Gogia, *Prog. Mater. Sci.*, 2009, **54**, 397–425.
- 3 K. Mitsutake and K. Niwaya, *N. Engl. J. Med.*, 2004, **351**, 193–195.
- 4 A. Trampuz and W. Zimmerli, *Injury*, 2006, **37**(suppl 2), S59–S66.
- 5 K. Ma, X. Cai, Y. Zhou, Y. Wang and T. Jiang, *Macromol. Biosci.*, 2017, **17**, 1600130.
- 6 M. B. Zaid, R. J. O'Donnell, B. K. Potter and J. A. Forsberg, *J. Am. Acad. Orthop. Surg.*, 2019, **27**, e977–e985.
- 7 W. J. Metsemakers, N. Emanuel, O. Cohen, M. Reichart, I. Potapova, T. Schmid, D. Segal, M. Riool, P. H. Kwakman, L. de Boer, A. de Breij, P. H. Nibbering, R. G. Richards, S. A. Zaai and T. F. Moriarty, *J. Controlled Release*, 2015, **209**, 47–56.
- 8 Y. Zhuang, L. Ren, S. Zhang, X. Wei, K. Yang and K. Dai, *Acta Biomater.*, 2021, **119**, 472–484.
- 9 J. T. Murphy, R. Walshe and M. Devocelle, *J. Theor. Biol.*, 2008, **254**, 284–293.
- 10 V. Alt, K. Kirchhof, F. Seim, I. Hrubesch, K. S. Lips, H. Mannel, E. Domann and R. Schnettler, *Acta Biomater.*, 2014, **10**, 4518–4524.
- 11 B. Nie, T. Long, H. Ao, J. Zhou, T. Tang and B. Yue, *Antimicrob. Agents Chemother.*, 2017, **61**, e01766.
- 12 G. T. Boo, D. W. Grijpma, T. F. Moriarty, R. G. Richards and D. Eglin, *Biomaterials*, 2015, **52**, 113–125.
- 13 M. Abdolhosseini, S. R. Nandula, J. Song, H. Hirt and S. U. Gorr, *Peptides*, 2012, **35**, 231–238.
- 14 J. M. Agetos, A. Sánchez-Pérez, P. Calo-Mata and T. G. Villa, *Biochem. Pharmacol.*, 2017, **133**, 117–138.
- 15 R. K. Thapa, D. B. Diep and H. H. Tønnesen, *Acta Biomater.*, 2020, **103**, 52–67.
- 16 V. Balhara, R. Schmidt, S. U. Gorr and C. Dewolf, *Biochim. Biophys. Acta*, 2013, **1828**, 2193–2203.
- 17 T. Diniz and M. I. Tavares, *J. Appl. Polym. Sci.*, 2003, **90**, 2955–2958.
- 18 X. Chen, H. Hirt, Y. Li, S. U. Gorr and C. Aparicio, *PLoS One*, 2014, **9**, e111579.
- 19 L. Zhou, Z. Lin, J. Ding, W. Huang, J. Chen and D. Wu, *Colloids Surf. B*, 2017, **160**, 581–588.
- 20 D. Buser, N. Broggini, M. Wieland, R. K. Schenk, A. J. Denzer, D. L. Cochran, B. Hoffmann, A. Lussi and S. G. Steinemann, *J. Dent. Res.*, 2004, **83**, 529–533.
- 21 A. Piattelli, L. Manzon, A. Scarano, M. Paolantonio and M. Piattelli, *Int. J. Oral Maxillofac. Implants*, 1998, **13**, 805–810.
- 22 D. Buser, T. Nydegger, T. Oxland, D. L. Cochran, R. K. Schenk, H. P. Hirt, D. Snétivy and L. P. Nolte, *J. Biomed. Mater. Res.*, 1999, **45**, 75–83.
- 23 F. Hizal, N. Rungraeng, J. Lee, S. Jun, H. J. Busscher, H. C. van der Mei and C. H. Choi, *ACS Appl. Mater. Interfaces*, 2017, **9**, 12118–12129.
- 24 Y. Zhou, T. Jiang, M. Qian, X. Zhang, J. Wang, B. Shi, H. Xia, X. Cheng and Y. Wang, *Biomaterials*, 2008, **29**, 461–474.
- 25 T. Jiang, Z. Zhang, Y. Zhou, Y. Liu, Z. Wang, H. Tong, X. Shen and Y. Wang, *Biomacromolecules*, 2010, **11**, 1254–1260.
- 26 K. V. Holmberg, M. Abdolhosseini, Y. Li, X. Chen, S. U. Gorr and C. Aparicio, *Acta Biomater.*, 2013, **9**, 8224–8231.
- 27 J. Zhang, J. Liu, C. Wang, F. Chen, X. Wang and K. Lin, *Bioact. Mater.*, 2020, **5**, 9–16.
- 28 R. A. Gittens, T. McLachlan, R. Olivares-Navarrete, Y. Cai, S. Berner, R. Tannenbaum, Z. Schwartz, K. H. Sandhage and B. D. Boyan, *Biomaterials*, 2011, **32**, 3395–3403.
- 29 J. Li, K. Yu, K. Qian, H. Cao, X. Lu and J. Sun, *Nanoscale Res. Lett.*, 2014, **9**, 172.
- 30 K. Ma, L. Gong, X. Cai, P. Huang, J. Cai, D. Huang and T. Jiang, *Int. J. Nanomed.*, 2017, **12**, 3665–3679.
- 31 C. S. Chang, T. M. Lee, C. H. Chang and J. K. Liu, *Clin. Oral Implants Res.*, 2009, **20**, 1178–1184.
- 32 D. V. Kilpadi and J. E. Lemons, *J. Biomed. Mater. Res.*, 1994, **28**, 1419–1425.
- 33 M. Textor, C. Sittig, V. Frauchiger, S. Tosatti and D. M. Brunette, in *Titanium in Medicine: Material Science, Surface Science, Engineering, Biological Responses and Medical Applications*, Springer Berlin Heidelberg, Berlin, Heidelberg, 2001, pp. 171–230, DOI: 10.1007/978-3-642-56486-4_7.
- 34 J.-C. Liu, M.-J. Xu, T. Lai and B. Li, *Ind. Eng. Chem. Res.*, 2015, **54**, 9733–9741.
- 35 Z. Ye, A. C. Kobe, T. Sang and C. Aparicio, *Nanoscale*, 2020, **12**, 20767–20775.
- 36 F. Wang, P. Huang, D. Huang, Y. Hu, K. Ma, X. Cai and T. Jiang, *J. Mater. Chem. B*, 2018, **6**, 2304–2314.
- 37 A. Rustom, R. Saffrich, I. Markovic, P. Walther and H. H. Gerdes, *Science*, 2004, **303**, 1007–1010.
- 38 H. H. Gerdes and R. N. Carvalho, *Curr. Opin. Cell Biol.*, 2008, **20**, 470–475.
- 39 G. Zhao, A. L. Raines, M. Wieland, Z. Schwartz and B. D. Boyan, *Biomaterials*, 2007, **28**, 2821–2829.
- 40 T. H. Lee, K. N. Hall and M. I. Aguilar, *Curr. Top. Med. Chem.*, 2016, **16**, 25–39.
- 41 L. Zhou, Y. Lai, W. Huang, S. Huang, Z. Xu, J. Chen and D. Wu, *Colloids Surf. B*, 2015, **128**, 552–560.
- 42 R. Kundu, *ChemMedChem*, 2020, **15**, 1887–1896.



43 P. Bastos, F. Trindade, J. da Costa, R. Ferreira and R. Vitorino, *Med. Res. Rev.*, 2018, **38**, 101–146.

44 M. D. Manniello, A. Moretta, R. Salvia, C. Scieuzzo, D. Lucchetti, H. Vogel, A. Sgambato and P. Falabella, *Cell. Mol. Life Sci.*, 2021, **78**, 4259–4282.

45 J. Lei, L. Sun, S. Huang, C. Zhu, P. Li, J. He, V. Mackey, D. H. Coy and Q. He, *Am. J. Transl. Res.*, 2019, **11**, 3919–3931.

46 J. C. Espeche, M. Martinez, P. Maturana, A. Cutro, L. Semorile, P. C. Maffia and A. Hollmann, *Arch. Biochem. Biophys.*, 2020, **693**, 108549.

47 Y. Luo and Y. Song, *Int. J. Mol. Sci.*, 2021, **22**, 11401.

48 F. Savini, M. R. Loffredo, C. Troiano, S. Bobone, N. Malanovic, T. O. Eichmann, L. Caprio, V. C. Canale, Y. Park, M. L. Mangoni and L. Stella, *Biochim. Biophys. Acta, Biomembr.*, 2020, **1862**, 183291.

49 Y. Yan, Y. Li, Z. Zhang, X. Wang, Y. Niu, S. Zhang, W. Xu and C. Ren, *Colloids Surf., B*, 2021, **202**, 111682.

50 M. Magana, M. Pushpanathan, A. L. Santos, L. Leanse, M. Fernandez, A. Ioannidis, M. A. Giulianotti, Y. Apidianakis, S. Bradfute, A. L. Ferguson, A. Cherkasov, M. N. Seleem, C. Pinilla, C. de la Fuente-Nunez, T. Lazaridis, T. Dai, R. A. Houghten, R. E. W. Hancock and G. P. Tegos, *Lancet Infect. Dis.*, 2020, **20**, e216–e230.

51 M. Tajbakhsh, A. Karimi, A. Tohidpour, N. Abbasi, F. Fallah and M. M. Akhavan, *J. Microbiol.*, 2018, **56**, 128–137.

52 A. Giedraitienė, A. Vitkauskienė, R. Naginienė and A. Pavilonis, *Medicina*, 2011, **47**, 137–146.

53 B. Lazzaro, M. Zasloff and J. Rolff, *Science*, 2020, **368**, eaau5480.

54 K. Sieradzki and A. Tomasz, *J. Bacteriol.*, 2003, **185**, 7103–7110.

55 A. Moretta, C. Scieuzzo, A. M. Petrone, R. Salvia, M. D. Manniello, A. Franco, D. Lucchetti, A. Vassallo, H. Vogel, A. Sgambato and P. Falabella, *Front. Cell. Infect. Microbiol.*, 2021, **11**, 668632.

56 X. Liu, Z. Li, X. Wang, Y. Chen, F. Wu, K. Men, T. Xu, Y. Luo and L. Yang, *Int. J. Nanomed.*, 2016, **11**, 6781–6794.

