RSC Advances



PAPER

View Article Online
View Journal | View Issue



Cite this: RSC Adv., 2022, 12, 15694

Simple simultaneous determination of moxifloxacin and metronidazole in complex biological matrices†

Amira F. El-Yazbi, (10) *a Faten M. Aboukhalil, a Essam F. Khamis, a Mohammed A. W. Elkhatib, b Mahmoud A. El-Sayed and Rasha M. Youssef

A simple, sensitive and rapid RP-HPLC method is presented, for the first time, for the simultaneous determination of moxifloxacin hydrochloride and metronidazole in different biological fluids including saliva and plasma without any matrix interference. The separation was performed using ACN and phosphate buffer (30:70% v/v) as the mobile phase on a Zorbax Eclipse Plus-C18 column attached to a guard column. The method was validated according to the FDA guidelines for bioanalytical method validation and was successfully applied for simultaneous determination of the studied drugs in saliva and plasma samples. The good precision and selectivity of the developed method allow it to be used for routine therapeutic drug monitoring of such drugs and it presents a simple and sensitive analytical tool for performing versatile pharmacokinetics and bioavailability studies. A DAD detector is valuable to determine each drug at its maximum wavelength to ensure high sensitivity. Determination of such a combination in saliva introduces a quick and non-invasive alternative to blood analysis.

Received 13th March 2022 Accepted 17th May 2022

DOI: 10.1039/d2ra01631a

rsc.li/rsc-advances

1. Introduction

Moxifloxacin hydrochloride (MX) is a synthetic fluroquinolone antibiotic with bactericidal effects (Fig. 1a). It binds to DNA gyrase (a bacterial enzyme) and inhibits DNA replication. It is active against Gram-positive bacteria such as *Micrococcus luteus* and *Staphylococcus aureus*, in addition to aerobic Gram-negative bacteria such as *Actinobacter lwoffli* and *H. influenza*. It is available as tablets, ophthalmic solutions and infusion bags for treatment of skin infections and respiratory infections such as pneumonia and bronchitis.¹

Metronidazole (MT) is a nitroimidazole drug as shown in Fig. 1b. It inhibits the synthesis of nucleic acid as a result of nitroso radicals' formation. It is active against anaerobes and effective for treatment of amebiasis, giardiasis, trichomoniasis and sexually transmitted disease.²

MX has a reduced activity against anaerobic pathogens. Therefore, a combination of MX with an antimicrobial agent active against anaerobes, such as MT, is of extreme importance for the treatment of mixed aerobic/anaerobic infections. The

combination therapy of MX and MT is effective for patients having intra-abdominal abscesses.³

The therapeutic effect of the drug is related to its concentration thus monitoring the drug concentration in biological fluids is very essential for therapeutic dose monitoring and pharmacokinetic studies. The method proposed in this work is intended for the determination of the investigated drugs in both saliva and plasma. Analysis of saliva has recently received considerable attention over the last decade for drug testing. It can provide a quick and non-invasive alternative to blood analysis and also can be an easier alternative to urine due to the lower complexity of its matrix.⁴ This facilitates sample preparation and permits the development of non-invasive simple and sensitive analytical methods for the accurate therapeutic drug monitoring of drugs that are well distributed in saliva.

It is well-reported that MX and MT are absorbed rapidly with high bioavailability without being affected by food or

Fig. 1 Chemical structure of (a) moxifloxacin hydrochloride and (b) metronidazole.

^aPharmaceutical Analytical Chemistry Department, Faculty of Pharmacy, Alexandria University, Egypt. E-mail: amira.elyazbi@alexu.edu.eg; Fax: +20 3 4873273; Tel: +20 3 4871317

^bDepartment of Pharmacology and Toxicology, Faculty of Pharmacy, Alexandria University, Egypt

[†] Electronic supplementary information (ESI) available. See https://doi.org/10.1039/d2ra01631a

infection.^{5,6} MX shows a good distribution between saliva and lung. Its maximum concentration ($C_{\rm max}$) in saliva is approximately double its $C_{\rm max}$ in plasma which imparts MX its potential activity against oropharynx pathogens. After single dose of MX, $C_{\rm max}$ was 3.2 $\mu {\rm g \ mL}^{-1}$.⁵ On the other hand, $C_{\rm max}$ for MX in human saliva is approximately double its $C_{\rm max}$ in plasma. While for MT, the $C_{\rm max}$ in human saliva is approximately equal to its $C_{\rm max}$ in plasma.⁷ After co-administration of single 500 mg and 250 mg MT doses, the $C_{\rm max}$ is 11.5 $\mu {\rm g \ mL}^{-1}$ and 6.2 $\mu {\rm g \ mL}^{-1}$, respectively.⁸ Moreover, MX is well absorbed in rat plasma reaching maximum concentration at 0.75 \pm 0.14 h with $C_{\rm max}$

value of 4.53 \pm 0.79 $\mu g~mL^{-1}$ after administration of 40 mg kg^{-1} of MX.9

Thorough literature review showed that this combination has been determined by very few analytical methods such as, spectrophotometry with chemometric analysis¹⁰ and simultaneous equations method.¹¹ However, to the best of our knowledge, there are still no reports on the pharmacokinetics parameters of the binary combination of MT and MX. For this reason, the purpose of this work is to develop, for the first time, a simple, fast and sensitive RP-HPLC method with dual diode array and fluorimetric detection for the simultaneous




Fig. 2 Typical HPLC chromatograms of a standard synthetic mixture containing 10 μg mL $^{-1}$ SAL, 20 μg mL $^{-1}$ MT and 20 μg mL $^{-1}$ MX (a) using DAD at 300 nm, (b) using DAD at 320 nm and (c) using FLD at $\lambda_{ex} = 300$ nm and $\lambda_{em} = 460$ nm.

determination of MX and MT in presence of salicylic acid (SAL) as internal standard (IS) in complex biological fluids that can be applied for the fast therapeutic drug monitoring of such combination.

Experimental 2.

2.1. Instrumentation

HPLC Agilent 1200 series formed of quaternary pump, autoinjector, vacuum degasser; Diode Array Detector G1315 C/D, multiple wavelength detector G1365 C/D and FLD detectors is used. The used software was agilent ChemStation (Agilent Technologies, Santa Clara, USA). The separation was performed using agilent Zorbax Eclipse plus C18 analytical column with the following dimensions 4.6 × 250 mm, 5 Micron. Zorbax Eclipse Plus C8 guard column was used to protect the analytical one (4.6 \times 12.5 mm, 5 Micron particle size). pH meter JENWAY (model 3505) was used for adjustment of pH of the solutions.

2.2. Materials and reagents

Reference moxifloxacin hydrochloride (99.58%), metronidazole (99.21%) and salicylic acid (99.08%) were purchased from Orchidia Pharmaceutical Industries, Al-Obour City, Egypt and Amriya Pharmaceuticals, Alexandria, Egypt, respectively. HPLC grade acetonitrile was obtained from Macron Fine Chemicals, USA while, sodium dihydrogen phosphate monohydrate, orthophosphoric acid 85% and sodium hydroxide were obtained from El-Nasr Chemical Industry Company, Egypt. Ethyl ether anhydrous was obtained from TEDIA Company.

2.3. Chromatographic conditions

The separation was achieved depending on isocratic system at flow rate 1 mL min⁻¹ using acetonitrile (ACN) and phosphate buffer pH 3 (sodium dihydrogen phosphate monohydrate adjusted with ortho-phosphoric acid) in ratio 30:70% v/v as mobile phase modified by 0.1% triethylamine. This mobile phase was filtered using 0.45 µm membrane filter. The analysis was performed with 10 µL injection volume at room temperature 25 °C. DAD detector is available to determine each drug at its maximum wavelength to ensure high sensitivity. Thus, determination of MX was performed using dual wavelength system where MX and MT were determined at 300 nm and at 320 nm, respectively. Also, DAD provided peak purity plots to ensure drug purity and absence of any interference. It was observed that MX and SAL (IS) produced response at FLD that was set at $\lambda_{ex}=300$ nm and $\lambda_{em}=460$ nm.

2.4. Preparation of solutions and construction of calibration graphs

2.4.1. Preparation of standard solutions. Stock solutions (1000 µg mL⁻¹) of MX, MT and SAL (IS) were prepared in HPLC grade ACN. Working solutions (500 µg mL⁻¹) were separately prepared using mobile phase as solvent. For preparation of calibrators and quality control samples, 20 µg mL⁻¹ working solution of MX, MT and SAL (IS) were prepared by dilution of the stock solution in ethyl ether.

2.4.2. Analysis of synthetic mixtures. Dilution of working solutions were performed in mobile phase at different ratios to obtain synthetic mixtures with different concentrations within the linearity ranges. These mixtures were valuable to assess reproducibility and repeatability of the proposed method.

2.4.3. Analysis of biological fluids

2.4.3.1. Analysis of saliva samples. In centrifuge tubes, samples were prepared by spiking 500 µL of human saliva with different concentrations of MX, MT and 10 μg mL⁻¹ SAL. Then 2 mL ethyl ether was added to saliva. All tubes were vortexed for 5 min then centrifuged 10 min at 2500 rpm. Using Pasteur pipettes, the upper layer of the organic solvent was transferred in small tubes that were subjected to evaporation by vacuum till dryness. Reconstitution of the formed residues was made with 500 µL of mobile phase directly before injection. Quality control samples and standard solutions were prepared and analyzed on each day. Peak area of investigated drugs to peak area of SAL (IS) was plotted against concentration of MX and MT to obtain calibration graph of each drug.

Low, medium and high-quality control samples (LQC, MQC and HQC) were prepared as 2, 10 and 20 µg mL⁻¹, respectively for both MX and MT by diode array detector and 2, 6 and 20 μg mL⁻¹, respectively, for MX by using FLD detector.

2.4.3.2. Analysis of plasma samples. Spiking of 200 µL human and rat plasma of each drug was performed in the presence of IS to prepare different concentrations, followed by addition of 2 mL ethyl ether to extract the investigated drugs. Then the procedure mentioned above under "Analysis of spiked saliva" was followed to construct the calibration graph.

LQC, MQC and HQC were prepared as 4, 10 and 20 $\mu \mathrm{g} \; \mathrm{mL}^{-1}$ respectively, for MX in rat plasma and 2, 10 and 20 $\mu g \text{ mL}^{-1}$ respectively, in human plasma. Also, the prepared concentrations of MT were 2, 10 and 20 $\mu g \text{ mL}^{-1}$ for LQC, MQC and HQC, respectively, in rat plasma and 1, 5 and 10 μg mL⁻¹ for LQC, MQC and HQC, respectively, in human plasma. For MX at FLD, the concentrations for LQC, MQC and HQC were 4, 6 and 10 μg mL⁻¹ and 2, 6 and 20 μg mL⁻¹ in rat and human plasma, respectively.

Table 1 Regression and statistical parameters for simultaneous determination of MX and MT in standard solutions using the proposed RP-HPLC method

Parameter	MX (DAD)	MX (FLD)	MT
Linearity range (μg mL ⁻¹)	0.2-100	0.1-100	0.2-80
$LOQ (\mu g mL^{-1})$	0.18	0.1	0.19
$LOD (\mu g mL^{-1})$	0.06	0.03	0.06
Intercept (a)	1.49×10^{-2}	1.13×10^{-2}	-3.41×10^{-2}
Slope (b)	1.60×10^{-1}	4.48×10^{-2}	7.44×10^{-1}
Correlation coefficient (r)	0.9997	0.9999	0.9999
$S_a^{\ a}$	3.96×10^{-2}	6.24×10^{-3}	1.06×10^{-1}
S_b^b S_b^2	8.83×10^{-4}	1.58×10^{-4}	3.05×10^{-3}
$S_{\rm b}^{\ 2}$	7.79×10^{-7}	2.48×10^{-8}	9.31×10^{-6}
$S_{y/x}^{c}$ F^d	1.00×10^{-1}	1.59×10^{-2}	2.62×10^{-1}
F^d	32 684.33	80 981.24	59 366.01
Significance F	2.47×10^{-17}	2.60×10^{-17}	9.01×10^{-17}

 $[^]a$ Standard deviation of the intercept. b Standard deviation of the slope. c Standard deviation of residuals. d Variance ratio.

Open Access Article. Published on 23 May 2022. Downloaded on 7/31/2025 9:17:35 AM.

This article is licensed under a Creative Commons Attribution-NonCommercial 3.0 Unported Licence.

Table 2 Intra-day and inter-day precision and accuracy for simultaneous determination of MX and MT in different biological fluids using the proposed RP-HPLC method

nn human saliva MX (FiD) MT MT MX (FiD) MT MX (FiD) MT MX (FiD) MY MX (FiD) MY MX (FiD) MX		Average% recovery $\pm\mathrm{SD}$	± SD		%RSD			$\%E_{ m r}$		
nun saliva 90.53 ± 13.01 90.62 ± 18.06 97.29 ± 11.09 14.31 19.92 11.40 9p precision 96.11 ± 4.58 89.44 ± 6.93 97.72 ± 3.67 4.77 7.74 3.76 9p precision 96.11 ± 4.58 89.44 ± 6.93 97.73 ± 3.67 4.77 7.74 3.76 103.60 ± 4.89 99.53 ± 2.73 100.57 ± 0.77 6.26 5.79 0.77 102.60 ± 1.99 101.20 ± 1.30 98.53 ± 4.16 0.44 1.29 0.77 101.49 ± 0.45 101.02 ± 1.30 98.73 ± 4.16 0.44 1.29 4.22 nan plasma 101.49 ± 0.45 101.02 ± 1.30 90.51 ± 5.52 1.99 1.35 0.66 pyrecision 95.60 ± 1.89 108.55 ± 15.19 90.51 ± 5.52 1.99 1.35 4.41 top precision 97.74 ± 1.82 10.58 ± 3.29 10.58 ± 4.42 1.23 4.41 ap precision 101.04 ± 1.97 100.44 ± 1.37 100.40 ± 1.37 1.05 3.37 4.51 ap precision 101.04 ± 1.97 100.45 ± 0	Conc. $(\mu \mathrm{g \ mL}^{-1})$	MX (DAD)	MX (FLD)	MT	MX (DAD)	MX (FLD)	MT	MX (DAD)	MX (FLD)	MT
90.93 ± 13.01 90.62 ± 18.06 97.29 ± 114.0 14.31 19.92 114.0 114.0 90.31 ± 13.01 90.52 ± 18.06 97.72 ± 114.0 14.31 19.92 17.74 ± 3.76 99.73 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 99.53 ± 5.46 98.20 ± 1.47 15.80 0.66 101.49 ± 0.45 101.02 ± 1.32 103.30 ± 6.98 11.47 15.90 17.49 1.20 0.66 11.02 ± 1.30 104.72 ± 1.32 103.30 ± 0.49 11.87 11.87 11.80 0.66 11.85 10.02 ± 1.32 10.02 ± 1.32 10.03 ± 0.42 11.87 10.04 ± 1.49 1	In human saliva Intra-day precision									
$\begin{array}{cccccccccccccccccccccccccccccccccccc$	LLQC	90.93 ± 13.01	90.62 ± 18.06	97.29 ± 11.09	14.31	19.92	11.40	-9.07	-9.38	-2.71
ap precision 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.10 97.52 ± 6.70 97.70 </td <td>ĽĄĆ</td> <td>96.11 ± 4.58</td> <td>89.44 ± 6.93</td> <td>97.73 ± 3.67</td> <td>4.77</td> <td>7.74</td> <td>3.76</td> <td>-3.90</td> <td>-10.56</td> <td>-2.27</td>	ĽĄĆ	96.11 ± 4.58	89.44 ± 6.93	97.73 ± 3.67	4.77	7.74	3.76	-3.90	-10.56	-2.27
op precision 103.60 ± 4.89 99.53 ± 5.46 98.20 ± 1.43 4.72 5.49 1.46 op precision 93.10 ± 10.77 114.26 ± 18.06 105.37 ± 0.69 11.47 15.80 0.66 nan plasma 102.69 ± 1.79 104.73 ± 13.72 103.04 ± 6.98 1.74 13.10 6.76 p precision 95.26 ± 1.89 104.74 ± 1.82 11.87 4.69 1.85 sp precision 95.26 ± 1.89 108.55 ± 15.19 90.51 ± 5.52 1.99 13.39 6.10 sp precision 97.82 ± 2.84 97.64 ± 1.33 108.97 ± 1.79 2.90 3.37 4.14 sp precision 101.04 ± 1.97 100.49 ± 1.33 108.97 ± 1.29 1.95 0.44 4.14 sp precision 104.73 ± 8.26 113.04 ± 13.04 91.86 ± 4.15 7.88 11.54 4.51 sp precision 100.66 ± 3.28 97.68 ± 3.29 106.48 ± 3.49 6.41 5.38 3.28 sp precision 102.36 ± 4.33 100.84 ± 3.39 106.48 ± 3.39 10.56 8.45 3.78 </td <td>MQC</td> <td>97.52 ± 6.10</td> <td>97.95 ± 2.73</td> <td>100.57 ± 0.77</td> <td>6.26</td> <td>2.79</td> <td>0.77</td> <td>-2.48</td> <td>-2.05</td> <td>0.57</td>	MQC	97.52 ± 6.10	97.95 ± 2.73	100.57 ± 0.77	6.26	2.79	0.77	-2.48	-2.05	0.57
apprecision 93.10 ± 10.77 114.26 ± 18.06 105.37 ± 0.69 11.47 15.80 0.66 102.69 ± 1.79 104.73 ± 13.72 103.30 ± 6.98 1.74 13.10 6.76 101.49 ± 0.45 101.02 ± 1.30 98.73 ± 4.16 0.44 1.29 4.22 101.49 ± 0.45 101.02 ± 1.30 98.73 ± 4.16 0.44 1.29 4.22 101.49 ± 0.45 101.02 ± 1.30 90.51 ± 5.52 1.99 13.99 6.10 108.31 ± 12.86 97.82 ± 2.84 97.68 ± 3.29 108.97 ± 1.79 2.90 3.37 1.64 97.72 ± 2.84 97.68 ± 3.29 106.83 ± 4.42 1.23 1.32 4.14 97.72 ± 2.84 97.68 ± 3.29 106.83 ± 4.42 1.23 1.34 4.14 ap precision 101.04 ± 1.97 100.45 ± 0.45 101.90 ± 4.22 1.95 6.41 4.14 ap precision 104.73 ± 8.26 103.97 ± 5.95 107.50 ± 10.08 3.26 3.37 9.38 102.35 ± 4.05 102.86 ± 3.95 107.50 ± 10.28 3.26 3.37 <td>НОС</td> <td>103.60 ± 4.89</td> <td>99.53 ± 5.46</td> <td>98.20 ± 1.43</td> <td>4.72</td> <td>5.49</td> <td>1.46</td> <td>3.60</td> <td>-0.47</td> <td>-1.80</td>	НОС	103.60 ± 4.89	99.53 ± 5.46	98.20 ± 1.43	4.72	5.49	1.46	3.60	-0.47	-1.80
93.10 ± 10.77 114.20 ± 18.06 105.37 ± 0.69 11.47 15.80 0.66 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.79 10.2.69 ± 1.20 10.2.69 ± 1.30 0.44 10.2.90 11.87 4.69 1.85 ± 1.80 0.54 ± 1.20 10.2.54 ± 1.20 10.2.54 ± 1.20 10.2.54 ± 1.20 10.2.44 2 1.2.3 1.32 1.32 1.34 1.41	Inter-day precision									
nan plasma 102.69 ± 1.79 10.43 ± 13.72 103.30 ± 6.98 1.74 13.10 6.76 nan plasma 101.49 ± 0.45 104.24 ± 1.82 103.30 ± 6.98 1.74 13.10 6.76 ann plasma 95.26 ± 1.89 101.02 ± 1.30 90.51 ± 5.52 1.99 13.39 6.10 ap precision 95.26 ± 1.89 108.55 ± 15.19 90.51 ± 5.52 1.99 13.39 6.10 ap precision 97.82 ± 2.84 97.68 ± 3.29 106.83 ± 4.42 1.23 1.64 4.14 ap precision 104.73 ± 8.26 113.04 ± 1.37 10.68 ± 4.42 1.23 1.64 4.14 ap precision 104.73 ± 8.26 113.04 ± 13.04 91.86 ± 4.15 7.88 11.54 4.51 plasma 100.66 ± 3.28 97.86 ± 3.29 106.48 ± 3.49 6.41 4.14 ap precision 94.12 ± 5.69 105.88 ± 8.95 106.48 ± 3.97 95.35 3.84 2.99 plasma 102.36 ± 4.43 95.10 ± 6.67 96.32 ± 2.35 4.33 3.69 1.13	LLQC	93.10 ± 10.77	114.26 ± 18.06	105.37 ± 0.69	11.47	15.80	99.0	06.9	14.26	5.37
nan plasma 101.49 ± 0.45 101.02 ± 1.30 98.73 ± 4.16 0.44 1.29 4.22 nan plasma 108.31 ± 12.86 95.47 ± 4.48 97.47 ± 1.82 11.87 4.69 1.85 ap precision 95.26 ± 1.89 108.53 ± 15.19 90.51 ± 5.52 1.99 13.99 6.10 97.82 ± 2.84 97.68 ± 3.29 108.97 ± 1.79 2.90 3.37 1.64 97.47 ± 1.20 100.49 ± 1.33 106.83 ± 4.42 1.23 1.32 4.14 ap precision 104.73 ± 8.26 113.04 ± 1.30 10.90 ± 4.22 1.95 0.44 4.14 ap precision 106.65 ± 3.28 97.68 ± 3.29 107.50 ± 10.08 3.26 3.37 9.38 plasma 100.66 ± 3.28 97.68 ± 3.95 106.48 ± 3.49 6.41 5.38 3.28 plasma Apprecision 94.12 ± 5.69 105.88 ± 8.95 106.35 ± 2.37 6.43 5.38 3.28 plasma Apprecision 94.12 ± 5.69 106.44 ± 3.95 106.35 ± 2.35 4.33 7.02 2.99 </td <td>гóс</td> <td>102.69 ± 1.79</td> <td>104.73 ± 13.72</td> <td>103.30 ± 6.98</td> <td>1.74</td> <td>13.10</td> <td>92.9</td> <td>2.69</td> <td>4.73</td> <td>3.30</td>	гóс	102.69 ± 1.79	104.73 ± 13.72	103.30 ± 6.98	1.74	13.10	92.9	2.69	4.73	3.30
nan plasma 95.47 ± 4.48 97.47 ± 1.82 11.87 4.69 1.85 ap precision 95.26 ± 1.89 108.55 ± 15.19 90.51 ± 5.52 1.99 13.99 6.10 ap precision 97.82 ± 2.84 97.68 ± 3.29 108.97 ± 1.79 2.90 3.37 1.64 97.82 ± 2.84 97.68 ± 3.29 106.83 ± 4.42 1.23 1.32 4.14 97.82 ± 2.84 97.68 ± 3.29 106.44 1.23 1.32 4.14 97.82 ± 2.84 97.68 ± 3.29 101.90 ± 4.22 1.95 0.44 4.14 40 precision 104.73 ± 8.26 113.04 ± 1.30 91.86 ± 4.15 7.88 11.54 4.51 plasma 102.38 ± 6.56 103.97 ± 5.59 106.88 ± 3.49 6.41 5.38 3.28 99.78 ± 0.55 102.86 ± 3.95 99.33 ± 2.97 0.55 3.84 2.99 plasma 4yp precision 94.12 ± 5.69 105.88 ± 8.95 103.35 ± 3.91 6.05 8.45 3.78 plasma 99.88 ± 1.35 106.48 ± 2.35 99.33 ± 2.9	MQC	101.49 ± 0.45	101.02 ± 1.30	98.73 ± 4.16	0.44	1.29	4.22	1.49	1.02	-1.27
$\begin{array}{cccccccccccccccccccccccccccccccccccc$	НОС	108.31 ± 12.86	95.47 ± 4.48	97.47 ± 1.82	11.87	4.69	1.85	8.31	-4.53	-2.53
$\begin{array}{cccccccccccccccccccccccccccccccccccc$	In human plasma									
$\begin{array}{cccccccccccccccccccccccccccccccccccc$	Intra-day precision									
$\begin{array}{cccccccccccccccccccccccccccccccccccc$	LLQC	95.26 ± 1.89	108.55 ± 15.19	90.51 ± 5.52	1.99	13.99	6.10	-4.74	8.55	-9.49
ay precision 97.47 ± 1.20 100.49 ± 1.33 106.83 ± 4.42 1.23 1.32 4.14 ay precision 101.04 ± 1.97 100.45 ± 0.45 101.90 ± 4.22 1.95 0.44 4.14 ay precision 104.73 ± 8.26 113.04 ± 13.04 91.86 ± 4.15 7.88 11.54 4.51 plasma 100.66 ± 3.28 107.50 ± 10.08 90.33 ± 2.97 0.55 3.84 2.99 plasma 94.12 ± 5.69 102.86 ± 3.95 99.33 ± 2.97 0.55 3.84 2.99 plasma 94.12 ± 5.69 105.88 ± 8.95 103.35 ± 3.91 6.05 8.45 3.78 99.78 ± 4.43 95.10 ± 6.67 90.52 ± 2.35 4.33 7.02 2.59 99.88 ± 1.35 100.03 ± 2.93 98.01 ± 1.27 5.01 2.92 1.29 ay precision 91.13 ± 3.36 101.47 ± 3.09 100.35 ± 1.11 1.35 3.05 1.11 99.88 ± 1.35 103.92 ± 4.90 97.31 ± 5.01 5.58 4.71 <td>róc</td> <td>97.82 ± 2.84</td> <td>97.68 ± 3.29</td> <td>108.97 ± 1.79</td> <td>2.90</td> <td>3.37</td> <td>1.64</td> <td>-2.18</td> <td>-2.32</td> <td>8.97</td>	róc	97.82 ± 2.84	97.68 ± 3.29	108.97 ± 1.79	2.90	3.37	1.64	-2.18	-2.32	8.97
ay precision 101.04 ± 1.97 100.45 ± 0.45 101.90 ± 4.22 1.95 0.44 4.14 ay precision 104.73 ± 8.26 113.04 ± 13.04 91.86 ± 4.15 7.88 11.54 4.51 plasma 100.66 ± 3.28 90.78 ± 0.55 102.86 ± 3.95 90.33 ± 2.97 0.55 3.37 9.38 plasma 99.78 ± 0.55 102.86 ± 3.95 90.33 ± 2.97 0.55 3.84 2.99 qy precision 94.12 ± 5.69 105.88 ± 8.95 103.35 ± 3.91 6.05 8.45 3.78 qy precision 94.12 ± 5.69 100.03 ± 2.93 98.01 ± 1.27 5.01 2.92 1.29 qy precision 94.12 ± 5.69 100.03 ± 2.93 98.01 ± 1.27 5.01 2.92 1.11 qy precision 91.13 ± 3.36 101.47 ± 3.09 100.35 ± 1.11 1.35 3.05 1.11 qy precision 91.13 ± 3.36 103.42 ± 4.90 95.97 ± 2.69 3.69 11.49 2.80 qy precision 91.13 ± 3.36 92.24 ± 9.02 92.34 ± 5.01 92.94 92.94 <td>MQC</td> <td>97.47 ± 1.20</td> <td>100.49 ± 1.33</td> <td>106.83 ± 4.42</td> <td>1.23</td> <td>1.32</td> <td>4.14</td> <td>-2.53</td> <td>0.49</td> <td>6.83</td>	MQC	97.47 ± 1.20	100.49 ± 1.33	106.83 ± 4.42	1.23	1.32	4.14	-2.53	0.49	6.83
ay precision 104.73 ± 8.26 113.04 ± 13.04 91.86 ± 4.15 7.88 11.54 4.51 100.66 ± 3.28 97.68 ± 3.29 107.50 ± 10.08 3.26 3.37 9.38 100.63 ± 6.56 103.97 ± 5.59 106.48 ± 3.49 6.41 5.38 3.28 99.78 ± 0.55 102.86 ± 3.95 99.33 ± 2.97 0.55 3.84 2.99 99.78 ± 0.55 102.86 ± 8.95 103.35 ± 3.91 6.05 8.45 3.78 qy precision 94.12 ± 5.69 105.88 ± 8.95 103.35 ± 3.91 6.05 8.45 3.78 qy precision 94.12 ± 5.69 105.04 ± 6.67 90.52 ± 2.35 4.33 7.02 2.59 99.88 ± 1.35 100.47 ± 3.09 100.35 ± 1.11 1.35 3.05 1.11 qy precision 91.13 ± 3.36 103.43 ± 11.89 95.97 ± 2.69 3.69 11.49 2.80 96.81 ± 5.40 101.29 ± 0.82 101.29 ± 1.17 1.62 8.82 1.15 101.29 ± 1.62 97.75 ± 8.62 101.29 ± 1.17 1.62 8.82 1.15	НОС	101.04 ± 1.97	100.45 ± 0.45	101.90 ± 4.22	1.95	0.44	4.14	1.04	0.45	1.90
plasma plasma $10.34 \pm 1.3.04$ 91.86 ± 4.15 7.88 11.54 4.51 plasma plasma 10.66 ± 3.28 10.50 ± 10.08 3.26 3.37 9.38 splasma 102.33 ± 6.56 103.97 ± 5.59 106.48 ± 3.49 6.41 5.38 3.28 gy.78 ± 0.55 102.86 ± 3.95 106.48 ± 3.97 0.55 3.84 2.99 gy.78 ± 0.55 102.86 ± 3.95 106.33 ± 2.97 0.55 3.84 2.99 sqy precision 94.12 ± 5.69 105.88 ± 8.95 103.35 ± 3.91 6.05 8.45 3.78 qy precision 99.88 ± 1.35 101.47 ± 3.09 100.35 ± 1.11 1.35 3.05 1.11 qy precision 91.13 ± 3.36 103.43 ± 11.89 95.97 ± 2.69 3.69 11.49 2.80 96.81 ± 5.40 103.92 ± 4.90 97.31 ± 5.01 5.58 4.71 5.15 101.29 ± 0.82 10.572 ± 10.20 98.21 ± 1.62 0.81 9.65 1.16 100.07 ± 1.62 97.75 ± 8.62 101.29 ± 1.17 1.62	Inter-day precision									
plasma 100.66 ± 3.28 97.68 ± 3.29 107.50 ± 10.08 3.26 3.37 9.38 plasma 102.33 ± 6.56 103.97 ± 5.59 106.48 ± 3.49 6.41 5.38 3.28 plasma 99.78 ± 0.55 102.86 ± 3.95 99.33 ± 2.97 0.55 3.84 2.99 qy precision 94.12 ± 5.69 105.88 ± 8.95 103.35 ± 3.91 6.05 8.45 3.78 qy precision 94.12 ± 5.69 105.88 ± 8.95 103.35 ± 2.35 4.33 7.02 2.59 gen th ± 5.01 100.03 ± 2.93 98.01 ± 1.27 5.01 2.92 1.29 qy precision 91.13 ± 3.36 101.47 ± 3.09 100.35 ± 1.11 1.35 3.05 1.11 qy precision 91.13 ± 3.36 103.43 ± 11.89 95.97 ± 2.69 3.69 11.49 2.80 qy precision 91.13 ± 3.36 105.72 ± 10.20 98.21 ± 1.62 9.65 4.71 5.15 qy precision 91.02.94 92.97 ± 2.69 3.69 11.49 2.65 1.65 qy precisio	LLQC	104.73 ± 8.26	113.04 ± 13.04	91.86 ± 4.15	7.88	11.54	4.51	4.73	13.04	-8.14
plasma 102.33 ± 6.56 103.97 ± 5.59 106.48 ± 3.49 6.41 5.38 3.28 plasma plasma (a) $precision$ 94.12 ± 5.69 102.88 ± 8.95 103.35 ± 3.91 6.05 8.45 3.78 102.36 ± 4.43 95.10 ± 6.67 90.52 ± 2.35 4.33 7.02 2.59 98.16 ± 5.01 100.03 ± 2.93 98.01 ± 1.27 5.01 2.92 1.29 qy precision 99.88 ± 1.35 101.47 ± 3.09 100.35 ± 1.11 1.35 3.05 1.11 qy precision 91.13 ± 3.36 103.43 ± 11.89 95.97 ± 2.69 3.69 11.49 2.80 96.81 ± 5.40 103.92 ± 4.90 97.31 ± 5.01 5.58 4.71 5.15 101.29 ± 0.82 105.72 ± 10.20 98.21 ± 1.62 0.81 9.65 1.65 100.07 ± 1.62 97.75 ± 8.62 101.29 ± 1.17 1.62 8.82 1.15	гóс	100.66 ± 3.28	97.68 ± 3.29	107.50 ± 10.08	3.26	3.37	9.38	99.0	-2.32	7.50
plasma 99.78 ± 0.55 102.86 ± 3.95 99.33 ± 2.97 0.55 3.84 2.99 plasma (ay precision) 94.12 ± 5.69 105.88 ± 8.95 103.35 ± 3.91 6.05 8.45 3.78 102.36 ± 4.43 95.10 ± 6.67 90.52 ± 2.35 4.33 7.02 2.59 98.16 ± 5.01 100.03 ± 2.93 98.01 ± 1.27 5.01 2.92 1.29 99.88 ± 1.35 101.47 ± 3.09 100.35 ± 1.11 1.35 3.05 1.11 ay precision 91.13 ± 3.36 103.43 ± 11.89 95.97 ± 2.69 3.69 11.49 2.80 96.81 ± 5.40 103.92 ± 4.90 97.31 ± 5.01 5.58 4.71 5.15 101.29 ± 0.82 105.72 ± 10.20 98.21 ± 1.62 0.81 9.65 1.65 100.07 ± 1.62 97.75 ± 8.62 101.29 ± 1.17 1.62 8.82 1.15	MQC	102.33 ± 6.56	103.97 ± 5.59	106.48 ± 3.49	6.41	5.38	3.28	2.33	3.97	6.48
$ay \ precision \\ 94.12 \pm 5.69 \\ 102.36 \pm 4.43 \\ 99.88 \pm 1.35 \\ ay \ precision \\ ay \ precision \\ ay \ precision \\ 101.29 \pm 0.82 \\ above boldship \\ beta boldship \\ beta boldship \\ column \\ co$	НОС	99.78 ± 0.55	102.86 ± 3.95	99.33 ± 2.97	0.55	3.84	2.99	-0.22	2.86	-0.67
tay precision 94.12 ± 5.69 105.88 ± 8.95 103.35 ± 3.91 6.05 8.45 3.78 102.36 ± 4.43 95.10 ± 6.67 90.52 ± 2.35 4.33 7.02 2.59 98.16 ± 5.01 100.03 ± 2.93 98.01 ± 1.27 5.01 2.92 1.29 99.88 ± 1.35 101.47 ± 3.09 100.35 ± 1.11 1.35 3.05 1.11 ay precision 91.13 ± 3.36 103.43 ± 11.89 95.97 ± 2.69 3.69 11.49 2.80 96.81 ± 5.40 103.92 ± 4.90 97.31 ± 5.01 5.58 4.71 5.15 101.29 ± 0.82 105.72 ± 10.20 98.21 ± 1.62 0.81 9.65 1.65 100.07 ± 1.62 97.75 ± 8.62 101.29 ± 1.17 1.62 8.82 1.15	In rat plasma									
4.12 ± 5.69 $105.88 ± 8.95$ $103.35 ± 3.91$ 6.05 8.45 3.78 $102.36 ± 4.43$ $95.10 ± 6.67$ $90.52 ± 2.35$ 4.33 7.02 2.59 $98.16 ± 5.01$ $100.03 ± 2.93$ $98.01 ± 1.27$ 5.01 2.92 1.29 $99.88 ± 1.35$ $101.47 ± 3.09$ $100.35 ± 1.11$ 1.35 3.05 1.11 ay precision $91.13 ± 3.36$ $103.43 ± 11.89$ $95.97 ± 2.69$ 3.69 11.49 2.80 $96.81 ± 5.40$ $103.92 ± 4.90$ $97.31 ± 5.01$ 5.58 4.71 5.15 $101.29 ± 0.82$ $105.72 ± 10.20$ $98.21 ± 1.62$ 0.81 9.65 1.65 $100.07 ± 1.62$ $97.75 ± 8.62$ $101.29 ± 1.17$ 1.62 8.82 1.15	Intra-day precision									
(a) precision 101.36 ± 4.43 95.10 ± 6.67 90.52 ± 2.35 4.33 7.02 2.59 98.16 ± 5.01 100.03 ± 2.93 98.01 ± 1.27 5.01 2.92 1.29 99.88 ± 1.35 101.47 ± 3.09 100.35 ± 1.11 1.35 3.05 1.11 99.88 ± 1.35 103.43 ± 11.89 95.97 ± 2.69 3.69 11.49 2.80 96.81 ± 5.40 103.92 ± 4.90 97.31 ± 5.01 5.58 4.71 5.15 101.29 ± 0.82 105.72 ± 10.20 98.21 ± 1.62 0.81 9.65 1.65 100.07 ± 1.62 97.75 ± 8.62 101.29 ± 1.17 1.62 8.82 1.15	LLQC	94.12 ± 5.69	105.88 ± 8.95	103.35 ± 3.91	6.05	8.45	3.78	-5.88	5.88	3.35
98.16 \pm 5.01 100.03 \pm 2.93 98.01 \pm 1.27 5.01 2.92 1.29 99.88 \pm 1.35 101.47 \pm 3.09 100.35 \pm 1.11 1.35 3.05 1.11 99.88 \pm 1.35 101.47 \pm 3.09 100.35 \pm 1.11 1.35 3.05 1.11 101.29 \pm 0.80 103.43 \pm 11.89 95.97 \pm 2.69 3.69 11.49 2.80 96.81 \pm 5.40 103.92 \pm 4.90 97.31 \pm 5.01 5.58 4.71 5.15 101.29 \pm 0.82 105.72 \pm 10.20 98.21 \pm 1.62 0.81 9.65 1.65 100.07 \pm 1.62 97.75 \pm 8.62 101.29 \pm 1.17 1.62 8.82 1.15	CÓC	102.36 ± 4.43	95.10 ± 6.67	90.52 ± 2.35	4.33	7.02	2.59	2.36	-4.90	-9.48
(a) precision 99.88 \pm 1.35 101.47 \pm 3.09 100.35 \pm 1.11 1.35 3.05 1.11 99.88 \pm 1.35 101.3 \pm 3.09 103.43 \pm 11.89 95.97 \pm 2.69 3.69 11.49 2.80 96.81 \pm 5.40 103.92 \pm 4.90 97.31 \pm 5.01 5.58 4.71 5.15 101.29 \pm 0.82 105.72 \pm 10.20 98.21 \pm 1.62 0.81 9.65 1.65 100.07 \pm 1.62 97.75 \pm 8.62 101.29 \pm 1.17 1.62 8.82 1.15	MQC	98.16 ± 5.01	100.03 ± 2.93	98.01 ± 1.27	5.01	2.92	1.29	-1.84	0.03	-1.99
tay precision91.13 \pm 3.36103.43 \pm 11.8995.97 \pm 2.693.6911.492.8096.81 \pm 5.40103.92 \pm 4.9097.31 \pm 5.015.584.715.15101.29 \pm 0.82105.72 \pm 10.2098.21 \pm 1.620.819.651.65100.07 \pm 1.6297.75 \pm 8.62101.29 \pm 1.171.628.821.15	НОС	99.88 ± 1.35	101.47 ± 3.09	100.35 ± 1.11	1.35	3.05	1.11	-0.12	1.47	0.35
$\begin{array}{cccccccccccccccccccccccccccccccccccc$	Inter-day precision									
$\begin{array}{cccccccccccccccccccccccccccccccccccc$	LLQC	91.13 ± 3.36	103.43 ± 11.89	95.97 ± 2.69	3.69	11.49	2.80	-8.87	3.43	-4.03
$ \begin{array}{cccccccccccccccccccccccccccccccccccc$	róc	96.81 ± 5.40	103.92 ± 4.90	97.31 ± 5.01	5.58	4.71	5.15	-3.19	3.92	-2.69
100.07 ± 1.62 97.75 ± 8.62 101.29 ± 1.17 1.62 8.82 1.15	MQC	101.29 ± 0.82	105.72 ± 10.20	98.21 ± 1.62	0.81	9.65	1.65	1.29	5.72	-1.79
	НОС	100.07 ± 1.62	97.75 ± 8.62	101.29 ± 1.17	1.62	8.82	1.15	0.07	-2.25	1.29

All animal experiments were conducted according to a protocol approved by Institutional Animal Care and Use Committee (IACUC), Alexandria University and in compliance with the Guide for Care and Use of Laboratory Animals of the Institute for Laboratory Animal Research of the National Academy of Sciences, USA. It was performed on twelve male Wistar rats weighing 290–340 g. The rats were kept at ambient temperature and humidity-controlled room, in a 12 hour light/dark cycle. The rats received oral dosage combining both MT (7.5 mg kg⁻¹) and MX (9.2 mg kg⁻¹) prepared in saline. Then they were anesthetized and blood samples were collected *via* retro-orbital bleeding in polypropylene tubes with anti-

coagulant EDTA solution at specified time (1 h). Preparation of plasma was performed after centrifugation for 10 min at 2500 rpm then it was kept at $-20\,^{\circ}\text{C}$ to be thawed directly before preparation. SAL (10 $\mu\text{g mL}^{-1}$) was added to 200 μL plasma. The samples were prepared and analyzed as the same procedure mentioned above.

2.5. Stability studies

Freeze and thaw stability for MX and MT was performed by applying three cycles of freeze and thaw. LQC and HQC samples were subjected to freezing $(-20 \,^{\circ}\text{C})$ for 1 h then thawed at room temperature (n=3). This cycle was repeated three times. The




Fig. 3 Typical HPLC chromatograms of (a) blank saliva using DAD at 300 nm, (b) 1.5 μ g mL⁻¹ MX and 1.5 μ g mL⁻¹ MT in saliva in the presence of 10 μ g mL⁻¹ SAL (IS) using DAD at 300 nm, (c) 1.5 μ g mL⁻¹ MX and 1.5 μ g mL⁻¹ MT in saliva in the presence of 10 μ g mL⁻¹ SAL (IS) using DAD at 320 nm and (d) 1.5 μ g mL⁻¹ MX and 1.5 μ g mL⁻¹ MT in saliva in the presence of 10 μ g mL⁻¹ SAL (IS) using FLD at $\lambda_{ex} = 300$ nm and $\lambda_{em} = 460$ nm.

This article is licensed under a Creative Commons Attribution-NonCommercial 3.0 Unported Licence.

Open Access Article. Published on 23 May 2022. Downloaded on 7/31/2025 9:17:35 AM

Paper

samples were measured after the third one. All obtained results were compared with the results of freshly prepared samples. While in short-term stability, LQC and HQC samples were left for 8 h at ambient temperature (25 °C). For post preparative stability, the prepared sample of LQC and HQC were stored for one day at 4 °C before analysis. The samples equivalent to LQC and HQC were prepared and stored for ten days at -20 °C for study of longterm stability. Finally, the stock solutions stability was determined at 25 °C at different time intervals. Moreover, the stocks were measured after freezing within the whole study period.

3. Results and discussion

3.1. Method development and optimization

In order to obtain the optimum chromatographic conditions for the sensitive and accurate analysis of the studied combination, we studied the factors that affect the separation of the investigated drugs with significant effect on system suitability parameters, such as, packing material, pH of phosphate buffer, organic modifier ratio and detection wavelengths.

- 3.1.1. Stationary phase. Different columns with different packing materials and dimensions were tested. For example; Zorbax Eclipse plus C18 analytical column (4.6 \times 250 mm), Zorbax SB C8 analytical column (4.6 \times 250 mm) and Zorbax Eclipse plus C18 analytical column (4.6 imes 150 mm). Zorbax Eclipse plus C18 (4.6 \times 250 mm, 5 micron) produced the best separation with short analysis time. It provided the most symmetric peaks in addition to good resolution between peaks. These columns were tested using mobile phase of phosphate buffer and acetonitrile in variable ratios and pH values from 3 to 6.
- **3.1.2. Modifier ratio.** Different percentages of acetonitrile (ACN) from 10% to 30% were examined. Increasing ACN% led to significant changes in retention time. It was obvious that ACN of 30% obtained the best peak shape with good resolution between the eluted peaks (Fig. S1a, ESI†). It produced accepted separation at short retention time.
- 3.1.3. Buffer pH. Different pH values (pH 3-6) of phosphate buffer were tested. It was obvious that high pH values produced distortion in peak shapes with significant changes in retention time accompanied with minor changes in the pH values higher than pH 4 with longer analysis time as shown in Fig. S1b.† While, pH 3 resulted in well resolved peaks with good shapes within a short analysis time. Therefore, pH 3 was selected for analysis.
- **3.1.4.** Detection wavelength. DAD provided an advantage of analysis of each drug at its maximum wavelength which ensured high sensitivity and selectivity of assay. So, application of HPLC-Dual wavelength system was performed to measure MX at 300 nm and MT at 320 nm. Moreover, it provided peak purity plots to ensure purity of investigated drugs. Also, using fluorescence detector within complex biological matrixes ensures high sensitivity and selectivity of the method for analysis of MX at $\lambda_{ex} = 300$ nm and $\lambda_{em} = 460$ nm.

3.2. System suitability parameters

The system suitability parameters are essential part for accurate chromatographic methods. They were valuable for assessment

of efficiency of separation, peak symmetry, and resolution of peaks. As shown in Table S1 (ESI†), the peaks of investigated drugs were well separated with high efficiency of separation with short analysis time, where, MT and MX were eluted at retention time of 3.22 and 3.99 min, respectively (Fig. 2).

3.3. Validation of the proposed method

Validation of the proposed method was performed depending on Bio-analytical Method Validation Guidance for Industry to determine MX and MT in biological fluids.12

- 3.3.1. Linearity and concentration ranges. The calibration graphs were constructed by plotting peak area ratio of the investigated drugs to IS against their corresponding concentrations. The drugs showed linear response with correlation coefficient >0.999, small significance F values and large F values. The linearity ranges of each drug were demonstrated in Table 1 for standard solutions and Table S2 (ESI†) for biological fluids to prove the sensitivity of the proposed methods.
- 3.3.2. Lower limit of quantitation (LLOQ). LLOQ were determined by analysis of five replicates of small concentrations of each analyte against blank of biological fluids. The lowest concentration of each analyte with high precision and accuracy were determined as LLOQ. Low LLOQ values were obtained by the proposed methods as demonstrated in Table S2† indicating the high sensitivity of the proposed methods.
- 3.3.3. Accuracy, precision and recoveries. The accuracy and precision were performed for the proposed method on the same day and on three different days to assess accuracy, intra-day precision and inter-day precision. Set of calibrators and quality control sample involving LLOQ, LQC, MQC and HQC have been analyzed with five replicates per each quality control level. The values obtained of %recovery, percentage error (% E_r) and percentage relative standard deviation (%RSD) were lower than 15% for concentrations higher that LLOQ and 20% for LLOQ. These values showed high accuracy and precision of the proposed method for synthetic mixtures (Table S3, ESI†) and for quality control samples of different biological fluids (Table 2).

Table 3 Results for the analysis of the synthetically prepared mixtures containing MX and MT in biological fluids using the proposed RP-HPLC method

Average % recovery \pm SD ^a				% RSD ^b			
MX			MX	MX			
DAD	FLD	MT	DAD	FLD	МТ		
Saliva 96.01 ± 4.97	106.06 ± 9.42	105.17 ± 10.67	5.17	8.89	10.14		
Human plasma 107.41 ± 3.20		105.72 ± 5.88	2.98	3.90	5.56		
Rat plasma 103.37 ± 7.81	103.34 ± 8.37	102.84 ± 9.22	7.56	8.10	8.96		

 $[^]a$ Average percentage recovery \pm standard deviation for five determinations. b Percentage relative standard deviation.

3.3.4. Robustness. Robustness is a measure of the capacity of the developed method to remain unaffected by small but deliberate variations in used conditions. It provides an indication of the reliability of the developed method during normal usage. Applying slight variations in some chromatographic conditions provided no significant effect on resolution, peak shape or responses of peak area, which ensured by accepted values of %recovery and small values of %RSD (Table S4, ESI†).

The involved parameters were buffer pH, buffer concentration and detection wavelength.

3.3.5. Stability. The stability studies were carried out on the stock solutions for each drug where the drugs showed high stability. Also, the studies were performed on QC samples with low and high concentration including long and short terms stability, freeze and thaw and post preparative stability. The obtained results were compared with freshly prepared solutions




Fig. 4 Typical HPLC chromatogram of (a) blank rat plasma using DAD at 300 nm, (b) 4 μ g mL $^{-1}$ MX and 4 μ g mL $^{-1}$ MT spiked in rat plasma in presence of 10 μ g mL $^{-1}$ SAL (IS) using DAD at 300 nm, (c) 4 μ g mL $^{-1}$ MX and 4 μ g mL $^{-1}$ MT spiked in rat plasma in presence of 10 μ g mL $^{-1}$ MX and 4 μ g mL $^{-1}$ MT spiked in rat plasma in presence of 10 μ g mL $^{-1}$ SAL (IS) using FLD at $\lambda_{ex}=300$ nm and $\lambda_{em}=460$ nm.

responses. As shown in Table S5 (ESI†) for spiked saliva and Table S6 (ESI†) for spiked plasma, $\%E_{\rm r}$ and %RSD values were within the accepted limits which proved the stability of drugs by application of different conditions.

3.3.6. Selectivity and matrix effect. During method development, it is essential to ensure that the determination is free of any interfering material of matrix component. The analysis of blank matrix (saliva, human plasma and rat plasma) from different sources showed that the biological matrix did not have any interference with the investigated drugs. Also, the purity of investigated drugs is ensured by peak purity plots of MX and MT as demonstrated in Fig. S2.† Moreover, detection of MX using FLD ensured the selectivity of the proposed method.

3.3.7. Sensitivity. The proposed method is highly sensitive method showing small values of LLOQ in different biological fluids (Table S2†). Moreover, detection of MX using FLD ensured the sensitivity of the proposed method.

3.4. Application to biological fluids

3.4.1. Analysis of saliva samples. The proposed RP-HPLC method was applied to determine the investigated drugs in saliva. Fig. 3 shows the MT and MX peaks determined in saliva without any interference from its matrix. The assay of drugs is

performed providing good %recoveries and %RSD as demonstrated in Table 3. This indicates that the proposed method allows the sensitive simultaneous determination of MX and MT in saliva.

3.4.2. Plasma analysis (*in vivo* **analysis).** The proposed method was successfully used for sensitive determination of the selected drugs in rats. Fig. 4 demonstrated well separated peaks of MX and MT with high selectivity and no interference with the plasma matrix.

After 1 h of co-administration of single dose of MX (9.2 mg kg⁻¹) and MT (7.5 mg kg⁻¹), analysis of the plasma samples showed well-separated peaks of MX and MT (Fig. 5), where the recovered concentrations of MX and MT were 4.19 \pm 0.45 µg mL⁻¹ and 4.21 \pm 0.64 µg mL⁻¹, respectively.

Moreover, in order to assess the applicability of the proposed method in human plasma, the investigated drugs added to human plasma were analyzed using the proposed method. Results showed very similar elution and retention characteristics to those of rat plasma (Fig. 6).

3.5. Comparison with reported methods

To the best of our knowledge, only two spectrophotometric studies were previously reported for the simultaneous analysis of MT and MX. ^{10,11} Table 4 demonstrates a detailed comparison




Fig. 5 Typical chromatograms of single dose of co-administrated MX and MT in rat plasma after 1 h spiked with IS (a) using DAD at 300 nm and (b) using DAD at 320 nm.




Fig. 6 Typical HPLC chromatograms of (a) blank human plasma using DAD at 300 nm, (b) 20 μ g mL $^{-1}$ MX and 20 μ g mL $^{-1}$ MT spiked in human plasma in presence of 10 μ g mL $^{-1}$ SAL (IS) using DAD at 300 nm, (c) 20 μ g mL $^{-1}$ MX and 20 μ g mL $^{-1}$ MT spiked in human plasma in presence of 10 μ g mL $^{-1}$ SAL (IS) using DAD at 320 nm and (d) 20 μ g mL $^{-1}$ MX and 20 μ g mL $^{-1}$ MT spiked in human plasma in presence of 10 μ g mL $^{-1}$ SAL (IS) using FLD at $\lambda_{ex}=300$ nm and $\lambda_{em}=460$ nm.

Table 4 Analytical comparison between the proposed RP-HPLC method and reported spectrophotometric methods for simultaneous determination of MX and MT

	Proposed RP-HPLC method			Reported method ¹⁰		Reported method ¹¹	
Parameter	MX (DAD)	MX (FLD)	MT	MX	МТ	MX	MT
$\lambda_{\rm nm}$ (nm)	300	$\lambda_{\mathrm{ex}} = 300$ $\lambda_{\mathrm{em}} = 460$	320	200-500		295.4	231.8
Linearity range (μg mL ⁻¹)	0.2-100	0.1-100	0.2-80	5-25	7-35	1-10	3-150
$LOQ (\mu g mL^{-1})$	0.18	0.1	0.19	5.00	7.00	1.00	3.00
Application	It was applied in variable biological fluids			It was applied only		It was applied only in	
	It is the most sensitive method			in pharmaceutical formulations		synthetic mixture	

of our proposed method and the reported ones. Chromatographic analysis offers many advantages for the analysis of various compounds. Thus, our proposed RP-HPLC method shows many advantages over the reported ones. It shows higher selectivity and sensitivity. Also, the proposed method provides peak purity plots to ensure the purity of each drug that allows the determination of the studied drugs in complex biological matrixes. Moreover, using the fluorescence detector enhances the selectivity and sensitivity of the method. The present work offers several applications that not available in other reported methods. It is applied for the simultaneous determination of MT and MX in various biological fluids including saliva, and plasma that permits the fast therapeutic drug monitoring of such combination and studying the pharmacokinetics parameters and bioavailability of both co-administered drugs.

Conclusion

Paper

In this work, a selective and sensitive HPLC method has been developed and validated for the simultaneous determination of MX and MT in different biological fluids using dual DAD and FLD. MX and MT were determined by the DAD at 300 nm and at 320 nm, respectively and MX was determined using FLD at $\lambda_{\rm ex}=300\,$ nm and $\lambda_{\rm em}=460\,$ nm. The proposed method was successfully applied for the simultaneous determination of MT and MX in saliva and plasma samples. Thus, this method is valuable for clinical studies and therapeutic monitoring of both drugs, simultaneously or as single components. Also, it introduces a facile analytical method for the fast therapeutic drug monitoring of both co-administered drugs.

Ethical statement

All animal experiments were conducted according to a protocol approved by our institutional animal care and use committee and in compliance with the Guide for Care and use of Laboratory Animals of the Institute for Laboratory Animal Research of the National Academy of Sciences, USA.

Conflicts of interest

There are no conflicts of interest to declare.

References

- 1 National Center for Biotechnology Information, *PubChem Compound Summary for CID 152946 MRJ*, 2021 from https://pubchem.ncbi.nlm.nih.gov/compound/Moxifloxacin.
- 2 National Center for Biotechnology Information, *PubChem Compound Summary for CID 4173 MRJ*, 2021 from https://pubchem.ncbi.nlm.nih.gov/compound/Metronidazole.
- 3 G. Weiss, P. Reimnitz, B. Hampel, E. Muehlhofer and H. Lippert, Moxifloxacin for the treatment of patients with complicated intra-abdominal infections (the AIDA Study), *J. Chemother.*, 2009, **21**(2), 170–180.
- 4 S. H. van den Elsen, L. M. Oostenbrink, S. K. Heysell, D. Hira, D. J. Touw, O. W. Akkerman, *et al.*, Systematic review of

- salivary versus blood concentrations of antituberculosis drugs and their potential for salivary therapeutic drug monitoring, *Ther. Drug Monit.*, 2018, **40**(1), 17.
- 5 Nord C. E., Pharmacology of moxifloxacin—absorption, distribution, metabolism and excretion, First International Moxifloxacin Symposium, Springer, 2000.
- 6 E. R. Pähkla, T. Koppel, M. Saag and R. Pähkla, Metronidazole concentrations in plasma, saliva and periodontal pockets in patients with periodontitis, *J. Clin. Periodontol.*, 2005, 32(2), 163–166.
- 7 S. Suryawati and B. Santoso, Pharmacokinetics of metronidazole in saliva, *Int. J. Clin. Pharmacol.*, *Ther. Toxicol.*, 1991, **29**(12), 474–478.
- 8 E. D. Ralph, J. T. Clarke, R. D. Libke, R. P. Luthy and W. M. Kirby, Pharmacokinetics of metronidazole as determined by bioassay, *Antimicrob. Agents Chemother.*, 1974, **6**(6), 691–696.
- 9 C. M. Culley, M. K. Lacy, N. Klutman and B. Edwards, Moxifloxacin: clinical efficacy and safety, *Am. J. Health-Syst. Pharm.*, 2001, **58**(5), 379–388.
- 10 G. P. Ertokuş and Ü. M. ÇELİK, Chemometric Analysis of Moxifloxacin and Metronidazole in Binary Drug Combinations With Spectrophotometric Method, *Eastern Anatolian Journal of Science*, 2017, 3(2), 30–37.
- 11 P. Patel, B. Suhagia and M. Patel, Simultaneous spectrophotometric determination of Moxifloxacin and Metronidazole in synthetic mixture by simultaneous equations method, *Indian Drugs*, 2005, 2(3), 155–157.
- 12 U. FDA, *Bioanalytical method validation guidance for industry*. US Department of Health and Human Services Food and Drug Administration Center for Drug Evaluation and Research and Center for Veterinary Medicine, 2018.
- 13 A. F. El-Yazbi, M. G. Guirguis, M. M. Bedair and T. S. Belal, Simultaneous Quantitation of Paracetamol and Lornoxicam in the Presence of Five Related Substances and Toxic Impurities by a Selective HPLC-DAD Method, *J. AOAC Int.*, 2022, 1–7.
- 14 A. F. El-Yazbi, F. M. Aboukhalil, E. F. Khamis, R. M. Youssef and M. A. El-Sayed, Greenness assessment of a stability indicating simple inexpensive high-performance thin-layer chromatography-dual wavelength method for simultaneous determination of mometasone furoate and salicylic acid in complex matrix using analytical eco-scale, *J. Planar Chromatogr.-Mod. TLC*, 2021, 34, 455–466.
- 15 A. F. El-Yazbi, N. E. Elashkar, H. M. Ahmed, W. Talaat and K. M. Abdel-Hay, Cost-effective green chromatographic method for the simultaneous determination of four commonly used direct-acting antiviral drugs in plasma and various pharmaceutical formulations, *Microchem. J.*, 2021, 168, 106512.
- 16 A. F. El-Yazbi, Green methods for the simultaneous analysis of pharmaceutical mixtures present in disparate concentration ranges: Application to antidiabetic mixtures, *Microchem. J.*, 2021, 164, 106073.
- 17 A. F. El-Yazbi, F. M. Aboukhalil, E. F. Khamis, R. M. Youssef and M. A. El-Sayed, Simultaneous determination of Mometasone Furoate and salicylic acid in complex matrix

- using green analytical method, *Microchem. J.*, 2021, **163**, 105900.
- 18 A. F. El-Yazbi, Y. Khalifa, M. A. W. Elkhatib and A. F. El-Yazbi, Green Analytical Method for the Determination of Sofosbuvir, Ledipasvir, Ribavirin and Complex Silymarin Flavonoids Simultaneously in Biological Fluids, *Microchem. J.*, 2021, 105964.
- 19 A. F. El-Yazbi, M. G. Guirguis, M. M. Bedair and T. S. Belal, Validated specific HPLC-DAD method for simultaneous estimation of paracetamol and chlorzoxazone in the presence of five of their degradation products and toxic impurities, *Drug Dev. Ind. Pharm.*, 2020, 46(11), 1853–1861.
- 20 E. I. El-Kimary, H. A. Khalil, T. S. Belal and A. F. El-Yazbi, Green stability-indicating capillary electrophoretic method for simultaneous determination of Lesinurad and Allopurinol in tablet dosage form: Degradation kinetics investigation, *Microchem. J.*, 2020, 158, 105199.
- 21 A. F. El-Yazbi, N. E. Elashkar, K. M. Abdel-Hay, W. Talaat and H. M. Ahmed, An Eco-friendly HPTLC-dual wavelength spectrodensitometric method for simultaneous determination of sofosbuvir and ledipasvir in biological and pharmaceutical samples: Stability indicating application, *Microchem. J.*, 2020, 154, 104584.
- 22 A. F. El-Yazbi, E. I. El-Kimary and R. M. Youssef, Hantzsch pre-column derivatization for simultaneous determination

- of alendronate sodium and its pharmacopoeial related impurity: Comparative study with synchronous fluorometry using fluorescamine, *J. Food Drug Anal.*, 2019, 27, 208–220.
- 23 F. Ibrahim, A. F. El-Yazbi, M. M. Wagih and M. A. Barary, Chromatographic determination of zonisamide, topiramate and sulpiride in plasma by a fluorescent 'turnon' chemosensor, *Bioanalysis*, 2017, 9(14), 1049–1064.
- 24 A. F. El-Yazbi, Comparative Validation of the Determination of Sofosbuvir in Pharmaceuticals by Several Inexpensive Ecofriendly Chromatographic, Electrophoretic, and Spectrophotometric Methods, *J. AOAC Int.*, 2017, **100**(4), 1000–1007.
- 25 A. F. El-Yazbi and A. El-Hawiet, Novel Chromatographic Methods for Simultaneous Quantification of Fish and Wheat Germ Oils Mixture in Pharmaceutical Dosage Forms, J. Chromatogr. Sci., 2017, 55, 497–507.
- 26 F. Ibrahim, A. F. El-Yazbi, M. A. Barary and M. M. Wagih, Sensitive inexpensive HPLC determination of four antiepileptic drugs in human plasma: application to PK studies, *Bioanalysis*, 2016, 8(21), 2219–2234.
- 27 A. F. El-Yazbi and R. M. Youssef, Eco-friendly HPTLC method for assay of Eszopiclone in pharmaceutical preparation: Investigation of its water-induced degradation kinetics, *Anal. Methods*, 2015, 7, 7590–7595.