


# Lab on a Chip

Devices and applications at the micro- and nanoscale

[rsc.li/loc](https://rsc.li/loc)







ISSN 1473-0197

**PAPER**

Pouyan E. Boukany *et al.*  
High-throughput mechanophenotyping of multicellular  
spheroids using a microfluidic micropipette aspiration chip


 Cite this: *Lab Chip*, 2023, 23, 1768

## High-throughput mechanophenotyping of multicellular spheroids using a microfluidic micropipette aspiration chip†

 Ruben C. Boot, <sup>a</sup> Alessio Roscani,<sup>a</sup> Lennard van Buren,<sup>b</sup> Samadarshi Maity, <sup>a</sup> Gijsje H. Koenderink <sup>b</sup> and Pouyan E. Boukany <sup>\*a</sup>

Cell spheroids are *in vitro* multicellular model systems that mimic the crowded micro-environment of biological tissues. Their mechanical characterization can provide valuable insights in how single-cell mechanics and cell-cell interactions control tissue mechanics and self-organization. However, most measurement techniques are limited to probing one spheroid at a time, require specialized equipment and are difficult to handle. Here, we developed a microfluidic chip that follows the concept of glass capillary micropipette aspiration in order to quantify the viscoelastic behavior of spheroids in an easy-to-handle, more high-throughput manner. Spheroids are loaded in parallel pockets *via* a gentle flow, after which spheroid tongues are aspirated into adjacent aspiration channels using hydrostatic pressure. After each experiment, the spheroids are easily removed from the chip by reversing the pressure and new spheroids can be injected. The presence of multiple pockets with a uniform aspiration pressure, combined with the ease to conduct successive experiments, allows for a high throughput of tens of spheroids per day. We demonstrate that the chip provides accurate deformation data when working at different aspiration pressures. Lastly, we measure the viscoelastic properties of spheroids made of different cell lines and show how these are consistent with previous studies using established experimental techniques. In summary, our chip provides a high-throughput way to measure the viscoelastic deformation behavior of cell spheroids, in order to mechanophenotype different tissue types and examine the link between cell-intrinsic properties and overall tissue behavior.

 Received 14th November 2022,  
 Accepted 16th February 2023

DOI: 10.1039/d2lc01060g

[rsc.li/loc](https://rsc.li/loc)

## Introduction

Physical properties like cellular mechanics are of undeniable importance in physiological processes such as morphogenesis,<sup>1</sup> tissue remodeling,<sup>2</sup> wound-healing<sup>3</sup> and cancer growth.<sup>4,5</sup> During these events, cells are collectively confined, squeezed, pushed or pulled upon, affecting their self-organization in time and space. The overall mechanical response to these forces, termed tissue mechanics, will shape the resulting tissue morphology. This response depends on the properties of the single cells alongside the interplay between cells across multiple length scales.<sup>6,7</sup> While the mechanical deformation of single cells mostly depends on their cytoskeleton, plasma membrane and nuclear stiffness,

tissue mechanics are defined through the linkage between cell adhesion molecules, the cytoskeleton and the extracellular environment.<sup>8,9</sup>

Cell spheroids have become a popular *in vitro* model to study tissue mechanics, as they replicate both the multicellular nature and three-dimensional (3D) micro-environment of *in vivo* tissues.<sup>10</sup> These spherical aggregates are made from immortalized cell lines or primary cells that adhere to each other and collectively round up. The resulting spheroid morphology and internal cell arrangement is defined by the interplay between cell-cell adhesion and cortical tension.<sup>7,11,12</sup> Probing spheroids with relevant physical forces therefore increases insight in how tissue composition and resulting mechanics relate to tissue sorting, cellular mechanosensing and cell invasion.<sup>6,13,14</sup>

Spheroid mechanics have been quantified using various techniques, probing either from within or from outside.<sup>15</sup> For example, hydrogel mechanosensors give information on the spatial distribution of mechanical stress within spheroids.<sup>16,17</sup> Cavitation rheology probes the internal elasticity and tissue interfacial tension by inducing a spherical cavity in the spheroid with a needle.<sup>18</sup> From

<sup>a</sup> Department of Chemical Engineering, Delft University of Technology, Delft, The Netherlands. E-mail: p.e.boukany@tudelft.nl

<sup>b</sup> Department of Bionanoscience, Kavli Institute of Nanoscience Delft, Delft University of Technology, Delft, The Netherlands

† Electronic supplementary information (ESI) available. See DOI: <https://doi.org/10.1039/d2lc01060g>



outside, the elastic modulus has been quantified by squeezing the spheroid between two “chopsticks” termed microweavers.<sup>19</sup> Atomic force microscopy (AFM) determines the viscoelastic response of a spheroid by indenting the surface with a nano-probe,<sup>20,21</sup> while tissue surface tensiometry (TST) squeezes the spheroid between two plates<sup>7,11,22–24</sup> and micropipette aspiration (MPA) aspirates a spheroid tongue in a glass capillary to look at the viscoelastic creep response.<sup>13,25–27</sup> Additionally, TST and MPA quantify a tissue surface tension, for which the analogy is made between round spheroids and liquid droplets.<sup>11,13,28</sup> This surface tension is directly related to tissue sorting, tissue spreading and energetic constraints on the size of spheroids.<sup>7,26,29</sup>

However, available techniques to quantify spheroid mechanical parameters such as the elastic modulus  $E$ , viscosity  $\eta$  and surface tension  $\gamma$  have a limited throughput. First, techniques such as AFM, microweavers, TST or MPA only probe one spheroid at a time. Second, the handling of nano-cantilevers, small tweezers or glass microcapillaries is a delicate, difficult and time-consuming task. The resulting low throughput is found in the previously mentioned studies probing spheroid mechanics, as these usually report a data set that ranges between a total number of  $\sim 5$  to 30 probed spheroids. As the technical challenges form a bottleneck on the size of data sets, it is difficult to quantify differences between various spheroid models using present techniques. Given the fact that biological variability tends to be rather large, mechanical phenotyping, for instance to compare different cancer types, or in-depth studies of the role of cytoskeletal components or specific (cancer) biomarkers in overall tissue behavior require an assay with higher throughput than what is currently available.

Microfluidic devices are widely used to measure the mechanical properties of single cells at high throughput.<sup>30–40</sup> Here, cell deformability is examined by letting large numbers of cells flow or migrate through narrow channels or micro-pillars. The chip's defined geometries are easily replicated into new chips, making this a highly reproducible set-up. Besides overall cell deformability, these devices are able to quantify more specific mechanical parameters such as both the cell's and nuclear elastic modulus and viscosity. For example, the design of a microfluidic array where single cells land in individual pockets and are aspirated *via* a pressure gradient enables high-throughput micropipette aspiration.<sup>34,35</sup> However, applying the same principles to study viscoelastic mechanics of spheroids or tissues, which requires microfluidic chips with larger channel dimensions, has remained unaddressed. While microfluidic devices exist that examine spheroid growth, functionally assess drugs or perform high-throughput compression of *Drosophila* embryos,<sup>41–43</sup> a high-throughput microfluidic chip to study spheroid viscoelastic mechanics in parallel and with high reproducibility does not exist to date.

Inspired by the microfluidic micropipette array for single cells,<sup>34,35</sup> we have designed a microfluidic device to perform MPA on multiple spheroids in parallel, thereby drastically

increasing the throughput. After each measurement, spheroids can be easily removed from the device by reversing the flow and aspirating them at the inlet, allowing for multiple experiments per chip. As the device is made from a mold, each chip has the exact same dimensions for the micropipette channels, which is much harder to obtain when pulling glass micropipettes for traditional MPA. The aspiration pressure is precise and easily controlled for each measurement by using hydrostatic pressure. With a custom-made Python script for automated image analysis, the creep length of aspirated spheroid tongues can be analyzed to derive the viscoelastic response of the spheroids. Our device can aspirate 8 spheroids in parallel per measurement for multiple runs per day, allowing for much larger data sets while providing the same information as traditional glass micropipette aspiration. Additionally, we show that our microfluidic device is sensitive enough to pick up mechanical differences between different spheroid models, making it a suitable device to mechanically phenotype different cellular systems.

## Materials and methods

### Cell culture

Human embryonic kidney 293 (HEK293T) cells were generously provided by the group of Dimphna Meijer (Department of Bionanoscience, Delft University of Technology). They were kept in Dulbecco's Modified Eagle Medium High Glucose (DMEM, Sigma) containing 4.5 g L<sup>-1</sup> glucose, L-glutamine but without sodium pyruvate, and supplemented with 10% Fetal Bovine Serum (FBS, Sigma) and 1% antibiotic-antimycotic solution (Gibco).

NIH3T3 embryonic mouse fibroblasts (ATCC CRL-1658) were kept in Dulbecco's Modified Eagle Medium High Glucose (DMEM, Sigma) containing 4.5 g L<sup>-1</sup> glucose, L-glutamine but without sodium pyruvate, and supplemented with 10% newborn calf serum (NCS, Sigma) and 1% antibiotic-antimycotic solution (Gibco).

Human mammary MCF10A cells (ATCC CRL-10317) were cultured in DMEM/F12 1:1 medium (Gibco) supplemented with 5% horse serum (Gibco), 0.5 g mL<sup>-1</sup> hydrocortisone (Sigma), 20 ng mL<sup>-1</sup> human epidermal growth factor (hEGF) (Peprotech), 100 ng mL<sup>-1</sup> cholera toxin (Sigma), 10 g mL<sup>-1</sup> insulin (Human Recombinant Zinc, Gibco) and 1% Penicillin–Streptomycin 100 $\times$  solution (VWR Life Science).

All cells were incubated at 37 °C with 5% CO<sub>2</sub> and subcultured at least twice a week.

### Spheroid fabrication

Spheroids were generated using a custom-designed microfabricated microwell array platform (which is available at: [https://github.com/RubenBoot/HighThroughput\\_Spheroid\\_MPA/blob/main/SpheroidMicrowellArray.dwg](https://github.com/RubenBoot/HighThroughput_Spheroid_MPA/blob/main/SpheroidMicrowellArray.dwg)), inspired by work from Minglin Ma's lab.<sup>44</sup> Following their protocol, two microwell array platforms were created using standard soft lithography at the Kavli Nanolab Delft to allow for the creation




of spheroids with different diameters. Using SU-8 2150 photoresist (Kayaku Advanced Materials) and a  $\mu$ MLA laserwriter (Heidelberg Instruments), the master wafers were designed to have an array of circular microposts. The first wafer had posts with a diameter of 200  $\mu\text{m}$  and height of  $220 \pm 20$   $\mu\text{m}$ , while the second wafer had posts with a diameter of 280  $\mu\text{m}$  and a height of  $300 \pm 30$   $\mu\text{m}$ . The master wafers were coated with trichloro(1*H*,1*H*,2*H*,2*H*-perfluorooctyl)silane (Sigma-Aldrich) to allow for easy demolding. Microwell arrays were molded from the wafers using polydimethylsiloxane (PDMS) (Sylgard 184, Dow Corning) and curing agent at a mixing ratio of 10:1 (w/w). The arrays were placed in a 12 well cell culture plate (Thermo Fisher Scientific) using rubber glue (Reprorubber), and then sterilized by thoroughly washing with ethanol and leaving under UV light overnight.

Before seeding cells, the arrays were coated with 1% (w/v) Pluronic® F127 (Sigma-Aldrich) solution to prevent cell adhesion to the PDMS. The pluronic solution was removed from the well after 45 minutes of incubating. A cell suspension with a concentration of  $\pm 1 \times 10^6$  cells in matching cell media, obtained through trypsinization, was deposited in the well with the coated array in order to form spheroids. It is important to note that the resulting spheroid dimensions not only depend on the chosen cell concentration but also on the duration of culture, cell adhesion and the cell type-specific proliferation rate. After deposition, cells divide over the microwells and settle at the bottom due to gravity, where they aggregate into spheroids overnight (Fig. S1†). The spheroids were cultured in the wells for either 2 or 3 days before aspiration experiments, changing the media every day. On the day of the experiment, spheroids were gently washed out of the microwells using the same media and brought into suspension.

### Design and fabrication of the microfluidic chip

The master wafer was created using standard soft lithography at the Kavli Nanolab Delft. The design is available at [https://github.com/RubenBoot/HighThroughput\\_Spheroid\\_MPA](https://github.com/RubenBoot/HighThroughput_Spheroid_MPA). The multi-layer design contains features with different heights, so had to be created in two separate photolithography steps using a  $\mu$ MLA laserwriter (Heidelberg Instruments). The final chip was designed as a combination of two slabs of PDMS, one slab with the aspiration channels 50  $\mu\text{m}$  in height plus the top half of the aspiration pockets  $\pm 150$   $\mu\text{m}$  in height (see Fig. 1A, PDMS slab 1), and the other slab containing the bottom half of the aspiration pockets (PDMS slab 2). The molds for both these slabs were fabricated together on one silicon wafer. To obtain this, the first step of the design was created by spinning SU-8 3050 to an average thickness of 50  $\mu\text{m}$ . For this, the SU-8 was first spun at 500 rpm for 10 seconds with an acceleration of 100 rpm per second and then at 3000 rpm for 30 seconds with an acceleration of 300 rpm per second. Then, the SU-8 was soft baked at 95  $^{\circ}\text{C}$  for 15 minutes, after which the laserwriter wrote the first layer. The wafer was post baked at 65  $^{\circ}\text{C}$  for 1 minute, then at 95  $^{\circ}\text{C}$  for 5 minutes and developed in SU-8 developer. The second layer



**Fig. 1** Overview of the microfluidic chip. (A) 3D schematic showing (left) the top view of the design and a close-up on the pockets and aspiration channels, and (right) a tilted side view showing the heights of the two separate PDMS slabs and the resulting height when bonded together. (B) Photograph of two actual devices, filled with red dye for visualization and a EU 1 cent coin for scale. (C) Schematic of the experimental set-up. Spheroids enter the PDMS chip at the inlet reservoir, which is a pipette cone, after which they are aspirated with hydrostatic pressure by changing the height  $\Delta h$  of the outlet vial that is mounted to a vertical rod. Note that the drawing is not to scale: in reality, the vertical rod was placed next to the microscope, requiring a 60 cm outlet tube to bridge the distance between the chip and the outlet vial. The experiment was captured with an inverted microscope using a 5 $\times$  objective. (D) Brightfield top view image of the aspiration pockets loaded with HEK293T cell spheroids aspirated at 700 Pa, scale bar 200  $\mu\text{m}$ .

was created with SU-8 2050 and spun to an average thickness of 150  $\mu\text{m}$ . It was first spun at 500 rpm for 10 seconds with an acceleration of 100 rpm per second, after which it was spun at 1200 rpm for 30 seconds at an acceleration of 300 rpm per second. It is important to note that the thickness was not equal over the whole wafer, as the resist covered both the first half of the design (consisting of the micropipette channel and part of the aspiration pocket) and the empty place where the second half of the aspiration pockets will be written. Therefore, one half of the aspiration pockets (see Fig. 1A, PDMS slab 2) resulted in a thickness of  $150 \pm 2$   $\mu\text{m}$ , while the other half containing half of the pocket plus the aspiration channel (PDMS slab 1) had a different thickness of  $175 \pm 25$   $\mu\text{m}$ . After spinning, the wafer was soft baked at 65



°C for 5 minutes and then at 95 °C for 30 minutes. The laserwriter wrote the second part of the design, after which the wafer was post baked at 65 °C for 5 minutes, at 95 °C for 12 minutes and then developed. The master wafer was coated with trichloro(1*H*,1*H*,2*H*,2*H*-perfluorooctyl)silane to allow for easy demolding. PDMS chips were created using Sylgard 184 at a curing agent ratio of 10 : 1. Individual chips were cut and holes were punched in only one slab of the design for the later introduction of tubing. Both halves of the design were plasma cleaned (Harrick Plasma) for two and a half minutes at 30 W to facilitate bonding, after which one half was slightly wetted with a droplet of distilled water to allow for better alignment. The two halves were put together and aligned using the alignment arrows at the border of the design and an optical microscope (ZEISS Primovert) to check the alignment (Fig. S2†). Finally, the chip was kept in the oven at 65 °C to bond overnight. The bonding overnight usually resulted in a slight misalignment between both halves of the device, potentially due to shrinking of the PDMS.<sup>45</sup> Across the chips used in this study ( $n = 14$ ), an average misalignment of  $16 \pm 1 \mu\text{m}$  (in the range of 10–22  $\mu\text{m}$ ) between the side walls of the pockets, and an average misalignment of  $12 \pm 1 \mu\text{m}$  (in the range of 5–21  $\mu\text{m}$ ) between the front of the pocket and the aspiration channel was measured from brightfield images using ImageJ. We tested by computer simulations that this slight misalignment did not affect our results (see Results and discussion section). After fabrication, chips could be stored indefinitely and used on the desired day for the experiment.

### Data acquisition

Before each experiment, the chip was filled with 1% Pluronic® F127 (Sigma) solution and left at room temperature to prevent cell adhesion to the PDMS walls. After 45 minutes, the pluronic solution was flushed from the chip using the culture media that matched the cell line used in the experiment. For this, vials with cell-free culture media were connected to the inlet and outlet of the microfluidic chip with PTFE 008T16-030-200 tubing (Diba Industries, inner diameter 0.3 mm, outer diameter 1.6 mm) and a pressure was applied to the media using an MFCS-EZ pressure controller (Fluigent). Once all the pluronic solution, PDMS debris particles and possible air bubbles were flushed out, the tube connected to the inlet was gently unplugged from the chip and a loading reservoir, being a shortened 1 ml pipette tip cut with a scalpel, was plugged into the inlet. Then, the vial connected to the outlet was disconnected from the pressure controller and mounted to a vertical translation stage (Thorlabs, VAP10) with a ruler on the side such that the pressure in terms of  $\text{cmH}_2\text{O}$  could be read off, to be able to exert a precise hydrostatic aspiration pressure when lowering the stage with the vial compared to the height of the reservoir. Slightly lowering the stage induced a minor flow in the chip towards the aspiration pockets, after which 20  $\mu\text{l}$  of spheroid suspension was pipetted into the reservoir.

Spheroids were guided by the flow and entered the aspiration pockets, after which the outlet tube was brought back to the height where no flow is present. For loading, the inlet hole needed to be punched close enough to the aspiration pockets. Otherwise, the volume of space in the loading bay of the chip would be too large and could not induce a high enough flow velocity to sweep the spheroids into the pockets. Spheroids then sedimented to the bottom instead and remained immobile. Once ready to perform the experiment, the vial was manually lowered to the chosen aspiration pressure (in this study 500, 700 and 1500 Pa, depending on the studied cell line), thereby inducing spheroid tongue aspiration. The inlet reservoir volume remained constant during the duration of the experiment, confirming that there was no leakage at the corners of the squared aspiration channels and spheroids fully blocked the channels. Whenever a spheroid did not fully block the channel, a clear flow was visible in the aspiration channel and the rare experiments where this occurred were not used for analysis.

Brightfield images of spheroid tongues entering the aspiration channels were captured on an inverted fluorescence microscope (Zeiss Axio-Observer) every 5 seconds for a total of 5 minutes using a 5 $\times$ /NA 0.16 air objective and ORCA Flash 4.0 V2 (Hamamatsu) digital camera with a resolution of 2048  $\times$  2048 px<sup>2</sup>. We ensured that the full aspiration curve of the tongue was captured by starting the image acquisition before lowering the outlet tube (Movie S1†). At the end of the experiment, spheroids were pushed out of the pockets and flown back towards the inlet by raising the outlet vial above the reservoir. There, they were removed from the chip by pipetting them up through the reservoir. This way, new spheroids could be inserted and a new measurement started with the same chip. All experiments were conducted at 37 °C and 5% CO<sub>2</sub> using a stage top incubator (ibidi). Chips were used for 4–5 successive experiments on average, and were always discarded after the final experiment of the day.

### COMSOL simulations on pressure distribution in the chip

The design of the device contains 8 parallel pockets, and thus pressure will redistribute once spheroids start clogging the flow in pockets. To examine the influence of this effect, the pressure distribution was computationally modeled in the 3D design of the device using the finite elements modeling software COMSOL Multiphysics 5.6. Considering the fluid flow to be laminar and following the Navier–Stokes equation,<sup>46</sup> the pressure distribution was modeled for two different cases: (1) all pockets are open, or (2) all pockets are clogged, except for one where fluid still flows through the aspiration channel.

The Hagen–Poiseuille equation tells us that the pressure drop over a tubular channel with laminar flow scales with the length of the channel and the inverse of the channel radius to the fourth power.<sup>47</sup> As the design of our device has much larger dimensions than typical microfluidic chips, the cross-




sectional area of the tubing connected to the chip now has the same order of magnitude as the cross-sectional area of the device. Therefore, the hydrodynamic resistance across the tubing is considerable and non-negligible as long as a flow is present in the device. It is therefore important to realize that the device is only able to accurately perform spheroid aspiration at a single defined step pressure once *all* pockets are filled with a spheroid, thus blocking the flow.

To examine what pressure spheroids experience in pockets when not all are filled yet, case (1) and (2) were modeled for the design of the device including a 60 cm long rectangular channel with a  $300 \times 300 \mu\text{m}$  cross section that mimics the outlet tube and corresponding pressure drop. The same average length of tubing was used during our experiments. Boundary conditions of 700 Pa at the left edge of the inlet region and 0 Pa at the end of the simulated outlet tube were installed, similar to lowering the outlet tube with  $7 \text{ cmH}_2\text{O}$ . These simulations generated the pressure distribution and corresponding fluid flow profile in the device.

### High-throughput analysis of spheroid tongue aspiration

The creep length of the spheroid tongues into the aspiration channels over time was extracted from the experimental images using Fiji (<https://imagej.net/software/fiji/>) and a custom-written Python script (which is available at: [https://github.com/RubenBoot/HighThroughput\\_Spheroid\\_MPA/blob/main/SpheroidAspiration\\_AnalysisScript.py](https://github.com/RubenBoot/HighThroughput_Spheroid_MPA/blob/main/SpheroidAspiration_AnalysisScript.py)). First, the brightfield aspiration time-lapse images were rotated to make the aspiration channels vertically oriented (with the tongue creeping upward over time), and then cropped using ZEN software (Zeiss). The cropped region captured the whole aspirated spheroid tongue and aligned the beginning of the channels with the bottom of the cropped images. The cropped time-lapse was saved as a JPEG stack, and converted to binary images using a threshold in Fiji (see bottom two frames in Fig. 2A). The threshold value was chosen manually to obtain a clear contrast between the aspirated protrusion edge and the surrounding empty channel. This binary stack was then imported in the Python script. Using the Fiji interface, the  $x$ -coordinates of pixels along the horizontal line were manually inserted in the Python script to indicate where the middle of all 8 aspiration channels was located. The script was set up to find the edge of the aspiration protrusion by checking the binary value of every pixel on the vertical line along these coordinates (from top to bottom) and recording the  $y$ -coordinate corresponding to the protrusion edge. This analysis was repeated for all images in the stack, returning the set of  $y$ -coordinates for all 8 channels and for each time step.  $Y$ -Coordinates (pixels) were converted into creep lengths ( $\mu\text{m}$ ) by subtracting the  $y$ -coordinate (which is the distance from the top of the image to the protrusion edge) from the total height of the image and multiplying this result with the pixel distance. In some experiments, cells at the protrusion edge of the spheroid tongue were loosely attached and would break off during aspiration, thus altering the protrusion edge



**Fig. 2** Automated data analysis using a custom-written Python script. (A) Preparation workflow before running the script. First, the 8 pocket-time series (scale bar  $200 \mu\text{m}$ ) is cropped to the constrictions with spheroid tongues, where the bottom of the cropped images is aligned with the precise start of the aspiration channels. Next, the cropped time series is converted to binary images by setting a threshold using ImageJ. Tongues are analyzed if they have a thresholded edge that does not change structure over time (due to cells breaking off) and belong to spheroids that remain round and retain a constant volume before and during aspiration, here indicated by a red star. (B) Plot of the creep length of the 5 HEK293T cell spheroids (indicated by the red stars in (A), with corresponding pocket numbers) as a function of time, aspirated at 700 Pa. The inset shows how the tongue length is calculated by subtracting the distance from the top of the cropped image to the protrusion edge from the total image height.

structure. Spheroids where this occurred were excluded from analysis, alongside the ones that did not remain round nor retained a constant volume before and during aspiration. All results were collected in a Microsoft Excel file alongside the time step per image.

### Statistical analysis

Statistical analysis was performed using Python and Microsoft Excel. Student  $t$ -tests were executed using the TTEST function in Excel and  $p$  values below 0.05 were considered to be significant. Python was used to do standard error propagation calculations on the fitting parameters to obtain error values. The average human error in reading off



the aspiration pressure when manually lowering the outlet vial is defined by the thickness of the lines on the ruler indicating the distances, which is 0.2 cmH<sub>2</sub>O (~20 Pa), and the error in dimensions of the aspiration channel is around 5 μm (measured with a Dektak stylus profiler). The error bars in the figures display the standard error of the mean unless indicated otherwise, and are always based on at least two independent experiments with separately prepared chips.

## Results and discussion

### High-throughput data extraction

In order to measure mechanics of cell spheroids in a high-throughput manner, we designed a microfluidic device to parallelize micropipette aspiration. The design consists of 8 parallel aspiration pockets that are 250 μm wide and 325 ± 27 μm tall, in order to be able to contain a single spheroid with a maximum diameter of 250 μm. Each pocket connects to an aspiration channel that is 50 μm wide and 50 ± 2.5 μm tall, chosen to be similar in size to pipette diameters used in previous glass micropipette spheroid aspiration studies (Fig. 1A and B).<sup>13,26</sup> We decided to implement 8 pockets as it was the maximum number of pockets that fit in the field of view of the 5× microscope objective. To mimic traditional micropipette aspiration as accurately as possible, the multilayer wafer mold was designed to create symmetric aspiration pockets with aspiration channels positioned at the *centerline* of the pockets. This way, spheroids were raised from the device bottom during aspiration (similar to the single cell aspiration device created by Lee *et al.*).<sup>35</sup> Two PDMS halves from the device mold, one side with half the pocket and the other with the other half of the pocket plus the aspiration channel, were aligned under a microscope using alignment arrows incorporated in the design (see ESI† Fig. S2) and bonded in the oven overnight.

The chip was flushed with media before starting an experiment. Then, a small volume of spheroid suspension (~20 μL) was added to the inlet reservoir. The spheroids were moved into the pockets through the flow induced by lowering a media reservoir connected to the outlet and mounted to a vertical rod, thereby exerting a precise hydrostatic pressure gradient that varied between 500 and 1500 Pa in our experiments, depending on the studied cell type (Fig. 1C). Once a spheroid arrived in the pocket, it blocked the flow through the aspiration channel, thus preventing other spheroids from entering the same pocket. Induced pressure gradients, by lowering the outlet media reservoir, were at first kept low enough (below ~200 Pa) to ensure that spheroids did not deform in the aspiration channels yet. Only when all pockets were loaded with spheroids was the outlet reservoir lowered to a level that induced the chosen step pressure for aspiration. Spheroid tongues started creeping in the aspiration channels as they were subjected to the pressure difference between the atmospheric pressure at the inlet reservoir and the hydrostatic pressure exerted by the outlet media tube (Fig. 1D). In this way, spheroids experienced the


aspiration force in a similar manner as for traditional micropipette aspiration, where spheroids are kept at atmospheric pressure and aspirated by applying an underpressure in a glass capillary.

We developed a custom-made Python script to analyze the spheroid tongue deformations in an efficient and fast manner. Before running the script, brightfield time-lapse image series were rotated, cropped and changed into binary images to focus on the creeping protrusions in the channels (Fig. 2A). To determine the creep length  $L(t)$ , the script derived the distance from the top of the binarized image to the protrusion edge, and subtracted this value from the total image height. The Python script thus extracted the creep lengths for the 8 aspiration channels and each step in time (Fig. 2B). We excluded spheroids that were not round before or did not remain constant in volume during aspiration.

### Pressure distribution in the chip

In contrast to microfluidic single cell aspiration chips, where channels have smaller dimensions than the used tubing,<sup>34,35</sup> the required cross-section for channels to flow undeformed spheroids in is as large as the tubing (with a diameter of 300 μm). Therefore, the pressure drop over the tubing is non-negligible in comparison to the pressure drop across the chip. This is a considerable problem, as slight changes in the length of the tubing may cause significant changes in the experienced pressure drop in the microfluidic device. To avoid these pressure drop effects, we designed the device in such a way that the entire flow is stopped once all pockets are filled with spheroids and the pressure gradient is fully defined by the hydrostatic pressure. However, spheroids that are loaded into the device and swept along by the induced flow will reach the aspiration pockets one at a time. Therefore, pressures will redistribute in the device for each spheroid that clogs an aspiration pocket and stops the flow through the adjacent aspiration channel. This might cause a problem, as the spheroids that already arrived in the pockets may experience an aspiration pressure that increases over time when additional pockets fill up with spheroids. Instead, all spheroids should experience a single step pressure to allow correct data analysis. To examine the effect of pressure redistribution, we performed 3D COMSOL simulations for two different cases: (1) all pockets are open, and (2) all pockets are clogged by spheroids except for one. A 60 cm long rectangular channel with a 300 × 300 μm cross section was incorporated into the 3D design to mimic the outlet tube, as it induces a considerable pressure drop, and a pressure gradient of 700 Pa was simulated across this total geometry. The model showed how the pressure gradient over the aspiration channel increases from ~50 Pa when all pockets are open to ~300 Pa when all but one are clogged (Fig. 3A). It is important to note that these values are dependent on the chosen boundary conditions, and in this context simply serve to estimate the extent of this effect. Additionally, we ran an identical simulation as case (2), only





**Fig. 3** Pressure distribution across aspiration pockets. (A) A 3D numerical simulation of the pressure distribution assuming the device is connected to a 60 cm long straight rectangular channel mimicking the outlet tube. A pressure gradient of 700 Pa is simulated for two different configurations: all pockets are open (left), and all pockets but one are clogged by spheroids (right). (B) Brightfield images from an HEK cell spheroid aspiration experiment, demonstrating how simultaneous aspiration starts as soon as all channels are clogged. At first (a), no pressure gradient is present yet and three pockets remain open, indicated by red arrows. Five seconds later (b), a pressure gradient of 500 Pa has been induced and spheroids flow into the pockets but do not clog everything yet. After five more seconds (c), all spheroids have clogged the pockets, thereby blocking further flow and ramping up the pressure to 700 Pa. Spheroids now experience the full induced hydrostatic pressure gradient as evident from the tongues all creeping simultaneously into the aspiration channels. Scale bar 200  $\mu\text{m}$ .

now both halves of the design were slightly misaligned by 20  $\mu\text{m}$  in both alignment directions, to test whether the variations between devices had any influence on the pressure drop across the aspiration channels (Fig. S3†). We found that the pressure changed by only a few pascal, demonstrating that the effect of slight misalignment on the aspiration pressure is negligible.

We concluded from these simulations that it is important to ensure that spheroids arrive at the pockets at approximately the same moment in time. Otherwise, spheroids might experience a

pressure that starts aspiration when *almost* all pockets are clogged but jumps once the final pocket clogs. In experiments, we discovered that spheroids all start aspirating at the same time as long as we punch the inlet hole close to the pockets and add sufficient spheroids to the inlet reservoir (Fig. 3B). To circumvent the effect of pressure redistribution, spheroids were gently flown towards the pockets by inducing a minor pressure gradient ( $\sim 150$  Pa) and were halted before the pockets by bringing the height of the outlet vial back to the starting point with no flow. Then, the final pressure gradient was induced by lowering the outlet vial again, and all spheroids that were floating near the pockets experienced the flow and filled up the remaining unclogged pockets at the same time. This ramped up the pressure, and the spheroids experienced the full pressure gradient, thus starting the measurement at the same time for all pockets.

### Validation of sensitivity and reproducibility when working at different aspiration pressures

To confirm the sensitivity and reproducibility of our device, we measured the deformability of HEK293T spheroids at two slightly different aspiration pressures of 500 and 700 Pa. The obtained creep data alongside visual confirmation showed that the aspirated spheroids displayed an initial elastic deformation followed by a viscous response (Fig. 4A). The tissue relaxation time  $\tau$  demarcates these two regimes and is given by  $\tau = \eta/E$ , with  $\eta$  being the viscosity and  $E$  the elastic modulus of the spheroid. We fitted this viscoelastic response to different spring-and-dashpot models and compared their coefficient of determination (see Fig. S4†). The modified Maxwell model, previously used in studies on spheroid micropipette aspiration,<sup>13,25,26</sup> proved to fit the data significantly better than other more simple models (Table S1†) and was therefore used to determine the mechanical parameters of our spheroid aspiration measurements. The model consists of four elements (shown in the inset of Fig. 4A): a dashpot  $\zeta_t$  in series with a modified Kelvin-Voigt element, which consists of a spring  $k_1$  in parallel with a dashpot  $\zeta_c$  and spring  $k_2$  in series. The creep length  $L(t)$  in the context of this model is given by:

$$L(t) = \frac{f}{k_1} \left( 1 - \frac{k_2}{k_1 + k_2} e^{-\frac{t}{\tau_c}} \right) + \frac{f}{\zeta_t} t \quad (1)$$


where  $f$  is the aspiration force,  $k_1$  is the spring constant for the elasticity of the spheroid,  $k_2$  relates to the initial increase in  $L(t)$ ,  $\zeta_t$  represents the viscous dissipation of the flowing tongue and  $\tau_c$  is the rising time of the elastic deformation  $\delta$ .

Here,  $\tau_c = \frac{\zeta_c(k_1 + k_2)}{k_1 k_2}$  with  $\zeta_c$  being a friction coefficient related to the rising time. Hence, the modified Maxwell model has four fitting parameters:  $\delta = f/k_1$ ,  $\dot{L}_\infty = f/\zeta_t$  (being the flow velocity at long timescales),  $\beta = k_2/(k_1 + k_2)$  and  $\tau_c$ .

As shown in Fig. 4B, the creep length  $L(t)$  of the spheroid tongue was smaller for measurements at 500 Pa in comparison to 700 Pa, demonstrating the sensitivity of our







**Fig. 4** Validation of sensitivity of the device to determine spheroid mechanical properties. (A) Comparison plot of the creep length versus time for one HEK cell spheroid aspirated at 700 Pa (orange) and another at 500 Pa (blue). The creep curves are fitted with the modified Maxwell model (black dashed lines), which is explained in the inset. The fast elastic deformation  $\delta$  equals the intercept of the linear viscous flow with the  $y$ -axis, and is shown to be larger when aspirating spheroids at larger pressures. (B) Brightfield images of two separate aspiration experiments for HEK cell spheroids, showing the creep length  $L(t)$  after 23 s of aspiration at 500 Pa (top) or 700 Pa (bottom), with snapshots focusing on the aspiration pocket (left, scale bar 50  $\mu\text{m}$ ) and zoomed in on the aspiration channel (right, scale bar 20  $\mu\text{m}$ ). (C) Histograms comparing the average elastic deformation  $\delta$  for HEK cell spheroids aspirated at different pressures (500 Pa and 700 Pa).  $***, p < 0.001$  with  $n = 56$  and 24 for 500 Pa and 700 Pa respectively. Error bars are SEM.

device. However, creep length can only be compared for a precise point in time. To quantify differences between the entire measurements, we compared the average elastic deformation  $\delta$  by fitting eqn (1) to the creep data. The average deformation  $\delta$  was significantly smaller for measurements performed at 500 Pa ( $\delta = 70 \pm 2 \mu\text{m}$ ) in comparison to 700 Pa ( $\delta = 90 \pm 4 \mu\text{m}$ ), as one would expect when exerting a smaller force (Fig. 4C). This demonstrates how the device is sensitive enough to work at small differences in pressure and create reproducible deformations in agreement with theoretical expectations.

### Measurements on spheroids with known mechanical properties


To demonstrate the high-throughput mechanical phenotyping capabilities of our device, we measured the mechanical properties of three different cell spheroid models that have been probed in previous studies.<sup>48</sup> In addition to our measurements performed with HEK293T cell spheroids, we studied two stiffer spheroid models made of either NIH3T3 fibroblasts or MCF10A cells. While the HEK293T cell spheroids were probed at 500 and 700 Pa, the two stiffer spheroid models were aspirated at a higher pressure of 1500 Pa as lower pressures would induce a slower deformation and require a longer time scale to analyze the full viscoelastic response (Fig. 5A).

In traditional micropipette aspiration, the aspiration force  $f$  of the pipette when considering spheroid volume conservation is given by:

$$f = \pi R_p^2 \Delta P, \quad (2)$$

where  $R_p$  is the radius of the pipette and  $\Delta P$  the applied underpressure in the pipette.<sup>49</sup> Previous studies on micropipette aspiration of spheroids have pointed out that

the actual pressure exerted on the spheroid equals the applied underpressure  $\Delta P$  minus a critical pressure  $\Delta P_c$  at which aspiration of the spheroid occurs. When aspirating at a pressure lower than  $\Delta P_c$ , the spheroid will not deform due to its inherent tissue surface tension  $\gamma$ .<sup>13,26</sup>



**Fig. 5** High-throughput comparison of spheroid mechanics for different cell lines. (A) Brightfield snapshots after 23 s of aspiration at 1500 Pa of a NIH3T3 (left) and MCF10A (right) spheroid, scale bars 50  $\mu\text{m}$ . (B) Comparison plot of creep length versus time for a HEK293T spheroid (purple) aspirated at 700 Pa, and a NIH3T3 (green) and MCF10A (orange) spheroid aspirated at 1500 Pa. The creep curves are fitted with the modified Maxwell model (black dashed lines). (C) Average elastic moduli and (D) average viscosity measured for the three different cell lines.  $***, p < 0.001$  and ns is nonsignificant with  $n = 80, 75$  and 34 for HEK293T, NIH3T3 and MCF10A respectively. Error bars are SEM.



To calculate this critical pressure, separate measurements of the retraction of the spheroid tongues are required when the pressure is brought back to zero (see ESI†). In contrast with traditional glass micropipette aspiration, where spheroids remain stuck in the pipette and the tongue slowly retracts, on our device spheroid tongues retracted so fast (displaying a mainly elastic response) that it was impossible to measure a viscoelastic retraction curve when the pressure was brought back to zero (see Movie S2†). This effect was consistent for all three cell lines, even when a small aspiration pressure of 100 Pa was kept during retraction to make certain that this fast retraction was not caused by backflow in the chip (see Movie S3†). The effect was also not governed by the surface treatment of the aspiration channels, as experiments performed on a chip that was not coated with pluronic F-127 still resulted in an elastic fast retraction that took tens of seconds (see Movie S4†). We therefore investigated the effect of accounting for  $\Delta P_c$  through traditional micropipette measurements on HEK293T spheroids (measuring both aspiration and retraction, see ESI†). We found that  $\Delta P_c$  ranged between 50 and 150 Pa (when aspirating at 500 Pa), which changes parameter values obtained by fitting aspiration curves by maximally 10–30% (Fig. S5 and Table S2†). Additionally, we saw that the spheroids deformed differently in the glass micropipettes in comparison to our device, displaying a lower elastic deformation  $\delta$  and flow velocity  $\dot{L}_\infty$ . One possible explanation for the discrepancy between our device and traditional MPA is that the chip has square aspiration channels instead of round capillaries, potentially influencing the creep of the tongue due to the different geometry or due to wall friction (PDMS instead of glass). Moreover, the pockets on our chip have rounded walls which the spheroids potentially push back on during relaxation, making them move slightly backwards in the pocket during retraction. We conclude that creep curves obtained with our microfluidic device slightly deviate from glass micropipette aspiration measurements. However, our device has the unique benefit of providing measurements at high throughput, thus allowing systematic comparisons between different cell types. The creep curves obtained with our microfluidic device showed that the two stiffer spheroid models made of NIH3T3 or MCF10A cells had a lower initial elastic deformation  $\delta$  and slower viscous flow  $\dot{L}_\infty$  than the more deformable HEK293T spheroids (Fig. 5B). The observed linear viscous response, where creep length does not reach a plateau after the elastic elongation but keeps increasing over time, has also been observed in previous aspiration studies on spheroids that used traditional glass micropipettes.<sup>13,26</sup> Where in a liquid, viscosity arises from friction between the molecules, for these spheroids the differences in viscous response potentially arise from differences in cell adhesion and cortical strength between the three cell types.<sup>11</sup> All curves were fitted with the modified Maxwell model to extract the relevant mechanical parameters. The first term in eqn (1) characterizes the elastic regime, with  $k_1 = \pi R_p E$ , while the second term represents the

flow at constant velocity  $\dot{L}_\infty$  for longer timescales, with  $\xi_t = 3\pi^2 \eta R_p$ .<sup>13,50</sup> As our microfluidic device, unlike traditional glass micropipettes, does not present cylindrical constrictions, a correction for rectangular constrictions needs to be implemented in regards to  $R_p$ , as previously done by Davidson *et al.*<sup>34,51</sup> The effective channel radius  $R_{\text{eff}}$  is now given by:

$$R_{\text{eff}}^4 = \frac{2}{3\pi} \frac{W \times H^3}{\left(1 + \frac{H}{W}\right)^2 \times f^*}, \quad (3)$$

with  $W$  and  $H$  being the width and height of the rectangular constriction and  $f^*$  being a function of the aspect ratio ( $H/W$ ), given by Son.<sup>51</sup> For  $W$  and  $H$  values of 50  $\mu\text{m}$ , this results in a corrected channel radius  $R_{\text{eff}} = 27 \mu\text{m}$ .

When fitting obtained creep data with eqn (1), extracting fitting parameters  $\delta$  and  $\dot{L}_\infty$  and plugging in eqn (2), the elastic modulus  $E$  is derived as:

$$E = \frac{R_{\text{eff}} \Delta P}{\delta}, \quad (4)$$

and the viscosity  $\eta$  as:

$$\eta = \frac{1}{3\pi \dot{L}_\infty} R_{\text{eff}} \Delta P, \quad (5)$$

where we thus chose not to include  $\Delta P_c$  as our chip is not able to capture retraction measurements.

The measured average elastic modulus and viscosity of the HEK293T cell spheroids were the same for different applied pressures of 500 and 700 Pa (Fig. S6†). The two data sets were not statistically significant from each other, allowing us to consider them as one population when comparing them with the other two spheroid models. For both NIH3T3 and MCF10A cell spheroids, the average elastic moduli were almost four times as large in comparison to the softer HEK293T cell spheroids, but were not significantly different from each other (Fig. 5C). Similarly, both models had a viscosity almost six times larger than for HEK293T spheroids but again were not significantly different from each other (Fig. 5D). Additionally, as all spheroids varied in size, we investigated a possible relationship between spheroid size and their mechanical parameters. However, the derived elastic moduli and viscosities proved to be independent of spheroid radius  $R_0$  (in the range of 60–125  $\mu\text{m}$ ) for all three spheroid models (Fig. S7†).

The obtained values are consistent with values measured in previous studies (Table 1). For HEK293T cell spheroids, we measured an average elastic modulus of  $\sim 210$  Pa which

**Table 1** Spheroid mechanical parameters for different cell lines, derived using the modified Maxwell model and performing a least squares regression of the experimental creep curves

Cell line	$E$ (Pa)	$\eta$ (kPa s)	$n$
HEK293T	210 ( $\pm 5$ )	12 ( $\pm 1$ )	80
NIH3T3	780 ( $\pm 30$ )	67 ( $\pm 3$ )	75
MCF10A	770 ( $\pm 30$ )	67 ( $\pm 5$ )	34



agrees with parallel-plate compression on HEK293 cell spheroids measuring a range of 200–400 Pa.<sup>52</sup> Our NIH3T3 spheroids had an average modulus of ~780 Pa, which falls within the range of 500–3500 Pa measured with colloidal probe AFM.<sup>53</sup> However, the MCF10A spheroids displayed an average modulus of ~770 Pa which is just below the range of  $1250 \pm 320$  Pa measured by squeezing MCF10A spheroids with microtweezers.<sup>19</sup> This discrepancy could be explained by the fact that this range was determined for only 6 spheroids in the microtweezer study, or potentially squeezing might induce different deformation behavior compared to aspiration. Additionally, the values obtained with our device for HEK293T cell spheroids at 500 Pa differed from the values obtained with traditional glass micropipette aspiration (see Table S2†). Aspiration with glass cylinder pipettes (at room temperature) resulted in an elastic modulus  $E$  that was approximately two-thirds larger and a viscosity  $\eta$  that was three times as large in comparison to the values measured with our microfluidic device (at 37 °C). Aside from the possible influence of temperature, this distinction is due to differences in dissipation and the rate of rearrangement of cells at the entry of the squared as opposed to cylindrical aspiration channels. Furthermore, we had to omit the critical pressure  $\Delta P_c$  from our derivations, affecting the absolute values of  $E$  and  $\eta$  with an estimated 10–30% based on our traditional MPA measurements on HEK293T cell spheroids. Nevertheless, our results together with previous studies show that relative differences in the mechanical properties of different cell types are much larger than the discrepancies that arise between microfluidic and glass MPA or between different measuring techniques when determining absolute values. Altogether, these results demonstrate that our device is well capable to measure the viscoelastic behavior of multicellular spheroids and determine their mechanical properties in agreement with other experimental techniques.

### Outlook and conclusions

We have developed a microfluidic chip that follows the principles of traditional micropipette aspiration to quantify the viscoelastic response of cell spheroids in an efficient manner, making it possible to make statistically meaningful comparisons between different experimental conditions. The chip performs viscoelastic creep measurements as soon as spheroids fill up the 8 parallel aspiration pockets and block further fluid flow. The design of the chip can in principle be adapted to obtain smaller or larger aspiration channels, but this will influence the overall volumetric flow rate through the chip and thus the ease of loading spheroids. Similarly, the number of pockets could be increased, though this would make it more difficult to load them. With the current geometry, our chip is able to obtain reproducible and accurate results and to detect differences upon small changes in pressure. Our results are in agreement with values from previous studies and demonstrate the high-throughput aspect of the chip in context of the much lower output traditional

glass micropipette aspiration: the chip is efficient and easy to use in contrast to the cumbersome re-use of glass pipettes, and provides large amounts of data in a smaller amount of time. Therefore, the microfluidic device presented here is a suitable technique to investigate the mechanics of a wide range of tissues, from embryonic to tumor, to provide mechanistic insights in important physiological processes such as tissue remodeling and cancer metastasis.

### Author contributions

R. C. B., S. M., G. H. K. and P. E. B. conceived the ideas and designed the experiment. R. C. B. and A. R. carried out the experiments and collected the data. R. C. B. and L. B. performed the traditional glass micropipette aspiration measurements. R. C. B., S. M. and L. B. performed the computational modeling and analysis. R. C. B. wrote the paper, and G. H. K. and P. E. B. edited it.

### Conflicts of interest

There are no conflicts to declare.

### Acknowledgements

R. C. B. and P. E. B. gratefully acknowledge funding from the European Research Council (ERC) under the European Union's Horizon 2020 research and innovation program (grant agreement no. 819424). A. R. was supported by an Erasmus+ Traineeship. G. H. K. gratefully acknowledges funding from the VICI project How cytoskeletal teamwork makes cells strong (project number VI.C.182.004) which is financed by the Dutch Research Council (NWO). The authors thank Dimphna Meijer for the HEK293T cells, and thank Timon Idema for helpful discussions.

### References

- 1 C. Hahn and M. A. Schwartz, *Nat. Rev. Mol. Cell Biol.*, 2009, **10**, 53–62.
- 2 T. Mammoto and D. E. Ingber, *Development*, 2010, **137**, 1407–1420.
- 3 A. Brugués, E. Anon, V. Conte, J. H. Veldhuis, M. Gupta, J. J. Muñoz, G. W. Brodland and B. Ladoux, *Nat. Phys.*, 2014, **10**, 683–690.
- 4 H. T. Nia, H. Liu, G. Seano, M. Datta, D. Jones, N. Rahbari, J. Incio, V. P. Chauhan, K. Jung, J. D. Martin, T. P. Padera, D. Fukumura, Y. Boucher, F. J. Hornicek, A. J. Grodzinsky, J. W. Baish, L. L. Munn, R. K. Jain, L. S. Laboratories, M. General, T. Medicine, B. Sciences, B. Engineering, H. S. Joao, O. Service, C. T. Oncology and M. G. Hospital, *Nat. Biomed. Eng.*, 2016, **1**, 1–25.
- 5 H. T. Nia, L. L. Munn and R. K. Jain, *Science*, 2020, **370**, eaaz0868.
- 6 K. Jakab, B. Damon, O. Doaga, V. Mironov, I. Kosztin, R. Markwald and G. Forgacs, *Dev. Dyn.*, 2008, **237**, 2438–2449.



- 7 R. A. Foty, C. M. Pflieger, G. Forgacs and M. S. Steinberg, *Development*, 1996, **122**, 1611–1620.
- 8 J. E. Marturano, J. D. Arena, Z. A. Schiller, I. Georgakoudi and C. K. Kuo, *Proc. Natl. Acad. Sci. U. S. A.*, 2013, **110**, 6370–6375.
- 9 N. R. Schiele, F. Von Flotow, Z. L. Tochka, L. A. Hockaday, J. E. Marturano, J. J. Thibodeau and C. K. Kuo, *J. Orthop. Res.*, 2015, **33**, 874–881.
- 10 D. Gonzalez-Rodriguez, K. Guevorkian, S. Douezan and F. Brochard-Wyart, *Science*, 2012, **82**, 910–917.
- 11 M. L. Manning, R. A. Foty, M. S. Steinberg and E. M. Schoetz, *Proc. Natl. Acad. Sci. U. S. A.*, 2010, **107**, 12517–12522.
- 12 G. W. Brodland, *J. Biomech. Eng.*, 2002, **124**, 188–197.
- 13 K. Guevorkian, M.-J. Colbert, M. Durth, S. Dufour and F. Brochard-Wyart, *Phys. Rev. Lett.*, 2010, **104**, 1–4.
- 14 Y. L. Han, A. F. Pegoraro, H. Li, K. Li, Y. Yuan, G. Xu, Z. Gu, J. Sun, Y. Hao, S. K. Gupta, Y. Li, W. Tang, H. Kang, L. Teng, J. J. Fredberg and M. Guo, *Nat. Phys.*, 2020, **16**, 101–108.
- 15 R. C. Boot, G. H. Koenderink and P. E. Boukany, *Adv. Phys.: X*, 2021, **6**, 1978316.
- 16 S. Mok, S. Al Habyan, C. Ledoux, W. Lee, K. N. MacDonald, L. McCaffrey and C. Moraes, *Nat. Commun.*, 2020, **11**, 1–11.
- 17 M. E. Dolega, M. Delarue, F. Ingremeau, J. Prost, A. Delon and G. Cappello, *Nat. Commun.*, 2017, **8**, 1–9.
- 18 A. Blumlein, N. Williams and J. J. McManus, *Sci. Rep.*, 2017, **7**, 1–10.
- 19 D. Jaiswal, N. Cowley, Z. Bian, G. Zheng, K. P. Claffey and K. Hoshino, *PLoS One*, 2017, **12**, 1–21.
- 20 M. Dolega, G. Zurlo, M. L. Goff, M. Greda, C. Verdier, J. F. Joanny, G. Cappello and P. Recho, *J. Mech. Phys. Solids*, 2021, **147**, 1–21.
- 21 V. Vyas, M. Solomon, G. G. D'Souza and B. D. Huey, *Cell. Mol. Bioeng.*, 2019, **12**, 203–214.
- 22 E. M. Schötz, R. D. Burdine, F. Jülicher, M. S. Steinberg, C. P. Heisenberg and R. A. Foty, *HFSP J.*, 2008, **2**, 42–56.
- 23 C. Norotte, F. Marga, A. Neagu, I. Kosztin and G. Forgacs, *EPL*, 2008, **81**, 1–6.
- 24 A. Mgharbel, H. Delanoë-Ayari and J. P. Rieu, *HFSP J.*, 2009, **3**, 213–221.
- 25 K. Guevorkian, D. Gonzalez-rodriguez, C. Carlier, S. Dufour and F. Brochard-Wyart, *Proc. Natl. Acad. Sci. U. S. A.*, 2011, **108**, 13387–13392.
- 26 M. S. Yousafzai, V. Yadav, S. Amiri, Y. Errami, S. Amiri and M. Murrell, *Phys. Rev. Lett.*, 2022, **128**, 48103.
- 27 V. Yadav, M. S. Yousafzai, S. Amiri, R. W. Style, E. R. Dufresne and M. Murrell, *Phys. Rev. Fluids*, 2022, **7**, 1–9.
- 28 R. A. Foty, C. M. Pflieger, G. Forgacs and M. S. Steinberg, *Phys. Rev. Lett.*, 1994, **72**, 2298–2301.
- 29 P. L. Ryan, R. A. Foty, J. Kohn and M. S. Steinberg, *Proc. Natl. Acad. Sci. U. S. A.*, 2001, **98**, 4323–4327.
- 30 T. N. Truongvo, R. M. Kennedy, H. Chen, A. Chen, A. Berndt, M. Agarwal, L. Zhu, H. Nakshatri, J. Wallace, S. Na, H. Yokota and J. E. Ryu, *J. Micromech. Microeng.*, 2017, **27**, 1–9.
- 31 M. Mak and D. Erickson, *Integr. Biol.*, 2013, **5**, 1374–1384.
- 32 M. Mak, C. A. Reinhart-King and D. Erickson, *Lab Chip*, 2013, **13**, 340–348.
- 33 H. W. Hou, Q. S. Li, G. Y. Lee, A. P. Kumar, C. N. Ong and C. T. Lim, *Biomed. Microdevices*, 2009, **11**, 557–564.
- 34 P. M. Davidson, G. R. Fedorchak, S. Mondésert-Deveraux, E. S. Bell, P. Isermann, D. Aubry, R. Allena and J. Lammerding, *Lab Chip*, 2019, **19**, 3652–3663.
- 35 L. M. Lee and A. P. Liu, *Lab Chip*, 2015, **15**, 264–273.
- 36 S. H. Au, B. D. Storey, J. C. Moore, Q. Tang, Y. L. Chen, S. Javaid, A. F. Sarioglu, R. Sullivan, M. W. Madden, R. O'Keefe, D. A. Haber, S. Maheswaran, D. M. Langenau, S. L. Stott and M. Toner, *Proc. Natl. Acad. Sci. U. S. A.*, 2016, **113**, 4947–4952.
- 37 H. A. Cognart, J. L. Viovy and C. Villard, *Sci. Rep.*, 2020, **10**, 1–14.
- 38 A. Raj, M. Dixit, M. Doble and A. K. Sen, *Lab Chip*, 2017, **17**, 3704–3716.
- 39 A. Adamo, A. Sharei, L. Adamo, B. Lee, S. Mao and K. F. Jensen, *Anal. Chem.*, 2012, **84**, 6438–6443.
- 40 Y. N. Chang, Y. Liang, W. Gu, J. Wang, Y. Qin, K. Chen, J. Li, X. Bai, J. Zhang and G. Xing, *ACS Omega*, 2019, **4**, 8318–8323.
- 41 P. Sabhachandani, V. Motwani, N. Cohen, S. Sarkar, V. Torchilin and T. Konry, *Lab Chip*, 2016, **16**, 497–505.
- 42 J. Ruppen, L. Cortes-Dericks, E. Marconi, G. Karoubi, R. A. Schmid, R. Peng, T. M. Marti and O. T. Guenat, *Lab Chip*, 2014, **14**, 1198–1205.
- 43 A. Z. Shorr, U. M. Sönmez, J. S. Minden and P. R. Leduc, *Lab Chip*, 2019, **19**, 1141–1152.
- 44 W. Song, C. K. Tung, Y. C. Lu, Y. Pardo, M. Wu, M. Das, D. I. Kao, S. Chen and M. Ma, *Soft Matter*, 2016, **12**, 5739–5746.
- 45 S. W. Lee and S. S. Lee, *Microsyst. Technol.*, 2008, **14**, 205–208.
- 46 P. Constantin and C. Foias, *Navier-stokes equations*, University of Chicago Press, 2020.
- 47 R. Liu, Y. Jiang, B. Li and L. Yu, *Microfluid. Nanofluid.*, 2016, **20**, 1–13.
- 48 Y. M. Efremov, I. M. Zurina, V. S. Presniakova, N. V. Kosheleva, D. V. Butnaru, A. A. Svistunov, Y. A. Rochev and P. S. Timashev, *Biophys. Rev.*, 2021, **13**, 541–561.
- 49 R. M. Hochmuth, *J. Biomech.*, 2000, **33**, 15–22.
- 50 K. Guevorkian, F. Brochard-Wyart and D. Gonzalez-Rodriguez, *Flow dynamics of 3D multicellular systems into capillaries*, Academic Press, New York, 2021, pp. 193–223.
- 51 Y. Son, *Polymer*, 2007, **48**, 632–637.
- 52 E. V. Koudan, A. A. Gryadunova, P. A. Karalkin, J. V. Korneva, N. Y. Meteleva, I. I. Babichenko, A. V. Volkov, S. A. Rodionov, V. A. Parfenov, F. D. Pereira, Y. D. Khesuani, V. A. Mironov and E. A. Bulanova, *Biotechnol. J.*, 2020, **15**, 1–12.
- 53 A. J. Jorgenson, K. M. Choi, D. Sicard, K. M. Smith, S. E. Hiemer, X. Varelas and D. J. Tschumperlin, *Am. J. Physiol.*, 2017, **312**, C277–C285.

