



Cite this: *Lab Chip*, 2023, 23, 2039

# A low-cost, label-free microfluidic scanning flow cytometer for high-accuracy quantification of size and refractive index of particles†

Riccardo Reale, <sup>a</sup> Giovanna Peruzzi, <sup>a</sup> Maryamsadat Ghoreishi, <sup>a</sup> Helena Stabile, <sup>b</sup> Giancarlo Ruocco <sup>a</sup> and Marco Leonetti <sup>a,c</sup>

Flow cytometers and fluorescence activated cells sorters (FCM/FACS) represent the gold standard for high-throughput single-cell analysis, but their usefulness for label-free applications is limited by the unreliability of forward and side scatter measurements. Scanning flow cytometers represent an appealing alternative, as they exploit measurements of the angle-resolved scattered light to provide accurate and quantitative estimates of cellular properties, but the requirements of current setups are unsuitable for integration with other lab-on-chip technologies or for point-of-care applications. Here we present the first microfluidic scanning flow cytometer ( $\mu$ SFC), able to achieve accurate angle-resolved scattering measurements within a standard polydimethylsiloxane microfluidic chip. The system exploits a low cost linearly variable optical density (OD) filter to reduce the dynamic range of the signal and to increase its signal-to-noise ratio. We present a performance comparison between the  $\mu$ SFC and commercial machines for the label free characterization of polymeric beads with different diameters and refractive indices. In contrast to FCM and FACS, the  $\mu$ SFC yields size estimates linearly correlated with nominal particle sizes ( $R^2 = 0.99$ ) and quantitative estimates of particle refractive indices. The feasibility of using the  $\mu$ SFC for the characterization of biological samples is demonstrated by analyzing a population of monocytes identified based on the morphology of a peripheral blood mononuclear cells sample, which yields values in agreement with the literature. The proposed  $\mu$ SFC combines low setup requirements with high performance, and has great potential for integration within other lab-on-chip systems for multi-parametric cell analysis and for next-generation point-of-care diagnostic applications.

Received 23rd December 2022,  
Accepted 7th February 2023

DOI: 10.1039/d2lc01179d

rsc.li/loc

## Introduction

High-throughput biological analysis at the single-cell level is a key tool for several research fields, including immunology, molecular biology, bacteriology, cancer biology, and infectious disease monitoring, as it enables the assessment of variability and heterogeneity of cellular properties within a biological population.<sup>1</sup> In recent years, distributions and variabilities of cellular properties (e.g. size, shape, refractive index, deformability) within a population have become increasingly relevant for basic research and diagnostic applications.<sup>2–5</sup> For example, the dimensional characterization of monocytes (a type of leukocytes commonly known for their role in inflammatory processes<sup>6</sup>) has recently been proposed as a

label-free indicator for point-of-care diagnostic applications. In particular, variations in the mean or width of the size distribution of a monocyte population have been highlighted as potential diagnostic tools for early-stages of sepsis<sup>7–10</sup> and COVID-19 infections.<sup>11–15</sup>

Flow cytometers (FCM) and fluorescence activated cells sorters (FACS) currently represent the gold standard for high-throughput single cell-analysis, as they can routinely analyze several thousand cells per second.<sup>1,16</sup> These systems use excitation lasers to stimulate the cells and collection apparatuses (usually a mix of photomultiplier tubes and photodiodes) to measure scattered and fluorescent light. Measured FCM/FACS signals include the forward scattered light (FSC), i.e. the total light scattered at low angle (usually up to 20°, with variation depending on the manufacturer<sup>17,18</sup>), the side scattered light (SSC), i.e. the light scattered at ~90°, and several fluorescence channels.<sup>19</sup> While scattered light correlates with intrinsic cell properties, fluorescence signals are generated by labels previously attached to cells with an immune-specific process, requiring extensive manipulation of the sample and not suitable for

<sup>a</sup> Center for Life Nano- & Neuro-Science, Italian Institute of Technology, Rome, Italy. E-mail: riccardo.reale@iit.it

<sup>b</sup> Department of Molecular Medicine, Sapienza University of Rome, Rome, Italy

<sup>c</sup> Soft and Living Matter Laboratory, Institute of Nanotechnology, Consiglio Nazionale delle Ricerche, 00185 Rome, Italy

† Electronic supplementary information (ESI) available. See DOI: <https://doi.org/10.1039/d2lc01179d>




point-of care (POC) diagnostic applications.<sup>20–22</sup> Even though FSC is commonly used to discriminate cells based on their size,<sup>16,19,23</sup> FSC is not proportional to either particle's diameter, cross-sectional area, or volume, due to the marked dependency that the low-angle scattering has with respect to the angular integration range and to its dependency on cell refractive index, among other factors.<sup>19,24,25</sup> Moreover, FSC is highly dependent on the optical components used in the system, which may vary greatly among manufacturers. As a result, FSC measurements don't provide quantitative information on cell size and, in the best-case scenario, can be used as a relative size estimate within the same experiment.<sup>19</sup> Similarly, SSC is commonly used to acquire qualitative information on cell granularity and on its internal complexity, but it fails to provide any quantitative information on intrinsic cell properties.<sup>1,19,24</sup>

Scanning flow cytometry (SFC) is an alternative high-throughput label-free approach for quantitative single-cell analysis, based on the angle-resolved measurement of scattered light.<sup>26,27</sup> Thanks to its design, an SFC system can measure light scattered by particles as a function of the scattering angle, while they flow through the measurement region. The output of an SFC measurement is the scattering profile over an angular range, which can be used to extract quantitative information on particle intrinsic properties. SFC

is particularly appealing for diagnostic applications, as it can reliably provide quantitative estimates of cell size and refractive index, as well estimates of cell shape and bacterial morphology.<sup>28–30</sup> Nevertheless, available SFC systems require highly specialized components (*e.g.* custom flow cells, rotating detectors, motorized iris diaphragms, or numerous photomultiplier tubes) which makes them unsuitable for integration with other lab-on-chip technologies or for point-of-care (POC) applications.<sup>31–33</sup>

Here, we describe a simple, low-cost microfluidic scanning flow cytometry ( $\mu$ SFC) system for the high-throughput measurement of size and refractive index of single cells. The proposed system uses a single photoreceiver to perform an SFC measurement within a standard polydimethylsiloxane (PDMS) microchip exploiting a linearly variable optical density (OD) filter to reduce the dynamic range of the signal and to increase its signal-to-noise ratio (SNR). The variable filter consists of a glass slide coated with metal with variable density, and is commercially available or can be easily manufactured. The  $\mu$ SFC estimates the time-of-flight (ToF) of particles in the channel to compensate for flow-induced velocity variations and significantly reduce size and refractive index estimates dispersion without the need for an additional laser or detector. First, we demonstrate the superior performance of the  $\mu$ SFC with respect to label-free FCM/FACS



**Fig. 1** Microfluidic scanning flow cytometry ( $\mu$ SFC) setup. a) 3D rendering and b) schematic drawing of the system. A cylindrical lens (leftmost element in a), not shown in b)) focuses the laser beam with a 40° angle onto the microfluidic channel, which lies on a plane 400  $\mu$ m apart from the focal plane of a microscope objective. Light scattered by flowing particles is collected by the objective and focused by a lens onto a slit placed in front of a detector. The virtual image of the slit selects a different subset of scattering angles at each position along particle trajectories. Different scattering angles are mapped as different position in the back-focal plane of the objective, and reach the detector at different times. A filter with a linearly variable optical density located in the back-focal plane of the objective reduces the dynamic range and increase the signal-to-noise-ratio of the measurement. c) Diagram and picture of the linearly variable filter.



for size and refractive index measurement on polymeric beads, and successively we show the applicability of the proposed approach to the analysis of a biological target (a monocytes population). The minimal setup requirements and high performance of the  $\mu$ SFC make it well suited for compact, POC diagnostic applications, as well as readily integrated into other lab-on-a-chip systems (e.g. in microfluidic impedance cytometers<sup>34,35</sup>) for multi-parametric cellular analysis or integrated real-time applications.<sup>36</sup>

## Working principle


### Angle-resolved scattering measurements

The  $\mu$ SFC can collect the angle-resolved scattered light using a single photoreceiver and is based on the setup previously described by Loken *et al.*<sup>27</sup> A laser beam is focused by a cylindrical lens onto the microchannel where particles flow (Fig. 1). A microscope objective collects the light scattered by particles, which is focused onto the photoreceiver sensor by a lens. The microchannel lies on an out-of-focus plane located at a 400  $\mu$ m distance with respect to the objective focal plane, resulting in each scattering angle being mapped into a spatial position in the back-focal plane of the objective, and allowing to select collected scattering angles using a slit placed on the lens focal plane (which creates its virtual image on the objective focal plane). In this configuration, a different subset of scattering angles is collected at each position along particle trajectories. As a result, different scattering angles will reach the detector at different times. The angular resolution is a function of the out-of-focus distance and of the slit width. When all other setup parameters are held constant, increasing the defocus distance will increase the angular resolution and decrease the signal intensity. The 400  $\mu$ m defocus distance is the result of an experimental optimization between these two parameters. The angular

range, *i.e.* the smallest and biggest angles that can be collected, depends on the objective numerical aperture. By placing the excitation laser beam at an angle with respect to the microchannel, it is possible to shift the center of the measured angular range.

If a standard microfluidic chip is used instead of a high-spec flow cytometry flow cell,<sup>32</sup> the signal-to-noise ratio (SNR) of the measurement is significantly reduced. In particular, the lack of a reliable and easy to implement 3D flow-focusing approach limits the flow-focusing to 2D. Thus, the channel size along the third dimension (usually the channel height) must be small (e.g. <40  $\mu$ m) to minimize dispersion (e.g. to ensure similar laser excitation). As a result, interfaces are closer to particles and interfere more with the measurement. Manufacturing problems of the PDMS chip (e.g. impurities, air bubbles or curing inhomogeneity) can also lead to signal degradation and increased noise. Additionally, it's usually difficult to measure the angle-resolved scattered light over an angular range >40° with a single photoreceiver, as the signal has a dynamic range larger than the photoreceiver's.

In the  $\mu$ SFC, we included a linearly variable OD filter in the back-focal plane (see Fig. 1-b) of the objective to reduce the dynamic range and increase the SNR. The filter has a transmittance that depends logarithmically on its linear coordinate, and can filter spatially separated light with different OD. As a result, it can be used to selectively filter the high-intensity light scattered at low-angle more than the low-intensity light scattered at high-angle. A signal filtered this way has several key advantages compared to an unfiltered signal: (i) the dynamic range of the signal is significantly reduced, resulting in less stringent hardware requirements for the photodetector; (ii) particle angular oscillations frequencies in the power spectrum are easier to identify thanks to a higher separation from the lower frequency components; (iii) the event start/end timepoints



**Fig. 2** Angular resolved scattering signal. a) Representative signal measured for a 10  $\mu$ m polystyrene beads after segmentation and low-pass denoising (20 kHz). Light scattered at different angles reaches the detector at different times. A thresholding approach (3 standard deviations from the baseline, dashed line) is used to identify event start/end timepoints (red vertical lines) and time-of-flight (event end–event start). The integral of the event used for refractive index estimation is shaded in grey. b) Fourier power spectrum of the measured signal. The frequency of the angular oscillations  $f_t$  is identified using a peak-finding procedure.



can be clearly identified, enabling an accurate estimate of particles ToF.

### Particle size estimation

During analysis, the signal is segmented into individual events, each containing the signal string of one particle. As shown in Fig. 2, each event exhibits the oscillations predicted by the theoretical Mie scattering model.<sup>37</sup> The angular oscillations frequency ( $f_\theta$ ), *i.e.* the number of oscillations of the scattering profile over a defined angular window, is an intrinsic particle scattering property which yields information on particle size.<sup>38–40</sup> In the proposed configuration, we can't directly measure  $f_\theta$ , because measurements are performed in the time domain and not in the angular domain. As a result, we can only measure  $f_t$ , *i.e.* the time-domain frequency of the oscillations. While  $f_\theta$  doesn't depend on any measurement-specific parameters,  $f_t$  depends on particle velocities resulting from the parabolic flow distribution. For example, if two particles with equal angular oscillations frequency are travelling at different velocities, the faster particle will appear to have higher oscillations frequency in the time-domain due to its higher velocity. It is possible to convert the time-domain frequency ( $f_t$ ) into the angular oscillations frequency ( $f_\theta$ ) by normalizing with the event time-of-flight (ToF) according to:

$$f_\theta = \frac{f_t}{2 \sin^{-1} \frac{NA}{n}} \frac{ToF}{F_s} \quad (1)$$

where ToF is expressed in samples,  $F_s$  is the measurement sampling rate, NA is the numerical aperture of the setup, and  $n$  is the refractive index of the medium. The use of eqn (1) requires an estimate of the event ToF, which is not usually available in traditional microfluidic setups. In fact, due to the high dynamic range of the scattering signal, the high-angle scattering (which has the lowest intensity) is buried in the baseline noise, and the event start/end time-points can't be reliably estimated (*cf.* Fig. S1 of the ESI†). In the proposed  $\mu$ SFC system, thanks to the variable OD filter and the resulting lower dynamic range, it's possible to reliably estimate event start/end time-points, ToF, and  $f_\theta$ .

### Particle refractive index estimation

Even though it's possible to quantitatively estimate particle refractive index from angle-resolved scattering data by solving the inverse light-scattering problem,<sup>41</sup> it's a time- and computation-intensive procedure which is not well suited for real-time or POC applications, where computational power might be limited. We propose a faster method to extract estimates of particle refractive indices with lower computational cost. It assumes that the integral of the signal, *i.e.* total amount of light scattered in the observed angular range, depends only on particle size and refractive index mismatch between particle and carrier fluid (just like an ideal spherical particle scattering according to the Mie

model,<sup>37</sup> *cf.* Fig. S2 of the ESI†). If the refractive index of the carrier fluid and the angular range measured are known, and if the particle size is determined using  $f_\theta$ , the signal integral becomes dependent only on the particle refractive index. Nevertheless, potential ToF differences between particles induced by the fluid can artificially increase the dispersion of the measurement (*e.g.* the integral of slower particles will appear higher because they spend more time in the measurement region). It's therefore best to normalize the signal integral by particle ToF, which is equivalent to computing the mean of the signal over the observed angular range. This normalization is enabled by the high SNR of the system, which allows estimating particle ToF. Refractive indices can then be mapped in a look-up table as a function of particle diameter and signal mean over the angular range of interest.

### Particle measurement range

The proposed system has been initially developed and calibrated for the analysis of particles with size similar to monocytes (*i.e.* 9–20  $\mu$ m). This measurement range includes many other relevant biological targets, including fibroblasts, epithelial cells, hepatic cells, mesenchymal stem cells, and glioblastomas, among others. Nevertheless, the measurement range can be adjusted to analyze biological targets outside of this range. Two parameters are mostly responsible for the definition of the measurement range: i) the laser wavelength and ii) the laser spot width. i) Just like in FCM/FACS systems, the wavelength of the excitation laser is a key parameter defining the lower detection limit, and shorter wavelength lasers are required to detect smaller events reliably. By adopting a laser with a shorter wavelength (*e.g.* 488 nm) or a multi-laser configuration, the lower detection limit can be reduced. ii) The excitation beam is focused by a cylindrical lens into a strip in the center of the channel. The laser strip should be wide enough to excite the whole target cell, but at the same time, the wider the strip, the higher the background noise. Therefore, a thinner laser spot is required in order to characterize smaller particles (it can be achieved either using a cylindrical lens with shorter focal length, or with a cylindrical telescope to expand the laser beam before the final focusing lens).

## Materials and methods

### Microfluidic chip

The  $\mu$ SFC was designed to be as simple as possible, to maximize integrability with other systems. The microfluidic chip is made of PDMS, and is fabricated with standard soft-lithography procedures using an SU-8 mold, and successively bonded to a glass microscopy slide after cleaning and plasma exposure.<sup>42</sup> The microfluidic channels are formed by a sample channel and by two lateral sheath flow channels, which are designed to have the same hydraulic resistance and to focus the sample in the middle of the channel. The inlet and outlet holes lie on a different line with respect to



the measurement region to avoid interactions between the incoming laser beam and the tubing coming out of the chip. The measurement region consists of a straight channel 120  $\mu\text{m}$  wide and 25  $\mu\text{m}$  high, and particles were focused within  $\pm 15 \mu\text{m}$  from the centre of the channel (*cf.* Fig. S3 of the ESI†).

### Optical setup

A 638 nm laser with a 200 mW output is used as the light source (Oxxius), which is focused onto the measurement region by a cylindrical lens (75 mm focal length), forming a strip 60  $\mu\text{m}$  wide. The incoming beam lies on the same plane formed by the microchannel and objective, and forms a 40° angle with the microchannel. The chip is mounted on a xyz-stage with micrometric regulation, which is used to align the setup and to accurately move the channel to the desired out-of-focus distance (400  $\mu\text{m}$ ). Light scattered by particles is collected using a 50× infinity corrected long working-distance microscope objective with a 0.55 numerical aperture (Edmund Optics) resulting in a measurement over the 15–65° angular range. A rectangular, continuously variable, metallic optical density filter (0.075 OD-units per mm, Thorlabs, Fig. 1) is placed in the back-focal plane of the objective. The collected light is focused with a lens (75 mm focal length) onto a 10  $\mu\text{m}$  slit placed in front of a variable gain photoreceiver (OE-200-SI, Femto), operated at a nominal gain of  $10^8$  V/W and a bandwidth of 200 kHz.

### Sample preparation

Beads with different diameters and refractive indexes in the range of common biological applications<sup>43,44</sup> were purchased from different vendors, as described in Table 1. Refractive indices provided by vendors were independently verified using the approach previously described by Bolognesi *et al.*<sup>45</sup> Healthy donors' peripheral blood mononuclear cells (PBMCs) were isolated by Lymphoprep (Nycomed) gradient centrifugation. The sample was composed by lymphocytes, monocytes, and debris, as verified by FCM (*cf.* Fig. S4 of the ESI†). During  $\mu\text{SFC}$  analysis, the lymphocytes were below the lower detection limit and only monocytes could be observed. Each particle population is dispersed in a 1× PBS buffer ( $n = 1.335$ ) at a concentration of approximately  $10^6$  particles per ml, and the same buffer is used as sheath flow. For  $\mu\text{SFC}$  analysis, samples are loaded in glass Hamilton syringes (1 ml

for the main flow and 10 ml for sheath flows) and pumped using syringe pumps (Harvard Instruments) at a flow rate of 10  $\mu\text{l min}^{-1}$ , with a 30  $\mu\text{l min}^{-1}$  sheath flow, leading to a nominal throughput of  $\sim 150$  particles per second.

### FCM/FACS data acquisition

Flow cytometry measurements were performed using a flow cytometer BD LSRFortessa (Becton Dickinson, BD Biosciences, USA) equipped with a 488 nm Solid Sapphire 488/50 laser and a cell sorter BD FACSARIAIII (Becton Dickinson, BD Biosciences, USA) equipped with a 488 nm Solid Sapphire 488/20. Samples were pumped with a flow rate of 30  $\mu\text{l min}^{-1}$ , resulting in a nominal throughput of  $\sim 150$  particles per second. FSC and SSC measurements were acquired using FACSDiva software (BD Biosciences, version 6.1.3).

### $\mu\text{SFC}$ data acquisition and analysis


Photoreceiver signals are continuously acquired for 1 minute at 500 kHz using a PCI express data acquisition card (PCIe-6321, National Instruments) and successively analyzed offline using a custom Matlab script. Before analysis, the acquired signal is filtered with a low-pass filter with 20 kHz cut-off frequency to remove high frequency noise. Successively, the signal is segmented into signal events containing single events using a peak-finding procedure. Data are selected and gated using the bivariate distribution  $f_\theta$  vs. average intensity (*cf.* Fig. S5 of the ESI†). Coincidences, *i.e.* the simultaneous presence of 2 or more particles in the measurement region, yield signals with higher intensities which are discarded by thresholding. Coincidences were on average  $\sim 13\%$  of total events measured, but could potentially be recovered by using a more sophisticated signal analysis (*e.g.* a Bayesian approach<sup>46</sup>) to increase system throughput. The start/end points of each event are determined as the first and last time-points where the signal crosses a threshold (3 standard deviations of the baseline noise). A fast-Fourier transform (FFT) is used to compute the frequency power spectrum, and peak-finding is used to find the angular oscillations frequency peak. Single-shell theoretical Mie scattering profiles are computed using Matlab MieScat package.<sup>47</sup>

**Table 1** Properties of beads used

Material	Vendor	Mean size [ $\mu\text{m}$ ]	Size CV	Refr. index
PS	Spherotech, Inc.	8.91	<10%	1.59
PS	Polysciences, Inc.	10	<8%	1.59
PS	Polysciences, Inc.	15.7	<9%	1.59
PMMA	Microparticles, GmbH	10.16	<5%	1.48
SiO <sub>2</sub>	Microparticles, GmbH	9.98	<5%	1.42

PS: polystyrene; PMMA: polymethylmethacrylate; SiO<sub>2</sub>: silica.





**Fig. 3**  $\mu$ SFC and FCM performance comparison. Polymeric beads with different sizes and refractive indices (Table 1) are used to compare the performance of  $\mu$ SFC and FCM for size estimation. a and b) Data are shown as bivariate distributions ( $>200$  events each) of (a) angular oscillations frequency  $f_\theta$  versus normalized integral for the  $\mu$ SFC, and (b) FSC versus SSC for the FCM. Markers represent distribution means, and shaded ellipses represent areas within one standard deviation of the mean. c and d) Linear regression of (c) angular oscillations frequency  $f_\theta$  and (d) FSC with nominal beads diameter. Markers represent distribution means and error bars represent one standard deviation. Black dashed line indicate the best-fit linear regression ( $R^2 = 0.99$  for the  $\mu$ SFC and  $R^2 = 0.40$  for the FCM). The red dotted line in (c) represents the regression with the theoretical Mie model ( $R^2 = 0.88$ ). Fit coefficients are reported in Table 2.

## Results

### Performance comparison of $\mu$ SFC and FCM/FACS for size estimation


Polymeric beads with different sizes and refractive indices were used to perform a direct performance comparison between  $\mu$ SFC and FCM/FACS for particle size estimation. Angle-resolved signals representative of each particle population measured with  $\mu$ SFC after de-noising can be found in Fig. S6 of the ESI†. Samples of polystyrene (PS,  $n = 1.59$ ) beads with different diameters (9, 10, and 15  $\mu\text{m}$ ) were used to assess the effects of size, whereas beads with same diameter (10  $\mu\text{m}$ ) made of silica ( $\text{SiO}_2$ ,  $n = 1.42$ ) or polymethyl-methacrylate (PMMA,  $n = 1.48$ ) were used to assess the effects of refractive indices (see Table 1). Each sample was measured and analyzed individually using  $\mu$ SFC and FCM. Results are shown in Fig. 3-a as bivariate distributions of angular oscillations frequency  $f_\theta$  versus average signal intensity for the  $\mu$ SFC, and in Fig. 3-b as FSC versus SSC for the FCM. Both the measured FSC and  $f_\theta$  monotonously increase with increasing diameter of PS particles, (FSC equal

to  $5.1 \pm 0.1 \times 10^4$ ,  $6.9 \pm 0.2 \times 10^4$ ,  $12.6 \pm 0.3 \times 10^4$  and  $f_\theta$  equal to  $0.28 \pm 0.01$ ,  $0.31 \pm 0.01$ ,  $0.55 \pm 0.02$  for 9, 10, and 15  $\mu\text{m}$  beads, respectively) showing that both systems yield consistent results for particle with the same refractive index. Nevertheless, when comparing measurements of beads with the same size made of different materials (10  $\mu\text{m}$  PS, PMMA, and  $\text{SiO}_2$  beads), FCM measurements are heavily affected by particles refractive indices (FSC equal to  $6.9 \pm 0.2 \times 10^4$ ,  $10.7 \pm 0.4 \times 10^4$ ,  $12.2 \pm 1.9 \times 10^4$ , respectively), whereas  $\mu$ SFC aren't ( $f_\theta$  equal to  $0.31 \pm 0.01$ ,  $0.31 \pm 0.01$ ,  $0.33 \pm 0.01$ , respectively). Additionally, FSC measurement errors depend on particles refractive indices, with larger standard deviations for lower refraction indices. This leads to coefficient of

**Table 2**  $\mu$ SFC fit results

FIT TYPE	$d = f_\theta P_0 + P_1$		$R^2$
	$P_0$	$P_1$	
Best fit	25.0	1.9	0.99
Mie model	28.3	-0.034	0.88





**Fig. 4**  $\mu$ SFC measurement of monocytes. a) Representative angle-resolved measurement of the monocyte population. b) Bivariate distribution of angular frequency  $f_\theta$  versus average signal intensity. The top axis is derived from the bead best fit.

variations (CV) for FSC measurement outside manufacturer's range for  $\text{SiO}_2$  particles (CV = 16%). These results indicate that the proposed  $\mu$ SFC significantly outperforms FCM for particle size estimation, and confirm that extra care should be used when analyzing FSC data for size estimation, as both mean values and standard deviations are heavily affected by particles refractive indices. The performance difference is further demonstrated if a linear regression is used to correlate measurements with nominal diameters of the beads populations, assessing the ability of each system to provide quantitative size estimates. As shown in Fig. 3-c and d, the best fit for  $\mu$ SFC data achieves an  $R^2 = 0.99$ , versus an  $R^2 = 0.40$  for FCM data (fit coefficients can be found in Table 2 –  $\mu$ SFC fit results). It is also worth noting that while FSC depends on several vendor-dependent parameters making difficult inter-machine comparisons,  $f_\theta$  is machine invariant, and can be compared with the theoretical Mie scattering model. As shown in Fig. 3-c, the fit of experimental  $\mu$ SFC data with the theoretical model yields an  $R^2 = 0.88$ . Small discrepancies between data and model predictions may origin from errors in the event length estimation, which may result in the angular range actually measured being different from the nominal. Similar results are obtained if a FACS machine is used for the analysis (Fig. S7 of the ESI†). In particular, FACS performs better than FCM on PMMA beads, but still fails to correctly characterize  $\text{SiO}_2$  beads, resulting in an  $R^2 = 0.54$ .

### Monocytes measurement

A monocyte population within a PBMC sample was analyzed with the  $\mu$ SFC to demonstrate the feasibility of the proposed system to perform biological measurements (lymphocytes were below current lower detection limit of  $\sim 9 \mu\text{m}$ ). An angle-resolved measurement representative of signals of the monocyte population is shown in Fig. 4-a. The signal has angular oscillations similar to the ones observed for

polymeric beads which can be used to extract size estimates. The signal analysis previously described is used to obtain the bivariate distribution of angular frequency  $f_\theta$  versus average signal intensity, as shown in Fig. 4-b. In particular, the monocyte population has an  $f_\theta = 0.46 \pm 0.04$ , which correspond to a cell diameter  $d = 13.1 \pm 1.1 \mu\text{m}$  using the theoretical model, or  $d = 13.6 \pm 1.00 \mu\text{m}$  using the beads best-fit regression. Both size estimates are within the range of healthy monocytes.<sup>48,49</sup> Similarly, the monocyte population has an average signal intensity equal to  $0.24 \pm 0.03 \text{ V}$ , which using the look-up table calibrated with beads (cf. Fig. S2-b of the ESI†), yields a refractive index  $\sim 1.54$ .

## Conclusions

We presented the first microfluidic scanning flow cytometer ( $\mu$ SFC), demonstrating its potential to characterize polymeric beads and a monocyte population. The low-cost  $\mu$ SFC achieves better performance than gold-standard FCM/FACS machines for particle size estimation and label-free analysis. Thanks to its simplicity, the system can be easily integrated within other lab-on-chip system and other point-of-care diagnostic applications. Future generation microchips will be potentially fabricated in materials with higher optical quality and feasible for mass production (e.g. glass or PMMA). Future work will also include the development of a machine learning approach for automated data analysis.<sup>50,51</sup>

## Ethical approval

Healthy donors' peripheral blood mononuclear cells (PBMCs) were collected and used from healthy volunteers after informed consent, in accordance with Italian laws and Research Ethics Committee guidelines at the Policlinico Umberto I Hospital (RIF.CE: 5192).



## Author contributions

Conceptualization: R. Reale, G. Ruocco, and M. Leonetti.  
Funding acquisition: G. Ruocco, and M. Leonetti.  
Investigation: R. Reale, G. Peruzzi, and M. Ghoreishi.  
Resources: H. Stabile. Original draft: R. Reale. Review & editing: R. Reale, G. Peruzzi, M. Ghoreishi, H. Stabile, G. Ruocco, and M. Leonetti.

## Conflicts of interest

There are no conflicts to declare.

## Acknowledgements

This work was supported by the Project LOCALSCENT, Grant PROT. A0375-2020-36549, Call POR-FESR “Gruppi di Ricerca 2020”. The research leading to these results has also been supported by European Research Council Synergy Grant ASTRA (no. 855923).

## References

- 1 K. M. McKinnon, *Curr. Protoc. Immunol.*, 2018, **120**, 5.1.1.
- 2 N. Komin and A. Skupin, *Curr. Opin. Syst. Biol.*, 2017, **3**, 154–160.
- 3 R. Avraham, N. Haseley, D. Brown, C. Penaranda, H. B. Jijon, J. J. Trombetta, R. Satija, A. K. Shalek, R. J. Xavier, A. Regev and D. T. Hung, *Cell*, 2015, **162**, 1309.
- 4 S. J. Altschuler and L. F. Wu, *Cell*, 2010, **141**, 559.
- 5 Y. Jo, J. Jung, J. W. Lee, D. Shin, H. Park, K. T. Nam, J. H. Park and Y. Park, *Sci. Rep.*, 2014, **4**, 1–6.
- 6 C. Shi and E. G. Pamer, *Nat. Rev. Immunol.*, 2011, **11**, 762–774.
- 7 L. Agnello, A. Iacona, B. L. Sasso, C. Scazzone, M. Pantuso, R. V. Giglio, C. M. Gambino, A. M. Ciaccio, G. Bivona, M. Vidali and M. Ciaccio, *Clin. Chem. Lab. Med.*, 2021, **59**, 1600–1605.
- 8 E. D. Crouser, J. E. Parrillo, G. S. Martin, D. T. Huang, P. Hausfater, I. Grigorov, D. Careaga, T. Osborn, M. Hasan and L. Tejedor, *J. Intensive Care*, 2020, **8**, 1–10.
- 9 P. Hausfater, N. Robert Boter, C. Morales Indiano, M. Cancellata de Abreu, A. M. Marin, J. Pernet, D. Quesada, I. Castro, D. Careaga, M. Arock, L. Tejedor and L. Velly, *Crit. Care*, 2021, **25**, 227–239.
- 10 L. Agnello, B. L. Sasso, M. Vidali, C. Scazzone, C. M. Gambino, R. V. Giglio, A. M. Ciaccio, G. Bivona and M. Ciaccio, *Int. J. Lab. Hematol.*, 2021, **43**, O183–O185.
- 11 D. Zhang, R. Guo, L. Lei, H. Liu, Y. Wang, Y. Wang, H. Qian, T. Dai, T. Zhang, Y. Lai, J. Wang, Z. Liu, T. Chen, A. He, M. O'Dwyer and J. Hu, *J. Leukocyte Biol.*, 2021, **109**, 13.
- 12 M. Cusinato, L. Hadcocks, S. Yona, T. Planche and D. Macallan, *Int. J. Lab. Hematol.*, 2022, 1029–1039.
- 13 G. Lippi, F. Sanchis-Gomar and B. M. Henry, *Int. J. Lab. Hematol.*, 2021, **43**, O161–O163.
- 14 M. Lorubbio, D. Tacconi, G. Iannelli, M. Feri, R. Scala, S. Montemerani, M. Mandò and A. Ognibene, *Clin. Biochem.*, 2022, **103**, 29–31.
- 15 L. Alsuwaidi, S. A. Heialy, N. Shaikh, F. A. Najjar, R. Seliem, A. Han and M. Hachim, *BMC Infect. Dis.*, 2022, **22**, 1–10.
- 16 C. B. Black, T. D. Duensing, L. S. Trinkle and R. T. Dunlay, *Assay Drug Dev. Technol.*, 2011, **9**, 13.
- 17 X. Mao, A. A. Nawaz, S. C. S. Lin, M. I. Lapsley, Y. Zhao, J. P. McCoy, W. S. El-Deiry and T. J. Huang, *Biomicrofluidics*, 2012, **6**, 024113.
- 18 Optics & Detection - Flow Cytometry Guide|Bio-Rad, <https://www.bio-rad-antibodies.com/flow-cytometry-optics-detection.html>, (accessed 30 November 2022).
- 19 H. M. Shapiro, *Practical Flow Cytometry*, Wiley, 2003.
- 20 E. C. Jensen, *Anat. Rec.*, 2012, **295**, 2031–2036.
- 21 Y. Suzuki, K. Kobayashi, Y. Wakisaka, D. Deng, S. Tanaka, C. J. Huang, C. Lei, C. W. Sun, H. Liu, Y. Fujiwaki, S. Lee, A. Isozaki, Y. Kasai, T. Hayakawa, S. Sakuma, F. Arai, K. Koizumi, H. Tezuka, M. Inaba, K. Hiraki, T. Ito, M. Hase, S. Matsusaka, K. Shiba, K. Suga, M. Nishikawa, M. Jona, Y. Yatomi, Y. Yalikun, Y. Tanaka, T. Sugimura, N. Nitta, K. Goda and Y. Ozeki, *Proc. Natl. Acad. Sci. U. S. A.*, 2019, **116**, 15842–15848.
- 22 M. Herbig, K. Tessmer, M. Nötzel, A. A. Nawaz, T. Santos-Ferreira, O. Borsch, S. J. Gasparini, J. Guck and M. Ader, *Sci. Rep.*, 2022, **12**, 1–17.
- 23 A. D. Stern, A. H. Rahman and M. R. Birtwistle, *Cytometry, Part A*, 2017, **91**, 14–24.
- 24 H. M. Shapiro, *Methods Mol. Biol.*, 2018, **1678**, 1–10.
- 25 C. Becker, J. Parker, M. Hechinger and R. Leif, in *ISAC XXI Congress*, San Diego, CA, USA, 2002.
- 26 V. P. Maltsev, *Rev. Sci. Instrum.*, 1999, **71**, 243.
- 27 M. R. Loken, R. G. Sweet and L. A. Herzenberg, *J. Histochem. Cytochem.*, 1976, **24**, 284–291.
- 28 A. I. Konokhova, A. A. Gelash, M. A. Yurkin, A. V. Chernyshev and V. P. Maltsev, *Cytometry, Part A*, 2013, **83**, 568–575.
- 29 I. V. Kolesnikova, S. V. Potapov, M. A. Yurkin, A. G. Hoekstra, V. P. Maltsev and K. A. Semyanov, *J. Quant. Spectrosc. Radiat. Transfer*, 2006, **102**, 37–45.
- 30 A. L. Litvinenko, V. M. Nekrasov, D. I. Strokotov, A. E. Moskalensky, A. V. Chernyshev, A. N. Shilova, A. A. Karpenko and V. P. Maltsev, *Anal. Methods*, 2021, **13**, 3233–3241.
- 31 D. Chen, X. Liu, J. Han, M. Jiang, Z. Wang and J. Qi, *Sensors*, 2019, **19**, 2243.
- 32 E. S. Yastrebova, A. L. Litvinenko, D. I. Strokotov, R. S. Vladimirov, K. V. Gilev, V. M. Nekrasov, A. A. Karpenko and V. P. Maltsev, *J. Opt.*, 2021, **23**, 105606.
- 33 M. Geiser, S. Bolay, F. Truffer and D. Roßkamp, *Opt. Express*, 2007, **15**, 2683–2690.
- 34 T. Sun and H. Morgan, *Microfluid. Nanofluid.*, 2010, **8**, 423–443.
- 35 C. Petchakup, H. Li and H. W. Hou, *Micromachines*, 2017, **8**, 87.
- 36 J. McGrath, R. Reale, C. Honrado, P. Bisegna, N. Swami and F. Caselli, in *23rd International Conference on Miniaturized Systems for Chemistry and Life Sciences (MicroTAS 2019)*, 2019.
- 37 G. Mie, *Ann. Phys.*, 1908, **330**, 377–445.
- 38 A. V. Romanov and M. A. Yurkin, *Opt. Laser Technol.*, 2022, **151**, 108047.



- 39 Y. Luo, J. Huang and A. Adibi, *Opt. Lett.*, 2007, **32**, 1171–1173.
- 40 K. A. Semyanov, P. A. Tarasov, A. E. Zharinov, A. V. Chernyshev, A. G. Hoekstra and V. P. Maltsev, *Appl. Opt.*, 2004, **43**, 5110–5115.
- 41 V. P. Maltsev and V. N. Lopatin, *Appl. Opt.*, 1997, **36**, 6102.
- 42 J. Friend and L. Yeo, *Biomicrofluidics*, 2010, **4**, 026502.
- 43 P. Y. Liu, L. K. Chin, W. Ser, H. F. Chen, C. M. Hsieh, C. H. Lee, K. B. Sung, T. C. Ayi, P. H. Yap, B. Liedberg, K. Wang, T. Bourouina and Y. Leprince-Wang, *Lab Chip*, 2016, **16**, 634–644.
- 44 Q. Zhang, L. Zhong, P. Tang, Y. Yuan, S. Liu, J. Tian and X. Lu, *Sci. Rep.*, 2017, **7**, 1–10.
- 45 G. Bolognesi, S. Bianchi, R. di Leonardo, H. Mushfique, J. Leach, H. Yin, M. Padgett and J. Cooper, *Opt. Express*, 2011, **19**, 19245–19254.
- 46 F. Caselli, A. de Ninno, R. Reale, L. Businaro and P. Bisegna, *IEEE Trans. Biomed. Eng.*, 2021, **68**, 340–349.
- 47 MatScat-File Exchange-MATLAB Central, <https://it.mathworks.com/matlabcentral/fileexchange/36831-matscat>, (accessed 5 December 2022).
- 48 J. Prinyakupt and C. Pluempitiwiriyaewej, *J. Geophys. Res. Planets*, 2015, **14**, 63.
- 49 K. Semyanov, A. Zharinov, P. Tarasov, M. Yurkin, I. Skribunov, D. van Bockstaele and V. Maltsev, *Optics of Biological Particles*, 2007, pp. 269–280.
- 50 F. Caselli, A. de Ninno, R. Reale, L. Businaro and P. Bisegna, in *MicroTAS 2020 - 24th International Conference on Miniaturized Systems for Chemistry and Life Sciences*, 2020.
- 51 C. Honrado, J. S. McGrath, R. Reale, P. Bisegna, N. S. Swami and F. Caselli, *Anal. Bioanal. Chem.*, 2020, **412**, 3835–3845.

