Lab on a Chip



PAPER

View Article Online



Cite this: Lab Chip, 2023, 23, 2623

Interfacing centrifugal microfluidics with linearoriented 8-tube strips and multichannel pipettes for increased throughput of digital assays†

Yu-Kai Lai, (10 ab Yu-Ting Kao, (10 b Jacob Friedrich Hess, (10 ab Silvia Calabrese, (10 a Felix von Stetten pab and Nils Paust ** **ab

We present a centrifugal microfluidic cartridge for the eight-fold parallel generation of monodisperse water-in-oil droplets using standard laboratory equipment. The key element is interfacing centrifugal microfluidics with its design based on polar coordinates to the linear structures of standard highthroughput laboratory automation. Centrifugal step emulsification is used to simultaneously generate droplets from eight samples directly into standard 200 µl PCR 8-tube strips. To ensure minimal manual liquid handling, the design of the inlets allows the user to load the samples and the oil via a standard multichannel pipette. Simulation-based design of the cartridge ensures that the performance is consistent in each droplet generation unit despite the varying radial positions that originate from the interface to the linear oriented PCR 8-tube strip and from the integration of linear oriented inlet holes for the multichannel pipettes. Within 10 minutes, sample volumes of 50 µl per droplet generation unit are emulsified at a fixed rotation speed of 960 rpm into 1.47×10^5 monodisperse droplets with a mean diameter of 86 µm. The overall coefficient of variation (CV) of the droplet diameter was below 4%. Feasibility is demonstrated by an exemplary digital droplet polymerase chain reaction (ddPCR) assay which showed high linearity ($R^2 \ge 0.999$) across all of the eight tubes of the strip.

Received 18th April 2023, Accepted 27th April 2023

DOI: 10.1039/d3lc00339f

rsc.li/loc

Introduction

Digital assays allow for precise and absolute quantification of target molecules in certain dynamic ranges without the need for reference standards.1 They are widely used for the quantification of nucleic acids,2 the detection of low levels of proteins,^{3,4} measurements of enzymatic activities⁵ and for single cell phenotypic analysis.6 In order to detect single molecules, the target molecules are split into thousands or even millions of small partitions.^{7,8} Each partition carries a discrete number of target molecules and the distribution follows Poisson statistics. Absolute quantification of the target molecules is assessed by counting the ratio of partitions that contain a target molecule. From this binary "yes" or "no" answer, the term "digital" is derived.

In recent years, emulsion-based approaches, which utilize water-in-oil (W/O) droplets as partitions, have become popular due to the possibility of high-throughput partitioning and flexible handling of the droplets.9 The droplets are generated by microfluidic devices¹⁰ based on different strategies, such as T-junction emulsification, 11 flowfocusing,12 or step-emulsification.13,14 In addition to this classification by the droplet generation methods, microfluidic devices can be categorized into on-chip¹⁵⁻¹⁸ and off-chip¹⁹⁻²⁷ systems. On-chip systems automate the complete workflow within a single microfluidic device; off-chip systems allow for dispensing of generated droplets into reaction tubes. Whereas integrated on-chip systems are typically limited in the degree of parallelization and require complex and costly instrumentation,28 off-chip systems allow for processing of subsequent workflows such as droplet visualization or thermal cycling²⁹⁻³¹ with high flexibility and theoretically in a highly parallel manner. If the workflows can be performed on standard laboratory equipment, a cost-efficient automation strategy that does not require additional lab space is

To date, several off-chip systems for droplet generation in reaction tubes have been developed. 19-27 Some require syringe pumps for flow control, 19-22 which makes it difficult

^a Hahn-Schickard, Georges-Koehler-Allee 103, 79110 Freiburg, Germany. E-mail: Nils.Paust@Hahn-Schickard.de; Tel: +49 761 203 73245

^b Laboratory for MEMS Applications, IMTEK - Department of Microsystems Engineering, University of Freiburg, Georges-Koehler-Allee 103, 79110 Freiburg,

[†] Electronic supplementary information (ESI) available. See DOI: https://doi.org/

Sample loading Emulsification Amplification Readout Analysis

Fig. 1 Automated workflow for versatile digital droplet PCR assays based on the introduced multi-sample emulsification cartridge. After sample preparation, high-throughput droplet generation is carried out on a standard benchtop centrifuge and the PCR is realized on a standard thermal cycler.

to parallelize droplet generation of multiple samples. Moreover, microfluidic setups (e.g., syringe pumps and tubing) often require special knowledge for operation, thus, adopting them to different laboratories is challenging. Commercial instruments, which can be easily applied in different laboratories, often come at high costs (e.g., QX 200[™] ~80 000–100 000 USD, Bio-Rad, USA). Off-chip systems based on centrifugal microfluidics can generate droplets on standard centrifuges without the need for specialized equipment.23-27 Current solutions typically consist of a cartridge or a syringe needle²⁷ that is placed into a reaction tube for droplet generation, which allows the user to emulsify one sample per tube. Schulz et al. 23 presented a cartridge using centrifugal step-emulsification, 32 which enables the generation of monodisperse droplets in 2 ml reaction tubes on standard centrifuges. However, similar to other off-chip solutions, this solution is not intended for increased throughput sample processing. Each cartridge must be loaded manually in the tubes, which results in an increase of manual pipetting steps with the rising number of samples.

This paper describes a novel "off-chip" cartridge design strategy that seamlessly adapts linear-oriented laboratory tools to centrifugal microfluidics, enabling increased throughput of digital assays by employing centrifugal step emulsification. The microfluidic cartridge, specifically designed according to this strategy, supplies droplets to a standard 200 µl PCR 8-tube strip, not only to enhance the throughput of sample processing but also to fit to the format of standard laboratory instrumentation. Up to 50 µl aqueous sample in each droplet generation unit (DGU) can be emulsified in fluorinated oil. A multichannel pipette supplies the oil and the sample. We developed a simulation-supported fluidic cartridge design strategy that connects the centrifugal microfluidic system in polar coordinates to both the linearoriented 8-tube strip and the linear-oriented multichannel pipette while maintaining similar flow conditions in each DGU to ensure the supply of monodisperse droplets to each of the eight tubes. The cartridge design strategy is validated through experimental evaluation of droplet generation and assay performance. The cartridge was manufactured by hot embossing. It is amenable to injection molding and insertion into the centrifuge is very simple. Thus, we introduce a costefficient and convenient tool for laboratory automation of digital assays.

Results and discussion

Automated workflow

The laboratory workflow based on the new multi-sample emulsification cartridge is depicted in Fig. 1. First, the sample and oil are loaded into the droplet generation cartridge by a standard multichannel pipette. The cartridge is installed on the centrifuge for emulsification directly into the standard PCR 8-tube strip. After PCR amplification in a thermal cycler, the emulsion is ready for downstream processing, such as image-based droplet analysis as demonstrated in this work, or demulsification within a sequencing workflow. PCR amplification in droplets has been adopted for amplifying rare target in sequencing workflows, e.g., single-cell whole-genome sequencing, 30,31,33 offering advantages like reduced competition and bias. The cartridge presented in this work is compatible with aforementioned sequencing workflow, making it potentially suitable for sequencing applications as well. However, investigating this aspect is beyond the scope of this work.

Working principle

For emulsification, the following steps must be carried out: (Fig. 2a-d):

- 1. Install two sets of 8-tube strips and the cartridges on the customized rotor.
 - 2. Put on the adaptors and screw on tight.
- 3. Load oil and sample into the inlets by a multichannel pipette.
- 4. Centrifuge 4 minutes for 25 μ l or 7 minutes for 50 μ l sample at 960 rpm.
 - 5. Remove adaptors and cartridges.
- 6. Collected emulsions at the bottom of 8-tube strips are ready for further steps (*e.g.*, thermal cycling).

The dimension of the customized rotor resembled the centrifugal radius of the commercial rotor F45-48-PCR rotor (Eppendorf AG, Hamburg, Germany).³⁴ In our current design, two cartridges (16 samples) are processed simultaneously, which balances the mass. A slightly adjusted design of the holder would allow processing of three cartridges (24 samples) in parallel. An asymmetric jigsaw-fit structure would provide a quick and fail-safe way to attach the cartridge to the rotor.

Lab on a Chip Paper Adaptor




Fig. 2 The workflow for droplet generation: (a) insert two 8-tube strips on the rotor, followed by two cartridges and adaptor. Parts are mounted by M5 screw on centrifuge (b). Then load sample and oil phase via the two different rows of inlets (c and d). (e) Illustration of the cartridge under rotation during droplet generation. Blue color represents sample phase and yellow color represents oil phase. Red box is the region of interest (ROI) for detail droplet generation. (f) Zoom-in of ROI from (e), depicting droplet generation and emulsion spun-out. (g) The side view of the droplet transfer into one of the eight collection tubes

The use of multichannel pipettes reduces the pipetting steps from 16 to 2 per 8 samples. The pipetting order of sample and oil does not make any difference in our workflow. On the contrary, the previously reported cartridges²⁴⁻²⁷ usually require the user to load the oil, insert the cartridge, and load the sample one after the other. The presented system can thus save time and reduce potential manual errors. In addition, the pipetting steps can be fully automated with a robotic pipetting station thanks to the compatibility of the cartridge with standard laboratory consumables.

The microfluidic cartridge consists of a single hotembossed plastic part and a sealing film. The dimensions of the cartridge are 76 \times 74 \times 3.5 mm³ ($l \cdot w \cdot h$). As shown in Fig. 2e, the eight DGUs are implemented with a constant pitch of 9 mm, which is the pitch dimension of standard multichannel pipettes and of 8-tube strips. The diameter of the inlet holes is 3 mm, which allows proper insertion of the pipette tips into the inlets for correct loading and reduces the risk of leakage. During centrifugation at a fixed rotational speed, each unit undergoes various centrifugal acceleration (see Table 1). The detailed fluidic design of a single DGU is described in the following chapter. Each DGU consists of an array of 8 nozzles, which are oriented towards the center of

Table 1 The parameter, r_{nzl} selection and conversion between linear and polar coordinate and relative centrifugal force. The $r_{\rm nzl}$ chosen for the design was rounded to the nearest 0.5 multiples for simplicity

DGU case	$\Delta x_{\rm nzl}$ (mm)	$\Delta y_{\rm nzl}$ (mm)	$r_{\rm nzl}$ (mm)	RCF (×g)
1	70	31.5	77.0	79.5
2	70	22.5	73.5	75.9
3	70	13.5	71.5	73.8
4	70	4.5	70.0	72.8
5	70	4.5	70.0	72.8
6	70	13.5	71.5	73.8
7	70	22.5	73.5	75.9
8	70	31.5	77.0	79.5

rotation. After droplet generation, the droplets are spun out of the cartridge via the outlet (Fig. 2f). Each DGU provides an extended outlet feature for droplet transfer on the backside of the cartridge that reaches into the tubes to ensure that generated emulsions from each DGUs are completely conveyed into the corresponding tube (Fig. 2g). The distance between the outlet and the inner surface of the tube strip is designed to be ~1.5 mm. The short distance allows for a smooth transfer of the emulsions to the tube thus avoiding a coalescence of droplets as they contact the inner surface of the tube. After centrifugation, generated emulsions sit at the bottom of the tube strip, ready for the next step.

Fluidic design

Each DGU consists of the following geometric components: an inlet for oil, an inlet for the sample, an oil channel, a sample channel, an array of eight step nozzles, a chamber for droplet generation, and an outlet (see Fig. 3a). The fluidic concept of a single DGU was described previously.²³ Since the single DGU from the previous work had shown excellent droplet monodispersity, we aimed to apply the same flow conditions on our system. Here, the main challenge was to provide the same flow conditions to each DGU. Since forces on a rotating system are oriented radially outwards, the design in centrifugal microfluidics is based on polar coordinates. To fit the linear structures to the design, different radial positions of the inlets and outlets and thus varying forces acting on these features cannot be avoided. To provide similar flow conditions to each DGU and thus monodisperse droplets, a simulation-supported design strategy has been applied that consists of the following steps. First, the array of eight nozzles and chambers are positioned towards the center of rotation to ensure that the direction of the centrifugal force is the same in each DGU. As illustrated in Fig. 2e, sample and oil inlets are placed at the same distance of x-direction for fitting the dimension of the multichannel pipette. The same applies for the outlets, to provide the interface to the 8-tube strip. Sample channel and oil channels are then derived by simulation to ensure that the desired flowrates from previous work²³ is realized in each DGU.

Paper Lab on a Chip




Fig. 3 (a) Fluidic design using DGU1 as an example. The origin of polar coordinates is set at the center of rotation. The values of Δx , Δy and r_{nzl} are listed in Table 1. (b) Model block representation of the fluidic network of single DGU in (a). The model consists of volume chamber, channel junction, and isoradial and radial channels. Ground symbol represents a connection to the atmospheric pressure. Numbers indicate the equivalent fluidic component shown in (a). The color boxes represent the filled fluid in the channel. Blue: dispersed phase. Yellow: oil phase. Green: emulsion phase. (c) Measured mean sample flowrate and validation of simulation model from individual DGUs. The dash line marks the mean flowrate in a case of ideal design. Error bars denote standard deviation from three performed cartridges.

Network simulation model

A MATLAB based fluidic network simulation with lumped elements was set up to predict flowrates and optimize channel lengths (Fig. 3b). In a network simulation describing centrifugal microfluidic structures, the fluidic components (see Fig. 3a), can be translated to basic elements such as radially or iso-radially oriented channels, volume chambers or T-junctions as illustrated in Fig. 3b and detailed by Schwarz et al.35 Each fluidic component is represented by a model block with corresponding physical functions. The sample and oil inlets are represented by volume chambers with fill level function $r(V_{spl})$ and $r(V_{oil})$, which are radial position functions in relation to the current fluid volumes inside the chambers. Channels with arbitrary orientation can be simplified to radial and isoradial channels.35 Each element consists of a physical function that derives the total pressure difference across the element and the flowrate Q through the element. At constant rotational speed, the predominant pressure difference is the sum of centrifugal force (Δp_c) and viscous dissipation (Δp_v) . The equations are given as follows:

$$\Delta p_{\rm c} = \frac{\rho}{2} \times \omega^2 \times \left(r_{\rm o}^2 - r_{\rm i}^2\right) \tag{1}$$

$$\Delta p_{\rm c} = -RQ \tag{2}$$

where ρ is the liquid density, ω the angular rotational speed, $r_{\rm i}$ and $r_{\rm o}$ the inner and outer radial positions of the liquid column in the corresponding element. R is the hydraulic resistance. For the channel with rectangular cross section, the R is given as:

$$R = \frac{12\eta L}{d^3 w \left(1 - \frac{192d}{\pi^5 w} \times \left(\tanh \frac{\pi w}{2d} + \frac{31}{32} \zeta(5) - 1\right)\right)}$$
(3)

where η is the dynamic viscosity, L the length of the filled liquid in the given element, and d and w are the depth and width of the element (d < w), respectively, and $\zeta(5)$ a constant. All abbreviations and subscripts are listed in the ESI,† Table S1. The radial position required in eqn (1) for each simply fluidic element is derived from $r = \sqrt{\Delta x^2 + \Delta y^2}$, where Δx and Δy are the x-distance and y-distance between the point of the interest and the center of rotation (see Fig. 3a for an example). The hydraulic resistances of the sample and oil channels, $R_{\rm spl}$ and $R_{\rm oil}$, are adjusted for each individual DGU to match the mean flowrates of all DGUs. To simplify the optimization and manufacturing of the cartridges, we fixed the cross-section dimension of the sample channel, oil channel and outlet channel among all DGUs, and only altered channel lengths (dimensions are listed in ESI,† Table S2). The fluidic properties, density and viscosity of the three phases, sample, oil and mixture phases (highlighted by color boxes in Fig. 3b), must be known to run the simulation. While the fluidic properties of the sample and the oil phase are known, the density and viscosity of the generated emulsion are not evident. Due to the conservation of mass and the assumption of incompressible fluids, the density of the emulsion can simply be calculated using the volume fraction:

$$\rho_{\rm eml} = \varphi \rho_{\rm spl} + (1 - \varphi) \rho_{\rm oil} \tag{4}$$

Lab on a Chip Paper

where φ is the mean sample volume fraction in the emulsion flow, denoted by $\varphi = \bar{Q}_{\rm spl}/(\bar{Q}_{\rm spl} + \bar{Q}_{\rm oil})$. $\bar{Q}_{\rm spl}$ and $\bar{Q}_{\rm oil}$ are the mean sample and mean oil flowrates, respectively. The effective viscosity for calculating the flow resistance of emulsions in microchannels, however, cannot be calculated analytically as, in addition to the volume fractions, the viscosity depends on complex flow patterns influenced by interfacial forces and by the size of the droplets. It has been observed in previous studies that the presence of droplets in a channel contributes to an additional pressure drop,³⁶ resulting in an effective viscosity higher than an average viscosity derived from volume fractions. Effective viscosities have extensively been studied37-39 and implemented in network simulations⁴⁰⁻⁴² previously. However, the derived values are not suitable for the scenario in our system in which droplets are much more compact and squeezed in the outlet channel (see ESI,† Fig. S1). To account for the additional pressure drop due to the emulsion's transport through the outlet channel, we estimated the effective viscosity by fitting parameters using experimental data from the previous study²² on which the current design is based (see ESI,† S2). The determination procedure of effective viscosity is described in detail in ESI,† S2. As a result, an effective viscosity of $\eta_{\rm eml}$ = 10.5 mPa s was used in this work. The mean flowrates of $\bar{Q}_{\rm spl}$ = 0.13 $\mu l \ s^{-1}$ and $\bar{Q}_{\rm oil}$ = 0.09 $\mu l \ s^{-1}$ for sample and oil flow, respectively, were chosen as design values for all DGUs with $\varphi = 0.591$ and $\rho_{\rm eml} = 1251.1$ kg m⁻³ (eqn (4)).

The result of the simulation study is a parameter set for the sample and oil channel length for each DGU as depicted in Table 2, while the cross-section areas of the sample and the oil channel are $62 \times 58 \mu m^2$ and $49 \times 37 \mu m^2$, respectively.

Validation of the simulation model and the manufactured cartridge

The manufactured cartridge turned out to have some deviations from the ideal design as listed in ESI,† S4. For the comparison of the simulation results with the experiments, the dimensions of the manufactured cartridge were measured and applied to the simulation.

Then, we experimentally evaluated the mean sample flowrate $Q_{\rm spl}$ and compared it with the results of the proposed simulation model to validate our design strategy.

Table 2 The sample and oil channel length

DGU case	$L_{ m spl}$ (mm)	$L_{ m oil}$ (mm)
1	15.28	16.14
2	15.87	15.75
3	16.11	14.75
4	16.28	14.91
5	16.40	14.39
6	16.72	13.96
7	16.84	13.34
8	17.03	13.19

50 µl of PCR buffer (1×) and 50 µl of oil were supplied to each DGU in the cartridge and then emulsified. The emulsification time from each DGU was determined from video sequences. We obtained $\bar{Q}_{\rm spl}$ by dividing the sample volume of 50 µl in each DGU by the measured emulsification time. The depth and width of the sample, oil, nozzles and outlet channels were measured by optical profilometry (the values are listed in ESI,† S4). The measured dimension parameters are used in the simulation model. In this case, the flowrate conditions are unknown and thus $\rho_{\rm eml}$ is unknown (eqn (4)). To obtain the correct results in which the solved mean flowrates $\bar{Q}_{\rm spl}$ and $\bar{Q}_{\rm oil}$, and $\rho_{\rm eml}$ fit eqn (4), one can use an iterative method: start a simulation with an arbitrary φ , e.g., 0.5, as an initial guess and then utilize the solved $\bar{Q}_{\rm spl}$ and $\bar{Q}_{\rm oil}$ as an input for the next simulation run. After two runs, the predetermined emulsion density matched the outcome of eqn (4) with a difference less than 0.001 kg m⁻³. After the integration of all manufacturing variances to the simulation, the model using manufactured dimensions (red circle, Fig. 3c) showed a good agreement with the experimental results (black square, Fig. 3c). The maximum variance within the experimental $\bar{Q}_{\rm spl}$ and compensated $\bar{Q}_{\rm spl}$ is ~2.2% in the case of DGU3. The accordance of the simulated and experimental results suggests that the variance of measured Q_{spl} is mainly attributed to tolerances of the manufacturing technology. In addition, the experimental result also supports our simulation-based design strategy, which integrates linearly arranged fluidic elements on the centrifugal microfluidic platform.

Droplet characterization

We investigated the performance of the cartridge by examining the diameter of the generated droplets. The PCR buffer (1x) was used as sample phase for droplet characterization. 50 µl of PCR buffer and oil per DGU were pipetted into the cartridge and then emulsified. After 7 minutes, all reaction mixes were completely emulsified, and no residual liquid was visible, neither in the sample inlets or channels, nor on the inner surface of the tube strip. As shown in Fig. 4a, the mean of the droplet diameter from each DGU has a slight variance. The largest mean droplet diameter (88.9 µm) from DGU8 is about 1.07-fold the smallest mean droplet diameter (82.7 µm) from DGU1. This variance originates from tolerances of the applied manufacturing technology. The cartridge has a relatively large area (76 × 74 mm²) compared, for example, to the previously presented microfluidic cartridges from Schulz et al. 23 (33.6 × 7.8 mm²). Higher milling tolerance is usually expected as the work piece is larger.

We investigated how manufacturing variance affected droplet diameter variation. First, the mean nozzle depth from each DGU was obtained by optical profilometry (the values are listed in the ESI,† S4). Dividing the mean droplet diameter by the mean nozzle depth of the corresponding

а Dronlet diameter Normalized diameter Droplet diameter (μm) 4.5 90 80 • 70 3.0 60 50 b Relative Frequency (%) 25 -20 $86.5 \pm 3.21 \mu m$ 15 -C.V: 3.7% 10 -5 75 80 70 90 100 105 110 85

Fig. 4 (a) Measured droplet diameter from individual DGUs. Red data show the normalized diameter derived by dividing the droplet diameter by the mean nozzle depth of the eight nozzles in each DGU. 1000 droplets were evaluated from each DGU. (b) Overall droplet size distribution, calculated from all eight DGUs (N = 8000). The red curve represents normal Gaussian distribution.

Diameter (µm)

DGU (red dots in Fig. 4a), the variance of the mean droplet diameter decreased to 1.03-fold between the largest value and the smallest value (DGU2 and DGU8). The overall nozzle dimension has a mean depth of 25.6 \pm 0.81 μm with CV < 3.5% (64 nozzles among 8 DGUs). Despite the slight differences of droplet diameters among DGUs, the overall droplet diameter distribution (Fig. 4b) revealed good droplet monodispersity (CV < 5%). Hence, we can consider that the droplet diameters from different DGUs are identical and can be used for digital droplet PCR ($d=86.5~\mu m$) with estimated 1.47 \times 10 5 generated droplets in each DGU.

ddPCR assay

To demonstrate compatibility of the cartridge biochemical assays, exemplary ddPCRs for the detection of the human cystic fibrosis transmembrane conductance regulator (CFTR) gene were performed in serial dilutions. Triplicates were run in each DGU for each dilution step and for every run, a fresh cartridge was used. Cartridges were loaded and processed as explained in the workflow description. The ddPCR was evaluated using an automated fluorescence microscope and image data analysis software. The concentration was calculated by applying Poisson statistics.⁴³ Evaluation of the ddPCR showed homogeneous, monodisperse droplet images with a clear differentiation between positive and negative droplets (see Fig. 5a). By plotting expected values against measured values, a very good concordance of template concentrations could be




Fig. 5 Analysis of the digital droplet polymerase chain reaction (ddPCR). (a) Exemplary microscopic fluorescence images of droplet monolayers after amplification. Serial dilution of target template (C1–C3) and a non-template control (NTC) are shown (using the presented cartridge). (b) Measured template concentrations per DGU. The dashed line represents the expected concentration. Error bars represent standard deviations of three repetitions. (c) Comparison of presented cartridge and commercial instrument is performed by plotting overall concentration against the expected concentration. Each data point represents the mean of concentration of three repetitions of the eight DGUs for the presented cartridge and four repetitions for the Stilla Naica system. The linear fit curve is represented by the solid line. RAW data are presented in the ESI;† S6.

observed for each serial dilution per DGU with R^2 values between $R^2 \ge 0.998$ and $R^2 \ge 0.999$ (Fig. 5b, ESI,† S6) as well as for the overall experimental performance (red line, Fig. 5c) with $R^2 \ge 0.999$.

Additional experiments were performed separately to verify quantitative accuracy against commercial ddPCR platform (Naica system, Stilla Technologies, Villejuif, France). These experiments showed comparable results in terms of linearity, with an R^2 value of ≥ 0.997 (Fig. 5c).

The good concordance between template concentration in the single DGU as well as for the overall set-up demonstrates that the cartridge performance did not interfere with introduced biomolecules nor impair assay performance. Therefore, it is suitable for increased throughput biochemical assays.

The presented cartridge is an open system without any lid during sample emulsification. Usually low DNA-concentrations are applied before amplification, and the risk of cross contamination can be considered low, similar to other common methods of DNA extraction such as the use of spin-columns. The accordance of the ddPCR result also supports our assumption.

Materials and methods

Microfluidic cartridge design and manufacturing

The microfluidic structure was designed in a computer-aideddesign software (Dassault Systèmes SE, SolidWorks, Vélizy-Villacoublay, France) and MATLAB based fluidic networking simulation (MathWorks Corp., Matlab, Natick, MA, USA) and manufactured by the Hahn-Schickard Lab-on-a-Chip Foundry.44 The PMMA (Evonik AG, Essen, Germany) master was milled by a CNC milling machine (KERN Evo mill, KERN Microtechnik GmbH, Eschenlohe, Germany). The milled surface was measured by confocal microscopy (DUO Vario, Confovis GmbH, Jena, Germany). Then, the positive stamp was produced. Polydimethylsiloxane (PDMS) (mixture of elastosil 607 and elastosil 675, 1:1, with monomer/ crosslinker ratios of 9:1 and 1:1 respectively) was poured onto the milled master, and centrifuged (Zentrifuge Rotanta 460 R, Hettich GmbH, Kirchlengern, Germany). The positive PDMS stamp was used to create replicates by hot embossing, transferring structures onto cyclic olefin copolymer substrate (COC) (TOPAS COC 5013, TOPAS Advanced Polymers GmbH, Raunheim, Germany). Finally, the cartridge was sealed manually using a pressure sensitive adhesive film (9795R diagnostic tape, 3M Corp., Saint Paul, MN, USA).

Centrifuge and customized rotor

The rotor was designed in Solidworks and printed by a commercially available 3D printer (Original Prusa i3, Prusa Czech Republic) using polyethylene terephthalate glycol filament (PETG, Prusa Research sro, Czech Republic).

To perform and observe droplet generation, a customized centrifuge was used (Euler player 2nd generation, BioFluidix GmbH, Freiburg im Breisgau, Germany).

Droplet generation and readout

For all experiments, fluorinated oil with surfactant dSurf (4% (w/w) in Novec[™] 7500, Fluigent, France) was used as a continuous phase. The PCR buffer (1x) was used as a dispersed phase in this paper with a measured viscosity of 1.65 mPa s (see ESI,† S1) and was prepared by mixing 2× ddPCR Supermix for probes (Bio-Rad Laboratories Inc., Hercules, CA, USA) with distilled water (UltraPure DNase/ RNase free, Invitrogen Corp., Carlsbad, CA, USA). Pipetting was performed using an Eppendorf Research Plus 8-channel pipette (Eppendorf AG, Hamburg, Germany) and epT.I.P.S. Standard 2-200 µL pipette tips (Eppendorf AG, Hamburg, Germany). No leakage was observed through all experiments. Droplet diameters were evaluated by the following procedure: bright field images were recorded by microscope Observer Z1 (Zeiss GmbH, Jena, Germany) and then further analyzed using MATLAB based build-in circle detection algorithm (see ESI,† S5).

ddPCR

The ddPCR experiment on the presented cartridge was set-up according to Schulz et al. 23 25 µl reaction mixes were pipetted into each DGU of the cartridge with a multichannel pipette (Eppendorf AG, Hamburg, Germany). After droplet generation, the emulsion in the 8-tube strip (Biozym Scientific GmbH, Hessisch Oldendorf, Germany) was covered with 15 µl mineral oil (Sigma Aldrich, St. Louis, MO, USA) and transferred to a thermal cycler (T 100 Thermal Cycler, Bio-Rad Laboratories Inc., Hercules, CA, USA). The cycling protocol was as follows: 5 min at 94 °C, 45 cycles (15 s at 94 °C and 60 s at 58 °C), and holding at 23 °C until readout. After cycling, droplets were transferred into cell counting chamber slides (Countess™, Invitrogen Corp., Carlsbad, CA, USA), where the droplets aligned in a monolayer. Positive and negative droplets ($N \approx 2000$ per sample) were counted using a fluorescence microscope (Lionheart LX, BioTek Instruments GmbH, Bad Friedrichshall, Germany) and the corresponding image analysis software package (Gen5).

A comparative experiment on a commercial ddPCR platform was performed using the Stilla Naica system (Stilla Technologies, Villejuif, France). The same assay was used with a minor adjustment: 5× PerfeCTa qPCR ToughMix (QuantaBio, Beverly, MA, USA) was used in the reaction mix instead of 2× ddPCR Supermix for probes. 25 µl reaction mixes were pipetted into sample inlets on Sapphire Chips (Stilla Technologies, Villejuif, France). The same thermal cycling protocol was applied after an additional step of partitioning droplets (40 °C for 12 min). 23 000 to 27 000 droplets were evaluated after thermal cycling using the corresponding readout instrument and software (Prism 6, Crystal Miner).

Conclusions

We present a microfluidic cartridge comprising eight "droplet generation units", DGUs, for increased throughput sample emulsification as a part of a parallelized and automated laboratory workflow. The automated workflow is suitable for digital droplet assay and can be adapted in any laboratory as it requires no specialized instruments. The user-friendly installation simplifies cartridge set up with minimal manual steps within a short period of time. Compared to existing solutions, 23-27 the use of eight channel pipettes reduces pipetting steps for loading sample and oil from 16 to 2. Using our simulation-based strategy, we derived an in silico design that provides consistent flow conditions for each DGU despite the challenges of interfacing the linearly oriented eight channel pipette and the eight tube strips to the centrifugal microfluidic chip designed in polar coordinates. We demonstrated that the cartridge enables the user to emulsify eight samples for ddPCR with good reproducibility, regardless of the DGU position. The presented cartridge design is tailored to the Eppendorf rotor for 6× PCR 8-tube strips. By applying the introduced design strategy, however, adaptation of the cartridge design to rotors with different

dimensions is straightforward. One critical parameter is the distance between the rotor's center and the installed PCR 8-tube strip, as this affects the orientation and radial position of each DGU on the cartridge (parameters shown in Fig. 3(a)). The second parameter is the rotor radius required to hold the cartridge. In the presented cartridge, the length (x-direction) of the DGU is 40 mm, and additional length is

necessary to achieve sufficiently large hydrostatic pressures at the start of centrifugation. Consequently, the minimum rotor radius should be no less than 50 mm. This also determines the centrifuge models compatible with the presented cartridge.

The current 3D printed rotor allows for the operation of 16 samples per run by processing two cartridges simultaneously. By adapting the rotor design, one can easily integrate at least three cartridges. It is possible that higher throughput can be achieved by using a customized rotor for 4×8 DGUs or by using a 96-spin plate rotor with stacked DGU units on top of a 96-well PCR plate (12×8 DGUs, not shown in this study). Furthermore, the current design required adaptors to fix the cartridges to the rotor and therefore, a customized centrifuge was used. Once the cartridge is transferred to injection molding, a cartridge can be designed to fit directly to the commercially available rotor and thereby become a powerful tool for cost-effective digital assay implementations.

Finally, we demonstrate the interfacing of centrifugal microfluidics with linear-oriented standard equipment. The design strategy is not limited to droplet generation. Arranging fluidic elements in a linear-oriented manner enables compact fluidic design and reduces the footprint, resulting in increased throughput. This approach could further extend parallelization for various applications in diagnostics and bioanalytical processes, including lateral flow strips,45 immunoassays46 and nucleic acid assays.47 This streamlining could offer a way to make the most of the major advantages of centrifugal microfluidics when handling small volumes in a highly precise manner. 48,49 Efficiency is further enhanced by the excellent technical benefits of linearoriented laboratory equipment for high-throughput laboratory automation.

Conflicts of interest

There are no conflicts to declare.

Acknowledgements

The authors thank the team at Hahn-Schickard Lab-on-a-Chip Foundry for manufacture of cartridges, engineer Moritz Boesenberg and Daniel Frejek for 3D printing, and Dr. Martin Schulz for discussion at an early stage of development. The authors also thank Tobias Lange and Tobias Gross for providing samples. The scientific work was funded by Federal Ministry of Education and Research (BMBF) within the projects FREEDOM (project number 13N15288).

References

- 1 A. S. Basu, Digital Assays Part I: Partitioning Statistics and Digital PCR, SAGE Publications Inc, *SLAS Technol.*, 2017, **22**, 369–386.
- 2 E. A. Ottesen, J. W. Hong, S. R. Quake and J. R. Leadbetter, Microfluidic Digital PCR Enables Multigene Analysis of Individual Environmental Bacteria, *Science*, 2006, 314, 1464–1467.
- 3 L. Cohen, N. Cui, Y. Cai, P. M. Garden, X. Li, D. A. Weitz and D. R. Walt, Single Molecule Protein Detection with Attomolar Sensitivity Using Droplet Digital Enzyme-Linked Immunosorbent Assay, ACS Nano, 2020, 14, 9491–9501.
- 4 J. U. Shim, R. T. Ranasinghe, C. A. Smith, S. M. Ibrahim, F. Hollfelder, W. T. S. Huck, D. Klenerman and C. Abell, Ultrarapid generation of femtoliter microfluidic droplets for single-molecule-counting immunoassays, ACS Nano, 2013, 7, 5955–5964.
- 5 Y. Obayashi, R. Iino and H. Noji, A single-molecule digital enzyme assay using alkaline phosphatase with a cumarin-based fluorogenic substrate, *Analyst*, 2015, **140**, 5065–5073.
- 6 Y. Zeng, R. Novak, J. Shuga, M. T. Smith and R. A. Mathies, High-performance single cell genetic analysis using microfluidic emulsion generator arrays, *Anal. Chem.*, 2010, 82, 3183–3190.
- 7 D. Witters, B. Sun, S. Begolo, J. Rodriguez-Manzano, W. Robles and R. F. Ismagilov, Digital biology and chemistry, *Lab Chip*, 2014, 14, 3225–3232.
- 8 B. Vogelstein and K. W. Kinzler, Digital PCR, Proc. Natl. Acad. Sci. U. S. A., 1999, 96, 9236–9241.
- 9 M. T. Guo, A. Rotem, J. A. Heyman and D. A. Weitz, Droplet microfluidics for high-throughput biological assays, *Lab Chip*, 2012, 12, 2146–2155.
- 10 K. S. Elvira, F. Gielen, S. S. H. Tsai and A. M. Nightingale, Materials and methods for droplet microfluidic device fabrication, *Lab Chip*, 2022, 22, 859–875.
- B. J. Hindson, K. D. Ness, D. A. Masquelier, P. Belgrader, N. J. Heredia, A. J. Makarewicz, I. J. Bright, M. Y. Lucero, A. L. Hiddessen, T. C. Legler, T. K. Kitano, M. R. Hodel, J. F. Petersen, P. W. Wyatt, E. R. Steenblock, P. H. Shah, L. J. Bousse, C. B. Troup, J. C. Mellen, D. K. Wittmann, N. G. Erndt, T. H. Cauley, R. T. Koehler, A. P. So, S. Dube, K. A. Rose, L. Montesclaros, S. Wang, D. P. Stumbo, S. P. Hodges, S. Romine, F. P. Milanovich, H. E. White, J. F. Regan, G. A. Karlin-Neumann, C. M. Hindson, S. Saxonov and B. W. Colston, High-Throughput Droplet Digital PCR System for Absolute Quantitation of DNA Copy Number, Anal. Chem., 2011, 83(22), 8604–8610.
- 12 L. Yobas, S. Martens, W.-L. Ong and N. Ranganathan, High-performance flow-focusing geometry for spontaneous generation of monodispersed droplets, *Lab Chip*, 2006, 6, 1073–1079.
- 13 F. Schuler, M. Trotter, M. Geltman, F. Schwemmer, S. Wadle, E. Domínguez-Garrido, M. López, C. Cervera-Acedo, P. Santibáñez, F. von Stetten, R. Zengerle and N. Paust, Digital droplet PCR on disk, *Lab Chip*, 2016, 16, 208–216.

14 M. Schulz, F. von Stetten, R. Zengerle and N. Paust, Centrifugal step emulsification: How buoyancy enables high generation rates of monodisperse droplets, *Langmuir*, 2019, 35, 9809–9815.

Lab on a Chip

- 15 W. Lyu, J. Zhang, Y. Yu, L. Xu and F. Shen, Slip formation of a high-density droplet array for nucleic acid quantification by digital LAMP with a random-access system, *Lab Chip*, 2021, 21, 3086–3093.
- 16 J. J. Agresti, E. Antipov, A. R. Abate, K. Ahn, A. C. Rowat, J.-C. Baret, M. Marquez, A. M. Klibanov, A. D. Griffiths and D. A. Weitz, Ultrahigh-throughput screening in drop-based microfluidics for directed evolution, *Proc. Natl. Acad. Sci. U. S. A.*, 2010, 107, 4004–4009.
- 17 V. Yelleswarapu, J. R. Buser, M. Haber, J. Baron, E. Inapuri and D. Issadore, Mobile platform for rapid sub-picogram-per-milliliter, multiplexed, digital droplet detection of proteins, *Proc. Natl. Acad. Sci. U. S. A.*, 2019, **116**, 4489–4495.
- 18 M. Nie, M. Zheng, C. Li, F. Shen, M. Liu, H. Luo, X. Song, Y. Lan, J.-Z. Pan and W. Du, Assembled Step Emulsification Device for Multiplex Droplet Digital Polymerase Chain Reaction, Anal. Chem., 2019, 91, 1779–1784.
- 19 B. Shi, W. Di, Y. Jiang, J. An and W. Wenming, Off-Chip Vertical Step Emulsification Droplets Preparation Device Applied for Droplet Digital PCR, Adv. Mater. Interfaces, 2020, 7, 2001074.
- 20 S. Zhao, Z. Zhang, F. Hu, J. Wu and N. Peng, Massive droplet generation for digital PCR via a smart step emulsification chip integrated in a reaction tube, *Analyst*, 2021, 146, 1559–1568.
- 21 Y. Li, H. Cherukury, L. Labanieh, W. Zhao and D.-K. Kang, Rapid Detection of β-Lactamase-Producing Bacteria Using the Integrated Comprehensive Droplet Digital Detection (IC 3D) System, *Sensors*, 2020, **20**, 4667.
- 22 T. J. Abram, H. Cherukury, C.-Y. Ou, T. Vu, M. Toledano, Y. Li, J. T. Grunwald, M. N. Toosky, D. F. Tifrea, A. Slepenkin, J. Chong, L. Kong, D. V. Del Pozo, K. T. La, L. Labanieh, J. Zimak, B. Shen, S. S. Huang, E. Gratton, E. M. Peterson and W. Zhao, Rapid bacterial detection and antibiotic susceptibility testing in whole blood using one-step, high throughput blood digital PCR, Lab Chip, 2020, 20, 477–489.
- 23 M. Schulz, S. Probst, S. Calabrese, A. R. Homann, N. Borst, M. Weiss, F. von Stetten, R. Zengerle and N. Paust, Versatile Tool for Droplet Generation in Standard Reaction Tubes by Centrifugal Step Emulsification, *Molecules*, 2020, 25, 1914.
- 24 Z. Chen, P. Liao, F. Zhang, M. Jiang, Y. Zhu and Y. Huang, Centrifugal micro-channel array droplet generation for highly parallel digital PCR, *Lab Chip*, 2017, 17, 235–240.
- 25 H. Yamashita, M. Morita, H. Sugiura, K. Fujiwara, H. Onoe and M. Takinoue, Generation of monodisperse cell-sized microdroplets using a centrifuge-based axisymmetric coflowing microfluidic device, *J. Biosci. Bioeng.*, 2015, 119, 492–495.
- 26 D. C. Shin, Y. Morimoto, J. Sawayama, S. Miura and S. Takeuchi, Centrifuge-based step emulsification device for simple and fast generation of monodisperse picoliter droplets, *Sens. Actuators, B*, 2019, 301, 127164.

- 27 W. Zhou, C. Liu, T. Zhang, K. Jiang, H. Li, Z. Zhang and Y. Tang, Low Cost, Easily-Assembled Centrifugal Buoyancy-Based Emulsification and Digital PCR, *Micromachines*, 2022, 13, 171.
- 28 J. Madic, A. Zocevic, V. Senlis, E. Fradet, B. Andre, S. Muller, R. Dangla and M. E. Droniou, Three-color crystal digital PCR, *Biomol. Detect. Quantif.*, 2016, 10, 34–46.
- 29 D. Pekin, Y. Skhiri, J.-C. Baret, D. Le Corre, L. Mazutis, C. B. Salem, F. Millot, A. El Harrak, J. Brian Hutchison, J. W. Larson, D. R. Link, P. Laurent-Puig, A. D. Griffiths and V. Taly, Quantitative and sensitive detection of rare mutations using droplet-based microfluidics, *Lab Chip*, 2011, 11, 2156–2166.
- 30 Y. Fu, F. Zhang, X. Zhang, J. Yin, M. Du, M. Jiang, L. Liu, J. Li, Y. Huang and J. Wang, High-throughput single-cell whole-genome amplification through centrifugal emulsification and eMDA, *Commun. Biol.*, 2019, 2, 1–10.
- 31 Y. Fu, C. Li, S. Lu, W. Zhou, F. Tang, X. Sunney Xie and Y. Huang, Uniform and accurate single-cell sequencing based on emulsion whole-genome amplification, *Proc. Natl. Acad. Sci. U. S. A.*, 2015, **112**, 11923–11928.
- 32 F. Schuler, F. Schwemmer, M. Trotter, S. Wadle, R. Zengerle, F. von Stetten and N. Paust, Centrifugal step emulsification applied for absolute quantification of nucleic acids by digital droplet RPA, *Lab Chip*, 2015, 15, 2759–2766.
- 33 X. Li, D. Zhang, W. Ruan, W. Liu, K. Yin, T. Tian, Y. Bi, Q. Ruan, Y. Zhao, Z. Zhu and C. Yang, Centrifugal-Driven Droplet Generation Method with Minimal Waste for Single-Cell Whole Genome Amplification, *Anal. Chem.*, 2019, 91, 13611–13619.
- 34 Eppendorf AG, Original instructions. Centrifuge 5804/5804 R Centrifuge 5810/5810 R, Rotor F-45-48-PCR 11.1.9, 2019, https://www.eppendorf.com/product-media/doc/en/330816/ Centrifugation_Operating-manual_Centrifuge-5804-5804-R_ Centrifuge-5810-5810-R.pdf.
- 35 I. Schwarz, S. Zehnle, T. Hutzenlaub, R. Zengerle and N. Paust, System-level network simulation for robust centrifugal-microfluidic lab-on-a-chip systems, *Lab Chip*, 2016, **16**, 1873–1885.
- 36 C. N. Baroud, F. Gallaire and R. Dangla, Dynamics of microfluidic droplets, *Lab Chip*, 2010, **10**, 2032.
- 37 B. J. Adzima and S. S. Velankar, Pressure drops for droplet flows in microfluidic channels, *J. Micromech. Microeng.*, 2006, **16**, 1504–1510.
- 38 V. Labrot, M. Schindler, P. Guillot, A. Colin and M. Joanicot, Extracting the hydrodynamic resistance of droplets from their behavior in microchannel networks, *Biomicrofluidics*, 2009, 3, 12804.
- 39 D. A. Sessoms, M. Belloul, W. Engl, M. Roche, L. Courbin and P. Panizza, Droplet motion in microfluidic networks: Hydrodynamic interactions and pressure-drop measurements, *Phys. Rev. E: Stat., Nonlinear, Soft Matter Phys.*, 2009, **80**, 16317.
- 40 A. Grimmer, M. Hamidović, W. Haselmayr and R. Wille, Advanced Simulation of Droplet Microfluidics, ACM J. Emerg. Technol. Comput. Syst., 2019, 15, 1–16.

41 T. Glawdel and C. L. Ren, Global network design for robust operation of microfluidic droplet generators with pressuredriven flow, *Microfluid. Nanofluid.*, 2012, 13, 469–480.

- 42 G. Tetradis-Meris, D. Rossetti, C. Pulido de Torres, R. Cao, G. Lian and R. Janes, Novel Parallel Integration of Microfluidic Device Network for Emulsion Formation, *Ind. Eng. Chem. Res.*, 2009, 48, 8881–8889.
- 43 S. Dube, J. Qin, R. Ramakrishnan and X. Wu, Mathematical Analysis of Copy Number Variation in a DNA Sample Using Digital PCR on a Nanofluidic Device, *PLoS One*, 2008, 3, e2876.
- 44 Lab-on-a-Chip Foundry, https://www.hahn-schickard.de/en/service-portfolio/production/lab-on-a-chip-foundry.
- 45 D. M. Kainz, S. M. Früh, T. Hutzenlaub, R. Zengerle and N. Paust, Flow control for lateral flow strips with centrifugal microfluidics, *Lab Chip*, 2019, 19, 2718–2727.

- 46 J. Park, V. Sunkara, T.-H. Kim, H. Hwang and Y.-K. Cho, Labon-a-Disc for Fully Integrated Multiplex Immunoassays, Anal. Chem., 2012, 84, 2133–2140.
- 47 Y. Chen, N. Zong, F. Ye, Y. Mei, J. Qu and X. Jiang, Dual-CRISPR/Cas12a-Assisted RT-RAA for Ultrasensitive SARS-CoV-2 Detection on Automated Centrifugal Microfluidics, *Anal. Chem.*, 2022, 94, 9603–9609.
- 48 P. Juelg, M. Specht, E. Kipf, M. Lehnert, C. Eckert, M. Keller, T. Hutzenlaub, F. von Stetten, R. Zengerle and N. Paust, Automated serial dilutions for high-dynamic-range assays enabled by fill-level-coupled valving in centrifugal microfluidics, *Lab Chip*, 2019, 19, 2205–2219.
- 49 J.-N. Klatt, M. Depke, N. Goswami, N. Paust, R. Zengerle, F. Schmidt and T. Hutzenlaub, Tryptic digestion of human serum for proteomic mass spectrometry automated by centrifugal microfluidics, *Lab Chip*, 2020, 20, 2937–2946.