


 Cite this: *Sens. Diagn.*, 2023, 2, 427

A sensitive electrochemiluminescence immunosensor for carbohydrate antigen 12-5 analysis based on dual-function SnO₂ nanoflowers†

 Xiangfei Zheng, Dongmiao Qin, Shenglan Hu, Weiming Mo and Biyang Deng *

Tin oxide nanoflowers (SnO₂NFs) are used as both an electrochemiluminescent (ECL) signal generator and a carrier of nanomaterials due to the advantages of excellent electrical conductivity, large specific surface area, and good optical properties. AuNPs/PEI-SnO₂NFs provide a larger binding site for the immobilization of secondary antibodies (CA12-5-Ab₂). The synthesized Ag@ZnONS composite with a large specific surface area and good biocompatibility was applied to immobilize more primary antibodies (CA12-5-Ab₁), which further enhanced the ECL signal and obviously improved the sensitivity and selectivity of the immunosensor. A novel signal amplification ECL immunosensor for detecting carbohydrate antigen 12-5 (CA12-5) was fabricated with AuNPs/PEI-SnO₂NFs as a signal probe and Ag@ZnONSs as the sensing matrix. The ECL intensity was linearly related to the logarithm of the CA12-5 concentration in the range from 1 × 10⁻⁵ to 50 ng mL⁻¹ under optimized conditions. The limit of detection was 2.6 fg mL⁻¹ (S/N = 3). The assay had good operational stability, specificity, and high sensitivity. The study was applied to analyze the CA12-5 level in human serum with satisfactory results, which could be applied in clinical immunoassays.

 Received 6th November 2022,
 Accepted 24th January 2023

DOI: 10.1039/d2sd00198e

rsc.li/sensors

1. Introduction

Ovarian cancer is the main malignant tumor among gynecological cancers and has a high incidence worldwide, posing a serious threat to women's lives.^{1,2} Women with ovarian cancer may not have any symptoms or only have slight symptoms up to the disease progression. Oncologists believe that detecting tumor markers is an important means of achieving early detection, diagnosis, treatment, and mortality reduction of ovarian cancer.³ Carbohydrate antigen 12-5 (CA12-5) is a significant protein marker in serum for the clinical diagnosis of ovarian cancer.⁴ The level of CA12-5 is used to monitor responses to chemotherapy, relapse, and disease progression in ovarian cancer patients. CA12-5 is also regarded as mucin 16 (MUC16);⁵ the level of CA12-5 in human serum is usually lower than 35 U mL⁻¹.⁶ In addition, CA12-5 can also be used as a biomarker in the clinical diagnosis of lung cancer, endometrial cancer, and breast cancer.^{7,8} Therefore, the high-sensitivity detection of CA12-5

has crucial clinical significance for these cancers, such as early screening, early diagnosis, and treatment.

Immunoassay methods based on antibody-antigen-specific binding have been widely used in CA12-5 detection. Currently, the reported immunoassay methods mainly include fluorescence immunoassay,⁹ electrochemical immunoassay,^{10,11} chemiluminescence immunoassay,¹² enzyme-linked immunosorbent assay,¹³ and surface plasma resonance immunoassay.¹⁴ These methods have some weaknesses, including low sensitivity, long analysis time, unstable substrates, and complex experimental procedures. Compared with the aforementioned methods, electrochemiluminescence (ECL) immunoassay has been generally applied in environmental, food, and biological detection due to its low background signal, high sensitivity, good selectivity, rapid response, wider dynamic range, and easier operating procedures.¹⁵⁻²⁰ Sandwich-type ECL immunosensors can achieve satisfactory sensitivity and selectivity. Wang's group designed a sandwich-type ECL immunosensor for determining the CA12-5 level using luminol nanocomposites (AuNPs-EGN-Luminol) labeling Ab₁ as the signal probe and glucose oxidase-encapsulated nano-gold hollow microspheres (GOx-GHS) as a sensing platform for fixing Ab₂.²¹

Luminol, semiconductor quantum dots (QDs), and Ru(bpy)₃²⁺ have been used as a traditional ECL luminophore,

State Key Laboratory for Chemistry and Molecular Engineering of Medicinal Resources, School of Chemistry and Pharmaceutical Sciences, Guangxi Normal University, Guilin 541004, China. E-mail: dengby16@163.com


† Electronic supplementary information (ESI) available. See DOI: <https://doi.org/10.1039/d2sd00198e>



which needs a suitable nanocarrier to load itself for improving luminous efficiency.^{22–24} Still, an efficient ECL luminescent material with excellent electrical conductivity, low toxicity, high stability, and good biocompatibility needs to be developed to assemble immunosensors. Metal oxide nanocrystals such as tin oxide crystals (SnO₂NCs) are promising nanomaterials. Their basic research and applications in sensors have been subjected to an area of concern in terms of chemical and biological analyses.²⁵ Tin oxide nanoflowers (SnO₂NFs) with a flower-like structure, which are formed by the aggregation of ultrathin SnO₂ nanosheets, have good conductivity, high load capacity, and large specific surface area. They are a new type of highly efficient ECL luminophore, which reacts with co-reactant K₂S₂O₈, resulting in cathode ECL emission.²⁶ However, pure SnO₂NFs were electronegative, and the water solubility was poor, limiting their application in biosensors. Polyethylenimine (PEI) is a cationic polyelectrolyte rich in amine, which has excellent hydrophilicity and easily forms complexes with metal ions.^{27,28} In this study, PEI was selected as the stabilizer and dispersant of SnO₂NFs. After the surface modification of SnO₂NFs, PEI not only improved the water solubility of SnO₂NFs but also endowed SnO₂NFs with abundant –NH₂ groups, which was conducive to connecting with other nanomaterials. The prepared PEI-functionalized SnO₂NFs (PEI-SnO₂NFs) were completely encapsulated by cations due to the cationic nature of PEI and could be combined with anionic nanomaterials by electrostatic adsorption.

High-conductivity nanomaterials need to be introduced to enhance the electron transmission capability of SnO₂NFs and amplify the ECL signals so as to improve the sensitivity of the immunosensor. Gold nanoparticles (AuNPs) have the advantages of excellent electrical conductivity, good biocompatibility, and easy coupling with antibodies,²⁹ thus improving the performance of biosensors. The –NH₂ groups of PEI-SnO₂NFs can be used as linkers to combine with AuNPs. The AuNP/PEI-SnO₂NF composites are obtained, and the ECL signal is significantly enhanced. Another important step in assembling the sandwich-type ECL immunosensor is to synthesize the capture probe for labeling a primary antibody (Ab₁). Zinc oxide (ZnO) can grow on various substrates, with high chemical stability and electrochemical activity and low cytotoxicity.³⁰ Compared with zinc oxide nanoparticles (ZnONPs), ZnO nanosheets (ZnONSs) have a larger specific surface area. However, ZnONSs have poor conductivity as an electrode material.³¹ The introduction of AgNPs into ZnONSs not only facilitates the immobilization of antibodies but also improves the electrical conductivity of ZnONSs, thus accelerating the electron transfer of immunosensors.

In this study, SnO₂NFs, as dual-function materials, were used as both an ECL luminophore and a carrier of AuNPs and biomolecules, greatly simplifying the assembly of the immunosensor. The ECL immunosensor for CA12-5 detection was prepared using an Ag@ZnONS-modified electrode as the



Scheme 1 Schematic illustration of ECL immunosensor preparation and CA12-5 detection.

sensing substrates for loading Ab₁ and the AuNP/PEI-SnO₂NF composite as the signal probe-labeled Ab₂. As shown in Scheme 1, SnO₂NFs were used as an ECL luminophore and a nanocarrier in the immunosensor. The PEI-functionalized SnO₂NFs improved the luminous efficiency. It also endowed SnO₂NFs with –NH₂ groups, which were further loaded with AuNPs to obtain the AuNP/PEI-SnO₂NF composite. The introduction of AuNPs could enhance the loading ability of Ab₂ and improved the conductivity of luminescent materials, promoting electron transfer and thereby greatly amplifying ECL signals. Ag@ZnONSs served as a sensing platform to immobilize Ab₁ and enhance the electrical conductivity of the electrode, which further improved the sensitivity of the immunosensor. The ECL sensor had some advantages, such as a wide linear range, high sensitivity, good stability, and repeatability. The study provided a new method for determining other biomarkers.

2. Experimental

2.1 Reagents and apparatus

The analytical-grade reagents were used. Sodium citrate dihydrate (C₆H₅Na₃O₇·2H₂O, 99%), *N*-hydroxysuccinimide, silver nitrate (AgNO₃, 99.8%), tin chloride pentahydrate (SnCl₄·5H₂O, 99%), polyvinyl pyrrolidone (PVP), polyethyleneimine (PEI, 99%), urea (99%), zinc nitrate hexahydrate Zn(NO₃)₂·6H₂O (99%), glucose, H₂HAuCl₄·3H₂O (99.8%), K₃[Fe(CN)₆] (99.5%), K₄[Fe(CN)₆]·3H₂O (99.5%), sodium borohydride (NaBH₄, 99%), and 1-ethyl-3-(3-dimethylaminopropyl) carbodiimide (EDC, 98%) were obtained from Aladdin Chemical Reagent Co., Ltd. (Nanning, China). Potassium chloride (KCl, 99.5%), potassium persulfate (K₂S₂O₈, 99.5%), sodium hydroxide (NaOH, 96%), Na₂HPO₄·12H₂O (99%), and NaH₂PO₄·2H₂O (99%) were



procured from Xilong Chemical Co., Ltd. (Guangzhou, China). Cancer antigen 12-5 (CA12-5), prostate-specific antigen (PSA), human chorionic gonadotropin (HCG), carcinoembryonic antigen (CEA), α -fetoprotein (AFP), squamous cell carcinoma antigen (SCCA), and bovine serum albumin (BSA) were purchased *via* Beijing Boosen Biotechnology Co., Ltd. (Beijing, China). The experimental medium was deionized water.

The composite properties were recorded using a LEO1530 field emission scanning electron microscope (SEM) (Carl Zeiss AG Co., Ltd., Germany), a Tecnai G2 F20 S-TWIN field emission transmission electron microscope (TEM) (FEI, USA), a Max-2200 PC power X-ray diffractometer (XRD) (Rigaku, Japan), a Spectrum Two Fourier-transform infrared spectrophotometer (FT-IR) (Perkin-Elmer, USA), an RF-5301 fluorescence spectrometer (Shimadzu, Japan), and a Cary-60 ultraviolet-visible (UV-vis) absorption spectrometer (Agilent Technologies, USA). ECL signals were recorded with an MPI-ECL analyzer (Xi'an Remex Electronic Science-Tech Co., Ltd., China). A CHI660 electrochemical workstation (Shanghai Chenhua Instrument Co., China) was used for electrochemical impedance spectroscopy (EIS). In this study, the reference electrode was an Ag/AgCl electrode, the auxiliary electrode was a platinum wire, and the working electrode was a modified glassy carbon electrode (GCE). A phosphate buffer solution (PBS, pH 7.4, 0.1 M) was prepared using stock solutions of 0.1 M NaH_2PO_4 and 0.1 M Na_2HPO_4 under definite volume ratios.

2.2 Preparation of SnO_2NFs

SnO_2NFs were prepared using a one-step hydrothermal approach, as reported earlier.³² For this, 1.128 g $\text{SnCl}_4 \cdot 5\text{H}_2\text{O}$ and 2.941 g $\text{C}_6\text{H}_5\text{Na}_3\text{O}_7 \cdot 2\text{H}_2\text{O}$ were mixed with 40 mL of 50% (V/V) anhydrous ethanol and stirred vigorously for 1 h to obtain a uniform solution. The mixture was transferred into a Teflon-lined autoclave and then heated at 180 °C for 12 h. The autoclave was spontaneously cooled to room temperature (RT). The products were obtained *via* centrifugation at 12 000 rpm for 10 min. The resulting product was rinsed using $\text{CH}_3\text{-CH}_2\text{OH}$ and deionized water. Finally, the solid SnO_2NFs product was obtained by drying at 60 °C overnight.

2.3 Synthesis of AuNP/PEI- SnO_2NF composites

Scheme 1 depicts the synthesis of AuNP/PEI- SnO_2NF composites. For this, 10 mg SnO_2NFs were dispersed in absolute ethanol solution (10 mL) and ultrasonicated for 30 min to obtain SnO_2NF solution. To modify the SnO_2NFs with amino groups, 5 mg PEI was mixed with the aforementioned solution and stirred vigorously for 6 h. The sample was centrifuged and rinsed with deionized water three times. The obtained PEI- SnO_2NFs and 2% HAuCl_4 (3000 μL) were added to deionized water (20 mL) and stirred for 5 min. Then, 0.007 g of polyethylene pyrrolidone was dispersed in the aforementioned suspension and stirred continually for 30 min to block the agglomeration of AuNPs. Further, 100 μL of

0.5 M $\text{C}_6\text{H}_5\text{Na}_3\text{O}_7$ (reductant) and NaBH_4 (reductant) were added to the aforementioned solution and stirred continually for 12 h. The solid product was rinsed with deionized water several times till the supernatant became colorless. The AuNP/PEI- SnO_2NF nanocomposites were obtained and dried at 60 °C overnight for subsequent use.

2.4 Synthesis of $\text{Ab}_2/\text{AuNPs}/\text{PEI-}\text{SnO}_2\text{NFs}$

Scheme 1 shows the preparation of $\text{Ab}_2/\text{AuNP}/\text{PEI-}\text{SnO}_2\text{NF}$ bioconjugate. First, 20 mg AuNPs/PEI- SnO_2NFs were mixed with 10 mL of PBS solution. Next, 1 mg mL^{-1} Ab_2 (10 μL) solution was mixed with the AuNP/PEI- SnO_2NF suspension and gently stirred for 6 h. Subsequently, 20 μL of 1% BSA was added to the aforementioned suspension and reacted for 1 h to impede nonspecific binding sites. $\text{Ab}_2/\text{AuNPs}/\text{PEI-}\text{SnO}_2\text{NFs}$ were collected *via* centrifugation and washed with PBS solution to eliminate unreacted substances. Finally, the obtained $\text{Ab}_2/\text{AuNP}/\text{PEI-}\text{SnO}_2\text{NF}$ signal probes were redispersed in 1000 μL of PBS solution and preserved at 4 °C till further use.

2.5 Preparation of the Ag@ZnONS composite

ZnONSs were synthesized as described previously.³³ For this, 1 mM zinc nitrate hexahydrate, 2 mM urea, and 500 mg polyethylene pyrrolidone were consecutively mixed with deionized water and stirred for 30 min. The mixture was transferred into a Teflon-lined autoclave and heated at 120 °C for 8 h. After the autoclave was naturally cooled to RT, the resulting white product was separated through centrifugation (10 000 rpm, 5 min). The ZnONSs were rinsed with anhydrous ethanol and deionized water three times and dried at 60 °C overnight.

An Ag@ZnONS composite was prepared by dissolving 0.03 g ZnONSs, 0.2 g PVP, and 0.1 g glucose in 20 mL of deionized water and magnetically stirred for 10 min at 50 °C. Then, 500 μL of 0.4 M AgNO_3 solution was mixed rapidly with the aforementioned mixture. Next, the mixture was stirred at 90 °C for 2 h to generate silver nanoparticles and modified on the surface of ZnONSs. The brown precipitate was collected by centrifugation and washed with ethanol and deionized water. Solid Ag@ZnONSs were obtained by drying at 60 °C overnight for later use. Further, 1 mg Ag@ZnONS was mixed with 1000 μL of PBS solution and stored at 4 °C until use.

2.6 Construction of the ECL sensor

The sensor was assembled as illustrated in Scheme 1. The electrode was treated by a previously described method.¹⁷ First, 8 μL of 1 mg mL^{-1} Ag@ZnONS was directly covered on the bare electrode surface and naturally dried. Afterward, 8 μL of 100 $\mu\text{g mL}^{-1}$ Ab_1 solution was immobilized onto the surface of GCE/Ag@ZnONSs and reacted at 4 °C for 12 h to obtain GCE/Ag@ZnONSs/ Ab_1 . Subsequently, 5 μL of 1% BSA solution was dropped and reacted at 37 °C for 50 min to prevent nonspecific adsorption. Then, 6 μL of different concentrations of CA12-5 antigen solutions were combined



with the GCE/Ag@ZnONSs/Ab₁/BSA and incubated at 37 °C for 50 min to specifically bind with Ab₁. Finally, 8 μL of Ab₂/AuNP/PEI-SnO₂NF (2 mg mL⁻¹) was incubated with GCE/Ag@ZnONSs/Ab₁/BSA/CA12-5 through specific binding at 37 °C for 1 h. In addition, the modified GCE was washed with PBS (pH 7.4) solution to eliminate excess antibodies. The obtained sandwich-type sensor was kept at 4 °C until use.

2.7 CA12-5 measurement

An electrochemiluminescence test was performed in a cell containing 1 mL of pH 7.4 PBS (0.1 M) including potassium persulfate (0.1 M) solution. The photomultiplier tube (PMT) voltage was set at 800 V. The working potential ranged from -2.0 to 0 V, and the scan rate was 100 mV s⁻¹. The GCE/Ag@ZnONSs/Ab₁/BSA/CA12-5/Ab₂/AuNP/PEI-SnO₂NF immune electrode was used for signal measurement.

3. Results and discussion

3.1 Characterization of the AuNPs/PEI-SnO₂NFs and Ag@ZnONSs

SEM and TEM were used to investigate the morphology of the nanocomposite. Fig. 1A and B show the SEM image of SnO₂NFs. The prepared SnO₂NFs displayed a hierarchic flower-like structure with a size distribution range of 1–2 μm. This structure originated from the assembly of nanosheets with an ultrathin thickness. A majority of nanosheets were vertically arranged in a special mode such that the edge of a nanosheet was connected with the side surface of the other nanosheets. The SnO₂NFs were well dispersed, and the surface was smooth with high growth density, indicating that SnO₂NFs had a large specific surface area. The morphology of the prepared AuNPs/PEI-SnO₂NFs was examined using SEM and is shown in Fig. 1C. It was clearly shown that the surface of PEI-SnO₂NFs was covered with large numbers of small particles. The morphology of AuNPs/PEI-SnO₂NFs was studied using a TEM. Fig. 1D shows that the uniformly dispersed AuNPs were adsorbed on the surface of the nanoflowers. The AuNPs were spherically shaped with different sizes and had an average diameter of 13 nm. The addition of PVP during the synthesis prevented the aggregation of AuNPs, indicating the successful preparation of AuNP/PEI-SnO₂NF nanocomposites. The TEM images of ZnONSs and the TEM and SEM images of Ag@ZnONSs are displayed in Fig. 1E–G. It was obviously seen that the ZnONSs had an ultrathin nanosheet structure and a large surface interface. A large number of AgNPs were loaded on the ZnONSs and had an average diameter of 25 nm. This indicated that AgNPs triumphantly grew on the ZnONSs surface, providing enough contact surfaces for biomolecules.

The SnO₂NFs (a), PEI-SnO₂NFs (b), and AuNPs/PEI-SnO₂NFs (c) were characterized using zeta potential and FT-IR spectroscopy. As displayed in Fig. 2A, the zeta potential value of SnO₂NFs was -33.5 mV. After PEI modification, the zeta potential value of PEI-SnO₂NFs increased to +39 mV, which belonged to high positive charges of the -NH₂ of PEI on the SnO₂NFs surface. When the PEI-SnO₂NFs were loaded with negatively charged AuNPs through electrostatic adsorption, the zeta potential value of the AuNPs/PEI-SnO₂NFs dropped to +11.3 mV. Fig. 2B shows the FT-IR spectra of SnO₂NFs, PEI-SnO₂NFs, and AuNPs/PEI-SnO₂NFs. It was observed that SnO₂NFs (curve a) had infrared peaks at 3430 cm⁻¹ and 1651 cm⁻¹, which depicted the stretching and bending vibrations of -OH. The main peak at 592 cm⁻¹ was attributed to the antisymmetric vibration of O-Sn-O.³⁴ The FT-IR spectrum of PEI-SnO₂NFs (curve b) had two new peaks at 2930 cm⁻¹ and 1120 cm⁻¹, corresponding to C-H stretching vibration and N-H bending vibration.³⁵ It indicated that PEI had successfully modified SnO₂NFs. The peak intensity of AuNPs/PEI-SnO₂NFs (curve c) reduced after AuNPs were adsorbed on the surface of PEI-SnO₂NFs. The zeta potential measurements and FT-IR spectra proved that the AuNP/PEI-SnO₂NF composites were successfully prepared.




Fig. 1 (A and B) SEM images of SnO₂NFs with different magnifications. (C) SEM and (D) TEM images of AuNPs/PEI-SnO₂NFs. (E) TEM images of ZnONSs. (F) TEM and (G) SEM images of Ag@ZnONSs.






Fig. 2 Zeta potential (A) and FT-IR spectra (B) of SnO₂NFs (a), PEI-SnO₂NFs (b), and AuNPs/PEI-SnO₂NFs (c).

XPS characterization was carried out to determine the elemental composition and chemical oxidation valence of the SnO₂NF survey spectrum (Fig. S1A[†]), Sn3d spectrum (Fig. S1B[†]), and O1s spectrum (Fig. S1C[†]). The details of XPS spectra are included in the ESI[†] (Fig. S1). The SnO₂NF, PEI-SnO₂NF, AuNP, and AuNP/PEI-SnO₂NF composites were investigated using UV-vis (Fig. S2A[†]). The crystal structures of AuNPs, SnO₂NFs, and AuNPs/PEI-SnO₂NFs (Fig. S2B[†]) were characterized using XRD. A detailed explanation of UV-vis and XRD spectra is shown in the ESI[†] (Fig. S2). The elemental analyses of the AuNPs/PEI-SnO₂NF (Fig. S3A[†]) and Ag@ZnONS (Fig. S3B[†]) composite were performed using EDS. A detailed explanation of EDS diagrams is displayed in the ESI[†] (Fig. S3).

3.2 Electrochemical and ECL characteristics of the ECL immunosensor

The Ag@ZnONS-modified GCE was used for the sensing platform to fix Ab₁, then GCE/Ag@ZnONS/Ab₁ was used to capture CA12-5 antigen, and Ab₂/AuNP/PEI-SnO₂NF signal

probe was used to fabricate a sandwich ECL sensor. The modification process of each step for the GCE was researched using CV in 5 mM K₃[Fe(CN)₆]/K₄[Fe(CN)₆] including 0.1 M KCl; the working voltage was between -0.2 and 0.6 V. As displayed in Fig. 3A, the unmodified GCE electrode showed a couple of symmetric redox peaks (curve a). After Ag@ZnONSs were covered on the bare electrode (curve b), the redox peak current decreased owing to the low electrical conductivity of ZnONSs. When Ab₁ was adsorbed to the GCE/Ag@ZnONS surface, the peak currents further decreased (curve c). This might be because the biomolecules hindered the electron transfer.³⁶ The redox currents were significantly reduced after continuous assembly of the BSA and CA12-5 antigens on the modified electrode, which was due to the hindering effect of insulating protein molecules on the transfer of electrons (curves c and d). In the end, when the GCE/Ag@ZnONSs/Ab₁/BSA/CA12-5 reacted with the Ab₂/AuNP/PEI-SnO₂NF signal probe, the peak current further decreased (curve f). The AuNPs increased the conductivity, but the protein molecules hindered the electronic transmission. The results indicated the successful assembly of the immunosensor.




Fig. 3 (A) CV and (B) EIS plot of the diverse modified GCEs: (a) bare GCE, (b) GCE/Ag@ZnONSs, (c) Ab₁/Ag@ZnONSs/GCE, (d) BSA/Ab₁/Ag@ZnONSs/GCE, (e) CA12-5/BSA/Ab₁/Ag@ZnONSs/GCE, and (f) Ab₂/AuNPs/PEI-SnO₂NFs/CA12-5/BSA/Ab₁/Ag@ZnONSs/GCE in 5 mM [Fe(CN)₆]^{3-/4-} solution including 0.1 M KCl. Fig. 3B inset: equivalent circuit for EIS Nyquist curves. Experimental conditions: CV scan rate, 0.1 V s⁻¹; working potential, -0.2 to 0.6 V; EIS amplitude, 5 mV; and EIS frequency, 0.01 to 10⁵ Hz.



The EIS was used to further prove the stepwise assembly procedure of the ECL sensor. It comprised a linear part and a semicircular part. The linear part was related to the diffusion-limiting procedure, and the semicircular part was related to the electron transfer resistance (R_{ct}). The equivalent circuit from EIS is shown in Fig. 3B (inset), which contains the Warburg impedance (Z_w), the resistance of solution (R_s), and the double layer capacitance (C_{dl}).³⁷ Fig. 3B shows the EIS characterization of different modified electrodes. Compared with the bare electrode (curve a), GCE/Ag@ZnONSs showed a big semi-circle, indicating a high R_{ct} (curve b). This was attributed to the low conductivity of ZnONSs. After Ab₁ (curve c), BSA (curve d) and CA12-5 antigens (curve e) were modified continuously on the modified GCE, and the diameter of the semi-circle gradually increased. This was due to the presence of the protein layer, which prevented the transfer of electrons.³⁸ Finally, the impedance value further increased when the Ab₂/AuNPs/PEI-SnO₂NFs were connected to the GCE/Ag@ZnONSs/Ab₁/BSA/CA12-5 to construct a sandwich-type ECL sensor (curve f). The EIS results were similar to CV, confirming the successful assembly.

3.3 ECL behavior and electrochemical

The ECL behavior and electrochemical of AuNPs/PEI-SnO₂NFs were investigated. The ECL signal-potential curves (Fig. 4A) and CV waves (Fig. 4B) of the bare electrode, SnO₂NF/GCE, PEI-SnO₂NF/GCE, and AuNP/PEI-SnO₂NF/GCE composites are shown in Fig. 4, which were measured in 0.1 M PBS (pH = 7.4) including K₂S₂O₈ (0.1 M). The bare GCE produced a low ECL signal with a value of 535 a.u. (Fig. 4A, curve a). This was due to the ECL signal of ¹(O₂)₂^{*} in the O₂/S₂O₈²⁻ platform. When the GCE was modified with SnO₂NFs detected in the S₂O₈²⁻ solution (curve b), the ECL signal was enhanced significantly to 5046 a.u. Compared with a bare GCE, the ECL signal of SnO₂NFs/GCE increased six-fold. In addition, the ECL emissions of GCE/SnO₂NFs showed no significant change under 12 successive cyclic potential scans; the relative standard deviation (RSD) was 0.6% with acceptable stability (Fig. 4A, inset). Fig. 4A (curve c) shows

that the ECL signal of GCE/PEI-SnO₂NFs was better than that of GCE/SnO₂NFs, and GCE/PEI-SnO₂NFs exhibited more positive reduction potential because the presence of PEI enhanced the water solubility of SnO₂NFs. When gold nanoparticles were adsorbed on the PEI-SnO₂NFs surface (curve d), the ECL intensity was further enhanced significantly, demonstrating that AuNPs could enhance the conductivity of luminescent materials and quicken the transfer of electrons between SnO₂NFs and electrodes, thus increasing the ECL intensity of SnO₂NFs.

As seen in Fig. 4B, there were no remarkable redox peaks of the bare electrode (curve a). After SnO₂NFs were modified on the bare GCE, a noticeable reduction peak was observed at -1.60 V (curve b). After the PEI was introduced into SnO₂NFs, the reduction peak potential of the CV curve (curve c) exhibited a slight positive shift to -1.54 V. Subsequently, the AuNPs/PEI-SnO₂NFs showed a more positive potential at about -1.39 V (curve d), the peak current was increased compared with other curve, indicating that the large surface area and good conductivity of AuNPs were beneficial to electron transport.

In this study, SnO₂NFs (luminophore) could react with S₂O₈²⁻ (co-reactant). With the cathodic polarization of the potential, SnO₂NFs on the electrode acquired an electron and were reduced to anion radical SnO₂NFs⁻ [eqn (1)]. S₂O₈²⁻ was reduced to the intermediate of S₂O₈^{3•-} [eqn (2)], and the strong oxidant SO₄⁻ was obtained *via* the decomposition of S₂O₈^{3•-} [eqn (3)]. Then, SO₄⁻ could react with SnO₂NFs⁻ to produce the excited state SnO₂NFs* [eqn (4)]. Hence, the unstable SnO₂NFs* returned to the ground state with obvious light emission [eqn (5)]. In accordance with the literature,²⁶ the ECL emission mechanism of SnO₂NFs is expressed as follows.







Fig. 4 (A) ECL intensity-potential plots and (B) cyclic voltammograms of the (a) GCE, (b) SnO₂NFs/GCE, (c) PEI-SnO₂NFs/GCE, and (d) AuNPs/PEI-SnO₂NFs/GCE. Insets: The stability of GCE/SnO₂NFs under continuous CV scanning for 12 cycles. Experimental conditions: electrode scanning from -0.2 to 0 V with pH 7.4 PBS (0.1 M) including potassium persulfate (0.1 M).





3.4 Optimization of analytical conditions

Some experiment conditions were optimized, such as the concentration of the co-reactant $\text{K}_2\text{S}_2\text{O}_8$, pH value, and concentrations of the $\text{Ag}@Z\text{nONSs}$ and Ab_1 to obtain the best ECL emission. The concentration of the $\text{K}_2\text{S}_2\text{O}_8$ had a great influence on the ECL sensor. As shown in Fig. 5A, the ECL signal gradually increased when the $\text{K}_2\text{S}_2\text{O}_8$ concentration was between 0.02 M and 0.1 M. When the $\text{K}_2\text{S}_2\text{O}_8$ concentration was 0.1 M, the ECL intensity was maximal. It was possible that more $\text{SO}_4^{\cdot-}$ was produced with the increase in the $\text{K}_2\text{S}_2\text{O}_8$ concentration, which promoted the generation of more excited states SnO_2NFs^* . However, when the $\text{K}_2\text{S}_2\text{O}_8$ concentration increased to more than 0.1 M, the ECL intensity decreased significantly, which was because excessive $\text{S}_2\text{O}_8^{2-}$ could generate charge repulsion. Therefore, 0.1 M $\text{K}_2\text{S}_2\text{O}_8$ was chosen as the best concentration for CA12-5 detection. The influence of pH was optimized to realize the best CA12-5 detection. As shown in Fig. 5B, the ECL signal increased gradually when the pH was increased from 6.0 to 7.4 and reached the maximum ECL intensity at pH 7.4. When the pH increased to more than 7.4, the ECL intensity started to decrease. This was because the protons were reduced readily under acidic conditions, leading to the suppression of the ECL signal. Besides, alkaline conditions could influence the activity of the antibodies and antigens, thus affecting the signal of ECL. Therefore, the ECL detection solution of pH at 7.4 was selected for the detection of CA12-5. The




Fig. 6 ECL emission-potential plots of the modified electrode in 0.1 mol L^{-1} $\text{K}_2\text{S}_2\text{O}_8$ solution (pH 7.4): (a) $\text{Ag}@Z\text{nONSs}/\text{GCE}$, (b) $\text{Ab}_1/\text{Ag}@Z\text{nONSs}/\text{GCE}$, (c) $\text{BSA}/\text{Ab}_1/\text{Ag}@Z\text{nONSs}/\text{GCE}$, (d) $\text{CA12-5}/\text{BSA}/\text{Ab}_1/\text{Ag}@Z\text{nONSs}/\text{GCE}$, (e) $\text{Ab}_2/\text{AuNPs}/\text{PEI-SnO}_2\text{NFs}/\text{CA12-5}/\text{BSA}/\text{Ab}_1/\text{Ag}@Z\text{nONSs}/\text{GCE}$, and (f) $\text{BSA}/\text{Ab}_1/\text{Ag}@Z\text{nONSs}/\text{GCE}$ without CA12-5 but dipped in $\text{Ab}_2/\text{AuNPs}/\text{PEI-SnO}_2\text{NF}$ suspension. The CA12-5 concentration was 0.1 ng mL^{-1} .

concentration of $\text{Ag}@Z\text{nONSs}$ for the assembly of the immunosensor was also researched. As shown in Fig. 5C, the ECL signal gradually increased with the increase in the concentration of $\text{Ag}@Z\text{nONSs}$ and then attained a maximal value at 1.0 mg mL^{-1} . Therefore, the best concentration of $\text{Ag}@Z\text{nONSs}$ was 1 mg mL^{-1} . In addition, the CA12-5- Ab_1 concentration was also optimized. As shown in Fig. 5D, the ECL intensity was the highest when the Ab_1 concentration reached 100 $\mu\text{g mL}^{-1}$, and then gradually tended to stabilize at a higher concentration. The low-concentration Ab_1 was insufficient to capture the CA12-5 antigen and Ab_2 signal probe, affecting the ECL intensity. As the concentration of




Fig. 5 (A) Effect of $\text{K}_2\text{S}_2\text{O}_8$ concentration, (B) pH of PBS solution, (C) $\text{Ag}@Z\text{nONSs}$ concentration, and (D) Ab_1 concentration on the ECL signal of the sensor toward 1×10^{-2} ng mL^{-1} CA12-5.






Fig. 7 (A) ECL intensities of the sensor for the determination of CA12-5 at diverse concentrations (a-i): 1×10^{-5} , 1×10^{-4} , 1×10^{-3} , 1×10^{-2} , 0.1, 1, 10, 25, and 50 ng mL^{-1} . (B) Plot between ECL intensity and $\log C_{\text{CA12-5}}$.

Ab_1 increased, more immunosensors were conjugated on the electrode. Therefore, the best concentration of the Ab_1 was $100 \mu\text{g mL}^{-1}$.

3.5 Analytical performance

The stepwise fabrication process of the immunosensor was recorded with ECL emission-potential in $0.1 \text{ mol L}^{-1} \text{ K}_2\text{S}_2\text{O}_8$ solution (pH 7.4). As shown in Fig. 6, firstly, the Ag@ZnONSs/GCE exhibited a good ECL emission (curve a), which was ascribed to the excellent conductivity and large specific surface area of Ag@ZnONSs . After successively modified with Ab_1 (curve b), BSA (curve c) and CA12-5 antigen (curve d), the corresponding ECL intensity decreased sequentially. When the $\text{Ab}_2/\text{AuNPs/PEI-SnO}_2\text{NFs}$ were adsorb on the $\text{CA12-5/BSA/Ab}_1/\text{Ag@ZnONSs/GCE}$, the ECL signal obviously increased (curve e). The ECL intensity of $\text{BSA/Ab}_1/\text{Ag@ZnONSs/GCE}$ without CA12-5 but dipped in $\text{Ab}_2/\text{AuNPs/PEI-SnO}_2\text{NFs}$ suspension was lower than the intensity of $\text{Ab}_2/\text{AuNPs/PEI-SnO}_2\text{NFs/CA12-5/BSA/Ab}_1/\text{Ag@ZnONSs/GCE}$.

However, the intensity of curve f was more than those of curves a, b and c. The change of the ECL signal value of the immunosensor with CA12-5 was remarkable, which demonstrated that the immunosensor exhibited good performance.

Under optimal conditions, the developed immunosensor was examined by incubating with different concentrations of the CA12-5 antigen. As shown in Fig. 7A, the ECL responses increased step by step in the CA12-5 concentration range of 1×10^{-5} to 50 ng mL^{-1} (curves a-i). The linear relation between the ECL signal (Y) and the logarithmic values of CA12-5 concentrations is plotted in Fig. 7B. The standard curve equation was $Y = 10900 + 1771.6 \log C_{\text{CA12-5}}$ with a correlation coefficient (R^2) of 0.9989. The limit of detection (LOD) was 2.6 fg mL^{-1} ($S/N = 3$).

Table 1 provides the comparison of this method with other previously reported various methods for CA12-5 detection.³⁹⁻⁴⁴ The proposed ECL immunoassay showed a lower LOD and broader linear range, indicating that the immunosensor had high sensitivity and excellent potential for the trace determination of CA12-5. The high sensitivity of this method was attributed to the following aspects: (1) $\text{SnO}_2\text{-NFs}$ had a large specific surface area and excellent conductivity, and hence could be used as both ECL emitters and nanocarriers. The $\text{AuNP/PEI-SnO}_2\text{NF}$ composites displayed stable and strong ECL intensity, which improved the performance of the immunosensor. (2) Ag@ZnONSs with fine conductivity provided a large surface area for loading more Ab_1 . (3) AuNPs could not only immobilize enough Ab_2 but also accelerate the transfer of electrons and increase the sensitivity of the developed ECL immunosensor.

Table 1 Comparison of methods for CA12-5 detection

Method	Linear range	LOD	Reference
$\text{CuBTC@MoS}_2\text{-AuNPs}$ electrochemistry	0.0005 to 500 U mL^{-1}	0.0005 U mL^{-1}	39
MOF-808/CNT electrochemistry	0.001 to 30 ng mL^{-1}	0.5 pg mL^{-1}	40
$\text{AuNPs-PAMAM/CDs FRET}$	1×10^{-6} to 1.0 ng mL^{-1}	0.5 fg mL^{-1}	41
$\text{AgInS}_2/\text{ZnS NCs ECL}$	5×10^{-6} to 0.005 U mL^{-1}	1×10^{-6} U mL^{-1}	42
$\text{NIR CdTe/CdS QDs-SiO}_2$ ECL	0.0001 to 10 U mL^{-1}	3.4×10^{-5} U mL^{-1}	43
Met-capped Au-Ag BNCs ECL	5×10^{-4} to 1 U mL^{-1}	5×10^{-5} U mL^{-1}	44
$\text{AuNPs/PEI-SnO}_2\text{NFs ECL}$	1×10^{-5} to 50 ng mL^{-1}	2.6 fg mL^{-1}	This study

3.6 Stability, reproducibility, and selectivity of the ECL immunosensor

The stability of the immunosensor was studied *via* testing CA12-5 (0.1 ng mL^{-1}) for 12 cycles of successive CV scanning. As exhibited in Fig. 8A, the ECL signals of the immunosensor displayed no remarkable changes (RSD was 1.8%), manifesting the acceptable stability of the immunosensor. The storage stability of the sensor was also studied *via* regular inspection of its relative activity. The signal of ECL of the freshly constructed immunosensor was used as the initial value. After storage of 7, 15, 23, and 30 days, the ECL intensity of the immunosensor for the determination of CA12-5 (0.1 ng mL^{-1}) was 97.4%, 92.5%, 87.8%, and 83.6% of the initial intensity of $\text{AuNPs/PEI-SnO}_2\text{NFs}$, respectively (Fig. 8B). Hence, the






Fig. 8 (A) ECL stability under successive scanning for 12 cycles in 0.1 M PBS (pH 7.4) including $K_2S_2O_8$ (0.1 M) for the determination of CA12-5 (0.1 ng mL^{-1}). (B) Stability of the ECL sensor when stored for 30 days (0.1 ng mL^{-1} CA12-5). (C) Reproducibility of the ECL immunosensor toward 0.1 ng mL^{-1} CA12-5 concentration. (D) Selectivity and specificity of the sensor to (1) CA12-5 (1 ng mL^{-1}), (2) AFP (1 ng mL^{-1}), (3) PSA (1 ng mL^{-1}), (4) CEA (1 ng mL^{-1}), (5) SCCA (1 ng mL^{-1}), (6) HCG (1 ng mL^{-1}), (7) CA12-5 (1 ng mL^{-1}) + AFP (100 ng mL^{-1}), (8) CA12-5 (1 ng mL^{-1}) + PSA (100 ng mL^{-1}), (9) CA12-5 (1 ng mL^{-1}) + CEA (100 ng mL^{-1}), (10) CA12-5 (1 ng mL^{-1}) + SCCA (100 ng mL^{-1}), and (11) CA12-5 (1 ng mL^{-1}) + HCG (100 ng mL^{-1}).

Table 2 Analytical results and recoveries for CA12-5 in human serum samples ($n = 6$)

Serum sample	Original (ng mL^{-1})	Added (ng mL^{-1})	Found (ng mL^{-1})	Recovery (%)	RSD (%)	Hospital (chemiluminescence, ng mL^{-1})
1	0.9200	1.000	1.910 ± 0.0400	99.00	2.1	0.9600
2	4.730	5.000	9.950 ± 0.200	104.4	2.0	5.100
3	9.860	10.00	20.00 ± 0.56	101.4	2.8	9.350
4	25.88	20.00	46.60 ± 1.63	103.6	3.5	26.40

immunosensor had long-term stability. The composite material of Ag@ZnONs and AuNPs/PEI-SnO₂NFs had good biocompatibility and provided a good microenvironment for the activity of immobilized antibodies. The results showed that the immunosensor had excellent storage stability.

CA12-5 (0.1 ng mL^{-1}) was tested using seven equally immune GCEs under the same conditions to assess the repeatability of the immunosensor. As shown in Fig. 8C, the RSD lower than 2.5% in the ECL signal was obtained, which demonstrated the superior reproducibility of the immunosensors. Besides, selectivity is an important indicator for analyzing biological samples using immunosensors. Some interfering substances were researched, such as alpha-fetoprotein (AFP), PSA, carcinoembryonic antigen (CEA), SCCA, and HCG, to investigate the selectivity of the proposed immunosensor. As shown in Fig. 8D, the signal of ECL of 1 ng mL^{-1} CA12-5 was well above that of AFP, PSA, CEA, SCCA, and HCG with the same concentration, and the intensity of the interferences was negligible. In addition, the intensity of

ECL with a mixture solution [CA12-5 (1 ng mL^{-1}) consisting of AFP (100 ng mL^{-1}), PSA (100 ng mL^{-1}), CEA (100 ng mL^{-1}), SCCA (100 ng mL^{-1}), and HCG (100 ng mL^{-1})] was nearly similar to the ECL response of CA12-5, demonstrating that the ECL immunosensor had acceptable selectivity and specificity to detect CA12-5.

3.7 Analyses of real samples

The recovery experiments were used for CA12-5 determination in serum samples by the standard addition method to investigate the clinical application potential of the proposed immunosensor. Various concentrations of sample solution were prepared by adding CA12-5 standard solutions to human serum (the Guilin Fifth People's Hospital). Then, the immunosensor was reacted with the aforementioned prepared sample solution. The analytical results are shown in Table 2. The recovery between 99.00% and 104.4% for CA12-5 with an RSD of 2.0–3.5% indicated that the ECL



immunosensor had the potential for determining CA12-5 in human serum.

4. Conclusions

This study used AuNPs/PEI-SnO₂NFs as labels and Ag@ZnONS-modified electrode as the substrate to construct a sensitive sandwich-type ECL immunosensor with signal enhancement. SnO₂NFs had a dual function as both an ECL luminophore and a carrier, simplifying the assembly of immunosensors. PEI-modified SnO₂NFs provided a good interface for the loading of AuNPs and Ab₂. The AuNP/PEI-SnO₂NF nanocomposite showed excellent ECL luminous performance that realized the signal amplification for the SnO₂NFs/K₂S₂O₈ system. Ag@ZnONSs were used as the sensing matrix with good conductivity, large specific surface area, and biocompatibility, which could immobilize sufficient Ab₁ and improved the sensitivity of immunosensors. The immunosensor displayed excellent selectivity, stability, good reproducibility, high sensitivity, and better detection linear range. The study might have enormous application potential in biological and clinical analyses.

Author contributions

Xiangfei Zheng: formal analysis, investigation, data curation, and writing – original draft. Dongmiao Qin: critical revision including writing – review & editing, as well as preparation of the article for publication. Shenglan Hu: validation and data curation. Weiming Mo: writing – review & editing. Biyang Deng: supervision, project administration, and funding acquisition.

Conflicts of interest

The authors declare no conflict of interest.

Acknowledgements

This study was financially supported by the National Natural Science Foundation of China (22164004) and the Guangxi Science Foundation of China (2022GXNSFDA035069).

References

- N. Razmi and M. Hasanzadeh, *TrAC, Trends Anal. Chem.*, 2018, **108**, 1–12.
- S. Zhang, Y. Huang, Y. Chen, S. Yan, H. Dai and J. Zhao, *Sens. Diagn.*, 2023, **2**, 140–146.
- L. Farzin, S. Sadjadi, M. Shamsipur, S. Sheibani and M. H. Mousazadeh, *Mater. Sci. Eng., C*, 2019, **97**, 679–687.
- M. Wu, S. Liu, F. Qi, R. Qiu, J. Feng, X. Ren, S. Rong, H. Ma, D. Chang and H. Pan, *Talanta*, 2022, **241**, 123254.
- L. Farzin and M. Shamsipur, *J. Pharm. Biomed. Anal.*, 2018, **147**, 185–210.
- C. Zhao, C. Li, M. Li, L. Qian, L. Wang and H. Li, *Sens. Actuators, B*, 2022, **367**, 132063–132072.
- Y. Fan, S. Shi, J. Ma and Y. Guo, *Biosens. Bioelectron.*, 2019, **135**, 1–7.
- L. Wu, Y. Sha, W. Li, S. Wang, Z. Guo, J. Zhou, X. Su and X. Jiang, *Sens. Actuators, B*, 2016, **226**, 62–68.
- S. Hamd-Ghadareh, A. Salimi, F. Fathi and S. Bahrami, *Biosens. Bioelectron.*, 2017, **96**, 308–316.
- M. Chen, R. Han, W. Wang, Y. Li and X. Luo, *Anal. Chem.*, 2021, **93**, 13555–13563.
- M. Wu, S. Liu, F. Qi, R. Qiu, J. Feng, X. Ren, S. Rong, H. Ma, D. Chang and H. Pan, *Talanta*, 2022, **241**, 123254–123260.
- I. Al-Ogaidi, H. Gou, Z. P. Aguilar, S. Guo, A. K. Melconian, A. K. Al-Kazaz, F. Meng and N. Wu, *Chem. Commun.*, 2014, **50**, 1344–1346.
- X. Ren, H. Wang, D. Wu, D. Fan, Y. Zhang, B. Du and Q. Wei, *Talanta*, 2015, **144**, 535–541.
- K. Zhang and X. Shen, *Analyst*, 2013, **138**, 1828–1834.
- G. Mo, X. He, C. Zhou, D. Ya, J. Feng, C. Yu and B. Deng, *Biosens. Bioelectron.*, 2019, **126**, 558–564.
- G. Mo, D. Qin, X. Jiang, X. Zheng, W. Mo and B. Deng, *Sens. Actuators, B*, 2020, **310**, 127852–127859.
- G. Mo, X. He, D. Qin, S. Meng, Y. Wu and B. Deng, *Biosens. Bioelectron.*, 2021, **178**, 113024–113031.
- P. Gai, X. Kong, L. Pu, M. Zhang, D. Zhu and F. Li, *Anal. Chem.*, 2021, **93**, 11745–11750.
- C. Wang, Q. Han, P. Liu, G. Zhang, L. Song, X. Zou and Y. Fu, *Anal. Chem.*, 2021, **93**, 12289–12295.
- J. Cao, L. Zhao, Y. Fu, X. Liu, S. Ren and Y. Liu, *Sens. Actuators, B*, 2021, **331**, 129427–129432.
- X. Wang, Z. Bian, C. Chu, X. Zheng, S. Ge, J. Yu, M. Yan and X. Song, *RSC Adv.*, 2014, **4**, 52796–52803.
- L. Shang, B. Shi, W. Zhang, L. Jia, R. Ma, Q. Xue and H. Wang, *Anal. Chem.*, 2022, **94**, 12845–12851.
- J. Adhikari, M. Rizwan and M. U. Ahmed, *Sens. Diagn.*, 2022, **1**, 968–976.
- C. Yang, Y. Tian, B. Wang, Q. Guo and G. Nie, *Sens. Actuators, B*, 2021, **338**, 129870.
- Y. Lei, Y. Zhuo, M. Guo, Y. Chai and R. Yuan, *Anal. Chem.*, 2020, **92**, 2839–2846.
- M. Jiang, P. Lu, Y. Lei, Y. Chai, R. Yuan and Y. Zhuo, *Electrochim. Acta*, 2018, **271**, 464–471.
- D. Qin, X. Jiang, G. Mo, J. Feng, C. Yu and B. Deng, *ACS Sens.*, 2019, **4**, 504–512.
- M. Y. Kim, H. Park, J. Y. Lee, D. J. Park, J. Y. Lee, N. V. Myung and K. H. Lee, *Chem. Eng. J.*, 2022, **431**, 134250–134261.
- D. Qin, S. Meng, Y. Wu, G. Mo and B. Deng, *Sens. Actuators, B*, 2022, **367**, 132176–132185.
- F. Haghayegh, R. Salahandish, M. Hassani and A. Sanati-Nezhad, *ACS Appl. Mater. Interfaces*, 2022, **14**, 10844–10855.
- H. Li, Y. Wei, Y. Zhang, F. Yin, C. Zhang, G. Wang and Z. Bakenov, *Ionics*, 2016, **22**, 1387–1393.
- Y. Liu, Y. Jiao, Z. Zhang, F. Qu, A. Umar and X. Wu, *ACS Appl. Mater. Interfaces*, 2014, **6**, 2174–2184.
- L. Zhu, W. Zeng and Y. Li, *Mater. Lett.*, 2018, **228**, 331–333.
- S. Pethaperumal and G. T. Mohanraj, *Surf. Interfaces*, 2022, **30**, 101872–101878.



- 35 X. Sun, L. Zhou and W. Zhao, *Bioelectrochemistry*, 2022, **145**, 108104–108110.
- 36 M. Sentic, M. Milutinovic, F. Kanoufi, D. Manojlovic, S. Arbault and N. Sojic, *Chem. Sci.*, 2014, **5**, 2568–2572.
- 37 C. Yang, Q. Guo, Y. Lu, B. Zhang and G. Nie, *Sens. Actuators, B*, 2020, **303**, 127246.
- 38 H. B. Habtamu, M. Sentic, M. Silvestrini, L. D. Leo, T. Not, S. Arbault, D. Manojlovic, N. Sojic and P. Ugo, *Anal. Chem.*, 2015, **87**, 12080–12087.
- 39 S. Li, C. Hu, C. Hu, C. Chen, J. Zhang, Y. Bai, C. Tan, G. Ni, F. He, W. Li and D. Min, *ACS Appl. Bio Mater.*, 2021, **4**, 5494–5502.
- 40 S. Biswas, Q. Lan, Y. Xie, X. Sun and Y. Wang, *ACS Appl. Mater. Interfaces*, 2021, **13**, 3295–3302.
- 41 S. Hamd-Ghadareh, A. Salimi, F. Fathi and S. Bahrami, *Biosens. Bioelectron.*, 2017, **96**, 308–316.
- 42 M. Yin, Y. Wang, X. Gao, S. Du, Y. Cheng, S. Yu, G. Zou and F. Xue, *Anal. Bioanal. Chem.*, 2021, **413**, 2207–2215.
- 43 H. Gao, Z. Zhang, Y. Zhang, H. Yu, S. Rong, L. Meng, S. Song, Y. Mei, H. Pan and D. Chang, *J. Electroanal. Chem.*, 2021, **886**, 115104.
- 44 L. Fu, X. Gao, S. Dong, H. Hsu and G. Zou, *Anal. Chem.*, 2021, **93**, 4909–4915.

