Sensors & Diagnostics



PERSPECTIVE

View Article Online



Cite this: Sens. Diagn., 2023, 2, 620

All-in-one terahertz taste sensor: integrated electronic and bioelectronic tongues

Jin Wang, ** Kenji Sakai and Toshihiko Kiwa

Taste sensors, also known as electronic tongues or bioelectronic tongues, are designed to evaluate food and beverages, as well as for medical diagnostics. These devices mimic the ability of the human tongue to detect and identify different tastes in liquid samples, such as sweet, sour, salty, bitter, and umami. In this study, a novel all-in-one terahertz taste sensor was proposed, which differs from traditional electrochemical approaches. This sensor utilizes terahertz technology for imaging and sensing chemical reactions on the terahertz semiconductor emitter surface. The surface can be functionalized with ionsensitive membranes, proteins, DNA aptamers, and organic receptors, enabling the detection of various substances, such as solution pH, physiological ions, sugars, toxic chemicals, drugs, and explosives. Terahertz taste sensors offer several advantages, including being label-free, high sensitivity and selectivity, rapid response, minimal sample consumption, and the ability to detect non-charged chemical substances. By integrating multiple receptors or sensing materials on a single chip, the all-in-one terahertz taste sensor has significant potential for future taste substance detection, nutrition evaluation, metabolite and drug monitoring, and biomarker sensing.

Received 10th February 2023, Accepted 11th March 2023

DOI: 10.1039/d3sd00038a

rsc.li/sensors

Taste sensors

Taster sensors, also known as electronic or bioelectronic tongues, are devices that mimic the human sense of taste by identifying and measuring different chemical compounds in a sample solution. Researchers began developing taste sensors in the late twentieth century to analyze the chemical compounds found in foods and beverages, which are difficult to measure or identify. These electronic tongues are generally potentiometric or amperometric devices or equipment with ion-selective electrodes,1 metal oxides,2,3 or artificial lipid polymer arrays, 4-8 which can detect or measure changes in pH and ion concentration as well as identify a wide range of taste substances or chemical compounds in liquid samples (milk, beer, tea, etc.) by measuring the electrical potential or current. Legin et al. developed a sensor array-based electronic tongue and pattern recognition algorithm for the qualitative measurement of several different beverages.9 Dias et al. successfully built 36 cross-sensibility sensors for the recognition of five basic taste substances. 10 Moreover, several sensor devices and equipment have been manufactured and commercialized for laboratory research and industrial applications. For example, the TS-5000Z taste sensing system

Graduate School of Interdisciplinary Science and Engineering in Health Systems. Okayama University, 3-1-1, Tsushima-Naka, Kitaku, Okayama, 700-8530, Japan. E-mail: wangjin@okayama-u.ac.jp

(Fig. 1A), developed by Professor Toko and Intelligent Sensor Technology Inc., Japan, detects five basic tastes, flavor properties (sharpness and richness), and taste masking properties (Fig. 1B). 11,12 Another commercial tongue consists of seven ion-selective field effect transistors (ISFET) developed by Astree II (Alpha MOS, Toulouse, France), which are also featured in taste evaluation, food quality control, and pharmaceutical industry applications.

In the early twenty-first century, another important development of the bioelectronic tongue was presented to analyze complex mixtures of chemical compounds and identify individual compounds that are not available using traditional methods. 13-16 Bioelectronic tongues are based on a variety of recognition elements, 17-23 such as nucleic acids, proteins, antibodies, cells, and taste receptors. A key feature of bioelectronic tongues is their high selectivity and sensitivity. This enables the identification of specific chemical compounds that may not be detectable using traditional analytical methods. Furthermore, bioelectronic tongues can be used for the detection of biomolecules, such as proteins and nucleic acids, using enzyme-, antibody-, or nucleic acid-based receptors. Qin et al. proposed a biohybrid tongue that combines a microelectrode array chip with mammalian gustatory epithelium to create a highly sensitive taste sensing system. The study quantified the interaction effects of sourness on sweetness, revealing that suppression of sweetness by sourness increases with the concentration of the sour stimulus (Fig. 2). Jeong et al. successfully developed an ultrasensitive artificial

В Liquid sample Working Reference electrode Taste substance (Modified material) Ag/ AgCL electrode Sensor KCl/Sat.AgCl membrane (inner solution) Taste sensing system Lipid/polymer membrane

Fig. 1 (A) Taste sensor (TS-5000Z) system, configuration, and lipid/polymer membrane. (B) TS-5000Z sensor response versus five basic tastes and astringency substances. Reprinted with permission from ref. 12, copyright 2023, Elsevier.

sweet taste bioelectronic tongue based on the T1R2 VFT of a human sweet taste receptor, which detects substances down to a limit of 0.1 fM and selectively discriminates sweet substances from other taste substances (Fig. 3).

Recent advancements in technology, materials science, computational biology, and artificial intelligence have enabled the incorporation of novel receptors and material structures into taste sensors. These include metallic nanostructure-based plasmonic sensors, 24,25 voltammetric electronic tongue with machine learning algorithms, 26 and electrochemical sensors.^{27,28} These advancements have significantly enhanced the performance of electronic and bioelectronic tongues, thereby offering more accurate and robust measurements.

Current challenges

Taste sensors are still considered a research area and are not widely employed in commercial applications. They have been used in laboratory settings to analyze a wide range of samples, including foods and beverages and medical and

Taste Nerves Human Gustatory System Biohybrid tongue

Fig. 2 Biohybrid tongue. (A) Human gustatory system. (B) Gustatory epithelium functionalized microelectrode array chip. Reprinted with permission from ref. 19, copyright 2022, American Chemical Society.

environmental samples. They have also been used for quality control in the food and beverage industry, as well as in medical diagnostics.

However, they have difficulties in detecting and selectively identifying non-charged taste substances such as glucose and caffeine, lack applicability for detecting nutrients, drugs, metabolites, and heavy ions at the same time, and have limitations in the design of bioreceptors for multiplex sensing.

There is still a need for further development and validation to make them more practical for commercial use. Scientists are still working on developing and improving electronic and bioelectronic tongues with the goal of making them more accurate, reliable, and practical for commercial use. This includes developing new sensors, improving signal processing and data analysis methods, and reducing the costs of the devices. The main challenges in the development and practical use of electronic and bioelectronic tongues include the following:

· Require more receptors or sensing materials for multianalyte measurement.




Fig. 3 Bioelectronic tongues. (A) Sweet taste receptor-based CNT-FET tongue. (B) Real-time sensor response corresponding to different concentrations of sucrose solutions. Reprinted with permission from ref. 23, copyright 2022, American Chemical Society.

Table 1 Main chemical compounds and properties of the five basic tastes

Taste	Chemical compounds	Properties
Sweet	Glucose, fructose, sucrose	Source of natural sugar and vitamins
Sour	Citric acid, lactic acid, acetic acid	Source of vitamin C
Salty	Sodium, potassium, other minerals	Maintain electrolyte balance
Bitter	Caffeine, quinine, alkaloids	Noxious compounds but some have beneficial effects
Umami	Amino acids, glutamate, aspartate	Provide essential amino acids and other nutrients

- Improve the accuracy and reliability with machine learning algorithms and data analysis approaches.
- · Reduce the cost of the sensor system or device equipment for commercialization.
- Simplify the sample preparation process and operation of the sensor system.
- · Integrate multiple sensors or obtain a more precise and comprehensive view.
 - Conduct standardization when using the taste sensors.

Sensing materials and receptors play a critical role in the development of advanced taste sensors. In recent years, there have been several widely used sensing materials or receptors for sensing taste chemical compounds in electronic bioelectronic tongues. These include ion-selective membranes, metal oxides, artificial lipid polymer membranes, and bioreceptors, including enzymes, antibodies, nucleic acids, and cells. For example, cell-based bioelectronic tongues can be highly selective but are also more expensive to produce and maintain, require more time to prepare, and have limited selectivity and specificity in detecting certain chemical compounds compared to other types of sensors.

As is well-known, thousands of chemical compounds contribute to taste, including sweet, sour, salty, bitter, and umami tastes. Table 1 shows a summary of the main substances and properties of the five basic tastes. Specific compounds that contribute to each taste can vary depending on the food or beverage being tasted. Measuring certain taste substances, such as noncharged glucose and caffeine, using potentiometric electronic tongues can be challenging. Therefore, further research is needed to develop new taste receptors or sensing materials coupled with nanotechnology and microfabrication techniques for advanced taste sensor development. For example, conductive polymers, organic and inorganic materials, nanomaterials, and artificial receptor molecularly imprinted polymers (MIP) have been widely developed for electronic tongue development.^{29–31} These sensing materials can be integrated into electronic devices such as field-effect transistors (FET), quartz crystal microbalances (QCM), surface plasmon resonance sensors (SPR), or cantilever sensors through an array-based approach and machine learning algorithms to detect and identify the target chemical compounds with high-throughput, simple, and low-cost measurements, overcoming the current challenges.³² Additionally, nanomaterials such as carbon nanotubes can be combined with bioreceptors, enzymes, antibodies, cells, or other biological molecules to develop more sensitive and specific taste sensors.

For bioelectronic tongue development, small receptors like peptides or aptamers over antibodies, cells, or enzymes have several advantages such as the following: they can be engineered to have specific binding properties and are versatile and easier to synthesize with low cost, which makes them ideal for use in biosensors and diagnostic assays. To better understand taste receptors for taste substance perception, in silico investigation is required. AlphaFold2 is a novel computational method that utilizes deep learning and neural networks to predict the structures of proteins from the amino acid sequences.33-35 It is a valuable tool for researchers in the fields of structural biology and bioinformatics because it eliminates the need for costly and time-consuming experimental methods to determine the protein structure. As shown in Fig. 4A and B, a new bitter taste receptor, TAS2R20, which is a G protein-coupled receptor (GPCR), was predicted and analyzed by molecular docking.36 Three typical bitter chemicals - quinine, humulone, and caffeine - interact with TAS2R20 with different binding affinities, and the binding residues are shown in Fig. 4C. This indicates how bitter molecules interact with taste proteins. Additionally, the binding motif was predicted, which can aid in the future design of short peptides for binding bitter substances using artificial neural networks (Fig. 4D and E).

Many studies have used computational algorithms to predict the structure and binding properties of peptides, including






Fig. 4 AlphaFold2 prediction of Homo sapiens TAS2R20 (taste receptor type 2 member 20) computational molecular docking simulation and design for bitter compound perception. (A) Typical bitter compounds. (B) TAS2R20 structure and molecular docking with bitter compounds. (C) Binding properties of a receptor-ligand complex. (D) Predicted binding motif via multiple Em for motif elicitation. (E) Molecular docking of a binding motif-ligand complex.

computational design of odorant-binding peptides or rational design of binding peptides from odorant-binding proteins (OBPs),37-40 which can be a very efficient way to identify potential candidates for further experimentation. In our previous study, we rationally designed a small receptor peptide from an anti-TNT monoclonal antibody in the complementarity determination region (CDR) (Fig. 5A and B), which is the antigen-binding region. Based on the molecular docking simulation and array-based fluorescence measurements (Fig. 5C), 15-mer amino acid peptides were selected as the TNTbinding peptides. 41,42 The small receptor exhibited high affinity $(K_d = 10.14 \text{ nM})$ for targeting TNT molecules and excellent selectivity. 43 These studies revealed a novel approach for the development of small receptors.

All-in-one terahertz taste sensor

Terahertz (THz) waves are electromagnetic radiation that fall between the microwave and infrared frequencies in the








Fig. 5 Rational design of a TNT binding peptide from an anti-TNT monoclonal antibody. (A) Structure of the TNT antibody, CDR and designed binding peptide. Reprinted with permission from ref. 42, copyright 2018, Elsevier. (B) Homology modelling and amino acid sequence analysis. (C) Experimental fluorescence results of the arraybased peptide measurements (left), docking simulation (middle), and surface plasmon resonance kinetic analysis (right). Reprinted with permission from ref. 41 and 43, copyright 2017, American Chemical Society. Copyright 2019, Elsevier.

electromagnetic spectrum. They have a frequency range of approximately 0.1 to 10 THz. THz waves have unique properties that make them useful for a variety of applications in different fields, such as spectroscopy, imaging, sensing, communications, materials science, etc. 44 Terahertz technology has been considered as a potential method for taste sensing, environmental monitoring, and medical diagnosis, as it offers a non-invasive, highly sensitive, selective, and label-free method for detecting chemical compounds and identifying their concentrations in a sample.

Our group developed an advanced terahertz chemical microscope (TCM). The TCM utilizes a semiconductor device, known as the sensing plate, to visualize the electrical potential distribution. 45-49 The sensing plate was constructed by depositing a thin silicon film on a sapphire substrate and then adding a silicon oxide layer on the top (Fig. 6A). The thickness of the silicon film was 500 nm and that of the silicon oxide film was several nanometers (Fig. 6B and C). The sensing plate was 10 mm square for easy handling; however, it can be fabricated to wafer size depending on the sensing area that needs to be measured. The defects near the boundary between the silicon and silicon oxide films caused the energy band to bend towards the boundary surface, creating a depletion layer electric field. The direction of the bend depended on the doping type of the silicon film. When a femtosecond laser pulse with photon energy equal to or greater than the silicon bandgap was applied to the sensor plate from the substrate side, the charge carriers within the silicon film were excited and accelerated by the depletion layer electric field. As shown in eqn (1), the amplitude intensity of the generated terahertz wave (E_{THz}) is proportional to the square root of the electric potential $(\sqrt{\phi})$. In this equation, E_1 represents the depletion layer, e represents the electric elementary charge, N represents the electric charge density per unit area, and ε_0 denotes the dielectric constant of vacuum. Therefore, various types of chemical reactions occurring on the sensing plate can be measured.

$$E_{\mathrm{THz}}(t) \propto E_{\mathrm{l}} = \sqrt{(2eN/\varepsilon_0)} \times \sqrt{\phi}$$
 (1)




Fig. 6 (A) Schematic of the energy band of the terahertz sensing plate. (B) Components of the terahertz sensing plate. (C) A photograph of the plastic holder for microwell formation and the sensing plate. Reprinted with permission from ref. 53, copyright 2023, Elsevier.

To address the challenges, an innovative solution - an allin-one terahertz taste sensor - has been proposed. This terahertz taste sensor can detect and quantify a diverse range of chemicals and biochemicals using multi-well arrays on a single semiconductor chip (sensing plate). It can visualize chemical and biochemical reactions with minimal sample volume consumption by using a terahertz chemical microscope system. This technology has shown a wide range of potential applications (Fig. 7A and B), including measuring physiological ions, the solution pH, non-charged molecules (Fig. 7C and D), sugars (Fig. 7E), toxic chemicals and drugs (Fig. 7F), enzyme kinetics, detecting cancer cells, analyzing the binding of antibodies and antigens in immunoassays, evaluating cosmetic products and




Fig. 7 (A) Multifunctional terahertz chemical microscope in a variety of applications. Reprinted with permission from ref. 54. Copyright 2022 MDPI (Basel, Switzerland). (B) Sensing plate functionalization methods. Reprinted with permission from ref. 54. Copyright 2021 MDPI (Basel, Switzerland). (C) Visualization of pH, Na⁺, K⁺, and Ca²⁺. Reprinted with permission from ref. 45 and 54. Copyright 2018 Optical Society of America and SPIE. Reprinted with permission from ref. 53. Copyright 2023, Elsevier. (D) DNA aptamer-based terahertz chip for non-charge cortisol detection with high sensitivity. Reprinted with permission from ref. 53, copyright 2023, Elsevier. (E) Sensitive detection of the sugar compound mannose. Reprinted with permission from ref. 45. Copyright 2019 MDPI (Basel, Switzerland). (F) TNT explosive (left) and toxic histamine detection (right). Reprinted with permission from ref. 53. Copyright 2023, Elsevier and reprinted with permission from ref. 55. Copyright 2022 MDPI (Basel, Switzerland).

batteries. 50–54 The terahertz sensing plate functionalized with ion-sensitive membranes, DNA aptamers, organic molecules, proteins, and antibodies. 53,55 measuring and visualizing the distribution electrochemical potential on the semiconductor sensing plate, the TCM can provide detailed information about chemical reactions. Furthermore, the TCM is a label-free, highly sensitive technology that can detect chemical and physical changes in the terahertz sensing plate. Unlike other established techniques such as FETs, SPR, QCMs, lightaddressable potentiometric sensors (LAPS), electrochemical sensors, the TCM does not have the limitations of Debye length design, molecular mass weight and charged molecule requirements or three electrodes requirements.⁵⁶⁻⁵⁹ For example, non-charged toxic and drug compounds have been successfully measured by simply modifying the semiconductor chip with a polymer and a DNA aptamer based on charge-transfer complex formation and conformational changes of the negatively charged DNA structure.⁵³ Moreover, as previously discussed, development of a wide range of bioreceptors and sensing materials through in silico or rational design, combined with the all-in-one terahertz taste sensor, could accelerate the development of multi-receptor arrays for detecting various types of chemicals.

However, the all-in-one terahertz taste sensor has some limitations or drawbacks, such as long-term stability problems of the laser power output and the non-uniform initial potential of the terahertz semiconductor chip caused by the manufacturing process. It requires specialized equipment and trained personnel to operate, which may limit its use under certain conditions. Therefore, future research and development could focus on maintaining the stability of the laser power output, improving the manufacturing process of the terahertz semiconductor chip, reducing the cost of the optical setup to make it more accessible, and integrating the sensor for ease of use. In addition, the development of advanced data processing or machine learning algorithms is achieve more reliable and accurate necessary to measurements.

In summary, the all-in-one terahertz taste sensor shows great potential for both academic research and industrial applications, and its non-invasive, label-free, and multi-analyte sensing approach will be a major advancement in taste sensing technology. It also has potential for use in environmental monitoring, nutrition evaluation, metabolite and drug measurement, biomarker sensing, and even for odorant identification, which could provide a comprehensive view and status and contribute to the future Food Data Bank (FDB).

Author contributions

J. W. wrote the manuscript. K. S. and T. K. reviewed the manuscript. All the authors provided permission for submission.

Conflicts of interest

There are no conflicts to declare.

Acknowledgements

This work was partially supported by a JSPS KAKENHI Grant-in-Aid for Early-Career Scientists (grant number: 21K14169).

References

- 1 L. Moreno, A. Merlos, N. Abramova, C. Jiménez and A. Bratov, Sens. Actuators, B, 2006, 116, 130-134.
- 2 Y. Kojima and Y. Hasegawa, AIP Conf. Proc., 2011, 1362, 195–196.
- 3 H.-M. Jeong, H.-C. Kwon, B. Xu, D. Jung, M. Han, D.-H. Kwon and S.-W. Kang, *Sens. Actuators, B*, 2020, **308**, 127661.
- 4 K. Toko, Mater. Sci. Eng., C, 1996, 4, 69-82.
- 5 K. Hayashi, M. Yamanaka, K. Toko and K. Yamafuji, Sens. Actuators, B, 1990, 2, 205-213.
- 6 Y. Kobayashi, M. Habara, H. Ikezazki, R. Chen, Y. Naito and K. Toko, Sensors, 2010, 10, 3411–3443.
- 7 Y. Tahara, A. Ikeda, Y. Maehara, M. Habara and K. Toko, Sensors, 2011, 11, 9878–9886.
- 8 J. Yoshimatsu, K. Toko, Y. Tahara, M. Ishida, M. Habara, H. Ikezaki, H. Kojima, S. Ikegami, M. Yoshida and T. Uchida, *Sensors*, 2020, **20**, 1–13.
- 9 A. Legin, A. Rudnitskaya, Y. G. Vlasov, C. Di Natale, F. Davide and A. D'Amico, Sens. Actuators, B, 1997, 44, 291–296.
- 10 L. A. Dias, A. M. Peres, A. C. A. Veloso, F. S. Reis, M. Vilas-Boas and A. A. S. C. Machado, *Sens. Actuators, B*, 2009, 136, 209–217.
- 11 X. Wu, Y. Tahara, R. Yatabe and K. Toko, *Anal. Sci.*, 2020, **36**, 147–159.
- 12 X. Wu and K. Toko, TrAC, Trends Anal. Chem., 2023, 158, 116874.
- 13 G. Valdés-Ramírez, M. Gutiérrez, M. del Valle, M. T. Ramírez-Silva, D. Fournier and J. L. Marty, *Biosens. Bioelectron.*, 2009, 24, 1103–1108.
- 14 M. Gutiérrez, S. Alegret and M. del Valle, *Biosens. Bioelectron.*, 2008, 23, 795–802.
- 15 M. Gutiérrez, S. Alegret and M. del Valle, *Biosens. Bioelectron.*, 2007, 22, 2171–2178.
- 16 J. Zeravik, A. Hlavacek, K. Lacina and P. Skládal, *Electroanalysis*, 2009, **21**, 2509–2520.
- 17 D. Ha, Q. Sun, K. Su, H. Wan, H. Li, N. Xu, F. Sun, L. Zhuang, N. Hu and P. Wang, Sens. Actuators, B, 2015, 207, 1136–1146.
- 18 Z. Qin, B. Zhang, L. Hu, L. Zhuang, N. Hu and P. Wang, Biosens. Bioelectron., 2016, 78, 374–380.
- 19 C. Qin, C. Chen, Q. Yuan, N. Jiang, M. Liu, Y. Duan, H. Wan, R. Li, L. Zhuang and P. Wang, *Anal. Chem.*, 2022, 94(19), 6976–6985.
- 20 Y. Hou, M. Genua, D. Tada Batista, R. Calemczuk, A. Buhot, P. Fornarelli, J. Koubachi, D. Bonnaffé, E. Saesen, C. Laguri, H. Lortat-Jacob and T. Livache, *Angew. Chem., Int. Ed.*, 2012, 51, 10394–10398.

- 21 H. S. Song, O. S. Kwon, S. H. Lee, S. J. Park, U. K. Kim, J. Jang and T. H. Park, Nano Lett., 2013, 13, 172-178.
- 22 H. S. Song, H. J. Jin, S. R. Ahn, D. Kim, S. H. Lee, U. K. Kim, C. T. Simons, S. Hong and T. H. Park, ACS Nano, 2014, 8, 9781-9789.
- 23 J. Y. Jeong, Y. K. Cha, S. R. Ahn, J. Shin, Y. Choi, T. H. Park and S. Hong, ACS Appl. Mater. Interfaces, 2022, 14, 2478-2487.
- 24 G. Macias, J. R. Sperling, W. J. Peveler, G. A. Burley, S. L. Neale and A. W. Clark, Nanoscale, 2019, 11, 15216-15223.
- S. Forest, T. Theórêt, J. Coutu and J. F. Masson, Anal. Methods, 2020, 12, 2460-2468.
- 26 Z. Yang, N. Miao, X. Zhang, Q. Li, Z. Wang, C. Li, X. Sun and Y. Lan, Food Control, 2021, 121, 107608.
- 27 Z. Chen, Q. Zhang, J. Shan, Y. Lu and Q. Liu, ACS Omega, 2020, 5, 27536-27545.
- 28 D. C. Braz, M. P. Neto, F. M. Shimizu, A. C. Sá, R. S. Lima, A. L. Gobbi, M. E. Melendez, L. M. R. B. Arantes, A. L. Carvalho, F. V. Paulovich and O. N. Oliveira Jr, Talanta, 2022, 243, 123327.
- 29 H. Sun, Z. H. Mo, J. T. S. Choy, D. R. Zhu and Y. S. Fung, Sens. Actuators, B, 2008, 131, 148-158.
- 30 A. Herrera-Chacon, A. González-Calabuig, I. Campos and M. del Valle, Sens. Actuators, B, 2018, 258, 665-671.
- 31 M. Podrazka, E. Báczyńska, M. Kundys, P. S. Jeleń and E. W. Nery, Biosensors, 2017, 8, 1-24.
- 32 M. Wang, X. Cetó and M. del Valle, Biosens. Bioelectron., 2022, 198, 113807.
- 33 M. Varadi, S. Anyango, M. Deshpande, S. Nair, C. Natassia, G. Yordanova, D. Yuan, O. Stroe, G. Wood, A. Laydon, A. Zídek, T. Green, K. Tunyasuvunakool, S. Petersen, J. Jumper, E. Clancy, R. Green, A. Vora, M. Lutfi, M. Figurnov, A. Cowie, N. Hobbs, P. Kohli, G. Kleywegt, E. Birney, D. Hassabis and S. Velankar, Nucleic Acids Res., 2022, 50, D439-D444.
- 34 J. Jumper, R. Evans, A. Pritzel, T. Green, M. Figurnov, O. Ronneberger, K. Tunyasuvunakool, R. Bates, A. Žídek, A. Potapenko, A. Bridgland, C. Meyer, S. A. A. Kohl, A. J. Ballard, A. Cowie, B. Romera-Paredes, S. Nikolov, R. Jain, J. Adler, T. Back, S. Petersen, D. Reiman, E. Clancy, M. Zielinski, M. Steinegger, M. Pacholska, T. Berghammer, S. Bodenstein, D. Silver, O. Vinyals, A. W. Senior, K. Kavukcuoglu, P. Kohli and D. Hassabis, Nature, 2021, 596, 583-589.
- 35 K. Tunyasuvunakool, J. Adler, Z. Wu, T. Green, M. Zielinski, A. Žídek, A. Bridgland, A. Cowie, C. Meyer, A. Laydon, S. Velankar, G. J. Kleywegt, A. Bateman, R. Evans, A. Pritzel, M. Figurnov, O. Ronneberger, R. Bates, S. A. A. Kohl, A. Potapenko, A. J. Ballard, B. Romera-Paredes, S. Nikolov, R. Jain, E. Clancy, D. Reiman, S. Petersen, A. W. Senior, K.

- Kavukcuoglu, E. Birney, P. Kohli, J. Jumper and D. Hassabis, Nature, 2021, 596, 590-596.
- 36 M. Mirdita, K. Schütze, Y. Moriwaki, L. Heo, S. Ovchinnikov and M. Steinegger, Nat. Methods, 2022, 19, 679-682.
- T. Wasilewski, B. Szulczyński, M. Wojciechowski, W. Kamysz and J. Gebicki, Sensors, 2019, 19(19), 4284.
- 38 A. J. M. Barbosa, A. R. Oliveira and A. C. A. Roque, Trends Biotechnol., 2018, 36, 1244-1258.
- 39 J. Wang, K. Sakai and T. Kiwa, Molecules, 2022, 27, 3917.
- 40 Y. Lu and Q. Liu, Sens. Diagn., 2022, 1, 1126-1142.
- 41 M. Okochi, M. Muto, K. Yanai, M. Tanaka, T. Onodera, J. Wang, H. Ueda and K. Toko, ACS Comb. Sci., 2017, 19, 625-632.
- 42 J. Wang, M. Muto, R. Yatabe, Y. Tahara, T. Onodera, M. Tanaka, M. Okochi and K. Toko, Sens. Actuators, B, 2018, 264, 279-284.
- 43 T. Komikawa, M. Tanaka, K. Yanai, B. R. G. Johnson, K. Critchley, T. Onodera, S. D. Evans, K. Toko and M. Okochi, Biosens. Bioelectron., 2020, 153, 112030.
- 44 M. Tonouchi, Nat. Photonics, 2007, 1, 97-105.
- 45 T. Kiwa, T. Kamiya, T. Morimoto, K. Fujiwara, Y. Maeno, Y. Akiwa, M. Iida, T. Kuroda, K. Sakai, H. Nose, M. Kobayashi and K. Tsukada, Photonics, 2019, 6(1), 10.
- 46 H. Murakami, N. Uchida, R. Inoue, S. Kim, T. Kiwa and M. Tonouchi, Proc. IEEE, 2007, 95, 1646-1657.
- K. Akimune, Y. Okawa, K. Sakai, T. Kiwa and K. Tsukada, Appl. Phys. Express, 2014, 7, 122401.
- 48 T. Kiwa, T. Kamiya, T. Morimoto, K. Sakai and K. Tsukada, Opt. Express, 2018, 26, 8232.
- 49 T. Kiwa, A. Tenma, S. Takahashi, K. Sakai and K. Tsukada, Sens. Actuators, B, 2013, 187, 8-11.
- 50 T. Kiwa, Y. Kondo, Y. Minami, I. Kawayama, M. Tonouchi and K. Tsukada, Appl. Phys. Lett., 2010, 96, 1-4.
- 51 F. Ahmed, A. Mahana, K. Taniizumi, J. Wang, K. Sakai and T. Kiwa, Jpn. J. Appl. Phys., 2021, 60, 027003.
- 52 M. Shimizu, R. Yamanaka, T. Teranishi, J. Wang, K. Sakai, K. Tsukada and T. Kiwa, Electr. Eng. Japan, 2021, 214, 1-7.
- 53 J. Wang, M. Ando, H. Nagata, S. Yoshida, K. Sakai and T. Kiwa, Biosens. Bioelectron., 2023, 220, 114901.
- 54 T. Kiwa, K. Sakai and K. Tsukada, SPIE Newsroom, 2013, 2-5.
- 55 J. Wang, K. Sato, Y. Yoshida, K. Sakai and T. Kiwa, Photonics, 2022, 9(1), 26.
- 56 M. del Valle, Electroanalysis, 2010, 22, 1539-1555.
- P. Vahdatiyekta, M. Zniber, J. Bobacka and T. P. Huynh, Anal. Chim. Acta, 2022, 1221, 340114.
- 58 L. Du, J. Wang, W. Chen, L. Zhao, C. Wu and P. Wang, Anal. Chim. Acta, 2018, 1022, 106-112.
- 59 T. Wagner, W. Vornholt, C. F. Werner, T. Yoshinobu, K. I. Miyamoto, M. Keusgen and M. J. Schöning, Phys. Med., 2016, 1, 2-7.