Analyst



PAPER View Article Online
View Journal | View Issue



Cite this: Analyst, 2024, 149, 5832

Rapid lung ventilation MRI using parahydrogeninduced polarization of propane gas†

Md Raduanul H. Chowdhury, ^a Clementinah Oladun, ^a Nuwandi M. Ariyasingha, ^a Anna Samoilenko, ^a Tarek Bawardi, ^a Dudari B. Burueva, ^b Oleg G. Salnikov, ^b Larisa M. Kovtunova, ^{b,c} Valerii I. Bukhtiyarov, ^b ^c Zhongjie Shi, ^d Kehuan Luo, ^d Sidhartha Tan, ^d Juri G. Gelovani, ^{a,e} Igor V. Koptyug, ^b Boyd M. Goodson f and Eduard Y. Chekmenev ^b *^a

Proton-hyperpolarized contrast agents are attractive because they can be imaged on virtually any clinical MRI scanner, which is typically equipped to scan only protons rather than heteronuclei (*i.e.*, anything besides protons, *e.g.*, 13 C, 15 N, 129 Xe, 23 Na, *etc.*). Even though the lifetime of the proton spin hyperpolarization is only a few seconds, it is sufficient for inhalation and scanning of proton-hyperpolarized gas media. We demonstrate the utility of producing hyperpolarized propane gas *via* heterogeneous parahydrogen-induced polarization for the purpose of ventilation imaging in an excised rabbit lung model. The magnetization of protons in hyperpolarized propane gas is similar to that of tissue water protons, making it possible to rapidly perform lung ventilation imaging with a 0.35 T clinical MRI scanner. Here, we demonstrate the feasibility of rapid (2 s) lung ventilation MRI in excised rabbit lungs using hyperpolarized propane gas with a 1 \times 1 mm² pixel size using a 50 mm slice thickness, and a 1.7 \times 1.7 mm² pixel size using a 9 mm slice thickness.

Received 29th July 2024, Accepted 22nd October 2024 DOI: 10.1039/d4an01029a

rsc.li/analyst

Introduction

In the presence of a magnetic field, nuclear spins gain population differences among their Zeeman energy levels, giving rise to a net magnetization of a given nuclear spin ensemble. The signal detected in NMR and its sister technique MRI is directly proportional to the net magnetization. In turn, the nuclear spin magnetization is directly proportional to the concentration of nuclear spins and the population difference of adjacent energy levels (termed spin polarization P). Therefore, the NMR signal scales linearly with P. Although P scales linearly with the applied magnetic field B_0 (main static magnetic

field for the NMR magnet or MRI scanner) under conditions

NMR hyperpolarization temporarily increases P by several orders of magnitude (in some cases order unity³⁻⁵), rendering correspondingly massive gains in spin magnetization and MRI signals.^{2,6-8} Hyperpolarized (HP) MRI is emerging as a safe molecular imaging technique because it relies on harmless, non-ionizing radiofrequency (RF) waves for detection. 9-14 Indeed, the first HP MRI contrast agent, hyperpolarized ¹²⁹Xe gas, was recently approved by the FDA for lung ventilation imaging to diagnose a wide range of pulmonary diseases. The resonance frequency of 129Xe gas is approximately 3.6 times lower than that of proton spins; thus, 129Xe lung images are naturally free from tissue background signals, making image interpretation streamlined and straightforward. 14-20 This important feature of HP 129Xe represents a double-edged sword because the dominating majority (over 99.5%) of clinical MRI scanners are set up to record images of protons only. Since clinical MRI scanners employ narrow-band electronics, they cannot readily perform a HP 129Xe MRI scan because 129Xe resonates at a vastly different (~3.6 times lower) frequency.

of vanishingly small P values, nuclear spin polarization remains low even within the strong magnetic fields of clinical MRI scanners (e.g., $P = 10^{-5}$ at 3 T for 1 H nuclei). As a result of this fundamental limitation, tissues become only slightly magnetized, which restricts most clinical MR imaging to water (and fat) protons with a high overall concentration of ~ 80 M.

^aDepartment of Chemistry, Integrative Biosciences (Ibio), Wayne State University, Karmanos Cancer Institute (KCI), Detroit, Michigan 48202, USA. E-mail: chekmenevlab@gmail.com

^bInternational Tomography Center, SB RAS, 3A Institutskaya St., Novosibirsk 630090. Russia

^cBoreskov Institute of Catalysis SB RAS, 5 Acad. Lavrentiev Pr., Novosibirsk 630090, Russia

^dDepartment of Pediatrics, Wayne State University, Detroit, Michigan 48202, USA ^eUnited Arab Emirates University, Al Ain, United Arab Emirates

^fSchool of Chemical & Biomolecular Sciences, Materials Technology Center, Southern Illinois University, Carbondale, IL 62901, USA

[†]Electronic supplementary information (ESI) available: Additional MRI images and additional experimental details (file type, PDF). See DOI: https://doi.org/10.1039/d4an01029a

Analyst Paper

Although, in principle, a clinical MRI scanner can be upgraded with a multi-nuclear capability for HP 129Xe scanning, such upgrades are complex and expensive (over \$0.5 M). Moreover, while all the leading vendors of clinical MRI scanners (i.e., Siemens, GE Healthcare and Philips) offer multi-nuclear capability, it is typically limited to their 3.0 T MRI scanners, which is a downside since 3.0 T MRI scanners are inherently more expensive compared to lower-field MRI scanners (furthermore, the owners of most lower-field MRI scanners are effectively locked out of the opportunity for multi-nuclear capability). Moreover, the production of HP 129Xe is slow (it can take ~0.5 hours per dose) and expensive (the hyperpolarizer device cost is >\$0.3 M). 21-28 Although continuous-flow Xe hyperpolarizers perform the process of 129Xe hyperpolarization continuously, the continuous-flow production process operates with a lean Xe mixture (typically 1-2% with the remaining fraction of buffering He gas).21 HP Xe is cryogenically frozen in such systems during the production over the course 30-60 minutes.²¹ So, although such a hyperpolarization process employs continuous flow, a batch of HP Xe is actually produced via the process of thawing of frozen Xe (following 30-60 minutes of accumulation) to maximize the concentration of HP 129Xe in the produced gas contrast agent and the resulting MRI signal after in vivo administration.²¹ These limitations of HP ¹²⁹Xe pose substantial translational challenges for making this life-saving technology broadly accessible.

One way to address this limitation is to deploy proton-HP contrast agents.^{29,30} For example, HP water was proposed.³¹ However, proton T_1 is typically 1–2 seconds in vivo, making the detection window (usually defined as several T_1 values, where T_1 is the exponential decay constant) insufficiently small compared to that of blood circulation (over 20 s); therefore, making the practical utility of this otherwise simple and streamlined technology possible is challenging in the clinical context.

However, unlike an intravenous injection of an HP contrast agent, requiring blood circulation for delivery to an organ of interest, inhalation of the HP contrast agent into the lungs can be accomplished quickly, i.e., on a time scale comparable to proton T_1 or faster. ^{32,33} Indeed, the feasibility of production of proton-hyperpolarized gases has been demonstrated for MRI applications in catalysis. 30,34 Moreover, the more recent feasibility demonstrations of clinical-scale production of proton-HP gases have rekindled interest in exploring their clinical utility as gas contrast media.^{7,8,33,35-41} In this paper, we employed proton-HP propane gas to demonstrate that high-resolution (1 × 1 mm² pixel size using 50 mm slice thickness) rapid ventilation lung MRI is feasible on a low-field clinical scanner.

Materials and methods

The HP propane gas employed in this study was produced via parahydrogen-induced polarization (PHIP) technique 42,43 (Fig. 1a). Briefly, parahydrogen (p-H₂, over 95% para-state;⁴⁴ 99.999% purity, Airgas) was mixed with propylene gas (>99% purity, Airgas, PP CP35) in a 1:1 ratio inside a high-pressure

0.7-liter aluminum storage tank. The produced mixture (10 bar pressure) was directed through a high-flow (0.15 standard liter per second (sLs) flow rate) reactor, containing a heterogeneous^{45,46} Rh/TiO₂ catalyst (0.2 g, 0.9 wt% Rh content⁴⁷), placed inside a 1/4" outer-diameter (OD) copper tubing (Fig. 1b). The reactant gas mixture was directed via Teflon tubing (1/8" OD, 1/16" inner diameter) through the catalytic reactor (T > 100 °C as a result of exothermic reaction), Fig. 1c. The temperature of the reactor was achieved by a combination of self-heating through the process of the exothermic reactor or applying heat to the reactor using a heat gun or an optional heater (both approaches yielded similar results in our studies). The reactor resided outside the main field of the MRI scanner at an estimated magnetic field below 1 mT. Pairwise p-H₂ addition to propylene results in the symmetry breaking of nascent p-H2 hydrogens, thus, converting the singlet spin order to the HP states of HA and HB sites (denoted by emphasizing colors in Fig. 1a). The estimated $P_{\rm H}$ of HP protons $H_{\rm A}$ and H_B in the produced propane gas was 0.5-1% at an estimated 40 mM gas-phase concentration of produced HP propane at 1 atm. 36 At 1% proton polarization, the anticipated polarization and signal enhancement is 8000 fold at 0.35 T, making it an equivalent to a 320 M thermally polarized proton species. Thus, the potential contrasting signal from HP gas $(P_{\rm H}$ = 1%) is anticipated to surpass that of water protons (ca. 80 M proton concentration in vivo) by several fold. In practice, the maximum attainable HP gas signal in the experiments reported here is attenuated by two primary effects. First, the HP state decays during the propane delivery from the reactor to the excised lungs. Second, the experiments are performed at 0.35 T, where partial cancellation of HA and HB resonances (with a chemical shift difference of 0.45 ppm (or \sim 7 Hz) and a 180° phase difference with respect to each other 48,49) is anticipated as $J_{H_A-H_B}$ (~7.2 Hz) and the corresponding septet of H_A and triplet of HB will have substantial spectral overlap and potential signal cancelation as HA and HB have an opposite

The excised rabbit lungs were obtained from animals that were euthanized under unrelated animal studies (Wayne State University IACUC approval # IACUC-23-01-5364). 50-53,73-75 Following the lung excision from a freshly euthanized animal, the lungs were transferred to the MRI scanner facility (~15 min) and connected to the Teflon tubing line using plastic adapting connectors and a rubber band. Next, the lungs were inflated with air using a tire pump, and placed under water to confirm that no substantial gas leaks were present. "Leaky" lungs were discarded (in approximately 50% cases), while the HP gas studies proceeded with non-"leaky" lungs. Following the placement of the lungs on a wetted bedding (to ensure that lungs were maintained wet), they were placed inside a 3.6-liter empty plastic container that was fitted inside the clinical knee detection RF coil provided by the vendor (Time Medical) of the 0.35 T clinical MRI. The plastic enclosure was utilized to prevent excessive drying of the lungs' tissue during experiments. The lungs were sprayed with water approximately every 0.5 hours to maintain their integrity. Overall, it mixture

tank

coppe

Paper

pairwise parahydrogen reactor rabbit lungs f 0.35 T clinical MRI scanner a addition to propylene catalyst **HP** propane propylene setup schematic b knee MRI coil catalytic optional pressure . regulator rabbit MRI lunas Teflon plastic connector

Fig. 1 (a) Schematic of pairwise p-H₂ addition to propylene to produce HP propane gas. (b) Schematic of the experimental setup showing the gas flow path through the catalytic reactor into excised rabbit lungs connected via Teflon tubing. (c) Close-up photograph of the catalytic reactor. (d) Photograph of representative excised rabbit lungs connected to a Teflon tubing. (e) Photograph of a knee MRI coil and the excised rabbit lung before positioning them inside the RF coil. (f) Annotated photograph of the experimental setup for conducting ventilation MRI imaging of excised rabbit lungs, utilizing a clinical knee RF coil, a tank with a gas mixture, and a catalytic reactor with a Rh/TiO2 catalyst. During the experiment, the lungs reside at the homogeneous field of the MRI scanner (0.35 T), the reactor resides at an estimated field of less than 1 mT, and the mixture storage tank resides at a magnetic field of less than 0.1 mT. The MRI scanner was shimmed using a brain phantom (placed inside the brain MRI coil) provided by the vendor prior to experimentations with HP propane gas and excised rabbit lungs.

was possible to utilize a set of intact lungs for 20-40 gas injections throughout a day of experiments. At the end of the day, the lungs were placed in the refrigerator and maintained at 5 °C. It was possible to re-use the excised lungs for an additional day following the removal from the refrigerator.

Rh/TiO₂

Teflon tubing

alass wool

optional

The HP propane gas exiting from the reactor was directed to the excised rabbit lungs. Following lung inflation with HP gas (1-1.5 s) with an estimated volume of 0.1-0.2 standard liters (sL), the gas flow was stopped.

A rapid 2D gradient echo (GRE) sequence was composed of 16 repeat back-to-back scans with a total scan time 1.7 s for each scan with a 96 × 96 imaging matrix. Other MR imaging parameters used in these experiments were $100 \times 100 \text{ mm}^2$ field of view (FOV), 30° slice-selective excitation RF pulse, 10.42 kHz spectral width, 17.8 ms and 20.27 ms repetition time (TR), and 8.5 ms and 9.75 ms echo time (TE), respectively. A separate (lower-resolution) MR scanning protocol was also utilized with a reduced imaging matrix to 64 × 64, resulting in the reduction of each scan from 1.7 s to 0.94 s with corresponding values of TR of 14.7 ms and TE of 6.97 ms.

The imaging sequence was initiated a few seconds before the administration of HP gas into the lungs to ensure the HP

propane gas ventilation would be recorded immediately upon gas delivery into the lungs (with the goal of minimizing polarization losses due to T_1 decay of the HP state). The sequence employed a slice thickness of 50 mm, effectively resulting in 2D projection imaging. Additional images with smaller slice thickness (e.g., 9 mm) were also acquired albeit at lower spatial resolution. Typically, the HP signal persisted for 1-2 scans due to rapid depolarization by 30° RF excitation pulses, but longer acquisitions can be made in principle to extend the scanning time window. The imaging data were additionally post-processed via customwritten MATLAB code to perform subtraction of the background signal from the thermally polarized excised lung tissue and the surrounding wet blanket (ESI†). For background signal subtraction, a difference image was computed between the image with inflated lungs containing HP gas (usually, scan #4 to #5) and an image acquired after inflation and HP state depolarization (typically, scan #5 or #6). Additionally, the mean SNR is calculated by dividing the average intensity of signal pixel by the root mean square (RMS) value of the noise obtained from a 8 × 8 matrix in the noise region (see the ESI† for examples).

Results and discussion

Analyst

Fig. 2 highlights the highest-resolution images that were obtained in this study. The excised rabbit lungs were ventilated and visualized with HP propane gas in the axial (Fig. 2a) and coronal projections (Fig. 2b). These images demonstrate good spatial resolution of lung ventilation MRI using HP propane gas. Indeed, Fig. 2a and b show high-resolution images with a pixel size of 1 × 1 mm² acquired using a field of view (FOV) of $100 \times 100 \text{ mm}^2$ and an imaging matrix size of 96×96 (with a slice thickness of 50 mm), with a mean SNR of 12-13 in the background-subtracted images. Each imaging series (Fig. 2a and b, respectively) shows the selected repeat scans over the same anatomical location: as the HP propane gas reaches the lungs, the image denoted as "inflation" exhibits the HP gas signals throughout the lungs in addition to the background signal from thermally polarized tissues. In most cases, the signal from HP gas persisted only in one image (additional data analysis for Fig. 2b is shown in Fig. S27†). However, in cases where a higher SNR was available (e.g., Fig. S13a†), the lung ventilation visualization was seen for more than one scan (in those cases, the thermally polarized background image

subtraction procedure employed the post-inflation image, where the HP state was deemed to be depolarized rather than employed the scan acquired immediately "after inflation"). The image denoted as "before inflation" corresponds to the image acquired from thermally polarized samples of the collapsed lungs (before gas injection) and the surrounding wet blanket. MRI scanning with HP propane was reproducible: the comparative repeat scans are shown in Fig. S1.† The feasibility of achieving high spatial resolution is important because better-quality images allow the detection of smaller ventilation defects, i.e., potentially enabling earlier diagnosis of lung diseases. Moreover, the recent studies with HP propane gas injected in excised pig lungs allowed achieving only a 4 × 4 mm² pixel size (using the same slice thickness and a similar MRI scanner setup) corresponding to 15-times lower spatial resolution.³⁶ The more-than-order-of-magnitude lower image resolution in the previous study is likely due to an overall slower inflation process with HP gas, rendering substantially greater depolarization losses in the previous study. Specifically, the inflation of the pig lungs with HP propane gas (0.7 sL) was performed over 5 seconds in the previous study, 36 and over 95% of HP state magnetization was likely lost due to




Fig. 2 Rapid slice-selective 2D GRE images of HP propane gas rapidly expanding in excised rabbit lungs, acquired by utilizing a 0.35 T MRI scanner and a knee RF coil. (a) Axial projection of the excised rabbit lungs recorded before inflation (scan #4), during inflation (scan #5), and after inflation (scan #6) with HP propane gas at 1.7 s temporal resolution. (b) Corresponding scans from coronal projection. The difference images for both axial and coronal projections were obtained as the difference between scan #5 and scan #6 (Fig. S22 and S23† show the screenshots of the MRI scans as visualized with a 0.35 T scanner without any additional data processing). The corresponding mean SNR values associated with each image were obtained as described in detail in the ESI.† The axial and coronal scans were acquired with a $100 \times 100 \text{ mm}^2$ FOV, a 50 mm slice thickness, a 30° RF excitation pulse, a 96×96 imaging matrix, and post-processing interpolation to 768×768 pixels. A total of 16 repeat scans were recorded in 27 s, with a repetition time (TR) of 17.8 ms and an echo time (TE) of 8.55 ms.

Paper Analyst

a fast decay constant (T_S was estimated between 1 and 2 s; T_S refers to the mono-exponential decay of long-lived spin states (LLSS), 32,33,54,55 which are known to exist at low magnetic fields for HP propane gas produced under ALTADENA conditions⁴⁹): work is in progress to perform precise mapping of HP state decay rates in excised rabbit lungs at 0.35 T. In this work, the rabbit lungs (injected gas volume was estimated at 0.1-0.2 sL) were injected in 1-1.5 s, i.e., approximately 5-7 times faster than in the previous experiments. While HP state depolarization losses have occurred in the experiments with the rabbit lungs during the gas administration too, these losses were likely at least an order of magnitude lower compared to the previous experiments with larger pig lungs;³⁶ we note that developing a more quantitative model of HP state polarization losses during gas delivery and imaging requires the quantitative knowledge of the HP state decay constant, which is not available to us at the moment (although active experiments are in progress and will be reported in the future). This fact underscores the need for high-capacity on-demand production of proton-hyperpolarized gas media that allows rapid gas delivery in the lungs to minimize depolarization losses. The depolarization losses can be potentially further mitigated using long-lived HP states at low magnetic fields^{54,55} and using larger molecules such as diethyl ether and butane.37,55,56

It should also be noted that there is a reduced background signal from thermally polarized inflated lung tissue (images acquired after the inflation and depolarization of the HP state). This observation clearly shows that the HP signal overshadows the background signal, thus making the background subtraction method more robust. We anticipate that the background signal will be stronger from the thermally polarized surrounding tissues in vivo, and the image subtraction would work adequately.

In a separate experiment, the imaging matrix was decreased to 64×64 resulting in an increased pixel size of 1.6×1.6 mm²,

using the same FOV of 100 × 100 mm² and the same slice thickness of 50 mm. This decrease of the imaging matrix size accelerated image acquisition to sub-second speed (0.94 s scan time), while yielding similar values of the mean SNR, Fig. S4† (one should not directly compare the SNR values per pixel size of images recorded with 64 × 64- and 96 × 96-imaging matrices as multiple parameters are changed at the same time including the signal acquisition time, voxel size and the effect of HP state depolarization). These results are important as they demonstrate the feasibility of performing a ventilation MRI scan with sub-second temporal resolution. Moreover, faster scan speed is also important to minimize HP propane gas polarization losses due to fast T_1 (ca. 0.8 s) and T_S (ca. 1-2 s, which have been shown to exist in HP propane gas at sufficiently low magnetic fields) at a clinically relevant pressure of 1 atm. Furthermore, the fast scan speed is also welcome from the future clinical perspective of performing the ventilation imaging scan on a single patient breath hold in a manner similar to that of the HP ¹²⁹Xe gas contrast agent. ¹⁶ Indeed, we have previously demonstrated the feasibility of recording 8 slices in 0.4 s on a HP propane gas phantom using a clinical 3 T GE Healthcare scanner³⁶ that has a substantially faster scan speed compared to the 0.35 T scanner employed for this study, demonstrating that comprehensive multi-slice 2D scanning of the lungs may be potentially feasible in less than 1 s. Sub-second ventilation imaging was also reproducible (Fig. S7†).

In a separate experiment shown in Fig. 3, the slice thickness was decreased to 9 mm versus 50 mm (Fig. 2), while the FOV was increased to $160 \times 160 \text{ mm}^2$, resulting in a pixel size of $1.7 \times 1.7 \text{ mm}^2$. The reduced slice thickness allows scanning a section of the lung (9 mm thick) versus effectively recording a projection scan with a 50 mm slice thickness. The results shown with 9 mm slice thickness are important as they demonstrate the feasibility of recording MRI scans with pixel size on a mm-scale versus effective 2D projection image scans.




Fig. 3 Rapid slice-selective 2D GRE images of HP propane gas (axial projection) rapidly expanding in excised rabbit lungs, acquired by utilizing a 0.35 T MRI scanner and a knee RF coil. Scan #2 was recorded before HP gas inflation, scan #3 was recorded during the inflation, and scan #4 was recorded after the HP gas inflation at 1.94 s temporal resolution. The difference image is obtained as the difference between scan #3 and scan #4. The corresponding mean SNR (see the ESI† for details) is provided. The scans were acquired with a $160 \times 160 \text{ mm}^2$ FOV, a 9 mm slice thickness, a 30° RF excitation pulse, a 96×96 imaging matrix, and post-processing interpolation to 768×768 pixels. A total of 16 repeat scans were recorded in 31 s, with a repetition time (TR) of 20.27 ms and an echo time (TE) of 9.75 ms. Fig. S24† shows the screenshot of the MRI scans as visualized with the 0.35 T scanner without any additional data processing.

Analyst

As expected, further decrease of the spatial resolution by increasing the pixel size to $2 \times 2 \text{ mm}^2$ (50 mm slice thickness) via expanding the FOV to 128 × 128 mm² and maintaining an imaging matrix size of 64 × 64 yielded sub-second MRI scans with a substantially improved mean SNR up to 28 in the axial projection (in the background subtraction images, Fig. S10†), which was also reproducible (with a corresponding mean SNR of 28, Fig. S13†). Further increase of the FOV and the pixel size to 160 × 160 mm² and 2.5 × 2.5 mm², respectively (while maintaining other parameters the same, i.e., matrix size of 64×64 and (50 mm slice thickness)), resulted in additional gains in the mean SNR of up to 43 (axial projection difference image in Fig. S16†). These results highlight the possibility of recording ventilation MRI scans with a high SNR over a large FOV, which is important in the context of future clinical translation of this MRI modality. Moreover, the results reported in Fig. S16a,† showing the image of the inflated lungs with a higher SNR (albeit at the expense of spatial resolution), clearly exhibited the following unexpected feature: the image recorded at inflation (SNR of 54) is followed by the image with an SNR of 26, i.e., clearly showing substantial residual magnetization left after the completion of the first MRI scan that was recorded after the gas injection was stopped (note the same overall shape of the lungs in two images, further supporting the fact that the HP gas was stopped, when they were recorded). The substantial amount of retained SNR is rather unexpected because a 30° RF excitation pulse is anticipated to depolarize the non-replenishable HP state by 10-fold in 16 excitation pulses (note that 64 RF pulses were applied in total to record that image); thus, one anticipates recording only one "goodquality" image with virtually no magnetization left after the completion of 64 slice-selective RF excitation pulses employed (Fig. S16a†). We rationalize this finding by the fact that HP HA and H_B sites are in the intermediate spin-spin coupling regime $(J_{H_A-H_R})$ is approximately equal to the chemical shift difference between HA and HB sites). As a result, long-lived spin states (LLSS) may be "partially" present at this field for HP propane gas produced via PHIP. We speculate that "partial" LLSS are mostly immune to the RF excitation pulses and may "effectively" replenish NMR-visible magnetization throughout acquisition MRI scans (and application of successive 64 RF excitation pulses), thus, resulting in the substantially lower depolarizing effect of the RF excitation pulses: however, detailed relaxation dynamics studies are certainly required to support this hypothesis.

Additional control experiments were performed with nonhyperpolarized non-hydrocarbon gas (Fig. 4). In these control experiments, we first recorded a ventilation scan with HP propane over a 128 × 128 mm² FOV and an imaging matrix size of 64 × 64: a clear visualization of the lungs is seen in the difference image in Fig. 4a, while no substantial signals are seen in the difference image in the corresponding data set when HP propane gas was replaced with N2 gas (Fig. 4b). These results are important as they clearly show that the ventilation image is obtained truly from HP propane inflating the lungs rather than from other MRI artifacts due to rapid lung expansion.

Future outlook

We envision that the obtained results could be additionally optimized in the future as follows. More SNR-efficient scanning approaches could be deployed (for example, compressed and parallel sensing, spiral k-space sampling, and echo-planar imaging³⁶) to increase the sensitivity and speed, potentially enabling multi-slice scanning in less than 1 s. 36,57 Furthermore, additional efforts could certainly be made to utilize water background suppression using already-established clinical protocols. 58,59 It should also be made clear that future in vivo studies will likely reveal a substantial thermally polarized background signal from the lung tissue and the surrounding tissues additionally compounded by the motion artifacts that are commonly present in lung imaging.⁵⁷ A number of already established approaches for background subtraction (for example, those employed in Dynamic Contrast Enhancement (DCE) MRI⁶⁰) and advanced motion correction techniques⁵⁷ have been developed to address these challenges in vivo. The required high scanning speed is a clear limitation of the proton-HP gas technology, and some MRI scanners may not be sufficiently fast, limiting the reach of this technology.

For future in vivo experiments, we envision the delivery of the gas via a mouthpiece with an overpressure exit path to ensure that the arriving gas can be either readily inhaled or safely exited without over-pressurization of the mouth, airways, and lungs. Noteworthily, portable handheld oxygen aluminum cans (e.g., Boost OxygenTM) employ such a mouthpiece and an oxygen pressure of 15 bar inside the can: no special training is required to operate such a device, which is sold without prescription in the US and other developed countries. We also envision that the exiting gas may be potentially scavenged by a carbon filter (active ongoing studies in our partnering laboratories). The can/tank that stores the mixture of unsaturated precursor and p-H₂ can be potentially pressurized to 10 bar or lower pressure as long as a sufficiently high flow of the HP gas stream is provided for HP propane gas inhalation: the key requirement is the pressure drop to atmospheric pressure at the mouthpiece for safe operation of such a device. We also envision that the can/tank, containing the mixture of the unsaturated precursor and p-H2 can be either placed inside the bore of the MRI scanner (using all-non-magnetic components and single-dose capacity) or placed in the equipment room of the MRI scanner (for a multi-dose capacity). It is also envisioned that HP propane gas can be potentially administered as a single bolus of pure HP gas or as a dilute mixture of HP propane gas with medical air for multi-inhalation protocols. Both approaches have their own merits. Noteworthily, a substantially more dense HP 129Xe gas is well tolerated in a singledose inhalation of up to 1 liter, 61,62 indicating that less dense propane can be also potentially similarly well tolerated. We anticipate no substantial difference between any change in lifetime of HP propane gas caused by dilution with oxygen, because dilution-induced changes in gas viscosity and molecular properties are anticipated to be negligible at physiologically relevant pressures and temperatures. 33,55,63,64

80

120-




Fig. 4 Rapid slice-selective 2D GRE images of HP propane gas rapidly expanding in excised rabbit lungs and the control study performed using non-hyperpolarized inert nitrogen gas, acquired by utilizing a 0.35 T MRI scanner and a knee RF coil. (a) Axial projection of the excised rabbit lungs recorded before inflation (scan #3), during inflation (scan #4), and after inflation (scan #5) with HP propane gas at a 1.1 s temporal resolution. (b) Corresponding control experiment, where HP propane gas was replaced with inert nitrogen gas that performs similar inflation, but carries no HP state in contrast to images shown in display a. The difference images for both HP propane and nitrogen gas experiments were obtained as the difference between scan #4 and scan #5. The corresponding mean SNR values associated with each image are reported. Note that the difference image for nitrogen gas inflation (control experiment) yields virtually no signal (as expected), and therefore, the SNR value is marked as N/A. The axial scans were acquired with a 128 x 128 mm² FOV, a 50 mm slice thickness, a 30° RF excitation pulse, a 64 x 64 imaging matrix, and post-processing interpolation to 768 x 768 pixels. A total of 16 repeat scans were recorded in 18 s, with TR = 17.19 ms and TE = 8.21 ms. Fig. S25 and S26† show the screenshots of the corresponding MRI scan series (shown in displays a and b, respectively) as visualized with the 0.35 T scanner without any additional data processing.

80

120

Despite the fact that the results reported here were obtained with a $P_{\rm H}$ of only 0.5-1%, while clinical studies with HP 129 Xe typically utilize 20-30% polarized 129Xe, high sensitivity of proton-hyperpolarized scanning becomes possible due to the higher gyromagnetic ratio and the greater natural abundance of protons compared to those of 129Xe: specifically, we anticipate that two HP propane gas protons will yield more than 10 times more SNR compared to HP 129Xe gas at the same nominal isotopic concentration and polarization. 58,59 The additional gains in PH are also potentially feasible through improved efficiency of pairwise p-H2 addition to propylene via a more advanced catalyst design.⁶⁵ HP propane gas could be quickly produced on demand at a low cost,³⁶ and it can be readily deployed on clinical MRI scanners without any modifications (including the low-field 0.35 T MRI scanner employed here), highlighting key advantages of this technology. On the other hand, the fast depolarization of the propane HP state demands establishing the HP agent production in close proximity to the MRI scanner, which is a clear disadvantage compared to HP 129Xe production, which can be performed far

80

x (mm)

80

away from the imaging suite. Moreover, HP 129Xe exhibits distinct in vivo chemical shifts (gas, tissue membrane, and redblood cell phases, and potentially more using the biosensor approach⁶⁶), whereas propane protons are expected to have no substantial chemical shift differences in those environments, thus, making measurements of gas perfusion in lungs16,67 using HP propane hardly possible.

80

120

We have employed 99% pure propylene and 99.999% hydrogen for these experiments, which we believe meet the purity requirements for GMP production of propane gas (typically 99%). The heterogeneous catalyst allows producing a catalyst-free stream of HP propane gas, although rigorous excipient testing of the key components (clean food-grade copper tubing, food-grade titania, rhodium metal, and brass connectors) employed in the hyperpolarizer fluid path design is required for future clinical translation. Propane gas is already regulated by the FDA and many other regulatory bodies (designated E944 food additive label in Europe), and propane is generally recognized as safe (GRAS). 68 Propane gas is approved for unlimited use in food applications as a propellant gas and as

an ingredient under GMP production. Propane gas is a nontoxic asphyxiant. Indeed, a randomized ninety-day inhalation toxicity study at 10 000 ppm gas-phase concentration revealed no observable systemic effects.⁶⁹ Propane gas was also safely administered in a concentration up to 20% for one minute. 70,71 While these previous studies serve as a good foundation that hyperpolarized propane gas can likely be safely administered for MRI scanning purposes, we note that there is currently no regulatory approval for such examination. We envision that the possible side effects of HP propane gas administration certainly need to be studied prior to clinical utilization of this potential inhalable contrast agent for functional lung imaging utilization. Moreover, careful MRI suite planning with respect to the dosing, administration, flammability, and scavenging using a carbon filter must be addressed to maintain the safety of the personnel and patients near equipment that is known to generate kV potentials and arcing.72

Conclusions

In summary, the feasibility of high-resolution ventilation proton MRI of HP propane in excised rabbit lungs has been demonstrated with high resolution (1 \times 1 \times 50 mm³ and 1.7 \times 1.7 \times 1 \times 9 mm³ voxel size) and ultrafast scan speed (1.7–1.9 s) using a clinical 0.35 T MRI scanner without any specialized hardware or pulse sequences. Moreover, the feasibility of subsecond scan speed has been demonstrated, albeit at a slightly reduced spatial resolution of a 1.6 \times 1.6 \times 50 mm³ voxel size. These results bode well for future biomedical translation of HP propane gas and other proton-hyperpolarized inhalable contrast media. 37

Ethical statement

All animal procedures were performed in accordance with the Guidelines for Care and Use of Laboratory Animals of "Wayne State University" and approved by the Animal Ethics Committee of "IACUC (23-01-5364)

Author contributions

Md Raduanul H. Chowdhury: software, data curation, formal analysis, visualization, and writing – original draft writing, review & editing; Clementinah Oladun: software, data curation, formal analysis, and visualization; Nuwandi M. Ariyasingha: software, investigation, data curation, formal analysis, and visualization; Anna Samoilenko: investigation; Tarek Bawardi: data curation, formal analysis, visualization, and writing – original draft writing, review & editing; Dudari B. Burueva: review & editing; Oleg G. Salnikov: review & editing and methodology; Larisa M. Kovtunova: review & editing; Zhongjie Shi: investigation and methodology; Kehuan Luo: investigation and

methodology; Sidhartha Tan: project administration, resources, supervision, and review & editing; Juri G. Gelovani: review & editing and funding acquisition; Igor V. Koptyug: supervision and review & editing; Boyd M. Goodson: project administration, funding acquisition, resources, and review & editing; Eduard Y. Chekmenev: conceptualization, data curation, formal analysis, funding acquisition, investigation, methodology, project administration, resources, software, supervision, validation, writing, and review & editing.

Data availability

All raw DICOM MRI scans that are presented in the main text and the ESI† can be found at https://doi.org/10.17632/dbkf9pfvkh.1.

Conflicts of interest

BMG and EYC declare stake ownership in XeUS Technologies, Ltd. EYC declares a stake of ownership in Vizma Life Sciences.

Acknowledgements

We thank Prof. Matthew S. Rosen (Harvard Medical School) and Prof. Rahim R. Rizi (University of Pennsylvania School of Medicine) for providing polling information about the number of clinical multi-nuclear MRI scanners in their institutions and the estimates for such clinical multi-nuclear scanners domestically in the United States. This work was supported by DOD CDMRP W81XWH-20-10576, W81XWH-20-10578, NSF CHE-2404387 and CHE-2404388, NHLBI R21HL154032 and F32HL160108, NINDS R01NS114972, R01NS117146, and R01NS130258. D. B. B. thanks the Russian Science Foundation (grant #23-23-00394) for financial support of the development of PHIP protocols. O. G. S. and I. V. K. thank the Russian Ministry of Science and Higher Education for access to research equipment. L. M. K. and V. I. B. thank the Ministry of Science and Higher Education of the Russian Federation (FWUR-2024-0032).

References

- 1 D. I. Hoult and R. E. Richards, *J. Magn. Reson.*, 1976, 24, 71-85.
- 2 J. Eills, D. Budker, S. Cavagnero, E. Y. Chekmenev, S. J. Elliott, S. Jannin, A. Lesage, J. Matysik, T. Meersmann, T. Prisner, J. A. Reimer, H. Yang and I. V. Koptyug, *Chem. Rev.*, 2023, 123, 1417–1551.
- 3 S. Jannin, A. Bornet, R. Melzi and G. Bodenhausen, *Chem. Phys. Lett.*, 2012, 549, 99–102.
- 4 P. J. Rayner, M. J. Burns, A. M. Olaru, P. Norcott, M. Fekete, G. G. R. Green, L. A. R. Highton, R. E. Mewis and

Paper

S. B. Duckett, *Proc. Natl. Acad. Sci. U. S. A.*, 2017, **114**, E3188–E3194.

- P. Nikolaou, A. M. Coffey, L. L. Walkup, B. M. Gust,
 N. Whiting, H. Newton, S. Barcus, I. Muradyan,
 M. Dabaghyan, G. D. Moroz, M. Rosen, S. Patz,
 M. J. Barlow, E. Y. Chekmenev and B. M. Goodson, *Proc. Natl. Acad. Sci. U. S. A.*, 2013, 110, 14150–14155.
- 6 J. Kurhanewicz, D. B. Vigneron, K. Brindle, E. Y. Chekmenev, A. Comment, C. H. Cunningham, R. J. DeBerardinis, G. G. Green, M. O. Leach, S. S. Rajan, R. R. Rizi, B. D. Ross, W. S. Warren and C. R. Malloy, *Neoplasia*, 2011, 13, 81–97.
- 7 D. A. Barskiy, A. M. Coffey, P. Nikolaou, D. M. Mikhaylov, B. M. Goodson, R. T. Branca, G. J. Lu, M. G. Shapiro, V.-V. Telkki, V. V. Zhivonitko, I. V. Koptyug, O. G. Salnikov, K. V. Kovtunov, V. I. Bukhtiyarov, M. S. Rosen, M. J. Barlow, S. Safavi, I. P. Hall, L. Schröder and E. Y. Chekmenev, Chem. Eur. J., 2017, 23, 725–751.
- 8 K. V. Kovtunov, I. V. Koptyug, M. Fekete, S. B. Duckett, T. Theis, B. Joalland and E. Y. Chekmenev, *Angew. Chem.*, *Int. Ed.*, 2020, **59**, 17788–17797.
- 9 S. J. Nelson, J. Kurhanewicz, D. B. Vigneron, P. E. Z. Larson, A. L. Harzstark, M. Ferrone, M. van Criekinge, J. W. Chang, R. Bok, I. Park, G. Reed, L. Carvajal, E. J. Small, P. Munster, V. K. Weinberg, J. H. Ardenkjaer-Larsen, A. P. Chen, R. E. Hurd, L. I. Odegardstuen, F. J. Robb, J. Tropp and J. A. Murray, Sci. Transl. Med., 2013, 5, 198ra108.
- 10 J. Kurhanewicz, D. B. Vigneron, J. H. Ardenkjaer-Larsen, J. A. Bankson, K. Brindle, C. H. Cunningham, F. A. Gallagher, K. R. Keshari, A. Kjaer, C. Laustsen, D. A. Mankoff, M. E. Merritt, S. J. Nelson, J. M. Pauly, P. Lee, S. Ronen, D. J. Tyler, S. S. Rajan, D. M. Spielman, L. Wald, X. Zhang, C. R. Malloy and R. Rizi, Neoplasia, 2019, 21, 1–16.
- 11 S. Fain, M. L. Schiebler, D. G. McCormack and G. Parraga, J. Magn. Reson. Imaging, 2010, 32, 1398–1408.
- 12 B. T. Chung, H.-Y. Chen, J. Gordon, D. Mammoli, R. Sriram, A. W. Autry, L. M. Le Page, M. M. Chaumeil, P. Shin, J. Slater, C. T. Tan, C. Suszczynski, S. Chang, Y. Li, R. A. Bok, S. M. Ronen, P. E. Z. Larson, J. Kurhanewicz and D. B. Vigneron, J. Magn. Reson., 2019, 309, 106617.
- 13 Z. J. Wang, M. A. Ohliger, P. E. Z. Larson, J. W. Gordon, R. A. Bok, J. Slater, J. E. Villanueva-Meyer, C. P. Hess, J. Kurhanewicz and D. B. Vigneron, *Radiology*, 2019, 291, 273–284.
- 14 B. Driehuys, Acad. Radiol., 2017, 24, 1-3.
- 15 L. L. Walkup and J. C. Woods, *NMR Biomed.*, 2014, 27, 1429–1438.
- 16 J. P. Mugler and T. A. Altes, *J. Magn. Reson. Imaging*, 2013, 37, 313–331.
- 17 B. M. Goodson, J. Magn. Reson., 2002, 155, 157-216.
- 18 M. S. Albert, G. D. Cates, B. Driehuys, W. Happer, B. Saam, C. S. Springer and A. Wishnia, *Nature*, 1994, 370, 199–201.

- 19 T. N. Matin, N. Rahman, A. H. Nickol, M. Chen, X. Xu, N. J. Stewart, T. Doel, V. Grau, J. M. Wild and F. V. Gleeson, *Radiology*, 2017, 282, 857–868.
- 20 J. T. Grist, M. Chen, G. J. Collier, B. Raman, G. AbuEid, A. McIntyre, V. Matthews, E. Fraser, L.-P. Ho, J. M. Wild and F. Gleeson, *Radiology*, 2021, 301, E353–E360.
- 21 A. S. Khan, R. L. Harvey, J. R. Birchall, R. K. Irwin, P. Nikolaou, G. Schrank, K. Emami, A. Dummer, M. J. Barlow, B. M. Goodson and E. Y. Chekmenev, *Angew. Chem., Int. Ed.*, 2021, 60, 22126–22147.
- 22 J. R. Birchall, M. R. H. Chowdhury, P. Nikolaou, Y. A. Chekmenev, A. Shcherbakov, M. J. Barlow, B. M. Goodson and E. Y. Chekmenev, *Molecules*, 2022, 27, 1327.
- 23 I. C. Ruset, L. L. Tsai, R. W. Mair, S. Patz, M. I. Hrovat, M. S. Rosen, I. Muradian, J. Ng, G. P. Topulos, J. P. Butler, R. L. Walsworth and F. W. Hersman, *Concepts Magn. Reson.*, Part B, 2006, 29B, 210–221.
- 24 G. Norquay, S. R. Parnell, X. J. Xu, J. Parra-Robles and J. M. Wild, *J. Appl. Phys.*, 2013, **113**, 9.
- 25 S. E. Korchak, W. Kilian and L. Mitschang, *Appl. Magn. Reson.*, 2013, 44, 65–80.
- 26 B. Driehuys, G. D. Cates, E. Miron, K. Sauer, D. K. Walter and W. Happer, *Appl. Phys. Lett.*, 1996, **69**, 1668–1670.
- 27 J. R. Birchall, R. K. Irwin, P. Nikolaou, A. M. Coffey, B. E. Kidd, M. Murphy, M. Molway, L. B. Bales, K. Ranta, M. J. Barlow, B. M. Goodson, M. S. Rosen and E. Y. Chekmenev, J. Magn. Reson., 2020, 319, 106813.
- 28 J. W. Plummer, K. Emami, A. Dummer, J. C. Woods, L. L. Walkup and Z. I. Cleveland, *J. Magn. Reson.*, 2020, 320, 106845.
- 29 L. S. Bouchard, K. V. Kovtunov, S. R. Burt, M. S. Anwar, I. V. Koptyug, R. Z. Sagdeev and A. Pines, *Angew. Chem., Int. Ed.*, 2007, 46, 4064–4068.
- 30 L. S. Bouchard, S. R. Burt, M. S. Anwar, K. V. Kovtunov, I. V. Koptyug and A. Pines, *Science*, 2008, 319, 442–445.
- 31 M. D. Lingwood, T. A. Siaw, N. Sailasuta, O. A. Abulseoud, H. R. Chan, B. D. Ross, P. Bhattacharya and S. Han, *Radiology*, 2012, **265**, 418–425.
- 32 D. A. Barskiy, O. G. Salnikov, A. S. Romanov, M. A. Feldman, A. M. Coffey, K. V. Kovtunov, I. V. Koptyug and E. Y. Chekmenev, *J. Magn. Reson.*, 2017, 276, 78–85.
- 33 N. M. Ariyasingha, O. G. Salnikov, K. V. Kovtunov, L. M. Kovtunova, V. I. Bukhtiyarov, B. M. Goodson, M. S. Rosen, I. V. Koptyug, J. G. Gelovani and E. Y. Chekmenev, J. Phys. Chem. C, 2019, 18, 11734–11744.
- 34 E. S. Kononenko, A. I. Svyatova, I. V. Skovpin, L. M. Kovtunova, E. Y. Gerasimov and I. V. Koptyug, *J. Phys. Chem. C*, 2022, 126, 14914–14921.
- 35 O. G. Salnikov, P. Nikolaou, N. M. Ariyasingha, K. V. Kovtunov, I. V. Koptyug and E. Y. Chekmenev, *Anal. Chem.*, 2019, 91, 4741–4746.
- 36 N. M. Ariyasingha, A. Samoilenko, J. R. Birchall, M. R. H. Chowdhury, O. G. Salnikov, L. M. Kovtunova, V. I. Bukhtiyarov, D. C. Zhu, C. Qian, M. Bradley,

- J. G. Gelovani, I. V. Koptyug, B. M. Goodson and E. Y. Chekmenev, *ACS Sens.*, 2023, **8**, 3845–3854.
- 37 N. M. Ariyasingha, M. R. H. Chowdhury, A. Samoilenko, O. G. Salnikov, N. V. Chukanov, L. M. Kovtunova, V. I. Bukhtiyarov, Z. Shi, K. Luo, S. Tan, I. V. Koptyug, B. M. Goodson and E. Y. Chekmenev, *Chem. Eur. J.*, 2024, 30, e202304071.
- 38 K. V. Kovtunov, A. S. Romanov, O. G. Salnikov, D. A. Barskiy, E. Y. Chekmenev and I. V. Koptyug, *Tomography*, 2016, 2, 49–55.
- 39 K. V. Kovtunov, D. A. Barskiy, A. M. Coffey, M. L. Truong, O. G. Salnikov, A. K. Khudorozhkov, E. A. Inozemtseva, I. P. Prosvirin, V. I. Bukhtiyarov, K. W. Waddell, E. Y. Chekmenev and I. V. Koptyug, *Chem. Eur. J.*, 2014, 20, 11636–11639.
- 40 B. Vuichoud, E. Canet, J. Milani, A. Bornet, D. Baudouin, L. Veyre, D. Gajan, L. Emsley, A. Lesage, C. Copéret, C. Thieuleux, G. Bodenhausen, I. V. Koptyug and S. Jannin, J. Phys. Chem. Lett., 2016, 7, 3235–3239.
- 41 B. Joalland, N. M. Ariyasingha, H. R. Younes, S. Nantogma, O. G. Salnikov, N. V. Chukanov, K. V. Kovtunov, I. V. Koptyug, J. G. Gelovani and E. Y. Chekmenev, *Chem. – Eur. J.*, 2021, 27, 2774–2781.
- 42 T. C. Eisenschmid, R. U. Kirss, P. P. Deutsch, S. I. Hommeltoft, R. Eisenberg, J. Bargon, R. G. Lawler and A. L. Balch, J. Am. Chem. Soc., 1987, 109, 8089–8091.
- 43 C. R. Bowers and D. P. Weitekamp, *Phys. Rev. Lett.*, 1986, 57, 2645–2648.
- 44 S. Nantogma, B. Joalland, K. Wilkens and E. Y. Chekmenev, *Anal. Chem.*, 2021, **93**, 3594–3601.
- 45 K. V. Kovtunov, I. E. Beck, V. I. Bukhtiyarov and I. V. Koptyug, *Angew. Chem., Int. Ed.*, 2008, 47, 1492–1495.
- 46 E. V. Pokochueva, D. B. Burueva, O. G. Salnikov and I. V. Koptyug, *ChemPhysChem*, 2021, 22, 1421–1440.
- 47 D. A. Barskiy, K. V. Kovtunov, E. Y. Gerasimov, M. A. Phipps, O. G. Salnikov, A. M. Coffey, L. M. Kovtunova, I. P. Prosvirin, V. I. Bukhtiyarov, I. V. Koptyug and E. Y. Chekmenev, J. Phys. Chem. C, 2017, 121, 10038–10046.
- 48 N. M. Ariyasingha, S. Nantogma, A. Samoilenko, O. G. Salnikov, N. V. Chukanov, L. M. Kovtunova, I. V. Koptyug and E. Y. Chekmenev, *J. Magn. Reson. Open.*, 2023, 16–17, 100135.
- 49 M. G. Pravica and D. P. Weitekamp, *Chem. Phys. Lett.*, 1988, 145, 255–258.
- 50 Z. Shi, K. Luo, S. Jani, M. February, N. Fernandes, N. Venkatesh, N. Sharif and S. Tan, *J. Neurosci. Res.*, 2022, 100, 2138–2153.
- 51 J. Vasquez-Vivar, Z. Shi, J.-W. Jeong, K. Luo, A. Sharma, K. Thirugnanam and S. Tan, *Redox Biol.*, 2020, **29**, 101407.
- 52 J. Vasquez-Vivar, Z. Shi, K. Luo, K. Thirugnanam and S. Tan, *Redox Biol.*, 2017, **13**, 594–599.
- 53 Z. Shi, N. Sharif, K. Luo and S. Tan, *Dev. Neurosci.*, 2024, 1-15, DOI: 10.1159/000538607.
- 54 K. V. Kovtunov, M. L. Truong, D. A. Barskiy, I. V. Koptyug, A. M. Coffey, K. W. Waddell and E. Y. Chekmenev, *Chem. – Eur. J.*, 2014, **20**, 14629–14632.

- 55 N. M. Ariyasingha, A. Samoilenko, M. R. Chowdhury, S. Nantogma, C. Oladun, J. R. Birchall, T. Bawardi, O. G. Salnikov, L. M. Kovtunova, V. I. Bukhtiyarov, Z. Shi, K. Luo, S. Tan, I. V. Koptyug, B. M. Goodson and E. Y. Chekmenev, *Chem. Biomed. Imaging*, 2024, 2, 698–710.
- 56 N. M. Ariyasingha, B. Joalland, H. R. Younes, O. G. Salnikov, N. V. Chukanov, K. V. Kovtunov, L. M. Kovtunova, V. I. Bukhtiyarov, I. V. Koptyug, J. G. Gelovani and E. Y. Chekmenev, *Chem. Eur. J.*, 2020, 26, 13621–13626.
- 57 A. Javed, R. Ramasawmy, V. Ozenne, P. Su, K. Chow and A. Campbell-Washburn, *Magn. Reson. Med.*, 2024, **92**, 173–185.
- 58 M. L. Truong, A. M. Coffey, R. V. Shchepin, K. W. Waddell and E. Y. Chekmenev, *Contrast Media Mol. Imaging*, 2014, 9, 333–341.
- 59 J. Wang, F. Kreis, A. J. Wright, R. L. Hesketh, M. H. Levitt and K. M. Brindle, *Magn. Reson. Med.*, 2018, **79**, 741–747.
- 60 S. Wu, M. L. Zuley, W. A. Berg, B. F. Kurland, R. C. Jankowitz, J. H. Sumkin and D. Gur, *Sci. Rep.*, 2017, 7, 2115.
- 61 B. Driehuys, S. Martinez-Jimenez, Z. I. Cleveland, G. M. Metz, D. M. Beaver, J. C. Nouls, S. S. Kaushik, R. Firszt, C. Willis, K. T. Kelly, J. Wolber, M. Kraft and H. P. McAdams, *Radiology*, 2012, 262, 279–289.
- 62 L. L. Walkup, R. P. Thomen, T. G. Akinyi, E. Watters, K. Ruppert, J. P. Clancy, J. C. Woods and Z. I. Cleveland, *Pediatr. Radiol.*, 2016, 46, 1651–1662.
- 63 G. J. Hirasaki and K. K. Mohanty, Fluid-Rock Characterization and Interactions in NMR Well Logging, Rice University, Houston, TX, 2001.
- 64 K. V. Kovtunov, M. L. Truong, D. A. Barskiy, O. G. Salnikov, V. I. Bukhtiyarov, A. M. Coffey, K. W. Waddell, I. V. Koptyug and E. Y. Chekmenev, *J. Phys. Chem. C*, 2014, 118, 28234– 28243.
- 65 E. W. Zhao, R. Maligal-Ganesh, C. Xiao, T.-W. Goh, Z. Qi, Y. Pei, H. E. Hagelin-Weaver, W. Huang and C. R. Bowers, Angew. Chem., Int. Ed., 2017, 56, 3925–3929.
- 66 Y. Wang and I. J. Dmochowski, Acc. Chem. Res., 2016, 49, 2179–2187.
- 67 I. Muradyan, J. P. Butler, M. Dabaghyan, M. Hrovat, I. Dregely, I. Ruset, G. P. Topulos, E. Frederick, H. Hatabu, W. F. Hersman and S. Patz, *J. Magn. Reson. Imaging*, 2013, 37, 457–470.
- 68 G. A. Burdock and I. G. Carabin, *Toxicol. Lett.*, 2004, **150**, 3–18.
- 69 R. H. McKee, D. Herron, M. Saperstein, P. Podhasky, G. M. Hoffman and L. Roberts, *Int. J. Toxicol.*, 2014, 33, 28S-51S.
- 70 C. F. Reinhardt, A. Azar, M. E. Maxfield, P. E. Smith and L. S. Mullin, *Arch. Environ. Health*, 1971, 22, 265–279.
- 71 N. R. Council, *High Magnetic Field Science and Its Application in the United States: Current Status and Future Directions*, The National Academies Press, Washington, DC, 2013.

Paper Analyst

- 72 Y. Lai and X. Wei, presented in part at the 20th ISMRM Conference, May 5-11, Melbourne, Australia,
- 73 Z. Shi, K. Luo, S. Deol and S. Tan, J. Neurosci. Res., 2022, **100**(12), 2154-2173.
- 74 G. Ambwani, Z. Shi, K. Luo, J.-W. Jeong and S. Tan, Dev. Neurosci., 2024, 1-13, DOI: 10.1159/000539212.
- 75 Z. Shi, J. Vasquez-Vivar, K. Luo, Y. Yan, F. Northington, M. Mehrmohammadi and S. Tan, Dev. Neurosci., 2019, 40(5-6), 534-546.