



Cite this: *Lab Chip*, 2024, 24, 2774

## High-resolution low-cost LCD 3D printing for microfluidics and organ-on-a-chip devices†

Houda Shafique,<sup>ab</sup> Vahid Karamzadeh,<sup>ib</sup> Geunyoung Kim,<sup>ab</sup> Molly L. Shen,<sup>ab</sup> Yonatan Morocz,<sup>ab</sup> Ahmad Sohrabi-Kashani<sup>ab</sup> and David Juncker<sup>id</sup>★<sup>ab</sup>

The fabrication of microfluidic devices has progressed from cleanroom manufacturing to replica molding in polymers, and more recently to direct manufacturing by subtractive (e.g., laser machining) and additive (e.g., 3D printing) techniques, notably digital light processing (DLP) photopolymerization. However, many methods require technical expertise and DLP 3D printers remain expensive at a cost ~15–30 K USD with ~8 M pixels that are 25–40 μm in size. Here, we introduce (i) the use of low-cost (~150–600 USD) liquid crystal display (LCD) photopolymerization 3D printing with ~8–58 M pixels that are 18–35 μm in size for direct microfluidic device fabrication, and (ii) a poly(ethylene glycol) diacrylate-based ink developed for LCD 3D printing (PLink). We optimized PLink for high resolution, fast 3D printing and biocompatibility while considering the illumination inhomogeneity and low power density of LCD 3D printers. We made lateral features as small as 75 μm, 22 μm-thick embedded membranes, and circular channels with a 110 μm radius. We 3D printed microfluidic devices previously manufactured by other methods, including an embedded 3D micromixer, a membrane microvalve, and an autonomous capillary circuit (CC) deployed for interferon-γ detection with excellent performance (limit of detection: 12 pg mL<sup>-1</sup>, CV: 6.8%). We made PLink-based organ-on-a-chip devices in 384-well plate format and produced 3420 individual devices within an 8 h print run. We used the devices to co-culture two spheroids separated by a vascular barrier over 5 days and observed endothelial sprouting, cellular reorganization, and migration. LCD 3D printing together with tailored inks pave the way for democratizing access to high-resolution manufacturing of ready-to-use microfluidic and organ-on-a-chip devices by anyone, anywhere.

Received 31st December 2023,  
Accepted 12th April 2024

DOI: 10.1039/d3lc01125a

rsc.li/loc

## Introduction

Microfluidics, through miniaturization in micrometer-sized vessels and microchannels, can reduce the fluid volumes required for analysis and synthesis to microliters and less, form the foundation for lab-on-a-chip devices, and are amenable to automation.<sup>1,2</sup> However, wider adoption of microfluidics and lab-on-a-chip devices in diagnostics, synthesis, and research is slowed by complex fabrication processes. Microfluidics emerged as a field thanks to cleanroom microfabrication inherited from the semiconductor industry relying on photolithography and using silicon or glass microfabrication methods which are dependent on capital cost intensive equipment. Soft lithography methods helped relieve the dependency on the

cleanroom as multiple replicates from a single microfabricated mold could be made in a common research lab, and greatly accelerating the adoption and dissemination of microfluidics for primarily research applications.<sup>3</sup> More recently, direct manufacturing methods have been introduced including subtractive ones such as laser ablation<sup>4</sup> or micromilling,<sup>5</sup> but offer limited relief due to drawbacks such as the need for expensive equipment, technical expertise, or provide limited resolution.

Additive manufacturing, and in particular 3D stereolithography (SLA) printing characterized by layer-by-layer UV patterning and photopolymerization of successive layers in a photocurable ink to build up a 3D printed object, has received considerable attention thanks to its affordability, high-resolution, and ease-of-use.<sup>6,7</sup> A layer is exposed to a digital pattern that solidifies the ink within a defined layer thickness; the layer then rises to allow uncured ink to fill the void, followed by digital photopolymerization of the new layer, and the process repeats iteratively. In an effort to clarify the terminology, we distinguish three methods of SLA and strategies to selectively expose ink within the layer: (i) laser scanning SLA operating with a galvanometer, (ii) digital light

<sup>a</sup> Biomedical Engineering Department, McGill University, Montreal, QC, Canada

<sup>b</sup> Victor Phillip Dahdaleh Institute of Genomic Medicine, McGill University, Montreal, QC, Canada

† Electronic supplementary information (ESI) available. See DOI: <https://doi.org/10.1039/d3lc01125a>



processing (DLP-SLA) that relies on a digital micromirror device and an optical system for projecting a pattern, and most recently (iii) masked SLA using a liquid crystal display (LCD) 3D printer where collimated light is directed through an LCD screen that digitally renders the design and photopolymerizes ink atop the LCD.

Laser SLA gained popularity thanks to high-resolution prototyping on a large print bed ( $335 \times 200 \times 300 \text{ mm}^3$ ) for microchannels ranging between 250–500  $\mu\text{m}$  with 30–140  $\mu\text{m}$  laser spot sizes.<sup>8</sup> Low-force SLA using a flexible vat reduces the adhesion force between formed layers and the bottom of the vat for intricate microfeature formation (e.g., separation membranes).<sup>9</sup> Additionally, many materials used in laser SLA are biocompatible,<sup>10</sup> but have largely been limited to commercial inks with proprietary formulations. Further, the single spot photopolymerization process with one or two lasers increases build times, especially for microfluidic devices that are generally blocks of solid ink with few voids that constitute the channels.

DLP 3D printing became widely adopted for microfluidics thanks to rapid and high-resolution fabrication with reported microchannels as small as  $18 \times 20 \mu\text{m}^2$ , 3D printer pixel sizes ranging from 2–40  $\mu\text{m}$ , and an illumination wavelength between 365–405 nm that can be used to photocure a wide range of materials.<sup>11–13</sup> The availability of open-source printers, online design repositories (e.g., Thingiverse, GrabCAD, Printables), tailored workflows (e.g., print-pause-print for multimaterial designs),<sup>14</sup> and custom ink formulations further increase the potential. The development of open-source inks such as those based on poly(ethylene glycol) diacrylate (PEGDA) for DLP 3D printing benefit from known compositions, which could help evaluate the impact of leachable and washable cytotoxic photosensitive components, and can be tailored and optimized for high-resolution embedded 3D printing, enhanced mechanical properties, low viscosity for fast printing, as well as for low protein adsorption and cytocompatibility.<sup>7,15,16</sup>

Altogether, high-precision 3D printers, custom inks, and direct 3D printable designs enables digital manufacturing, i.e., the seamless and automated fabrication from digital file to final product with minimal post-processing. Digital manufacturing of microfluidic components has been possible early on, and now extended to the fabrication of fully functional systems based on capillary flow.<sup>7</sup> Indeed, as capillary microfluidics can operate without peripherals,<sup>17</sup> and complex fluidic algorithms could be structurally encoded into so-called capillary circuits (CCs),<sup>18,19</sup> our group showed digital manufacturing of functional systems in the form of CCs. Thanks to custom intrinsically hydrophilic inks, ready-to-use CCs systems, could thus be printed using DLP 3D printers.

However, the capital cost of common research-grade microfluidic DLP 3D printers (~15–30 K USD) constitute a significant entry barrier for many potential users. Furthermore, while the pixel numbers have increased, with many printers culminating at  $3840 \times 2160 \cong 8 \text{ M pixels}$ , the trade-off between print resolution and build area has not

been resolved for microfluidics which require small pixel size, and hence small build areas, but come at the cost of limited manufacturing throughput.

LCD photopolymerization 3D printers retail for as little as ~150–600 USD, with pixel numbers of 4 K (>8 M pixels), 8 K (>33 M pixels), and up to 12 K (>58 M pixels), and pixel size of 18–50  $\mu\text{m}$ , thus outperforming DLP 3D printers both in terms of number of pixels and affordability. LCD 3D printers utilize an array of discrete light-emitting diodes (LEDs) that can now be mounted at high density (i.e., chip-on-board, COB) and that are collimated by an optical system (e.g., COB lens and Fresnel lens) then pass through an LCD screen to reach the vat bottom. The number of pixels has been growing exponentially, and with a range of pixel sizes that extend to smaller dimensions, thus offering both higher density and larger print areas, and the capacity to print high resolution structures such microfluidics on large print beds. However, in a recent study, Caplins *et al.* report illumination non-uniformity due to variable irradiance and spectral differences in discrete LEDs resulting in inconsistent prints.<sup>20</sup> Furthermore, the 50% transmittance loss of LCD screens by the crossed polarizers further reduces the irradiance of LCD 3D printers ( $2\text{--}3 \text{ mW cm}^{-2}$ ) compared to their DLP counterparts ( $5\text{--}100 \text{ mW cm}^{-2}$ ). Printing more voxels per time requires higher irradiance as the rate of printing for a given ink is limited by the power density of the light source.<sup>21</sup> Lastly, LCD screens degrade rapidly at low wavelengths and are thus limited to >400 nm illumination, which reduces material selection and ink efficiency.<sup>12,20,22,23</sup> Prior work has shown success in leveraging LCD 3D printing for microfluidic master mold fabrication,<sup>24–27</sup> but the potential for throughput manufacturing on large build plates and direct LCD 3D printing of open and embedded microchannels has not been shown.

Here, we present high-resolution fabrication of embedded and open microfluidic devices using low-cost LCD 3D printing with a custom formulated low-viscosity PEGDA-based ink that cures using low irradiance and minimizes the effect of illumination variability on curing depth. The lateral and vertical resolution of open and embedded structures are characterized using a series of test structures, and showcases high fidelity and dimensionally accurate printing down to a resolution in the tens of micrometers. The biocompatibility of the ink is validated based on an ISO standard for cell toxicity. Three microfluidic devices are manufactured by LCD 3D printing and characterized: (1) a microfluidic mixer previously made by laser micromachining, (2) membrane microvalves commonly made by replica molding, and (3) CCs previously made by DLP 3D printing. Finally, we demonstrate LCD 3D printing for microfluidic organ-on-a-chip (OoC) devices and for mass production. An OoC for co-culture of spheroids separated by an endothelial barrier was designed, printed and tested within 2 weeks. OoC devices are further used to illustrate large area printing in a well plate format, and for mass production by 3D printing thousands of OoC devices in a single run.



# Materials and methods

## Materials

**Ink materials.** Poly(ethylene glycol) diacrylate (PEGDA)-250 (Cat. #475629, lot #MKCS0146, Sigma-Aldrich, Oakville, Ontario, Canada), diphenyl(2,4,6-trimethylbenzoyl)phosphine oxide (TPO) (Cat. #415952, lot #MKCK2346, Sigma-Aldrich, Oakville, Ontario, Canada), 2-isopropylthioxanthone (ITX) (Cat. #I067825G, lot #ZNNQE-KT, TCI America, Portland, Oregon, United States), pentaerythritol tetraacrylate (PETTA) (Cat. #408263, lot #MKCR5556, Sigma-Aldrich, Oakville, Ontario, Canada).

**Other chemicals.** 3-(Trimethoxysilyl)propyl methacrylate (Cat. #M6514, lot #SHBG7600V, Sigma-Aldrich, Oakville, Ontario, Canada), fluorescein sodium salt (Cat. #46960, lot #2082530, Sigma-Aldrich, Oakville, Ontario, Canada), isopropyl alcohol (IPA) (Fisher Scientific, Saint-Laurent, Quebec, Canada).

**Immunoassay.** Purified mouse monoclonal IgG anti-human interferon- $\gamma$  capture antibody (Cat. #MAB2852, lot #FIO1022021, R&D Systems, Minneapolis, Minnesota, United States), biotinylated affinity purified goat IgG anti-human interferon- $\gamma$  detection antibody (Cat. #BAF285, lot #ZX2721071, R&D Systems, Minneapolis, Minnesota, United States), recombinant human interferon- $\gamma$  protein (Cat. #285-IF, lot #RAX2422031, R&D Systems, Minneapolis, Minnesota, United States), Pierce streptavidin poly-horseradish peroxidase (pHRP) (Cat. #21140, lot #XJ360080, Thermo Fisher Scientific, Waltham, Massachusetts, United States), SIGMAFAST 3,3'-diaminobenzidine tablets (Cat. #D4293, lot #SLCG5357, Sigma-Aldrich, Oakville, Ontario, Canada), bovine serum albumin (BSA) (Cat. #001-000-162, lot #162191, Jackson ImmunoResearch Labs, West Grove, Pennsylvania, United States), BSA-biotin (Cat. #A8549, Sigma-Aldrich, Oakville, Ontario, Canada), Tween 20 (Cat. #P7949, lot #SLBX0835, Sigma-Aldrich, Oakville, Ontario, Canada).

All assay reagents were prepared using 1 $\times$  phosphate-buffered saline (PBS) (pH  $\sim$  7.4) supplemented with 0.05% Tween 20 and 5% BSA. All other solutions were prepared using water from a Milli-Q system (resistivity: 18 M $\Omega$  cm; Millipore).

## Ink preparation

The 3D printing ink was based on a low molecular weight PEGDA-250 supplemented with 0.5% (wt/wt) diphenyl(2,4,6-trimethylbenzoyl)phosphine oxide (TPO) photoinitiator, 1.5% (wt/wt) 2-isopropylthioxanthone (ITX) photoabsorber, and 2% (wt/wt) pentaerythritol tetraacrylate (PETTA) crosslinker. The reagents were mixed in a 500 mL amber glass bottle under magnetic stir for at least 2 h before use and stored at room temperature thereafter.

## 3D printing of microfluidic chips

The microfluidic chips were designed either in AutoCAD (Autodesk) or Fusion 360 (Autodesk), then exported as an STL

file for slicing in a third-party software, CHITUBOX, at a layer thickness of 20 or 50  $\mu$ m. The slices were uploaded to the Elegoo Mars 3 Pro, Elegoo Mars 4 Ultra, Elegoo Saturn 2, or Elegoo Saturn 3 Ultra (ELEGOO, Shenzhen, China) masked stereolithography LCD 3D printers with a 405 nm light source. Print settings for all the devices presented here are given in Table S1.† The printed chips were washed on the build plate to remove excess uncured resin with IPA and dried with compressed air or nitrogen, followed by 1 min of UV curing (CureZone, Creative CADWorks, Concord, Ontario, Canada). Embedded devices were ready for use following UV curing; meanwhile, open channel CCs were sealed with a pressure adhesive tape (9795R microfluidic tape, 3M, Perth, Ontario, Canada) to encapsulate the microchannels.

## FTIR-ATR spectroscopy

Functional group characterization was done using FTIR-ATR spectroscopy (Nicolet 6700/Smart iTR, Thermo Scientific) on both uncured and cured PLInk samples. The measurement was performed on 8  $\mu$ L of uncured PLInk followed by curing the sample with a 405 nm illumination wavelength and grinding into powder format, then loading 500 mg of the powder to measure the spectra of the cured sample. Functional groups were identified based on peak positions to monitor the difference after photopolymerization.

## Cure depth, light penetration depth, and absorbance measurements

To determine the penetration depth of light, 50  $\times$  75  $\times$  1 mm<sup>3</sup> glass slides were first cleaned with IPA, then silanized *via* liquid phase deposition by immersing a glass slides in a solution of 2% 3-(trimethoxysilyl)propyl methacrylate prepared in toluene for at least 2 h or overnight. The slides were then cleaned in fresh toluene and dried with compressed nitrogen. The treated glass slides were placed directly on the 3D printer LCD screen; with the UV illumination on, the power intensity was read through the glass using a UV light meter with a 405 nm probe (Model 222, G&R Labs, Santa Carla, California, United States) to be 2.23 mW cm<sup>-2</sup>. Then, 8  $\mu$ L of uncured ink was placed on the glass slide, and the UV light was illuminated at different exposure times and repeated for each ink formulation. Following exposure, the glass was cleaned with IPA to remove excess uncured ink, dried with compressed nitrogen, and the cure depth of the formulation was measured using a stylus profilometer (DektakXT, Bruker, Billerica, Massachusetts, United States) that was configured to measure using a 12.5  $\mu$ m probe radius with a 3 mg force to scan a 7 mm region in 20 s. The cure depths were recorded in Vision64 and the average height was measured according to the International Organization for Standardization (ISO) 4287 protocol after 2-point leveling to record the baseline.

The light absorbance of each photocurable ink was measured using a spectrophotometer (NanoDrop@ND-1000, NanoDrop Technologies, Wilmington, Delaware, United



States). A blank reading was performed using MilliQ water, followed by recording the light absorption spectra with 2  $\mu\text{L}$  of ink solution at a 0.1 mm path length.

### 3D printer emission spectra measurement

The emission spectrum of the 3D printer was measured using an extended range (200–1000 nm) charge-coupled device spectrometer (CCS200, Thorlabs) *via* a 200  $\mu\text{m}$ -diameter bifurcated optical fiber (BFY200HF2, Thorlabs). Spectral data was recorded with a 500  $\mu\text{s}$  integration time and analyzed in the ThorSpectra software to obtain the emission spectra and relative intensities.

### Viscosity measurement

The viscosity of 3D printing inks was measured using a vibrating rod viscometer (SV-10, A&D Company, Limited). A 50 mL sample was loaded into the viscometer receptacle; the sensing rods were lowered into the receptacle until they were fully immersed in the inks, then the rods were set to vibrate at a frequency of 30 Hz to measure the resistance to flow. All measurements were done at 21  $^{\circ}\text{C}$ .

### Cell culture and cytotoxicity assay

To assess the cytotoxicity of the ink formulation, a cytocompatibility assay was performed in compliance with ISO 10993-5:2009 standards. The cells used in this study were kindly provided by Dr. Arnold Hayer of McGill University,<sup>28</sup> and they were grown and passaged according to ATCC's recommendations and cultured in EGM-2 media. Briefly, 8  $\times$  3 mm<sup>2</sup> (diameter  $\times$  thickness) rings were 3D printed and washed for 72 h with 70% ethanol with daily refresh of ethanol and then washed with PBS for 48 h to remove any unreacted photoactive components. The rings were then co-cultured with mCherry-labelled human umbilical vein endothelial cells (HUVECs) seeded at a density of 100 000 cells per well in a 24-well plate. Quantitative cell viability measurements were performed every 24 h over a total of 72 h using the PrestoBlue™ cell viability reagent. HUVECs seeded at an identical density were cultured alongside the ring co-culture as a control and used to establish 100% cell viability for each time point. Both the control and co-culture conditions were imaged every 24 h over a total of 72 h using a Ti2 inverted microscope and analyzed using NIS-Element (Nikon, Japan) for all biological replicates.

### Numerical simulation of concentration fields in the micromixer

The concentration field of the micromixer was calculated by the finite element method using COMSOL Multiphysics v.5.6 (COMSOL, Inc., Burlington, Massachusetts, United States). The diffusion coefficient of fluorescein ( $4.25 \times 10^{-10} \text{ m}^2 \text{ s}^{-1}$ ) was applied to solve the steady-state concentration field of fluorescein at a flow rate of 0.1 mL min<sup>-1</sup>. The concentration field was sliced into cross-sections to obtain the splitting and recombining stream profiles along the length of a mixing unit.

### Fluidic demonstrations

To visually assess the fluid flow in the microfluidic chips, a 2% solution of food dye in MilliQ water was prepared and loaded in the chips. For the micromixer, a 10  $\mu\text{M}$  solution of fluorescein was prepared in MilliQ water. In the case of the ELISA-chips, the devices were assessed with a solution of 2% food dye in 1 $\times$  PBS buffer containing 0.05% Tween 20.

### Fluorescent imaging through microfluidic chips

To facilitate microscopy imaging of fluorescent solutions in the chips, micromixer devices were mounted to a glass slide by UV photopolymerizing a drop of uncured resin between the chips and a plain glass slide ( $25 \times 75 \times 1 \text{ mm}^3$ ) for 40 s. The device was printed with cylindrical ports connected to a programmable syringe pump (Kd Scientific KDS250) *via* Tygon E-3603 tubing to flow solutions into the micromixer at known flow rates. Fluorescent images were acquired using a Nikon Ti2 inverted fluorescence microscope using NIS elements. Flow profiles were analyzed in ImageJ2 Ver. 2.9.0/1.53t (public domain software, National Institute of Health, Bethesda, Maryland, United States).

### Flow rate measurements through the microvalve

To assess the functionality of the microvalve, black dyed water was flown continuously through the flow channel inlet and collected in a beaker at the outlet. Meanwhile, a pressure gauge (MA059, MEASUREMAN) and an air pressure regulator (850-AC, ControlAir Inc.) were used to control the air pressure from a compressed air source directed at the control channel. The pressure regulator was used to adjust the control pressure and the liquid collected in the outlet beaker was massed after a known collection time (*i.e.*, 10 s) on a digital analytical balance (XS204, Mettler-Toledo) to determine the flow rate. The air pressure in the control channel was increased in  $\sim 3 \text{ kPa}$  increments until the outlet flow rate neared 0  $\mu\text{L s}^{-1}$ , indicating that the valve was closed.

### Contact angle measurement

The static contact angle was measured by placing 2  $\mu\text{L}$  of either water, PBST 0.05%, or PBST 0.1% supplemented with 2% liquid food dye on a  $10 \times 10 \times 2 \text{ mm}^3$  3D printed part and then photographed from the side using a digital camera (Panasonic Lumix DMC-GH3K) with a macro lens (M.Zuiko Digital ED 60 mm F2.8 Macro). The images were imported to ImageJ where the Contact Angle plugin was used to measure the static contact angle.

### ELISA-chip nitrocellulose assay preparation

The assay was designed based on lateral flow nitrocellulose membranes (Vivid 120, no. VIV1202503R; Pall Corporation, Port Washington, USA) that were cut to  $3 \times 12 \text{ mm}^2$  (width  $\times$  length) with a pointed base using a film cutter (Cameo 3, Silhouette Portrait, Lindon, USA). The membranes were then spotted using an inkjet spotter (sciFLEXARRAYER SX,





Scienion) with a  $2.5 \times 1 \text{ mm}^2$  (width  $\times$  length) test and control line spaced 5 mm apart. The test line was spotted with  $100 \mu\text{g mL}^{-1}$  of anti-human IFN- $\gamma$  antibody in a  $0.22 \mu\text{m}$  filtered 1X PBS buffer by programming the release of 350 pL droplets in a  $25 \times 4$  line array; spotting was done over 40 passes, wherein each pass covered alternating positions on the line array to allow for spots to dry between passes. Similarly, the control line was spotted with  $50 \mu\text{g mL}^{-1}$  of BSA-biotin with 8 passes covering alternating positions for each pass. The spotted membranes were dried at  $37^\circ\text{C}$  for 1 h, then blocked in a solution of 0.1% Tween 20 in  $1\times$  PBS supplemented with 5% BSA by dipping and wetting the membranes in a tray containing excess blocking buffer placed on an orbital shaker at 100 rpm for 1 h at room temperature. The blocked membranes were left to dry at  $37^\circ\text{C}$  for 1 h, followed by overnight storage in an air-tight container with desiccant at  $4^\circ\text{C}$ , then used within 48 h of protein spotting. The nitrocellulose assay strip was mounted onto the chip and sandwiched between 4 absorbent pads (Electrophoresis and Blotting Paper, Grade 320, Ahlstrom-Munksjo Chromatography) laser cut to be  $10 \times 2.4 \text{ mm}^2$  (width  $\times$  length). The drainage channel of the chip was mounted with a  $1.5 \times 3.5 \text{ mm}^2$  (width  $\times$  length) glass fiber (G041 SureWick, Millipore Sigma), then both the glass fiber and nitrocellulose strips were clamped into place using a custom 3D printed compressive clip.

### Assay protocol

The assay solutions for IFN- $\gamma$  detection were prepared in a wash and diluent buffer of 0.05% Tween 20 in  $1\times$  PBS supplemented with 5% BSA. Recombinant human IFN- $\gamma$  protein was spiked in the buffer at concentrations of 0,  $10^0$ ,  $10^1$ ,  $10^2$ ,  $10^3$ ,  $10^4$ ,  $10^5$ , and  $10^6 \text{ pg mL}^{-1}$ , followed by preparing reagent solutions including anti-human IFN- $\gamma$  gamma biotinylated antibody at  $1 \mu\text{g mL}^{-1}$  and streptavidin-poly-horseradish peroxidase (pHRP) at  $25 \mu\text{g mL}^{-1}$ . To prepare the assay substrate solution, SIGMAFAST™ DAB (3,3'-diaminobenzidine) tablets were dissolved in 5 mL Milli-Q water, then  $0.22 \mu\text{m}$  filtered prior to running the assay.

### Organ-on-a-chip seeding and maintenance

Spheroids were formed by seeding and culturing  $\sim 10\,000$  cells in a 96-well ultra-low attachment plate. Then, a five-day-old MDA-MB-231 breast cancer spheroid stained with 5-(and-6)-carboxyfluorescein diacetate succinimidyl ester (CFDA-SE) (Invitrogen, Massachusetts, United States) and a five-day-old IMR-90 lung fibroblast spheroid stained with cell tracker deep red (Invitrogen, Massachusetts, United States) was embedded in  $\sim 5 \mu\text{L}$  of 50% Matrigel DMEM solution and placed into the cell seeding chambers of the organ-on-a-chip device. The device was incubated at  $37^\circ\text{C}$  for 30 min for gelation, then the central cell seeding chamber was loaded with a 50% Matrigel EGM-2 solution containing  $\sim 200\,000$  mCherry-labelled HUVECs and placed on ice while the channels filled *via* capillary flow. Following filling, the device

was again incubated at  $37^\circ\text{C}$  for an additional 1 h for gelation. Finally, the media reservoirs were each loaded with  $\sim 100 \mu\text{L}$  of EGM-2 media and the device was imaged daily.

### Videos and image processing

3D images of the microfluidic devices were obtained by micro-computed tomography ( $\mu\text{CT}$ ) (SkyScan 1172, Bruker, Kontich, Belgium) at a pixel size of  $8 \mu\text{m}$ . Images were reconstructed using CT Analyzer (CTAn v.1.18, Bruker, Kontich, Belgium) and orthogonal projections were visualized and measured in ImageJ2 Ver. 2.9.0/1.53t (public domain software, National Institute of Health, Bethesda, Maryland, United States). 3D microscopy images were taken with a stereomicroscope (SteREO Discovery.V20, Zeiss, Baden-Württemberg, Germany). Videos and images were recorded to characterize flow using dyed water on either a Panasonic Lumix DMC-GH3K with a macro lens (M.Zuiko Digital ED 60 mm F2.8 Macro) or Sony  $\alpha 7\text{R III}$  camera also with a macro lens (Sony FE 90 mm F2.8 Macro G OSS Lens). Assay nitrocellulose membranes were imaged using a flatbed scanner (Epson Perfection V600) with the SilverFast 8 software at 600 dpi in a 48 bit RGB format, then imported to ImageJ2 for 16 bit grayscale colorimetric line intensity readouts. The readouts were normalized to rescale the colorimetric intensity from 0–65 535 gray values to 0–1 relative signal intensities.

## Results and discussion


Fig. 1 shows the process flow including a low-cost LCD 3D printer, a custom PEGDA-based ink (PLInk) optimized for LCD 3D printing, some microfluidic devices fabricated in this study, and the device design that closes the rapid prototyping cycle.

### Design of ink for LCD 3D printing

Embedded microfluidic channels are designed as narrow voids in a block of solid ink. To create these voids, the design of an ink formulation consists of monomers, light-responsive additives (*i.e.*, photoinitiator to catalyze the reaction, photoabsorbers to absorb excess energy), and crosslinkers. Polymerization must proceed efficiently layer-by-layer, *i.e.*, within the defined thickness of each layer while both avoiding under-polymerization of the current layer, and over-polymerization of uncured ink in voids of the preceding layers.

The design of PLInk was based on our prior ink formulations for 385 nm DLP 3D printing,<sup>7,16</sup> and adapted for LCD-based photopolymerization by considering the light heterogeneity, low irradiance, and 405 nm illumination wavelength. Based on our prior inks, PEGDA-250 was selected as the monomer due to its low viscosity, low protein adsorption, inherent cytocompatibility, and compatibility with solvents such as isopropyl alcohol for efficient removal of uncured ink in embedded microchannels. Diphenyl(2,4,6-trimethylbenzoyl)phosphine oxide (TPO) was selected again as the photoinitiator due to its low cytotoxicity and an





**Fig. 1** Low-cost LCD 3D printing of microfluidic devices and the rapid prototyping cycle. The workflow including (1) manufacturing on low-cost LCD 3D printers using a custom PEGDA-based ink (PLInk) optimized for LCD 3D printing, (2) directly manufactured microfluidic devices that can be tested and characterized, and (3) inform design improvement for the next iteration. Pixel sizes of 18–50  $\mu\text{m}$  and print areas of up to  $218 \times 122 \text{ mm}^2$  afford high print resolution over large areas. Photoinks optimized for LCD 3D printing with reduced sensitivity to light heterogeneity and low viscosity enable the direct manufacture of microfluidic chips including open and embedded microchannels with a lateral resolution  $<100 \mu\text{m}$  and vertical features as thin as 22  $\mu\text{m}$  in  $<45 \text{ min}$ . Scale bars = 500  $\mu\text{m}$ .


activation peak between 380–425 nm, as well as 2-isopropylthioxanthone (ITX) as the photoabsorber due to its broad absorbance peak between 350–425 nm, and known optical transparency, unlike other photoabsorbers such as 2-nitrophenyl phenyl sulfide (NPS), Sudan-1, or UV absorbing dyes with poor cytocompatibility and yellow-orange tints. The concordance between the activation range of TPO and absorbance range of ITX allowed us to model our ink design with the assumption that the absorbance remains consistent over the photopolymerizable region, even where the PLInk absorbance spectrum cuts off before the tail end of the 3D printer emission spectrum, ESI† Fig. S1. In a case where the photoinitiator range extended beyond the region attenuated by the photoabsorber and within the 3D printer emission spectrum, a polychromatic configuration that accounts for a changing absorbance could be considered.<sup>12,29</sup> Finally, due to the low irradiance of LCD 3D printers, we added pentaerythritol tetraacrylate (PETTA) crosslinker to increase

reactivity (discussed further below). Each of these ink components individually met suitability for a 405 nm illumination source, Fig. 2a.

We confirmed photocuring by Fourier transform infrared-attenuated total reflectance (FTIR-ATR) spectroscopy on uncured and 405 nm cured PLInk samples. A broader peak at  $1200 \text{ cm}^{-1}$  was observed for the cured ink, consistent with carbon-carbon bond formation between adjacent PEGDA and PETTA acrylate groups, ESI† Fig. S2.

Next, to mitigate the effects of light inhomogeneity, we sought to characterize the photopolymerization of the ink as function of total energy dosage and varying ITX photoabsorber concentration from 0 to 1.5%; the latter being the maximal concentration at which ITX could readily be dissolved. The fabrication of embedded microchannels, *i.e.*, voids, is predicated on precise control and understanding of the (measured) cure depth,  $C_d$ , to both avoid cross-linking of uncured ink trapped inside the microchannel while ensuring





**Fig. 2** PLInk optimization and characterization for LCD 3D printing of surface and embedded microstructures and microfluidics. (a) Formulation of PLInk containing PEGDA-250, TPO, ITX and PETTA. (b) Jacob's working curve showing the cure depth as a function of the energy dose for 1.5% ITX and yielding  $D_p = 56.5 \mu\text{m}$  and  $E_c = 2.31 \text{ mJ cm}^{-2}$ . (c)  $D_p$  and  $E_c$  values derived as in (b) for different concentrations of ITX showing a plateau for ITX > 0.75%. (d) Pillar printed area compared to nominal area of  $0.2 \times 0.2 \text{ mm}^2$  as function of crosslinker concentration. Data shows mean  $\pm$  standard deviation (STD) of four replicates. (e) 3D printed pillars showing features as small as  $\sim 3 \times 3$  pixels (nominal printer pixel size =  $28.5 \times 28.5 \mu\text{m}^2$ ). Scale bar =  $100 \mu\text{m}$ . (f) Monolithic circular microchannels with a corresponding  $\mu\text{CT}$  scan with channel cross-sections radii of  $\sim 125 \mu\text{m}$  (left) and  $\sim 110 \mu\text{m}$  (right) (nominal printer pixel size =  $35 \times 35 \mu\text{m}^2$ ). Scale bar =  $250 \mu\text{m}$ . (g)  $\mu\text{CT}$  image of a  $22 \mu\text{m}$  thick embedded membrane (nominal printer pixel size =  $18 \times 18 \mu\text{m}^2$ ). Scale bar =  $500 \mu\text{m}$ . (h) Cell viability at different time points of HUVECs expressing actin-mCherry co-cultured with 3D printed PLInk rings. Data shows mean  $\pm$  STD of three biological replicates. Scale bar =  $100 \mu\text{m}$ .

curing of the working layer.  $C_d$  is experimentally measurable and varies as function of the total energy,  $E$ , according to Jacob's working curve:<sup>12,22,30,31</sup>

$$C_d = D_p \ln\left(\frac{E}{E_c}\right)$$

where  $D_p$  is the penetration depth of light at which the light intensity of incident light is reduced by a factor  $1/e$  derived from Beer-Lambert's law,<sup>22</sup> and  $E_c$  is the critical energy dosage corresponding to the minimum required energy to initiate photopolymerization.  $E$  is simply  $t_e$ , the exposure time multiplied by  $P$ , the irradiance:

$$E = t_e \times P$$

The LCD 3D printer irradiance was measured to be  $2.23 \text{ mW cm}^{-2}$ . The thickness of polymerized ink for 1.5% ITX as function of energy dosage was fitted with Jacob's curve to derive both  $D_p = 56.5 \mu\text{m}$  and  $E_c = 2.31 \text{ mJ cm}^{-2}$ , Fig. 2b. The same experiment was repeated for varying concentrations of ITX, and the resulting  $D_p$  and  $E_c$  derived, Fig. 2c. As expected,  $D_p$  decreased with increasing ITX concentration. Interestingly,  $E_c$  also decreased with increasing photoabsorber, suggesting that ITX contributes not only to light adsorption, but also to more effective photopolymerization of the ink. We also observed that the slope was flattened for higher ITX concentration, meaning that the variation in the thickness of photopolymerized ink as a result of light illumination inhomogeneity would be





minimized. Hence, high ITX concentrations were optimal for LCD 3D printing. While we observed a lower plateau in both  $D_p$  and  $E_c$  once ITX concentrations reached 0.75%, we chose 1.5% as the optimal condition to minimize susceptibility to variations in ITX concentration.

To balance precision, material sensitivity and print speed, and while considering printer pixel size, we set the print layer thickness (and model slicing) to 20 or 50  $\mu\text{m}$ . This satisfied the requirement for printing embedded microchannels of slice layer thickness =  $0.3\text{--}1 \times D_p$  formulated by Nordin and colleagues.<sup>22</sup> PLInk also allowed for rapid photopolymerization with an exposure time  $t_e$  of 1.3–1.8 s and 3.5 s for a  $C_d$  of  $\sim 20 \mu\text{m}$  and  $\sim 50 \mu\text{m}$ , respectively. As an example of the benefits of lower  $D_p$ , for a change in energy dosage of 7–9  $\text{mJ cm}^{-2}$ , the layer thickness variation with 1.5% and 0.02% ITX would be  $\sim 20 \mu\text{m}$  and  $\sim 80 \mu\text{m}$ , respectively, ESI† Fig. S3. Commercial inks typically favor a high  $D_p$  ( $>179 \mu\text{m}$ ),<sup>32</sup> which has the advantage of printing thicker layer slices and faster print times, but are inadequate for printing embedded microchannels and susceptible to variable cured thickness with a non-uniform light source.

To improve printing fidelity, we supplemented the PLInk formulation with PETTA with four additional acrylate groups to increase the availability of polymerizable groups and speed up the formation of an interconnected polymer network. We empirically adjusted the PETTA concentration by measuring the printed area of  $0.2 \times 0.2 \text{ mm}^2$  pillars with a  $28.5 \mu\text{m}$  pixel size LCD 3D printer. Incomplete photopolymerization was visualized by tracking underfilling of the nominal pillar shape and by the distortion or bending of the pillars.<sup>33</sup> The PETTA concentration was increased until the nominal XY pillar area matched the 3D printed design, which was achieved at a value of 2%, Fig. 2d. Pillar printing confirmed suitable mechanical stability of the print without collapse and good dimensional accuracy, as illustrated with an array of  $\sim 3 \times 3$  pillars, Fig. 2e.

### PLInk performance characterization

PLInk is based on PEGDA-250 with a comparatively low viscosity of  $\sim 16 \text{ mPa s}$ , thus making it suitable for microchannel fabrication. Indeed, following photoexposure of a layer, the retraction of a relatively flat print attached to the vat bottom will create suction force; next, ink needs to flow into the growing gap, and immediately flow out of the gap as the print is lowered back onto the vat bottom to expose the next layer, all of which would benefit from a low viscosity ink. Notably, despite its high viscosity ( $\sim 700 \text{ mPa s}$  at  $25^\circ\text{C}$ ),<sup>34</sup> the addition of PETTA at low concentrations did not impact the native viscosity of PEGDA-250, and thus vastly outperformed commercial inks ( $\sim 200\text{--}500 \text{ mPa s}$ ) in this respect, ESI† Table S2. The low viscosity also facilitates printing of fine features as it reduced the risks of mechanical failure caused by suction and adhesion to the vat bottom. Coupling low viscosity and low  $D_p$ , the cured PLInk was smooth with a surface roughness of  $\sim 500 \text{ nm}$ , which

suggests favourability for intricate microchannel fabrication, ESI† Section S1 and Fig. S4.

To assess suitability for microfluidic device fabrication, we evaluated the resolution of the designed PLInk formulation by printing open channels with decreasing size and were able to print features as small as  $\sim 75 \times 75 \mu\text{m}^2$  with a  $35 \mu\text{m}$  pixel size LCD 3D printer, ESI† Fig. S5. We performed  $\mu\text{CT}$  scans of the device to evaluate the printing accuracy; we measured the printed open channel size and found it to be within 2.8% of the nominal dimension.

To assess our ability to 3D print embedded microchannels, we similarly evaluated the printing of progressively smaller rectangular and circular channels that were embedded a depth at least ten times greater than the  $D_p$ . Embedded rectangular channels down to  $\sim 170 \times 220 \mu\text{m}^2$  (width  $[W] \times$  height  $[H]$ ) were printed using a  $35 \mu\text{m}$  pixel size LCD 3D printer, ESI† Fig. S6. The smallest rectangular embedded conduits were within 2.7% of their nominal size. We found that high aspect ratio ( $H/W > 1$ ) channels were limited by the pixel resolution of the LCD screen, *i.e.*, typically 3–4 pixels because of scattering, non-parallel illumination, and possible photoinitiator diffusion.<sup>35,36</sup> Meanwhile, the height of low aspect ratio microchannels ( $H/W < 1$ ) was limited by the optical penetration (the shortest embedded channel  $\sim 2.3 \times D_p$ ).<sup>12</sup> Circular conduits are notably of interest to minimize capillary edge flow (also called filaments),<sup>7</sup> and embedded conduits with circular cross-section and radius as small as  $\sim 110 \mu\text{m}$  were printed with a dimensional accuracy within 1.5% of the nominal dimension, Fig. 2f and S7.† A shallow  $D_p$  also benefits the printing of thin embedded membranes due to fine control over the cured thickness and a sharp transition between cured and uncured layers. Vertical embedded channels designed with a series of ever thinner membranes were 3D printed down to a thickness as low as  $\sim 22 \mu\text{m}$  within a single exposure to demonstrate free-standing membrane fabrication, Fig. 2g and S8.†

Further, we evaluated the cytocompatibility of the ink by co-culturing 3D printed PLInk with human umbilical vein endothelial cells (HUVECs) according to the ISO 10993-5:2009 standard for implantable medical devices. A primary cell line was selected due to specific but rigorous culturing conditions for cells with high sensitivity to their environment and a limited passage number. We 3D printed  $8 \times 3 \text{ mm}^2$  (diameter  $\times$  thickness) rings and thoroughly washed any unreacted photoactive elements (details in the Methods), then co-incubated the PLInk rings with cells in a single well with shared media for 72 h.<sup>15</sup> After 72 h, we found  $>80\%$  cell viability, meeting the threshold for a cytocompatible material and demonstrating suitability for cell culture microfluidic device fabrication, Fig. 2h.

In summary, the optimized and low viscosity PLInk formulation for LCD 3D printing was found to be suitable for high-resolution and dimensionally accurate printing of smooth structures including posts, open and embedded microchannels, embedded membranes, and to be






cytocompatible, making it amenable for a broad range of applications, and notably in microfluidics as explored below.

### LCD 3D printing of embedded microfluidic mixer

We re-designed a recently published micromixer implementing the Baker's transformation and made by direct laser micromachining of two separate parts for direct LCD 3D printing.<sup>37</sup> In the original design, the micromixer was assembled from two substrate devices with open conduits made by direct laser machining, chemical wet etching, and bonding, thus forming a closed, interconnected weaving flow path. For LCD 3D printing, the micromixer was designed as a single digital model of an embedded micromixer including (i) overhanging wedges to split the fluidic streams horizontally and progressively, and (ii) an embedded pillar to create an interface before vertically recombining the two fluidic streams, Fig. 3a. The micromixer was designed with  $310 \times 310 \mu\text{m}^2$  cross-sections

that interweaved, merged into a large conduit of  $900 \times 900 \mu\text{m}^2$ , and then split again, and so on. The shaped pillars measured  $203 \mu\text{m}$  ( $\cong 6$  pixels, nominal 3D printer pixel size =  $35 \times 35 \mu\text{m}^2$ ) along its longest dimension and  $306 \mu\text{m}$  ( $\cong 9$  pixels) along its widest dimension, Fig. 3b. The pillar extended across the full height of the mixer (45 layers of  $20 \mu\text{m}$  each), which necessitated a sufficiently high energy dosage to fully crosslink the pillar as well as overhanging, suspended structures and preserve their integrity during the build plate movements, while at the same time preventing photopolymerization of PLInk in the embedded weaving conduits.  $\mu\text{CT}$  scans of the 3D printed device confirmed the fidelity and integrity of the pillar that narrowed down to  $\sim 45\text{--}72 \mu\text{m}$  ( $\cong 1\text{--}2$  pixels) in width before recombining the fluidic streams, and of the internal overhangs and the sharp edges that split, guide, and merge the fluidic streams, Fig. 3c(i). A finite element method numerical simulation that solved the steady-state concentration field of two fluidic streams illustrated the importance of the



**Fig. 3** Embedded microfluidic mixer. (a) Schematic representation of two fluidic channels combined into a single mixing unit. (b) Stereomicroscope image of four 3D printed mixing units showing overhangs and pillar formation in an embedded device. Scale bar =  $500 \mu\text{m}$ . (c) (i)  $\mu\text{CT}$  cross-sections of the microchannel with (ii) a corresponding numerical simulation showing the mixing principle based on horizontal stream splitting and vertical stream recombining. Scale bar =  $500 \mu\text{m}$ . (d) Stereomicroscope image of a single mixing unit showing yellow and blue fluidic streams split into ever thinning striations by the microarchitecture. Arrow shows the direction of flow. Scale bar =  $300 \mu\text{m}$ . (e) Mixing of  $10 \mu\text{M}$  fluorescein with clear water at a flow rate  $\cong 0.1$ ,  $1$ , and  $10 \text{ mL min}^{-1}$ , corresponding to a Reynold's number  $\cong 1.85$ ,  $18.5$ , and  $185$ , respectively. Arrow shows the direction of flow. Scale bar =  $500 \mu\text{m}$ .




microarchitecture to horizontally split and vertically recombine the flows for cross-sections along the length of the mixing unit, Fig. 3c(ii).

Owing to the transparency of the device, mixing could be visually tracked through the entire height of the channel. Using water with yellow and blue dyes allowed for visual tracking of the mixing and the observation of striations as the streams folded and recombined within the micromixer, Fig. 3d. We further assessed the mixing performance with a fluorescent dye (10  $\mu\text{M}$  fluorescein) in one of the streams and tracked the fluorescence intensity along the length of the mixer by fluorescence microscopy. The progression from two separate streams to complete mixing was visible from the intensity profile that progressed from a step function to a flat, homogeneous distribution of the dye across the width of the micromixer, Fig. 3e and S9.† When investigated over a range of laminar flow rates (0.01–10  $\text{mL min}^{-1}$ ), we observed the efficiency of mixing decreased with increasing flow rates, as expected because the time for diffusive mixing decreases. Interestingly, we observed that for flow rates  $>1 \text{ mL min}^{-1}$

the mixing efficiency did not decrease, but instead improved again, which we attribute to inertial effects and recirculation. The mixing performance was concordant with the laser-manufactured mixer and the Baker's transformation principle.<sup>37,38</sup> Across three replicate devices, we quantified the mixing efficiency to be 92–99%, confirming the successful printing and operation of the 3D printed device, ESI† Fig. S10, Section S2. The micromixer illustrates the potential of LCD 3D printing for producing complex embedded structures that are not easily manufactured by more traditional micromachining methods.

### LCD 3D printing of an embedded membrane microvalve

Next, we 3D printed an embedded membrane microvalve. Elastomeric microvalves made of polydimethylsiloxane (PDMS) and manufactured by replica molding have been widely used and adopted for microfluidics.<sup>39,40</sup> Recently, direct manufacturing of embedded free-standing membranes by 3D printing has been demonstrated using DLP 3D



**Fig. 4** Embedded membrane microvalve. (a) Microvalve schematic and (b) photograph showing flow channel, control channel, a  $\sim 43 \mu\text{m}$  thick membrane, ridged valve seat forming a separation wall with a  $100 \mu\text{m}$  gap to the membrane at the centre (nominal 3D printer pixel size =  $28.5 \times 28.5 \mu\text{m}^2$ ). Actuation of the membrane by pressurization in the control channel leads to deflection onto the valve seat and closure of the flow channel. (c) Orthogonal  $\mu\text{CT}$  views of the 3D printed membrane and valve seat. Scale bars =  $500 \mu\text{m}$ . (d) Top view images of the open and closed valve with schematics showing cross-sections of the valve according to the labels in (b). With an open valve shown on the left, the black water flows through the channel, while for a closed valve shown on the right, the pneumatically deflected membrane is sealed onto the valve seat and stops black water flow. Scale bars =  $500 \mu\text{m}$ . (e) Flow rate in the flow channel as a function of the pressure in the control channel showing the gradual closing of the valve and flow stop at  $\sim 41 \text{ kPa}$ . Data points are measurement collected from three different devices. Line is a guide to the eye.



printers.<sup>30,41</sup> We LCD 3D printed an embedded membrane with a valve seat modelled based on existing ones comprising a 40  $\mu\text{m}$ -thick membrane with a diameter of 1.7 mm and  $\sim 100$   $\mu\text{m}$  above a 500  $\mu\text{m}$ -wide ridged valve seat. An embedded control channel overlaid orthogonally above the membrane and the valve seat in the flow channel was used for membrane actuation by pressurization, Fig. 4a and b. All the channels were 3D printed with a unique inlet and outlet to facilitate precursor ink removal and avoid post-processing fabrication steps. The valve seat in the form of a thin curved ridge improved printability compared to a solid 'bowl' shape that might lead to incomplete ink removal while providing reliable valve closure upon actuation.  $\mu\text{CT}$  images of the valve revealed a fully released, free-standing membrane, Fig. 4c, ESI† Video S1. The measured thickness on the  $\mu\text{CT}$  images (with 8  $\mu\text{m}$  pixel resolution) was  $\sim 43$   $\mu\text{m}$ , closely matching the design.

Water spiked with a black dye was flown through the microvalve to visually assess whether the valve was open (*i.e.*, flow channel junction was visually black) or closed (*i.e.*, junction visually clear). The valve was designed to be open at rest, and as the compressed air pressure was increased in the control channel, the membrane deflected to form a seal with the valve seat, interrupting the flow of the black water, Fig. 4d.

The mechanical properties of 3D printed PLInk were assessed by tensile testing yielding a Young's modulus of  $68 \pm 3$  MPa, ESI† Fig. S11. Compared to elastomeric membranes, PLInk's Young's modulus was  $\sim 10\times$  higher than PDMS; therefore, a thin ( $\sim 40$ – $50$   $\mu\text{m}$ ), 1.7 mm diameter membrane was predicted to deflect  $\sim 100$   $\mu\text{m}$  at a control pressure of  $\sim 45$  kPa to seal the valve, ESI† Section S3. The control pressure was increased incrementally while the flow was monitored and flow stop observed at  $\sim 41$  kPa, Fig. 4e. The experimental valve closing pressure was thus in good agreement with the prediction, and the variation could be attributed to imprecision in the gap between the membrane and the valve, in the thickness of the membrane, or incomplete curing of the membrane that might make it more pliable. Overall, both the reproducibility of the closing pressure across all valve replicates, and the agreement to theory were consistent. While we did not assess the durability under cyclical stress loading, the durability of 3D printed membranes based on low molecular weight PEGDA inks was demonstrated by Folch and colleagues,<sup>41</sup> suggesting that the PLInk membrane will also be suitable, or could be made suitable, for cyclical loading. These results indicate that LCD 3D printing can be used for making thin, compliant, and mechanically actuated embedded elements such as membrane microvalves.

### LCD 3D printing of an ELISA-on-a-chip capillary circuit – an ELISA-chip

CCs operate by structurally encoding fluidic operations using capillary valves for fluidic operation and capillary flow


for self-filling, and function thanks to a controlled, moderate hydrophilicity. We previously developed hydrophilic inks for DLP 3D printing of functional CCs with embedded channels and with contact angles with water  $\sim 35^\circ$  owing to the use of hydrophilic acrylic or methacrylic acid additives.<sup>7</sup> The contact angle with water of PLInk was  $\sim 65$ – $70^\circ$ , which while being moderately hydrophilic, was too high for reliable capillary self-filling, ESI† Fig. S12a. The photopolymerization of acrylic or methacrylic acid groups competes with crosslinking by PEGDA acrylate groups, and thus requires higher light energy doses, which would lead to much higher exposure times for low irradiance LCD 3D printers. Previously, plasma activation had also been used, but depends on access to a plasma chamber, and only provides temporarily hydrophilicity for select materials.<sup>18,19</sup> Hence, instead of increasing the surface energy of the microchannels, we opted to reduce the surface tension of the aqueous solutions by adding surfactants (*i.e.*, Tween 20) and thereby reducing the contact angle to as low as  $\sim 46^\circ$  with 0.05% Tween 20 and  $\sim 31^\circ$  with 0.1% Tween 20, thus meeting the requirements for CC operation, ESI† Fig. S12b and c. Considering that the use of surfactants in immunoassays is common to reduce non-specific binding, their addition to the solutions does not compromise the suitability of CCs for typical biological applications.

To illustrate the reliability of LCD 3D printing, we designed a CC with a microfluidic chain reaction (MCR)<sup>18</sup> implementing an ELISA-on-a-chip akin to the ones made previously using DLP 3D printing of open microchannels followed by sealing with a hydrophobic pressure adhesive transparent cover.<sup>18,19</sup> The ELISA-chip was developed for a new target, with adjusted geometries for LCD 3D printing, and importantly with a reduced time-to-result while maintaining high sensitivity, Fig. 5a. The target was interferon (IFN)- $\gamma$ , a cytokine critical to the immune response against a wide range of infections,<sup>42</sup> and which is notably used in the IFN- $\gamma$  release assays as a biomarker for tuberculosis infection.<sup>43,44</sup> The microfluidic assay was based on a classical ELISA sandwich immunoassay using a capture antibody, a biotinylated detection antibody, and a streptavidin-enzyme conjugate (poly-horseradish peroxidase, pHRP). While in conventional well-plate ELISAs soluble substrates are used, for on-chip applications with a nitrocellulose membrane and under active flow conditions, precipitating substrates are required for localized accumulation of the enzymatically oxidized substrate, such as 3,3'-diaminobenzidine tetrahydrochloride (DAB), in the presence of pHRP and hydrogen peroxide, Fig. 5b. A nitrocellulose membrane spotted with an anti-IFN- $\gamma$  capture antibody was connected to the ELISA-chip that encoded an 8-step assay for automated, sequential flow of wash buffers and reagents. As in the DLP 3D printed ELISA-chip design, functions for on-chip aliquoting were integrated to facilitate the operations for untrained users, Fig. 5c.

The lower limit of detection of the previous ELISA-chip<sup>19</sup> outperformed rapid tests (*e.g.*, lateral flow assays), but the







**Fig. 5** LCD 3D printed microfluidic ELISA chip. (a) ELISA-on-a-chip designed for LCD 3D printing with structurally encoded sequential delivery of reagents to autonomously perform an assay and coupled with a built-in chip-to-assay connection. (b) ELISA workflow showing the sandwich immunoassay designed for the detection of IFN- $\gamma$  by sequentially delivering assay reagents and wash buffer to a nitrocellulose membrane pre-spotted with anti-human IFN- $\gamma$  capture antibody; (1) sample containing IFN- $\gamma$ , (2) biotinylated anti-human IFN- $\gamma$  detection antibody, (3) streptavidin-conjugated enzyme pHRP, and (4) enzyme substrate in the presence of hydrogen peroxide to generate the colorimetric readout. (c) Autonomous CC workflow for on-chip reagent aliquoting by metering the correct reagent volumes and drainage of the excess, followed by (d) MCR-based sequential delivery of assay reagents with wash steps in between. Arrows show the direction of flow. Scale bar = 5 mm. (e) Binding curve of the on-chip assay for the detection of IFN- $\gamma$  with a limit of detection of 12 pg mL<sup>-1</sup> (CV: 6.8%) across triplicate chips for each tested concentration point; line shows a 4-point logistic fit.

assay time was longer at 1 h 15 min. Thus, we sought to reduce the assay time for the LCD 3D printed ELISA-chip. The incubation times were structurally encoded by the volume of reagents that flowed over the test zone (see discussion on assay optimization below for further details), the capillary pressure of the pump (*i.e.*, absorbent pad and glass fiber conjugate pad backing the nitrocellulose membrane), and the flow resistance of the functional connections that linked each reservoir to the main channel. The capillary pressure coming from an absorbent pad backing the nitrocellulose membrane was the same as a single pump was used to wick all the reagents. Compared to

our previous ELISA-chip design that also had a glass fiber conjugate pad mounted the nitrocellulose and served both as a fluidic connection to the chip and an immediate capillary pump to wick reagents over the nitrocellulose, the glass fiber was considered a source of analyte loss due to protein adsorption over the assay run time. To remedy these limitations, we connected the nitrocellulose membrane to the ELISA-chip directly. Without the glass fiber, the chip-to-assay connection was re-designed as a gradual opening with a weak stop valve designed to break when the liquid front arrived at the end of the channel and wetted the nitrocellulose membrane; pre-wetting with buffer bridged the ELISA-chip's



liquid interface with the absorbent pad, and facilitated a connection to the capillary pump that subsequently began to wick the reagents over the nitrocellulose assay test zone. Finally, to adjust the flow rate, we increased the functional connection cross-sections to  $200 \times 200 \mu\text{m}^2$  across the entire chip. These changes reduced reagent loss and provided a suitable flow speed for consistent fluidic performance, which allowed us to reduce reagent volumes and the time-to-result to 48 min, Fig. 5d, ESI† Video S2.

We evaluated the flow reproducibility of the new LCD 3D printed ELISA-chip by timing each of the sequential steps in three replicate chips, Table 1. The flow of sample, which contains only limited concentration of analyte is the most critical step when considering assay reproducibility and LOD, and the one that necessitated high reproducibility. Other steps, such as detection antibody and enzyme are provided in excess concentration and hence variation of flow time is not expected to significantly affect the assay result. Likewise, precise incubation time for wash steps are not as critical as long as reagents are flowed and flushed across the nitrocellulose membrane. The comparatively high variability for the DAB incubation time could arise as a result of the precipitate formed on the test strip, especially at higher concentrations of IFN- $\gamma$ , which could affect the flow properties of the strip.

The assay portion of the ELISA-chip was optimized using a design of experiments approach, which enabled the optimization of multiple assay parameters simultaneously since the optimal concentration of one parameter would dictate the optimal of another in a classical sandwich immunoassay, and served to establish the relative contribution of each parameter.<sup>45</sup> We evaluated a capture antibody spotting concentration of 50, 100 and  $200 \mu\text{g mL}^{-1}$  and both a detection antibody and pHRP concentration of 1, 5, and  $25 \mu\text{g mL}^{-1}$  at a fixed sample concentration of  $100 \text{ ng mL}^{-1}$ . Using the Taguchi method for design of experiments,<sup>46</sup> the selection led to nine experiments to determine significantly impacting assay factors, ESI† Table S3. From the results, we evaluated the significance of each factor using analysis of variance and found that the capture antibody concentration was a significant parameter ( $p < 0.05$ ) for the assay performance, and the weighted contribution of the capture antibody concentration was found to be 47%, which was higher than the other factors, *i.e.*, detection antibody

(25%), and pHRP (24%), ESI† Table S4. Altogether, this indicated that a to reduce assay time while preserving the sensitivity, capture antibody spotting density needed to be increased. We kept the reagent volumes relatively low, *i.e.*, sample volume was  $75 \mu\text{L}$  and took  $\sim 14$  min to flow, and ensured that all the reagent were being delivered to the nitrocellulose membrane with no losses on a connecting glass fiber; meanwhile, we increased the spotting density of capture antibody from our original ELISA-chip by nearly 10-fold, resulting in  $0.7 \mu\text{L}$  of  $100 \mu\text{g mL}^{-1}$  capture antibody spotted on a thin  $3 \times 1 \text{ mm}^2$  (width  $\times$  length) line on the nitrocellulose membrane. Taking the relative contribution of each parameter into consideration, the optimal IFN- $\gamma$  assay involved flowing  $45 \mu\text{L}$  (at a fixed flow rate and fixed time) of the detection antibody at  $1 \mu\text{g mL}^{-1}$  and of pHRP at  $25 \mu\text{g mL}^{-1}$ . The assay wash steps volume and time were minimized to reduce assay run time while preventing pre-mixing of reagents in the main channel.

Following both optimization of the fluidic performance and the nitrocellulose assay, we evaluated the ELISA-chip over a wide concentration range of IFN- $\gamma$  and achieved a limit of detection as low as  $12 \text{ pg mL}^{-1}$ . With a 6.8% CV, our LCD 3D printed ELISA-chip showed consistent performance, Fig. 5e and S13.† Using PLInk and based on the cost of research-grade materials, assay reagents and assembly components of the nitrocellulose assay, the ELISA-chip costs  $<4$  USD per device, ESI† Table S5. The low capital cost and low material cost enable affordable fabrication of autonomous ELISA-chip devices globally, especially in low- and middle-income countries with limited access to traditional manufacturing and a high incidence of infectious disease. These results indicate the suitability of low-cost LCD 3D printing for the fabrication of ready-to-use CC chips that automate complete assays with lab-grade accuracy and short time-to-result.

### LCD 3D printing of organ-on-a-chip devices

We 3D printed an organ-on-a-chip (OoC) device that included embedded microchannels. Microfabricated OoC devices have largely been fabricated by soft lithography for manufacturing high resolution microstructures; replica molded devices fabricated with PDMS offer biocompatibility, tunable elasticity, and gas permeability.<sup>47</sup> Recently, DLP 3D printed OoC and cell culture devices have been introduced with complex 3D architectures not feasible by replica molding.<sup>16,48</sup> However, the biocompatibility of photoinks including the photopolymer and additives such as the photoinitiator and photoabsorber must be validated, and in some case poorly cytocompatible phototoxins can be leached out for applications requiring cell seeding, cell reorganization and migration.<sup>11,15,29</sup>

In the case of PLInk, biocompatibility was assessed after washing and sterilizing the 3D printed devices in 70% ethanol and PBS for 5 days before cell seeding (see PLInk performance metrics above and Materials and methods for more

**Table 1** LCD 3D printed ELISA-chip steps, reagent, volume, and timing of automated assay

Reagent	Vol. [ $\mu\text{L}$ ]	Time $\pm$ STD [s]
1 Sample (IFN- $\gamma$ )	75	$875 \pm 16$ (CV: 1.9%)
2 Wash buffer, PBST 0.05% + 5% BSA	15	$152 \pm 21$ (CV: 14.2%)
3 Biotinylated detection antibody	45	$513 \pm 21$ (CV: 3.9%)
4 Wash buffer	15	$144 \pm 19$ (CV: 12.8%)
5 Streptavidin-pHRP	45	$518 \pm 34$ (CV: 6.5%)
6 Wash buffer	15	$146 \pm 17$ (CV: 17.5%)
7 Enzyme substrate DAB	45	$504 \pm 52$ (CV: 10.3%)
8 Wash buffer	5	$79 \pm 25$ (CV: 31.6%)



details). Then, to evaluate its potential for an LCD 3D printed cell culture device, we developed a new multi-OoC spheroid design to monitor the interaction between two organ compartments. *In vivo*, organ–organ segregation and communication are maintained by endothelial barriers and vascular flow, respectively. To mimic native physiology, we designed two shallow spheroid seeding compartments that were separated with a 400  $\mu\text{m}$  wide vascular barrier channel in the middle. To maintain interconnectivity between spheroid compartments and the vascular barrier, we added capillary stop valves designed as parallel embedded microchannels that preserve compartmentalization during seeding subsequently forming an open conduit between the two reservoirs and the central vascular barrier, Fig. 6a and b. To increase the surface area of crosstalk, capillary stop valves with a  $200 \times 200 \mu\text{m}^2$  square cross-section were stacked as a  $3 \times 4$  array between the vascular barrier and each of the two organ compartments on either side; thus, each OoC included 24 embedded stop valves. A cross-section visualized by stereomicroscopy shows the 3D printed structures, Fig. 6c.

The spacing and dimensions of the OoC devices were designed to match the overall footprint of an industry-standard 384-well plate with 4 inlets per OoC, *i.e.*, two for seeding the spheroids and one to seed the endothelial cells and one for air to exit, hence 192 OoC units and 768 inlets in total per well plate. The large build area of the LCD 3D printer could accommodate up to three OoC plates that could be printed within <1.5 h, Fig. 6d. To validate the fluidic operation, the wells were loaded with gelatin solutions spiked with red and green dye to mimic an extracellular matrix loaded into the seeding chamber, which were found to be effectively compartmentalized with 100% yield on three separate plates, ESI† Fig. S14. The cost based on research-grade materials used for PLInk synthesis of an individual OoC device and of a 384-well plate OoC are estimated at ~0.10 USD and ~20 USD, respectively, ESI† Table S6. Cost savings could readily be achieved by increasing the void space on the plate, as is common for injection molded well plates.

To demonstrate cell culture compatibility of the multi-OoC design, the side chambers of the OoC device were seeded with a Matrigel-embedded five-day-old IMR-90 lung fibroblast spheroid on the left and a Matrigel-embedded five-day-old MDA-MB-231 breast cancer spheroid on the right. The central channel was loaded with HUVEC-embedded Matrigel solution to form the vascular barrier and the remainder of the device was loaded with media. To demonstrate reproducible seeding of the devices, three devices 3D printed separately were seeded and imaged, showing rounded spheroids isolated in their respective chambers and the endothelial cells in the middle channel, ESI† Fig. S15. Then, one OoC device was monitored for a time course of 5 days, revealing maturation of the model with endothelial cell tightening and reorganizing into a vascular structure. Initial sprouting of the endothelial cells towards the spheroids and some migration of the breast cancer cells from the spheroid towards the vascular barrier were

observed, Fig. 6e and S16†, indicating that cells can be introduced in LCD 3D printed microchannels for multi-day studies and they begin self-assembling into a micro-physiological system. While the current study is a proof-of-concept of an LCD 3D printed OoC device, future iterations would benefit from optimizing the ink for cell culture applications, and from optimizing the valve and central chamber geometry to ensure connectivity of endothelial sprouts with the spheroids.

### LCD 3D printers and digital mass manufacturing


We explored the potential for LCD 3D printers to mass produce OoC devices in a stacked array of  $19 \times 18 \times 10$  ( $X \times Y \times Z$ ) devices connected by short breakable supporting struts providing mechanical stability during printing. The OoCs were printed with a 50  $\mu\text{m}$  layer thickness which decreased print times. 3420 OoC devices were manufactured in an 8 h print run, Fig. 6f, and individual devices could be retrieved by breaking the struts. Given the vertical print range, we foresee that it would be possible to make >10 000 OoC devices in 24 h on a single 3D printer without user intervention. In consideration of the low capital cost of ~500 USD for a 12K LCD 3D printer, the possibility to start printing immediately upon receipt of a digital design file, and the minimal user intervention needed, the use of multiple such 3D printers could be attractive for on-demand mass production applications.

## Conclusion

We presented the use of low-cost photopolymerization LCD 3D printing for the fabrication of microfluidic devices using PLInk, optimized for rapid polymerization under low irradiance, 405 nm illumination, and reliable printing of embedded microchannels (and thin membranes) despite illumination inhomogeneity. The effect of ITX photoabsorber concentration on ink photopolymerization and notably  $D_p$  and  $E_c$  was characterized for 20  $\mu\text{m}$ -layer-by-layer printing. Posts with lateral resolution of 75  $\mu\text{m}$ , embedded membranes 22  $\mu\text{m}$  thin, and embedded microchannels with rectangular cross-sections of  $170 \times 220 \mu\text{m}^2$  and round cross-sections with 110  $\mu\text{m}$  radius were 3D printed. Further, we demonstrated that microfluidic devices previously made by other methods, such as laser machining, replica molding and DLP 3D printing, can now be fabricated using LCD 3D printing, including an embedded micromixer, a membrane microvalve, and an ELISA-chip for IFN- $\gamma$  detection. We also 3D printed an OoC device and demonstrated high throughput manufacturing by fabricating a 384-well plate format OoC within 1.5 h. Finally, we 3D printed 3420 OoC devices in an 8 h print run, and considering the vertical print capability, we anticipate the possibility of printing >10 000 OoC devices in a 24 h print run with just one 3D printer, demonstrating the potential for on-demand mass production. We note that the OoC devices and plates were conceptualized, designed and manufactured within 2 weeks







**Fig. 6** LCD 3D printed OoC device. (a) Design of the OoC device with two organ compartments seeded with a lung fibroblast spheroid (red) and breast cancer spheroid (green) and separated by a vascular barrier channel seeded with endothelial cells (orange). (b) Image of 3D printed OoC device loaded with dyed gel mimicking Matrigel. (c) Stereomicroscope images of OoC device cross-section showing the  $200 \times 200 \mu\text{m}^2$  embedded multi-layer stop valves. Scale bar =  $500 \mu\text{m}$ . (d) 384-well plate configuration of the OoC device printed in one run. The inset shows red and green dyed gel solutions confined by the capillary valve to their respective compartments. The yield was 100% across 3 plates with each 192 OoC devices. (e) Maximal intensity projection of confocal fluorescent microscopy z-stack over  $\sim 200 \mu\text{m}$  with  $10 \mu\text{m}$  increments featuring an OoC seeded with a five-day-old IMR-90 lung fibroblast spheroid (red) and a five-day-old MDA-MB-231 breast cancer spheroid (green), separated by a vascular barrier channel with HUVEC-embedded gel (orange), imaged at day 0 and day 5. The 3D printed stop-valve features are outlined in white dashed lines. Scale bar =  $300 \mu\text{m}$ . (f) Mass manufacturing of OoC devices connected by breakable struts printed as an array of  $19 \times 18 \times 10 = 3420$  devices ( $X \times Y \times Z$ ). Inset shows a zoomed-in view of stacked OoC devices. Scale bar =  $5 \text{ mm}$ .



following reviewers' comments, which illustrates the advantages and potential of 3D LCD printing (and PLink) for rapidly exploring new ideas and concepts for microfluidic and OoC devices.

Future work could explore the concurrent variation of TPO photoinitiator, ITX photoabsorber, and PETTA crosslinker concentration to better understand their interplay with regards to  $D_p$ ,  $E_c$ , and printing accuracy, and choose the optimal mixture based on a specific application and criteria. Photoactive components that better match LCD 3D printer light spectrum, especially at higher wavelength between ~420–450 nm could help improve photocuring efficiency and reduce exposure time; however, increased light adsorption at higher wavelength is expected to come at the expense of yellow-orange tinted devices compared to visually overall neutral and transparent PLink. In addition, mapping the light heterogeneity of the LCD 3D printers by the end user, and the tools to do that, would open the door to digital correction of the illumination heterogeneity by programming the 3D printer, and further improve the resolution achievable both with commercial and custom photoinks. Finally, low-cost commercial inks that are easily accessible to the end-user (e.g., water-washable inks), but which are often viscous and suffer from overly large  $D_p$  for high-resolution microfluidic 3D printing, might be improved simply by supplementing them with additives. The addition of (i) solvents such as PEGDA-250 to reduce the viscosity, (ii) photoabsorbers to reduce the light penetration, and (iii) photoinitiators to reduce the exposure time could all be explored. The biocompatibility of commercial inks needs to be evaluated and possibly improved by testing different washing and leaching conditions towards removing residual toxic components from the cured parts.

We may expect that 3D stereolithography printer manufacturers driven by technological advances in LCDs (and light engines), and market pressure, will continue to increase pixel numbers and concomitantly reduce pixel size, all while preserving the affordability of LCD 3D printers, which will further increase their appeal and adoption. We foresee that some of the greatest opportunities lie in improving the photoinks for LCD (and more generally stereolithography) 3D printing, which are in their infancy. While here we showed the application of LCD 3D printing to microfluidics that were primarily designed based on prior manufacturing technologies, opportunities arise to re-design and ideate microfluidic systems and OoC devices that leverage the strength of LCD and stereolithography 3D printing.

Digital manufacturing by LCD 3D printing is as simple as downloading a file and printing it, thus circumventing the need for specialized machinery and advanced training, while enabling customizability and rapid design iterations by the end user. The advent of low-cost and easy-to-use 3D printers compared to traditional manufacturing methods enables the fabrication of open and embedded microscopic features by anyone, anywhere, thus democratizing access to high-resolution fabrication and reducing the entry barrier for many

potential users. The combination of low-cost, high-resolution 3D printers, and readily 3D printable designs enable the realization of low-cost and distributed digital manufacturing.

## Data availability

3D design files are uploaded to Thingiverse and Printables (<https://www.thingiverse.com/junckerlab/collections> and [https://www.printables.com/@JunckerLab\\_743461](https://www.printables.com/@JunckerLab_743461)).

Data not presented in the article or ESI† material will be available upon request.

## Author contributions

V. K. designed the ink formulation, V. K. and H. S. characterized the ink and assessed printability for LCD 3D printing. M. L. S. performed the biocompatibility assay. H. S., G. K., and Y. M. printed and characterized the micromixer. G. K. and H. S. printed and characterized the microvalve. A. S.-K. contributed to the embedded sequential delivery chip. H. S. designed and optimized the ELISA-chips. H. S. and M. L. S. designed and optimized the organ-on-a-chip devices. H. S. prepared the figures and analyzed the data. H. S., V. K., and D. J. prepared the initial draft of the manuscript, with feedback from all the authors. H. S. and D. J. edited and reviewed the manuscript drafts. D. J. provided funding, and conceptualized, supervised, and administered the project.

## Conflicts of interest

The authors have no conflicts of interest to declare.

## Acknowledgements

We acknowledge Yongjun Xiao for assistance with  $\mu$ CT at the Centre for Bone and Periodontal Research, Mohsen Ketabi and Gwénaél Chamoulaud for assistance in profilometry, viscometry, and FTIR-ATR at the NanoQAM Research Center, and Lucie Riffard for assistance in tensile testing at the McGill Structures and Composite Materials Laboratory. H. S. acknowledges a CGS-M CIHR, an FRQNT master's research scholarship and a McGill BME recruitment award. V. K. acknowledges an FRQNT doctoral research scholarship. G. K. acknowledges an FRQNT doctoral research scholarship and a McGill BME recruitment award. Y. M. acknowledges a McGill BME recruitment award. A. S.-K. acknowledges FRQNT postdoctoral fellowship. D. J. acknowledges support from a Canada Research Chair in Bioengineering.

## References

- 1 G. M. Whitesides, *Nature*, 2006, **442**, 368–373.
- 2 E. K. Sackmann, A. L. Fulton and D. J. Beebe, *Nature*, 2014, **507**, 181–189.
- 3 G. M. Whitesides and A. D. Stroock, *Phys. Today*, 2001, **54**, 42–48.



- 4 G. Trotta, A. Volpe, A. Ancona and I. Fassi, *J. Manuf. Process.*, 2018, **35**, 107–117.
- 5 D. J. Guckenberger, T. E. De Groot, A. M. Wan, D. J. Beebe and E. W. Young, *Lab Chip*, 2015, **15**, 2364–2378.
- 6 A. Naderi, N. Bhattacharjee and A. Folch, *Annu. Rev. Biomed. Eng.*, 2019, **21**, 325–364.
- 7 V. Karamzadeh, A. Sohrabi-Kashani, M. Shen and D. Juncker, *Adv. Mater.*, 2023, **35**, 2303867.
- 8 B. Heidt, R. Rogosic, S. Bonni, J. Passariello-Jansen, D. Dimech, J. W. Lowdon, R. Arreguin-Campos, E. Steen Redeker, K. Eersels, H. Diliën, B. van Grinsven and T. J. Cleij, *Phys. Status Solidi A*, 2020, **217**, 1900935.
- 9 Z.-X. Low, Y. T. Chua, B. M. Ray, D. Mattia, I. S. Metcalfe and D. A. Patterson, *J. Membr. Sci.*, 2017, **523**, 596–613.
- 10 B. Carnero, C. Bao-Varela, A. I. Gómez-Varela, E. Álvarez and M. T. Flores-Arias, *Mater. Sci. Eng., C*, 2021, **129**, 112388.
- 11 A. P. Kuo, N. Bhattacharjee, Y. S. Lee, K. Castro, Y. T. Kim and A. Folch, *Adv. Mater. Technol.*, 2019, **4**, 1800395.
- 12 H. Gong, B. P. Bickham, A. T. Woolley and G. P. Nordin, *Lab Chip*, 2017, **17**, 2899–2909.
- 13 J. L. Saorin, M. D. Diaz-Alemán, J. De la Torre-Cantero, C. Meier and I. Pérez Conesa, *Appl. Sci.*, 2021, **11**, 3197.
- 14 Y. T. Kim, A. Ahmadianyazdi and A. Folch, *Nat. Protoc.*, 2023, **18**, 1243–1259.
- 15 C. Warr, J. C. Valdoz, B. P. Bickham, C. J. Knight, N. A. Franks, N. Chartrand, P. M. Van Ry, K. A. Christensen, G. P. Nordin and A. D. Cook, *ACS Appl. Bio Mater.*, 2020, **3**, 2239–2244.
- 16 V. Karamzadeh, M. Shen, F. Lussier and D. Juncker, *bioRxiv*, 2023, preprint, DOI: [10.1101/2023.11.27.568937](https://doi.org/10.1101/2023.11.27.568937).
- 17 A. Olanrewaju, M. Beaugrand, M. Yafia and D. Juncker, *Lab Chip*, 2018, **18**, 2323–2347.
- 18 M. Yafia, O. Ymbern, A. O. Olanrewaju, A. Parandakh, A. Sohrabi Kashani, J. Renault, Z. Jin, G. Kim, A. Ng and D. Juncker, *Nature*, 2022, **605**, 464–469.
- 19 A. Parandakh, O. Ymbern, W. Jogia, J. Renault, A. Ng and D. Juncker, *Lab Chip*, 2023, **23**, 1547–1560.
- 20 B. W. Caplins, C. I. Higgins, T. J. Kolibaba, U. Arp, C. C. Miller, D. L. Poster, C. J. Zarobila, Y. Zong and J. P. Killgore, *Addit. Manuf.*, 2023, **62**, 103381.
- 21 P. Somers, A. Münchinger, S. Maruo, C. Moser, X. Xu and M. Wegener, *Nat. Rev. Phys.*, 2024, **6**(2), 99–113.
- 22 H. Gong, M. Beauchamp, S. Perry, A. T. Woolley and G. P. Nordin, *RSC Adv.*, 2015, **5**, 106621–106632.
- 23 N. Guselnikov, P. Lazarev, M. Paukshto and P. Yeh, *J. Soc. Inf. Disp.*, 2005, **13**, 339–348.
- 24 I. Ahmed, K. Sullivan and A. Priye, *Biosensors*, 2022, **12**, 652.
- 25 V. G. Colin, T. Travers, D. Gindre, R. Barillé and M. Loumagne, *Int. J. Adv. Manuf. Tech.*, 2021, **116**, 537–544.
- 26 M. G. Mohamed, H. Kumar, Z. Wang, N. Martin, B. Mills and K. Kim, *J. Manuf. Mater. Process.*, 2019, **3**, 26.
- 27 A. Vedhanayagam, M. Golfetto, J. L. Ram and A. S. Basu, *Micromachines*, 2023, **14**, 1519.
- 28 A. Hayer, L. Shao, M. Chung, L.-M. Joubert, H. W. Yang, F.-C. Tsai, A. Bisaria, E. Betzig and T. Meyer, *Nat. Cell Biol.*, 2016, **18**, 1311–1323.
- 29 M. Boaks, C. Roper, M. Viglione, K. Hooper, A. T. Woolley, K. A. Christensen and G. P. Nordin, *Micromachines*, 2023, **14**, 1589.
- 30 H. Gong, A. T. Woolley and G. P. Nordin, *Lab Chip*, 2016, **16**, 2450–2458.
- 31 P. F. Jacobs, *Rapid prototyping & manufacturing: fundamentals of stereolithography*, Society of Manufacturing Engineers, 1992.
- 32 Y. Xu, F. Qi, H. Mao, S. Li, Y. Zhu, J. Gong, L. Wang, N. Malmstadt and Y. Chen, *Nat. Commun.*, 2022, **13**, 918.
- 33 R. Paul, Y. Zhao, D. Coster, X. Qin, K. Islam, Y. Wu and Y. Liu, *Nat. Commun.*, 2023, **14**, 4520.
- 34 N. Rekowska, J. Huling, A. Brietzke, D. Arbeiter, T. Eickner, J. Konasch, A. Riess, R. Mau, H. Seitz and N. Grabow, *Pharmaceutics*, 2022, **14**, 628.
- 35 S. M. Montgomery, F. Demoly, K. Zhou and H. J. Qi, *Adv. Funct. Mater.*, 2023, **33**, 2213252.
- 36 S. M. Montgomery, C. M. Hamel, J. Skovran and H. J. Qi, *Extreme Mech. Lett.*, 2022, **53**, 101714.
- 37 J. Qi, W. Li, W. Chu, J. Yu, M. Wu, Y. Liang, D. Yin, P. Wang, Z. Wang, M. Wang and Y. Cheng, *Micromachines*, 2020, **11**, 213.
- 38 W. Ruijin, L. Beiqi, S. Dongdong and Z. Zefei, *Sens. Actuators, B*, 2017, **249**, 395–404.
- 39 N. Li, C. H. Hsu and A. Folch, *Electrophoresis*, 2005, **26**, 3758–3764.
- 40 B. Mosadegh, C.-H. Kuo, Y.-C. Tung, Y.-s. Torisawa, T. Bersano-Begey, H. Tavana and S. Takayama, *Nat. Phys.*, 2010, **6**, 433–437.
- 41 Y.-S. Lee, N. Bhattacharjee and A. Folch, *Lab Chip*, 2018, **18**, 1207–1214.
- 42 G. Kak, M. Raza and B. K. Tiwari, *Biomol. Concepts*, 2018, **9**, 64–79.
- 43 M. Pai, C. M. Denking, S. V. Kik, M. X. Rangaka, A. Zwerling, O. Oxlade, J. Z. Metcalfe, A. Cattamanchi, D. W. Dowdy and K. Dheda, *Clin. Microbiol. Rev.*, 2014, **27**, 3–20.
- 44 M. Pai and M. Behr, *Microbiol. Spectrum*, 2016, **4**(5), DOI: [10.1128/microbiolspec.TB2-0023-2016](https://doi.org/10.1128/microbiolspec.TB2-0023-2016).
- 45 W. Luo, M. Pla-Roca and D. Juncker, *Anal. Chem.*, 2011, **83**, 5767–5774.
- 46 G. Taguchi, S. Chowdhury and Y. Wu, *Taguchi's quality engineering handbook*, 2004.
- 47 D. Huh, B. D. Matthews, A. Mammoto, M. Montoya-Zavala, H. Y. Hsin and D. E. Ingber, *Science*, 2010, **328**, 1662–1668.
- 48 I. Salmon, S. Grebenyuk, A. R. Abdel Fattah, G. Rustandi, T. Pilkington, C. Verfaillie and A. Ranga, *Lab Chip*, 2022, **22**, 1615–1629.

