

PAPER

View Article Online
View Journal | View Issue

Cite this: *Nanoscale Adv.*, 2024, 6, 3644

Core-shell niobium(v) oxide@molecularly imprinted polythiophene nanoreceptors for transformative, real-time creatinine analysis†

Zohaib Saddique,^a Maleeha Saeed,^a Muhammad Faheem,^b Sadia Z. Bajwa,^c Adnan Mujahid^{ID}^a and Adeel Afzal^{ID}^{*a}

Creatinine, a byproduct of muscle metabolism, is typically filtered by the kidneys. Deviations from normal concentrations of creatinine in human saliva serve as a crucial biomarker for renal diseases. Monitoring these levels becomes particularly essential for individuals undergoing dialysis and those with kidney conditions. This study introduces an innovative disposable point-of-care (PoC) sensor device designed for the prompt detection and continuous monitoring of trace amounts of creatinine. The sensor employs a unique design, featuring a creatinine-imprinted polythiophene matrix combined with niobium oxide nanoparticles. These components are coated onto a screen-printed working electrode. Thorough assessments of creatinine concentrations, spanning from 0 to 1000 nM in a redox solution at pH 7.4 and room temperature, are conducted using cyclic voltammetry (CV), differential pulse voltammetry (DPV), and electrochemical impedance spectroscopy (EIS). The devised sensor exhibits a sensitivity of 4.614 $\mu\text{A cm}^{-2} \text{ nM}^{-1}$, an impressive trace level limit of detection at 34 pM, and remarkable selectivity for creatinine compared to other analytes found in human saliva, such as glucose, glutamine, urea, tyrosine, etc. Real saliva samples subjected to the sensor reveal a 100% recovery rate. This sensor, characterized by its high sensitivity, cost-effectiveness, selectivity, and reproducibility, holds significant promise for real-time applications in monitoring creatinine levels in individuals with kidney and muscle-related illnesses.

Received 9th April 2024

Accepted 23rd May 2024

DOI: 10.1039/d4na00300d

rsc.li/nanoscale-advances

Introduction

Creatinine, a metabolic byproduct of muscle activity during typical human functions, results from the catabolism of creatine and phosphocreatine, releasing energy in the form of adenosine triphosphate (ATP). Produced by muscles, transported through the bloodstream, filtered by the kidneys, and eventually eliminated through urine, abnormal levels of creatinine are indicative biomarkers for the optimal functioning of these organs.¹ The normal concentration of creatinine in human saliva typically falls within the range of 4.4 to 17.7 μM , subject to influences such as gender, body weight, and physiological conditions.^{2,3} Fluctuations in salivary creatinine levels play a pivotal role as biomarkers for renal diseases; elevated concentrations ($>28 \mu\text{M}$) signal kidney disorders, while lower concentrations ($<2 \mu\text{M}$) signify muscular dysfunction.^{3,4}

Given the significance of creatinine as a biomarker for conditions such as renal failure, chronic kidney disorder (CKD), and muscle dystrophy, it ranks as the second most analyzed biomolecule.⁵ Currently, there is no point-of-care (PoC) device that can readily detect creatinine in blood or saliva samples. The importance of timely creatinine monitoring in humans, particularly through PoC diagnostics, is highlighted by global fatalities due to renal failure. In this context, a human saliva sample is more advantageous in terms of convenience, non-invasiveness, and lower risk over urine and serum samples for PoC diagnostics.^{6,7} Additionally, the collection of saliva is easy and does not require a skilled individual, while posing a minimal risk of contamination or disease transmission.^{8,9}

Although Jaffe's method has traditionally been used for creatinine detection, its drawbacks, including interference from other molecules, the need for skilled operators, higher costs, and incompatibility with PoC diagnostics, have prompted the exploration of alternative methods.¹⁰ For instance, Bai *et al.*¹¹ developed a PoC colorimetric metabolomic sensor for different metabolites, including creatinine. Over recent decades, electrochemical methods have been developed for creatinine quantification in physiological fluids.^{12–14} These methods stand out for their excellent sensitivity and rapid response, offering considerable potential for real-time applications. However, it is important to overcome the challenges of non-selectivity and

^aSensors and Diagnostics Lab, School of Chemistry, University of the Punjab, Quaid-i-Azam Campus, Lahore, 54590, Pakistan. E-mail: adeel.chem@pu.edu.pk

^bSchool of Chemistry and Chemical Engineering, Jiangsu University, Zhenjiang, Jiangsu, 212013, P. R. China

^cNational Institute for Biotechnology and Genetic Engineering, PO Box 577, Jhang Road, Faisalabad, 38000, Pakistan

† Electronic supplementary information (ESI) available. See DOI: <https://doi.org/10.1039/d4na00300d>


interferences in complex mixtures of nanomaterial-based electrochemical sensors.³

Molecularly imprinted polymers (MIP) serve as synthetic counterparts to enzymes and natural antibodies, offering biomimetic selectivity, greater stability, and cost-effectiveness.¹⁵ MIPs are suitable for developing PoC diagnostics, as they provide low cost and long-term results.^{16,17} Incorporating nanomaterials into MIP-based sensors enhances sensitivity, aligning the strengths of nanomaterials and MIP for optimal results.^{18,19} In our earlier works, we employed $g\text{-C}_3\text{N}_4$,¹⁹ $\text{MoO}_3\cdot\text{V}_2\text{O}_5$,²⁰ or V-TiO_2 ²¹ for the enhancement of the sensitivity of electrochemical sensors. Pristine metal oxide nanoparticles with higher metal oxidation states, such as Nb_2O_5 , offer preferential reactivity toward hard bases such as creatinine, containing an R-NH_2 groups, due to their hard acid characteristics.²²

The merits and demerits of electrochemical sensors for creatinine detection are extensively discussed in prior works, outlining key limitations and proposing future recommendations.^{3,5} Alizadeh *et al.*²³ used an innovative approach to design MIP with a Cu(II) -creatinine complex as the template. Thus-fabricated electrochemical sensor could detect creatinine down to 59 nM threshold detection limit (LOD) with a wide range linear response. Recently, Ashakirin *et al.*²⁴ reported an amperometric $\text{Cu}_2\text{O@MIP}$ nanocomposite creatinine sensor, with a 22 nM LOD, and $2.16 \mu\text{A cm}^{-2} \text{ nM}^{-1}$ sensitivity.

Herein, the design of $\text{Nb}_2\text{O}_5\text{@MIP}$ nanoreceptors for real-time monitoring of salivary creatinine levels is presented. $\text{Nb}_2\text{O}_5\text{@MIP}$ nanoreceptors are synthesized *in situ* via a chemical oxidative polymerization of thiophene to form inverse core-shell niobium(v) oxide nanoparticles at conductive, molecularly imprinted polythiophene (MIP). The as-synthesized $\text{Nb}_2\text{O}_5\text{@MIP}$ nanoreceptors are fabricated on cost-effective, disposable,

screen-printed electrodes (SPE) for the electrochemical recognition of creatinine and analysis of human saliva samples (Fig. 1).

To understand the composition-structure-sensing properties relations, pristine MIP, non-imprinted polythiophene (NIP), and $\text{Nb}_2\text{O}_5\text{@NIP}$ sensors are also fabricated. These sensors are exposed to varying concentrations (0–1000 nM) of creatinine, with responses recorded through different electrochemical techniques including cyclic voltammetry (CV), differential pulse voltammetry (DPV), and electrochemical impedance spectroscopy (EIS). The reliability of the sensor for real-time applications is validated against different human saliva samples as a complex mixture of various molecules.

Experimental section

Synthesis

To prepare core-shell $\text{Nb}_2\text{O}_5\text{@MIP}$ nanoreceptors, 410 mg of niobium(v) oxide nanoparticles, and 164 mg of sodium dodecyl sulfate (SDS, $\geq 98.5\%$, SigmaAldrich) were dissolved in 20 mL of deionized water and were homogenized under ultrasound sonication for 45 min. After that, 117 mg of creatinine (anhydrous, $\geq 98\%$, SigmaAldrich) and 840 μL of thiophene (Th, $>98\%$, SigmaAldrich) were added to the solution. The reaction mixture was again homogenized for 30 min to self-organize the monomer, the template, and Nb_2O_5 nanoparticles. Separately, another solution containing 826 mg of the catalyst, copper sulfate pentahydrate ($\text{CuSO}_4\cdot 5\text{H}_2\text{O}$, SigmaAldrich), was prepared in 2 mL of the oxidant, hydrogen peroxide (H_2O_2 , 35%, SigmaAldrich). Thus, *in situ*, oxidative polymerization was initiated by the addition of the later solution to the reaction mixture containing self-organized monomers, template, and Nb_2O_5 nanoparticles, which changed the color of the reaction




Fig. 1 A schematic of the synthesis and fabrication of $\text{Nb}_2\text{O}_5\text{@MIP}$ nanoreceptors on the surface of disposable gold screen-printed electrodes (SPE). The $\text{Nb}_2\text{O}_5\text{@MIP}$ sensor is applied for real-time monitoring of salivary creatinine, a biomarker for chronic kidney disorder (CKD), using electrochemical impedance spectroscopy and voltammetric measurements shown in the inset.



mixture to dark brown. The resulting reaction mixture was heated at 50 °C for 7 hours under continuous magnetic stirring. Dark brown precipitates were obtained which were washed with deionized water and ethanol to remove unreacted species. The product (Nb₂O₅@MIP) was oven-dried at 60 °C for 12 hours. Similarly, Nb₂O₅@NIP nanoreceptors were prepared without creatinine (the template). Also, pristine MIP was prepared as a reference without the inclusion of Nb₂O₅ nanoparticles, while NIP was prepared without the addition of Nb₂O₅ nanoparticles and creatinine (the template).

Characterization

Powder X-ray diffraction (PXRD) analysis of Nb₂O₅ nanoparticles was performed on a Siemens D5005 X-ray diffractometer, equipped with Cu K α ($\lambda = 1.5418$ Å) irradiation, employing a scanning rate of 5° min⁻¹. Nb₂O₅ nanoparticles and Nb₂O₅@NIP and Nb₂O₅@MIP nanoreceptors were analyzed by Fourier transform infrared (FTIR) spectroscopy using a Nicolet-410-FTIR spectrophotometer, across 4000 to 650 cm⁻¹. Transmission electron micrographs (TEM) of Nb₂O₅@MIP nanoreceptors were obtained with a HITACHI-2100 PLUS TEM, at an acceleration voltage of 120 kV. Scanning electron micrographs (SEM) were obtained with a HITACHI SU8010 scanning electron microscope.

Sensor fabrication

Four different electrochemical sensors were fabricated using NIP, MIP, Nb₂O₅@NIP, and Nb₂O₅@MIP nanoreceptors. First, as-synthesized nanoreceptors were immobilized in a 1% polystyrene (packaging waste, local) solution in tetrahydrofuran (THF; $\geq 99.9\%$, anhydrous, inhibitor-free, SigmaAldrich). The amount of nanoreceptors was optimized to be 1 mg mL⁻¹. The immobilization mixture was sonicated for 30 min to homogenize the suspension. All samples were fabricated on disposable, Italsens screen-printed gold (Au) electrodes (gold SPE), provided by PalmSens BV, The Netherlands. The diameter of the working electrode was 3 mm, with a geometrical area of 7.07 mm². Before the fabrication of each sensor, SPEs were washed with water, ethanol, and acetone to remove any contaminants. Subsequently, 5 μ L of the nanoreceptors' suspension was drop-coated on the electrode surface. All devices were air-dried at room temperature and stored in sealed containers for measurements. Before electrochemical characterization and sensor measurements, MIP and Nb₂O₅@MIP sensors were thoroughly washed with deionized water to remove the template. The removal of template, creatinine molecules, is characterized by FTIR spectroscopy, as shown in Fig. S1.†

Electrochemical characterization and sensor measurements

The electrochemical characterization and sensor responses of fabricated creatinine sensors were measured with a PalmSens 4, a USB-powered, compact potentiostat, operated *via* PSTrace Windows software, provided by PalmSens BV, The Netherlands. For this purpose, firstly, a redox solution was prepared by dissolving equimolar concentrations (2.5 mM) of K₃[Fe(CN)₆]·3H₂O (ACS reagent, 98.5–102.0%, SigmaAldrich) and

K₃[Fe(CN)₆] (ACS reagent, $\geq 99.0\%$, SigmaAldrich) in phosphate buffer saline (PBS, pH 7.4, BioPLUS), prepared in deionized water. To enhance electron transfer, 0.05 M potassium chloride (KCl; analytical grade, local) was also added to the redox solution.

EIS was used to investigate the interfacial electrochemical characteristics of different nanoreceptors. EIS measurements were performed in the frequency range of 1 Hz to 100 kHz with DC potential (E_{dc}) of 0.1 V and AC potential (E_{ac}) of 0.006 V while recording 50 frequencies per 9.8 decades. The electrochemical responses of the NIP, MIP, Nb₂O₅@NIP, and Nb₂O₅@MIP sensors were investigated by CV and DPV when exposed to varying concentrations of creatinine ranging from 0 to 1000 nM. Different concentrations (0–1000 nM) of creatinine were prepared in the above-mentioned redox solution. CV measurements were conducted in the potential range of -0.8 to $+0.8$ V at a scan rate of 100 mV s⁻¹ with a potential step (E step) of 10 mV. To optimize the scan rate and determine the mechanisms, the effect of different scan rates (10–100 mV) was also studied. DPV responses of electrochemical creatinine sensors were recorded in a potential window of -0.5 to $+0.5$ V with a scan rate of 100 mV s⁻¹. Peak current shifts with changing potential due to oxidation-reduction of [Fe(CN)₆]^{3-/4-} redox pairs in PBS were used as bases for CV and DPV analysis.

From these electrochemical measurements, the sensitivity, limit of detection (LOD), and limit of quantification (LOQ) were calculated by employing standard methods, as described elsewhere.²⁰ The selectivity was tested by exposing the Nb₂O₅@MIP sensor to 100 nM of common analytes present in human saliva, such as urea, uric acid, tyrosine, glutamine, guanine, spermine, *etc.* To determine the reproducibility of Nb₂O₅@MIP sensors, three devices were separately fabricated and tested at a known concentration of creatinine. The mean values of sensor response and standard deviation were subsequently calculated.

Pre-treatment and analysis of human saliva

Real saliva samples were obtained from three healthy members of the Sensors and Diagnostics Laboratory, with their prior consent. Saliva samples were collected in 2 mL clean plastic vials *via* passive drooling. The samples were cooled down to 4 °C by placing them in an ice bath. To remove food particles, the samples were centrifuged at 4000 rpm. The supernatant was collected and treated in an ultrasonic bath to break down long-chain muco-polypeptides. These molecules impart viscosity and make handling difficult. Each sample was subsequently filtered through a 220 nm nylon syringe filter to remove large protein bodies and other irrelevant macromolecules. 400 μ L of the saliva samples were diluted with 800 μ L of the redox solution, to realize the same concentrations of the [Fe(CN)₆]^{3-/4-} redox pairs. Then, the saliva samples were tested to determine the initial concentration of creatinine in each sample. Separately, all saliva samples were spiked with 200 nM creatinine to investigate the differences in creatinine concentration and estimate the percent recovery of creatinine, as described elsewhere.²⁵ During the measurements, the saliva samples were stored below 8 °C.



Results and discussion

Fabrication of Nb₂O₅@MIP sensors

Fabrication of the Nb₂O₅@MIP sensor is monitored by EIS and CV. The interfacial properties of the sensors with changes in electrochemical characteristics before and after modification of the bare, gold SPE with Nb₂O₅@MIP nanoreceptors are investigated. Firstly, modification of the SPE exponentially reduces charge transfer resistance (R_{ct}) from 14.74 K Ω to 203 Ω , as demonstrated by the Nyquist plots (Fig. S2a†). Lower R_{ct} represents lower resistance at the Nb₂O₅@MIP-electrolyte interface leading to faster electron transfer and improved electrocatalytic properties. A similar trend is observed in the Bode magnitude plot (Fig. S2b†), featuring variation in absolute impedance as a function of frequency. It exhibits higher absolute impedance for bare electrodes as compared to Nb₂O₅@MIP modified SPE owing to the absence of ionic interactions on the bare electrode surface.²⁶ These results are attested by Bode phase angle plots (Fig. S2c†), where the Nb₂O₅@MIP modified electrode shows deviation toward 40° demonstrating capacitive properties with surface defects and leakage current.²⁵ These defect sites pertain to imprints and cavities that may act as diffusion channels in Nb₂O₅@MIP nanoreceptors, promoting the electrocatalytic activity for sensitive detection of the target analyte. This is manifested in the greater voltammetric response of the Nb₂O₅@MIP modified electrode (Fig. S2d†). CV results corroborate earlier conclusions from EIS analysis.

The electroactive surface area (A) and heterogeneous rate constant (K^0) emerge as pivotal factors distinguishing the electrocatalytic features, based on the presence of interaction sites and it is calculated through the Randles–Ševčík equation (eqn (1)).¹⁹

$$I_p = 2.69 \times 10^5 n^{\frac{3}{2}} A D^{\frac{1}{2}} C v^{\frac{1}{2}} \quad (1)$$

The heterogeneous rate constant, on the other hand, is calculated by following equation (eqn (2)).²⁷

$$K^0 = RT/R_{ct} A C F^2 \quad (2)$$

The calculated electroactive area for Nb₂O₅@MIP is 43.87 $\times 10^{-3}$ cm², which is 24.5-fold higher than 1.79 $\times 10^{-3}$ cm² for the bare electrode (Fig. S2e†). The heterogeneous rate constant, an indicator of the electron transfer kinetics at the electrode–electrolyte interface, signifies faster electron transfer for the Nb₂O₅@MIP-modified electrode. Calculated at 1.20 $\times 10^{-2}$ cm s^{−1}, it also outperforms the bare electrode at 4.04 $\times 10^{-3}$ cm s^{−1}, as illustrated in (Fig. S2f†). In summary, these results confirm the fabrication of Nb₂O₅@MIP nanoreceptors and their superior electrocatalytic properties.

Microstructure and surface morphology

The revelations from TEM micrographs, depicted in Fig. 2a–d, reveal the intriguing architecture of Nb₂O₅@MIP nanostructures. These images suggest the existence of an inverted core–shell structure of Nb₂O₅@MIP. However, Nb₂O₅ nanoparticles do not form a complete shell due to the presence and removal of template molecules by a thorough washing step. 2D-SEM images

of the sensor surface, shown in Fig. 2e and f, demonstrate a consistent distribution of irregularly shaped clusters of Nb₂O₅@MIP. Moving to the 3D perspective (Fig. 2g), a highly rough surface of the Nb₂O₅@MIP-modified electrode can be seen, an attribute desired for an effective electrocatalytic interface, maximizing the surface area available for the electrode–analyte interactions. The surface profile (Fig. 2h) illustrates the presence of peaks and valleys, while the size distribution of surface features is presented in the histogram (Fig. 2i). This topographical intricacy aligns with the insights from the Bode phase angle plot, where the observed current leakage and surface defects, attributed to molecular imprinting cavities, are comprehensible. In principle, Nb₂O₅@MIP-modified electrodes present morphology and structural features finely tuned for electrocatalytic applications.

Structural characterization

In Fig. 3a and b, the FTIR spectra of Nb₂O₅, Nb₂O₅@NIP, and Nb₂O₅@MIP are presented. The distinctive fingerprint region for polythiophene, spanning 600–1650 cm^{−1}, unfolds peaks corresponding to ν C–C ring stretching vibrations, ν C _{α} –C _{α} rings resonance absorption, and ν C=C symmetric vibration at approximately 1072 cm^{−1}, 1123 cm^{−1}, and around 1650 cm^{−1}, respectively.^{28–30} Nb₂O₅ introduces a less intense, co-linear Nb–O–Nb stretching peak at around 800 cm^{−1}, evident in its FTIR spectrum. The inclusion of Nb₂O₅ in Nb₂O₅@NIP and Nb₂O₅@MIP manifests as a less intense Nb–O–Nb stretching peak at approximately 800 cm^{−1}.^{31,32} The remainder of the spectra for Nb₂O₅@NIP and Nb₂O₅@MIP are anticipated to be consistent, given their equivalent quantities of monomers and other reacting reagents. However, a weak ν C–N stretching is observed around 1215 cm^{−1} in Nb₂O₅@MIP, which is attributed to the presence of creatinine molecules in the polymeric matrix.³³ The absence of this peak in Nb₂O₅@NIP is therefore justified. Furthermore, washing off the template or creatinine molecules leads to a significant decrease in the peak intensity (Fig. S1†), which also confirms the removal of template from the polymeric matrix.

In the XRD pattern (Fig. 3c), Nb₂O₅ unveils a monoclinic crystal system with space group $P1\ 2/m\ 1$, indexed as JCPDS. 96-210-7338. Distinct diffraction peaks at 17.30°, 19.18°, 22.54°, 23.75°, 24.44°, 26.55°, 28.32°, 31.59°, 32.23°, 33.09°, 35.24°, 36.018°, and 38.88° 2θ , along with their hkl Miller indices (103), (403), (105), (011), (504), (112), (604), (213), (114), (513), (308), (700), and (610), respectively, delineate the crystal structure.^{34,35} Employing the Debye–Scherrer equation, the crystallite size of Nb₂O₅ is calculated to be 40 nm. The monoclinic system's expanded surface area, coupled with the nanoscale size, positions Nb₂O₅ as a highly sought-after semiconductor for enhancing the sensitivity of electrochemical sensors.^{36,37}

Effect of scan rate

CV was performed to examine the impact of scan rate (10–100 mV s^{−1}) on anodic (I_{pa}) and cathodic (I_{pc}) peak currents, elucidating mass transfer mechanisms, and electron transfer kinetics at the electrode–electrolyte interface. Fig. 3d–k portray the escalating redox peak current with increasing scan rate when different sensors are exposed to a redox solution in PBS at






Fig. 2 Morphological characterization of Nb₂O₅@MIP: (a–d) TEM images (scale bar indicates 200 nm); (e) SEM image of the sensor surface (scale bar indicates 2 μm); (f) 2D, and (g) 3D high-resolution colored SEM image of the sensor surface; (h) surface profile, and (i) surface topography maps with (j) height scale (z-axis).

pH 7.4. Linearity in the increase in I_{pa} with ascending values of the scan rate square root ($\nu^{1/2}$) is mathematically expressed, confirming our premise that this process is diffusion-limited.^{38,39} The results of the linear regression analysis for different sensors are given as follows:

Pristine NIP:

$$I_{pa} = (0.345 \pm 0.0245) \nu^{1/2} + (3.079 \pm 0.181); (R^2 = 0.96)$$

Pristine MIP:

$$I_{pa} = (1.215 \pm 0.084) \nu^{1/2} - (0.648 \pm 0.627); (R^2 = 0.96)$$

Nb₂O₅@NIP:

$$I_{pa} = (4.189 \pm 0.312) \nu^{1/2} - (2.108 \pm 2.31); (R^2 = 0.95)$$

Nb₂O₅@MIP:

$$I_{pa} = (11.74 \pm 0.244) \nu^{1/2} - (29.46 \pm 1.8); (R^2 = 0.99)$$

A similarly linear but opposite trend is observed in the case of I_{pc} , which confirms the reversibility of the electron-transfer reactions, and their diffusion-limiting nature. Furthermore, the positive shift in oxidation peak potential and the negative shift in reduction peak potential with increasing scan rate indicate






Fig. 3 (a) Functional group characterization of Nb_2O_5 , $\text{Nb}_2\text{O}_5\text{@NIP}$, and $\text{Nb}_2\text{O}_5\text{@MIP}$ through FTIR spectra, and (b) magnified view of FTIR spectra to demonstrate the presence of Nb_2O_5 nanoparticles. (c) XRD pattern of Nb_2O_5 nanocrystals. Effect of scan rate (d, f, h and j) and the respective calibration curves (e, g, i and k) showing anodic, and cathodic peak currents (I_{pa} and I_{pc}) as a function of the square root of the scan rate for (d and e) NIP, (f and g) MIP, (h and i) $\text{Nb}_2\text{O}_5\text{@NIP}$, and (j and k) $\text{Nb}_2\text{O}_5\text{@MIP}$, respectively. CV scans were recorded by exposing the electrodes to a redox solution with different scan rates (10–100 mV s^{-1}). All measurements are performed in a redox solution containing $[\text{Fe}(\text{CN})_6]^{3-/4-}$ (2.5 mM), KCl (0.05 M) in PBS (pH 7.4).

the occurrence of quasi-reversible redox reactions at the surface of the modified electrodes.⁴⁰

Electrochemical characterization

The deployment of EIS is a powerful tool for delving into the interfacial interactions of the fabricated sensor materials. All samples are tested to reveal their distinctive interfacial properties. In Fig. 4a–c, a side-by-side comparison of NIP and MIP is presented through Nyquist, Bode magnitude, and Bode phase plots. In the Nyquist plot, where real (Z') and imaginary impedance (Z'') are plotted, a large semi-circle area appears in the case of NIP, signifying an elevated polarization resistance at the electrode–electrolyte interface (R_{ct}), quantified at a substantial 6.64 k Ω . Contrarily, the MIP showcases a smaller semi-circle area, an outcome of the introduction of diffusion channels and selective interaction sites that substantially reduce R_{ct} to 1.21 k Ω . The NIP coatings are generally devoid of imprinted cavities and diffusion channels, whereas the MIP surface possesses imprints or cavities generated by the removal of template molecules, which improves interfacial electron

transfer.^{41,42} The Bode magnitude plot reveals a similar trend, *i.e.*, showing a significantly lower absolute impedance for MIP compared to NIP. The Bode phase angle plot also introduces a distinct perspective, with MIP displaying a more pronounced deviation from 90° compared to NIP, which indicates the presence of defects caused by molecularly imprinted cavities leading to a higher leakage current.

To further improve the electrochemical characteristics, core-shell $\text{Nb}_2\text{O}_5\text{@NIP}$ and $\text{Nb}_2\text{O}_5\text{@MIP}$ electrodes were designed. The +5 oxidation state of Nb in Nb_2O_5 boosts robust electron transfer properties,^{43,44} evident in the exponential decrease in R_{ct} from 6.64 k Ω of NIP to 254 Ω for $\text{Nb}_2\text{O}_5\text{@NIP}$. Bode plots corroborate these changes, highlighting improved electronic conduction with the addition of semiconducting particles. A comparison of the electrochemical impedance and electroactive surface area of variously modified electrodes is provided in Fig. 4h and i, respectively. The best results are achieved with $\text{Nb}_2\text{O}_5\text{@MIP}$, where higher electroactive area and superior diffusion synergize with the fast electron-transferring semiconductor. As illustrated in Nyquist and Bode plots in Fig. 4d–f,






Fig. 4 EIS analysis: (a) Nyquist plot, (b) Bode magnitude plot, and (c) Bode phase angle plot of NIP and MIP; (d) Nyquist plot, (e) Bode magnitude plot, and (f) Bode phase angle plot of $\text{Nb}_2\text{O}_5@\text{NIP}$ and $\text{Nb}_2\text{O}_5@\text{MIP}$; (g) Randles circuit employed for fitting of EIS data; (h) a comparison of absolute impedance of different electrodes in a redox solution; and (i) a comparison of electroactive surface area of different electrodes. All measurements are performed in a redox solution containing $[\text{Fe}(\text{CN})_6]^{3-/4-}$ (2.5 mM), KCl (0.05 M) in PBS (pH 7.4).

the most significant disparities in impedance occur in the lower frequency region, indicative of a diffusion-controlled process.⁴⁵ Notably, $\text{Nb}_2\text{O}_5@\text{MIP}$ exhibits the lowest impedance, emphasizing its exceptional electrocatalytic properties and better suitability for sensor applications.

CV analysis of creatinine sensors

Expanding this investigation, CV measurements are conducted at a scan rate of 100 mV s^{-1} , across diverse creatinine concentrations (0–1000 nM). The CV responses from the fabricated

sensors exhibit a discernible increase in the anodic peak current (I_{pa}) with an upswing in creatinine concentration, a visual representation of which is shown in Fig. S3.† This CV response aligns with the EIS findings, where an elevation in creatinine concentration induces a reduction in the impedance of the interfacial layer. Consequently, an incremental surge in the I_{pa} is witnessed in tandem with the increasing creatinine concentrations, owing to the electrochemical oxidation of creatinine.^{20,21} The sensors demonstrate commendable linearity within the range of 0.97–0.99 coefficient of determination (R^2),



with the highest CV I_{pa} sensitivity observed in the case of $Nb_2O_5@MIP$ at $2.32 \times 10^{-2} \mu A nM^{-1}$, covering a linear dynamic range from 0 to 1000 nM. The CV response of the $Nb_2O_5@MIP$ sensor is substantially high compared to other sensors: NIP = $3.64 \times 10^{-3} \mu A nM^{-1}$; MIP = $4.71 \times 10^{-3} \mu A nM^{-1}$; and $Nb_2O_5@NIP$ = $9.87 \times 10^{-3} \mu A nM^{-1}$. Thus, these results validate the earlier assertion regarding the superior performance of $Nb_2O_5@MIP$ nanoreceptors.

DPV analysis and sensor properties

In pursuit of a comprehensive understanding and validation of the sensor responses, DPV measurements were thoroughly performed across various concentrations of creatinine (0–1000 nM). The resulting DPV responses, along with their respective calibration plots for the fabricated sensors, are shown in Fig. 5. DPV is a more reliable technique, offering a true perspective to distinguish between the fabricated sensors based on sensitivity, limit of detection (LOD), limit of quantification (LOQ), and linearity.⁴⁶ Calculations for LOD and LOQ are performed using the linear regression method, expressed as sensitivity = slope of the calibration curve; LOD = $3.3 \times (\text{standard deviation/sensitivity})$ and LOQ = $10 \times (\text{standard deviation/sensitivity})$, as outlined elsewhere.⁴⁷

In the pursuit of identifying the paramount sensor, key parameters such as heightened sensitivity, minimal LOD, elevated LOQ, and robust linearity emerge as pivotal benchmarks. Notably, the $Nb_2O_5@MIP$ sensor exhibits superior resolutions across varying concentrations of creatinine. It also boasts the highest DPV sensitivity at $4.614 \mu A cm^{-2} nM^{-1}$, coupled with an impressively low LOD of 34 pM, and LOQ of 103

pM, while maintaining a good linearity with an R^2 value of 0.93. These figures of merit are compared with the NIP, MIP, and $Nb_2O_5@NIP$ sensors in Table 1 along with examples from recently published data on electrochemical creatinine sensors. Thus, the $Nb_2O_5@MIP$ sensor not only emerges as the epitome of performance among the fabricated sensors but also lays the foundation for subsequent examinations. These examinations encompass measurements of selectivity, reproducibility, and real sample analyses, as discussed below.

Selectivity and reproducibility

In the realm of practical application of these sensors, the selectivity of the device is of paramount significance, especially given the intricate composition of saliva. The $Nb_2O_5@MIP$ sensor is exposed to prevalent analytes found in saliva, including urea, uric acid, glutamine, tyrosine, and others. The results are illustrated in Fig. 6a. Impressively, the sensor exhibits a remarkable selectivity for creatinine, showcasing a response that is 243% more pronounced compared to its closest contender – urea. Conversely, the sensor displays minimal responses (<30%) to other analytes.

In Fig. 6b, the reproducibility of the $Nb_2O_5@MIP$ sensors is illustrated. Three distinct disposable, gold SPEs are crafted using $Nb_2O_5@MIP$ nanoreceptors. Subsequently, all devices are exposed to equivalent concentrations of creatinine under ambient conditions, and their voltammetric responses are recorded in triplicate. Notably, the current responses across these devices consistently align within the range of 45–48 μA , with a mean value of $46.41 \pm 1.68 \mu A$. This affirms the outstanding reproducibility of the $Nb_2O_5@MIP$ sensors.





Fig. 5 DPV scans and corresponding calibration plots of (a and b) NIP, (c and d) MIP, (e and f) $Nb_2O_5@NIP$, and (g and h) $Nb_2O_5@MIP$ sensors at varying concentrations of creatinine (0–1000 nM) in a redox solution containing $[Fe(CN)_6]^{3-/4-}$ (2.5 mM), KCl (0.05 M) in PBS (pH 7.4).



Table 1 Performance comparison of the Nb₂O₅@MIP nanoreceptors-modified sensor with recently reported electrochemical creatinine sensors

Material	Method	Linear range	LOD	Sensitivity	Recovery	Ref.
PS/V:TiO ₂	CV, DPV	1–25 μ M	184 nM	30.95 μ A cm ⁻² μ M ⁻¹	—	21
PAA-Cu(II)/Cu ₂ O	CV	200–100 000 μ M	6.5 μ M	—	84.9%	48
CdSe-QD/IL	DPV	0.442–8840 μ M	229 nM	28.77 μ A μ M ⁻¹	94%	49
Cu ₂ O	CV	79.6–132.7 μ M	602 nM	—	—	50
Cu NPs	DPV	3.0–20.0 mM	0.084 mM	0.132 μ A mM ⁻¹	99%	51
NIP	CV, DPV	0–1000 nM	368 pM	0.707 μ A cm ⁻² nM ⁻¹	—	This work
MIP			244 pM	1.065 μ A cm ⁻² nM ⁻¹	—	
Nb ₂ O ₅ @NIP			223 pM	2.126 μ A cm ⁻² nM ⁻¹	—	
Nb ₂ O ₅ @MIP			34 pM	4.614 μ A cm ⁻² nM ⁻¹	100%	

**Fig. 6** (a) Selectivity of the Nb₂O₅@MIP sensor, when exposed to equivalent concentrations of various salivary analytes; (b) reproducibility of the Nb₂O₅@MIP sensor; and real-time sample analysis: (c) actual and spiked human saliva samples obtained from three different individuals, and (d) percent recovery of the spiked creatinine levels in human saliva samples. All measurements are performed in triplicate under ambient conditions.

Real-time salivary creatinine analysis

Human saliva stands as a complex amalgamation of various chemical species, capable of reacting and influencing the electrochemical detection of creatinine. To ascertain the reliability of the Nb₂O₅@MIP sensor in real-world applications, different samples, both actual saliva samples and those spiked with 200 nM creatinine are introduced to the sensor. Fig. 6c and

d presents the results of real-time salivary creatinine analyses. The recovery of the sensor is meticulously calculated using the formula: recovery (%) = $\{(C_{\text{total}} - C_{\text{initial}})/C_{\text{spiked}}\} \times 100$.¹⁹ The recovery outcomes for three real-time human saliva samples spiked with a known concentration of creatinine demonstrate 100.6% recovery, with a standard deviation of 9.1%. Hence, the creation of a point-of-care electrochemical sensor utilizing



Nb₂O₅@MIP nanoreceptors holds the potential to facilitate the monitoring of individuals with kidney-related diseases, offering a valuable tool for the timely diagnosis of renal failure.

Conclusions

In summary, this work introduces a disposable, easy-to-use salivary creatinine-detecting device — a sensor integrating Nb₂O₅ into creatinine-imprinted polythiophene, *i.e.*, Nb₂O₅@-MIP, meticulously designed for the rapid identification of minute creatinine levels in human saliva. The sensor undergoes rigorous characterization through FTIR spectroscopy, XRD, TEM, and SEM techniques. Electrochemical evaluation, encompassing EIS, CV, and DPV analyses, reveals the sensor's exceptional features. EIS exhibits remarkable charge transfer, featuring minimal electrode–electrolyte interfacial resistance and indicating the diffusion-controlled nature of the redox process. CV and DPV measurements further demonstrate the potential of the Nb₂O₅@MIP combination. DPV, in particular, demonstrates heightened sensitivity, achieving a sub-100 pM LOD — a remarkable feat for detecting trace creatinine levels in human saliva. The sensor exhibits considerable selectivity, and excellent reproducibility, and delivers outstanding creatinine recovery performance when subjected to real saliva samples. Hence, the Nb₂O₅@MIP sensor is an unparalleled PoC diagnostic device, providing an innovative solution for monitoring trace creatinine levels in individuals with kidney ailments and those experiencing muscle dysfunction.

Author contributions

Z. S. performed the synthesis and fabrication of electrocatalytic materials. Z. S. and M. S. designed the experiments, investigated sensor characteristics, and analyzed the data. M. F. carried out XRD and microscopic characterizations. Z. S. performed real sample analysis and wrote the first draft of the manuscript. S. Z. B. and A. M. revised the manuscript. A. A. conceived the idea, acquired funding, provided resources, supervised the research work, edited the first draft, and revised the manuscript.

Conflicts of interest

The authors declare no competing financial interests.

Acknowledgements

A. A. acknowledges the financial support provided by the Royal Society of Chemistry, UK (Research Grant: R21-9209614762).

References

- 1 C. S. Pundir, P. Kumar and R. Jaiwal, *Biosens. Bioelectron.*, 2019, **126**, 707–724.
- 2 V. Mani, T. Beduk, W. Khushaim, A. E. Ceylan, S. Timur, O. S. Wolfbeis and K. N. Salama, *TrAC, Trends Anal. Chem.*, 2021, **135**, 116164.
- 3 Z. Saddique, M. Faheem, A. Habib, I. UlHasan, A. Mujahid and A. Afzal, *Diagnostics*, 2023, **13**, 1737.
- 4 S. Kalasin, P. Sangnuang, P. Khownarumit, I. M. Tang and W. Surareungchai, *ACS Biomater. Sci. Eng.*, 2020, **6**, 5895–5910.
- 5 R. R. Kumar, M. O. Shaikh and C.-H. Chuang, *Anal. Chim. Acta*, 2021, **1183**, 338748.
- 6 P. Dasgupta, V. Kumar, P. R. Krishnaswamy and N. Bhat, *ACS Omega*, 2020, **5**, 22459–22464.
- 7 P. Delanaye, E. Cavalier and H. Pottel, *Nephron*, 2017, **136**, 302–308.
- 8 C.-Z. Zhang, X.-Q. Cheng, J.-Y. Li, P. Zhang, P. Yi, X. Xu and X.-D. Zhou, *Int. J. Oral Sci.*, 2016, **8**, 133–137.
- 9 D.-S. Ciou, P.-H. Wu, Y.-C. Huang, M.-C. Yang, S.-Y. Lee and C.-Y. Lin, *Sens. Actuators, B*, 2020, **314**, 128034.
- 10 M. Jaffé, *Biol. Chem.*, 1886, **10**, 391–400.
- 11 S. Bai, P. Gonzalez-Vasquez, C. Torres-Calzada, S. MacKay, J. Cook, Y. Khaniani, G. Davies, U. Singh, P. Kovur, J. Chen and D. S. Wishart, *Biosens. Bioelectron.*, 2024, **253**, 116186.
- 12 Y. Dong, X. Luo, Y. Liu, C. Yan, H. Li, J. Lv, L. Yang and Y. Cui, *Talanta*, 2022, **248**, 123592.
- 13 R. K. R. Kumar, M. O. Shaikh, A. Kumar, C.-H. Liu and C.-H. Chuang, *ACS Appl. Nano Mater.*, 2023, **6**, 2083–2094.
- 14 G. Álvarez Menéndez, O. Amor-Gutiérrez, A. Costa García, M. Funes-Menéndez, C. Prado, D. Miguel, P. Rodríguez-González, A. González-Gago and J. I. García Alonso, *Clin. Chim. Acta*, 2023, **543**, 117300.
- 15 J. J. BelBruno, *Chem. Rev.*, 2019, **119**, 94–119.
- 16 J. W. Lowdon, H. Diliën, P. Singla, M. Peeters, T. J. Cleij, B. van Grinsven and K. Eersels, *Sens. Actuators, B*, 2020, **325**, 128973.
- 17 M. Javanbakht, S. E. Fard, A. Mohammadi, M. Abdouss, M. R. Ganjali, P. Norouzi and L. Safaraliev, *Anal. Chim. Acta*, 2008, **612**, 65–74.
- 18 A. Afzal, S. Feroz, N. Iqbal, A. Mujahid and A. Rehman, *Sens. Actuators, B*, 2016, **231**, 431–439.
- 19 M. Saeed, Z. Saddique, A. Mujahid and A. Afzal, *Biosens. Bioelectron.*, 2024, **247**, 115899.
- 20 Z. Saddique, T. Siddique, A. Shehzadi, Q. Ul-Ain, A. Ali, N. Iqbal, A. Mujahid and A. Afzal, *Compos. Commun.*, 2023, **42**, 101673.
- 21 T. Sawaira, A. Jamil, S. Aziz, A. Mujahid, T. Hussain and A. Afzal, *Compos. Commun.*, 2023, **37**, 101398.
- 22 N. Manjula, S. Pulikkutty, T.-W. Chen, S.-M. Chen, C. H. Fan, M. A. Ali and F. M. Al-Hemaid, *Colloids Surf., A*, 2022, **637**, 128277.
- 23 T. Alizadeh and Z. Mousavi, *Microchim. Acta*, 2022, **189**, 393.
- 24 S. Nur Ashakirin, M. H. M. Zaid, M. A. S. M. Haniff, A. Masood and M. F. Mohd Razip Wee, *Measurement*, 2023, **210**, 112502.
- 25 S. Athar, I. Zaman, A. Liaqat and A. Afzal, *ACS Appl. Polym. Mater.*, 2023, **5**, 10438–10445.
- 26 J. Sopoušek, J. Věžník, J. Houser, P. Skládál and K. Lacina, *Electrochim. Acta*, 2021, **388**, 138616.
- 27 E. Pradeepa, Y. Arthoba Nayaka and H. R. Sahana, *Anal. Biochem.*, 2024, **687**, 115451.



- 28 X.-G. Li, J. Li and M.-R. Huang, *Chem.-Eur. J.*, 2009, **15**, 6446–6455.
- 29 R. Liu and Z. Liu, *Chin. Sci. Bull.*, 2009, **54**, 2028–2032.
- 30 A. Khan, A. Habib and A. Afzal, *Beilstein J. Nanotechnol.*, 2020, **11**, 1190–1197.
- 31 S. Nagarajan, V. Raman and N. Rajendran, *Mater. Chem. Phys.*, 2010, **119**, 363–366.
- 32 A. Pawlicka, M. Atik and M. A. Aegerter, *Thin Solid Films*, 1997, **301**, 236–241.
- 33 S. Shariati and G. Khayatian, *Microfluid. Nanofluid.*, 2022, **26**, 30.
- 34 J. H. Lim and J. Choi, *J. Ind. Eng. Chem.*, 2009, **15**, 860–864.
- 35 H. Park, D. Lee and T. Song, *J. Power Sources*, 2019, **414**, 377–382.
- 36 R. Shao, Z. Cao, Y. Xiao, H. Dong, W. He, Y. Gao and J. Liu, *RSC Adv.*, 2014, **4**, 26447–26451.
- 37 L. Durai and S. Badhulika, *IEEE Sens. J.*, 2021, **21**, 4152–4159.
- 38 H. Essousi, H. Barhoumi, S. Karastogianni and S. T. Girousi, *Electroanalysis*, 2020, **32**, 1546–1558.
- 39 E. Sharafi and S. Sadeghi, *New J. Chem.*, 2023, **47**, 500–514.
- 40 A. M. Abdel-Aziz, H. H. Hassan and I. H. A. Badr, *ACS Omega*, 2022, **7**, 34127–34135.
- 41 W. Liu, Y. Ma, G. Sun, S. Wang, J. Deng and H. Wei, *Biosens. Bioelectron.*, 2017, **92**, 305–312.
- 42 W. Zheng, M. Zhao, W. Liu, S. Yu, L. Niu, G. Li, H. Li and W. Liu, *J. Electroanal. Chem.*, 2018, **813**, 75–82.
- 43 C. Nico, T. Monteiro and M. P. F. Graça, *Prog. Mater. Sci.*, 2016, **80**, 1–37.
- 44 Z. Zhao, Z. Xue, Q. Xiong, Y. Zhang, X. Hu, H. Chi, H. Qin, Y. Yuan and H. Ni, *Sustainable Mater. Technol.*, 2021, **30**, e00357.
- 45 V. Ratautaite, M. Nesladek, A. Ramanaviciene, I. Baleviciute and A. Ramanavicius, *Electroanalysis*, 2014, **26**, 2458–2464.
- 46 B. J. Privett, J. H. Shin and M. H. Schoenfish, *Anal. Chem.*, 2010, **82**, 4723–4741.
- 47 U. Arshad, A. Mujahid, P. Lieberzeit, A. Afzal, S. Zafar Bajwa, N. Iqbal and S. Roshan, *RSC Adv.*, 2020, **10**, 34355–34363.
- 48 S. Kalasin, P. Sangnuang, P. Khownarumit, I. M. Tang and W. Surareungchai, *ACS Biomater. Sci. Eng.*, 2020, **6**, 1247–1258.
- 49 S. Hooshmand and Z. Es'haghi, *Anal. Chim. Acta*, 2017, **972**, 28–37.
- 50 C.-H. Chen and M. S. Lin, *Biosens. Bioelectron.*, 2012, **31**, 90–94.
- 51 S. Jankhunthod, K. Kaewket, P. Termsombut, C. Khamdang and K. Ngamchuea, *Anal. Bioanal. Chem.*, 2023, **415**, 3231–3242.

