


 Cite this: *RSC Adv.*, 2024, 14, 34279

Development of an ultra-sensitive laser stimulated fluorescence system for simultaneous detection of amino acids

 Megha Naik and Ajeetkumar Patil *

An ultra-sensitive, high-performance liquid chromatography-based laser-stimulated fluorescence detection system was developed and validated for the simultaneous detection of 20 derivatized amino acids. Dansyl chloride was used as a derivatizing agent, and key derivatization parameters, such as reaction time and temperature, were optimized to enhance sensitivity and reproducibility. The majority of amino acids showed a relative standard deviation of less than 5%, indicating the reliability of the approach. The method demonstrated excellent sensitivity for all 20 amino acids, with detection limits ranging from 4.32 to 85.34 femtomoles. It also exhibited good linearity, with regression (R^2) values greater than 0.98 for the amino acids. The system's performance was tested on human serum, and the eluted amino acids were identified. This method has great potential for analyzing amino acids in various body fluids and can be used in various clinical applications. It is ultra-sensitive, reliable, user-friendly, and cost-effective, offering a valuable tool for diagnosing and managing amino acid-related disorders.

 Received 4th July 2024
 Accepted 19th October 2024

DOI: 10.1039/d4ra04845h

rsc.li/rsc-advances

Introduction

Amino acids (AAs) are the structural units of proteins and intermediate metabolites that drive subsequent biosynthetic activities.^{1,2} AAs are essential for the human body's growth, development, and maintenance, and their proper intake through a balanced diet is crucial for optimal health.³ Some of the AAs, such as arginine, glutamine, glutamate, glycine, proline, and branched chain AAs, have significant effects on blood flow,⁴ nutrient transport and metabolism in animal cells,⁵ intestinal microbial growth and metabolism, antioxidative actions,⁶ as well as innate and cell-mediated immune responses.⁷ Functional AAs show immense potential in preventing and treating infectious diseases (including viral infections), metabolic disorders (such as obesity, diabetes, and cardiovascular diseases), lactation failure, fetal and postpartum growth restriction, male and female infertility, and organ dysfunctions.⁸ Therefore, a qualitative and quantitative assessment of AAs in body fluids is crucial, and it is important to use a sensitive and reliable system to ensure accurate results.

Several analytical techniques are available for the assessment of AAs, such as AA analyzers, nuclear magnetic resonance (NMR), gas chromatography (GC), high-performance liquid chromatography (HPLC) with UV/fluorescence detectors, and LC-mass spectrometry (LC-MS).^{9,10} The conventional AA analysis technique involves ion exchange chromatography with post-

column derivatization using a ninhydrin reagent. However, this approach has limitations such as poor retention for acidic molecules, low specificity, and longer analysis time.¹¹ NMR spectroscopy is frequently utilized for AA profiling due to its quick sample analysis, high repeatability, and precise quantification, though its sensitivity is relatively limited.^{12,13} While GC-MS provides high sensitivity and resolution for AAs, the complexity of sample preparation and derivatization makes HPLC-based methods more preferred.¹⁴ HPLC combined with MS offers the highest sensitivity and selectivity compared to other techniques.^{15,16} Nevertheless, the high instrumentation and operation costs, along with the complex sample preparation, are notable challenges.^{17,18} HPLC with ultraviolet (UV) and fluorescence (FL) detection is often preferred due to its low cost, ease of use, and adequate sensitivity. Compared to UV detectors, HPLC with fluorescence detection provides increased sensitivity and precision, making it a more favorable choice for analyzing AAs.

The derivatization process is typically carried out to identify AAs using HPLC with UV/FL/MS detectors because AAs generally lack potent chromophores or fluorophores necessary for direct detection.¹⁶ For derivatization, several reagents were employed, such as dansyl chloride (Dns-Cl), dabsyl chloride (Dbs-Cl), fluorescamine, and fluorenyl-methyl chloroformate (FMOC), phenylisothiocyanate (PITC), aminoquinolyl-N-hydroxylsuccinimidylcarbamate (AQC) and *o*-phthalaldehyde (OPA) have been used.^{19–21} Though OPA derivatization is often preferred due to its simplicity and flexibility, it does not react with secondary AAs, and the derivatives are often unstable.²² Other reagents, such as AQC, FMOC-Cl, and PITC, also have

Department of Atomic and Molecular Physics, Manipal Academy of Higher Education (MAHE), Manipal, India-576 104. E-mail: ajeetkumar.p@manipal.edu; ajeetkumar@gmail.com



limitations. PITC derivatives are said to be unstable, whereas interference from the hydrolysis products of AQC and FMOC-Cl hinders the detection of AAs.^{23,24} Dns-Cl derivatization is widely used for investigating compounds containing primary and secondary amines, including AAs, biogenic amines, and polyamines.^{25,26}

Despite the available methods, there is a need for a sensitive, cost-effective, and reliable technique for analyzing AAs. The present study describes the development of an ultrasensitive HPLC-coupled laser-stimulated fluorescence (LSF) detection system for qualitative and quantitative analysis of AAs using Dns-Cl as a derivatizing agent. This technique offers several advantages in terms of simplicity, excellent sensitivity, and cost-effectiveness compared to similar techniques.

Materials and methods

Reagents

AA standards kit ($\geq 99.0\%$) [alanine (Ala), aspartic acid (Asp), cysteine (Cys), cystine (Cys-Cys), glutamine (Gln), glutamic acid (Glu), glycine (Gly), histidine (His), hydroxyproline (Hyp), isoleucine (Ile), leucine (Leu), lysine (Lys), methionine (Met), phenylalanine (Phe), proline (Pro), serine (Ser), threonine (Thr), tryptophan (Trp), tyrosine (Tyr), and valine (Val)], sodium carbonate (99.9%), sodium bicarbonate ($\geq 99.7\%$), Dns-Cl ($\geq 99\%$ HPLC), ammonium hydroxide (30–33% NH_3 in H_2O , puriss.) (Sigma-Aldrich, U.S); HPLC grade water, acetic acid, acetonitrile (Merck Life Science, Germany); methanol (LCMS grade, RCI Labscan Limited, Thailand), and sodium acetate (99.9%) (Sisco Research Laboratories Pvt. Ltd India).

Derivatization procedure

The Dns-Cl (37 mM) solution was prepared in acetonitrile, and stock solutions of individual standard AAs were prepared using HPLC-grade water. The stock solutions were diluted to working concentrations using bicarbonate buffer (pH 9.2). Dns labeling occurs only in alkaline conditions because amino groups react as free bases, not conjugate acids. Dns-Cl reacts with nucleophiles like water to produce dansyl sulfonic acid (Dns-OH) under an OH-catalyzed reaction. A pH of 9.5 is recommended as a threshold to prevent hydrolysis.^{27,28} Therefore, a sodium bicarbonate buffer with pH 9.2 was used for the reaction.

To the solution of standard AAs, 25 μL of Dns-Cl was added and kept at room temperature for an hour. Fig. 1 shows the reaction mechanism of Dns-Cl with AA. The excess of Dns-Cl present in the reaction mixture competes with the formation of dansyl AAs (Dns-AA) and converts them to dansylamide (Dns-NH₂) and other byproducts.²⁹ Therefore, after one hour of reaction, the excess of Dns-Cl was quenched using an ammonium hydroxide (NH_4OH) solution to prevent the conversion of Dns-AA to Dns-NH₂.

Absorption and fluorescence studies of Dns-Cl

The absorption spectrum of Dns-Cl was recorded using JASCO-UV 3600. Fig. 2a depicts the absorption spectra of Dns-Cl in ACN. The absorption maxima were found to be 240 and 315 nm,




Fig. 1 Reaction of Dns-Cl and AA.




Fig. 2 (a) Absorption spectra of Dns-Cl, (b) fluorescence emission spectra of Dns-Cl.

respectively. In addition, the fluorescence spectra of Dns-Cl were recorded using a Jasco FP8300 fluorescence spectrometer. The fluorescence emission peak of Dns-Cl was observed at 530 nm when excited at a wavelength of 325 nm, as shown in Fig. 2b. In this study, an excitation source with a 325 nm wavelength was utilized, which aligns with the absorption characteristics of Dns-Cl. This choice of excitation wavelength also helps to avoid interference from autofluorescence exhibited by AAs, which is typically observed at wavelengths below 300 nm. Subsequently, the emission wavelength of 530 nm was chosen for detecting Dns-AA using the HPLC-LSF technique discussed below. By carefully choosing the excitation and emission parameters, the precision and selectivity of Dns-AA detection were enhanced, making it an excellent analytical approach.

Experimental setup

In earlier work, a sensitive and reliable HPLC coupled-fluorescence-based detection tool was reported for studying the proteins in various body fluids, including serum, saliva, tissue, and cellular samples. This method has a detection limit of femtomoles and is helpful for various biomedical applications.^{30–34} The studies were subsequently extended to develop an HPLC-LSF system for detecting AAs. Several modifications were made in the system and analysis procedure to obtain qualitative and quantitative assessments of AAs with high sensitivity.

The two primary components of the experimental setup are the HPLC unit for Dns-AA separation and the LSF unit for its detection, as depicted in Fig. 3.






Fig. 3 Schematic of HPLC-LSF system.

Chromatographic separation

The separation part consists of an HPLC system (Agilent 1200 series) with a G1322A degasser, G1311A pump, and a manual injector (model no. 7725i, Rheodyne, USA) coupled to a reversed-phase ZORBAX 300Extend-C18 selectivity HPLC column 5 μm , 250 mm \times 4.6 mm i.d. column (No. 770995-902, Agilent, USA). The column's effluent was directed into a UV-grade quartz capillary flow cell (75 μm inner diameter, HP CE Capillary, US). The capillary is connected to PEEK tubing (part no. 5042-6462, 1.58 mm outer diameter, 175 μm inner diameter, Agilent, USA) using finger-tight fittings. The flow cell is precisely positioned on precession mounts for consistent excitation and fluorescence collection. The derivatized solution was injected into the manual injector (7725i, Rheodyne, USA) fitted with a 20 μL sample loop.

Chromatographic condition

DNS-AAAs were eluted using a mobile phase consisting of a sodium acetate buffer (83 mM, pH 5.94) (A) and methanol (B). The elution process followed a gradient starting with 60% solvent A, decreasing to 55% in 26 min, 55% to 40% in 35 min, 40% to 33% in 48 min, and 0% in 60 min. The 100% B was continued for 10 minutes, extending the total run time to 70 minutes. A blank was recorded after every run to make sure the column was clean before each sample run.

LSF detection system

The LSF technique was employed to detect AAs. A continuous-wave, He-Cd laser (KIMMON KOHA, IK3083R-D, JAPAN) emitting at a wavelength of 325 nm was used to effectively excite Dns-AA. The beam was focused onto the capillary using a biconvex lens (focal length (f) = 5 cm). An optical chopper (SR540, Stanford Research Systems, Inc., USA) was used to chop the laser beam constantly at a frequency of 20 Hz to provide a reference for lock-in detection. As depicted in Fig. 2, the fluorescence signal from the sample was collected at an angle of 90° to lessen the background signal. After collimating the emitted signal using a biconvex lens [focal length (f) = 3 cm], an additional biconvex $f/4$ lens with a 10 cm focal length was used to refocus the fluorescence beam on the monochromator (No. 0051-08-06, Micro HR, Horiba Jobin Yvon, USA) to match the monochromator's F value, which is adjusted at a 530 nm wavelength. A high pass filter was positioned directly before the

monochromator slit to minimize stray light at other wavelengths. The fluorescence was detected by a photomultiplier tube (R750, Hamamatsu Inc., Hamamatsu, Japan) operated at -850 V, coupled through a preamplifier (Signal Recovery Model 5113, USA), which is connected to a lock-in amplifier (Signal Recovery Model 7265, USA). The chromatograms were obtained using a computer connected to the lock-in amplifier's output. The excitation and collection optics used along with the He-Cd LSF system were more effective at collecting the total emitted radiation, and the choice of parameters such as PMT voltage, the gain of lock-in amplifier, the choice of suitable and stable derivatizing agent, and optimization of reaction conditions all contributed to an improved detection limit.

Results and discussion

In clinical laboratories, the accurate measurement of AA analysis is vital, and it relies on the specificity and sensitivity of the technique. This study evaluated the stability, repeatability, and sensitivity of the HPLC-LSF system as key performance parameters.

The derivatization procedure was modified to achieve the best possible detection. Many studies have used Dns-Cl to detect AAs in various samples. The commonly used dansyl derivatization approach for detecting AAs involves heating the sample at 60–80 °C for 30–60 minutes or leaving it at room temperature for 24 hours.^{35,36} However, these conditions can lead to side reactions and the formation of byproducts, affecting the accuracy of the analysis. To improve the reliability and efficiency of the derivatization process, the conditions were optimized to allow the reaction to take place at room temperature within an hour. This offers several advantages over higher temperatures, as higher temperatures can cause sample degradation or necessitate additional heating equipment. After derivatization, 20 μL of the solution was injected into the HPLC system, and the chromatogram of dansylated AAs was recorded.

Fig. 4a shows the chromatogram of a typical blank recorded to ensure that the column was clean and no residues were present that could interfere with the analysis of AAs. A mixture




Fig. 4 (a) Blank, (b) chromatogram of Dns-Cl in bicarbonate buffer. (c) Chromatogram of the dansylated standard AA mixture in bicarbonate buffer (* – residual peaks).



of 20 AAs without derivatization was injected to assess the potential interference from the autofluorescence of AAs. It was observed that no fluorescence signal was detected when excited at 325 nm, and the resulting chromatogram of the mixture was indistinguishable from that of a blank run. Fig. 4b represents the chromatogram of Dns-Cl in bicarbonate buffer, while Fig. 4c shows the chromatogram of a standard AA mixture. The chromatographic conditions were optimized to prevent nonspecific peaks from interfering with the quantification of AAs, thus ensuring the reliability of the analysis.

A total of 20 AAs were detected, and out of these, the histidine dansylation resulted in two peaks. T4-hydroxyproline and




Fig. 5 Chromatograms of standard AAs (after baseline correction) at different concentrations: (a) Asp & Glu, (b) Gln, (c) Ser, (d) Thr & Gly, (e) Ala & Tyr, (f) Pro, (g) Val & Met, (h) Trp, (i) Ile & Leu, (j) Phe, (k) Cys, (l) Lys and (m) His.




Fig. 6 Calibration curves of standard AAs: (a) Asp, (b) Glu, (c) Gln, (d) Ser, (e) Thr, (f) Gly, (g) Ala, (h) Tyr, (i) Pro, (j) Val, (k) Met, (l) Trp, (m) Ile, (n) Leu, (o) Phe, (p) Cys, (q) Lys and (r) His.



the first of the two His peaks invariably co-eluted, posing difficulty in quantification due to the combined peak eluting as a shoulder to a significant peak formed by dansyl hydroxy product. The quantification was done using the second His peak. Fig. 5 displays the overlaid chromatograms for various AA concentrations. The solution of standard AAs was measured at decreasing concentrations until a peak three times higher than the baseline was observed. The AA peaks were quantified using a standard calibration method, which involved constructing five- to six-point calibration curves for each AA. Fig. 6 shows the calibration curves for the AAs plotted using the intensity of the AA peak of the chromatogram against their respective concentrations. The limit of detection (LOD), and limit of quantification (LOQ) were calculated using standard error and slope using the equation, $LOD = 3.3 \times \text{standard error/slope}$ and $LOQ = 10 \times \text{standard error/slope}$.³⁷ Table 1 shows the retention time (RT), LOD, LOQ, relative standard deviation, and R^2 values for 18 AAs, calculated using the calibration curves. It was observed that, most of the AAs analyzed have detection limits below 50 fmol. The standard curve R^2 value for 18 AAs analyzed was 0.98 or higher. This shows good linearity over the concentration range examined, indicating that the technique can accurately quantify AAs over a wide dynamic range.

The repeatability was assessed at two different concentration levels (200 and 1000 fmol) to ensure consistent and accurate results from the system. Table 2 summarizes the intra-day and inter-day repeatability for AA detection using the HPLC-LSF system. Intra-day precision was evaluated by performing five repeated injections on the same day. Most AAs exhibited a % RSD of less than 6% at both concentrations, indicating high precision within a single day. Inter-day precision measures the consistency of a method over an extended period. The inter-day precision was assessed using the same analytical conditions over three consecutive days ($n = 5$). The %RSD for AAs remained below 6%, indicating good repeatability for analyzing AAs over extended periods.

The efficacy and reliability of the suggested approach were assessed using correlation uncertainty and predicted values. Table 3 shows the correlation of uncertainty and the expected concentration of AAs using the leave-one-out approach. By employing known concentrations of the AAs in this procedure, a calibration curve was plotted, and one AA concentration can be omitted as a test sample. The predicted concentration (C_p) of the sample is determined from the ratio of signal intensity of the analyte concentration with corrected intercept value ($y - c$) and the slope (m) from the calibration curve.

Evaluation of the blue applicability grade index (BAGI) and greenness assessment

The field of green analytical chemistry focuses on reducing the use of harmful solvents in the extraction and sample preparation process and minimizing the energy consumption of analytical instruments.³⁸ The green analytical procedure index (GAPI) aids in assessing the environmental impact of analytical techniques by considering various parameters such as the types of reagents or solvents used, energy consumption, waste production, and the potential for materials recycling and reuse.³⁹ ComplexGAPI was utilized to assess our technique, demonstrating the method's green feature by providing a graphical representation of a pentagon. The representation includes three colours (green, red, and yellow) to indicate environmental impact and severity. The GAPI pictogram in Fig. 7a indicates that the proposed method largely follows green chemistry guidelines, as shown by the significant yellow and green colours. This method uses green solvents like methanol and sodium acetate, simple sample preparation, and treatment procedures. Additionally, the low flow rate (0.2 mL min^{-1}) used to separate AAs significantly reduces solvent, waste, and energy consumption in the LC system. The only red region in the Pentagon represents waste generation of more than 10 mL, which occurs because each run requires approximately 12 mL of

Table 1 Detection limits and quantification limits of Dns-AAs

Sl. No.	Amino acid	RT (seconds)	Standard error ($\times 10^{-4}$)	Slope ($\times 10^{-4}$)	LOD (fmol)	LOQ (fmol)	Linear range (fmol)	R^2
1	Asp	1676.8	7.5	0.29	85.34	258.62	100–5000	0.989
2	Glu	1706.7	11.1	0.46	79.63	241.30	100–5000	0.999
3	Gln	2134	29.4	2.1	46.20	140.00	100–5000	0.999
4	Ser	2328	2.7	0.59	15.1	45.76	20–5000	0.993
5	Thr	2516.7	10	1.3	25.38	76.92	100–5000	0.999
6	Gly	2576	3.2	0.42	25.14	76.19	100–5000	0.999
7	Ala	2737	5.4	0.835	21.34	64.67	100–5000	0.999
8	Tyr	2750	1.35	0.76	5.86	17.76	20–5000	0.997
9	Pro	3069.6	38	3.8	33.00	100.00	100–2000	0.999
10	Val	3186	3.6	2.8	4.24	12.85	20–2000	0.984
11	Met	3225	8.1	2	13.36	40.50	20–2000	0.996
12	Trp	3386	38	4.3	29.16	88.37	100–2000	0.993
13	Ile	3467	4.6	3	5.06	15.33	20–2000	0.995
14	Leu	3506	16	1.9	27.78	84.21	100–2000	0.999
15	Phe	3616.6	28	4	23.10	70.00	100–2000	0.987
16	Cys + Cys–Cys	3748	80	4.3	60.39	186.04	100–2000	0.992
17	Lys	4136.6	11.9	1.7	23.10	70.00	100–2000	0.983
18	His	4342	1.1	0.46	7.89	23.91	100–5000	0.998





Table 2 Intra-day and inter-day precision

Amino acids	Average RT (seconds)	Intra-day repeatability			Inter-day repeatability				
		For 1000 fmol		For 200 fmol		For 1000 fmol		For 200 fmol	
		Mean intensity \pm SD (10^{-3})	%RSD	Mean intensity \pm SD (10^{-3})	%RSD	Mean intensity \pm SD (10^{-3})	%RSD	Mean intensity \pm SD (10^{-3})	%RSD
Asp	1662 \pm 18	33 \pm 1.4	4.28	5.2 \pm 0.29	5.55	33 \pm 1.7	5.25	5.2 \pm 0.26	5.08
Glu	1694 \pm 13	50 \pm 2.1	4.34	7.7 \pm 0.24	3.08	51 \pm 2.3	4.49	7.3 \pm 0.29	3.93
Gln	2123 \pm 16	202 \pm 4.4	2.21	25.0 \pm 1.41	5.65	213 \pm 5.7	2.70	25.3 \pm 1.5	6.03
Ser	2318 \pm 17	85 \pm 3.9	4.67	5.1 \pm 0.25	4.82	85 \pm 4.2	4.88	5.2 \pm 0.25	4.80
Thr	2509 \pm 9	105 \pm 3.6	3.46	22.2 \pm 0.96	4.31	105 \pm 5.0	4.76	22.0 \pm 1.0	4.54
Gly	2567 \pm 18	34 \pm 1.6	4.86	7.0 \pm 0.38	5.47	35 \pm 1.5	4.32	7.0 \pm 0.47	6.72
Ala	2728 \pm 20	82 \pm 2.6	3.16	12.4 \pm 0.63	5.07	84 \pm 2.5	2.98	14.3 \pm 0.57	4.02
Tyr	2743 \pm 17	79 \pm 2.2	2.87	14.4 \pm 0.50	3.47	83 \pm 3.0	3.69	14.6 \pm 0.57	3.93
Pro	3065 \pm 16	398 \pm 4.4	1.12	68.0 \pm 2.88	4.24	393 \pm 5.7	1.46	67.3 \pm 3.05	4.53
Val	3182 \pm 14	210 \pm 6.1	2.91	41.0 \pm 1.15	2.81	210 \pm 10	4.76	41.3 \pm 1.52	3.69
Met	3221 \pm 14	204 \pm 5.5	2.68	30.4 \pm 1.00	3.29	203 \pm 5.7	2.84	31.0 \pm 1.0	3.22
Trp	3384 \pm 15	402 \pm 10.9	2.72	76.4 \pm 2.21	2.90	386 \pm 11.5	2.99	76.6 \pm 3.05	3.98
Ile	3467 \pm 17	320 \pm 12.2	3.82	47.8 \pm 2.38	4.98	316 \pm 15.2	4.82	47.0 \pm 2.6	5.63
Leu	3505 \pm 16	209 \pm 7.4	3.54	31.4 \pm 0.95	3.05	225 \pm 8.6	3.85	32.0 \pm 1.0	3.12
Phe	3616 \pm 16	397 \pm 6.7	1.68	69.8 \pm 1.50	2.15	386 \pm 11.5	2.98	66.6 \pm 2.88	4.33
Cys + Cys-Cys	3737 \pm 15	232 \pm 8.3	3.60	61.6 \pm 2.87	4.66	230 \pm 10	4.34	61.6 \pm 2.88	4.68
Lys	4136 \pm 4	185 \pm 5.0	2.70	29.8 \pm 1.00	3.35	183 \pm 5.7	3.15	28.6 \pm 1.15	4.02
His	4343 \pm 0.5	74 \pm 4.2	5.65	7.5 \pm 0.47	6.25	75 \pm 4.1	5.52	7.5 \pm 0.5	6.66

Table 3 Correlation uncertainty and predicted concentration of AAs by leave one out approach

Sl. No.	Amino acids	Known concentration (C_k) fmol	Predicted concentration (C_p) [$x = (y - c)/m$] fmol	Correlation uncertainty (%) $(C_k - C_p /C_p) \times 100\%$
1	Asp	1000	938.30	3.13
2	Glu	1000	1042.52	3.19
3	Gln	2000	1932.85	3.47
4	Ser	2000	2004.61	1.91
5	Thr	500	490.4	1.95
6	Gly	500	469.76	5.16
7	Ala	2000	1923.75	3.96
8	Tyr	500	466.66	5.63
9	Pro	1000	1015.58	1.53
10	Val	1000	944.59	4.51
11	Met	500	465.5	6.38
12	Trp	1000	1006.97	7.62
13	Ile	500	469	6.53
14	Leu	2000	1866.84	10.7
15	Phe	1000	977.5	2.04
16	Cys + Cys-Cys	1000	1055.68	5.08
17	Lys	500	455.55	7.40
18	His	2000	2019.35	0.95




Fig. 7 (a) The GAPI pictogram and (b) the BAGI index pictogram.

solvents. Despite this, the proposed method is predominantly green and environmentally friendly. To address waste generation, solvent recovery and method optimization can be employed.

The practicality of the analytical procedure was evaluated using BAGI software. It assesses a range of parameters, such as the type of analysis, methods for sample preparation, the total number of steps, samples examined in an hour, the volume of sample, necessary preconcentration, instruments required, and automation level.⁴⁰ It yields a score that typically ranges from 25 to 100, with a score of 100 indicating exceptional applicability and a score of less than 25 suggesting lower relevance. The total score obtained for the HPLC-LSF method using the BAGI software is 70, as depicted in Fig. 7b. This result meets the specified acceptance threshold. By utilizing derivatization, our analytical method allowed for the simultaneous analysis of 20 AAs. Less than 5 μL of sample is adequate for measuring AAs in biological fluids, ensuring sufficient sensitivity and minimal biological waste.

Detection of AAs in serum sample

AA analysis of a human serum sample was conducted to demonstrate the practical application of the usefulness of the Dns derivatized AAs with the above-mentioned HPLC-LSF

system. The blood sample was collected from a healthy volunteer. Ethical clearance was obtained for the study from the institutional ethical committee of Kasturba Hospital, Manipal (IEC 651/2020), and written informed consent was obtained from the participant. The serum sample was obtained by centrifuging the blood sample at 2000 rpm for 10 min. 1 μL of serum sample was diluted using sodium bicarbonate buffer (74 μL) and derivatized following the method described in the previous section. Fig. 8 depicts the chromatogram of the AA profile of human serum, and 15 AAs were found in the sample.

The study results emphasize that the in-house developed HPLC-LSF system is capable of simultaneously detecting 20 AAs at lower concentrations in the range of fmol. This high sensitivity and precision of the system enables accurate identification of AAs, making it a suitable option for AA profiling in biological fluids such as cerebrospinal fluid,⁴¹ saliva,⁴² and tears¹⁵ where AA levels are typically very low (of the order of picomole or below). This technique requires minimal sample volume (less than 5 μL), which is advantageous when working with precious biological samples. For instance, when studying




Fig. 8 Chromatogram of a serum sample.



AA-related disorders in infants, a limited volume of biological fluid is available for testing. In such cases, the HPLC-LSF technique can be used to analyze AAs with very small sample volumes, reducing the necessity for large blood draws and ensuring sufficient sensitivity.

Conclusion

The study demonstrated the effectiveness of the HPLC-LSF system for the simultaneous detection of 20 AAs. Optimization of the derivatization conditions significantly enhanced sensitivity and reliability in AA detection. The method exhibits excellent sensitivity, with a LOD ranging from 4.32 to 85.34 fmols. Furthermore, the repeatability and reliability of the system were validated using intra-day and inter-day precision tests.

The findings underscore the efficiency and reliability of the proposed method in thoroughly examining AA levels in biological fluids. This method holds significant potential for clinical applications, especially in monitoring disease progression, assessing treatment efficacy, and facilitating early diagnosis of AA-related disorders. In addition to clinical diagnostics, this method offers promising applications in biochemistry, pharmaceutical research, and environmental monitoring for accurate measurement of AAs.

Data availability

The data supporting the findings of this study are part of a patent application submitted to the Indian Patent Office (Application No. 202341044833 dated 4th July 2023) and data may be shared upon request.

Conflicts of interest

The present work has been filed for intellectual property rights in the Indian Patent Office, Chennai (Application No. 202341044833 dated 4th July 2023).

Acknowledgements

Ms Megha Naik gratefully acknowledges the Dr T. M. A. Pai scholarship provided by the Manipal Academy of Higher Education. The authors thank the Manipal Academy of Higher Education for the MAHE-Intramural Grant – MAHE/CDS/PHD/IMF/2019.

References

- W. Xu, C. Zhong, C. Zou, B. Wang and N. Zhang, *Amino Acids*, 2020, **52**, 1071–1088.
- H. Kaspar, K. Dettmer, W. Gronwald and P. J. Oefner, *Anal. Bioanal. Chem.*, 2009, **393**, 445–452.
- E. Aliu, S. Kanungo and G. L. Arnold, *Ann. Transl. Med.*, 2018, **6**, 471.
- B. Tan, X. Li, G. Wu, X. Kong, Z. Liu, T. Li and Y. Yin, *Amino Acids*, 2012, **43**, 2481–2489.
- A. Suryawan, H. V. Nguyen, R. D. Almonaci and T. A. Davis, *Amino Acids*, 2013, **45**, 523–530.
- S. Katayama and Y. Mine, *J. Agric. Food Chem.*, 2007, **55**, 8458–8464.
- J. Y. Kim, R. C. Burghardt, G. Wu, G. A. Johnson, T. E. Spencer and F. W. Bazer, *Biol. Reprod.*, 2011, **84**, 62–69.
- G. Wu, *Amino Acids*, 2013, **45**, 407–411.
- G. K. Jayaprakasha and B. S. Patil, *Talanta*, 2016, **153**, 268–277.
- R. Mandrioli, L. Mercolini and M. A. Raggi, *Anal. Bioanal. Chem.*, 2013, **405**, 7941–7956.
- M. M. Hasan, M. H. Razu, S. Akter, S. A. Mou, M. Islam and M. Khan, *RSC Adv.*, 2024, **14**, 22292–22303.
- D. Lane, R. Soong, W. Bermel, P. Ning, R. D. Majumdar, M. Tabatabaei-Anaraki, H. Heumann, M. Gundy, H. Bönisch, Y. L. Mobarhan, M. J. Simpson and A. J. Simpson, *ACS Omega*, 2019, **4**, 9017–9028.
- T. Tchpilov, K. Meyer and M. G. Weller, *Methods Protoc.*, 2023, **6**, 11.
- Y. Alnouti, I. L. Csanaky and C. D. Klaassen, *J. Chromatogr. B: Anal. Technol. Biomed. Life Sci.*, 2008, **873**, 209–217.
- C. X. Du and Z. Huang, *RSC Adv.*, 2019, **9**, 36539–36545.
- R. Zhao, B. Huang, G. Lu, S. Fu, J. Ying and Y. Zhao, *Int. J. Mol. Sci.*, 2023, **24**, 7332.
- S. Kambhampati, J. Li, B. S. Evans and D. K. Allen, *Plant Methods*, 2019, **15**, 1–12.
- G. Sharma, S. V. Attri, B. Behra, S. Bhisikar, P. Kumar, M. Tadgeja, S. Sharda, P. Singhi and S. Singhi, *Amino Acids*, 2014, **46**, 1253–1263.
- S. M. Eid, M. A. Farag and S. Bawazeer, *ACS Omega*, 2022, **7**, 31106–31114.
- S. A. Cohen, B. A. Bidlingmeyer and T. L. Tarvin, *Nature*, 1986, **320**, 769–770.
- Y. Sakaguchi, T. Kinumi, T. Yamazaki and A. Takatsu, *Analyst*, 2015, **140**, 1965–1973.
- S. A. Cohen, in *Amino Acid Analysis Protocols*, Humana Press, New Jersey, 2000, vol. 159, pp. 039–047.
- H. K. Mayer and G. Fiechter, *Anal. Bioanal. Chem.*, 2013, **405**, 8053–8061.
- M. J. González-Castro, J. López-Hernández, J. Simal-Lozano and M. J. Oruña-Concha, *J. Chromatogr. Sci.*, 1997, **35**, 181–185.
- C. Molins-Legua, P. Campíns-Falcó, A. Sevillano-Cabeza and M. Pedrón-Pons, *Analyst*, 1999, **124**, 477–482.
- N. Parris and D. Gallelli, *J. Liq. Chromatogr.*, 1984, **7**, 917–924.
- C. Gros and B. Labouesse, *Eur. J. Biochem.*, 1969, **7**, 463–470.
- W. R. Gray, *Methods Enzymol.*, 1972, **25**, 121–138.
- K. Stephens, PhD thesis, Ball State University, 1986.
- S. Bhat, A. Patil, L. Rai, V. B. Kartha and S. Chidangil, *Sci. World J.*, 2012, **2012**, 1–7.
- A. Patil, S. Bhat, L. Rai, V. B. Kartha and S. Chidangil, *AIP Conf. Proc.*, 2011, **1349**, 226–227.
- A. Patil, K. S. Choudhari, V. K. Unnikrishnan, N. Shenoy, R. Ongole, K. M. Pai, V. B. Kartha and S. Chidangil, *J. Biomed. Opt.*, 2013, **18**, 101317.



- 33 A. Patil, V. Prabhu, K. S. Choudhari, V. K. Unnikrishnan, S. D. George, R. Ongole, K. M. Pai, J. K. Shetty, S. Bhat, V. B. Kartha and S. Chidangil, *J. Biomed. Opt.*, 2010, **15**, 067007.
- 34 S. Bhat, A. Patil, L. Rai, V. B. Kartha and C. Santhosh, *J. Chromatogr. B: Anal. Technol. Biomed. Life Sci.*, 2010, **878**, 3225–3230.
- 35 X. Kang, J. Xiao, X. Huang and Z. Gu, *Clin. Chim. Acta*, 2006, **366**, 352–356.
- 36 R. Minocha and S. Long, *J. Chromatogr. A*, 2004, **1035**, 63–73.
- 37 R. V. John, T. Devasia, S. S. Adigal, J. Lukose, V. B. Kartha and S. Chidangil, *J. Chromatogr. B: Anal. Technol. Biomed. Life Sci.*, 2023, **1217**, 123616.
- 38 M. Sajid and J. Plotka-Wasyłka, *Talanta*, 2022, **238**, 123046.
- 39 J. Plotka-Wasyłka and W. Wojnowski, *Green Chem.*, 2021, **23**, 8657–8665.
- 40 N. Manousi, W. Wojnowski, J. Plotka-Wasyłka and V. Samanidou, *Green Chem.*, 2023, **25**, 7598–7604.
- 41 C. M. Jones, M. Smith and M. J. Henderson, *Ann. Clin. Biochem.*, 2006, **43**, 63–66.
- 42 H. Masoudi Rad, M. Rabiei, A. Sobhani, M. Sadegh Khanjani, M. Rahbar Taramsar and E. Kazemnezhad Leili, *J. Oral Rehabil.*, 2014, **41**, 759–767.

