Chemical Science



EDGE ARTICLE

View Article Online
View Journal | View Issue



Cite this: Chem. Sci., 2024, 15, 18170

dll publication charges for this article have been paid for by the Royal Society of Chemistry

Received 3rd July 2024 Accepted 11th October 2024

DOI: 10.1039/d4sc04404e

rsc.li/chemical-science

On-demand controlled bidirectional DNAzyme path for ultra-sensitive heavy metal ion detection†

Jing Xu,^a Yujin Li,^{ab} Futing Wang,^b Xinqi Luo,^a Wei Zhang,^f Yifan Lyu,^b Hongfen Yang,^e Ren Cai (1) *b and Weihong Tan (1) bcd

A bidirectional self-powered biosensor is constructed for the quasi-simultaneous detection of Pb^{2+} and Hg^{2+} based on $MoS_2@CuS$ heterostructures as an accelerator and hybridization chain reaction (HCR) as a signal amplification strategy. $MoS_2@CuS$ heterostructures significantly facilitate electron transfer between glucose and bioelectrodes, thereby greatly improving the detection signal of self-powered biosensors. This novel biosensor employs the unique sequences of DNAzymes to isolate Pb^{2+} and Hg^{2+} by the cleavage effect and thymine (T)– Hg^{2+} –thymine (T) structures, respectively. In the process, Pb^{2+} cuts the sequence of DNAzyme at the bioanode to trigger glucose oxidation to monitor Pb^{2+} . The asformed $T-Hg^{2+}-T$ structures activate HCR to reduce $[Ru(NH_3)_6]^{3+}$ to detect Hg^{2+} at the biocathode. It is noteworthy that this biosensor not only realizes Pb^{2+} or Hg^{2+} detection in a single-electrode, respectively, but also can quasi-simultaneously detect both Pb^{2+} and Hg^{2+} in the bioanode and the biocathode. The novel self-powered biosensor identifies Pb^{2+} in the range of 10^6 fM to 10 fM with a limit of detection (LOD) of 3.1 fM and Hg^{2+} in the range of 10^6 fM to 1 fM with an LOD of 0.33 fM.

Introduction

Heavy metal ion pollution is a severe hazard for human health because of the extremely low biodegradability of most heavy metals. At present, many novel detection methods are employed to monitor heavy metal ions, such as surface enhanced Raman spectroscopy (SERS), atomic absorption spectroscopy, inductively coupled plasma mass spectrometry, fluorescence spectroscopy, colorimetric method, and electrochemical methods. For example, an on-site visual approach was designed for toxic Cr(vi) detection based on an

imidazolium-functionalized conjugated polymer; ¹⁴ an aptamer-functionalized fluorescent DNA sensor was used to monitor $Cu(\pi)$ in living tumor cells; ¹⁵ an ultra-sensitive and selective electrochemical method was utilized for $Cd(\pi)$ detection by the synergistic activation of phosphorus and the orbital coupling effect of Fe-doped CoP. ¹⁶

Compared with the above methods, self-powered biosensors are widely used in the detection of biomolecules. In addition to eliminating the need for external power, these sensors exhibit strong anti-interference ability, ultra-sensitive detection capability, and highly identifiable specificity. Total However, there has been no report of self-powered biosensors with the bidirectional detection to monitor heavy metal ions. The major reason is the difficulty in forming double bioelectrodes to simultaneously achieve linear detection. Therefore, a bidirectional self-powered biosensor toward multiple heavy metal ions detection is urgently needed.

This work reports a bidirectional self-powered biosensor for the quasi-simultaneous detection of Pb²⁺ and Hg²⁺ based on MoS₂@CuS heterostructures as a co-accelerator and a hybridization chain reaction (HCR) as a signal amplification strategy (Scheme 1). Unique sequences of DNAzymes isolate Pb²⁺ and Hg²⁺ by the cleavage effect and thymine (T)–Hg²⁺–thymine (T) structures, respectively.^{25–29} In the process, Pb²⁺ cuts the sequence of the DNAzyme at the bioanode to trigger glucose oxidation to monitor Pb²⁺. The as-formed T–Hg²⁺–T structures activate HCR to reduce $[Ru(NH_3)_6]^{3+}$ to detect Hg²⁺ at the biocathode. As a result, this biosensor not only monitors Pb²⁺ or Hg²⁺ through the current changes caused by the oxidation–

^aCollege of Chemistry and Chemical Engineering, Xinyang Normal University, Xinyang 464000. China

^bMolecular Science and Biomedicine Laboratory, State Key Laboratory for Chemo/Bio-Sensing and Chemometrics, College of Material Science and Engineering, College of Chemistry and Chemical Engineering, College of Biology, Hunan University, Changsha, Hunan 410082, China. E-mail: cairen@hnu.edu.cn


^{&#}x27;The Cancer Hospital of the University of Chinese Academy of Sciences (Zhejiang Cancer Hospital), Hangzhou Institute of Medicine, Chinese Academy of Sciences, Hangzhou, Zhejiang 310022, China

⁴Institute of Molecular Medicine, Renji Hospital, Shanghai Jiao Tong University School of Medicine, College of Chemistry and Chemical Engineering, Shanghai Jiao Tong University, Shanghai 200240, China

^eHunan Key Laboratory of Typical Environmental Pollution and Health Hazards, School of Public Health, Hengyang Medical School, University of South China, Hengyang 421001, China

^fCollege of Life Science, Xinyang Normal University, Dabie Mountain Laboratory, Xinyang 464000, China

[†] Electronic supplementary information (ESI) available. See DOI https://doi.org/10.1039/d4sc04404e



Scheme 1 (a) The synthetic process of MoS_2 @CuS heterostructures. (b and c) The detection process of Pb^{2+} and Hg^{2+} of the bidirectional self-powered biosensor.

reduction reactions in a single-electrode, respectively, but also could quasi-simultaneously detect both Pb^{2+} and Hg^{2+} by observing the open circuit voltage (E^{ocv}) in the bioanode and the biocathode. The bidirectional biosensor also exhibits considerable practicability for detecting Pb^{2+} and Hg^{2+} in tap water.

Results and discussion

In the TEM images (Fig. 1a and S1†), uniform Cu₂O templates, CuS nanocubes, and MoS₂@CuS heterostructures (size of ~500 nm) are observed, and there are many small nanosheets on the surface of MoS2@CuS heterostructures. The selected area electron diffraction pattern (SAED) confirms the polycrystalline feature (inset, Fig. 1a), which is consistent with the X-ray diffraction (XRD) pattern (Fig. 1b). All peaks are indexed to MoS₂ (JCPDS: 37-1492) and CuS (JCPDS: 06-0462).30,31 After adding Au nanoparticles (AuNPs, ~20 nm, 5 mM, Fig. S2†), many AuNPs are observed on the surface of MoS2@CuS heterostructures (Fig. 1c). The elemental mapping images and X-ray photoelectron spectra (XPS) confirm that Au, Cu, Mo, and S are present (Fig. 1d and S3†).32 In addition, UV-vis spectra and FTIR spectroscopy confirmed the successful recombination of heterostructures. For MoS2@CuS heterostructures, the FTIR absorption peaks at 1605 and 895 cm⁻¹ were to the vibration of Mo-O bond and Mo-S bond, respectively (Fig. S4 and S5†).33 Finally, the Brunauer-Emmett-Teller (BET) specific surface of MoS₂@CuS heterostructures is calculated to be 30.8 cm² g⁻¹ with aperture size of \sim 5 nm (Fig. S6†).

The assembly process of the bioanode and biocathode was explored by cyclic voltammetry (CV) and differential pulse

voltammetry (DPV). Monitoring the maximum current value, the optimal recognition and cleavage time of Pb²⁺ on the DNAzyme was 60 min (Fig. 2a). After DNAzyme was cut by Pb²⁺, it was split into two sequences of single strand DNAs: S1 and S2, leaving as-split S2 on the surface of AuNPs/MoS₂@CuS/CP by Au–S bonding. When S3-GOD is added, the as-split S2 hybridizes with S3-GOD by complementary base pairing to form the bioanode. The optimal hybridization time between S2 and S3-GOD is 50 min (Fig. 2b). In the biocathode, the optimal incubation time among Hg²⁺, S5, and S4 is 30 min to form S5–Hg²⁺/S4/AuNPs/MoS₂@CuS/CP by the T–Hg²⁺–T structure (Fig. 2c). Finally, the as-formed T–Hg²⁺–T structure further combines with H1 and H2 from the HCR process to generate a multiduplex structure of H1, H2, and S5 to form H1, H2/S5–Hg²⁺/S4/AuNPs/MoS₂@CuS/CP.

Next, $[Ru(NH_3)_6]^{3^+}$ inserts into the multi-duplex structure of H1, H2, and S5 on the surface of H1, H2/S5–Hg²⁺/S4/AuNPs/MoS₂@CuS/CP (Fig. 2d) with optimal concentration of 500 μ M $[Ru(NH_3)_6]^{3^+}$. These assembly processes of the bioanode and biocathode were further verified by gel electrophoresis (Fig. 2e). Line M represents molecular weight standards. For the bioanode, a fading, shallow band with low molecular weight appears when Pb²⁺ (lane 2) is incubated with DNAzyme (lane 1). After S3-GOD is added, S3-GOD hybridizes with as-split S2 to form many new bright bands (lane 3). For the biocathode, there is no visible band for S4 (lane 4) or S4 + S5 (lane 5), respectively. When the T-Hg²⁺–T (*i.e.*, S4–Hg²⁺–S5) structure is formed, a higher bright band (lane 6) is observed. Finally, when the T–Hg²⁺–T structure triggers the HCR reaction, a new band is observed (lane 7). These results are consistent with circular dichroic chromatography




Fig. 1 Characterization of MoS₂@CuS heterostructures: (a) TEM image and SAED (insert); (b) XRD. (c) TEM image and (d) EDX images of AuNPs/ MoS₂@CuS heterostructures.

(Fig. S7†). In the sensing mechanism, the Pb²⁺ could cleave the RNA sites and change the DNAzyme structure; the thyminethymine mismatch base pair (T-T, H1-H2) can capture the Hg²⁺ to form a stable neutral T-Hg²⁺-T base pair.

Electrochemical impedance spectroscopy (EIS) was used to study the preparation process of the bioanode (Fig. 3a). The resistance $(R_{\rm et})$ value of CP (curve A) is higher than that of AuNPs/MoS₂@CuS/CP (curve B), confirming the good electrical conductivity of AuNPs/MoS₂@CuS. After DNAzyme, Pb²⁺, and S3-GOD are sequentially incubated with AuNPs/MoS₂@CuS/CP, the $R_{\rm et}$ increases from 14 Ω to 30 Ω (Fig. 3a). For the biocathode (Fig. 3b), H1, H2/S5-Hg²⁺/S4/AuNPs/MoS₂@CuS/CP shows the highest Ret after S4, S5 and Hg²⁺, and H1, H2 are modified the biocathode. In addition, the ζ-potential measurements confirmed the successful assembly of the bioanode and biocathode (Fig. S8†). CV and linear scanning voltammetry (LSV) were performed to verify the catalytic reaction of GOD on the bioanode. In PBS with 5 mM glucose, a distinct oxidation peak at -0.48 V is observed (blue line), which corresponds to the splitting of DNAzyme to S1 and S2 by Pb²⁺, and the as-split S2

triggers the oxidation of glucose (Fig. 3c and d). The feasibility of the biocathode for Hg²⁺ detection was investigated by DPV. In PBS with 500 μ M [Ru(NH₃)₆]³⁺, as-adsorbed [Ru(NH₃)₆]³⁺ acquires electrons and is reduced to [Ru(NH₃)₆]²⁺ at the biocathode. When AuNPs/MoS2@CuS (black line), S4 (blue line), S5, Hg²⁺ (green line), and H1, H2 (red line) are successively coated on the CP surface, H1, H2/S5-Hg²⁺/S4/AuNPs/MoS₂@-CuS/CP exhibits the largest peak current value (Fig. 3e). Finally, the current of DPV was increased when the biocathode combined with the HCR strategy (Fig. S9†), which confirmed the signal amplification effect of HCR. These results prove the efficient signal amplification strategy from HCR for the biocathode.

In Fig. 4a and b, there is a linear relationship between the peak current with increasing scan rate in the range of 60 to 200 mV s⁻¹ (peak potential difference ≤ -20 mV), which proves that the as-prepared bioanode and biocathode exhibit accurate identification ability.34,35 Here, one biosensor was constructed based AuNPs/CP as the bioanode and the biocathode; the other one was fabricated based on AuNPs/MoS2@CuS/CP as the




Fig. 2 The optimization of the hybridization reaction times: (a) incubation between Pb²⁺ and DNAzyme; (b) incubation between S3-GOD and assplit S2; (c) incubation among S4, S5 and Hg²⁺; (d) the optimization of $[Ru(NH_3)_6]^{3+}$ concentration. (e) Gel electrophoresis analysis.




Fig. 3 (a) EIS of the bioanode: bare CP (A), AuNPs/MoS2@CuS/CP (B), DNAzyme/AuNPs/MoS2@CuS/CP (C), DNAzyme/AuNPs/MoS2@CuS/CP (D), Pb²⁺/BSA/DNAzyme/AuNPs/MoS₂@CuS/CP (E), S3-GOD/Pb²⁺/DNAzyme/AuNPs/MoS₂@CuS/CP (F) (inset: the circuit diagram). (b) EIS of the bio $cathode: bare CP (A), AuNPs/MoS_2@CuS/CP (B), S4/AuNPs/MoS_2@CuS/CP (C), BSA/S4/AuNPs/MoS_2@CuS/CP (D), S5-Hg^2+/S4/AuNPs/MoS_2@CuS/CP (D), S5-Hg^2+/AuNPS/MoS_2@CuS/CP ($ CP (E), H1, H2/S5-Hg²⁺/S4/AuNPs/MoS₂@CuS/CP (F) (inset: the circuit diagram). (c) CV and (d) LSV curves of the bioanode in PBS (pH 7.4) without glucose and with 5 mM glucose in the presence of Pb²⁺. (e) DPV curves of the assembly process of the biocathode in PBS (pH 7.4) with [Ru(NH₃)₆]³⁺.




Fig. 4 (a and b) CV of S3-GOD/Pb²⁺/DNAzyme/AuNPs/MoS₂@CuS/CP (*i.e.*, bioanode) (yellow to red lines) and H1, H2/S5 Hg^{2+} /S4/AuNPs/MoS₂@CuS/CP (*i.e.*, biocathode) (pink to blue lines) with different scan rates (ν): (a–h) 60, 80, 100, 120, 140, 160, 180, 200 mV s⁻¹. (c) E^{ocv} and (d) power output of the self-powered biosensor constructed of: (A) both S3-GOD/Pb²⁺/DNAzyme/AuNPs/MoS₂@CuS/CP (the biocathode) and H1, H2/S5-Hg²⁺/S4/AuNPs/MoS₂@CuS/CP (the biocathode), and (B) both S3-GOD/Pb²⁺/DNAzyme/CP (the biocanode) and H1, H2/S5-Hg²⁺/S4/CP (the biocathode).

bioanode and the biocathode. The $E^{\rm ocv}$ of the biosensor based on AuNPs/MoS₂@CuS/CP is about 0.58 V, which is 0.13 V higher than that of the biosensor based on AuNPs/CP (Fig. 4c). Similarly, the maximum power output of the biosensors based on AuNPs/MoS₂@CuS/CP and AuNPs/CP is 49.5 μ W cm⁻² and 33.2 μ W cm⁻², respectively (Fig. 4d), which proves high-efficiency conduction of MoS₂@CuS heterostructures.

For the single detection of Pb²⁺, the peak intensity of glucose oxidation at -0.48 V increases with increasing Pb²⁺ concentration under optimal conditions in a CV test at the bioanode (Fig. S10†). Moreover, there is a good linear relationship between the logarithm of Pb²⁺ concentration and current intensity in the range of 10 fM \sim 1 nM. The equation is: $I=0.01\times \log C_{\rm Pb}^{2+}+0.20$ (correlation coefficient, $R^2=0.995$), and the detection of limit (LOD) is calculated to be 3.1 fM (S/N=3). Analogously, the current response at -0.18 V is attributed to the reduction of [Ru(NH₃)₆]³⁺ in the DPV test at the biocathode, which gradually increases with increasing Hg²⁺ concentration (Fig. S11†). A linear correlation between the logarithm of Hg²⁺

concentration and current intensity in the range of 1 fM to 1 nM was fitted by the regression equation: $I = 0.1 \times \log C_{\rm Hg}^{2+} + 2.4$ ($R^2 = 0.995$), and LOD is calculated to be 0.32 pM (S/N = 3).

For the bidirectional detection of Pb^{2+} and Hg^{2+} , E^{ocv} of the novel biosensors was investigated at the double electrodes (Fig. 5). For instance, Hg²⁺ concentration is determined to be 1 nM at the biocathode in a tap water sample (Fig. 5a and b). A good linear relationship between E^{ocv} and the logarithm of Pb²⁺ concentration is calculated: $E^{\text{ocv}} = 0.078 \log C_{\text{Pb}}^{2+} + 1.21 (R^2 =$ 0.994) with the LOD of 3.1 fM (S/N = 3) in the range of 10 fM to 10^6 fM. Here, the E^{ocv} increases with the increment of Pb²⁺ concentration at the bioanode. In another example, Pb2+ concentration is determined to be 1 nM at the bioanode in the same tap water sample (Fig. 5c and d). There is a linear relationship between $E^{
m ocv}$ and the logarithm of ${
m Hg^{2+}}$ concentration in the range of 1 fM to 10^6 fM: $E^{\text{ocv}} = 0.083 \times \log C_{\text{Hg}}^{2+} + 1.34 (R^2)$ = 0.995), and the LOD is calculated to be 0.33 fM (S/N = 3). Here, the E^{ocv} increases with the increment of Hg^{2+} concentration at the biocathode. Compared with previously reported

sensors (Tables S2 and S3†), the bidirectional self-powered biosensor shows wider detection ranges and lower LODs for Pb²⁺ and Hg²⁺ detection. In order to further visualize the on-site detection of Pb²⁺ and Hg²⁺, our capacitor amplification strategy was explored by smart Bluetooth data transfer with a smartphone for real-time read-out of test results (Fig. S12 and see Movie S1†).^{36–38} In the test process, the mobile phone reads out the instantaneous current when different concentrations of Pb²⁺ and Hg²⁺ are in an unknown water sample, furthering realizing on-site and real-time detection (Fig. S13 and S14†).

The signal responses of the self-powered biosensor toward Pb²⁺ and Hg²⁺ detection remained at 83.5% and 82.1% of

original signal after the biosensors were stored at 4 °C for 15 days, further confirming the long-time stability (Fig. S15†). The recyclability of one as-designed biosensor was tested by analyzing different concentrations of Pb²⁺ and Hg²⁺. All test results maintained within 5% of the error, which confirmed good recyclability for the biosensor (Fig. S16†). To confirm the specificity, there was no E^{ocv} signal when six possible interferences, *i.e.*, Na⁺, K⁺, Mg²⁺, Cu²⁺, Co²⁺, and blank water sample, were detected by the as-designed biosensor. The solution with Pb²⁺ or Hg²⁺ showed significantly high E^{ocv} , thus verifying the good specificity of the bidirectional biosensor (Fig. S17†). In addition, the relative standard deviations (RSDs) were




Fig. 5 (a) E^{ocv} of different concentrations of Pb²⁺ and (b) relationship between E^{ocv} and Pb²⁺ concentration; (c) E^{ocv} of different concentrations of Hg²⁺ and (d) relationship between E^{ocv} and Hg²⁺ concentration. (e and f) The repeatability of the self-powered biosensor.

measured to be \leq 5% in five parallel experiments, indicating the good repeatability of the novel biosensor (Fig. 5e and f). To further test the practical application, Pb²⁺ and Hg²⁺ with the concentrations of 1, 10, 50, and 100 nM in tap water and lake water were prepared to perform a labelled recovery experiment. The RSDs values were calculated to be the range from 3.7% to 5.3% for Pb²⁺ and 2.1% to 5.9% for Hg²⁺; and the recovery rates were in the range from 96.4% to 105.0% for Pb²⁺ and 97.0% to 104.8% for Hg²⁺ (Tables S4 and S5†). These results demonstrate that the bidirectional self-powered biosensor possesses good selectivity, stability, and reproducibility, and could be used for practical sample detection.

Conclusions

In summary, a bidirectional self-powered biosensor is designed to simultaneously monitor Pb2+ at the bioanode and Hg2+ at the biocathode. Here, the MoS2@CuS heterostructures are successfully prepared to accelerate the electron transfer between glucose and the bioelectrode to increase the power output. Moreover, the novel biosensor realizes Pb²⁺ detection by the cutting effect of Pb2+ toward DNAzyme to trigger glucose oxidation at the bioanode, and monitors Hg^{2+} by forming unique T-Hg²⁺-T structures to activate HCR for the reduction of [Ru(NH₃)₆]³⁺ at the biocathode. The self-powered biosensor could not only perform linear detection of Pb²⁺ at the bioanode or Hg²⁺ at the biocathode, but also quasi-simultaneously realizes linear detection for both Pb2+ and Hg2+ at the bioanode and the biocathode. The bidirectional biosensor exhibits wide detection ranges of 10^6 fM $\sim \! 10$ fM toward Pb²⁺ and 10^6 fM $\sim \! 1$ fM toward Hg^{2+} , respectively; and the LODs of Pb^{2+} and Hg^{2+} are as low as 3.1 fM and 0.33 fM, respectively. This biosensor possesses highly efficient, ultra-sensitive, highly selective, and good reproducible detection of Pb2+ and Hg2+. Meanwhile, smart and portable mobile phones and capacitors are well used for on-site direct reading detection. This work opens a new direction for designing novel bidirectional biosensors to detect multiple heavy metal ions.

Data availability

All underlying data are available in the published article itself and its ESI.†

Author contributions

J. X., F. W., X. L., W. Z., and H. Y. directed the project. Y. L performed the main experimental work. F. L provided experimental equipment. R. C., and W. T. provided some constructive suggestions for the experiments. All the authors discussed the experimental results.

Conflicts of interest

There are no conflicts to declare.

Acknowledgements

This work was supported by the Joint Fund Project for Technology Research and Development Plan of Henan Technology (Young Scientists, 235200810056), Hunan Provincial Natural Science Foundation of China (2024JJ5348 and 2021JJ20020), and National Natural Science Foundation of China (22004032). The authors also acknowledge the great support from the Dabie Mountain Laboratory, Nanhu Scholars Program for Young Scholars of XYNU and Analysis Testing Center of Xinyang Normal University for materials characterization.

References

- 1 S. H. Chen, J. J. Zhu, P. H. Li, Y. F. Sun, M. Yang and X. J. Huang, In-situ growth of zero-valent iron in FeOx/Mn₃O₄ to improve the surficial redox for high-efficient electrocatalysis of Pb(II), *Chem. Eng. J.*, 2022, **430**, 132959.
- 2 A. Amalraj, R. Pavadai and P. Perumal, Recyclable target metal-enhanced fluorometric naked eye aptasensor for the detection of Pb²⁺ and Ag⁺ ions based on the structural change of CaSnO₃@PDANS-constrained GC-rich ssDNA, *ACS Omega*, 2021, **6**(45), 30580–30597.
- 3 J. Xu, Y. J. Li, Y. B. Liu, X. Wu and K. J. Huang, Ultra-sensitive self-powered sensor based on the hollow Rubik's cube substrate and multiple signal amplification strategy for real-time heavy metal detection, *Sens. Actuators, B*, 2023, **394**, 134377.
- 4 M. Rana, M. Balcioglu, N. M. Robertson, M. S. Hizir, S. Yumak and M. V. Yigit, Low picomolar, instrument-free visual detection of mercury and silver ions using low-cost programmable nanoprobes, *Chem. Sci.*, 2017, **8**(2), 1200–1208.
- 5 T. T. X. Ong, E. W. Blanch and O. A. H. Jones, Surface enhanced Raman spectroscopy in environmental analysis, monitoring and assessment, *Sci. Total Environ.*, 2020, 720, 137601.
- 6 İ. Narin, M. Soylak, L. Elçi and M. Doğan, Determination of trace metal ions by AAS in natural water samples after preconcentration of pyrocatechol violet complexes on an activated carbon column, *Talanta*, 2000, 52(6), 1041–1046.
- 7 T. Van Acker, S. Theiner, E. Bolea-Fernandez, F. Vanhaecke and G. Koellensperger, Inductively coupled plasma mass spectrometry, *Nat. Rev. Methods Primers*, 2023, 3(1), 52.
- 8 H. Yuan, T. Ren, Q. Luo, Y. Huang, Y. Huang, D. Xu, X. Guo, X. Li and Y. Wu, Fluorescent wood with non-cytotoxicity for effective adsorption and sensitive detection of heavy metals, *J. Hazard. Mater.*, 2021, **416**, 126166.
- 9 J. Liu, L. Zhang, W. Zeng, L. Zhang, N. He and Z. Lu, High-throughput quantitative detection of triple-negative breast cancer-associated expressed miRNAs by rolling circle amplification on fluorescence-encoded microspheres, *Chin. Chem. Lett.*, 2023, 34(9), 108141.
- 10 R. Pavadai, A. Amalraj, S. Subramanian and P. Perumal, High catalytic activity of fluorophore-labeled Y-shaped DNAzyme/3D MOF-MoS₂NBs as a versatile biosensing platform for the simultaneous detection of Hg²⁺, Ni²⁺, and

Ag⁺ ions, ACS Appl. Mater. Interfaces, 2021, **13**(27), 31710-31724.

Edge Article

- 11 Y. Liu, T. Li, G. Yang, Y. Deng, X. Mou and N. He, A simple AuNPs-based colorimetric aptasensor for chlorpyrifos detection, *Chin. Chem. Lett.*, 2022, 33(4), 1913–1916.
- 12 M. Yang, P. Li, S. Chen, X. Xiao, X. Tang, C. Lin, X. Huang and W. Liu, Heavy metal detection: nanometal oxides with special surface physicochemical properties to promote electrochemical detection of heavy metal ions, *Small*, 2020, **16**(25), 2070140.
- 13 G. Yang, Z. Li, R. Usman, Z. Chen, Y. Liu, S. Li, H. Chen, Y. Deng, Y. Fang and N. He, DNA walker induced "signal on" fluorescence aptasensor strategy for rapid and sensitive detection of extracellular vesicles in gastric cancer, *Chin. Chem. Lett.*, 2024, 109930.
- 14 Z. Li, S. Shen, S. Hussain, B. Cui, J. Yao, Y. Su, Y. Wang, Y. Hao and R. Gao, Imidazolium-functionalized conjugated polymer for on-site visual and ultrasensitive detection of toxic hexavalent chromium, *Anal. Chem.*, 2024, 96(13), 5150–5159.
- 15 Y. Mou, P. Yin, M. Chen, C. Wei, Y. Zhang, J. Zhang, Y. Zhao, X. Luo and Y. Wang, Engineering of an aptamer-functionalized fluorescent DNA sensor for Cu(II) responding in living tumor cells, *Anal. Chem.*, 2023, 95(21), 8348–8356.
- 16 Z. Gao, Y. Li, P. Li, Y. Yang, Y. Zhao, M. Yang, S. Chen, Z. Song and X. Huang, Synergistic activation of P and orbital coupling effect for ultra-sensitive and selective electrochemical detection of Cd(II) over Fe-doped CoP, *J. Hazard. Mater.*, 2024, 463, 132842.
- 17 Y. J. Li, J. Xu, Y. Y. Hou, J. Q. Lu, K. J. Huang and R. Cai, 3D hierarchically electrode combined with DNA circuit strategy powered highly sensitive sensing devices, *Sens. Actuators, B*, 2024, **401**, 134963.
- 18 C. Gu, L. Bai, L. Pu, P. Gai and F. Li, Highly sensitive and stable self-powered biosensing for exosomes based on dual metal-organic frameworks nanocarriers, *Biosens. Bioelectron.*, 2021, **176**, 112907.
- 19 C. Gu, P. Gai, X. Kong, T. Hou and F. Li, Self-powered biosensing platform based on "Signal-On" enzymatic biofuel cell for DNA methyltransferase activity analysis and inhibitor screening, *Anal. Chem.*, 2020, 92(7), 5426–5430.
- 20 J. Xu, Y. Liu, K. Huang, Y. Hou, X. Sun and J. Li, Real-time biosensor platform based on novel sandwich graphdiyne for ultrasensitive detection of tumor marker, *Anal. Chem.*, 2022, **94**(49), 16980–16986.
- 21 Y. Hou, J. Xu, W. Xie, K. Huang, X. Tan, B. Zhao, S. Zhang and M. Gao, 3D DNA walker recognition-driven homogeneous dual-mode sensing strategy based on enzyme biofuel cell for ultrasensitive detection of HER2, *Sens. Actuators, B*, 2023, **376**, 132998.
- 22 J. Fang, C. Yuan, J. Li, J. Li, T. Yang, Y. Guo, D. Wang, J. Xue, W. Fu and G. Xie, An enzyme-powered, three-dimensional lame DNA walker, *Biosens. Bioelectron.*, 2021, 177, 112981.
- 23 J. Xu, Y. Li, F. Wang, H. Yang, K. Huang, R. Cai and W. Tan, A smartphone-mediated "all-in-one" biosensing chip for

- visual and value-assisted detection, *Anal. Chem.*, 2024, **96**(39), 15780–15788.
- 24 A. Arunjegan, P. Rajaji, S. Sivanesan and P. Panneerselvam, A Turn-ON fluorometric biosensor based on ssDNA immobilized with a metal phenolic nanomaterial for the sequential detection of Pb(II) and epirubicin cancer drug, *RSC Adv.*, 2021, **11**(20), 12361–12373.
- 25 Y. Miyake, H. Togashi, M. Tashiro, H. Yamaguchi, S. Oda, M. Kudo, Y. Tanaka, Y. Kondo, R. Sawa, T. Fujimoto, T. Machinami and A. Ono, MercuryII-mediated formation of thymine-HgII-thymine base pairs in DNA duplexes, *J. Am. Chem. Soc.*, 2006, 128(7), 2172–2173.
- 26 A. Ono and H. Togashi, Highly selective oligonucleotide-based sensor for Mercury(II) in aqueous solutions, *Angew. Chem., Int. Ed.*, 2004, **43**(33), 4300–4302.
- 27 S. W. Santoro and G. F. Joyce, A general purpose RNA-cleaving DNA enzyme, *Proc. Natl. Acad. Sci. U. S. A.*, 1997, 94(9), 4262–4266.
- 28 Y. Fang, Y. Wang, L. Zhu, H. Liu, X. Su, Y. Liu, Z. Chen, H. Chen and N. He, A novel cartridge for nucleic acid extraction, amplification and detection of infectious disease pathogens with the help of magnetic nanoparticles, *Chin. Chem. Lett.*, 2023, 34(8), 108092.
- 29 R. Huang, L. He, L. Jin, Z. Li, N. He and W. Miao, Recent advancements in DNA nanotechnology-enabled extracellular vesicles detection and diagnosis: A mini review, *Chin. Chem. Lett.*, 2023, 34(6), 107926.
- 30 M. Li, L. Wang, Y. Qian and J. Du, Facile synthesis of MoS₂/CuS nanoflakes as high performance electrocatalysts for hydrogen evolution reaction, *Int. J. Hydrogen Energy*, 2022, 47(8), 5319–5325.
- 31 H. Liu, Y. He, H. Zhang, S. Wang, K. Cao, Y. Jiang, X. Liu and Q. S. Jing, Heterostructure engineering of ultrathin SnS₂/Ti₃C₂T_x nanosheets for high-performance potassium-ion batteries, *J. Colloid Interface Sci.*, 2022, **606**, 167–176.
- 32 J. W. Stiles, A. L. Soltys, X. Song, S. H. Lapidus, C. B. Arnold and L. M. Schoop, Unlocking high capacity and fast Na⁺ diffusion of H_xCrS₂ by proton-exchange pretreatment, *Adv. Mater.*, 2023, 35(10), 2209811.
- 33 A. Amalraj, R. Pavadai, S. Subramanian and P. Perumal, Fabrication of multi-functional CuO@PDA-MoS₂ mediated dual-functional fluorescence aptamer for the detection of Hg²⁺ ions and chloramphenicol through desulfurization cleavage reaction and exonuclease I activity, *Appl. Surf. Sci.*, 2022, **602**, 154222.
- 34 H. Chen, X. Ma, X. Zhang, G. Hu, Y. Deng, S. Li, Z. Chen, N. He, Y. Wu and Z. Jiang, Novel aerosol detection platform for SARS-CoV-2: Based on specific magnetic nanoparticles adsorption sampling and digital droplet PCR detection, *Chin. Chem. Lett.*, 2023, 34(1), 107701.
- 35 Z. Guo, B. Jin, Y. Fang, L. Jin, S. Li, Y. Deng, Z. Chen, H. Chen, Y. Zhang, R. Usman and N. He, Automated screening of primary cell-based aptamers for targeting and therapy of pancreatic cancer, *Chin. Chem. Lett.*, 2024, 35(2), 108528.
- 36 F. Wang, P. Wang, H. Yang, R. Cai and W. Tan, Self-powered biosensing system with multivariate signal amplification for

Chemical Science

- real-time amplified detection of PDGF-BB, *Anal. Chem.*, 2023, **95**(44), 16359–16365.
- 37 Z. Guo, B. Jin, Y. Fang, Y. Deng, Z. Chen, H. Chen, S. Li, F. W. N. Chow, P. H. M. Leung, H. Wang, L. Cai and N. He, Selected aptamer specially combing 5-8F cells based on automatic screening instrument, *Chin. Chem. Lett.*, 2022, 33(9), 4208–4212.
- 38 F. Wang, H. Luo, Y. Hou, Y. Ya, K. J. Huang, S. Li, L. Wang, Y. Song and X. Tan, Dual microRNAs-driven enzyme biofuel cells: Visualization biosensing system with capacitor assistance output signal amplification, *Sens. Actuators, B*, 2022, 365, 131970.