



Cite this: *Analyst*, 2025, **150**, 1022

DOI: 10.1039/d5an90011e
 rsc.li/analyst

Correction: Infrared spectromicroscopy of biochemistry in functional single cells

Luca Quaroni *^a and Theodora Zlateva^b

Correction for 'Infrared spectromicroscopy of biochemistry in functional single cells' by Luca Quaroni et al., *Analyst*, 2011, **136**, 3219–3232, <https://doi.org/10.1039/C1AN15060J>.

The authors regret that there were multiple errors in the original article.

There was an error in Table 1 where column 4 row 2 reported the same value as column 5 row 2 (3.6 au), instead of the correct one (1.206 au). The correct version of Table 1 is presented here.

Table 1 Vibrational spectroscopic data of H₂O and D₂O, from ref. 54. ω_0 : peak frequency, expressed in wavenumber; ϵ_0 molar absorption coefficient; $A_{10\ \mu\text{m}}$: absorbance for a 10 μm pathlength; l_{opt} : optimal path length, at which the S/N ratio is maximized. ν_S : symmetric stretching mode; ν_{AS} : antisymmetric stretching mode; δ : bending mode; $\delta + \eta$: combination bending and libration mode

Vibrational mode	ω_0/cm^{-1}	$\epsilon_0/\text{M}^{-1}\ \text{cm}^{-1}$	$A_{10\ \mu\text{m}}/\text{au}$	$l_{\text{opt}}/\mu\text{m}$
$\nu_S, \text{H}_2\text{O}; \nu_{AS}, \text{H}_2\text{O}$	3404.0	99.9 ± 0.8	5.53	0.8
$\delta_{\text{H}_2\text{O}}$	1643.5	21.8 ± 0.3	1.206	3.6
$\delta + \eta_{\text{H}_2\text{O}}$	2127.5	3.50 ± 0.1	0.194	22.4
$\nu_S, \text{D}_2\text{O}; \nu_{AS}, \text{D}_2\text{O}$	2504.0	71.5 ± 0.4	3.94	1.1
$\delta_{\text{D}_2\text{O}}$	1209.4	17.4 ± 0.2	0.962	4.5
$\delta + \eta_{\text{D}_2\text{O}}$	1555.0	1.91 ± 0.05	0.105	41.2

There is also an error in a sentence in the section "Qualitative analysis: detection of analytes in static samples" on page 3228 of the original article. The sentence "Most of these applications were performed using a conventional global light source, ..." should be replaced with "Most of these applications were performed using a conventional thermal light source, ...".

The Royal Society of Chemistry apologises for these errors and any consequent inconvenience to authors and readers.

^aSwiss Light Source, Paul Scherrer Institut, Villigen-PSI, CH-5232, Switzerland. E-mail: luca.quaroni@psi.ch

^bDepartment of Biochemistry & Saskatchewan Cancer Center, Saskatoon, SK, S7N 5E5, Canada

