



Cite this: *Lab Chip*, 2025, 25, 1744

Label-free differentiation of living *versus* dead single yeast cells using broadband electrical impedance spectroscopy†

Amirhossein Favakeh,^a Amir Mokhtare,^a Mohammad Javad Asadi,^b James C. M. Hwang^{bc} and Alireza Abbaspourrad^{id} *^a

The use of the intrinsic electrical properties of a single cell by broadband electrical impedance spectroscopy (EIS) as a label-free and non-invasive method to monitor cellular and intracellular features is an emerging field. Here, we present a novel EIS-based sheathless microfluidic platform with an integrated coplanar waveguide to probe the interior of a single cell. This platform allows for precise single-cell trapping by dielectrophoresis, hydrodynamic focusing, and sensing the electrical properties of the trapped single cell. We measured the impedance characteristics of a single *Schizosaccharomyces pombe* (fission) yeast cell by a single frequency sweep (30 kHz to 6GHz) in a stagnant sucrose solution using two-port scattering (S) parameters. The measurements revealed a clear distinction between the cytoplasm impedance of live *versus* dead cells at 3 GHz. This platform could provide real-time monitoring of cellular electrical responses to chemical and physical antagonists for diagnostic purposes.

Received 13th January 2025,
Accepted 23rd February 2025

DOI: 10.1039/d5lc00043b

rsc.li/loc

Introduction

Single-cell biophysical characterization has gained considerable attention among biologists who seek to understand the mechanisms and processes that take place within this small unit of life.¹ Biophysical characterization of a single cell provides a wealth of information, including viability,² morphology,³ mechanical properties,⁴ and physiological state,^{5,6} information that is valuable in various fields of research such as food,⁷ medicine chemistry,⁸ biology,⁹ and environmental monitoring.¹⁰

Interest in identifying the electric properties of cells has gained recent attention using various methods.¹¹ Dielectric-based spectroscopy techniques such as impedance flow cytometry (IFC) have been developed to characterize cells in a high throughput manner, such as the impedance flow cytometry framework for single-cell analysis proposed by Feng *et al.*^{12,13} Alternatively, impedance-based spectroscopic techniques, such as electrical impedance spectroscopy (EIS) have been developed in which most of the electric field passes through the cells.¹⁴ EIS of single cells stands out as a

biosensing, label-free, and noninvasive technique. These biosensors are frequency dependent such that dispersion of a frequency spectrum can be allocated to different subcellular parts, revealing information about the cell's complex biophysical characterization.^{15,16} One promising tool that has been used to measure cell properties is a resonator.¹⁷ The main drawback of resonators is that they can only be used within a limited range of frequencies and are not easily tunable.² Conversely, coplanar waveguides (CPW) are suitable for use in a wide range of frequencies,^{18,19} and are sensitive to the cell impedance measurements.

EIS measurements are sensitive to the ion concentration between the electrodes and the cell,^{20,21} therefore, signals barely pass through the cell membrane at low frequency.²² To diminish this effect and to avoid electrode polarization complications at low frequencies, broadband impedance spectroscopy has been used to characterize living cells due to its dampened dispersion of ion movement at gigahertz frequencies.²² Regardless of the type of spectroscopy, sensitivity remains a concern.

To boost the sensitivity in single-cell measurement, low-cost microfluidic channels have been integrated with microwave-based sensing structures.^{23–26} These devices have increased the cell-to-sample volume ratios, and stabilized fluid flow for cell analysis.²⁷ As such, cells are guided through the channels to the electrode sensing area to analyze their biophysical characteristics.²⁸ Li *et al.* have been able to differentiate a bulk concentration of live and heat-killed *Escherichia coli* (*E. coli*) by impedance spectroscopy in the

^a Food Science Department, College of Agriculture and Life Sciences (CALS), Cornell University, Ithaca 14853, New York, USA. E-mail: alireza@cornell.edu

^b School of Electrical and Computer Engineering, Cornell University, Ithaca, New York 14853, USA

^c Department of Materials Science and Engineering, Cornell University, Ithaca, New York 14853, USA

† Electronic supplementary information (ESI) available. See DOI: <https://doi.org/10.1039/d5lc00043b>



frequency range of 0.5 to 20 GHz using a CPW integrated with a microfluidic platform.²¹ However, the main disadvantage of bulk measurement is that it does not reveal detailed information about a single cell within the population.

To monitor and analyze a single cell continuously, methods including hydrodynamic capture,^{26,29} optical tweezers,³⁰ and dielectrophoretic (DEP) trapping^{27,31} have been used to immobilize the cell. DEP, which is a non-uniform electric field that causes a polarization effect on the cell, offers a high throughput and simple means for single-cell trapping. Accordingly, most of the electrical signal can be distributed throughout the trapped cell. Therefore, the electrical properties of a single cell can be measured within well-defined impedance matching.³² Recently, Ferguson *et al.* successfully trapped a single muscle cell between two sensing sensors of a CPW using DEP.³³ They used broadband EIS in a sheath flow microfluidic channel to measure the cell impedance characteristics in a frequency range of 9 kHz to 9 GHz. In addition to cell trapping, the separation of viable and non-viable cells by the differences in electrical conductivity is another advantage of DEP.³⁴ Zhao *et al.* separated live from dead yeast cells using low frequency adjustment in DEP.³⁵ Therefore, the benefit of broadband electrical measurement is its wide frequency range. It can separate live cells from dead cells, capture them at low frequencies, and characterize them at high frequencies.

Yeast species are one of the cell types that have been investigated by EIS.^{3,36} EIS has been used to check the yeast's status in solutions with different concentrations of live and dead cells at frequencies ranging from 1 MHz to 3 GHz.³⁷ Yang *et al.* identified the live yeast from dead yeast in an in-flow radio frequency (RF) sensor using microfluidics.³⁸ Afterward, an improved in-flow interferometer using a microstrip line sensing electrode for different single yeast types was introduced by Osterberg *et al.*² They analyzed different single yeast-type viability and impedance signals at two frequencies within microstrip electrodes. They found that because the dead cell membrane is permeable to its surrounding media, the permittivity of the dead cell resembles that of the media. In addition to narrowband measurements, investigating the electrical properties of a single yeast cell by broadband EIS provides a comprehensive evaluation of the cell's electrical properties. Regardless of whether narrowband or broadband EIS is used, passing multiple yeast cells under the microstrip electrode simultaneously, or measuring the electrical properties of a single yeast while the cell is in a vertical position, or not in contact with the electrodes, still causes errors during the measurement. Fluid flow can also cause the cell to move and change orientation during measurements, which causes a fluctuation in the impedance measurement.³⁹

Therefore, we developed a label-free and noninvasive broadband EIS biosensor with sufficient bandwidth in combination with a microfluidic channel capable of hydrodynamic focusing to effectively trap and measure a

single *Schizosaccharomyces pombe's* (yeast cell) electrical properties. In our experiments, the cells are initially focused on a streamline at the center of the microfluidic channel where the CPW electrode gaps exist. Afterward, the cell can be captured precisely between the CPW electrode gap using DEP. Then, we measured the ratio of transmitted power (ΔS_{21}) and reflected power (ΔS_{11}) over a broadband frequency band of 30 kHz to 6 GHz to study the electrical properties of a single living or a single heat-treated (dead) yeast cell. The impedance data was then confirmed by fitting scattering (S) parameters to the equivalent circuit suggested for the system.

This is the first report where a single non-adherent yeast cell was successfully focused, trapped, and the EIS properties measured in a single experiment. This is the first system that successfully demonstrates a significant advance in high-throughput single-cell measurement (one cell per minute). This is also the first system to use EIS to continuously measure and target single-cell electrical properties in a zero-flow sucrose solution to differentiate living from dead cells and potentially observe cell growth. We demonstrated a substantial difference between the electrical impedance properties of a living vs. dead single yeast cell without using any fluorescent probes. The novelty of our platform for performing EIS measurements lies in using broadband frequency while the yeast cell is held motionless when in contact with the electrodes. Further evaluation of this platform and method of EIS measurement is also a valuable reference for other electrode configurations in single-cell EIS measurements.

Materials and methods

Milli-Q grade water (Millipore Inc., Bedford, MA) has been used in all of the solutions and cleaning steps.

Yeast culture and sample preparation

Yeast cells (*Schizosaccharomyces pombe*) were used in this experiment. They are cultured using a standard method.⁴⁰ The cells are streaked on a sterilized 9 cm diameter plastic Petri dish containing solidified YES 225 agar media (YES 225 AGAR Catalog #2012-500 Sunrise Science Products, Knoxville, TN). After incubation at 30 °C for about 48 hours, a colony of yeast was washed and re-suspended in a low conductive isotonic solution of 8.5 w/w% sucrose (Sigma Aldrich CAS No.: 57-50-1, St. Louis, MO) and 0.3 w/w% dextrose (Sigma Aldrich CAS No.: 50-99-7, St. Louis, MO) to prevent any osmotic pressure shocks while keeping the cells alive. This solution has been used in electro-manipulation media because it has a low electrical conductivity^{22,41} and can be used to keep cells alive for extended periods.⁴²

To render the cells nonviable, we heat-treated them by placing a portion of the yeast solution in hot water at 80 °C for 5 minutes. After this, the cells were washed with the sucrose/dextrose solution. This step was repeated several times to ensure the removal of the ions that leaked from cells.



To increase cell adhesion on the electrodes, a biocompatible protein, lectin ($50 \mu\text{g ml}^{-1}$) (Sigma-Aldrich L2380-1MG, St. Louis, MO), was injected into the microchannel to coat the substrate and electrodes and the whole device was kept at 4°C overnight.⁴³ Cell viability was checked in a separate experiment in the presence of 0.23 v/v% methylene blue dye (Ward's Science, 470301-808, West Henrietta, NY) inside the solution (Fig. 1a and b). The dye penetrated the heated treated cells staining the dead cells blue. To detect the cells in a label-free manner while they were subjected to the electric field, we did not use methylene blue in our measurements. Eventually, live and heat-treated cells were mixed equally and injected through the microchannel for final measurements.

CPW design and fabrication

A 50Ω CPW was designed as previously reported.⁴⁴ To fine-tune and verify our calculations, we simulated the CPW with three sets of transmission lines using ANSYS high-frequency structure simulator (HFSS) software (Fig. S2†). The CPW has a $16 \mu\text{m}$ gap between the signal and ground electrodes. The signal electrode is $200 \mu\text{m}$ wide, and the electrodes taper to a $3 \mu\text{m}$ gap to focus the electric field inside the cell (Fig. 2a). Accordingly, the ratio between the cell and detection volume, and the consequent impedance change increased. In the fabrication process, we chose fused silica as a substrate due to its biocompatibility with biological materials, its transparency for observing cell manipulation, and its low loss tangent in the wide frequency range. Gold was chosen to be the metal layer of the sensor also because of its biocompatibility. The fabrication process involved a $1 \mu\text{m}$ photoresist AZ nLOF 2020 was spin-coated on the $500 \mu\text{m}$ thick 4 inch fused silica substrate. Then, metal layers (20 nm Ti/500 nm Au) were deposited on the substrate by an electron beam evaporator, followed by the lift-off metallization process to remove the undesired metal layers and the rest of the photoresist structure (Fig. 2b). Afterward, the PDMS microfluidic channel was bonded to the CPW using plasma bonding to ensure a leak-proof bond (Fig. 2c). The microfluidic channel's

length, width, and height were 5 mm, 0.27 mm , and $20 \mu\text{m}$, respectively. Finally, the CPW input and output were terminated by SMA coaxial connectors.

Experimental setup

For visualization purposes, the device-under-test (DUT) was placed on top of a Nikon Eclipse Ti-E inverted fluorescence microscope (Fig. S3†). An Andor camera was then connected to the microscope for optical and electrical measurement. The DUT is connected to a Hewlett Packard 8753D vector network analyzer (VNA) through a pair of SMA connectors. All electrical signals are generated and measured in the S parameters by the VNA, from 30 kHz to 6 GHz. The VNA is calibrated using the short, open, load, and thru (SOLT) standards of a Maury-Microwave-2.92 mm calibration kit prior to measurement to avoid any errors by the connections and cables between the VNA and DUT. We set the intermediate frequency (IF) bandwidth to 30 Hz and the number of points to 801 in VNA.

Additionally, we used oscillatory microfluidics to initially focus the cells to the center of the microfluidic channel width⁴⁵ (Fig. 2d). Each inlet and outlet of the DUT microfluidic channel was coupled with funnel-shaped reservoirs connected to 3-way solenoid valves (Spex VapLock, Vernon Hills, IL) to switch the direction of the fluid inside the channel. Cells were loaded in the inlet reservoir and brought into the channel until the outlet reservoir filled and was at the same fluid level as the inlet reservoir. Once the cells are inside the channel and are detected *via* microscope, a lower oscillation pressure and frequency is applied to solenoid valves switching the fluid flow direction between the two ports. Flow oscillations push and pull the cells back and forth until a single cell approaches the CPW gap and is captured by the DEP between the CPW signal sensor gap. Fluid flow and oscillation frequency are controlled by a pressure controller (Fluigent, LineUp Push-Pull Flow EZ Le Kremlin-Bicêtre, France) module and function generator (Siglent SDG1032X Arbitrary Waveform, Solon, Ohio), respectively. With this method, an operator can precisely control the speed and location of the cells inside the microfluidic channel by controlling the




Fig. 1 Comparison between (a) live and (b) dead yeast in the presence of methylene blue 0.23 v/v% inside the sucrose/dextrose solution. The scale bar is the same for both images.



frequency, duty cycle (on (t_{on}) and off (t_{off}) periods), and applied pressure to the solenoid valves. The main advantage of this trapping system in the context of EIS measurements is that by focusing the cells on a center line within the channel, it provides an easy and reproducible method to trap a single cell between the CPW sensor gap and immediately bring the fluid flow to zero by switching the

solenoid valves to the closed state. Therefore, a single cell can be captured, characterized, and released with a controllable trap-and-release system, and then repeated to measure the next target cell (Movie S1†). In our experiments, we applied a signal with a frequency of 5 Hz, a duty cycle of 0.5, and a pressure of 0.2 mbar to move the cells between each EIS measurement.




Fig. 2 (a) Schematic top view of CPW (b) microfluidic device integrated with CPW fabrication process (c) description of sealing the microfluidic channel on the substrate. (d) Schematic of the oscillatory system.



Equivalent circuit

The equivalent circuit of the coplanar waveguide (CPW) without trapped yeast consists of five components: a lumped section under the microfluidic channel; two separate distributed areas under the air and PDMS (Fig. 3a); a series capacitor (C_s); and two shunt capacitors (C_G). Characteristic impedance (Z_0 , Z_{PDMS}) and electrical lengths (θ_0 , θ_{PDMS}) represent the distributed sections for air and PDMS. The loss of the entire CPW is accounted for by frequency-dependent resistances R_0 . The S parameters measured by the VNA are characteristic of the microfluidic channel using air and sucrose (fluid of known permittivity). The shunt capacitance is defined by air or sucrose, depending on whether the channel is filled or unfilled. We defined a single-shell spherical model for yeast. This model considers a parallel circuit of capacitance and resistance for each cell layer. As such, the yeast equivalent has four circuit elements, including yeast membrane resistance (R_{YM}), yeast membrane capacitance (C_{YM}), cytoplasm resistance (R_{CP}), and cytoplasm capacitance (C_{CP}) (Fig. 3b). To measure yeast characteristics by S parameters from the VNA, the cell circuit is expanded by

considering the CPW with a trapped yeast (Fig. 3c). When a single yeast is trapped between the CPW electrode gap, the yeast equivalent circuit is placed in parallel with C_s because the measurement provides information for both the cell and the solution between the CPW. The characteristic parameter values of the whole system, such as C_G , remain constant and are taken from the equivalent circuit data, while the microfluidic channel is filled with sucrose solution (Fig. 3a).

When a single cell is trapped between the two sharp electrode points, the majority of the RF passes through the cell's cytoplasm. Accordingly, no extra sub-circuits are needed for the field to bypass through the layers of the yeast cell or to cross the electrode polarization layer. The equivalent circuits were implemented in advanced design system (ADS) software. First, the parameters were estimated in the circuit by initial guesses. Then, the S parameters were fitted to the experimental results. By adjusting the values, we optimized each parameter. In brief, the de-embedding process was carried out step by step for the device, then the isotonic solution, and finally, for trapped cells. To minimize the least squares error, we used gradient optimization.




Fig. 3 (a) Equivalent circuit of the coplanar waveguide (CPW) comprising distributed sections under PDMS and air (b) equivalent circuit of the yeast with 4 equivalent circuit elements comprising yeast membrane resistance (R_{YM}), yeast membrane capacitance (C_{YM}), cytoplasm resistance (R_{CP}), and cytoplasm capacitance (C_{CP}). (c) Equivalent circuit of the coplanar waveguide with a yeast-trapped between the CPW signal electrode gaps.



The reflection and transmission coefficient for a cell trapped in a series configuration can be defined as:²²

$$S_{11} = \frac{Z - Z_0}{Z + Z_0} \quad (1)$$

$$S_{21} = \frac{2Z_0}{Z + 2Z_0} \quad (2)$$

where Z_0 is the impedance characteristics of the CPW, and Z is the cell impedance, which is defined as $Z = R + \frac{1}{j\omega C}$ where R is resistance, C is capacitance, and ω is angular frequency; therefore:

$$S_{11} = \frac{R}{R + 2Z_0 + j2\omega CRZ_0} \quad (3)$$

$$S_{21} = \frac{2Z_0(1 + j\omega CR)}{R + 2Z_0 + j2\omega CRZ_0} \quad (4)$$

Data acquisition and measurement protocol

All of our measurements were done at room temperature. For the initial verification of the system, we measured the impedance of air and an isotonic solution of sucrose/dextrose with known permittivity. The impedance measurement across these fluids was recorded by the VNA using S parameters between the frequency range of 30 kHz to 6 GHz with the power of -18 dBm (Fig. 4). For verification, we used an optimization module in ADS software to simulate the reflection coefficient (S_{11}) and the transmission coefficient (S_{21}) and to predict the equivalent circuit elements.

There was a good agreement between the experiment results (Fig. 4 – solid lines) and the simulation (Fig. 4 – dotted lines) except at low (<7 MHz) and high frequencies (>3 GHz). The reason for the discrepancy at very low frequency is due to the double layer effect between the cell and the electrodes at low

Table 1 Equivalent circuit parameters of the CPW without a cell

Section	Parameter	Sucrose	Air
Microfluidic channel	C_s (pF)	3.5	0.6
	C_G (fF)	48.28	4.18
CPW under air	Characteristic impedance Z (Ω)	53	
	Length@1 GHz θ air ($^\circ$)	1.19	
	Loss R_0 (Ω)	$10^{-9}f$ (GHz) + 5	
CPW under PDMS	Characteristic impedance Z (Ω)	46	
	Length@1 GHz θ PDMS ($^\circ$)	0.3	

frequency,⁴⁶ and the equivalent circuit cannot justify the water dispersion at high frequency. In sum, the signals were not efficiently captured by R_0 . Therefore, there is no agreement between simulation and experimental results at low and high frequencies. However, as long as it does not affect the change between the S parameter measurement, we kept R_0 in the equivalent circuit for simplicity.²² Besides, the difference between S_{21} measurements is more prominent than S_{11} , which shows the sensitivity of S_{21} in this configuration with respect to S_{11} . The values of the equivalent circuit parameters shown in this figure are listed in Table 1.

To ensure that our results are valid, we measured the background signal (measurement in the absence of yeast) every five minutes consecutively until the difference between the reflection and transmission line S parameters stabilized less than 0.001 dB. Cells inside the channel started to move toward the CPW within ~ 10 μm steps by the oscillatory system, introduced inside the Experimental setup section. When the target single cell was close to the tapered gap between the CPW electrodes, we immediately brought the fluid flow to zero by switching the solenoid valves to the closed state, and the single cell was trapped between the electrodes by DEP along its long axis (Movie S2†). We captured the live cell by DEP at 4 MHz and 0 dBm and




Fig. 4 Comparison between measured reflection and transmission coefficients when the microfluidic channel was filled with air or a sucrose/dextrose solution consisting of 8.5 w/w% sucrose and 0.3 w/w% dextrose with simulation. Note: the reflection coefficient for air and sucrose overlapped for both the experiment and simulation.




Fig. 5 Micrograph of trapped yeast by dielectrophoresis (DEP) between the coplanar wave guide (CPW) electrode.



visually confirmed it using a microscope (Fig. 5). In the presence of DEP, the yeast cells moved toward the electrodes and were immobilized by the CPW. For the heat-killed yeast samples, due to the attraction of the dead yeast to the lower frequency, we captured single dead yeast by DEP at 30 kHz and 0 dBm.

After trapping the yeast, the VNA was quickly switched from the hold-frequency (4 MHz for live cell, 30 kHz for dead cell) to the sweep-frequency mode (30 kHz to 6 GHz) to measure the yeast's biophysical characteristics without the interference of the DEP. The commands were applied to the VNA using the Python code with the data analyzer. Two port scattering measurements were performed to obtain the single-cell electrical properties. The power was adjusted to -18 dBm, which is below the required power for reversible electroporation⁴⁷ and to avoid any unwanted heating effect from microwave frequency.⁴⁸ The cell remains immobilized between the CPW sensing gap, due to the coated lectin on the surface. The changes in S parameters were calculated and used to derive yeast biophysical characteristics using rapidly successive S parameters recorded with and without yeast. The yeast signal is defined as the difference between the reflection S_{11} and transmission coefficients S_{21} measured between the CPW with and without yeast trapped. With this type of measurement, the net effect of trapped yeast between the CPW will be acquired even with the calibration of the VNA shift with time. Transmission and reflection changes ($\Delta|S_{21}|$, and $\Delta|S_{11}|$) are calculated as follows:

$$\Delta|S_{11}| = 10 \log|S_{11}^{w/cell}| - 10 \log|S_{11}^{w/o cell}| \quad (5)$$

$$\Delta|S_{21}| = 10 \log|S_{21}^{w/cell}| - 10 \log|S_{21}^{w/o cell}| \quad (6)$$

By calculating these parameters from S -parameter measurements, we measured the cell membrane's and cytoplasm's electrical properties. To illustrate the repeatability of the measurement and analysis, this process was repeated for different numbers of individual live or dead single yeast cells. The measurements were repeated three

times for each cell (live and dead). It is worth noting that we measured the background signal immediately after each single cell measurement due to the background signal shift.

Results and discussion

The impedance of ten live and twenty dead single yeast, individually, was measured while the single cell was trapped between the CPW signal electrodes. The averaged reflection S_{11} and transmission ΔS_{21} magnitude changes were measured for both live and dead yeasts (Fig. 6). To extract the electrical impedance values for the single yeast cell, the S parameters spectrum was fitted with the averaged experimental data based on the equivalent circuit changes with and without yeast trapped (Fig. 3). There was good agreement until ~4 GHz, which showed the membrane and cytoplasm's electrical characteristics. Above 4 GHz, the suggested equivalent circuit is no longer valid as additional subcircuits inside the cell-equivalent circuit are required to define the nucleus's electrical properties. However, as the applied frequency range was insufficient for this purpose, we avoided adding additional subcircuits into the cell-equivalent circuit.

The cell-equivalent-circuit values were optimized based upon the sensitivity of each parameter with its most prominent position on the simulation *versus* the frequency spectra as indicated by the arrows (Fig. 7). The direction of the arrows was intended to qualitatively show the incremental impact of each cell-equivalent circuit parameter at the frequency where the impact is most significant, that is, the positions of the arrows denote the frequency bands where each parameter is most sensitive. The values were optimized based on the trends, emphasizing different impedance parameters at various frequencies. Spectra for both the reflection changes and the transmission changes will undergo simultaneous adjustments during parameter tuning. Thus, it is essential for both spectra to reach an equilibrium to avoid overfitting in one spectrum and an off-target measurement in the other, which helps us to reach accurate data. At low frequency (kHz), signals can barely pass through




Fig. 6 (a) Reflection and (b) transmission magnitude changes for live and dead single yeast trapped between coplanar waveguide (CPW).






Fig. 7 Equivalent circuit parameters sensitivity with respect to the frequency in S-parameter spectrum for (a) reflection and (b) transmission changes up to 4 GHz. The circuit elements comprise yeast membrane resistance (R_{YM}), yeast membrane capacitance (C_{YM}), cytoplasm resistance (R_{CP}), and cytoplasm capacitance (C_{CP}). The direction of the arrows indicates each parameter's incremental effect at the frequency where the effect is most significant.

the cell membrane due to the high value of membrane resistance and, therefore R_{YM} is dominant. By increasing the frequency, due to the short-circuited yeast membrane capacitance, the membrane becomes transparent to the signal, and the cytoplasm's electrical properties can be calculated. The signal passes through the membrane (MHz) and R_{CP} is dominated in that frequency range. The dominance of, C_{YM} , and C_{CP} appeared in the MHz and GHz frequency range. R_{YM} 's dominance position region displays a more scattered region than the other equivalent parameter values (Fig. 6). This is due to the effect of solution conductivity at low frequencies. Although we used a low conductivity solution to decrease this effect at low frequency in our measurements, this effect is still noticeable. We calculated the membrane and cytoplasm electrical properties based on the ADS software's gradient optimization technique; their parameter values are listed in Table 2.

We gently killed the yeast cells by heating them which compromised the cell membrane integrity. No morphological changes between live and dead cells were discernible visually because the dead cell membrane was not completely broken down. Therefore, the inherent difference between live and dead cells was identified by EIS. Heated cells release most of their ion content, such as, K^+ , Mg^{2+} and Ca^{2+} into the solution through the perforations made in the membrane during the gentle heating process, as such, the permeability of the yeast membrane increases.⁴⁹ Also, washing the cell with sucrose/dextrose solution will cause additional ions to be washed away, and therefore only a limited number of ions remain inside the cell membrane. Because the membrane is more permeable, however, the cell is porous and does fill with the solution surrounding it. As a result, the conductivity

of the dead yeast membrane is going to be close to the sucrose/dextrose solution.⁵⁰ Thus, the resistance of the dead cell membrane to the live cell membrane increases. Here, the amount of R_{YM} increases from 0.17 ± 0.08 M Ω to 0.29 ± 0.08 M Ω . Cell membrane capacitance, which shows the dielectric property of the cell and cell integrity, remains constant while the cell is alive. However, it decreases due to cell death and drops to zero.⁵¹ The amount of C_{YM} drops from 0.3 ± 0.1 pF for live yeast to 0.023 ± 0.03 pF for dead yeast. The higher value for live yeast C_{YM} with respect to the dead cell reflects the membrane integrity of the live cell.

Changes in a dead cell's cytoplasm's electrical properties are a function of the media around the cell. Because the permeability of the dead cell membrane has increased, the media around the cell reaches the cytoplasm. Therefore, the conductivity of the cytoplasm is assumed to be close to the suspension media. This is consistent with literature reports where the resistance and capacitance of a dead cell cytoplasm in a buffer solution have been shown to decrease due to a greater amount of ion concentrations inside the solution with respect to the cell cytoplasm.⁵² While the electrical properties of the cytoplasm within a dead cell in a sucrose solution have been shown to increase due to the entrance of the non-conductive solution into the cell.⁵³ In a separate experiment, we showed that the methylene blue dye penetrates the dead yeast cells indicating that the cell membrane was permeable after heat treatment of the cells (Fig. 1b). As a result, sucrose/dextrose solution is expected to substitute by ionic content inside the dead yeast cytoplasm, which allows the conductivity as well as the permittivity of the cytoplasm to approach the sucrose/dextrose solution value. Therefore, the conductivity of the dead yeast cytoplasm decreases.

As the cytoplasm permittivity is lower than the sucrose/dextrose solution,⁵⁴ the permittivity inside the dead yeast cytoplasm increases as sucrose/dextrose solution concentration increases inside the dead cell. Thus, the dead cell cytoplasm resistance and capacitance due to the presence of sucrose/dextrose solution inside the cell increases, which is consistent with our observations. Here, the resistance of a dead yeast cytoplasm increases from 0.1

Table 2 Equivalent-circuit parameters for live/dead yeast

Parameter	Live yeast value	Dead yeast value
Membrane resistance R_{YM} (M Ω)	0.17 ± 0.08	0.29 ± 0.08
Membrane capacitance C_{YM} (pF)	0.3 ± 0.1	0.023 ± 0.030
Cytoplasm resistance R_{CP} (M Ω)	0.10 ± 0.04	0.24 ± 0.10
Cytoplasm capacitance C_{CP} (fF)	3.6 ± 0.1	10.0 ± 0.1
C_s Live/Dead (pF)	3.483	3.5



$\pm 0.04 \text{ M}\Omega$ to $0.24 \pm 0.1 \text{ M}\Omega$. Likewise, for the dead cell, as $C_{\text{CP/Live yeast}} < C_{\text{CP/Dead yeast}} < C_{\text{sucrose/dextrose}}$ due to the presence of sucrose/dextrose solution inside the dead cell cytoplasm, the capacitance of the dead yeast cytoplasm increases from $3.6 \pm 0.1 \text{ fF}$ to $10 \pm 0.1 \text{ fF}$. The experimental data for live cells are in general agreement with the estimated theoretical values available in ESI† section. A comparison of our data and the reported literature can be found in Table S1.†

When live cells are trapped between the electrodes, due to the loading effect, the amount of $C_{\text{s Live}}$ has to be lower than $C_{\text{s}} = 3.5 \text{ pf}$. In this research, based on the optimization, we consider $C_{\text{s Live}} = 3.483 \text{ pF}$, while $C_{\text{s Dead}} = C_{\text{s}} = 3.5 \text{ pf}$ due to the presence of sucrose/dextrose solution inside the cell and the loading effect can be neglected. The major difference between the dead and live cell signals appears in MHz and GHz, where the cytoplasm dominates in behavior with respect to the applied frequency (Fig. 7). This shows that, at higher frequency (GHz), microwave frequency can penetrate through the cell, as the membrane cannot act as an insulator anymore, and the difference between the intracellular properties will be more detectable. In this research, the noticeable difference appears at $\sim 3 \text{ GHz}$, where the cytoplasm capacitance dominates. Reflection and transmission phase changes for live and dead single yeast have been shown in Fig. S4.†

Based on the S parameters measurements, we showed that the change in cytoplasm structure can be identified within RF. The proposed system can speed up a continuous method for long-term measuring the target cell EIS measurement and checking the consecutive cell electrical measurement and viability. It provides a potential condition for investigating yeast growth between the CPW electrodes with the addition of an incubator and maintainable environmental growth conditions. Future work will focus on the EIS measurement of single-cell growth between the CPW to measure the structural changes during this period.

Conclusion

An innovative electrical-impedance sheathless microfluidic platform combined with CPW has been introduced to guide the cells with hydrodynamic focusing in a microfluidic channel and DEP was used to trap the single cell between the CPW electrodes. The electrical properties of the membrane and cytoplasm of single fission yeast *S. pombe* were investigated in a wide frequency range of 30 kHz to 6 GHz non-invasively. The electrical properties of the cell were measured through the suggested equivalent circuit simulation. We measured the cytoplasm's small capacitance to be approximately 3.6 fF. Then, the cell viability was successfully distinguished through impedance measurement. For dead cells, the resistance increases and the capacitance decreases due to the diffusion of the sucrose/dextrose solution through the perforated cell membrane. Also, the resistance and capacitance of the

cytoplasm of the dead cell increases because of an increase in the permeability of the membrane and the influx of the surrounding media (sucrose/dextrose) to the cell cytoplasm. Noticeable differences between the viable and nonviable single yeast appeared at 3 GHz, where cytoplasm capacitance dominates. This system offers a rapid, precise, noninvasive, and label-free method for the characterization of a single cell, which can open the doors to use for any type of cell and different electrode configurations regardless of the shapes and geometries for cell viability discrimination, real-time single-cell monitoring, and rapid, precise single-cell analysis.

Data availability

Data for this publication is available at <https://doi.org/10.5281/zenodo.12863742>.

Author contributions

A. F. was responsible for conceptualization, methodology, formal analysis, data curation, investigation, device design, device fabrication, software, writing – original draft, and review & editing; A. M. was responsible for conceptualization, methodology, investigation, device design, device fabrication, and writing – review & editing; M. J. A. was responsible for conceptualization device design, methodology, formal analysis, software, and writing – review & editing; J. C. M. H. was responsible for supervision, writing – review & editing; A. A. was responsible for project administration, funding acquisition, writing – review & editing, and supervision.

Conflicts of interest

The authors have no conflicts of interest to declare.

Acknowledgements

We would like to express our gratitude toward Dr. Qian Chen from the Department of Biological Sciences, University of Toledo, for sending the yeast strain, and Dr. Kelley Donaghy for editing this manuscript. This work was performed in part at the Cornell NanoScale Science & Technology Facility (CNF), a member of the National Nanotechnology Coordinated Infrastructure NNCI, which is supported by the National Science Foundation (Grant NNCI-2025233). This work made use of the facility of the Cornell Center for Materials Research (CCMR) with support from the National Science Foundation Materials Research Science and Engineering Centers (MRSEC) program (DMR-1719875). Additional funding was provided by the United States Army Combat Capabilities Development Command (proposal number: 78597-EL).

References

- 1 A. Schmid, H. Kortmann, P. S. Dittrich and L. M. Blank, *Curr. Opin. Biotechnol.*, 2010, **21**, 12–20.



- 2 J. A. Osterberg, N. Dahal, R. Divan, C. S. Miller, D. Moline, T. P. Caldwell, X. Yu, S. W. Harcum and P. Wang, *IEEE Trans. Microwave Theory Tech.*, 2021, **69**, 1875–1886.
- 3 N. Haandbæk, S. C. Bürgel, F. Heer and A. Hierlemann, *Lab Chip*, 2014, **14**, 369–377.
- 4 Y. Feng, H. Chai, W. He, F. Liang, Z. Cheng and W. Wang, *Small Methods*, 2022, **6**, 2200325.
- 5 X. Zhang, A. Hatamie and A. G. Ewing, *Curr. Opin. Electrochem.*, 2020, **22**, 94–101.
- 6 L. Huang, P. Zhao and W. Wang, *Lab Chip*, 2018, **28**, 2359–2368.
- 7 E. P. Preece, F. J. Hardy, B. C. Moore and M. Bryan, *Harmful Algae*, 2017, **61**, 31–45.
- 8 M. Barreiros Dos Santos, R. B. Queirós, Á. Galdes, C. Marques, V. Vilas-Boas, L. Dieguez, E. Paz, R. Ferreira, J. Morais, V. Vasconcelos, J. Piteira, P. P. Freitas and B. Espiña, *Biosens. Bioelectron.*, 2019, **142**, 111550.
- 9 H. E. Kubitschek, *Nature*, 1958, **182**, 234–235.
- 10 D. C. Grulke, N. A. Marsh and B. A. Hills, *Br. J. Exp. Pathol.*, 1973, **54**, 884–891.
- 11 M. Mansor and M. Ahmad, *Int. J. Mol. Sci.*, 2015, **16**, 12686–12712.
- 12 Y. Feng, J. Zhu, H. Chai, W. He, L. Huang and W. Wang, *Small*, 2023, **19**, 2303416.
- 13 Y. Feng, Z. Cheng, H. Chai, W. He, L. Huang and W. Wang, *Lab Chip*, 2022, **22**, 240–249.
- 14 J. C. M. Hwang, *IEEE Microwave Magazine*, 2021, pp. 78–87.
- 15 H. Morgan, T. Sun, D. Holmes, S. Gawad and N. G. Green, *J. Phys. D: Appl. Phys.*, 2007, **40**, 61–70.
- 16 T. Sun and H. Morgan, *Microfluid. Nanofluid.*, 2010, **8**, 423–443.
- 17 Y. Cui, W. F. Delaney, T. Darroudi and P. Wang, *Sci. Rep.*, 2018, **8**, 497.
- 18 R. Bashir, *Adv. Drug Delivery Rev.*, 2004, **56**, 1565–1586.
- 19 G. M. Whitesides, *Nature*, 2006, **442**, 368–373.
- 20 S. Patel, D. Showers, P. Vedantam, T.-R. Tzeng, S. Qian and X. Xuan, *Biomicrofluidics*, 2012, **6**, 034102.
- 21 H. Li, C. Multari, C. Palego, X. Ma, X. Du, Y. Ning, J. Buceta, J. C. M. Hwang and X. Cheng, *Sens. Actuators, B*, 2018, **255**, 1614–1622.
- 22 X. Ma, X. Du, H. Li, X. Cheng and J. C. M. Hwang, *IEEE Trans. Microwave Theory Tech.*, 2018, **66**, 3690–3696.
- 23 C. Yi, Q. Zhang, C.-W. Li, J. Yang, J. Zhao and M. Yang, *Anal. Bioanal. Chem.*, 2006, **384**, 1259–1268.
- 24 H. Chai, Y. Feng, J. Zhu, X. Meng, F. Liang, J. Bai and W. Wang, *ACS Sens.*, 2023, **8**, 2681–2690.
- 25 X. Ma, X. Du, L. Li, H. Li, X. Cheng and J. C. M. Hwang, *IEEE J. Electromagn. RF Microw. Med. Biol.*, 2020, **4**, 37–44.
- 26 Y. Feng, L. Huang, P. Zhao, L. Fei and W. Wang, *Anal. Chem.*, 2019, **91**, 15204–15212.
- 27 A. Denzi, C. Merla, M. Casciola, J. C. M. Hwang, X. Cheng, F. Apollonio and M. Liberti, in *2016 38th Annual International Conference of the IEEE Engineering in Medicine and Biology Society (EMBC)*, IEEE, Orlando, FL, USA, 2016, pp. 4232–4235.
- 28 C. Honrado, P. Bisegna, N. S. Swami and F. Caselli, *Lab Chip*, 2021, **21**, 22–54.
- 29 D. Malleo, J. T. Nevill, L. P. Lee and H. Morgan, *Microfluid. Nanofluid.*, 2010, **9**, 191–198.
- 30 B. Xu, Y. Shi, Z. Lao, J. Ni, G. Li, Y. Hu, J. Li, J. Chu, D. Wu and K. Sugioka, *Lab Chip*, 2018, **18**, 442–450.
- 31 J. Chen, Y. Zheng, Q. Tan, E. Shojaei-Baghini, Y. L. Zhang, J. Li, P. Prasad, L. You, X. Y. Wu and Y. Sun, *Lab Chip*, 2011, **11**, 3174.
- 32 H. Li, A. Denzi, X. Ma, X. Du, Y. Ning, X. Cheng, F. Apollonio, M. Liberti and J. C. M. Hwang, *IEEE Trans. Microwave Theory Tech.*, 2017, **65**, 3503–3511.
- 33 C. Ferguson, N. Pini, X. Du, M. Farina, J. M. C. Hwang, T. Pietrangelo and X. Cheng, *Anal. Chim. Acta*, 2021, **1173**, 338678.
- 34 M. Farasat, E. Aalaei, S. K. Ronizi, A. Bakhshi, S. Mirhosseini, J. Zhang, N.-T. Nguyen and N. Kashaninejad, *Biosensors*, 2022, **12**, 510.
- 35 K. Zhao, Larasati, B. P. Duncker and D. Li, *Anal. Chem.*, 2019, **91**, 6304–6314.
- 36 M. Zhang, G. Huo, J. Bao, T. Markovic, P. Van Dijk and B. Nauwelaers, *Chemosensors*, 2022, **10**, 318.
- 37 A. Zhang, D. Kawashima, H. Obara and M. Takei, *IEEE Trans. Microwave Theory Tech.*, 2021, **69**, 4678–4688.
- 38 Y. Yang, H. Zhang, J. Zhu, G. Wang, T.-R. Tzeng, X. Xuan, K. Huang and P. Wang, *Lab Chip*, 2010, **10**, 553.
- 39 D. Spencer and H. Morgan, *Lab Chip*, 2011, **11**, 1234.
- 40 F. Sherman, in *Methods in Enzymology*, Elsevier, 2002, vol. 350, pp. 3–41.
- 41 W. M. Arnold, A. G. Gessner and U. Zimmermann, *Biochim. Biophys. Acta, Gen. Subj.*, 1993, **1157**, 32–44.
- 42 Y. Ning, C. Multari, X. Luo, C. Palego, X. Cheng, J. C. M. Hwang, A. Denzi, C. Merla, F. Apollonio and M. Liberti, *IEEE Trans. Microwave Theory Tech.*, 2014, **62**, 1905–1911.
- 43 D. Sinha, D. Ivan, E. Gibbs, M. Chetluru, J. Goss and Q. Chen, *J. Cell Sci.*, 2022, **135**, jcs259046.
- 44 G. Gonzalez, *Microwave Transistor Amplifiers: Analysis and Design*, 1984.
- 45 B. R. Mutlu, J. F. Edd and M. Toner, *Proc. Natl. Acad. Sci. U. S. A.*, 2018, **115**, 7682–7687.
- 46 H. P. Schwan, *Ann. N. Y. Acad. Sci.*, 1968, **148**, 191–209.
- 47 Y. Ning, X. Ma, C. R. Multari, X. Luo, V. Gholizadeh, C. Palego, X. Cheng and J. C. M. Hwang, in *2015 IEEE MTT-S International Microwave Symposium*, IEEE, Phoenix, AZ, USA, 2015, pp. 1–3.
- 48 T. Chretiennot, D. Dubuc and K. Grenier, *IEEE Trans. Microwave Theory Tech.*, 2013, **61**, 972–978.
- 49 S. Guyot, P. Gervais, M. Young, P. Winckler, J. Dumont and H. M. Davey, *Environ. Microbiol.*, 2015, **17**, 2982–2992.
- 50 M. Kriegmaier, M. Zimmermann, K. Wolf, U. Zimmermann and V. L. Sukhorukov, *Biochim. Biophys. Acta, Gen. Subj.*, 2001, **1568**, 135–146.
- 51 C. Palego, C. Merla, Y. Ning, C. R. Multari, X. Cheng, D. G. Molinero, G. Ding, X. Luo and J. C. M. Hwang, in *2013 IEEE MTT-S International Microwave Symposium Digest (MTT)*, IEEE, Seattle, WA, USA, 2013, pp. 1–3.
- 52 H.-L. Gou, X.-B. Zhang, N. Bao, J.-J. Xu, X.-H. Xia and H.-Y. Chen, *J. Chromatogr. A*, 2011, **1218**, 5725–5729.



- 53 Y. Ning, C. Multari, X. T. Luo, F. P. Cristiano, D. F. Molinero, S. Cheng, X. Hwang and C. M. James, *Proc. 43rd Eur. Microw. Conf.*, 2013, p. 478.
- 54 A. Denzi, C. Merla, C. Palego, A. Paffi, Y. Ning, C. R. Multari, X. Cheng, F. Apollonio, J. C. M. Hwang and M. Liberti, *IEEE Trans. Biomed. Eng.*, 2015, **62**, 1595–1603.

