Lab on a Chip



PAPER

View Article Online
View Journal | View Issue



Cite this: Lab Chip, 2025, 25, 4410

Recovery of phenotypically sorted cells using droplet-digital microfluidics†

Zhiyang Deng,^{ab} James M. Perry,^b Marian Weiss, [©] Robert Genth,^c Alexis Autour,^d Christoph A. Merten [©] * and Steve C. C. Shih [©] * * and Steve C. C. Shih * and Steve C. C.

Droplet microfluidics has become a ubiquitous and powerful tool for high-throughput phenotypic screening at the single-cell level. Large numbers of cells can be sorted for a variety of functions, including the secretion of antibodies with tailored properties. The recovery of cells from sorted droplets is still very poor compared to droplet sorting, usually being limited to around 50% of all sorted hits. Here, we present a fully integrated droplet-digital microfluidic platform for the isolation and the recovery of rare single cells and applied our system to antibody discovery. From our work, we have achieved an 18-fold increase in the recovery rate of individual cells and beads from droplets, as compared to conventional methods. We believe that the combination of high-throughput droplet generation with the on-demand control features of digital microfluidics will improve the number of characterized hits in single-cell -omics, antibody screens, directed evolution of enzymes, and beyond.

Received 28th April 2025, Accepted 7th July 2025

DOI: 10.1039/d5lc00415b

rsc.li/loc

Introduction

Single cell analysis has become a frequently used method for assessing genomes, transcriptomes, or proteomes which leads to important advances in developing cancer therapies, 1-3 antibody discovery,4,5 and deciphering neural pathways and brain disorders.^{6,7} Droplet microfluidics has gained increasing popularity for such studies, 8,9 given its highthroughput and multi-parameter measurement capabilities. Fluorescence-activated droplet sorting (FADS) has been frequently used as a high-throughput screening tool for single-cell studies. 10-12 FADS is known to apply less stress to the sorted cells, enabling higher cell viability for postprocessing.¹³ Additionally, encapsulating secreted analytes in droplets has been crucial for many genotype-phenotype applications like enzyme-based strain development and antibody discovery. 12,14 However, droplet-based systems can face a critical challenge - sample loss during the recovery of the contents from processed droplets. 15-19 Conventionally, recovering contents from droplets is a complex multi-step process. Clausell-Tormos et al. pooled cell-containing droplets

There are currently two methods for sample recovery, recovery from bulk (pooled) droplets^{25–27} or from individual

in a centrifuge tube, broke the emulsion, and recovered the contents from those droplets. In this standard droplet recovery protocol, a recovery rate of 20-30% was reported after 24-hour incubation of the droplets. 15 Sun et al. reported an efficiency of 50% while sorting and collecting positive hits. Furthermore, only 60% of the collected droplets can be recovered for downstream analysis, contributing to an overall workflow efficiency of only 10-20%.17 Gaa et al. used a commercial system to recover contents from sorted droplets into well plates and were only able to achieve a recovery rate of less than 53%.19 The sample loss issue can be more daunting when working with sensitive and rare targets such as antibody-secreting B cells or circulating tumor cells (CTCs). Having a critical role in immunotherapy, B cells make up 10% to 20% of all lymphocytes. However, only 0.1% to 2% of isolated B cells can produce functional antigen-specific antibodies.20 CTCs are informative for cancer studies; however, they are extremely rare, often less than 100 CTCs can be found per 10^7 leukocytes and 5×10^9 erythrocytes in 1 mL of whole blood.21 Furthermore, since applications like clonal expansion,²² single-cell profiling,²³ and cytokine detection24 are downstream of cell sorting and recovery, the sample loss during collection can lead to loss of crucial data or significant time of experimental work. Hence, new methods for single-cell isolation and recovery of viable cells from processed droplets are in demand for field of single-cell analysis advancing the microfluidics.

^a Department of Electrical and Computer Engineering, Concordia University, Montréal, Canada. E-mail: steve.shih@concordia.ca

^b Centre for Applied Synthetic Biology, Concordia University, Montréal, Canada

^c VERAXA Biotech GmbH, Heidelberg, Germany

d Institute of Bioengineering, School of Engineering, École Polytechnique Fédérale de Lausanne (EPFL), Lausanne, Switzerland. E-mail: christoph.merten@epfl.ch

 $[\]dagger$ Electronic supplementary information (ESI) available. See DOI: https://doi.org/10.1039/d5lc00415b

Lab on a Chip

(single) droplets. 28-30 For pooled recovery, droplets are usually collected in a tube, followed by a break of emulsions to extract the contents.²⁷ This method is useful for recovering a large, processed population; however, there is a loss of resolution such that deconvoluting hits from different droplets is challenging and there is a risk of crosscontamination between droplets. Single-cell screens, such as antibody discovery campaigns, 12 could greatly benefit from droplet recovery, to directly match a droplet phenotype (e.g. strongest binding or functional effect) with a particular gene sequence of the corresponding hit. For example, Cole et al. have designed a continuous and programmable system integrating a droplet sorter with a cell printer for precise dispensing of single droplets, termed printed (PDM).²⁸ microfluidics This system dielectrophoretic trapping to position individual droplets into nanowell arrays, facilitating high-throughput droplet collection. While the structured printing substrate facilitates high-throughput collection of cells, the fixed nanowell configuration creates limitations on subsequent cell culture and analysis. Another approach by Nan et al., developed an on-demand droplet collection (ODC),²⁹ that enables continuous and quantitative collection of target droplets into microtubes mounted on three different rotary platforms. However, this platform is technically complex and still does not allow easy transfer of individual hits into any reservoir of choice, including e.g. industrially used microtiter plates. Josephides et al. reported the use of the Cyto-Mine droplet microfluidics, single-cell sorting and system.³¹ This platform achieved dispensing performance, with 98.93% accuracy in dispensing sorted droplets into microtiter plates and high cell viability after recovery. However, despite its effectiveness, Cyto-Mine is a closed, proprietary system with limited flexibility, making it difficult to modify or customize for specific applications or alternative workflows. Here, we introduce a newly constructed droplet-digital (D²) microfluidic platform for the isolation and recovery of rare single cells from processed droplets. In this method, single-plate digital microfluidics (DMF) is used in combination with a droplet channel layer. By integrating droplet-based microfluidics with DMF, droplets can be generated and manipulated in an ondemand and addressable manner.¹⁴ Using the open system, we are able to address any part of the workflow where it is difficult to recover rare single cells from individual droplets and transfer them into conventional culture plates. The workflow consists of generating droplets and sorting the droplets based on fluorescence intensity, followed by collecting the contents of the sorted droplets with a µLscale volume recovery buffer, in a controllable, precise, and automated manner. To demonstrate the utility of our platform, we have performed recovery experiments with single beads, single HEK 293 cells, and single specific antibody-secreting hybridoma cells at typical and low ratios²⁰ to model rare events, establishing it as a versatile

Methods

Reagents, materials, and equipment

Microfabrication materials. Microfluidic device fabrication materials included a high resolution 25 400 dpi transparent photomask (Artnet Pro Inc., Bandon, OR, USA), AZ-1500 positive photoresist coated chromium glass substrates (Telic, Valencia, CA, USA), Si wafers (Silicon Valley Microelectronics Inc., Santa Clara, CA, USA), MF-321 developer, CR-4 chromium etchant (OM Group, Cleveland, OH, USA), AZ-300 T photoresist stripper (Integrated Micro Materials, Argyle, TX, USA), SU-82075 photoresists and SU-8 developer (Kayaku Advanced Materials, Westborough, MA, USA), polydimethylsiloxane (PDMS) (DOW Silicones Corporation, Midland, MI, USA), chlorotrimethylsilane (Sigma-Aldrich, Oakville, ON, Canada) and polylactic acid (PLA) (Shop3D, Mississauga, ON, CA). General-use solvents and chemicals used for microfluidic chip fabrication, including acetone (cleanroom lab grade) and isopropanol (IPA, cleanroom lab grade), were purchased from Sigma-Aldrich (Oakville, ON, Canada). De-ionized (DI) water has a resistivity of 18 M Ω cm at 25 °C.

PDMS surface treatment and device operation reagents included 3 M Novec HFE 7500 engineering fluid (M.G. Chemicals, Burlington, ON, Canada), 008-fluoro-surfactant dissolved in HFE7500 (20 g of 5% wt) (RAN Biotechnologies, Beverly, MA, USA), Aquapel (Amazon, Canada) and OptiPrep density gradient medium (Sigma-Aldrich, Canada).

Microfabrication equipment. Microfabrication equipment included a Tergeo plasma cleaner (PIE Scientific LLC, Union City, CA, USA), a Quintel Q-4000 mask aligner (Neutronix Quintel, Morgan Hill, CA, USA), a Laurell spin coater (model WS-650Mz-8NPPB, Laurell Technologies Corporation, North Wales, PA, USA), and a UV-KUB 3 mask aligner (KLOÉ, Montpellier, FRA).

Cell culture and antibody screening reagents. The cell culture medium was purchased from Wisent Inc. (Saint-Jean-Baptiste, QC, Canada). Cell culture reagents include Iscove's modified Dulbecco's media (IMDM) for culturing hybridoma cells, IMDM with no phenol red, Dulbecco's modified Eagle's medium (DMEM) for the target cells, fetal bovine serum (FBS), recombinant mouse IL-6 (carrier-free), Geneticin (G418), penicillin/streptomycin and phosphate buffer saline (PBS) (Ca²⁺/Mg²⁺ free). FluoSpheresTM polystyrene microspheres (10 μm, blue-green fluorescent (430/465)), Alexa FluorTM 488 goat antimouse IgG (H+L) and CellTraceTM calcein red-orange (AM) dyes, and SYTOXTM blue dead cell stain were purchased from Thermo Fisher Scientific (Mississauga, ON, Canada). The cell dissociation solution (non-enzymatic, 1×) was purchased from Sigma-Aldrich (Oakville, ON, Canada).

Fluidic system. Gastight 500 μ L and 2.5 mL glass syringes were purchased from Hamilton (Reno, NV, USA). Fittings and tubing: 1/4– 28×10 –32 PEEK adapter, 10–32 PEEK union assembly, finger tight micro ferrule 10–32 coned for 1/32'' OD and PEEK tubing (1/32'' diameter) were purchased from IDEX Corporation (Lake Forest, Illinois, USA). BD 1 mL and 3 mL syringes were purchased from Fisher Scientific (Waltham,

tool for a range of single-cell studies.

Paper

MA, USA). Tygon tubing was purchased from VWR (Radnor, PA, USA). Stainless steel connection needles for the BD plastic syringes were purchased from McMaster-CARR (Elmhurst, Illinois, USA). A low-pressure Nemesys pump system (Cetoni, Korbussen, DE) was used for flow control.

Optical setup, data acquisition and control system. The construction of the electrode control system has been previously published by our lab 10,14 and others. 32 Briefly, the hardware requirements contain a function generator (33210A, Keysight Technologies, Santa Rosa, CA, USA), an amplifier (PZD700A, Advanced Energy Inc, Denver, CO, USA), a DC power supply (GPE-4323, GW Instek America Corp, Montclair, CA, USA), a stacked control board of optocoupler switches, and a microcontroller (Teensy 4.1, Sherwood, OR, USA). The microcontroller was connected to a port expander that delivers the potential to the contact pads of the device through optocoupler switches. Droplet manipulations were achieved by application of AC potentials (140 V_{rms}) at 15 kHz. The optical setup for sorting consisted of an Olympus IX73 inverted microscope (Evident Scientific, Mississauga, ON, mounted on a vibration-dampening bench Canada) (Thorlabs, Newton, NJ, USA), a photon counting head, a filter block, a SMA connector (Hamamatsu Photonics H11890, Township, NJ, USA), an X-Cite XLED1 multi-triggering LED illumination system (Excelitas Technologies Corp., Waltham, MA, USA), bandpass filters, an attenuator filter, fiber optics (Thorlabs, Newton, NJ, USA), and a Hamamatsu ORCA-Flash 4.0 camera (Township, NJ, USA). Fluorescence images of the single cells in well plates were taken using an EVOS M5000 imaging system (Thermo Fisher Scientific, Waltham, MA, USA). Spectrometry data acquisition, PMT control, and droplet actuation were performed through a USB interface using our Python-based graphical user interface, as described in previous work.¹⁴ The software is open-source and available download at https://bitbucket.org/shihmicrolab/ uflowcontrol. ESI† Movie S1 shows the software interface presenting the operation of the software and live sorting of a single cell droplet.

Microfluidic device fabrication

The fabrication of a droplet-digital (D^2) microfluidic device followed a similar protocol to that in Ahmadi *et al.* ¹⁴ The device consisted of three layers: a Cr-electrode layer, a dielectric layer, and a PDMS channel layer that were patterned by photolithography and soft-lithography methods. The first layer followed typical protocols for electrode fabrication. ^{10,14,32,33} Briefly, chromium-coated 50 \times 75 mm glass slides (Telic, Valencia, CA, USA) with an S1811 positive resist were UV exposed (7 s at 38–50 mW cm⁻²), then developed in MF-321 developer, etched with a CR-4 chromium etchant, and stripped with an AZ-300 T photoresist stripper. For the dielectric layer, 1 mL of PDMS (1:10 w/w ratio, curing agent to prepolymer) was poured at the center of the Cr-patterned glass substrate, and the device was spin-coated at 4500 rpm for 90 s. Chromium contact pads were covered with Scotch 3 M tape prior to spin coating

for easy removal of the excessive PDMS. The tape was removed after spin-coating and the device was baked at 90 °C for 1 h, resulting in a dielectric layer of \sim 7 μm .

For the sorter PDMS channel layer, we used our previous fabrication protocol. 10 SU-8 2075 was spin-coated onto a 100 mm silicon wafer (10 s - 500 rpm, 30 s - 2000 rpm, 10 s -500 rpm) to obtain a 100 µm thick layer, followed by a baking and exposure cycle according to the manufacturer's datasheet. A second layer (for the optical fibers) containing SU-8 2075 was spin-coated on top of the non-developed first layer (10 s - 500 rpm, 30 s - 2000 rpm, 10 s - 500 rpm) to obtain a 100 µm thick layer. After pre-exposure baking, the second layer mask was aligned and exposed (UV-KUB 3, KLOÉ, Montpellier, France), followed by baking and development according to the manufacturer's datasheet. The master mold was exposed to chlorotrimethylsilane vapour that was situated in a desiccator for 45 min, to prevent PDMS from permanently sticking to the mold. Next, PDMS (1:10 w/ w ratio curing agent to prepolymer) was degassed and poured over the mold. To prevent PDMS shrinkage, the PDMS was incubated at room temperature to cure for at least 48 h. PDMS parts were cut to the designed size with a razor blade. Inlets for oil and aqueous phases were made using a 1.5 mm biopsy puncher (World Precision Instruments, FL, USA). The withdraw/waste outlet and recovery buffer inlet were made using a 0.75 mm biopsy puncher. The droplet-digital collection site was made with a 3 mm biopsy puncher and a semicircle was excised with a razor blade. 1.5 mm inlets were fitted with 0.06-inch OD Tygon transparent tubing and 0.75 mm inlets were fitted with 1/32" OD PEEK tubing. The PDMS was cleaned with Scotch tape to remove dust before assembly. The PDMS channel layer and the substrate were treated with oxygen plasma for 15 s, bonded together, and baked at 60 °C for 30 min. The PDMS channel containing the sorter channel was manually aligned with the dielectriccoated electrodes under a microscope. Device channels were then treated with Aquapel for 1 min, and then the device was air-dried and baked at 85 °C for 20 min. A flat, cleaved, multi-mode optical fiber (200 µm core, 0.49 N.A.) was prepared for droplet detection (Thorlabs, NJ, US). The fiber was fixed with Scotch tape on the microscope stage.

Experimental sample preparation

HEK293 cell culture and preparation. Two HEK293 cell lines, wild type (WT) and GFP-expressing variant, were donated by Dr. Elena Kuzmin and Dr. Alisa Piekny lab and cultured in DMEM supplemented with 10% heat-inactivated FBS, 1% penicillin–streptomycin (P/S) in a cell incubator at 37 °C and in 5% CO₂. For cell maintenance, cells were passaged every two days at 80% confluency by washing with PBS and trypsinization, and then added to a final volume of 15 mL DMEM medium, in T-75 treated flasks.

For HEK cell experiments, the supernatant was discarded and cells were washed two times with 5 mL PBS. 2 mL of trypsin was added to the cell flask followed by 5 min

Lab on a Chip

Open Access Article. Published on 22 July 2025. Downloaded on 11/24/2025 3:17:50 PM

incubation in a cell incubator in 5% CO2 and at 37 °C. Next, 8 mL of DMEM (+ 10% FBS, 1% P/S) was added to the flask to inactivate trypsin and to wash the remaining cells off the bottom of the flask. 5 mL of the cell suspension was transferred to a 15 mL tube for centrifugation (1000 rpm, 5 min). Lastly, the supernatant in the tube was discarded and the cell pellet was resuspended in 5 mL of PBS and placed

For "spiked ratio" experiments, the GFP-expressing HEK cells and WT HEK cells were both adjusted to a final concentration of 1×10^6 cells per mL. The cell lines were mixed in two ratios (1:400 and 1:1000 - GFP:WT), at a total volume of 1 mL per tube. Next, to mimic low-volume rare cell inputs, tubes with 10000 cells and tubes with 100000 cells were prepared for different ratios and topped up to 400 µL with PBS. Lastly, 80 μL of OptiPrep was added to each 400 μL cell suspension tube, followed by a gentle vortex. For microfluidic sorting, the different ratios of cell suspensions (480 μL) were loaded into 1 mL BD plastic syringes. 600 μL of GFP-expressing and 600 µL of WT cells were loaded into 1 mL syringes as the positive and negative controls.

Cell culture and preparation of hybridoma and target cells. Two hybridoma cell lines were provided by the National Research Council Canada (NRC) and were cultured in IMDM supplemented with 10% heat-inactivated FBS, mouse interleukin-6 (1 ng mL⁻¹), and 1% penicillin-streptomycin in a cell incubator at 37 °C and in 5% CO2. Cells were passaged every three days at a final concentration of 8×10^5 cells per mL in 15 mL of medium in T-25 treated flasks. Cells were used for experiments when the cells were in the midlogarithmic and under passage 12.

Target cells were provided by the NRC. These cells were cultured in DMEM (+ 10% FBS) and G418 (0.4 mg mL⁻¹) in a cell incubator in 5% CO2 and at 37 °C. Cells were passaged every three days at 80% confluency. Passaging and preparation followed the same protocol as above for hybridoma cells, except that a pre-warmed (37 °C) non-enzymatic cell-dissociation solution was used in lieu of trypsin to preserve the expression of the surface receptor antigen.

For sorting experiments, the cells were resuspended in the IMDM (no phenol red) cell culture medium containing 10% FBS and mouse interleukin-6 (1 ng mL⁻¹) without antibiotics at a final concentration of 1×10^6 cells per mL. Specific-bindingantibody hybridoma producers were stained with CellTraceTM calcein-AM red-orange following the manufacturer's protocol. The target cells were resuspended in the IMDM (no phenol red) medium containing 10% FBS, mouse interleukin-6 (1 ng mL⁻¹), and Alexa-488-conjugated goat-anti-mouse antibodies (2.5 µg mL^{-1}) – i.e., no antibiotic – at a final concentration of 3 × 10⁶ cells per mL.

Preparation of fluorescent beads. For experiments using fluorescent beads (blue-green fluorescent, 10 µm diameter), the beads were washed three times with PBS and diluted in PBS to 1 × 10⁶ beads per mL before loading into a 1 mL syringe. 16% Optiprep was added to prevent beads from clumping.

D² microfluidic platform setup

Bead and HEK cell experiments. A D2 device was used to perform bead and cell encapsulation, sorting, collection, and recovery. The chip included two inlets for the flow-focusing droplet generator, a pair of sorting electrodes, four pairs of collection electrodes, a fiber channel connected to a PMT, a withdraw/waste channel, a recovery medium channel, a waste channel, and a collection channel for sorted droplets. A syringe (2.5 mL gastight) was connected to a withdrawal port and another for the recovery buffer injection (500 µL gastight). Both were fitted with 1/32" OD, 0.127 mm ID PEEK tubing. A syringe with HFE oil (2.5 mL gastight) and with beads/cells (1 mL BD plastic) was fitted with 0.06-inch OD and 0.02-inch ID Tygon tubing. Syringes were installed on a low-pressure neMESYS pump system (Cetoni, Korbussen, DE) for automation control.

The optical fiber was inserted into the channel for the fiber on the device and was fixed in a 3D printed holder, and clamped in place with a pogo pin PCB providing contact with the electrode pads. The holder base plate was custom-made to fit on the scanning stage (XYZ Tango, Marzhauser, Wetzlar, DE) of an inverted epi-fluorescence microscope (Olympus IX73, Olympus, Montreal, QC, Canada). The fluorescent excitation light (λ_{ex} = 470 nm) was produced by an X-cite XLED (Ocean Optics, Ottawa, ON, Canada). The SMA end of the detection fiber was coupled to a photon counting head (Hamamatsu Photonics H11890, Township, NJ, USA). The flow inside the microfluidic channel was observed under a 4× or 10× objective under bright-field illumination. The spectrometer (photon counting head), fluid flow, and electrode actuation were controlled using an in-house automation system and graphical user interface.14

Hybridoma and target cell experiments. A co-encapsulation chip (Fig. S1†) and a D² microfluidic device were used to perform an antibody-producing hybridoma cell screening and recovery assay. The co-encapsulation device consisted of a flowfocusing generator with two cell inlets for hybridoma cells and target cells, respectively. For co-encapsulation, a syringe for withdrawal (3 mL BD plastic, for droplet collection and off-chip incubation), a syringe filled with HFE oil (2.5 mL gastight) and two syringes with cells (1 mL BD plastic) were fitted with 0.06inch OD and 0.02-inch ID Tygon tubing. Syringes were installed on a low-pressure neMESYS pump system (Cetoni, Korbussen, DE). For sorting, collection and recovery, the setup was the same as described in the above section.

D² microfluidic operation for bead/cell recovery

General workflow. The D² microfluidic operation included droplet generation, sorting, collection and recovery. All the steps were performed with the in-house sorting software (uflow¹⁴). For bead and HEK cell recovery experiments, fluidic operation on the device followed four steps: first, a 2.5 mL syringe containing 2% 008-fluoro-surfactant HFE oil was connected to an oil inlet and was injected at a flow rate of 0.15 µL s⁻¹. Second, after priming all the channels with oil

Paper

and to ensure that all the droplets travel into the waste channel by default, a syringe for withdrawal was connected to the waste channel outlet and maintained at $-0.15~\mu L~s^{-1}$. Third, beads or cells were loaded into a 1 mL BD plastic syringe and introduced to the aqueous inlet of the chip at a flow rate of 0.05 $\mu L~s^{-1}$. Fourth, 300 μL recovery medium (PBS for beads, DMEM for HEK cells and IMDM for antibody-secreting cells) was loaded into a 500 μL gastight syringe, and the syringe was connected to the recovery medium channel inlet and injected at a flow rate of 1 $\mu L~s^{-1}$ to make a 3 μL recovery medium.

After droplets were generated at ~50 Hz, 470 nm wavelength light was cast at the detection fiber area. All the droplets prior to sorting were excited by the light and bead/ cell-containing droplets emitted green fluorescence signals (~520 nm); counts were recorded using the photon counting head. For single bead experiments, the sorting threshold was characterized and set between 7×10^3 counts and 2×10^4 counts (thresholding strategy shown in Fig. S2†). For singlecell experiments, the sorting threshold was set between 1 \times 10^3 counts and 2.5×10^3 counts. When the single bead/cell droplet passed the fiber region, a fluorescence peak signal was measured by the PMT system and shown on the uflowcontrol user interface (Fig. S2†). If the fluorescence counts were above the threshold of droplets containing one target and below the threshold line of droplets with two or more targets, our automation system triggered the electrodes at the sorting junction to sort the single-bead/cell droplets. After each actuation of the sorting electrode, a 3 µL recovery medium was generated at a flow rate of 1 µL s⁻¹ and 'parked' at the sorting channel outlet. After a 7 s delay post-sorting, the sorted 1 nL droplet arrived at the sorting channel outlet and was in contact with the recovery medium via actuation by the software. The mixture was transferred into a 96-well plate using a P20 pipette, and the pipette was set to 5 µL to collect all the media. Each manual transfer operation, from the device to the well, took ~10 seconds. The tubing method was similar to the D² workflow except that, after sorting, 6 cm Tygon tubing was connected to the sorting outlet and the end of the outlet tubing was directly guided to a well of a 96well plate. For our device design, we incorporated six pairs of electrodes in the sorting channel to arrest the sorted droplets via actuation. ESI† Movie S2 shows a sorted single-cell containing a droplet holding and merging with the recovery medium.

Fluorescent bead recovery. Before the experiment, the well plate was filled with 100 μ L per well of recovery medium (PBS). D^2 and tubing methods both followed the general workflow described above. For single recovery, 50 sorted droplets were collected and recovered into different individual wells. In the process of recovering 50 droplets into a well plate, 50 pipetting operations after merging the sorted droplet and the buffer were required using the D^2 method. For the tubing method, we manually moved the plate and guided the tubing outlet to individual wells after each sorting event. For pooled recovery, 50 sorted droplets were collected

and recovered into one well together, followed by a limiting dilution to achieve one bead/cell per well.³⁴ In the process of recovering 50 droplets to a well plate, one pipetting operation was required using the D² method; for the tubing method, removing the tubing from the device outlet and aligning it vertically to allow constituents containing the droplets (and oil) to flow into a well were required. After experiments, the well plates were stored in a mammalian cell incubator until downstream analysis.

GFP HEK-293 cell recovery. Before the experiment, the well plate was filled with 100 µL per well of recovery medium (DMEM). The D² method followed the general workflow and single recovery process described above. After experiments, the well plates were stored in a mammalian cell incubator until downstream analysis. For the viability test, 70 single cells were sorted/recovered into individual wells of a 96-well plate. A conditioned culture medium (150 µL, 30% supernatant from spent eGFP HEK cell culture filtered with a 0.22 µm filter and mixed with 70% fresh DMEM, 10% FBS and 1% penicillin-streptomycin added) was used for the first 24 hours post sorting, and a fresh medium (50 µL, 10% FBS and 1% penicillin-streptomycin added) was added every 48 hours to maintain the cells. From the starting time point (0 hour), after every 24 hours (within 7 days), SYTOX blue dead cell stain was added to 10 wells as per the manufacturer's protocol. The viability was calculated as the number of wells with non-blue cells per 10 wells. A dead cell stained with blue dye is shown in Fig. S3.† The proliferation images were taken every 24 hours using an EVOS M5000 imaging system.

Antibody-secreting hybridoma cell screening and recovery

Prior to experiments, cells were washed three times with neat IMDM to remove FBS from the media. Before cell encapsulation, we prepared a hybridoma cell mixture consisting of two cell lines: secreting monoclonal antibodies (mAbs) that bind to the receptor antigen expressed on the membrane of target cells and the other producing nonbinding mAbs in 1:400 and 1:1000 ratios. The specific mAbs producing hybridoma cell and target cell lines were stained with calcein-AM red-orange (shown as red) and CellTrace blue (shown as blue), respectively, as per the manufacturer's instructions. The target cells (3 \times 10⁶ cells per mL⁻¹) and goat anti-mouse Alexa Fluor 488 secondary antibodies (2.5 μ mL⁻¹) along with Optiprep (11%, w/v) were introduced from one inlet in the co-encapsulation chip at a flow rate of 0.3 μ L s⁻¹. The hybridoma cell mixture (1 \times 10⁶ cells per mL) along with Optiprep (11%, w/v) was introduced from another inlet in the co-encapsulation chip at a flow rate of 0.3 µL s⁻¹. HFE 7500 oil with 2% 008-fluoro-surfactant (RAN Biotechnologies) was used to produce droplets at a flow rate of 0.8 µL s⁻¹. A withdrawing syringe (BD 1 mL plastic) was connected to the chip outlet and set at a flow rate of 1.4 µL s⁻¹ to collect droplets, and the syringe was incubated for 2 hours in a cell incubator at 37 °C and in 5% CO2. The droplets were reinjected into the sorting and recovery chip using our D²

Lab on a Chip Paper

method at a flow rate of 0.02 µL s⁻¹, the spacing oil was set at a flow rate of 0.15 µL s⁻¹ and the waste syringe was set at -0.12 μL s⁻¹. The sorting and recovery were performed using in-house software (uflow) and the sorting threshold was set between 1.8×10^4 counts and 3×10^4 counts for the specific antibody-binding assay. The throughput of the sorting was ~50 Hz with a monodisperse droplet size of ~1 nL. After each sorting event, a 3 µL IMDM (10% FBS, 1 ng mL⁻¹ IL-6) recovery medium was generated at the collection channel outlet and a delay (7-9 seconds) was set before actuating the electrodes three times (500 ms per actuation) to merge the samples. The sorted droplet and recovery medium mixtures were individually transferred using a pipette to a 96-well plate for downstream analysis. The well plate was preloaded with 100 μL of IMDM (10% FBS, 1 ng mL⁻¹ IL-6, 1% penicillinstreptomycin) and immediately moved to an incubator after the experiment. The well plate was imaged using an EVOS M5000 imaging system.

Data processing and analysis

Data and statistical analyses were performed using GraphPad Prism 10. The image of the outlet tubing was taken using a phone camera. The images of 'stuck' droplets were taken using an inverted microscope (Olympus IX73) with brightfield illumination. The images of single beads and cells were taken using an EVOS M5000 imaging system. Bead/cell counting was done using Fiji (Image J) in the following steps: 1) open the image in Fiji software, 2) adjust the brightness & contrast by clicking 'auto', 3) apply a median filter to filter out noises (radius 5.0 pixels), and 4) find the maxima and output the number of the counts. The recovery rate was calculated by dividing the recovered (maxima) counts (Fiji) by the sorted counts (uflow):

$$Recovery \ rate \ (\%) = \frac{Recovered \ counts}{Sorted \ counts} \times 100$$

The applied electrical signal was measured using an oscilloscope. All in-house codes were previously described by our group. 14

Statistical analysis

Statistical analysis was performed using descriptive statistics (mean, standard deviation) and an unpaired t-test (p-value) with Prism V10.1.1 (GraphPad) for Fig. 2-4.

Results and discussion

Droplet-digital based microfluidics for sample collection

Droplet-based microfluidics has been frequently used for singlecell studies; however, such technologies are challenging with collection and recovery, resulting in unwanted sample loss.²⁹ To address this, we devised a new droplet-digital (D2) microfluidic strategy which has been reported as an efficient platform for on-demand merging, mixing and sorting of the droplets operations that were challenging with standard droplet-inchannel microfluidics.14 Many cell-based workflows generally require sorting, and fluorescence-activated droplet sorting was reported as a powerful tool for this purpose³⁵ (Fig. 1A). There are commercial systems (Cyto-Mine, 31 LiveDrop, 36 Atrandi Biosciences) that are effective in sorting and recovery; however, these are closed systems and are difficult to use for low-hit, rare, cell ratios (Table S1†). The D2 system can overcome these limitations using an open design for the recovery region, enabling direct pipetting of individually sorted droplets from the device to well plates.

We have previously described a droplet-to-digital platform that uses two devices connected by a capillary to transfer droplets from the generator to the digital part of the platform for screening cells.37 This technique, however, suffers from the requirement of tubing to transfer droplets, which can cause droplet losses and create difficulty to track a particular droplet. In addition, there is a mismatch in droplet numbers from generation (1000s) to droplet screening (only a few droplets at a time).

Here, we report a fully integrated droplet-digital screening platform, also being able to alleviate the droplet mismatch between the two systems. As shown in Fig. 1B(i and ii), are the steps for the droplet recovery. The D² microfluidic device is controlled by our in-house automation system (uflow, top view) and the device consists of three layers (side view). To ensure reliable droplet collection using the D2 method, we started with (iii) encapsulation of cells/beads in ~1 nL droplets, (iv) sorting of single cell/bead droplets using our fiber-based detection, 10 and (v) on-demand control to merge and to recover the sorted single-cell/bead-containing droplet. For the collection, the outlet was made with a 3 mm puncher and cut to a semicircular shape with one side opening as shown in the top view, offering direct access to pipetting samples from the collection channel. Sorted droplets were guided to the merging area by the oil flow of $\sim 0.05~\mu L~s^{-1}$. Two $200 \times 200~\mu m$ merging electrodes were patterned at the interface between the sorting channel and the merging area. The merging area contains a 3 μL recovery buffer that can merge with the sorted 1 nL droplet. After merging, these droplets were collected via manual pipetting and directly transferred to a well plate for downstream analysis. A key difference from previous dropletdigital interface work³⁷ is that our system reported here maximizes the capabilities of the droplet sorter and digital microfluidics to control the collection.

Comparison between D² and the conventional method for fluorescent bead recovery

Pooling droplets allows for the collection and grouping of desired droplet samples for many different applications. 10,12,18,25,38 For example, Debs et al. used a pooling strategy to recover sorted droplets for functional single-cell hybridoma screening, and after sorting, the authors pooled cell-containing droplets and performed droplet breakup for secreted antibody measurements.²⁵ As an initial test, we compared the traditional collection with D2 collection of fluorescent beads by pooling the sorted droplets.




Fig. 1 Traditional system (FADS) and D^2 device for droplet sorting and recovery workflows. (A) Traditional droplet microfluidic system for sorting/ recovering single-cell targets from a pooled cell population. All droplets fulfilling the sorting criteria are sorted into the same reservoir of hits, making their separation for individual analysis challenging and prone to sample loss. (B) D^2 device. The contact pads are connected to the inhouse automation control system (top view). A zoomed-in top view of the semicircular outlet with electrodes for droplet collection (left) and side view of the device (right). Droplet collection starts with D^2 sorting (i) followed by D^2 droplet recovery (ii). The dashed line around the single cell indicates that the sorted droplet has been merged with the recovery media. These operations are shown as images from a movie: droplet generator (iii), DMF sorter (iv) and droplet merger between the single cell droplet with the buffer droplet (v).

Droplet generation and sorting are the same for both techniques, except that the processed droplets are pooled and then transferred to a centrifuge tube or a well plate through tubing or using the on-demand D² collector. To maximize droplet collection with the traditional techniques, we followed other studies for collection in which protocols

such as removing the tubing from the device after sorting and orienting the tubing to vertically drain the liquid to the collection tube by gravity.³⁵ As shown in Fig. 2A is the recovery rate comparing the two different methods. We measured the number of pooled beads that can be recovered in the well-plate after the same number of sorting events.




Fig. 2 Comparison between D^2 and tubing for multi-bead recovery. (A) Recovery rate of recovering X number of beads into one well after X sorting events. (X = 5, 10, 20, 50). (B) Images of the tubing outlet indicating that the sorted droplets were 'trapped' at the tubing 'arc'. (C) An illustration of the PDMS-tubing interface (top) and an image of sorted droplets being 'trapped' at the connection junction (bottom). All error bars represent the \pm range; ns indicates no significant difference. Statistical analysis was performed using an unpaired t-test.

The D² method achieved an average collection recovery rate of ~93.5% compared to 78.5% with the traditional tubing method. Moreover, the significant difference is in the variation of the recovery (as shown by the error bars). The traditional case showed a large range of recovery rates, especially for the higher number of beads, while the D² case showed more reproducible collection regardless of the number of events. The tubing-based collection method showed a high standard deviation of 26.1%, reflecting substantial inconsistencies in recovery, as opposed to the D² method, which showed a lower standard deviation of 5.29% (Table S2 \dagger). These results describe that the D² system effectively eliminates the variability introduced by tubing collection. Particularly when collecting 50 beads, the D² system achieved a fivefold reduction in standard deviation, highlighting its robustness and consistency in sample handling. We speculate that the tubing configuration, as depicted in Fig. 2B and C, creates an "arc" for transferring processed droplets downstream, leading to a static area where droplets can accumulate. The commonly used HFE oil is heavier than the aqueous droplets, and thus the droplets tend to remain at the top region of the tubing (Fig. 2B). Additionally, in a conventional tubing setup for pooled transfer, the sorted droplets are in contact with each other, which occasionally results in droplet coalescence or crosstalk. In the current system, the droplets are isolated or "parked" individually, preventing any fusion or diffusive exchange. Moreover, when the number of sorted droplets is low, the flow rate is insufficient to displace them out of the tubing, leading to retention within the arc. Retention is also a problem at the interface between the PDMS and the tubing (Fig. 2C). Since commonly used tubing materials (Tygon or PEEK) have a tubing wall thickness greater than the droplet

diameter ($<100~\mu m$), sorted droplets became trapped along the tubing wall. However, in the D^2 system, since no tubing is required, the above two causes of sample loss will be mitigated. The new system is capable of easily collecting the pooled droplets and transferring them to well-plates with high fidelity.

Instead of pooling our droplet samples, we evaluated similar steps for collecting and transferring contents from individual cell-containing droplets. Individual cell-containing droplet collection is commonly required for enrichment of desired products or mutants from a mixed population.³⁹ For single recovery, we recovered the contents from droplets after each sorting event using tubing or D^2 collection and transferred them into individual wells of a 96-well plate, resulting in one bead per well. Interestingly, we observed a significantly higher mean recovery rate (\sim 87%) than the tubing method (\sim 5%) for single bead recovery (Fig. 3A). After collection, we acquired images of single cells in the well plate, confirming that we are able to transfer the single beads with our new method (Fig. 3B).

Low-input rare single-cell recovery using the D² platform

The recovery of low and rare cells is interesting for many applications, 40 but the collection of such cells has always been challenging. 49 We constructed a novel microfluidic system for high recovery rates, high viability cell sorting and recovery. Therefore, we conducted a series of experiments using our platform to sort cell-containing droplets and collect single cells in multi-well plates to assess the D² performance. By evaluating the recovery rates and post-sorting viability of GFP-expressing HEK 293 cells, we aimed to determine whether the D² method can recover the rare population.

Paper Lab on a Chip




Fig. 3 Comparison between D^2 and traditional tubing technique for single-bead recovery. (A) Recovery rate of isolating/recovering 50 beads into 50 individual wells, recovering one droplet at a time to one well (single recovery). All error bars represent \pm SD; *, **, ***, and **** represent p-values below 0.05, 0.01, 0.001 and 0.0001, respectively. Statistical analysis was performed using an unpaired t-test. (B) Images of nine wells filled with single beads from a 96-well plate, scale bar: 200 μ m. Zoomed-in views of the recovered beads are shown in the white boxes.

For this experiment, GFP-expressing HEK 293 cells were prepared at two different cell numbers, 10⁴ and 10⁵ cells, and in each set, different ratios (1:400 and 1:1000) of GFP to WT cells were spiked. Briefly, we performed cell sorting and collection/recovery into a 96-well plate, and validated the recovery rates and viability changes within 7 days post sorting/recovery for the D² system. The comparative data describing the recovery rate for the GFP cells using the D² method are shown in Fig. 4A. Compared to previously reported recovery rates from 20 to 30%, ¹⁵ the D² method shows excellent recovery rates of over 80% for both cell inputs (10⁴, 10⁵) and cell ratios (1:400, 1:1000). As shown in Fig. 4B is an example spectrometer plot in the D² workflow; the green line indicates the sorting threshold and peaks

above the threshold were sorted, counted, recovered and transferred to a well plate. Here, using our 1:1000 dilution ratio, six GFP target cells were sorted, counted and recovered during ~ 1 h of operation. Images of recovered GFP cells in the well plates suggest that the D^2 method is suitable for collecting single cells for downstream analysis.

A unique strength of the D² method is that it uses low electric fields to sort and collect the cells, which minimizes the electric field-induced damage to the cells. Instead of using conventional dielectrophoresis (DEP) with injected electrodes, ^{12,41} we used the D² binary sorter by patterning electrodes beneath the channel, which requires lower operation voltage for sorting. ⁴² The DEP sorting systems usually require voltages ranging from 200 V to 1200 V, ^{43,44} while the D² sorter uses a lower voltage of less than




Fig. 4 Low-input rare single-cell recovery using D^2 . (A) Recovery rate of D^2 , using different spike-in ratios. (B) The graph (left) shows the D^2 sorting threshold (green line) and sorting conditions (total cells, spike-in ratios). Images of six recovered single cells in a well plate (right), scale bar: 20 μ m. (C) Single HEK cell viability over 7 days using D^2 (left). Bright-field and fluorescence images of 7-day single HEK cell proliferation, sorted and recovered by the D^2 system (right), scale bar: 100 μ m.

Lab on a Chip Paper

140 V that can preserve the cell's health. To test this, single HEK cells were sorted, recovered, and cultured in well plates. As shown in Fig. 4C, we compared our results to previously reported postsorting viability (~95% for MCF-7 cells, 45 ~79% for Jurkat cells and ~90% for HEK293T cells¹⁵), and found that the cells that were collected with D^2 are similar in terms of viability (~90%) and also showed proliferation over seven days. In addition, the cells collected by the D² method exhibit spreading, a positive indicator of viability, suggesting that the platform effectively preserves cell health post-isolation and is well-suited for downstream functional assays.

Application: rare antibody-secreting cell recovery

Droplet-based microfluidics has been reported as a useful screening tool for finding specific antibody-secreting cells. 12,19 Hybridoma technology is one of the most common methods to produce monoclonal antibodies. It fuses antibody-secreting B cells with immortal myeloma cell lines to create a hybrid cell line that can be expanded to produce monoclonal antibodies.⁴⁶ However, the antigen-specific B cells in the peripheral blood of immunized human donors are very rare, and the ratio of such cells is usually in the range of 0.1-2%.20 Therefore, recovery of these rare hits is critical, yet challenging. There is a dropletbased platform for directly identifying and isolating antibodyencoding genes from rare screening hits, but this setup does not allow recovery of live cells for downstream expansion.⁴⁷

To showcase the D² system's capability of recovering sorted rare targets, we performed a specific antibody binding assay. We prepared hybridoma cell line mixtures consisting of two cell lines: one producing specific-binding monoclonal antibodies (mAbs) that bind to the cell-surface-antigen-overexpressing target cell line, and the other producing non-specific mAbs which will not bind to the surface of the target cells. In these experiments, we mixed Alexa Fluor 488 secondary antibodies with a plain IMDM solution containing target cells to be able to investigate specific mAbs binding through fluorescence intensity detection. The specific and non-specific antibody-secreting hybridoma cells were mixed in ratios of 1:400 (0.25%) and 1:1000 (0.1%) to match the low ratio of specific antibody-secreting B cells. Only the specific hybridoma cells were stained with calcein AM red and the target cells were stained with CellTrace blue. The hybridoma mixture and target cells were then introduced into a co-encapsulation device to form double-cell droplets (5% to 10%, based on Poisson statistics), as shown in Fig. 5A. Images of recovered cells in well plates and their corresponding green fluorescence intensities are shown in Fig. 5B. In this workflow, we sorted based on the green fluorescence intensity. All droplets contain the same amount of




Fig. 5 Rare specific-antibody-secreting cell recovery. (A) An illustration of the co-encapsulation chip. (B) Schematic showing the different cell combinations in droplets that correlate to the detected fluorescence signal (see arrows). A spectrometer plot showing different fluorescence intensity peaks. Sorting was triggered based on the threshold shown as the green line. Microscopy images of one hybridoma cell with one target cell, one hybridoma cell, and one target cell were taken at four different color channels to show the cell staining strategy: hybridoma Ab-producer (red), target cell (blue), and secondary antibody (green). Scale bars: 10 μm. (C) Plot showing the droplet recovery rate (%) for spiked 'rare' 1:400 and 1:1000 ratios (specific antibody-secreting cells: non-specific antibody-secreting cells). All error bars represent \pm SD.

Lab on a Chip Paper

fluorescent secondary antibodies. In the presence of a specificantibody-secreting hybridoma cell and a target cell, there will be an accumulation of fluorescent secondary antibodies on the target cell surface, resulting in a sharp, high green fluorescence intensity peak within the droplet. Droplets with only antibodysecreting cells emitted medium fluorescence peaks and droplets with only target cells emitted background fluorescence intensity. The fluorescence threshold for the green channel was set at 18 000 photon counts, which correlates to the peaks generated by the specific antibody binding to the surface of target cells. Using this threshold, we were able to obtain specific-mAb-secreting hybridoma and target cell doublet combinations. We verified the number of recovered specific antibody-secreting hybridomatarget cell combinations in well plates with the number of sorted droplet counts in our software. The recovery rates of 1:400 and 1:1000 spike-in ratios are shown in Fig. 5C. For the 1:400 ratio, we have a mean recovery rate of around 88%, and for the 1:1000 ratio, the mean recovery rate is around 78% and is generally better than the recovery rates of ~30% for recovering cells from droplets reported in the literature. 15 The enhanced performance can be attributed to the on-demand, selective collection of droplets based on phenotypic or fluorescence readouts, which significantly reduces false positives and minimizes sample loss due to nonspecific or bulk collection. In contrast, typical commercial systems like Cyto-Mine rely on pre-set gating, passive flow valve-based mechanisms, which are inherently limited in collection precision and are susceptible to high levels of stochastic loss, particularly during droplet transfer via tubing or plate deposition.

Conclusion

We have reported a droplet-digital microfluidic platform for the efficient isolation of rare single cells in phenotypic assays. The platform can generate droplets, sort them based on fluorescence intensity, and collect the contents of the sorted droplets with a µL-scale volume recovery buffer, in a controllable and automated manner. To demonstrate the utility of our platform, we have performed recovery experiments with single fluorescent beads, single HEK 293 cells, and single specific antibody-secreting hybridoma cells at low ratios (1:400 to 1:1000). For these experiments, we were able to obtain significantly higher recovery rates (>80%), as compared to previously published systems, 15,17-19,48 establishing it as a versatile tool for typical applications in single-cell analysis, especially for rare cell isolation/expansion and single-cell antibody screens. Lastly, the design of the recovery region can potentially be adapted to any droplet microfluidic devices where precise recovery of processed droplets is required, making it a convenient module for different droplet-based systems.

Data availability

The automated script and software for sorting cells were used to support this manuscript and are available on our shared bitbucket website (see the Methods section). We have

included the CAD files for the devices on our Open Science Framework website https://osf.io/jmxc6/. The raw data associated with this manuscript are available from the corresponding author upon request.

Author contributions

CAM and SCCS envisaged the project and coordinated the study. ZD, CAM, and SCCS designed the experiments and analyzed the data. JMP, MW and RG helped with the design of the microfluidic device (droplet generator, sorter geometry). ZD designed (sorter electrodes, recovery region) and fabricated all devices, and carried out all experiments except for the automation software (uflow) established previously. MW and AA helped with troubleshooting the antibody assay. ZD, CAM, and SCCS wrote the manuscript which was revised and reviewed by all authors.

Conflicts of interest

There are no conflicts to declare.

Acknowledgements

The authors would like to thank the Human Health Therapeutics Research team at National Research Council Canada for providing cell lines. ZD thanks VERAXA Biotech GmbH for funding and technical support when designing the device and antibody assay. The authors thank the Natural Sciences and Engineering Research Council (NSERC; RGPIN-2021-04101), the Canadian Foundation of Innovation (CFI JELF; 35649), and the Swiss National Science Foundation (SNSF; 222922) for funding. ZD thanks the Department of Electrical and Computer Engineering, Concordia University for FRS funding. SCCS thanks Concordia University for a Research Chair.

References

- 1 S. H. Gohil, J. B. Iorgulescu, D. A. Braun, D. B. Keskin and K. J. Livak, Nat. Rev. Clin. Oncol., 2021, 18(4), 244-256.
- 2 L. Rebuffet, J. E. Melsen, B. Escalière, D. Basurto-Lozada, A. Bhandoola, N. K. Björkström, Y. T. Bryceson, R. Castriconi, F. Cichocki, M. Colonna, D. M. Davis, A. Diefenbach, Y. Ding, M. Haniffa, A. Horowitz, L. L. Lanier, K.-J. Malmberg, J. S. Miller, L. Moretta, E. Narni-Mancinelli, L. A. J. O'Neill, C. Romagnani, D. G. Ryan, S. Sivori, D. Sun, C. Vagne and E. Vivier, Nat. Immunol., 2024, 25(8), 1474-1488.
- 3 P.-H. Li, X.-Y. Kong, Y.-Z. He, Y. Liu, X. Peng, Z.-H. Li, H. Xu, H. Luo and J. Park, Mil. Med. Res., 2022, 9(1), 52.
- 4 A. Pedrioli and A. Oxenius, Trends Immunol., 2021, 42(12), 1143-1158.
- 5 J. Van Lent, J. Breukers, K. Ven, L. Ampofo, S. Horta, F. Pollet, M. Imbrechts, N. Geukens, K. Vanhoorelbeke, P. Declerck and J. Lammertyn, Lab Chip, 2021, 21(19), 3627-3654.
- 6 M. Piwecka, N. Rajewsky and A. Rybak-Wolf, Nat. Rev. Neurol., 2023, 19(6), 346-362.

7 T. J. Walter, R. K. Suter and N. G. Ayad, Neurobiol. Dis., 2023, 184, 106201.

Lab on a Chip

- 8 K. Matuła, F. Rivello and W. T. S. Huck, Adv. Biosyst., 2020, 4(1), 1900188.
- 9 X. Xu, J. Wang, L. Wu, J. Guo, Y. Song, T. Tian, W. Wang, Z. Zhu and C. Yang, Small, 2020, 16(9), 1903905.
- 10 C. Leal-Alves, S. Dumont, Z. Deng, S. Alkhaldi, Z. Leung, M. Oeser and S. C. C. Shih, Adv. Mater. Technol., 2025, 10(6), 2401209.
- 11 A. R. Abate, T. Hung, P. Mary, J. J. Agresti and D. A. Weitz, Proc. Natl. Acad. Sci. U. S. A., 2010, 107(45), 19163-19166.
- 12 N. Shembekar, H. Hu, D. Eustace and C. A. Merten, Cell Rep., 2018, 22(8), 2206-2215.
- 13 W. Verbist, J. Breukers, S. Sharma, I. Rutten, H. Gerstmans, L. Coelmont, F. Dal Dosso, K. Dallmeier and J. Lammertyn, Lab Chip, 2024, 24(7), 2107-2121.
- 14 F. Ahmadi, H. Tran, N. Letourneau, S. R. Little, A. Fortin, A. N. Moraitis and S. C. C. Shih, Small, 2024, 20(26), 2308950.
- 15 J. Clausell-Tormos, D. Lieber, J.-C. Baret, A. El-Harrak, O. J. Miller, L. Frenz, J. Blouwolff, K. J. Humphry, S. Köster, H. Duan, C. Holtze, D. A. Weitz, A. D. Griffiths and C. A. Merten, Chem. Biol., 2008, 15(5), 427-437.
- 16 R. Murugan, K. Imkeller, C. E. Busse and H. Wardemann, Eur. J. Immunol., 2015, 45(9), 2698-2700.
- 17 C. Sun, L. Liu, L. Pérez, X. Li, Y. Liu, P. Xu, E. A. Boritz, J. I. Mullins and A. R. Abate, Nat. Biomed. Eng., 2022, 6(8), 1004-1012.
- 18 R. Ding, K. C. Hung, A. Mitra, L. W. Ung, D. Lightwood, R. Tu, D. Starkie, L. Cai, L. Mazutis, S. Chong, D. A. Weitz and J. A. Heyman, RSC Adv., 2020, 10(45), 27006-27013.
- 19 R. Gaa, E. Menang-Ndi, S. Pratapa, C. Nguyen, S. Kumar and A. Doerner, mAbs, 2021, 13(1), 1978130.
- 20 A. P. Kodituwakku, C. Jessup, H. Zola and D. M. Roberton, Immunol. Cell Biol., 2003, 81(3), 163-170.
- 21 P. Paterlini-Brechot and N. L. Benali, Cancer Lett., 2007, 253(2), 180-204.
- 22 D. R. Parks, V. M. Bryan, V. T. Oi and L. A. Herzenberg, Proc. Natl. Acad. Sci. U. S. A., 1979, 76(4), 1962-1966.
- 23 S. Ma, B. Zhang, L. M. LaFave, A. S. Earl, Z. Chiang, Y. Hu, J. Ding, A. Brack, V. K. Kartha, T. Tay, T. Law, C. Lareau, Y.-C. Hsu, A. Regev and J. D. Buenrostro, Cell, 2020, 183(4), 1103-1116.e1120.
- 24 K. Portmann, A. Linder, N. Oelgarth and K. Eyer, Cells Rep. Methods, 2023, 3(7), 100502.
- 25 B. E. Debs, R. Utharala, I. V. Balyasnikova, A. D. Griffiths and C. A. Merten, Proc. Natl. Acad. Sci. U. S. A., 2012, 109(29), 11570-11575.
- 26 M. Karbaschi, P. Shahi and A. R. Abate, Biomicrofluidics, 2017, 11(4), 044107.
- 27 R. Zilionis, J. Nainys, A. Veres, V. Savova, D. Zemmour, A. M. Klein and L. Mazutis, Nat. Protoc., 2017, 12(1), 44-73.

- 28 R. H. Cole, S.-Y. Tang, C. A. Siltanen, P. Shahi, J. Q. Zhang, S. Poust, Z. J. Gartner and A. R. Abate, Proc. Natl. Acad. Sci. U. S. A., 2017, 114(33), 8728-8733.
- 29 L. Nan, M. Y. A. Lai, M. Y. H. Tang, Y. K. Chan, L. L. M. Poon and H. C. Shum, Small, 2020, 16(9), e1902889.
- 30 M. Weiss, S. Hasan, R. Genth, M. Mollah, E. Robert, A. Gil and L. Hufnagel, Lab Chip, 2025, 25(4), 600-612.
- 31 D. Josephides, S. Davoli, W. Whitley, R. Ruis, R. Salter, S. Gokkaya, M. Vallet, D. Matthews, G. Benazzi, E. Shvets, F. Gesellchen, D. Geere, X. Liu, X. Li, B. Mackworth, W. Young, Z. Owen, C. Smith, D. Starkie, J. White, B. Sweeney, M. Hinchliffe, S. Tickle, D. J. Lightwood, M. Rehak, F. F. Craig and D. Holmes, SLAS Technol., 2020, 25(2), 177-189.
- 32 R. Fobel, C. Fobel and A. R. Wheeler, Appl. Phys. Lett., 2013, 102(19), 193513.
- 33 M. Abdelgawad, M. W. L. Watson and A. R. Wheeler, Lab *Chip*, 2009, **9**(8), 1046–1051.
- 34 L. W. G. Strijbosch, W. A. Buurman, R. J. M. M. Does, P. H. Zinken and G. Groenewegen, J. Immunol. Methods, 1987, 97(1), 133-140.
- 35 J.-C. Baret, O. J. Miller, V. Taly, M. Ryckelynck, A. El-Harrak, L. Frenz, C. Rick, M. L. Samuels, J. B. Hutchison, J. J. Agresti, D. R. Link, D. A. Weitz and A. D. Griffiths, Lab Chip, 2009, 9(13), 1850-1858.
- 36 K. Pernod, N. Coppieters, Q. Graillet, A. Wagner, C. Arnaud and S. van Loo, D-Sire & ModaFlow for High-Throughput Antibody Discovery, 2025.
- 37 S. C. C. Shih, P. C. Gach, J. Sustarich, B. A. Simmons, P. D. Adams, S. Singh and A. K. Singh, Lab Chip, 2015, 15(1), 225-236.
- 38 F. Lan, B. Demaree, N. Ahmed and A. R. Abate, Nat. Biotechnol., 2017, 35(7), 640-646.
- 39 P. Hu, W. Zhang, H. Xin and G. Deng, Front. Cell Dev. Biol., 2016, 4, 116.
- 40 Y. Chen, P. Li, P. H. Huang, Y. Xie, J. D. Mai, L. Wang, N. T. Nguyen and T. J. Huang, Lab Chip, 2014, 14(4), 626-645.
- 41 L. Mazutis, J. Gilbert, W. L. Ung, D. A. Weitz, A. D. Griffiths and J. A. Heyman, Nat. Protoc., 2013, 8(5), 870-891.
- K. Samlali, C. L. Alves, M. Jezernik and S. C. C. Shih, Microsyst. Nanoeng., 2022, 8(1), 123.
- 43 S. S. Schütz, T. Beneyton, J.-C. Baret and T. M. Schneider, Lab Chip, 2019, 19(13), 2220-2232.
- 44 O. Caen, S. Schütz, M. S. S. Jammalamadaka, J. Vrignon, P. Nizard, T. M. Schneider, J.-C. Baret and V. Taly, Microsyst. Nanoeng., 2018, 4(1), 33.
- 45 L. Chen, Y. Xu, L. Zhou, D. Ma, R. Zhang, Y. Liu and X. Mi, Microsyst. Nanoeng., 2025, 11(1), 10.
- 46 S. Mitra and P. C. Tomar, J. Genet. Eng. Biotechnol., 2021, 19(1), 159.
- 47 A. Autour and C. A. Merten, Proc. Natl. Acad. Sci. U. S. A., 2024, **121**(37), e2405342121.
- 48 R. Murugan, K. Imkeller, C. E. Busse and H. Wardemann, Eur. J. Immunol., 2015, 45(9), 2698-2700.