



Cite this: DOI: 10.1039/d6cc01897a

Rapid bacterial diagnostics at the point of care: emerging electrochemical and Raman-based approaches

 Petr Jakubec,^a Václav Ranc,^{ab} Milan Kolář,^c Ondřej Zítka,^{de} David Panáček^a and Michal Otyepka^{af}

Bacterial infections impose a major clinical and economic burden, which is likely to increase further as antimicrobial resistance rises, reinforcing the need for early detection to accelerate and better target therapy. Conventional culture-based diagnostics are well established, but they are too slow to inform early therapeutic decisions. This review frames the growing need for rapid bacterial diagnostics in the context of acute infection, antimicrobial stewardship, and point-of-care testing. We focus on emerging electrochemical biosensing and Raman-based approaches and compare them with established laboratory methods, including culture, MALDI-TOF MS, nucleic acid amplification tests, sequencing, and antimicrobial susceptibility testing, emphasising where central laboratories remain essential and where near-patient testing could alter clinical management. We highlight clinically relevant targets, recognition strategies, signal generation, materials design, manufacturability, and performance in complex biological matrices. Rather than focusing only on analytical sensitivity, we discuss the practical factors that will determine translation, including reproducibility, workflow integration, validation in real samples, and regulatory readiness. Taken together, the most clinically promising rapid bacterial diagnostics are those that deliver reliable, actionable information within the time window in which treatment decisions can still meaningfully influence outcomes. In this context, the most promising point-of-care systems are not those intended to replace central laboratories, but those designed to complement them by enabling earlier, better-targeted, and more judicious antimicrobial therapy.

 Received 28th March 2026,
 Accepted 29th May 2026

DOI: 10.1039/d6cc01897a

rsc.li/chemcomm

1. Clinical need and diagnostic landscape

Bacterial infections still place a heavy burden on healthcare systems, not only because of the illness and deaths they cause, but also because of the cost of managing them.¹ Treating these

infections effectively depends on identifying the causative pathogen, and doing so early enough for the result to actually shape clinical decisions. That is why “rapid” is best understood in clinical, not just technical, terms. In sepsis and other acute infections, rapid may mean minutes or a few hours, since even modest delays in appropriate therapy can cause serious consequences.^{2–7} In less urgent situations, such as urinary or respiratory tract infections, a same-day response may be sufficient to support targeted treatment and avoid unnecessary empirical prescribing. Conventional culture-based diagnostics, however, generally take 24–72 hours to deliver definitive identification and susceptibility results, making them too slow for many early treatment decisions.^{2,5,6} This gap is especially obvious in severe infections such as sepsis, where early recognition and timely antimicrobial therapy are closely tied to patient survival.^{2–6} The burden is not only clinical. These infections often lead to longer hospital stays, intensive care admission, and costly management of complications.^{3–5} Faster microbiological diagnostics therefore matter at two levels: they can improve outcomes for individual patients and ease pressure on healthcare systems. Just as importantly, earlier

^a Regional Centre of Advanced Technologies and Materials, Czech Advanced Technology and Research Institute (CATRIN), Palacký University Olomouc, Šlechtitelů 27, Olomouc 779 00, Czech Republic. E-mail: p.jakubec@upol.cz, michal.otyepka@upol.cz

^b Institute of Molecular and Translation Medicine, Faculty of Medicine and Dentistry, Palacký University Olomouc, Hnevotinska 5, Olomouc CZ-779 00, Czech Republic

^c Department of Microbiology, Faculty of Medicine and Dentistry, Palacký University Olomouc, Hněvotinská 3, 775 15 Olomouc, Czech Republic

^d Department of Chemistry and Biochemistry, Mendel University in Brno, Zemědělská 1, Brno, 613 00, Czech Republic

^e Faculty of Electrical Engineering and Communication, Department of Microelectronics, Brno University of Technology, Technická 3058/10, 616 00 Brno, Czech Republic

^f IT4Innovations, VSB-Technical University of Ostrava, Ostrava-Poruba, Czech Republic



pathogen identification makes it easier to move away from broad-spectrum empirical treatment toward more targeted therapy that is, in most cases, more appropriate.^{8,9} That point matters even more in the context of antimicrobial resistance (AMR). Empirical antibiotic treatment is often unavoidable, particularly when a patient is critically ill, but the longer broad-spectrum therapy is continued without microbiological clarification, the greater the selective pressure it creates. That, in turn, supports the emergence and spread of AMR.^{8,9} Rapid diagnostics can help interrupt this pattern by shortening the time to identification, enabling earlier de-escalation where justified, and, ideally, providing at least some information on likely susceptibility or resistance. Seen this way, rapid bacterial diagnostics are not just diagnostic tools; they are also part of the practical machinery of stewardship.^{8–11} Even so, central microbiology laboratories remain indispensable. Culture still underpins the field because it yields viable isolates for confirmatory identification, epidemiological work, and phenotypic antimicrobial susceptibility testing. Modern instrumental laboratory methods such as matrix-assisted laser desorption/ionisation time-of-flight mass spectrometry (MALDI-TOF MS), molecular assays, sequencing, and rapid phenotypic antimicrobial susceptibility testing (AST) have accelerated key steps in

the workflow, but they have not eliminated the need for centralised facilities or expert interpretation. The point of point-of-care (POC) and near-patient testing, then, is usually not to replace the central laboratory altogether. It is to bring actionable information closer to the clinical encounter. In the emergency department, intensive care unit, outpatient clinic, or field setting, an earlier result may influence triage, the start or narrowing of therapy, infection-control measures, or the decision to escalate to more comprehensive laboratory testing.^{12–14} These pressures shape what next-generation rapid bacterial diagnostics need to deliver.¹⁵ Ideally, they should work directly from clinical samples with little preparation, produce results within a genuinely useful timeframe, and remain reliable in complex biological matrices. They also need to balance analytical performance with practical concerns such as ease of use, robustness, cost, and compatibility with stewardship goals. Where possible, they should go beyond simple organism detection and provide resistance-related information that can guide treatment more precisely.¹⁶ In the end, the most useful rapid diagnostics will not necessarily be the fastest in purely analytical terms, but the ones that deliver dependable, clinically relevant information at the point where it can still change what happens next.



Petr Jakubec

Petr Jakubec received his PhD in Physical Chemistry from Palacký University Olomouc in 2013. He is currently a researcher at the Czech Advanced Technology and Research Institute, Palacký University Olomouc, where he leads the Synthesis and Characterization of Advanced Materials group. His research focuses on electrochemistry, advanced carbon-based materials, biosensors, and their applications in the detection of

bioactive molecules and pollutants, as well as in energy-storage systems. He has co-authored more than 60 scientific papers and has contributed to several national and international research projects.



Václav Ranc

Václav Ranc received his PhD in Analytical Chemistry in 2009 from Palacký University Olomouc, Czech Republic, under the supervision of Prof. Karel Lemr. From 2008 to 2010, he held the position of Assistant Professor at the Department of Analytical Chemistry, Palacký University Olomouc. In 2011, he joined the group of Prof. Rainer at the Department of Medicine, University of Fribourg, Switzerland, as a Research

Scientist. He subsequently obtained his habilitation in Physical Chemistry from Palacký University Olomouc. Since 2012, he has been affiliated with Palacký University, where he currently serves as Associate Professor and Head of the Nanomaterials in Medicine research group at the Institute of Molecular and Translational Medicine (IMTM), Faculty of Medicine and Dentistry and senior researcher at the Regional Centre of Advanced Technologies and Materials division at CATRIN. He is the Principal Investigator of Horizon Europe project STRIKE (MSCA Doctoral Network). He has authored or co-authored over 85 scientific publications, holds an H-index of 29, and is co-author of 3 international patents. His research is directed towards the development of plasmonic nanosensors based on Surface-Enhanced Raman Spectroscopy, synthesis of functionalized nanomaterials, and stain-free microscopic techniques encompassing AFM, confocal Raman, and infrared microscopy.



Scope of this review. This review focuses on electrochemical biosensing and Raman-based optical approaches, especially

surface-enhanced Raman scattering (SERS), as two emerging routes for rapid bacterial diagnostics at the point of care.^{17,18}



Milan Kolář

Prof. MUDr. Milan Kolář, PhD, is a Czech clinical microbiologist and academic affiliated with the Faculty of Medicine and Dentistry, Palacký University Olomouc, and the University Hospital Olomouc, where he holds senior roles in both research and clinical practice. He has held major leadership positions in clinical microbiology, contributing extensively to the implementation of antimicrobial

stewardship programmes. His current scientific interests focus on bacterial infections, antimicrobial resistance, novel antibacterial strategies, and the development of advanced antimicrobial materials, including nanomaterials.



Ondřej Zítka

Ondřej Zítka is Associated Professor, Deputy Head of Department of Chemistry and Biochemistry, Mendel University in Brno where he leads the Research Group of Biomarkers. He received his PhD in Chemistry (2012) from Mendel University in Brno, and RNDr. (2014) in Biochemistry from Charles University in Prague. He was awarded by the prize of Minister of education, youth and sports of Czech Republic. His

current research interests cover bioanalytical chemistry, magnetic particle separations, bioelectrochemistry, and mass spectrometry. In the last 10 years, he has focused on the development of diagnostic methods in the field of immunoassay and molecular diagnostics, and is also the founder and chief scientific officer of the startup NANTEO. He has (co-)authored 204 original articles or reviews in ISI indexed journals. His h-index according to WOS is 40.



David Panáček

David Panáček received his PhD in Materials Chemistry from Palacký University Olomouc, Czech Republic. His research focuses on the design and application of advanced graphene-based materials, including functional graphene derivatives and single-atom-engineered systems. He has contributed to the development of nanomaterials for antimicrobial applications, electrochemical sensing,

biosensing, and plant monitoring. His current work explores printed and wearable sensor technologies for agriculture, with an emphasis on scalable, material-efficient, and biocompatible platforms for real-time plant diagnostics. More recently, his research has expanded toward graphene-based inks, electrode functionalization, and integrated sensing systems that connect fundamental materials chemistry with practical technologies for environmental, biomedical, and agricultural applications.



Michal Otyepka

Michal Otyepka serves as Head of CATRIN-MAT at Palacký University Olomouc and Professor & Head of Nanolab at IT4Innovations, VŠB-TUO, Czech Republic. His research bridges nanomaterials chemistry, molecular modeling, and advanced functional materials. He is recognized for his pioneering contributions to the discovery and chemistry of fluorographene and for establishing fluorographene as a

versatile precursor for the synthesis of a broad family of functional graphene derivatives. His work led to the development of graphene-based materials, which have found applications in electrochemical sensing, biosensing, energy technologies, and catalysis. His research combines advanced experimental approaches with computational chemistry and molecular simulations to uncover structure–property relationships in complex molecular and nano-scale systems. More recently, his research has expanded toward single-atom materials and graphene-based inks for high-performance electrochemical sensors and biosensors, contributing to the translation of fundamental nanomaterials research into practical technologies. Prof. Otyepka has authored more than 400 scientific publications and several patents.



They both combine short analysis times with portable instrumentation, potential for miniaturisation, and compatibility with disposable or low-cost formats. Electrochemical methods provide quantitative output through simple electronics and readily manufacturable electrode platforms, whereas Raman offers molecular fingerprinting and, particularly in SERS, strong signal enhancement. In both cases, performance is governed by the chemistry of the sensing interface; in electrochemical platforms by receptor immobilisation, surface charge, electron transfer, antifouling layers, and matrix-tolerant electrode design, and in SERS by reproducible plasmonic substrates with controlled hotspot density, analyte access, surface functionalization, and signal normalisation. Other rapid diagnostic modalities, including nucleic-acid amplification tests (NAATs), sequencing, MALDI-TOF MS, and broader optical screening methods, are discussed only as benchmarks or contextual comparators.

2. Electrochemical and Raman-based POC strategies for rapid bacterial detection

2.1 Electrochemical approaches

Electrochemical biosensing is one of the most compatible strategies for POC bacterial diagnostics.^{19,20} Electrochemical measurements can be performed using compact, low-power,

and inexpensive instrumentation that can be readily integrated into disposable sensors and microfluidic platforms (Fig. 1A).^{21,22} Another key advantage is the inherently quantitative nature of the signal, since current, potential, charge, or impedance can be directly digitised and correlated with bacteria (or related analyte) concentration.²³ These characteristics make electrochemical biosensors attractive for rapid testing near the patient, where low cost, portability, and operational simplicity are essential. Practical performance is shaped by the choice of bacterial target, the selectivity of the biorecognition layer, the robustness of the signal-generation mechanism, and the manufacturability of the electrode platform.²⁴ Electrochemical sensing has evolved from proof-of-concept demonstrations toward more translational questions such as matrix tolerance, storage stability, antifouling behaviour, reproducibility, and compatibility with clinical workflows.^{25,26}

2.1.1 Detection targets. Electrochemical assays can be designed for several targets, including whole cells, cell lysates, secreted biomarkers, and nucleic acids or their amplification products (Fig. 1B).^{27,28} Whole-cell detection is appealing because it addresses the actual presence of the pathogen and may require only limited sample preparation.²⁸ At the same time, it is often complicated by non-specific adsorption, low bacterial abundance, and the difficulty of distinguishing closely related organisms in complex samples. Lysate-based detection²⁹ improves access to intracellular components and can generate stronger or more reproducible signals, but it adds a lysis step and may




Fig. 1 Schematic overview of electrochemical POC strategies for bacterial pathogen detection. (A) A representative workflow is shown from the collection of clinically relevant samples, including urine, blood, and wound swabs, to rapid readout using a portable POC device and disposable electrode. (B) The main analytical targets include whole bacteria, lysate-derived cell components, secreted biomarkers, and nucleic acids. (C) Selectivity at the electrode surface is achieved through different biorecognition layers, including antibodies, aptamers, and peptides. (D) The figure highlights representative transducing materials used at the electrode interface, including graphene, carbon nanotubes, MXene, and polymers. (E) Scalable electrode fabrication approaches, including screen printing and inkjet printing, illustrate the manufacturing routes that support translation toward disposable electrochemical diagnostic platforms. The figure was created using BioRender.com and Adobe Illustrator.



sacrifice information about cell viability.³⁰ Secreted biomarkers, such as toxins, enzymes, metabolites, or quorum-sensing molecules, offer another attractive route because they are soluble and can often be accessed more quickly than intact cells. Their presence may also reflect ongoing bacterial activity. Still, these analytes are not always uniquely linked to viable pathogen burden, which limits straightforward interpretation.³¹ Nucleic acid detection generally offers the highest specificity, since it can be directed toward species-, strain, or resistance-associated sequences.³² In many cases, though, this comes with the need for upstream amplification, with electrochemistry functioning as the final readout stage. This is particularly relevant in CRISPR-assisted formats, where sequence recognition is coupled to an electrochemical reporter system.^{33,34} In practice, the selection of the target should be guided by the intended clinical use of the assay,³⁵ rather than solely by the lowest reported limit of detection.³⁶ Whole-cell, lysate-based, and secreted-biomarker assays are useful for rapid detection, triage, or monitoring of bacterial activity.^{37,38} Nucleic-acid assays provide more specific information by detecting species-, strain-, or resistance-associated sequences, which is highly therapeutically relevant information; however, the presence of a resistance gene does not necessarily prove phenotypic resistance.³⁹ Assays intended to support susceptibility-related decisions must therefore measure the bacterial response to antibiotic exposure more directly, for example, by assessing changes in viability, metabolism, impedance, or redox activity after a brief antibiotic exposure. Once the target and intended clinical role are clear, the next design question is how to build a recognition interface capturing the target reliably at the electrode surface.⁴⁰

2.1.2 Biorecognition strategies. Selectivity at the electrode surface is usually achieved through antibodies, aptamers, peptides, lectins, bacteriophages, or nucleic acid probes, as shown in Fig. 1C.^{41,42} Antibody-based electrochemical biosensors remain widely used because they offer established affinity toward bacterial surface structures or secreted toxins. Their use in POC settings, however, is not without drawbacks. Orientation at the surface can be difficult to control, proteins may denature during storage or immobilisation, and cross-reactivity can reduce analytical specificity.⁴³ Aptamers are often presented as an attractive alternative because they are produced synthetically, are straightforward to chemically modify, and can be integrated well into electrochemical signal designs.⁴⁴ They are especially useful in redox-labelled formats. Even so, their performance in complex clinical matrices is often less impressive than in buffered model systems, which remains an important practical limitation.⁴⁵ Peptides and lectins provide additional recognition options, particularly in settings where broader-spectrum capture, lower cost, or greater robustness is preferred over absolute specificity.^{46,47} Bacteriophages and phage-derived proteins are also highly interesting for whole-cell sensing, since they bring biologically evolved recognition and may preferentially bind viable bacteria.⁴⁸ For molecular assays, nucleic acid probes still dominate, as they allow selective targeting of pathogen-specific genes and resistance determinants.⁴⁹ Across all of these strategies, the central

translational question is not simply how strongly the receptor binds its target, but whether the recognition layer remains stable, selective, and manufacturable under realistic POC storage and operating conditions.⁵⁰ Thus, receptor performance is inseparable from the electrode material, because surface chemistry governs loading, orientation, electron transfer, antifouling behaviour, and stability.

2.1.3 Electrode and interfacial materials. Recent advances in nanomaterials have expanded the range of available materials that can support immobilisation of the biorecognition element while maintaining sufficient conductivity of the electrode transducing layer.⁵¹ In this context, graphene-based materials, MXenes,^{52,53} carbon nanotubes,⁵⁴ metallic nanoparticles,^{55,56} metal oxides,^{57,58} conducting polymers,^{59,60} and hybrid composites⁶¹ (Fig. 1D) have been explored to improve receptor binding, electron transfer, catalytic amplification, antifouling behaviour, or mechanical robustness. Graphene and its derivatives are particularly attractive because they combine a high accessible surface area, favourable electronic properties, and chemically tunable surfaces.^{62–64} Controlled functionalization further expands this material portfolio by enabling dual-functionalized, Janus-type, or doped derivatives such as nitrogen-containing analogues.^{65–68} Carboxyl- and ethynyl-functionalized derivatives provide a suitable platform for bioconjugation strategies, including carbodiimide coupling and click chemistry, which can be used to immobilise biorelevant systems such as antibodies, aptamers, or nucleic acid probes.^{69,70} However, graphene-based materials should be viewed as one option rather than a default solution. Printed carbon inks, carbon black, AuNP-modified electrodes, PEDOT:PSS or polypyrrole layers, magnetic-particle-assisted formats, and MXene-containing composites^{71–73} may be equally or more appropriate depending on the target organism, sample matrix, readout mode, and intended point-of-care workflow. From a translational perspective, the decisive question is therefore not whether a platform uses a particular nanomaterial, but whether the electrode material supports a stable, reproducible, matrix-tolerant, and manufacturable sensing interface. Actual performance remains largely dependent on surface chemistry, antifouling design, storage stability, and batch-to-batch reproducibility, variables that deserve more attention than they often receive in proof-of-concept studies.^{74,75}

2.1.4 Signal generation and readout modes. Electrochemical bacterial sensors rely on several main readout modes, most notably voltammetric or amperometric, potentiometric, and impedimetric approaches. Voltammetric and amperometric methods are widely used because they provide sensitive and quantitative current-based signals while relying on relatively simple electronics.³⁷ These formats are well-suited to redox mediators, enzyme labels, and catalytic amplification strategies, and they can be implemented readily in disposable sensor platforms.⁷⁶ Potentiometric sensing is encountered less often, but it can be useful when bacterial recognition leads to measurable changes in potential through ion flux, pH shifts, or membrane-associated phenomena. Its low power demand is



attractive, although in practice, issues such as signal drift and reference-electrode instability can limit robustness.^{37,77} Impedimetric sensing is particularly appealing for POC use because it often allows label-free detection.⁷⁸ When bacteria or biomolecules bind to the electrode surface, they alter the interfacial electrical properties, making it possible to monitor the event directly without additional reporters.^{79,80} The drawback is that impedance signals are also highly sensitive to matrix effects and non-specific fouling.⁸¹ More generally, the choice between label-free and labelled sensing reflects a familiar trade-off; label-free systems simplify the workflow, whereas labelled or catalytic approaches provide stronger and more robust signals, albeit at the cost of greater assay complexity.

2.1.5 Electrode platforms and manufacturing. Electrode fabrication becomes a critical consideration in translating electrochemical biosensors from laboratory demonstrations to manufacturable devices (Fig. 1E). Screen-printed electrodes (SPEs) constitute the most practical platform for POC electrochemical sensing because they are inexpensive, scalable, and readily adapted into disposable integrated devices.⁸² SPEs can be produced on ceramic, plastic, or flexible substrates and easily configured as complete three-electrode systems.⁸³ Inkjet-printed electrodes offer finer digital patterning and more efficient use of functional materials, which is particularly attractive when working with advanced inks, including expensive nanomaterials and biocomponents.^{84–86} They are well-suited to form-factor optimisation, rapid prototyping and customised device architectures.⁸⁷ Flexible and wearable formats are gaining more attention, especially for wound-related sensing, but they bring a different set of requirements, including mechanical robustness, resistance to moisture, and stable performance under repeated deformation.^{88,89}

2.1.6 Representative case studies and benchmarking. The electrochemical sensing literature contains many encouraging examples of bacterial detection, but direct comparison between studies is often difficult. Differences in target choice, sample matrix, preparation workflow, and reporting standards make it hard to judge which approaches are genuinely closer to clinical use.³⁷ For that reason, benchmarking should focus less on record-low limits of detection alone and more on metrics that are meaningful in practice.⁴¹ Urine is one of the more promising matrices for electrochemical sensing, partly because it is relatively accessible and urinary tract infections remain a major use case for rapid triage.^{90,91} Blood is far more demanding: bacterial concentrations are often low, while matrix interference is high, so direct detection usually requires enrichment, lysis, or molecular amplification.⁹² Swab and wound samples are also attractive for planar or flexible electrochemical devices, but their heterogeneity and often polymicrobial nature make interpretation more complicated.^{93,94} Across all of these sample types, the most informative studies are those that report time-to-result from sample to answer, validate performance in real matrices, and compare results directly against culture or PCR. Reproducibility, false-positive behaviour, storage stability, and workflow simplicity are equally important. In the end, electrochemical POC diagnostics are most convincing when they

combine reliable recognition chemistry, manufacturable electrode platforms, matrix-tolerant interfaces, and benchmarking that is clinically realistic rather than merely analytically impressive.^{95,96}


2.2 Raman-based optical approaches

2.2.1 Realistic role in POC bacterial workflows. Optical methods offer a complementary approach to bacterial detection, providing spatially resolved, chemically informative signals without direct electrical contact with the sample. Raman spectroscopy, especially SERS, combines molecular specificity with the potential for miniaturised, portable instrumentation (Fig. 2A). Their main value is not simply rapid readout, but the ability to extract biochemical fingerprints, discriminate bacterial phenotypes, and support multiplexed analysis when combined with suitable substrates and data-processing workflows. Moreover, optical signals are spatially resolvable. This spatial resolution enables integrated “lab-on-a-chip” architectures that simultaneously map morphology and chemical concentration, providing diagnostic depth previously reserved for centralised pathology labs.⁹⁷ Among optical methods, Raman spectroscopy is a rapid technique that often eliminates the need for complex, time-consuming laboratory analyses.⁹⁸ It also provides precise measurements, enabling ultrasensitive multiplexed detections required for many clinical applications.^{99,100} Additionally, Raman spectroscopy can be simply combined with other optical and non-optical methods to enable accessible sample handling and processing of clinical patient samples without the need for complex wiring, often using disposable optics or measuring kits.¹⁰¹ Furthermore, its integration with microscopy enables high spatial resolution (<1 μm) and hyperspectral imaging, providing one more level of information.

2.2.2 Fingerprints, strengths, and limitations. The specificity of a bacterial Raman spectrum is a function of stoichiometric proportions of biomolecules, including proteins, lipids, carbohydrates, and nucleic acids.¹⁰² The spectral region of interest can be divided into two discrete regions: the high-wavenumber region and the fingerprint region, separated by a silent region (Fig. 2B). The high-wavenumber region (2800–3800 cm^{-1}) is dominated by $-\text{CH}_x$, $-\text{N}-\text{H}_x$, and $-\text{O}-\text{H}$ stretching vibrations. Spectral data from this region offer remarkable precision in discriminating between cell wall types (Gram-positive vs. Gram-negative), with an accuracy approaching 100%.¹⁰³ On the other hand, the fingerprint region (400–1800 cm^{-1}) is where key diagnostic markers reside. The phenylalanine ring-breathing mode at 1003 cm^{-1} serves as a stable internal reference for protein content. This band, together with bands characteristic of dipicolinate and amino acids, was used to study bacterial endospores.¹⁰⁴ The DNA spectral band at 785 cm^{-1} (phosphorus diester O–P–O bonds) tracks, for example, nucleic acid fragmentation during antibiotic exposure.¹⁰⁵ The amide I band (1650–1670 cm^{-1}) provides essential insight into secondary protein structures and overall cellular viability.¹⁰⁶ The entire Raman spectrum can be used to discriminate between *Escherichia coli* strains using a combination of Raman microscopy and chemometric methods.¹⁰⁷




A. Optical Diagnostic Workflow at the Point of Care.




B. Raman & SERS Technologies for a Detection of Pathogens

B1. Raman spectroscopy of bacteria



- Resolution of bacterial strains
- Robustness / speed
- Utilisation of modern smartphones
- Simple sample preparation

B2. SERS enhancement mechanism



C. Portable Raman & Data Analysis in Action

C1. Selection of portable Raman systems



C2. Examples of a field sampling scenarios



C3. Workflow for a detection of pathogens



C4. Integration of spectral data into smartphones/handhelds



C5. Comparison of spectral data of pathogens




Fig. 2 Schematic overview of Raman/SERS-based point-of-care strategies for bacterial pathogen detection. (A) A simplified workflow is shown from sample collection and preprocessing to spectral acquisition and data analysis. (B) The figure highlights the basis of bacterial Raman detection and the signal enhancement achieved by SERS, together with representative substrate and sampling formats. (C) Portable Raman instrumentation, field and near-patient sampling scenarios, smartphone-assisted analysis, and spectral discrimination of representative pathogens illustrate the potential of these approaches for rapid decentralised testing. The figure was created using BioRender.com and Adobe Illustrator.

Similar results were also obtained for nine clinical species of bacteria isolated from clinical samples,¹⁰⁸ or 11 bacterial mutants of *E. coli* MDS42.¹⁰⁹ Despite its non-destructive nature and single-cell resolution, spontaneous Raman scattering is inefficient: only approximately 1 in 10⁷ photons undergo the Raman effect. This necessitates long acquisition times and sophisticated baseline correction to mitigate the high autofluorescence background inherent in biological complex matrices. These limitations explain why SERS has become the more practically attractive Raman-based strategy for bacterial diagnostics in complex or low-biomass samples.

2.2.3 SERS as the practical Raman route for real samples.

SERS addresses the intrinsic weakness of the Raman effect by utilising the localised surface plasmon resonance (LSPR) of noble metal nanostructures (Fig. 2B). When an analyte is adsorbed onto or placed near the formed plasmonic “hot-spots”, the Raman signal is amplified by factors of up to 10¹⁰. This allows the detection of single bacterial cells or viral particles directly from complex matrices, including whole blood,¹¹⁰ tissue samples,¹¹¹ and even wastewater.¹¹² The evolution of SERS substrates has shifted from simple stochastic colloidal systems to complex 3D architectures, including structured surfaces fabricated *via* electron-beam lithography or glancing-angle deposition.^{113–115} These architected models provide uniform hotspot distribution and precisely regulated nanogaps. Moreover, 3D scaffolds, such as metallic foams and vertically aligned nanorod arrays, present other interesting approaches.^{116–118} These hierarchical structures extend hotspot

density into the third dimension, thereby prolonging photon lifetimes through multiple scattering events. When integrated into flexible polymer or cellulose substrates, these “swab-and-sense” platforms enable conformal, *in situ* detection of bacteria on complex, non-planar surfaces such as wound beds or food products.¹¹⁹ Furthermore, by employing surface cell-imprinting, these matrices can be engineered with bio-specific cavities tailored for target bacterial strains, effectively mitigating matrix interference and enhancing selectivity.¹²⁰ To move beyond qualitative detection, advanced calibration models are required. Such models are usually based on a set of Internal Standards (IS).¹²¹ Incorporating SERS probability models allows for signal normalisation and enables qualitative analysis of a single molecular target over a 9-order-of-magnitude concentration range.¹²² The next applicable calibration strategy to achieve quantitative information is Digital SERS, which is primarily used at ultra-low concentrations where intensity-based metrics usually fail.^{123,124} Digital SERS treats detection as a binary “on/off” event across a pixelated substrate, deriving concentration from the fraction of active hotspots *via* a Poisson distribution. The last example of a quantitative approach is an application of a SERS performance factor (SPF).¹²⁵ This metric reflects intrinsic sensitivity by comparing the slopes of concentration-dependent SERS and standard Raman curves.

2.2.4 Label-free vs. tag-based SERS. The choice between label-free and tag-based SERS usually involves a fundamental trade-off between speed, simplicity, and specificity. Label-free profiling measures the intrinsic Raman signature of the



pathogen. It is rapid and low-cost, providing a direct readout of the cellular physiological state (*e.g.*, metabolic activity or antibiotic stress).¹²⁶ However, because the SERS enhancement effect decays rapidly with distance from the surface (r^{-6}), the signal is dominated by the outermost cell wall components. This similarity across strains usually requires a powerful machine learning or more advanced AI models to extract these small discriminative features.^{127,128} Tag-Based SERS, including immuno and aptamer based techniques utilizes plasmonic cores functionalized with a Raman reporter and a recognition element, such as Immuno-SERS, which employs monoclonal antibodies in sandwich architectures to isolate targets from complex backgrounds,^{129,130} offering unparalleled specificity, or aptamer-SERS, which uses synthetic DNA/RNA “chemical antibodies”.¹³¹ Aptamers are considerably more stable and cheaper in production than antibodies, might be stored at room temperature, can be fully engineered to undergo conformational changes upon target binding. This change initiates a displacement of the Raman reporter, which results in a highly quantifiable “signal-on” or “signal-off” transition. While tag-based methods excel in multiplexing (by using reporters with distinct, non-overlapping peaks), they increase assay complexity and also preparation time. From a clinical perspective, label-free methods are primarily suitable for fingerprint identification and phenotype distinction,¹³² whereas labelled SERS is better suited for targeted pathogen identification or multiplex detection of selected markers.^{133,134} Resistance information can be approached indirectly through resistance-associated spectral phenotypes, reporter panels, or antibiotic-induced changes in bacterial metabolism and viability, but these outputs should not be treated as equivalent to phenotypic AST unless they are benchmarked against susceptibility testing in clinically representative samples.¹³⁵

2.2.5 Portable Raman and data analysis. The synergy between handheld hardware and chemometric processing drives the transition of Raman diagnostics to the field (Fig. 2C). Handheld spectrometers, while rugged, often suffer from lower laser powers and limited integration times. Sophisticated data-driven tools are usually required to compensate for these hardware limitations, including machine learning methods such as support vector machines (SVMs) and Random Forests, as well as convolutional neural networks (CNNs). Recent reviews highlight the integration of SERS with artificial intelligence and microfluidic lab-on-chip formats as a practical route toward point-of-care diagnostics. AI can help interpret complex spectra, while microfluidics can standardise enrichment, mixing, washing, and sample handling in compact devices.¹³⁶ Portable Raman systems are, at the same time, getting smaller and smarter, approaching pocket-sized form factors at decreasing cost.^{137,138} To address the limitations of the spectral intensity and interferences, Botto *et al.* utilised unsupervised Partial Least Squares Discriminant Analysis (PLS-DA) regression coefficients, combined with accuracy score maximisation in stratified k-fold cross-validation, to analyse 17 isolates belonging to the genera *Pseudomonas* spp., *Stenotrophomonas* spp., and *Erwinia* spp.¹³⁹ Similarly, to utilize a portable Raman system, Thomsen *et al.* developed a procedure

allowing for the analysis of antibiotic-resistant and susceptible phenotypes by utilizing a spectral transformer (ST) machine learning (ML) model.¹⁴⁰ Interestingly, Zhao developed a working portable Raman system and in a combination with a frequently available paper-based SERS substrate (originally aimed for a tobacco packaging) and used them to detect *Staphylococcus aureus* and *Shigella flexneri* in a pork meat.¹⁴¹ These examples demonstrate a trend in Raman detection of pathogens towards simple, field-manageable sample preparation, enabled by advanced signal processing and miniaturisation of hardware. Machine-learning-assisted Raman and SERS can help to reveal weak multivariate spectral differences that are difficult to assign peak by peak, especially in label-free bacterial profiling and phenotype discrimination.^{127,132,142} Its diagnostic value, however, depends on diverse training data and validation beyond a single instrument, strain set, or controlled matrix.¹⁴³ Future studies should therefore include external test sets, cross-site testing, calibration transfer between instruments and substrate batches, standardised preprocessing, and transparent uncertainty reporting.¹⁴³ Machine learning is best viewed as an interpretive layer, not a substitute for robust substrates, chemistry, and clinically realistic benchmarking.^{136,144} At present, Raman/SERS is most credible in controlled workflows using cultured or enriched material, defined matrices, or standardised preprocessing. Direct POC analysis of low-burden, heterogeneous clinical specimens remains largely preclinical because matrix interference, bacterial recovery, substrate reproducibility, and cross-site transfer still limit routine use.¹⁴⁵

2.2.6 Complementary optical modalities in the context of Raman-based workflows. Although this review focuses on Raman, several complementary optical methods help define the broader landscape of POC testing. Colourimetric assays, including smartphone-assisted platforms, are attractive for simple screening because they provide a visually accessible readout, but they generally offer lower molecular specificity.^{146,147} Fluorescence methods can be sensitive and spatially informative, although they usually require labelled reagents and careful background control.¹⁴⁸ SPR/LSPR platforms provide label-free monitoring of binding events, but their performance in clinical matrices can be limited by nonspecific adsorption and refractive index changes.¹⁴⁹ DHM and related phase-imaging approaches are better suited for monitoring bacterial growth, segmentation, or microfluidic culture behaviour than for direct molecular identification.¹⁵⁰ These methods are therefore mentioned here as complementary technologies, rather than as a second major focus of the review.

2.3 Comparative perspective: electrochemical vs. Raman POC strategies

Electrochemical and Raman platforms both offer routes toward rapid POC bacterial diagnostics, but they address this goal through different analytical strengths, practical constraints, and clinical use cases. Electrochemical systems align well with low-cost, near-patient testing because the hardware is simple, low-power, and readily adaptable to disposable sensors. The



main weakness is connected with real samples, where fouling, non-specific adsorption, interference, and low bacterial counts can make performance less reliable, particularly in blood samples.^{151,152} Raman approaches provide richer molecular information useful for label-free profiling, multiplexed analysis, and in some cases phenotypic discrimination,¹⁵³ but this added information depends on reproducible substrates, signal normalisation, spectral preprocessing, and robust data analysis.^{154,155} For this reason, electrochemical platforms are currently most plausible for accessible matrices such as urine, wound exudate, or swabs, whereas Raman is strongest when molecular specificity or multiplexed readout is worth the added complexity. In bloodstream infections, both approaches still face major preanalytical barriers and should be viewed as triage or complementary tools to culture and phenotypic AST, not as replacements for them.¹⁵⁶ The intended use, therefore, has to be defined by the matrix and the decision context. Urine and wound or swab samples are the nearest-term settings for low-cost electrochemical screening or swab-compatible SERS; respiratory samples require stronger preprocessing because mucus, viscosity, and host background complicate both readouts.^{157–159} In sepsis or bloodstream infection, especially in ICU or emergency-department workflows, greater device complexity may be acceptable if the result is rapid and actionable,¹⁶⁰ whereas field settings favour simple, robust, reagent-stable formats. Clinical actionability also requires more than binary pathogen detection.^{156,160} For stewardship, the most useful platforms should identify resistance determinants, monitor antibiotic-induced changes in viability or metabolism,³⁹ or provide early susceptibility-related information. Electrochemical assays may do this through resistance-gene detection or viability/metabolite readouts, while Raman can capture antibiotic-response fingerprints or resistance-associated spectral phenotypes.¹⁶¹ Both methods should therefore be benchmarked by their ability to support escalation, de-escalation, or earlier susceptibility-informed treatment, not by detection sensitivity alone. To make the practical comparison between the discussed technologies more explicit, Table 1 summarises their main analytical targets, typical performance ranges, operational strengths, and translational limitations. A related cross-cutting issue is amplification, which influences both analytical sensitivity and practical point-of-care feasibility. In nucleic-acid assays, PCR and qPCR remain highly sensitive, whereas isothermal methods such as LAMP or RPA are better suited to decentralised testing because they reduce instrumentation requirements.¹⁶² Electrochemical platforms may further improve sensitivity through enzyme labels, redox cycling, catalytic nanoparticles, nanostructured electrodes,⁷⁶ or CRISPR-coupled reporter cleavage,³³ while Raman approaches rely on plasmonic signal enhancement, reporter tags, internal standards, and digital SERS formats.^{131,153,154} These strategies can substantially lower detection limits, but they may also increase assay time, reagent complexity, washing steps, contamination risk, and susceptibility to matrix inhibition.¹⁶³

Table 1 Representative benchmarking parameters for electrochemical and Raman/SERS point-of-care bacterial diagnostics^a

Technology	Analytical target	Representative LOD range	Typical time-to-result	Common sample types	Clinical role/key strengths	Main limitations
Electrochemical whole-cell sensors	Intact bacteria captured by an antibody, aptamer, peptide, lectin, or phage	$\sim 10^1\text{--}10^4$ CFU mL ⁻¹	$\sim 10\text{--}60$ min, longer if enrichment is needed	Urine, wound swabs, food samples, diluted blood	Quantitative readout, low-cost electronics, disposable electrodes	Matrix fouling, non-specific adsorption, and limited validation in true clinical cohorts
Electrochemical nucleic-acid sensors	Species or resistance genes, often after amplification	$\sim 10\text{--}10^3$ genome copies or equivalent	$\sim 30\text{--}120$ min	Blood after extraction, urine, and respiratory samples	High specificity, resistance-marker detection possible	Extraction/amplification increases complexity and contamination risk
Electrochemical secreted-biomarker sensors	Toxins, enzymes, metabolites, quorum-sensing molecules	Highly analyte-dependent; often nM-pM or equivalent bacterial burden	$\sim 10\text{--}60$ min	Urine, wound fluid, culture supernatants	Can indicate bacterial activity	Biomarker may not map directly to viable pathogen burden
Spontaneous Raman	Intrinsic bacterial biochemical fingerprint	Usually requires isolated/enriched cells or a high bacterial load	Seconds to minutes per spectrum; longer, including preprocessing	Colonies, enriched samples, urine or sputum after processing	Label-free, single-cell information, phenotypic insight	Weak scattering, fluorescence background, demanding data analysis
Label-free SERS	Enhanced intrinsic bacterial surface signals	Single-cell to $\sim 10^2\text{--}10^5$ CFU mL ⁻¹ depending on substrate and matrix	$\sim 10\text{--}60$ min	Urine, swabs, food, environmental samples, selected blood workflows	Rapid spectral fingerprinting, potential strain discrimination	Substrate variability, matrix interference, and limited cross-site validation
Tag-based SERS	Antibody/aptamer-mediated pathogen capture with Raman reporters	Often $\sim 10\text{--}10^5$ CFU mL ⁻¹ under optimized conditions	$\sim 30\text{--}120$ min	Urine, serum/plasma after preparation, swabs, food matrices	High specificity, multiplexing	More assay steps, higher reagent complexity, batch-to-batch reproducibility

^a Ranges are representative rather than absolute; matrix, preprocessing, target organism, assay format, enrichment, and comparator method strongly affect reported performance.



3. Benchmarking, validation, and clinical translation of rapid bacterial diagnostics

The value of a rapid bacterial diagnostic is not determined solely by how quickly it produces a signal, but by how quickly it provides information that is actually useful in a clinical decision-making context relative to the existing microbiology workflow.¹⁶⁴ In practice, the most meaningful benchmark is therefore not an idealised proof-of-concept measured in buffer, but the established sequence of culture, species identification, and antimicrobial susceptibility testing (AST) that still underpins most treatment decisions. A strong, rapid test should be judged by whether it shortens the path to meaningful action, whether that means targeted therapy, de-escalation, triage, or escalation to central laboratory testing, rather than by analytical sensitivity alone.¹⁶⁵

3.1 Laboratory gold standards for bacterial pathogen identification

Conventional culture remains the clinical reference standard because it recovers viable organisms, supports phenotypic identification, and enables downstream AST. Its weakness, of course, is time, as shown in Fig. 3. Enrichment, subculture, and susceptibility testing usually take at least overnight, and often longer, while contamination, low pathogen burden, and prior antibiotic exposure can further reduce recovery or complicate interpretation.¹⁶⁵ MALDI-TOF MS has substantially changed routine laboratory identification by allowing rapid and accurate species-level assignment once isolated colonies, positive blood-culture material, or otherwise enriched biomass are available.^{166,167} Even so, direct-from-sample analysis remains much more difficult, especially in blood and similarly complex matrices, where low bacterial abundance, host background, and mixed infections can all interfere with performance.¹⁶⁸ NAATs, particularly PCR and qPCR, are now

the most mature rapid alternatives for targeted pathogen detection, while isothermal amplification methods provide an important bridge toward simpler, lower-instrumentation testing closer to the patient.^{169,170} Sequencing and metagenomic approaches offer the broadest diagnostic scope and can reveal polymicrobial or uncultivable organisms,¹⁷¹ but in most cases, they still depend on specialised sample preparation, strict contamination control, and substantial bioinformatic support. For that reason, they currently fit more naturally into advanced laboratory workflows than into routine POC use.¹⁷² AST cuts across all of these approaches. Genotypic assays can rapidly identify known resistance determinants, but phenotypic AST remains the clinically decisive endpoint because it reflects expressed susceptibility rather than genetic potential alone, even if that usually comes at the cost of time.¹⁵⁶

3.2 Clinical benchmarking

Clinical benchmarking merits at least the same level of attention as analytical sensitivity.^{173,174} To make such benchmarking transparent, rapid bacterial diagnostic platforms should be compared using a common set of quantitative and operational descriptors, including the analytical target, the clinically relevant sample matrix, the reported limit of detection, the total sample-to-answer time, the sample preparation requirements, the comparator method, and the level of validation (Table 1).¹⁷⁵ For bacterial diagnostics, reported limits of detection are difficult to compare because they may be expressed as CFU mL⁻¹, genome copies, colony equivalents, or molecular biomarker concentrations, and because values obtained in buffer often differ from those obtained in urine, blood, wound exudate, sputum, or food-related matrices.^{176,177} Therefore, representative performance ranges are more informative than isolated best-case values. Electrochemical bacterial sensors commonly report 10¹–10⁴ CFU mL⁻¹ in simplified or moderately complex matrices, with assay times ranging from minutes to 1–2 h, depending on enrichment, labelling, or




Fig. 3 Overview of major diagnostic modules for bacterial pathogen identification across different time scales. NAATs provide rapid targeted detection within minutes to hours, MALDI-TOF MS enables fast species identification from cultured material, culture with phenotypic AST remains the clinically decisive but slower reference method, and sequencing/metagenomics offers the broadest scope at the cost of greater workflow complexity and longer turnaround. The timeline highlights the trade-off between speed, diagnostic breadth, and laboratory requirements. The figure was created using BioRender.com and Adobe Illustrator.



amplification.^{18,177} SERS can reach single-cell to 10^2 – 10^5 CFU mL⁻¹ under favourable conditions, but performance depends on substrate reproducibility, capture efficiency, acquisition time, and chemometric robustness.^{178,179} These values should be read as indicative and matrix-dependent. Reported sensitivity, accuracy, or classification performance is meaningful only when the experimental context is clear.^{180–182} Many electrochemical and Raman studies demonstrate excellent analytical performance in buffer, spiked samples, enriched cultures, or carefully controlled matrices, but these results do not necessarily translate directly to routine clinical specimens.^{152,183–185} Real-world applicability depends on additional factors, including bacterial recovery from the sample, matrix-induced fouling or optical background, operator-dependent preprocessing, assay failure rate, and validation against clinically accepted comparator methods. Therefore, high-performance claims are most meaningful when supported by testing in relevant biological matrices, transparent sample-to-answer workflows, and independent validation using clinically representative samples.^{174,175} The influential studies rely on prospective cohorts, or at least carefully curated retrospective ones, together with blinded analysis, transparent comparator definitions, and prevalence-aware reporting of sensitivity, specificity, positive predictive value, and negative predictive value.¹⁸⁰ Just as importantly, they report operational characteristics such as time-to-result from sample to answer, failure rate, number of operator steps, and any measurable effect on treatment escalation, de-escalation, or triage.¹⁸¹ In many cases, these parameters say more about real clinical usefulness than an exceptionally low detection limit ever could. Future studies should report, at a minimum, the intended clinical use, the sample matrix and volume, the matrix-specific LOD, the total sample-to-answer time, the preprocessing and user steps, the comparator method, the assay failure rate, and the validation level.^{24,176} They should also state whether the assay is intended for pathogen detection, resistance marker identification, or susceptibility-related decision support.¹³⁵ This makes it easier to separate proof-of-concept sensitivity from clinical actionability, and to judge detection limits alongside workflow complexity, matrix tolerance, and decision relevance.

For spectral and optical platforms that rely on chemometrics or machine learning, internal cross-validation alone is insufficient.¹⁸² More convincing validation should include independent external test sets, cross-site generalisation, calibration transfer between instruments or substrate batches, uncertainty reporting, and transparent reporting frameworks such as TRIPOD + AI when predictive models are involved.¹⁷⁴ These details may seem technical, but they are exactly what determine whether a platform is robust enough to move beyond a single well-controlled study.

3.3 Matrix effects and mitigation strategies

Biological matrices shape rapid bacterial diagnostics at every step, from pathogen recovery and target accessibility to signal generation and interpretation.¹⁸⁶ This is particularly important for point-of-care workflows, where extensive purification or specialist troubleshooting is usually unrealistic. Matrix interference is one reason why assays that perform well in buffer often lose reliability in clinical samples.^{187,188} The challenge is

matrix-specific. Urine is accessible and often contains clinically relevant bacterial loads, but variation in pH, ionic strength, salts, proteins, urea, and debris can affect receptor binding, electrode response, nanoparticle stability, and optical background.¹⁸⁹ Blood is much more complicated.^{151,190} Bacterial counts may be very low, while blood cells, plasma proteins, haemoglobin, coagulation factors, and host DNA create strong interference. Respiratory samples, sputum, wound exudates, and swabs add viscosity, mucus, tissue debris, polymicrobial content, and uneven bacterial distribution.¹⁹¹ Limit of detection should therefore be reported in the matrix that matches the intended clinical use, not only in buffer or simplified media. In electrochemical biosensors, matrix effects commonly appear as electrode fouling, non-specific adsorption, altered charge-transfer resistance, pH or ionic-strength shifts, inhibition of redox or enzymatic reactions, and reduced access to the recognition layer.¹⁵² This is particularly problematic for label-free impedimetric systems, where non-specific interfacial changes can mimic binding.¹⁸³ Mitigation usually relies on antifouling and blocking layers, PEG-based¹⁹² or zwitterionic interfaces,^{193,194} dilution or washing, selective capture, magnetic preconcentration, filtration, or microfluidic separation.^{195–197} In Raman, matrix effects appear as autofluorescence, broad backgrounds, optical scattering, competitive adsorption, nanoparticle aggregation, and inconsistent contact between bacteria and hotspots.^{163,184} Proteins, salts, mucus, or host debris can mask label-free SERS signals, while tag-based assays may suffer from reduced capture efficiency or non-specific reporter binding.¹⁸⁵ Practical mitigation includes clarification, enrichment, immunomagnetic or aptamer-mediated capture, controlled washing, filtration, microfluidic concentration, more robust substrate functionalization, internal Raman standards, standardised preprocessing, and external validation of chemometric models.^{198–200} Matrix assessment should therefore include bacterial recovery, matrix-specific LOD, background signal, false-positive behaviour, sample volume, preprocessing, failure rate, and workflow compatibility.¹⁷⁶

3.4 Manufacturing, regulation, and real-world deployment

Ultimately, translation depends on whether a sensing concept can be turned into a reproducible sample-to-answer product.²⁰¹ For bacterial diagnostics, that usually means integrating lysis, enrichment, fluid handling, biosafety, and contamination control into a disposable cartridge or another tightly controlled workflow.²⁰² A platform that performs well only after extensive manual preprocessing may still have value in a central laboratory, but it is unlikely to function as a true POC device. Manufacturing, therefore, needs to be considered early rather than treated as a late-stage engineering detail.²⁰³ Electrochemical systems must demonstrate scalable electrode fabrication, stable interfacial chemistry, and batch-level quality control,²⁰⁴ whereas Raman/SERS and related optical methods must address substrate reproducibility, signal standardisation, and shelf-life without relying on cold-chain storage.²⁰⁵ Usability matters just as much. Low operator burden, effective error-proofing, minimal maintenance, and secure integration into



laboratory information systems or electronic health records often determine whether a device is actually suitable for use in the emergency department, ICU, outpatient clinic, or field setting.^{206,207} Regulatory evidence should likewise be built in from the start rather than assembled afterwards. In Europe, the IVDR requires evidence for scientific validity, analytical performance, and clinical performance in relation to the intended use.²⁰⁸ In parallel, manufacturing should be developed within an appropriate medical-device quality system, typically ISO 13485,²⁰⁹ with traceability, risk management, and post-market surveillance embedded throughout the product lifecycle. Overall, the rapid bacterial diagnostics most likely to achieve clinical adoption will not be those that simply report the lowest limits of detection, but those that combine honest benchmarking in real samples with reproducible manufacturing, usable workflows, and a credible route to implementation.

4. Future perspectives

Rapid bacterial diagnostics will have the greatest clinical impact not when they are merely fast in analytical terms, but when they provide information early enough to change treatment decisions, support triage, and improve antimicrobial stewardship. In that sense, the most useful point-of-care and near-patient tests are unlikely to replace central microbiology laboratories, but rather to complement them by delivering earlier actionable results while culture, MALDI-TOF MS, molecular assays, sequencing, and phenotypic AST continue to provide confirmation, broader characterisation, and definitive susceptibility data.

Among emerging decentralised approaches, electrochemical biosensors remain especially attractive because they align naturally with the practical requirements of point-of-care use, including portability, low cost, straightforward electronics, and compatibility with disposable formats. Their future success, however, will depend less on achieving ever lower detection limits than on solving more translational problems, such as matrix tolerance, antifouling behaviour, storage stability, reproducible surface functionalization, and scalable manufacturing. Raman-based strategies, particularly SERS, offer complementary strengths through their rich molecular specificity and potential for multiplexed detection in complex samples. Also in this domain, the key challenge is not proof-of-concept sensitivity alone, but the development of robust substrates, transferable calibration strategies, and data-analysis workflows that remain reliable across instruments, batches, and clinical settings. Across both fields, future progress is also likely to be accelerated by artificial intelligence (AI), particularly through improved data analysis, predictive modelling, and agentic systems that support interpretation and decision-making.^{210,211} At the same time, AI may help lower practical barriers to adoption by enabling more intuitive handling of complex outputs, for example, through natural language processing-based interfaces that make advanced diagnostic systems easier to use in routine clinical practice.¹⁴⁵

Across both fields, the next phase of progress will depend on more rigorous and clinically honest benchmarking. Studies

that use real patient samples, report time-to-result from sample to answer, compare performance directly with standard-of-care methods, and address usability, failure rates, and workflow integration will be far more valuable than demonstrations in simplified laboratory media. More broadly, the technologies most likely to succeed will be those designed around specific clinical decisions rather than around abstract analytical performance. The future of rapid bacterial diagnostics, therefore, lies not in a single universal platform but in well-validated, application-focused systems that combine robust sensing with manufacturability, regulatory readiness, and clear clinical utility.

5. Conclusions

This review highlights the clinical rationale for rapid bacterial diagnostics and examines two technically mature routes toward decentralised implementation: electrochemical biosensing and Raman-based detection, particularly SERS. Both approaches offer genuine advantages for POC use. Electrochemical platforms excel in their portability, low cost, and quantitative signal output. Raman/SERS offer their rich molecular specificity and multiplexing capability. However, neither field has yet produced devices that consistently meet the standards required for clinical adoption, namely robust performance in complex biological matrices, validated clinical utility, and regulatory approval. Progress will require a shift in focus from proof-of-concept demonstrations toward honest, clinically grounded benchmarking. The most impactful future diagnostics will not be those with the lowest detection limits in buffered model systems, but those that deliver reliable, actionable results in real patient samples, integrate seamlessly into existing clinical workflows, and provide information specific enough to guide treatment. In that context, electrochemical biosensors and Raman-based platforms are best understood not as replacements for central laboratory diagnostics, but as complementary tools that can provide earlier, more targeted information at the point of clinical decision-making. Their success will ultimately be measured not by analytical performance alone, but by their ability to improve patient outcomes and also support the responsible use of antibiotics.

Author contributions

P. J. contributed to conceptualisation, writing of the original draft, review and editing, visualisation, and funding acquisition. V. R., M. K., O. Z., D. P., and M. O. contributed to the writing of the original draft and to the review of the manuscript. M. O. contributed to conceptualisation, manuscript review and editing, and funding acquisition.

Conflicts of interest

M. O. discloses his share in InSiliBio (France), focused on biosimulations, and ATOMIVER (Czechia), focused on supercapacitor electrode materials, companies.



Data availability

No new primary research data were generated or included as part of this review article.

Acknowledgements

The work was supported from ERDF/ESF project TECHSCALE (No. CZ.02.01.01/00/22_008/0004587), project SALVAGE (no. CZ.02.01.01/00/22_008/0004644), and ERDF/ESF project 'Interdisciplinary Approaches to the Prevention and Diagnosis of Viral Diseases' (CZ.02.01.01/00/23_021/0008856). The authors also acknowledge financial support of the European Union under the REFRESH – Research Excellence For REgion Sustainability and High-tech Industries (CZ.10.03.01/00/22_003/0000048) via the Operational Programme Just Transition, project National Institute for Cancer Research (Programme EXCELES, ID Project No. LX22NPO5102)—Funded by the European Union—Next Generation EU, and project IGA_LF_2026_021. During the preparation of this manuscript, the authors used ChatGPT 5.5 to assist with language editing and improving readability. All content was subsequently reviewed and revised by the authors, who take full responsibility for the accuracy and integrity of the work.

References

- J. A. Homer, R. M. Johnson, R. A. Koelln, A. D. Moorhouse and J. E. Moses, *Nat. Rev. Bioeng.*, 2024, **3**, 213–229.
- M. Singer, C. S. Deutschman, C. W. Seymour, M. Shankar-Hari, D. Annane, M. Bauer, R. Bellomo, G. R. Bernard, J.-D. Chiche, C. M. Coopersmith, R. S. Hotchkiss, M. M. Levy, J. C. Marshall, G. S. Martin, S. M. Opal, G. D. Rubenfeld, T. Van Der Poll, J.-L. Vincent and D. C. Angus, *JAMA*, 2016, **315**, 801.
- K. E. Rudd, S. C. Johnson, K. M. Agesa, K. A. Shackelford, D. Tsoi, D. R. Kievlan, D. V. Colombara, K. S. Ikuta, N. Kissoon, S. Finfer, C. Fleischmann-Struzek, F. R. Machado, K. K. Reinhart, K. Rowan, C. W. Seymour, R. S. Watson, T. E. West, F. Marinho, S. I. Hay, R. Lozano, A. D. Lopez, D. C. Angus, C. J. L. Murray and M. Naghavi, *Lancet*, 2020, **395**, 200–211.
- M. Van Den Berg, F. E. Van Beuningen, J. C. Ter Maaten and H. R. Bouma, *J. Crit. Care*, 2022, **71**, 154096.
- A. Cassini, B. Allegranzi, C. Fleischmann-Struzek, T. Kortz, R. Markwart, H. Saito, et al., *Global Report on the Epidemiology and Burden of Sepsis: Current Evidence, Identifying Gaps and Future Directions*, World Health Organization, Geneva, 1st edn, 2020.
- L. Evans, A. Rhodes, W. Alhazzani, M. Antonelli, C. M. Coopersmith, C. French, F. R. Machado, L. McIntyre, M. Ostermann, H. C. Prescott, C. Schorr, S. Simpson, W. J. Wiersinga, F. Alshamsi, D. C. Angus, Y. Arabi, L. Azevedo, R. Beale, G. Beilman, E. Bellay-Cote, L. Burry, M. Cecconi, J. Centofanti, A. Coz Yataco, J. De Waele, R. P. Dellinger, K. Doi, B. Du, E. Estensoro, R. Ferrer, C. Gomersall, C. Hodgson, M. H. Møller, T. Iwashyna, S. Jacob, R. Kleinpell, M. Klompas, Y. Koh, A. Kumar, A. Kwizera, S. Lobo, H. Masur, S. McLaughlin, S. Mehta, Y. Mehta, M. Mer, M. Nunnally, S. Oczkowski, T. Osborn, E. Papanthanasoglou, A. Perner, M. Puskarich, J. Roberts, W. Schweickert, M. Seckel, J. Sevransky, C. L. Sprung, T. Welte, J. Zimmerman and M. Levy, *Intensive Care Med.*, 2021, **47**, 1181–1247.
- B. Reddy, U. Hassan, C. Seymour, D. C. Angus, T. S. Isbell, K. White, W. Weir, L. Yeh, A. Vincent and R. Bashir, *Nat. Biomed. Eng.*, 2018, **2**, 640–648.
- O. J. Dyar, B. Huttner, J. Schouten and C. Pulcini, *Clin. Microbiol. Infect.*, 2017, **23**, 793–798.
- M. Kolar, M. Htoutou Sedlakova, K. Urbanek, P. Mlynarcik, M. Roderova, K. Hricova, K. Mezerova, P. Kucova, J. Zapletalova, K. Fiserova and P. Kurfurst, *Antibiotics*, 2021, **10**, 93.
- S. A. Dunbar, M. Benavides and C. Gardner, *Front. Cell. Infect. Microbiol.*, 2025, **15**, 1725920.
- R. Banerjee and R. Patel, *JAC-Antimicrob. Resist.*, 2022, **5**, dlad018.
- P. K. Dabla and A. Dabas, *World J. Crit. Care Med.*, 2025, **14**, 1–6.
- I. Michael, D. Kim, O. Gulenko, S. Kumar, S. Kumar, J. Clara, D. Y. Ki, J. Park, H. Y. Jeong, T. S. Kim, S. Kwon and Y.-K. Cho, *Nat. Biomed. Eng.*, 2020, **4**, 591–600.
- F. Mazur, A. D. Tjandra, Y. Zhou, Y. Gao and R. Chandrawati, *Nat. Rev. Bioeng.*, 2023, **1**, 180–192.
- G.-R. Han, A. Goncharov, M. Eryilmaz, S. Ye, B. Palanisamy, R. Ghosh, F. Lisi, E. Rogers, D. Guzman, D. Yigci, S. Tasoglu, D. Di Carlo, K. Goda, R. A. McKendry and A. Ozcan, *Nat. Commun.*, 2025, **16**, 3165.
- S. Singh, A. Numan and S. Cinti, *Anal. Chem.*, 2022, **94**, 26–40.
- J.-W. Tang, X.-R. Wen, Y.-W. Liao and L. Wang, *Nanomedicine*, 2025, **20**, 701–706.
- E. Koçak, C. Ozkul, B. Bozal-Palabiyik, İ. Süslü and B. Uslu, *Electroanalysis*, 2023, **35**, e202300106.
- A. Parihar, S. Yadav, M. A. Sadique, P. Ranjan, N. Kumar, A. Singhal, V. Khare, R. Khan, S. Natarajan and A. K. Srivastava, *Transl. Med.*, 2023, **8**, e10481.
- S. Li, X. Zuo, M. D. Carpenter, R. Verdusco and C. M. Ajo-Franklin, *Nat. Rev. Bioeng.*, 2024, **3**, 30–49.
- F. Mariani, I. Gualandi, W. Schuhmann and E. Scavetta, *Microchim. Acta*, 2022, **189**, 459.
- F. Kurul, D. Aydogan, D. Topcu, F. Kocer, Ü. Akar, A. Koşar, A. E. Cetin and S. N. Topkaya, *npj Biosens.*, 2026, **3**, 18.
- J. Wang, *Analytical Electrochemistry*, Wiley, 1st edn, 2000.
- N. Docherty, D. Macdonald, A. Gordon, A. Dobrea, V. Mani, Y. Fu, S. Pang, M. Jimenez and D. K. Corrigan, *Chem. Commun.*, 2025, **61**, 13359–13377.
- P. Rivera, F. Della Pelle, J. Stonyte, W. D. C. Martins Antunes De Melo, A. Abouhaggar and R. Pauliukaite, *ACS Sens.*, 2025, **10**, 9183–9202.
- R. K. Talreja, H. Sable, V. Chaudhary, S. Kadian, M. Singh, M. Kumar, J. Kishore, V. Chaudhary and A. Khosla, *ECS Sens. Plus*, 2024, **3**, 041602.
- S. Ahuja, M. S. Kumar, R. Nandeshwar, K. Kondabagil and S. Tallur, *Sci. Rep.*, 2022, **12**, 8814.
- A. Ahmed, J. V. Rushworth, N. A. Hirst and P. A. Millner, *Clin. Microbiol. Rev.*, 2014, **27**, 631–646.
- H. M. Man, C. Omar, M. Freisa, D. Bouville, T. Baptiste, A.-M. Haghiri-Gosnet, H. Jacquier, I. Le Potier and J. Gamby, *ACS Electrochem.*, 2025, **1**, 886–896.
- J. Lee, B. Chua and A. Son, *Microchem. J.*, 2024, **207**, 111865.
- S. Ahmadi, F. R. Sedaghat, M. Y. Memar and M. Yekani, *Clin. Chim. Acta*, 2025, **565**, 120020.
- R. R. Singh, J. G. Tigga, S. Kosma and M. K. Patel, *Microchem. J.*, 2026, **220**, 116435.
- D. A. Thai and Y. Liu, *Microsyst. Nanoeng.*, 2025, **11**, 155.
- J. Cui, X. Cheng, X. Guo, D. Li, Z. Xu, W. Qian, Z. Wang and Z. Dai, *TRAC, Trends Anal. Chem.*, 2026, **196**, 118638.
- W. Zieliński and I. Grabowska, *Talanta*, 2026, **301**, 129323.
- E. Cesewski and B. N. Johnson, *Biosens. Bioelectron.*, 2020, **159**, 112214.
- S. Ding, X. Chen, B. Yu and Z. Liu, *Chem. Commun.*, 2024, **60**, 9513–9525.
- A. Khoshroo, M. Mavaei, M. Rostami, B. Valinezhad-Saghezi and A. Fattahi, *BiolImpacts*, 2022, **12**, 567–588.
- D. Kim and S. Yoo, *Chemosensors*, 2022, **10**, 53.
- D. C. Spencer, T. F. Paton, K. T. Mulroney, T. J. J. Inglis, J. M. Sutton and H. Morgan, *Nat. Commun.*, 2020, **11**, 5328.
- Y. Zhou, Q. Wang, T. Xiang and X. Chen, *Nano TransMed*, 2025, **4**, 100078.
- M.-J. Kim, J. A. Lasalde-Ramírez, W. Heng and W. Gao, *Nat. Biotechnol.*, 2026, DOI: 10.1038/s41587-026-03050-2.
- S. Gao, J. M. Guisán and J. Rocha-Martin, *Anal. Chim. Acta*, 2022, **1189**, 338907.
- M. Al Mamun, Y. A. Wahab, M. A. M. Hossain, A. Hashem and M. R. Johan, *TRAC, Trends Anal. Chem.*, 2021, **145**, 116458.



- 45 X. Li, S. Liu, X. Huang, C. Yao, J. Chen, L. Gao, C. Zhou, Y. Wu, J. Liu, M. Li, N. Zhao, H. Chen, S. Huang and X. Xie, *Microsyst. Nanoeng.*, 2025, **11**, 241.
- 46 V. E. Abrantes-Coutinho, A. O. Santos, R. B. Moura, F. N. Pereira-Junior, L. H. Mascaro, S. Morais and T. M. B. F. Oliveira, *Colloids Surf., B*, 2021, **208**, 112148.
- 47 M. O. Sabry, D. Li and L. C. Z. Chan, *Sens. Bio-Sens. Res.*, 2025, **48**, 100794.
- 48 F. Li, X. Jiang, Y. Zhang, R. Yao, H. Zhu, L. Fan and G. Li, *Trends Food Sci. Technol.*, 2026, **168**, 105492.
- 49 X. Meng, D. O'Hare and S. Ladame, *Biosens. Bioelectron.*, 2023, **237**, 115440.
- 50 X. Wang, J. Zhou and H. Wang, *Cell Rep. Phys. Sci.*, 2024, **5**, 101801.
- 51 N. Kalita, S. Gogoi, S. D. Minter and P. Goswami, *ACS Meas. Sci. Au*, 2023, **3**, 404–433.
- 52 M. Ali, S. C. Barman, I. Steer, A. H. Alshareef and D. Alsulaiman, *Adv. Mater. Technol.*, 2026, e02120.
- 53 S. C. Barman, Y. Jin, J. K. El-Demellawi, S. Thomas, N. Wehbe, Y. Lei, M. K. Hota, X. Xu, E. A. Hasan, O. F. Mohammed, O. M. Bakr, D. Alsulaiman and H. N. Alshareef, *Commun. Mater.*, 2025, **6**, 31.
- 54 G. A. Rivas, M. D. Rubianes, M. C. Rodríguez, N. F. Ferreyra, G. L. Luque, M. L. Pedano, S. A. Miscoria and C. Parrado, *Talanta*, 2007, **74**, 291–307.
- 55 P. Sivasubramanian, M. Kumar, C.-L. Chen, V. S. Kiran-kumar, M. S. Samuel and J.-H. Chang, *Environ. Nanotechnol. Monit. Manag.*, 2024, **22**, 101023.
- 56 A. A. Saei, J. E. N. Dolatabadi, P. Najafi-Marandi, A. Abhari and M. de la Guardia, *TrAC, Trends Anal. Chem.*, 2013, **42**, 216–227.
- 57 M. Ye and H. Huang, *Talanta*, 2026, **298**, 128994.
- 58 S. Kumar, S. D. Bukkitgar, S. Singh, Pratibha, V. Singh, K. R. Reddy, N. P. Shetti, C. Venkata Reddy, V. Sadhu and S. Naveen, *ChemistrySelect*, 2019, **4**, 5322–5337.
- 59 A. Dube, S. J. Malode, A. N. Alodhayb, K. Mondal and N. P. Shetti, *Talanta Open*, 2025, **11**, 100395.
- 60 N. Aydemir, J. Malmström and J. Travas-Sejdic, *Phys. Chem. Chem. Phys.*, 2016, **18**, 8264–8277.
- 61 H. Zeng, Y. Xie, T. Liu, Z. Chu, E. Dempsey and W. Jin, *Sens. Diagn.*, 2024, **3**, 165–180.
- 62 P. Jakubec, D. Panáček, M. Nalepa, M. Rossetti, R. Álvarez-Diduk, A. Merkoçi, M. Vasjari, L. Kulla and M. Otyepka, *ChemElectroChem*, 2025, **12**, e202400660.
- 63 X. Yu, H. Cheng, M. Zhang, Y. Zhao, L. Qu and G. Shi, *Nat. Rev. Mater.*, 2017, **2**, 17046.
- 64 Z. Jiang, B. Feng, J. Xu, T. Qing, P. Zhang and Z. Qing, *Biosens. Bioelectron.*, 2020, **166**, 112471.
- 65 H. Barès, A. Bakandritsos, M. Medved, J. Ugolotti, P. Jakubec, O. Tomanec, S. Kalytchuk, R. Zbořil and M. Otyepka, *Carbon*, 2019, **145**, 251–258.
- 66 D. D. Chronopoulos, M. Medved, G. Potsi, O. Tomanec, M. Scheibe and M. Otyepka, *Chem. Commun.*, 2020, **56**, 1936–1939.
- 67 A. Kouloumpis, D. D. Chronopoulos, G. Potsi, M. Pykal, J. Vlček, M. Scheibe and M. Otyepka, *Chem. – Eur. J.*, 2020, **26**, 6518–6524.
- 68 D. Zaoralová, V. Hrubý, V. Šedajová, R. Mach, V. Kupka, J. Ugolotti, A. Bakandritsos, M. Medved and M. Otyepka, *ACS Sustainable Chem. Eng.*, 2020, **8**, 4764–4772.
- 69 J. Janek, Z. Fohlerova, I. Dědek, V. Hrubý, D. Panáček, J. Hubalek, R. Havlík, R. Zbořil, M. Otyepka and P. Jakubec, *Microchim. Acta*, 2025, **192**, 823.
- 70 J. M. R. Flauzino, M. Nalepa, D. D. Chronopoulos, V. Šedajová, D. Panáček, P. Jakubec, P. Kührová, M. Pykal, P. Banáš, A. Panáček, A. Bakandritsos and M. Otyepka, *Small*, 2023, **19**, 2207216.
- 71 Y. Hu, F. Wang, H. Ye, J. Jiang, S. Li, B. Dai, J. Li, J. Yang, X. Song, J. Zhang, Y. Xie, L. Gao and D. Yang, *npj Flex. Electron.*, 2024, **8**, 30.
- 72 N. Gupta, D. Kumar, A. Das, S. Sood and B. D. Malhotra, *New J. Chem.*, 2023, **47**, 10930–10941.
- 73 E. Frantz, D. Han and A. Steckl, *ECS Meet. Abstr.*, 2021, **MA2021-01**, 1416.
- 74 M. J. Russo, M. Han, P. E. Desroches, C. S. Manasa, J. Dennaoui, A. F. Quigley, R. M. I. Kapsa, S. E. Moulton, R. M. Guijt, G. W. Greene and S. M. Silva, *ACS Sens.*, 2021, **6**, 1482–1507.
- 75 A. Dhaffouli, *Microchem. J.*, 2026, **220**, 116637.
- 76 J. Wu, H. Liu, W. Chen, B. Ma and H. Ju, *Nat. Rev. Bioeng.*, 2023, **1**, 346–360.
- 77 H. Karimi-Maleh, Y. Orooji, F. Karimi, M. Alizadeh, M. Baghayeri, J. Rouhi, S. Tajik, H. Beitollahi, S. Agarwal, V. K. Gupta, S. Rajendran, A. Ayati, L. Fu, A. L. Sanati, B. Tanhaei, F. Sen, M. Shabani-Nooshabadi, P. N. Asrami and A. Al-Othman, *Biosens. Bioelectron.*, 2021, **184**, 113252.
- 78 N. Radha Shanmugam, J. Rainbow, J. Lee, P. Jolly and D. E. Ingber, *Small Sci.*, 2025, **5**, e202500380.
- 79 Z. Yousefniaeyjahromi, L. Sembranti, F. Vivaldi, A. Bonini, N. Poma, A. Tavanti and F. Di Francesco, *Microchem. J.*, 2025, **208**, 112595.
- 80 L. C. T. Shoute, G. N. Abdelrasoul, Y. Ma, P. A. Duarte, C. Edwards, R. Zhuo, J. Zeng, Y. Feng, C. L. Charlton, J. N. Kanji, S. Babiuk and J. Chen, *Microsyst. Nanoeng.*, 2023, **9**, 3.
- 81 G. Rosati, M. Urban, L. Zhao, Q. Yang, C. De Carvalho Castro, E. Silva, S. Ronaldo, C. Parolo, E. P. Nguyen, G. Ortega, P. Fornasiero, A. Paccagnella and A. Merkoçi, *Biosens. Bioelectron.*, 2022, **196**, 113737.
- 82 P. Lakhara, V. Chaudhary, A. Jha, R. Singh, P. Kush and P. Kumar, *Mater. Today Chem.*, 2022, **26**, 101129.
- 83 P. M. Kalligiosfyri, A. Miglione and S. Cinti, *ECS Sens. Plus*, 2025, **4**, 010601.
- 84 M. Nalepa, L. Kulla, P. Jakubec, V. Hrubý, D. Panáček, I. Dědek, M. Vasjari and M. Otyepka, *Adv. Mater. Technol.*, 2026, **11**, e01765.
- 85 D. Panáček, V. Kupka, M.-A. Nalepa, I. Dědek, R. Álvarez-Diduk, S. Olenik, J. Flauzino, J. Zdražil, P. Jakubec, L. Zdražil, L. Spichal, K. K. Sonigara, R. Zbořil, M. Pumera, A. Merkoçi, J. Wang, N. De Diego, F. Güder and M. Otyepka, *Nat. Commun.*, 2026, **17**, 2009.
- 86 D. Panáček, M. Urban, A. Silvestri, I. Dědek, M. Nalepa, A. Merkoçi, M. Prato and M. Otyepka, *Small*, 2026, **22**, e13028.
- 87 M.-A. Nalepa, D. Panáček, I. Dědek, P. Jakubec, V. Kupka, V. Hrubý, M. Petr and M. Otyepka, *Biosens. Bioelectron.*, 2024, **256**, 116277.
- 88 M. Jose, M. T. Vijjapu, L. Neumaier, L. Rauter, A. H. Chakkunny, D. Corzo, R. Thoelen, A. Picard, J. Kosel and W. Deferme, *npj Flex. Electron.*, 2025, **9**, 46.
- 89 D. Kireev, S. Kutagulla, J. Hong, M. N. Wilson, M. Ramezani, D. Kuzum, J.-H. Ahn and D. Akinwande, *Nat. Rev. Mater.*, 2024, **9**, 906–922.
- 90 S. Yang, F. G. Berto, J. Denstedt, H. Tang and J. Zhang, *Adv. Sens. Res.*, 2024, **3**, 2400051.
- 91 A. Sharma, A. Agrawal, K. K. Awasthi, K. Awasthi and A. Awasthi, *Mater. Lett. X*, 2021, **10**, 100077.
- 92 V. Templier, T. Livache, S. Boisset, M. Maurin, S. Slimani, R. Mathey and Y. Roupioz, *Sci. Rep.*, 2017, **7**, 9457.
- 93 Q. Zeng, G. Shi and M. Zhang, *Adv. Sens. Res.*, 2024, **3**, 2200018.
- 94 H. Meng, W. Zhong, K. Ma, J. Su, L. Ma, Y. Hao, Y. Jiang, X. Liu, X. Fu and C. Zhang, *Eng. Regen.*, 2024, **5**, 186–198.
- 95 M. Pan, Y. Zhao, J. Qiao and X. Meng, *Folia Microbiol.*, 2024, **69**, 283–304.
- 96 S. Menon, M. R. Mathew, S. Sam, K. Keerthi and K. G. Kumar, *J. Electroanal. Chem.*, 2020, **878**, 114596.
- 97 T. Jiang, X. He, X. Du, S. Xiang, H. He, C. Ge, X. Zhang and Y. Xu, *Microchem. J.*, 2025, **213**, 113833.
- 98 R. R. Jones, D. C. Hooper, L. Zhang, D. Wolverson and V. K. Valev, *Nanoscale Res. Lett.*, 2019, **14**, 231.
- 99 L. M. Moreira, L. Silveira Jr., F. V. Santos, J. P. Lyon, R. Rocha, R. A. Zângaro, A. B. Villaverde and M. T. T. Pacheco, *Spectroscopy*, 2008, **22**, 1–19.
- 100 T. Dai, Z. Xiao, D. Shan, A. Moreno, H. Li, M. Prakash, N. Banaei and J. Rao, *ACS Sens.*, 2023, **8**, 3264–3271.
- 101 Z. Han, H. Liu, J. Meng, L. Yang, J. Liu and J. Liu, *Anal. Chem.*, 2015, **87**, 9500–9506.
- 102 L. Wang, W. Liu, J.-W. Tang, J.-J. Wang, Q.-H. Liu, P.-B. Wen, M.-M. Wang, Y.-C. Pan, B. Gu and X. Zhang, *Front. Microbiol.*, 2021, **12**, 683580.
- 103 A. B. Walter, E. Haugen, A. S. Rourke-Funderburg and A. K. Locke, *Anal. Chem.*, 2025, **97**, 28270–28280.
- 104 J. De Gelder, P. Scheldeman, K. Leus, M. Heyndrickx, P. Vandenabeele, L. Moens and P. De Vos, *Anal. Bioanal. Chem.*, 2007, **389**, 2143–2151.
- 105 P. R. Carey, G. R. Whitmer, M. J. Yoon, M. N. Lombardo, M. Pusztaí-Carey, H. Heidari-Torkabadi and T. Che, *J. Phys. Chem. B*, 2018, **122**, 6377–6385.
- 106 R. Li, D. Dhankhar, J. Chen, A. Krishnamoorthi, T. C. Cesario and P. M. Rentzepis, *IEEE Access*, 2019, **7**, 23549–23559.



- 107 S. Sil, R. Mukherjee, N. S. Kumar, S. Aravind, J. Kingston and U. K. Singh, *Def. Life Sci. J.*, 2017, **2**, 435.
- 108 F. S. De Siqueira, E. Oliveira, A. M. Da Silva, M. T. T. Pacheco, H. E. Giana and L. Silveira, *Lasers Med. Sci.*, 2021, **36**, 289–302.
- 109 J. Zahn, A. Germond, A. Y. Lundgren and M. T. Cicerone, *J. Biophotonics*, 2022, **15**, e202100274.
- 110 O. Vaculík, S. Bernatová, K. Rebrošová, O. Samek, L. Šilhan, F. Růžička, M. Šerý, M. Šiler, J. Ježek and P. Zemánek, *Biomed. Opt. Express*, 2023, **14**, 6410.
- 111 Q. Gao, R. Liu, Y. Wu, F. Wang and X. Wu, *Acta Biomater.*, 2024, **189**, 559–573.
- 112 L. K. Herndon, Y. Zhang, F. Safir, B. Ogunlade, H. B. Balch, A. B. Boehm and J. A. Dionne, *Nano Lett.*, 2025, **25**, 1250–1259.
- 113 J. Wei, Z. Liu, J. Huang, X. Zhong, Q. Zhang, Y. Sun, Y. Wan, Y. Chen, P. Duan and C. Han, *Opt. Mater.*, 2023, **145**, 114463.
- 114 X. Zhang, X. Cui, X. Liu, M. Li, J. Luo and H. Ge, *Colloids Surf., A*, 2026, **741**, 140220.
- 115 S. Das, K. Saxena, J.-C. Tinguely, A. Pal, N. L. Wickramasinghe, A. Khezri, V. Dubey, A. Ahmad, V. Perumal, R. Ahmad, D. N. Wadduwage, B. S. Ahluwalia and D. S. Mehta, *ACS Appl. Mater. Interfaces*, 2023, **15**, 24047–24058.
- 116 S. Wang, F. Li, Z. Luan, L. Li, S. Xi, Z. Du and X. Zhang, *ACS Sens.*, 2024, **9**, 2402–2412.
- 117 C. Qu, Y. Li, G. Li, X. Wang, M. Su and H. Liu, *ACS Appl. Mater. Interfaces*, 2024, **16**, 32824–32835.
- 118 K. Li, Y. Yang, C. Xu, Y. Ye, L. Huang, L. Sun, Y. Cai, W. Zhou, Y. Ge, Y. Li, Q. Zhang, Y. Wang and X. Liu, *Sens. Actuators, B*, 2023, **380**, 133381.
- 119 Y. Zhou, W. Zhang, S. Cheng, Y. Xing, J. Wang, F. Yu and R. Wang, *Talanta*, 2025, **293**, 128089.
- 120 Y. Yang, C. Zeng, J. Huang, M. Wang, W. Qi, H. Wang and Z. He, *Biosens. Bioelectron.*, 2022, **215**, 114524.
- 121 X. Li, J. Hu, D. Zhang, X. Zhang, Z. Wang, Y. Wang, Q. Chen and P. Liang, *Talanta*, 2024, **271**, 125650.
- 122 H. Zhang, L. Yang, M. Zhang, H. Wei, L. Tong, H. Xu and Z. Li, *Nano Lett.*, 2024, **24**, 11116–11123.
- 123 P. Wen, F. Yang, H. Zhao, S. Li, Y. Xu and L. Chen, *ACS Sens.*, 2024, **9**, 6167–6173.
- 124 P. Wen, F. Yang, H. Zhao, Y. Xu, S. Li and L. Chen, *Anal. Chem.*, 2024, **96**, 1454–1461.
- 125 X. Yue, S. Yan, T. Gao, S. Pu, H. Tang, X. Pei, Z. Tian, X. Wang, B. Ren and G. Liu, *Anal. Chem.*, 2024, **96**, 17517–17525.
- 126 A. Pistiki, O. Ryabchikov, A. Wagenhaus, T. W. Bocklitz, S. Deinhardt-Emmer, B. Löffler, P. Rösch and J. Popp, *Anal. Chem.*, 2026, **98**, 6523–6531.
- 127 W. Liu, J.-W. Tang, J.-Y. Mou, J.-W. Lyu, Y.-W. Di, Y.-L. Liao, Y.-F. Luo, Z.-K. Li, X. Wu and L. Wang, *Front. Microbiol.*, 2023, **14**, 1101357.
- 128 F. Li, Y.-T. Si, J.-W. Tang, Z. Umar, X.-S. Xiong, J.-T. Wang, Q. Yuan, A. C. Y. Tay, E. G. Chua, L. Zhang, B. J. Marshall, W.-X. Yang, B. Gu and L. Wang, *Comput. Struct. Biotechnol. J.*, 2024, **23**, 3379–3390.
- 129 M. Benešová, S. Bernatová, F. Mika, Z. Pokorná, J. Ježek, M. Šiler, O. Samek, F. Růžička, K. Rebrošová, P. Zemánek and Z. Pilát, *Biosensors*, 2023, **13**, 182.
- 130 A. D. Vasilyeva, L. V. Yurina, E. G. Evtushenko, E. S. Gavrilina, V. B. Krylov, N. E. Nifantiev and I. N. Kurochkin, *Sensors*, 2025, **25**, 4376.
- 131 L. Jin, S. Wang, Q. Shao and Y. Cheng, *Spectrochim. Acta, Part A*, 2022, **267**, 120625.
- 132 C.-S. Ho, N. Jean, C. A. Hogan, L. Blackmon, S. S. Jeffrey, M. Holodniy, N. Banaei, A. A. E. Saleh, S. Ermon and J. Dionne, *Nat. Commun.*, 2019, **10**, 4927.
- 133 Y. Pang, N. Wan, L. Shi, C. Wang, Z. Sun, R. Xiao and S. Wang, *Anal. Chim. Acta*, 2019, **1077**, 288–296.
- 134 L. Jin, X. Cai, F. Ren and J. Yang, *Sens. Actuators, B*, 2024, **405**, 135356.
- 135 N. E. Dina, M. A. Tahir, S. Z. Bajwa, I. Amin, V. K. Valev and L. Zhang, *Biosens. Bioelectron.*, 2023, **219**, 114843.
- 136 Z. Tas, F. Ciftci, K. Icoz and M. Unal, *Spectrochim. Acta, Part A*, 2025, **339**, 126285.
- 137 F. Zeng, W. Duan, B. Zhu, T. Mu, L. Zhu, J. Guo and X. Ma, *Anal. Chem.*, 2019, **91**, 1064–1070.
- 138 T. Mu, S. Li, H. Feng, C. Zhang, B. Wang, X. Ma, J. Guo, B. Huang and L. Zhu, *IEEE J. Sel. Top. Quantum Electron.*, 2019, **25**, 1–6.
- 139 C. Sacco Botto, C. Orecchio, C. D'Errico, E. Alladio, E. Noris and M. Vincenti, *Spectrochim. Acta, Part A*, 2026, **344**, 126701.
- 140 B. L. Thomsen, J. B. Christensen, O. Rodenko, I. Usenov, R. B. Grønne-mose, T. E. Andersen and M. Lassen, *Sci. Rep.*, 2022, **12**, 16436.
- 141 H. Zhao, D. Zheng, H. Wang, T. Lin, W. Liu, X. Wang, W. Lu, M. Liu, W. Liu, Y. Zhang, M. Liu and P. Zhang, *Int. J. Mol. Sci.*, 2022, **23**, 7340.
- 142 J.-W. Tang, Q.-H. Liu, X.-C. Yin, Y.-C. Pan, P.-B. Wen, X. Liu, X.-X. Kang, B. Gu, Z.-B. Zhu and L. Wang, *Front. Microbiol.*, 2021, **1–14**.
- 143 S. Guo, J. Popp and T. Bocklitz, *Nat. Protoc.*, 2021, **16**, 5426–5459.
- 144 S. Sloan-Dennison, G. Q. Wallace, W. A. Hassanain, S. Laing, K. Faulds and D. Graham, *Nano Converg.*, 2024, **11**, 33.
- 145 Y. Li, J. Xu, X. Yi, X. Li, Y. Luo, A. Glidle, P. Summersgill, S. Allen, T. Ryan, X. Liu, W. Yu, X. Chu, S. Chen, Q. Zhang, X. Xu, X. Hua, Q. Yang, J. Reboud, Y. Yu, W. E. Huang, J. M. Cooper and H. Yin, *Nat. Commun.*, 2025, **17**, 283.
- 146 J. Ki, I. H. Kwon, J. Lee, J. Lim, S. Jang, S. U. Son, S. B. Seo, S. Y. Oh, T. Ryan, J. Jung, K. G. Lee, J. Hwang and E.-K. Lim, *J. Hazard. Mater.*, 2023, **460**, 132398.
- 147 L. Zheng, C. Zheng, W. Wang, F. Huang, Y. Jiang, J. Lu and Y. Lou, *Colloids Surf., B*, 2025, **250**, 114541.
- 148 S. Kausshal, S. Shrivastava, Y.-R. Yun, Y. Park, T. Thanh-Quy Nguyen, M. Meeseepong, E. Lee, B. Jeon, M. B. Gu, S. Yang and N.-E. Lee, *ACS Sens.*, 2022, **7**, 2188–2197.
- 149 X. Huang, L. Chen, W. Zhi, R. Zeng, G. Ji, H. Cai, J. Xu, J. Wang, S. Chen, Y. Tang, J. Zhang, H. Zhou and P. Sun, *Anal. Chem.*, 2023, **95**, 13101–13112.
- 150 H. Kamel, J. Schmid, M. R. Nizami, I. Alekseenko, F. Hausladen, D. Claus, R. Wittig and D. P. Kelly, in *Optics and Photonics for Advanced Dimensional Metrology III*, ed. P. J. De Groot, P. Picart and F. Guzman, SPIE, Strasbourg, France, 2024, p. 51.
- 151 E. C. Wilkerson, K. L. Singampalli, J. Li, D. D. Dixit, X. Jiang, D. H. Gonzalez and P. B. Lillehoj, *Anal. Bioanal. Chem.*, 2023, **415**, 3983–4002.
- 152 E. Jarosińska, Z. Zambrowska and E. Witkowska Nery, *ACS Omega*, 2024, **9**, 4572–4580.
- 153 B. Zhu, J. Chen, B. Wang, H. Zhou, R. Xiao, Z. Gao and Y. Wang, *Anal. Chem.*, 2025, **97**, 20834–20842.
- 154 C. Yan, *iScience*, 2025, **28**, 112759.
- 155 H. Wei, W. Leng, J. Song, M. R. Willner, L. C. Marr, W. Zhou and P. J. Vikesland, *Anal. Chem.*, 2018, **90**, 3227–3237.
- 156 G. Resznetnik, K. Hammond, S. Mahshid, T. AbdElFatah, D. Nguyen, R. Corsini, C. Caya, J. Papenburg, M. P. Cheng and C. P. Yansouni, *Nat. Commun.*, 2024, **15**, 9719.
- 157 H. Çelik, B. B. Caf and G. Çebi, *Indian J. Microbiol.*, 2024, **64**, 894–909.
- 158 S.-W. Huang, Y.-F. Wu, T. Ahmed, S.-C. Pan and C.-M. Cheng, *Trends Biotechnol.*, 2024, **42**, 74–90.
- 159 H.-J. Park, A. Wool, J. M. Cha, K.-S. Lee and M.-Y. Lee, *Sci. Rep.*, 2018, **8**, 16508.
- 160 A. M. Peri, M. D. Chatfield, W. Ling, L. Furuya-Kanamori, P. N. A. Harris and D. L. Paterson, *Clin. Infect. Dis.*, 2024, **79**, 502–515.
- 161 Y.-Y. Han, Y.-C. Lin, W.-C. Cheng, Y.-T. Lin, L.-J. Teng, J.-K. Wang and Y.-L. Wang, *Sci. Rep.*, 2020, **10**, 12538.
- 162 Q. Bi, M. Liu, L. Yan, J. Cheng, Q. Sun, Y. Dai and L. Zou, *Front. Microbiol.*, 2025, **1–15**.
- 163 M. Xie, Y. Zhu, Z. Li, Y. Yan, Y. Liu, W. Wu, T. Zhang, Z. Li and H. Wang, *Talanta*, 2024, **268**, 125281.
- 164 A. M. Peri, A. Stewart, A. Hume, A. Irwin and P. N. A. Harris, *Curr. Infect. Dis. Rep.*, 2021, **23**, 12.
- 165 S. S. Arbefeville, T. T. Timbrook and C. D. Garner, *J. Antimicrob. Chemother.*, 2024, **79**, i2–i8.
- 166 P. Lasch, W. Beyer, A. Bosch, R. Borriss, M. Drevinek, S. Dupke, M. Ehling-Schulz, X. Gao, R. Grunow, D. Jacob, S. R. Klee, A. Paauw, J. Rau, A. Schneider, H. C. Scholz, M. Stämmler, L. T. Thanh Tam, H. Tomaso, G. Werner and J. Doellinger, *Sci. Data*, 2025, **12**, 187.
- 167 A. Cuénod, M. Aerni, C. Bagutti, B. Bayraktar, E. S. Boz, C. B. Carneiro, C. Casanova, A. T. Coste, P. Damborg, D. W. Van Dam, M. Demirci, P. Drevinek, O. Dubuis, J. Fernandez, G. Greub, J. Hrabak, G. Hürkal Yigitler, J. Hurych, T. G. Jensen, G. Jost, G. A. Kampinga, S. Kittl, C. Lammens, C. Lang, R. Lienhard,



- J. Logan, C. Maffioli, I. Mareković, M. Marschal, J. Moran-Gilad, O. Nolte, M. Oberle, M. Pedersen, V. Pflüger, S. Pranghofer, J. Reichl, R. J. Rentenaar, A. Riat, B. Rodríguez-Sánchez, C. Schilt, A.-K. Schlotterbeck, J. Schrenzel, S. Troib, E. Willems, M. Wootton, D. Ziegler and A. Egli, *Clin. Microbiol. Infect.*, 2023, **29**, 190–199.
- 168 A. Cuénod, F. Foucault, V. Pflüger and A. Egli, *Front. Cell. Infect. Microbiol.*, 2021, **11**, 646648.
- 169 S. Kundu, R. Varshney and S. Sulabh, *Gene Genome Ed.*, 2024, **7**, 100032.
- 170 P. Srivastava and D. Prasad, *3 Biotech*, 2023, **13**, 200.
- 171 W. Gu, X. Deng, M. Lee, Y. D. Sucu, S. Arevalo, D. Stryke, S. Federman, A. Gopez, K. Reyes, K. Zorn, H. Sample, G. Yu, G. Ishpuniani, B. Briggs, E. D. Chow, A. Berger, M. R. Wilson, C. Wang, E. Hsu, S. Miller, J. L. DeRisi and C. Y. Chiu, *Nat. Med.*, 2021, **27**, 115–124.
- 172 A. Elbehiry and A. Abalkhail, *Diagnostics*, 2025, **15**, 1991.
- 173 R. D. Riley, L. Archer, K. I. E. Snell, J. Ensor, P. Dhiman, G. P. Martin, L. J. Bonnett and G. S. Collins, *BMJ*, 2024, **384**, e074820.
- 174 G. S. Collins, K. G. M. Moons, P. Dhiman, R. D. Riley, A. L. Beam, B. Van Calster, M. Ghassemi, X. Liu, J. B. Reitsma, M. Van Smeden, A.-L. Boulesteix, J. C. Camaradou, L. A. Celi, S. Denaxas, A. K. Denniston, B. Glocker, R. M. Golub, H. Harvey, G. Heinze, M. M. Hoffman, A. P. Kengne, E. Lam, N. Lee, E. W. Loder, L. Maier-Hein, B. A. Mateen, M. D. McCradden, L. Oakden-Rayner, J. Ordish, R. Parnell, S. Rose, K. Singh, L. Wynants and P. Logullo, *BMJ*, 2024, **385**, e078378.
- 175 P. M. Bossuyt, J. B. Reitsma, D. E. Bruns, C. A. Gatsonis, P. P. Glasziou, L. Irwig, J. G. Lijmer, D. Moher, D. Rennie, H. C. W. de Vet, H. Y. Kressel, N. Rifai, R. M. Golub, D. G. Altman, L. Hooft, D. A. Korevaar, J. F. Cohen and STARD Group, *Clin. Chem.*, 2015, **61**, 1446–1452.
- 176 J.-F. Masson, *ACS Sens.*, 2020, **5**, 3290–3292.
- 177 K. Karagianni, T. Leontidou, M. Constantinou and C. Andreou, *Analyst*, 2025, **150**, 3762–3787.
- 178 Z. Nie, Z. Huang, Z. Wu, Y. Xing, F. Yu and R. Wang, *Spectrochim. Acta, Part A*, 2025, **336**, 126051.
- 179 L. Rodriguez, Z. Zhang and D. Wang, *Anal. Sci. Adv.*, 2023, **4**, 81–95.
- 180 T. F. Monaghan, S. N. Rahman, C. W. Agudelo, A. J. Wein, J. M. Lazar, K. Everaert and R. R. Dmochowski, *Medicina*, 2021, **57**, 503.
- 181 L. P. Murray and C. R. Mace, *Anal. Chim. Acta*, 2020, **1140**, 236–249.
- 182 J. J. Workman, *Appl. Spectrosc.*, 2018, **72**, 340–365.
- 183 G. L. Goh, M. F. Tay, J. M. Lee, J. S. Ho, L. N. Sim, W. Y. Yeong and T. H. Chong, *Adv. Electron. Mater.*, 2021, **7**, 2100043.
- 184 D. Cialla-May, A. Bonifacio, T. Bocklitz, A. Markin, N. Markina, S. Fornasaro, A. Dwivedi, T. Dib, E. Farnesi, C. Liu, A. Ghosh and J. Popp, *Chem. Soc. Rev.*, 2024, **53**, 8957–8979.
- 185 M. E. Berry, H. Kearns, D. Graham and K. Faulds, *Analyst*, 2021, **146**, 6084–6101.
- 186 M. L. Chiu, W. Lawi, S. T. Snyder, P. K. Wong, J. C. Liao and V. Gau, *SLAS Technol.*, 2010, **15**, 233–242.
- 187 P. L. Voyvodic, I. Conejero, K. Mesmoudi, E. Renard, P. Courtet, D. I. Cattoni and J. Bonnet, *Sci. Rep.*, 2022, **12**, 13785.
- 188 M. Parihar, N. W. N., Sahana, R. Biswas, B. Dehury and N. Mazumder, *RSC Adv.*, 2025, **15**, 29267–29283.
- 189 C. Hwang, W.-J. Lee, S. D. Kim, S. Park and J. H. Kim, *Biosensors*, 2022, **12**, 1020.
- 190 S. Li, H. Zhang, M. Zhu, Z. Kuang, X. Li, F. Xu, S. Miao, Z. Zhang, X. Lou, H. Li and F. Xia, *Chem. Rev.*, 2023, **123**, 7953–8039.
- 191 S. L. Taylor, C. R. Brooks, L. Elms, S. K. Manning, A. Richard, J. Burmanje, J. Douwes and G. B. Rogers, *Respir. Res.*, 2025, **26**, 209.
- 192 O. Y. F. Henry, J. L. A. Sanchez and C. K. O'Sullivan, *Biosens. Bioelectron.*, 2010, **26**, 1500–1506.
- 193 Z. Xu, R. Han, N. Liu, F. Gao and X. Luo, *Sens. Actuators, B*, 2020, **319**, 128253.
- 194 T. Feng, W. Ji, Y. Zhang, F. Wu, Q. Tang, H. Wei, L. Mao and M. Zhang, *Angew. Chem., Int. Ed.*, 2020, **59**, 23445–23449.
- 195 Y. L. Wu, J. Jia, J. Das, K. T. Riordan, C. D. Flynn, Y. Wang, S. O. Kelley and T. W. Odom, *J. Am. Chem. Soc.*, 2025, **147**, 7868–7874.
- 196 P.-H. Lin and B.-R. Li, *Analyst*, 2020, **145**, 1110–1120.
- 197 C. Muñoz-San Martín, M. Gamella, M. Pedrero, A. Montero-Calle, R. Barderas, S. Campuzano and J. M. Pingarrón, *Sens. Actuators, B*, 2020, **307**, 127623.
- 198 W. R. Premasiri, Y. Chen, P. M. Williamson, D. C. Bandara, C. Pyles and L. D. Ziegler, *Anal. Bioanal. Chem.*, 2017, **409**, 3043–3054.
- 199 R. Hunter, A. N. Sohi, Z. Khatoon, V. R. Berthiaume, E. I. Alarcon, M. Godin and H. Anis, *Sens. Actuators, B*, 2019, **300**, 126907.
- 200 Y. Wang, Q. Li, R. Zhang, K. Tang, C. Ding and S. Yu, *Microchim. Acta*, 2020, **187**, 290.
- 201 J. Loo, H. C. Kwok, C. C. H. Leung, S. Y. Wu, I. L. G. Law, Y. K. Cheung, Y. Y. Cheung, M. L. Chin, P. Kwan, M. Hui, S. K. Kong and H. P. Ho, *Biosens. Bioelectron.*, 2017, **93**, 212–219.
- 202 M. L. Sin, K. E. Mach, P. K. Wong and J. C. Liao, *Expert Rev. Mol. Diagn.*, 2014, **14**, 225–244.
- 203 A. A. Akhlaghi, H. Kaur, B. R. Adhikari and L. Soleymani, *ECS Sens. Plus*, 2024, **3**, 011601.
- 204 G. Zhou and H. Liu, *Biosensors*, 2026, **16**, 112.
- 205 Y. Zhao, A. Kumar and Y. Yang, *Chem. Soc. Rev.*, 2024, **53**, 1004–1057.
- 206 P. J. Garcia, P. You, G. Fridley, D. Mabey and R. Peeling, *Lancet Glob. Health*, 2015, **3**, e257–e258.
- 207 M. D. M. Castro, H. M. Ismail, C. A. Montenegro-Quinonez, E. I. Reipold, S. Shilton, C. Denkinger and S. Yerlikaya, *BMJ Open*, 2025, **15**, e092774.
- 208 European Parliament and Council of the European Union, Regulation (EU) 2017/746 of the European Parliament and of the Council of 5 April 2017 on in Vitro Diagnostic Medical Devices and Repealing Directive 98/79/EC and Commission Decision 2010/227/EU, 2017.
- 209 ISO 13485 Medical Devices, <https://www.iso.org/standard/59752.html>, (accessed 25 May 2026).
- 210 M. Otyepka, M. Pykal and M. Otyepka, *Appl. Mater. Today*, 2025, **47**, 102981.
- 211 M. M. Li, B. Y. Reis, A. Rodman, T. Cai, N. Dagan, R. D. Balicer, J. Loscalzo, I. S. Kohane and M. Zitnik, *Nat. Med.*, 2026, **32**, 439–448.

