



Cite this: DOI: 10.1039/d5sd00236b

Recent advances in multiplexed electrochemiluminescence immunoassays

 Yu Liu,^a Chunyan Liu,^a Zhihui Dai ^{abc} and Weiliang Guo ^{*ab}

Electrochemiluminescence (ECL) immunoassays have emerged as a powerful method for multiplex biomarker detection due to their advantages of low background noise, high sensitivity, and ease of miniaturization, holding significant applications in precision medicine and *in vitro* diagnostics (IVD). This article systematically reviews recent advances in ECL-based multiplex immunoassays, focusing on two primary strategies for achieving multiplex detection. One is the spectrum-resolved strategy that enables differentiation of multiple targets by utilizing multicolor ECL luminophores, such as metal complexes and quantum dots, which emit light at distinct wavelengths. The other is the spatially resolved strategy that employs technologies including microarrays, microfluidic chips, and encoded microbeads to physically separate detection sites, allowing high-throughput analysis with just one luminophore. This review summarizes innovative designs under both strategies aimed at enhancing detection performance, discusses their applications in the diagnosis of diseases such as cancer and cardiovascular disorders, and addresses current challenges related to analytical performance, system integration, and clinical use. Finally, prospects for ECL technology in IVD and precision medicine are outlined.

 Received 28th December 2025,
Accepted 13th March 2026

DOI: 10.1039/d5sd00236b

rsc.li/sensors

1. Introduction

Biomarkers are widely recognized as biochemical indicators of disease initiation and progression and are therefore crucial for timely and precise diagnosis.^{1,2} However, conventional

immunoassays are typically limited to measuring only one biomarker at a time (termed as single-plex immunoassays). The quantification of biomarkers always depends on a single readout signal, rendering the results susceptible to environmental interference, and consequently resulting in false-positive or false-negative errors.^{3,4} In contrast, multiplex immunoassays can measure multiple biomarkers in a single run and significantly improve diagnostic accuracy and efficiency. Moreover, they offer higher detection throughput, lower sample consumption, and reduced assay costs.⁵ Therefore, multiplex immunoassays have garnered widespread attention in clinical bioanalysis.^{6,7}

^a Collaborative Innovation Center of Biomedical Functional Materials, School of Chemistry and Materials Science, Nanjing Normal University, Nanjing 210023, China. E-mail: guowl@njnu.edu.cn

^b State Key Laboratory of Analytical Chemistry for Life Science, Nanjing University, Nanjing 210023, China

^c School of Chemistry and Molecular Engineering, Nanjing Tech University, Nanjing 211816, China



Yu Liu

Yu Liu received her B.Sc. degree in Chemistry from Anhui Normal University in 2025. She is currently a M.Sc. candidate in Analytical Chemistry at Nanjing Normal University. Her main research interests include electrochemiluminescence imaging and bioassays.



Chunyan Liu

Chunyan Liu received her B.Sc. degree in Chemistry from Nanjing Normal University in 2024. She is currently a M.Sc. candidate in Analytical Chemistry at Nanjing Normal University. Her main research interests focus on molecular crystal-based electrochemiluminescence analysis.



In recent years, numerous efforts have focused on developing novel multiplex immunoassays by integrating techniques such as radioisotopic,⁸ optical,^{9–11} spectroscopic,¹² and electrochemical methods.¹³ Electrochemiluminescence (ECL) refers to a light-emitting process in which the excited-state species is generated *via* high energy electron transfer reactions at the electrode surface.^{14–16} In comparison with the methods that rely on external light excitation such as fluorescence and surface plasmon resonance, ECL reactions do not require an external light source for excitation, thus avoiding interference from light scattering and offering outstanding advantages of high detection sensitivity and near-zero background.¹⁷ When compared with chemiluminescence that takes place in a mixed solution, ECL reactions occur near the electrode surface and are highly sensitive to interfacial processes, exhibiting excellent spatiotemporal control.¹⁸

A wide variety of ECL luminophores, ranging from aromatic hydrocarbons and metal complexes to nanomaterials, have been synthesized.^{19–21} However, tris(2,2'-bipyridyl) ruthenium(II) ($\text{Ru}(\text{bpy})_3^{2+}$) remains the most extensively employed in fundamental studies. In the presence of tri-*n*-propylamine (TPPrA) as a co-reactant, $\text{Ru}(\text{bpy})_3^{2+}$ exhibits high and stable ECL emission in aqueous solutions, thus making ECL one of the most powerful commercial analytical techniques.^{14,22} For instance, over 100 distinct biomarkers (*e.g.*, cancer, inflammatory, or cardiac biomarkers, as well as hormones) can be determined based on ECL immunoassay devices commercialized by Roche Diagnostics and Meso Scale Discovery (MSD).

This review offers a comprehensive overview of ECL immunoassays, designed specifically for researchers new to the field. We begin with a brief introduction to the fundamental principles of ECL, followed by a summary of recent progress in both spectrally and spatially resolved strategies for constructing multiplexed ECL immunoassays. Key achievements in enhancing analytical performance are also highlighted. Finally,


we discuss the current challenges and future perspectives for the advancement of ECL immunoassays.

2. Fundamentals of ECL

2.1 Annihilation pathway

The annihilation pathway represents one of the earliest elucidated and most classical mechanisms for ECL generation. Early detailed studies of annihilation ECL focused primarily on polycyclic aromatic hydrocarbons such as rubrene and 9,10-diphenylanthracene (DPA), as well as metal complexes like tris(bipyridine)ruthenium(II) ($\text{Ru}(\text{bpy})_3^{2+}$).²³ In this scheme, the oxidized and reduced species are typically generated sequentially at two different electrode potentials, followed by electron transfer between these species to produce an excited state.²⁴

In general, the annihilation mechanism can be summarized as follows. The potential of the working electrode is switched rapidly between two values. At the cathode, luminophore A is reduced to $\text{A}^{\cdot-}$ (eqn (1)), while at the anode, luminophore D is oxidized to $\text{D}^{\cdot+}$ (eqn (2)). These two species annihilate near the electrode surface to generate the excited-state species A^* (eqn (3)), which then emits light (eqn (4)). Depending on the identity of the reactants, the process is classified as “homogeneous annihilation” (A and D are the same species) or “heterogeneous annihilation” (A and D are different species).^{25,26} For example, in the homogeneous annihilation of $\text{Ru}(\text{bpy})_3^{2+}$, oxidized $\text{Ru}(\text{bpy})_3^{3+}$ and reduced $\text{Ru}(\text{bpy})_3^+$ react to produce the excited state $\text{Ru}(\text{bpy})_3^{2+*}$ with concomitant light emission. A representative heterogeneous annihilation process is the cross-reaction between the DPA anion radical and the *N,N,N',N'*-tetramethyl-*p*-phenylenediamine (TMPD) cation radical.²⁷



Zhihui Dai

Zhihui Dai received her Ph.D. degree in Analytical Chemistry from Nanjing University in 2004, and since then, she has been engaged in the research of nanomaterial-based biosensing. She was a professor at the School of Chemistry and Materials Science, Nanjing Normal University from 2009 to 2022. Currently, she works at the School of Chemistry and Molecular Engineering, Nanjing Tech University.



Weiliang Guo

Weiliang Guo received his B.Sc. degree in Applied Chemistry from Anhui University in 2013 and Ph.D. degree in Chemistry from Zhejiang University in 2018. After two years at Zhejiang University as a postdoc, he joined the School of Chemistry and Materials Science, Nanjing Normal University as an associate professor in 2021 and was promoted to full professor in 2025. His research interests are focused on spatially resolved electrochemistry, electrochemiluminescence, and their applications in bioassays.







Depending on the relationship between the reaction enthalpy change and the excited state energy (eqn (5)), where $-\Delta H_{\text{ann}}$ is the enthalpy of ion annihilation and E_p is the peak potential for electrochemical oxidation or reduction, the annihilation pathway can be further classified into the singlet-state route (S-route) and the triplet-state route (T-route).²⁸

$$-\Delta H_{\text{ann}} = E_p(D^+/D) - E_p(A/A^-) - 0.157 \quad (5)$$

When $-\Delta H_{\text{ann}} > E_s$, the system is energy-sufficient and follows the S-route, directly producing the singlet excited state $^1A^*$ (eqn (6)), which then emits light, as observed in the DPA annihilation system.^{29,30}







When $-\Delta H_{\text{ann}} < E_s$, the system is energy-deficient and follows the T-route. Here, the triplet excited state $^3A^*$ is initially formed (eqn (7)), followed by its conversion to $^1A^*$ via triplet-triplet annihilation (eqn (8)), as exemplified by the TMPD⁺/DPA⁻ system.³¹



2.2 Co-reactant pathway

While the annihilation pathway represents a milestone in elucidating the fundamental principles of ECL, the reliance on non-aqueous solvents restricts its application in bioanalysis. In contrast, the co-reactant pathway plays a crucial role in commercial ECL immunoassays due to its exceptional efficiency and compatibility in aqueous media. In this case, the use of a co-reactant can significantly enhance the ECL generation from the luminophore. Based on the electrochemical processes occurring near the electrode and the types of active intermediates generated, the co-reactant mechanisms can be accordingly categorized into two main pathways, namely “oxidative–reduction” and “reductive–oxidation”.³²

Oxidative–reductive co-reactants generate strong reducing intermediates upon electrochemical oxidation, which then reduce the oxidized luminophore to form the excited state. This type of co-reactant is the most widely used, with typical examples including oxalate ions ($C_2O_4^{2-}$) and TPrA. Oxalate, the first discovered co-reactant, undergoes bond cleavage under anodic oxidation conditions to produce the strongly reducing $CO_2^{\cdot-}$ (eqn (9)). Simultaneously, the luminophore $Ru(bpy)_3^{2+}$ is oxidized to $Ru(bpy)_3^{3+}$ (eqn (10)). Subsequently, $CO_2^{\cdot-}$ reduces $Ru(bpy)_3^{3+}$ to the excited state $Ru(bpy)_3^{2+*}$ (eqn (11)), which then emits light (eqn (12)).²²



TPrA, as a co-reactant in commercial ECL systems, exhibits a more complex oxidation process involving multiple parallel reaction pathways.^{33,34} Initially, TPrA undergoes electrochemical oxidation at the electrode surface to form the radical cation TPrA⁺ (eqn (13)), which rapidly undergoes deprotonation to yield the strongly reducing intermediate TPrA' (eqn (14)).³⁵



Fig. 1 summarizes several proposed reaction mechanisms. For instance, this reducing intermediate can directly reduce $Ru(bpy)_3^{3+}$ to generate the excited state (Fig. 1a). Simultaneously, TPrA' can also reduce $Ru(bpy)_3^{2+}$ to $Ru(bpy)_3^+$, which subsequently undergoes an annihilation reaction with $Ru(bpy)_3^{3+}$ to produce light emission (Fig. 1b). Fig. 1c illustrates the generation of the excited state $Ru(bpy)_3^{2+*}$ via the so-called low oxidation potential pathway, where only TPrA, but not $Ru(bpy)_3^{2+}$, undergoes electrochemical oxidation. In contrast, in the catalytic pathway TPrA⁺ and TPrA' are generated via homogeneous oxidation between electrogenerated $Ru(bpy)_3^{3+}$ and TPrA, rather than through direct electrochemical oxidation of TPrA at the electrode surface (Fig. 1d). Notably, this pathway is favored when the concentration of $Ru(bpy)_3^{2+}$ is high.³⁶

Reductive–oxidative co-reactants generate strong oxidizing intermediates upon electrochemical reduction, which then oxidize the reduced luminophore to yield the excited state. A typical example is peroxydisulfate ($S_2O_8^{2-}$), which undergoes cathodic reduction to generate $SO_4^{\cdot-}$ (eqn (15)), possessing an extremely high oxidation potential (3.15 V vs. SCE). $Ru(bpy)_3^{2+}$ is first reduced to $Ru(bpy)_3^+$ (eqn (16)) and is subsequently oxidized by $SO_4^{\cdot-}$ to the emissive excited state (eqn (17)).³⁷



At present, a deeper understanding of ECL mechanisms is achieved, facilitated by various advanced techniques including optical microscopy, mass spectrometry, and spectroscopic measurement. For example, electrochemistry coupled mass spectrometry (ECMS) enables the capture and






Fig. 1 (a)–(d) Proposed mechanism for $\text{Ru}(\text{bpy})_3^{2+}/\text{TPrA}$ ECL excited state formation and light emission.

identification of short-lived intermediates (*e.g.*, radical cation $\text{TPrA}^{\bullet+}$ and neutral radical TPrA^{\bullet}) involved in ECL reactions, providing direct evidence for co-reactant pathways and transforming previously proposed ECL mechanisms into experimentally validated findings.^{38,39} Furthermore, both ECL-based microscopy (ECLM) and ECL self-interference spectroscopy (ECLIS) allow the observation of the spatial distribution of the ECL-emitting layer, and have been employed to decipher ECL reaction mechanisms. By measuring the thickness of the emission layer, the switch of the dominant ECL pathway can be resolved and regulated *via* the luminophore concentration or the concentration ratio of the luminophore to the co-reactant.^{40–42}

3. Strategies for multiplexed ECL detection

3.1 Spectrum-resolved strategy

The spectrum-resolved multiplexed ECL immunoassay represents a significant technological advancement beyond conventional single-analyte detection. It involves constructing specific molecular luminophores through the integration of biological recognition units and ECL luminophores with distinguishable emission wavelengths, thereby allowing for

the simultaneous, high-throughput detection of multiple analytes in complex biological samples. The exploration of spectral resolution strategies begins with the use of structurally well-defined metal complexes. Richter characterized the spectral properties of inorganic metal complexes including $\text{Ru}(\text{bpy})_3^{2+}$ derivatives, $\text{Os}(\text{phen})_2(\text{dppene})^{2+}$ (phen = 1,10-phenanthroline; dppene = *cis*-1,2-bis(diphenylphosphino)ethylene), and $\text{Ir}(\text{ppy})_3$ (ppy = 2-phenylpyridine). By judicious modification of ligands with bipyridine and phenanthroline derivatives, emission wavelengths were precisely tuned across the 500–700 nm range, yielding a versatile luminophore panel spanning blue, green, and red spectral regions.^{43,44} The studies have paved the way for achieving multiplex detection in spectrum-resolved ECL immunoassays.

In spectrum-resolved multiplex ECL immunoassays, the spectral characteristics of the luminophores directly determine the number of targets that can be simultaneously detected. However, the major limitation of molecular luminophores is their broad full-width at half-maximum, which results in substantial spectral overlap and constrains the multiplexing capacity of spectrum-resolved ECL immunoassays. Su *et al.* developed a potential-modulated spectrum-resolved strategy based on novel ruthenium(II) and iridium(III) complexes, thus constructing a dual-resolved ECL system for






Fig. 2 Potential-resolved and spectrum-resolved ECL strategy for multiplex immunoassays. (a) Schematic procedure for the simultaneous detection of three tumor markers. (b) ECL detection platform and the principle of selectively activating three luminophores at different applied potentials. Reprinted with permission from ref. 45. Copyright 2018 American Chemical Society.

multiplex immunoassays (Fig. 2a). They synthesized $\text{Ru}(\text{bpy})_2(\text{dvpby})^{2+}$ (bpy = 2,2'-bipyridine; dvpby = 4,4'-bis(4-vinylphenyl)-2,2'-bipyridine, $\lambda_{\text{max}} = 636$ nm) and $\text{Ir}(\text{dFCF}_3\text{ppy})_2(\text{dtbbpy})^+$ (dFCF_3ppy = 3,5-difluoro-2-[5-(trifluoromethyl)-2-pyridinyl]-phenyl; dtbbpy = 4,4'-bis(*tert*-butyl)-2,2'-bipyridine, $\lambda_{\text{max}} = 491$ nm), achieving a 145 nm spectral separation, and further realized potential-regulated switch on/off emission of $\text{Ir}(\text{ppy})_3$ ($\lambda_{\text{max}} = 526$ nm), thus allowing for the potential-resolved ECL emission. As illustrated in Fig. 2b, applying

sequential potentials at +0.74 V, +1.33 V, and +1.75 V selectively activated each luminophore, enabling both spectral and potential resolution. Building on this principle, these three luminophores were loaded into polystyrene beads to encode detection antibodies for carcinoembryonic antigen (CEA), alpha-fetoprotein (AFP), and beta-human chorionic gonadotropin (β -hCG). This ultimately enabled the simultaneous identification of all three cancer markers in a single sample and scan.⁴⁵



Quantum dots (QDs) demonstrate the theoretical capability for simultaneous discrimination of four to five or even more targets, owing to their extremely narrow emission peaks, tunable wavelengths, and co-excitation capacity. As one of the earliest labeling materials applied in spectrum-resolved ECL immunoassays, QDs exhibit emission spectra covering the visible, near-infrared, and even far-infrared regions.⁴⁶ This excellent spectral tunability provides a crucial material foundation for the development of ECL-based multiplex detection strategies. Miao *et al.* developed dual-stabilizer-capped CdSe ($\lambda_{\text{max}} = 550$ nm) and CdTe ($\lambda_{\text{max}} = 776$ nm) nanocrystals as ECL labels for the simultaneous detection of AFP and CEA, with the limit of detection (LOD) of 10 fg mL^{-1} and 1 pg mL^{-1} , respectively. Using ammonium persulfate ($(\text{NH}_4)_2\text{S}_2\text{O}_8$) as a cathodic ECL co-reactant, the ECL signal was generated during cathodic potential scanning. The resulting ECL spectra from the two immobilized luminophores effectively eliminated cross-talk, thereby overcoming the spectral overlap inherent in traditional multiplex ECL assays.⁴⁷

Furthermore, Zou *et al.* reported a spectrally resolved three-color ECL multiplex immunoassay. Firstly, they synthesized dual-stabilizer-capped CdSe ($\lambda_{\text{max}} = 550$ nm),

CdTe ($\lambda_{\text{max}} = 650$ nm), and CdTe ($\lambda_{\text{max}} = 776$ nm) nanocrystals. By applying a potential in a buffer containing $(\text{NH}_4)_2\text{S}_2\text{O}_8$ and collecting emission intensities across wavelengths *via* a spectrometer, they successfully achieved simultaneous detection of three tumor markers, namely CEA, prostate-specific antigen (PSA), and AFP, on a single electrode interface, with LODs of 1, 10, and 0.01 pg mL^{-1} , respectively. With the ongoing development of new monochromatic ECL luminophores covering broader spectral ranges, this strategy could potentially be extended to detect four or more targets simultaneously.⁴⁸

To fundamentally solve the problems of low efficiency and poor stability of conventional QDs, innovative core/shell/shell structures (CdSe core–CdS intermediate layer–ZnS outer layer) were developed by Su and coworkers. These QDs maintained near-unity photoluminescence quantum yields ($>90\%$) and narrow emission peaks in aqueous solution. Their ECL efficiency was enhanced by six orders of magnitude compared to the conventional $\text{Ru}(\text{bpy})_3^{2+}$ emitter, and the relative standard deviation of the luminescence intensity remained below 1% over 125 consecutive potential scan cycles, demonstrating excellent operational stability. They precisely controlled the CdSe core size to prepare three-color




Fig. 3 (a) Normalized PL spectra and ECL photographs of green-, yellow-, and red-emitting CdSe/CdS/ZnS core/shell/shell QDs. (b) ECL-potential spectra of a mixed solution containing the three colors of QDs. (c–e) ECL spectra and corresponding photographs (insets) of the three-component mixed system at three different applied potentials. Reprinted with permission from ref. 49. Copyright 2020 American Chemical Society.



quantum dots emitting at 549 nm (green), 592 nm (yellow), and 643 nm (red), with emission wavelength intervals of approximately 50 nm (Fig. 3a). As illustrated in Fig. 3b, by mixing these luminophores in a phosphate-buffered solution containing $K_2S_2O_8$ and applying a cathodic potential sweep (-0.9 V to -1.3 V) on a single working electrode, the ECL emissions of the different QDs could be sequentially and selectively excited. The resulting signals were spectrally well-resolved with negligible crosstalk (Fig. 3c–e), providing a practical strategy for high-throughput detection of multiple biomarkers.⁴⁹

Due to the restricted spectral range of the visible region (400–750 nm), spectral multiplexing is constrained by the limited number of resolvable channels. The development of near-infrared (NIR) luminophores can extend the detectable spectral window, thereby increasing the number of resolvable channels for multiplex detection. Moreover, NIR

luminophores offer minimal autofluorescence, significantly reducing background noise and thus enhancing the signal-to-noise ratio compared to conventional UV-visible detection. However, most aqueous NIR ECL luminophores, such as graphene oxide and CdTe QDs, typically emit below 800 nm and suffer from substantial spectral overlap in multiplexed detection, compromising both sensitivity and accuracy. To address this issue, Shen *et al.* reported the fabrication of an innovative NIR ECL luminophore based on bovine serum albumin-stabilized silver nanoclusters (BSA-Ag NCs). When coupled with TiO_2 nanoparticles as a co-reactant accelerator, the luminophore enabled highly efficient anodic ECL emission at 904 nm in aqueous medium to surpass the 900 nm threshold. By tagging anti-cancer antigen 125 (anti-CA125), anti-carbohydrate antigen 19-9 (anti-CA19-9) and anti-cardiac troponin I (anti-cTnI) secondary antibodies with CdSe NCs ($\lambda_{max} = 553$ nm), CdTe NCs ($\lambda_{max} = 676$ nm) and




Fig. 4 (a) Spectrum-resolved ECL MIA toward samples containing CA125, CA19-9 and cTnI at different concentrations. Adapted with permission from ref. 50. Copyright 2020 Elsevier B.V. (b) Schematic illustration of the potential- and spectrum-resolved dual-mode ECL measurement process for the simultaneous detection of AFP, CA125, and CA19-9. Adapted with permission from ref. 52. Copyright 2024 Elsevier B.V.



the new BSA-Ag NCs ($\lambda_{\text{max}} = 904 \text{ nm}$), respectively, they achieved three baseline-resolved emission peaks in a single potential sweep, thus demonstrating a three-color, spectrally resolved ECL multiplex immunoassay on a single electrode surface (Fig. 4a).⁵⁰

Spectrum-resolved measurement using spectrometers often suffers from sensitivity loss due to the slits in the monochromator. Photomultiplier tubes (PMTs), by contrast, provide highly sensitive detection and enable spectral resolution without a sensitivity trade-off, when combined with optical filters. Zou *et al.* achieved spectrally resolved multiplexed ECL detection by combining two PMTs with a dichroic mirror. ECL emissions from two QDs, namely, CdSe (550 nm) and CdTe (790 nm), were completely separated by the dichroic mirror and measured by the two PMTs, respectively. As a result, this dual-band ECL detection platform exhibited LODs of 1 pg mL^{-1} for AFP and 1 mU mL^{-1} for CA125, avoiding spectral overlap and improving detection sensitivity.⁵¹ Very recently, Shen *et al.* reported a potential and spectrum dual-resolved ECL immunoassay based on a custom three-channel bandpass filter wheel (Fig. 4b). In a single cathodic sweep, perylene tetracarboxylic acid@carboxylic carbon nanotubes (710 nm) were first activated at lower potentials (0–0.8 V) for AFP detection. After rotating the filter wheel, CdSe (550 nm) and CdTe (750 nm) quantum dots were sequentially read out for spectrum-resolved detection of CA125 and CA19-9. The assay achieved detection limits of 0.53 fg mL^{-1} , 0.058 mU mL^{-1} , and 0.11 mU mL^{-1} for AFP, CA125, and CA19-9, respectively, which matched well with the performance of single-plex immunoassays. This platform eliminates the need for spectrometers and offers low-cost and high-throughput ECL detection.⁵²

In addition to the metal complexes and quantum dots discussed above, a variety of other materials have been synthesized for spectrum-resolved ECL immunoassays. For instance, metal-free carbon-based nanomaterials have attracted considerable attention due to their good

biocompatibility and tunable photoelectric properties.⁵³ Table 1 summarizes the recent advances in spectrum-resolved multiplex ECL immunoassays employing various luminophores. The analytical performances of these assays, including the limit of detection and linear ranges, are also compared.

3.2 Spatially resolved strategy

Besides the spectrum-resolved strategy, the spatially resolved strategy offers a promising alternative for achieving multiplexed immunoassays in a single run, when combined with the ECL imaging technology.⁵⁵ This strategy mainly lies in physically separating capture antibodies for different analytes onto independent, addressable spatial locations or carriers, enabling multiplexed detection using only a single and universal ECL luminophore. It exhibits three remarkable advantages: (i) it avoids spectral overlap and signal crosstalk caused by multiple luminophores. (ii) It reduces system complexity by eliminating the need to optimize specific labels and detection conditions for each analyte. (iii) It can overcome the throughput limitations imposed by spectral channels, because the multiplexing capacity depends on the number of precisely encoded spatial addresses. Therefore, this strategy serves as a straightforward and reliable means of expanding the multiplexing capacity of ECL immunoassays.⁵⁶

Microarrays are commonly used substrates for immobilizing different capture antibodies at defined locations, thus achieving spatial separation of detection sites. A typical example is the work reported by Pfeifer *et al.*, who constructed 50 spatially separated detection spots on a 4 mm diameter electrode (Fig. 5a–c), allowing for the simultaneous determination of two triplex biomarker panels for traumatic brain injury and cardiac diseases, and achieving detection limits of 1–33 pg mL^{-1} with good reproducibility and specificity (cross-reactivity <6%).⁵⁷ In addition, a triple ECL

Table 1 Summary of spectrum-resolved ECL multiplex immunoassays, including ECL luminophores and emission wavelengths, target biomarkers, limit of detection, and linear ranges

No.	ECL luminophores & $\lambda_{\text{em max}}$	Targets	Limit of detection	Linear range	Ref.
1	CdSe QDs; 550 nm	AFP	10 fg mL^{-1}	$50\text{--}1.0 \times 10^5 \text{ fg mL}^{-1}$	47
	CdTe QDs; 776 nm	CEA	1 pg mL^{-1}	$10\text{--}1.0 \times 10^4 \text{ pg mL}^{-1}$	
2	CdSe QDs; 550 nm	CEA	1.0 pg mL^{-1}	$0.01\text{--}10 \text{ ng mL}^{-1}$	48
	CdTe QDs; 650 nm	PSA	10.0 pg mL^{-1}	$0.1\text{--}100 \text{ ng mL}^{-1}$	
3	CdTe QDs; 776 nm	AFP	0.01 pg mL^{-1}	$0.1\text{--}100 \text{ pg mL}^{-1}$	50
	CdSe QDs; 553 nm	CA125	0.035 mU mL^{-1}	$0.1\text{--}1.0 \times 10^3 \text{ mU mL}^{-1}$	
	CdTe QDs; 676 nm	CA19-9	0.087 mU mL^{-1}	$0.3\text{--}1.0 \times 10^3 \text{ mU mL}^{-1}$	
4	BSA-Ag NCs; 904 nm	cTnI	0.016 pg mL^{-1}	$0.03\text{--}100 \text{ pg mL}^{-1}$	51
	CdSe QDs; 550 nm	AFP	1 pg mL^{-1}	$5\text{--}5.0 \times 10^3 \text{ pg mL}^{-1}$	
5	CdTe QDs; 790 nm	CA125	1 mU mL^{-1}	$5\text{--}1.0 \times 10^3 \text{ mU mL}^{-1}$	52
	CdSe QDs; 550 nm	CA125	0.058 mU mL^{-1}	$0.1\text{--}300 \text{ mU mL}^{-1}$	
6	CdTe QDs; 750 nm	CA19-9	0.11 mU mL^{-1}	$0.3\text{--}300 \text{ mU mL}^{-1}$	53
	PTCA@CNTs; 710 nm	AFP	0.53 fg mL^{-1}	$0.003\text{--}10 \text{ pg mL}^{-1}$	
7	CNNs; 438 nm	CA19-9	0.083 U mL^{-1}	$0.01\text{--}80 \text{ U mL}^{-1}$	54
	S-CNNs; 510 nm	Mesothelin	0.91 ng mL^{-1}	$0.001\text{--}50 \text{ } \mu\text{g mL}^{-1}$	
7	Zn-AIE-AuNCs; 485 nm	CEA	0.3 pg mL^{-1}	$1.0\text{--}5 \times 10^4 \text{ pg mL}^{-1}$	54
	CIS@ZnS nanocrystals; 775 nm	P53	0.5 pM	$1.0\text{--}5 \times 10^4 \text{ pM}$	






Fig. 5 (a) Schematic illustration of the disposable microfluidic cartridge and fluidic channel design. (b) Photograph of the cartridge held in hand. (c) Photograph of the cartridge loaded into the drawer of the ECL reader prior to measurement. Reprinted with permission from ref. 57. Copyright 2022 American Association for the Advancement of Science. (d) Electrode layout with an eight-channel circular array, a common cathode, a signal chamber, and an assay chamber. (e) Chip assembly: plastic fluidic layer bonded to a PCB, integrated magnet array, and pogo pins for electrical connection. (f) Signal detection: applied potential drives the ECL reaction in the signal chamber for real-time imaging using a CCD camera. Reprinted with permission from ref. 61. Copyright 2022 American Association for the Advancement of Science.

detection array was constructed by the combined use of disposable screen-printed carbon electrodes and a portable ECL reader, which enables the simultaneous detection of three biomarkers, namely GFAP, h-FABP, and S100 β , without cross-reactivity. In comparison with conventional imaging techniques such as CT and MRI, this approach offers advantages including simple operation, fast detection, and low cost, making it particularly suitable for resource-limited environments such as accident sites and remote areas.⁵⁸

The aforementioned studies usually employed nano-spotting or 3D printing to precisely immobilize or spatially separate capture antibodies on the substrate electrode. Alternatively, Cui *et al.* reported a different approach based on a single-electrode configuration for microarray fabrication, achieving simultaneous quantitative detection of acute myocardial infarction (AMI) biomarkers, such as cTnI, h-FABP, and copeptin. Validation with 260 clinical serum samples demonstrated 100% sensitivity and specificity, significantly surpassing the conventional high-sensitivity cardiac troponin T (hs-cTnT) assay, and reducing the misdiagnosis rate by over 20% within the critical 3–12 hour window after symptom onset. It thus provides a reliable measurement tool for the early and accurate diagnosis of AMI.⁵⁹

Furthermore, microfluidic chips enable automated sample and reagent delivery, making them a popular platform for multiplexed ECL immunoassays. Rusling and coworkers

fabricated a monolithic 3D-printed microfluidic chip integrated with an automated micro-pump controller. The assay can complete the simultaneous detection of eight prostate cancer biomarkers in human serum in only 25 minutes, with detection limits as low as 78–110 fg mL⁻¹, providing a powerful tool for rapid clinical multi-parameter diagnosis.⁶⁰ As illustrated in Fig. 5d–f, a novel sensing system termed the ECLipse (ECL in paired signal electrode) chip was constructed, in which the biorecognition and ECL generation processes were physically separated *via* bipolar electrodes (BPEs). In this modular design, the triangular tips of eight BPEs were positioned within a single signal chamber containing ECL luminophores, while biorecognition reactions occurred in eight independent assay chambers, thus allowing for the simultaneous detection of eight targets in a single run. Meanwhile, this chip achieved detection limits of 15–230 fg mL⁻¹ and an over 7000-fold increase in detection sensitivity compared to conventional ELISA.⁶¹

In contrast to planar microarrays or microfluidic chips that rely on precisely patterned electrodes or physically isolated channels to achieve spatial resolution, spot-free microbead-based arrays offer a fundamentally different and highly versatile approach for multiplexed ECL immunoassays. Rather than immobilizing capture antibodies on predefined two-dimensional substrates, microbeads function as addressable carriers for antibody immobilization. By encoding microbeads with fluorescent dyes, shapes, or diameters, bead-based ECL






Fig. 6 (a) Schematic of the bead-based multiplexed ECL immunoassay using a $\text{Ru}(\text{bpy})_3^{2+}/\text{TPrA}$ system with spatial resolution. Reprinted with permission from ref. 62. Copyright 2009 American Chemical Society. (b) Comparison of ECL generation efficiency between GMBs and MMBs under different pathways. Reprinted with permission from ref. 63. Copyright 2023 American Chemical Society.

immunoassays can achieve high-order multiplexing. As shown in Fig. 6a, Sojic and Walt demonstrated a bead-based microarray that combines fluorescence encoding with single-bead ECL imaging. Fluorescence encoding was achieved by doping beads with varying Eu^{3+} dye concentrations and conjugating them to specific capture antibodies, allowing each bead to be identified by its characteristic fluorescence intensity. After loading onto a gold-coated optical fiber bundle electrode microwell array, the encoded beads were subjected to a sandwich immunoassay using an ECL luminophore (SA–Ru complex), ultimately enabling spatially resolved ECL imaging at the single-bead level.⁶²

As previously reported, the main limitation in conventional bead-based ECL immunoassays originates from the low oxidation potential pathway.⁴⁰ Because the ECL-emitting region around the bead is determined by the diffusion distance of TPrA^+ , and its lifetime is only 0.2 ms, ECL generation is thus limited to a zone within 3 μm of the substrate electrode surface. To overcome this constraint, gold-coated microbeads (GMBs) were employed as conductive carriers and for constructing proximity ECL generation (Fig. 6b). As demonstrated, GMBs can function as spherical ultramicroelectrodes for direct co-reactant oxidation on their surface, thus shortening radical diffusion distances and enhancing ECL signals. When using CRP as a model analyte, the GMB-based ECL immunoassay

achieved a 4.96-fold increase in ECL intensity and a 21.7-fold enhancement in turnover frequency over conventional magnetic microbeads, consequently improving the detection limit from 0.35 ng mL^{-1} to 0.12 ng mL^{-1} . Moreover, because the ECL generation at GMBs is no longer restricted by the diffusion distance of TPrA^+ , the size-encoded multiplex immunoassay is eventually established for four AMI biomarkers, namely CRP, cTnI, FABP, and Myo.⁶³

As discussed above, the spatially resolved strategy can be implemented using planar microarrays, microfluidic chips, or encoded bead-based microarrays. Table 2 summarizes representative spatially resolved multiplexed ECL immunoassays, detailing their spatial encoding principles, multiplexing capacity, key innovations, detection performance, and commercial availability. These platforms enable multiplexing ranging from dual-plex to eight-plex detection and exhibit a sensitivity improvement of several orders of magnitude over conventional ELISA, with detection limits reaching fg mL^{-1} levels.

4. Conclusions and perspectives

In summary, this review has surveyed recent advances in ECL-based multiplexed immunoassays, with a focus on two primary strategies for achieving multiplex detection, namely,



Table 2 Summary of representative spatially resolved ECL immunoassays, detailing the spatial encoding principles, multiplexing capacity, innovations, detection performance, and commercial availability

Technology platform	Multiplexing capacity	Key innovation	Detection performance	Commercial availability	Ref.
Spatially resolved ECL immunoassay	3-Plex (2 panels of biomarkers)	Low-cost 3D-printed chip, physical isolation of detection zones	1–33 pg mL ⁻¹ , cross-reactivity < 6%	Partially available	57
Screen-printed electrode array	3-Plex	Portable ECL reader suitable for POC settings	GFAP: 6.94 pg mL ⁻¹ , h-FABP: 1.35 pg mL ⁻¹ , S100β: 15.73 pg mL ⁻¹	Fully available	58
Triplex diagnostic microchip	3-Plex	Chemiluminescence immuno-gold nanoassemblies (ciGold)	0.014–0.053 ng L ⁻¹ , clinical validation showed 100% sensitivity	Partially available	59
3D-printed microfluidic array	8-Plex	Automated integration of 3D-printed microfluidic chip	78–110 fg mL ⁻¹ , completed in 25 minutes	Partially available	60
ECLipse platform	8-Plex	Separation of biorecognition and signal generation regions using bipolar electrodes	15–230 fg mL ⁻¹ , sensitivity increased by 7000-fold	Partially available	61
Bead-based array	Multiplex (variable)	Single-bead-level ECL imaging, fluorescence encoding for bead identification	Single-molecule-level sensitivity	Fully available	62
Gold microbead-encoded platform	4-Plex	Proximity ECL, spherical ultramicroelectrodes generating reactive radicals autonomously	CRP: 0.12 ng mL ⁻¹	Fully available	63
Ultra-sensitive microfluidic immunoarray	4-Plex	Streptavidin–poly-HRP signal amplification	0.013–0.13 fg mL ⁻¹ , dynamic range spanning five orders of magnitude	Partially available	64
CRISPR-ECL array	3-Plex	Integration of the CRISPR/Cas13a system with a 3D-printed microfluidic array	Sub-fg mL ⁻¹ level sensitivity	Partially available	65
Spatial-potential resolved BPE-ECL system	2-Plex	Non-precious metal electrocatalyst CuHCF for signal amplification	miRNA-122: 40 fM; CEA: 3 fg mL ⁻¹	Fully available	66

the spectrum-resolved strategy and spatially resolved strategy. The former enables differentiation of multiple targets using multicolor luminophores, such as metal complexes and quantum dots, which emit at distinct wavelengths. The latter achieves multiplexing by physically separating detection sites using microarrays, microfluidic chips, or encoded microbeads. When combined with the ECL imaging technology, this strategy requires only a single ECL luminophore, thus overcoming the limitation of multiplexing capacity imposed by spectral overlap.

Thanks to the remarkable advantages of low background noise, high sensitivity, and high throughput, ECL immunoassays have emerged as a powerful method for multi-biomarker detection *in vitro* diagnostics and point-of-care testing. Despite significant progress in recent years, some challenges remain to be addressed. For example, although a variety of ECL luminophores have been synthesized, only Ru(bpy)₃²⁺ and its derivatives have found practical application. This limited selection severely restricts the clinical use of spectrum-resolved multiplexed immunoassays. In addition, although numerous innovative ECL assays based on spectrum- and/or spatially resolved strategies have been established at the methodological level, their validation using real samples containing complex matrices remains limited. Furthermore, standardized protocols for multiplexed ECL immunoassays are still lacking.

Future efforts should focus on improving the spatiotemporal resolution of ECL detection that offers a promising pathway to significantly improve analysis speed in multiplexed immunoassays. Nowadays, electron multiplying

charge coupled devices (EMCCDs) enable high-speed and real-time imaging of ECL emissions from single entities, while achieving microsecond temporal resolution and spatial resolution down to tens of nanometers.^{67,68} However, the imaging time for ECL immunoassays typically ranges from seconds to minutes. Achieving rapid signal acquisition would substantially reduce total assay time, which is a critical improvement for time-sensitive applications such as emergency diagnostics and point-of-care testing.

On the other hand, the application of artificial intelligence to ECL imaging and spectral analysis offers new opportunities for improving detection accuracy. Machine learning algorithms can assist in identifying spatial features, discriminating signals from background noise, and resolving overlapping emission spectra from multiple luminophores, thereby enabling reliable quantification. The combination of high-speed spatiotemporal imaging with intelligent data processing is expected to advance multiplexed ECL immunoassays toward rapid point-of-care testing, higher throughput, improved sensitivity and specificity, and greater automation.

Author contributions

Y. Liu: writing – original draft, investigation, visualization, writing – review & editing; C. Liu: writing – original draft, investigation; Z. Dai: supervision, project administration; W. Guo: writing – review and editing, supervision, project administration, funding acquisition. All authors have read and agreed to the published version of the manuscript.



Conflicts of interest

The authors declare no conflicts of interest.

Data availability

No primary research results, software or code have been included and no new data were generated or analysed as part of this review.

Acknowledgements

This work was financially supported by the National Natural Science Foundation of China (22374078 and 22522405) and the Qing Lan Project of Jiangsu Province of China.

References

- 1 A. D. Warren, S. T. Gaylord, K. C. Ngan, M. Dumont Milutinovic, G. A. Kwong, S. N. Bhatia and D. R. Walt, *J. Am. Chem. Soc.*, 2014, **136**, 13709–13714.
- 2 A. Roointan, T. Ahmad Mir, S. Ibrahim Wani, M. Ur Rehman, K. K. Hussain, B. Ahmed, S. Abraham, A. Savardashtaki, G. Gandomani, M. Gandomani, R. Chinnappan and M. H. Akhtar, *J. Pharm. Biomed. Anal.*, 2019, **164**, 93–103.
- 3 S. Wang, L. Ge, Y. Zhang, X. Song, N. Li, S. Ge and J. Yu, *Lab Chip*, 2012, **12**, 4489–4498.
- 4 W. Miao and A. J. Bard, *Anal. Chem.*, 2004, **76**, 7109–7113.
- 5 J. Tian, L. Zhou, Y. Zhao, Y. Wang, Y. Peng and S. Zhao, *Talanta*, 2012, **92**, 72–77.
- 6 B. S. Munge, T. Stracensky, K. Gamez, D. DiBiase and J. F. Rusling, *Electroanalysis*, 2016, **28**, 2644–2658.
- 7 P. S. Pakchin, S. A. Nakhjavani, R. Saber, H. Ghanbari and Y. Omid, *TrAC, Trends Anal. Chem.*, 2017, **92**, 32–41.
- 8 R. Liu, S. Zhang, C. Wei, Z. Xing, S. Zhang and X. Zhang, *Acc. Chem. Res.*, 2016, **49**, 775–783.
- 9 Z. Yang, C. Zong, F. Yan and H. Ju, *Talanta*, 2010, **82**, 1462–1467.
- 10 S. Otsuki and M. Ishikawa, *Biosens. Bioelectron.*, 2010, **26**, 202–206.
- 11 Z. Cao, H. Li, C. Lau and Y. Zhang, *Anal. Chim. Acta*, 2011, **698**, 44–50.
- 12 J. Li, J. Wang, Y. S. Grewal, C. B. Howard, L. J. Raftery, S. Mahler, Y. Wang and M. Trau, *Anal. Chem.*, 2018, **90**, 10377–10384.
- 13 Z. Liu, Q. Rong, Z. Ma and H. Han, *Biosens. Bioelectron.*, 2015, **65**, 307–313.
- 14 W. Miao, *Chem. Rev.*, 2008, **108**, 2506–2553.
- 15 L. Hu and G. Xu, *Chem. Soc. Rev.*, 2010, **39**, 3275–3304.
- 16 H. Qi and C. Zhang, *Anal. Chem.*, 2020, **92**, 524–534.
- 17 M. M. Richter, *Chem. Rev.*, 2004, **104**, 3003–3036.
- 18 Y. Wang, W. Guo, Q. Yang and B. Su, *J. Am. Chem. Soc.*, 2020, **142**, 1222–1226.
- 19 R. E. Visco and E. A. Chandross, *J. Am. Chem. Soc.*, 1964, **86**, 5350–5351.
- 20 F. Kanoufi and A. J. Bard, *J. Phys. Chem. B*, 1999, **103**, 10469–10480.
- 21 M.-F. Sun, J.-L. Liu, Y.-Q. Chai, J. Zhang, Y. Tang and R. Yuan, *Anal. Chem.*, 2019, **91**, 7765–7773.
- 22 M. M. Richter, *Chem. Rev.*, 2004, **104**, 3003–3036.
- 23 K. S. V. Santhanam and A. J. Bard, *J. Am. Chem. Soc.*, 1965, **87**, 139–140.
- 24 R. E. Visco and E. A. Chandross, *J. Am. Chem. Soc.*, 1964, **86**, 5350–5351.
- 25 D. L. Maricle and A. Maurer, *J. Am. Chem. Soc.*, 1967, **89**, 188–189.
- 26 R. Bezman and L. R. Faulkner, *J. Am. Chem. Soc.*, 1972, **94**, 6324–6330.
- 27 W. Miao, *Chem. Rev.*, 2008, **108**, 2506–2553.
- 28 P. Bertonecello and R. J. Forster, *Biosens. Bioelectron.*, 2009, **24**, 3191–3200.
- 29 R. S. Glass and L. R. Faulkner, *J. Phys. Chem.*, 1981, **85**, 1160–1165.
- 30 R. Pyati and M. M. Richter, *Annu. Rep. Prog. Chem., Sect. C: Phys. Chem.*, 2007, **103**, 12–78.
- 31 E. L. Ritchie, P. Pastore and R. M. Wightman, *J. Am. Chem. Soc.*, 1997, **119**, 11920–11925.
- 32 H. S. White and A. J. Bard, *J. Am. Chem. Soc.*, 1982, **104**, 6891–6895.
- 33 P. J. Smith and C. K. Mann, *J. Org. Chem.*, 1969, **34**, 1821–1826.
- 34 R. Y. Lai and A. J. Bard, *J. Phys. Chem. A*, 2003, **107**, 3335–3340.
- 35 R. D. Gerardi, N. W. Barnett and S. W. Lewis, *Anal. Chim. Acta*, 1999, **378**, 1–41.
- 36 W. Miao, J.-P. Choi and A. J. Bard, *J. Am. Chem. Soc.*, 2002, **124**, 14478–14485.
- 37 R. Memming, *J. Electrochem. Soc.*, 1969, **116**, 785.
- 38 R. Qiu, X. Zhang, H. Luo and Y. Shao, *Chem. Sci.*, 2016, **7**, 6684–6688.
- 39 J. Liu, K. Yu, H. Zhang, J. He, J. Jiang and H. Luo, *Chem. Sci.*, 2021, **12**, 9494–9499.
- 40 M. Sentic, M. Milutinovic, F. Kanoufi, D. Manojlovic, S. Arbault and N. Sojic, *Chem. Sci.*, 2014, **5**, 2568–2572.
- 41 Y. Wang, W. Guo, Q. Yang and B. Su, *J. Am. Chem. Soc.*, 2020, **142**, 1222–1226.
- 42 Y. Wang, J. Ding, P. Zhou, J. Liu, Z. Qiao, K. Yu, J. Jiang and B. Su, *Angew. Chem., Int. Ed.*, 2023, **62**, e202216525.
- 43 D. Bruce, M. M. Richter and K. J. Brewer, *Anal. Chem.*, 2002, **74**, 3157–3159.
- 44 D. Bruce and M. M. Richter, *Anal. Chem.*, 2002, **74**, 1340–1342.
- 45 W. Guo, H. Ding, C. Gu, Y. Liu, X. Jiang, B. Su and Y. Shao, *J. Am. Chem. Soc.*, 2018, **140**, 15904–15915.
- 46 Z. Guo, T. Hao, S. Du, B. Chen, Z. Wang, X. Li and S. Wang, *Biosens. Bioelectron.*, 2013, **44**, 101–107.
- 47 G. Zou, X. Tan, X. Long, Y. He and W. Miao, *Anal. Chem.*, 2017, **89**, 13024–13029.
- 48 J. Zhou, L. Nie, B. Zhang and G. Zou, *Anal. Chem.*, 2018, **90**, 12361–12365.
- 49 Z. Cao, Y. Shu, H. Qin, B. Su and X. Peng, *ACS Cent. Sci.*, 2020, **6**, 1129–1137.
- 50 L. Yu, M. Li, Q. Kang, L. Fu, G. Zou and D. Shen, *Biosens. Bioelectron.*, 2021, **176**, 112934.



- 51 F. Zhang, Y. He, K. Fu, L. Fu, B. Zhang, H. Wang and G. Zou, *Biosens. Bioelectron.*, 2018, **115**, 77–82.
- 52 J. Ge, J. Dou, X. Yu, Y. Sun, H. Song and D. Shen, *Microchem. J.*, 2024, **199**, 110117.
- 53 Y. Lv, Z. Zhou, Y. Shen, Q. Zhou, J. Ji, S. Liu and Y. Zhang, *ACS Sens.*, 2018, **3**, 1362–1367.
- 54 X. Gao, X. Liu, Y. Zeng, Q. Zhang, B. Zhang and G. Zou, *Anal. Chem.*, 2022, **94**, 15801–15808.
- 55 J. D. Hoheisel, *Nat. Rev. Genet.*, 2006, **7**, 200–210.
- 56 Y. Zhang, W. Liu, S. Ge, M. Yan, S. Wang, J. Yu, N. Li and X. Song, *Biosens. Bioelectron.*, 2013, **41**, 684–690.
- 57 M. Jović, D. Prim, G. Paciotti and M. E. Pfeifer, *Lab Chip*, 2025, **25**, 5428–5438.
- 58 M. Jović, D. Prim, E. Saini and M. E. Pfeifer, *Biosensors*, 2022, **12**, 172.
- 59 M. Guo, D. Du, J. Wang, Y. Ma, D. Yang, M. A. Haghghatbin, J. Shu, W. Nie, R. Zhang, Z. Bian, L. Wang, Z. J. Smith and H. Cui, *Chem. Biomed. Imaging*, 2023, **1**, 179–185.
- 60 K. Kadimisetty, S. Malla, K. S. Bhalerao, I. M. Mosa, S. Bhakta, N. H. Lee and J. F. Rusling, *Anal. Chem.*, 2018, **90**, 7569–7577.
- 61 Y. K. Cho, H. Kim, A. Bénard, H.-K. Woo, F. Czubyko, P. David, F. J. Hansen, J. I. Lee, J. H. Park, E. Schneck, G. F. Weber, I.-S. Shin and H. Lee, *Sci. Adv.*, 2022, **8**, eabq4022.
- 62 F. Deiss, C. N. LaFratta, M. Symer, T. M. Blicharz, N. Sojic and D. R. Walt, *J. Am. Chem. Soc.*, 2009, **131**, 6088–6089.
- 63 X. Yang, J. Hang, W. Qu, Y. Wang, L. Wang, P. Zhou, H. Ding, B. Su, J. Lei, W. Guo and Z. Dai, *J. Am. Chem. Soc.*, 2023, **145**, 16026–16036.
- 64 L. Dhanapala, A. L. Jones, P. Czarnecki and J. F. Rusling, *Anal. Chem.*, 2020, **92**, 8021–8025.
- 65 K. Hiniduma, P. I. T. De Silva, R. Canete, P. Vora, H. Gunathillaka, O. Clement, S. M. Shawky, J. L. Rouge, I. M. Mosa, D. C. Steffens, K. Manning, D. Breno and J. F. Rusling, *Biosens. Bioelectron.*, 2025, **289**, 117855.
- 66 H. Li, Q. Cai, Z. Li, G. Jie and H. Zhou, *Biosens. Bioelectron.*, 2024, **255**, 116258.
- 67 J. Dong, Y. Lu, Y. Xu, F. Chen, J. Yang, Y. Chen and J. Feng, *Nature*, 2021, **596**, 244–249.
- 68 D. Ye, Y.-C. Kong, C.-H. Xu, Z. Ma, Y. Xiong, H. Zhao and W. Zhao, *J. Am. Chem. Soc.*, 2026, **148**(1), 1457–1464.

