

# ChemComm

Accepted Manuscript



This is an *Accepted Manuscript*, which has been through the RSC Publishing peer review process and has been accepted for publication.

*Accepted Manuscripts* are published online shortly after acceptance, which is prior to technical editing, formatting and proof reading. This free service from RSC Publishing allows authors to make their results available to the community, in citable form, before publication of the edited article. This *Accepted Manuscript* will be replaced by the edited and formatted *Advance Article* as soon as this is available.

To cite this manuscript please use its permanent Digital Object Identifier (DOI®), which is identical for all formats of publication.

More information about *Accepted Manuscripts* can be found in the [Information for Authors](#).

Please note that technical editing may introduce minor changes to the text and/or graphics contained in the manuscript submitted by the author(s) which may alter content, and that the standard [Terms & Conditions](#) and the [ethical guidelines](#) that apply to the journal are still applicable. In no event shall the RSC be held responsible for any errors or omissions in these *Accepted Manuscript* manuscripts or any consequences arising from the use of any information contained in them.

## COMMUNICATION

# Detection of Ovarian Cancer Biomarker CA-125 by Chemiluminescence Resonance Energy Transfer to Graphene Quantum Dots

Cite this: DOI: 10.1039/x0xx00000x

Received 00th January 2012,  
Accepted 00th January 2012Israa Al-Ogaidi,<sup>†,a,b</sup> Honglei Gou,<sup>†,b</sup> Zoraida P. Aguilar,<sup>c</sup> Shouwu Guo,<sup>d</sup> Alice K. Melconian,<sup>a</sup> Abdul Kareem A. Al-kazaz,<sup>a</sup> Fanke Meng,<sup>b</sup> and Nianqiang Wu<sup>\*,b</sup>

DOI: 10.1039/x0xx00000x

www.rsc.org/

**An immunoassay has been developed for detection of ovarian cancer biomarker CA-125 by utilizing the chemiluminescence resonance energy transfer to graphene quantum dots. This biosensor shows a wide linear range from 0.1 U/mL to 600 U/mL with a limit of detection of 0.05 U/mL for CA-125 in a buffer solution.**

Ovarian cancer is among the fatal gynecological malignancies in women with a total 22,240 new cases and 14,030 deaths anticipated in U.S. in 2013 [1]. It accounts for about 3% of all cancers among women and 5% of cancer deaths, which causes more deaths than any other cancer of female reproductive system [1]. Nowadays, two clinical methods are normally used for diagnosis of ovarian cancer, that is, the transvaginal ultrasonography and the blood measurement of tumor marker CA-125. CA-125 is a biomolecule found in serum, which is associated with ovarian cancer. It has a threshold of 35 U/mL in normal human serum [2]. Elevated level of CA-125 has been found in 80–90% of women with advanced-stage ovarian cancer, however, that is only 47% in patients with early-stage cancer [3]. Although single protein biomarker such as CA-125 is not ideal for early detection, the combination of CA-125 with other protein biomarkers is effective in early detection of ovarian cancer [4,5]. Therefore, CA-125 test has been approved by the U.S. Food and Drug Administration (FDA) for monitoring the chemotherapy response, determining the prognosis and examining the recurrence, as well as the early-stage detection of ovarian cancer [6].

Enzyme-linked immunosorbent assay (ELISA) is generally used for detection of biomarkers [7,8], such as CA-125. Despite its high sensitivity, it requires several separation steps with strict washing processes, which is a time-consuming and tedious process. Additionally, various methods based on electrochemistry [9–13], surface plasmon resonance [13,14], fluorescence [15] and mass spectrometry [16] have been developed for detection of CA-125. However, most of these methods still need thorough washing steps. Moreover, owing to the complexity of the processes involved in these methods, professional skills still are needed for operation and

data interpretation, which makes it impractical for point-of-care (POC) diagnosis. Therefore, it is necessary to develop simple and effective portable sensors for CA-125 detection.

It is worth noting that optical biosensors have been constructed using the resonance energy transfer processes (RET) to gold nanoparticles [17–23] or to graphene oxide (GO) [24–28]. Our previous studies have shown that fluorescence quenching by ultra-small gold nanoparticles or by GO follows a nanometal surface energy transfer (NSET) mechanism instead of a Förster resonance energy transfer (FRET) process [17,18,24,28]. Compared to FRET, NSET has a longer effective energy-transfer distance [17,28], which provides the flexibility for construction of biosensors. However, for most of NSET- or FRET-based sensors, an external excitation source is required to excite the energy transfer processes from the donor (made of organic dyes or quantum dots) to the acceptor. To eliminate the external excitation source, the donor made of organic dyes or quantum dots can be replaced with a chemiluminescence reagent, which enables the chemiluminescence resonance energy transfer (CRET) process [28–34]. Since an excitation laser is not needed for CRET-based sensors, the signal interferences from the background fluorescence that could be triggered by an external excitation source can be avoided, which can lower the signal-to-noise ratio and improve the sensitivity of sensors [28,29,31].

In this study, a simple and sensitive sensor is developed for detection of ovarian cancer biomarker CA-125. In this sensor, graphene quantum dots (GQDs) are immobilized on a functionalized glass chip. Signal transduction is based on the CRET from the chemiluminescence reagent soluble in the aqueous assay to the GQDs on the solid chip. The use of GQDs as the energy acceptor avoids the photo-bleaching problem, which usually is associated with organic dyes. In addition, the use of GQDs enables the NSET mechanism [28], which does not require the spectral overlap between the energy donor and the acceptor. This provides more flexibility for selection of the energy donors. Moreover, GQDs are typically hydrophilic, and rich in carboxylic acid moiety, which is convenient for bioconjugation. The capture antibody is linked to the GQDs on a transparent solid substrate, which has a potential for

further development of high-throughput and automated sensor chips. Furthermore, less washing steps are needed during operation of the sensing system, which greatly simplifies the operation of the sensing assay.

The assembly of the chemiluminescent immuno-chip and the detection principle were shown in Figure 1. The detailed characterization results obtained from each step can be found in Supplementary Information. The amino-modified glass chips were silanized with the 3-aminopropyl-trimethoxysilane (APTMS) layer

(Figure 1(a)). The GQDs were then immobilized via the electrostatic attraction to the positively charged amine group in the well array confined in the polydimethylsiloxane (PDMS) stencil (Figure 1(b)). The capture antibody (cAb) that was specific to the CA-125 antigen was covalently linked to the GQDs through amide conjugation [35] (Figure 1(c)). Bovine serum albumin (BSA) was then used to block the unreacted sites on the glass surface, forming the GQDs~cAb chip (Figure 1(d)).




Figure 1 Scheme of the assembly of the immunoassay and the detection principle.

When CA-125 antigen is absent in the immunoassay (Figure 1(h)), The horseradish peroxidase (HRP) enzyme catalyzes the production of the reactive oxygen species (ROS) from the  $H_2O_2$ , which oxidizes luminol to singlet dianion, generating the excited electrons. When the electrons jumps from the excited state to the ground state (that is, the singlet dianion becomes the ground-state dianion), chemiluminescence is emitted [36]; and the intensity of the emitted blue light is recorded with a fluorescence plate reader. In the absence of CA-125 antigen, the HRP-labeled antibody (Ab~HRP) is far away from the GQDs due to the electrostatic repulsion between the Ab~HRP and the capture antibody (cAb) because they had the same charge polarity. The chemical reaction is catalyzed by the HRP; and the reaction products are close to the HRP. Hence the dianion is also far away from the GQDs. As a result, there is no strong interaction between the dianion and the GQDs.

When CA-125 antigen is present in the immunoassay (Figure 1(e)), the antibody-antigen complex is generated, forming the GQDs~cAb+CA-125. To enable the detection, this complex is exposed to the Ab~HRP to form the sandwich structure, GQDs~cAb+CA-125+Ab~HRP (Figure 1(f)). Owing to the sandwich structure, HRP is in proximity to the GQDs. Dianion catalyzed by the HRP enzyme is close to the GQDs, which enables the resonance energy transfer from dianion to the GQDs, quenching the chemiluminescence (Figure 1(g)). The higher the concentration of antigen captured, the higher the concentration of Ab~HRP bound near the GQDs, which causes more energy transfer events that cause

the chemiluminescence quenching. Hence, the chemiluminescence intensity is inversely proportional to the CA-125 concentration.




Figure 2 Chemiluminescence measured from the immunoassay containing various concentrations of CA-125 antigen, showing the linear range.

Based on the above sensing principle, a calibration curve (Figure S5) was established from the immunoassay containing various concentrations of CA-125 antigen as  $y=0.197x+0.290$ , ( $R^2=97.0\%$ ), where  $x$  is the logarithmic concentration of antigen; and  $y=(C_0-C)/C_0$ . Herein  $C$  and  $C_0$  are the chemiluminescence intensity of the assay in the presence and in the absence of CA-125, respectively. The

normalized chemiluminescence intensity ( $y$ ) exhibited a linear correlation over a wide range of CA-125 concentration from 0.1 U/mL to 600 U/mL (Figure 2). The limit of detection (LOD) was estimated to be 0.05 U/mL at a signal-to-noise ratio of 3. The LOD value was far below the threshold of CA-125 (35 U/mL) in normal human serum [2]. It was comparable to the LOD values obtained from ELISA (0.05-0.3 U/mL [7]), and better than the electrochemical sensors (1.8 U/mL [10], 1.29 U/mL [11], 0.5 U/mL [12]) and the SPR sensor (0.4 U/mL [14]).

In order to evaluate the sensor performance in real-world samples, the chemiluminescence signal was recorded after capture of the antigen from the sample matrix containing 50% human blood plasma and 50% PBS buffer (Figure S6). The calibration curve showed a linear range from 0.1 U/mL to 600 U/mL even though the slope was a little lower than that obtained in the PBS buffer ( $y=0.183x+0.245$ ,  $R^2=99.6\%$ ), giving a LOD of 0.08 U/mL at a signal-to-noise ratio of 3.

In summary, a rapid and sensitive immunoassay was developed for detection of CA-125 via chemiluminescence resonance energy transfer to the graphene quantum dots. The immunoassay exhibited a LOD as low as 0.05 U/mL and a wide linear range from 0.1 U/mL to 600 U/mL for CA-125 in the buffer solution. It is an excellent substitute for FRET-based assay because it eliminates the need for an external excitation source. This immunoassay exhibited a similar performance in the presence of 50% blood plasma with a LOD of 0.08 U/mL. The as-developed sensing platform can be further fabricated in an array chip-based device for high throughput multiplex detection.

The resource and facilities used were partially supported by NSF (EPS 1003907), the West Virginia University Research Corporation, and the West Virginia EPSCoR Office.

## Notes and references

<sup>a</sup>Department of Biotechnology, College of Science, University of Baghdad, Baghdad, Iraq.

<sup>b</sup>Department of Mechanical and Aerospace Engineering, West Virginia University, Morgantown, WV 26506-6106, USA.

E-mail: nick.wu@mail.wvu.edu; Fax: +1-304-293-6689

<sup>c</sup>Zystein, LLC., PO Box 10071, Fayetteville, Arkansas 72703, USA

<sup>d</sup>Key Laboratory for Thin Film and Microfabrication of the Ministry of Education, Research Institute of Micro/Nano Science and Technology, Shanghai Jiao Tong University, Shanghai 200240, China

†These authors contributed to this work equally

- American Cancer Society, *Cancer Facts & Figures 2013*, Atlanta, GA, USA, 2013.
- A. Santillan, R. Garg, M. L. Zahurak, G. J. Gardner, R. L. Giuntoli II, D. K. Armstrong and R. E. Bristow, *J. Clin. Oncol.*, 2005, **23**, 9338-9343.
- L. Nguyen, S. J. Cardenas-Goicoechea, P. Gordon, C. Curtin, M. Momeni, L. Chuang and D. Fishman, *Women's Health*, 2013, **9**, 171-187.
- V. Nossov, M. Amneus, F. Su, J. Lang, J. M. Janco, S. T. Reddy, R. Farias-Eisner, *Am. J. Obstet. Gynecol.*, 2008, **199**, 215-223.
- V. Nossov, M. Amneus, F. Su, J. Lang, J. M. Janco, S. T. Reddy, R. Farias-Eisner, *Am. J. Obstet. Gynecol.*, 2008, **199**, 215-223.
- B. J. D. Rein, S. Gupta, R. Dada, A. Agarwal, J. Safi and C. Michener, *J. Oncol.*, 2011, **2011**, 475983.
- N. Scholler, M. Crawford, A. Sato, C. W. Drescher, K. C. O'Brian, N. Kiviat, G. L. Anderson and N. Urban, *Clin. Cancer. Res.*, 2006, **12**, 2117-2124.
- Y. W. Kim, S. M. Bae, H. Lim, Y. J. Kim and W. S. Ahn, *PLoS One*, 2012, **7**, e44960.
- I. Al-Ogaidi, Z. P. Aguilar, S. Suri, H. L. Gou and N. Q. Wu, *Analyst*, 2013, **138**, 5647-5653.
- D. Tang, R. Yuan and Y. Chai, *Anal. Chim. Acta*, 2006, **564**, 158-165.
- Z. Dai, F. Yan, J. Chen and H. Ju, *Anal. Chem.*, 2003, **75**, 5429-5434.
- S. Viswanathan, C. Rani, S. Ribeiro and C. Delerue-Matos, *Biosens. Bioelectron.*, 2012, **33**, 179-183.
- S. Suwansa-ard, P. Kanatharana, P. Asawatreratanakul, B. Wongkittisuksa, C. Limsakul and P. Thavarungkul, *Biosens. Bioelectron.*, 2009, **24**, 3436-3441.
- K. Zhang and X. Shen, *Analyst*, 2013, **138**, 1828-1834.
- H. Z. Wang, H. Y. Wang, R. Q. Liang and K. C. Ruan, *Acta. Biochim. Biophys. Sin (Shanghai)*, 2004, **36**, 681-686.
- F. Weiland, K. Fritz, M. K. Oehler and P. Hoffmann, *Int. J. Mol. Sci.*, 2012, **13**, 9942-58.
- M. Li, S. K. Cushing, Q. Wang, X. Shi, L. A. Hornak, Z. Hong and N. Q. Wu, *J. Phys. Chem. Lett.*, 2011, **2**, 2125-2129.
- M. Li, Q. Wang, X. Shi, L. A. Hornak and N. Q. Wu, *Anal. Chem.*, 2011, **83**, 7061-7065.
- X. Huang, T. Lan, B. Zhang and J. Ren, *Analyst*, 2012, **137**, 3659-66.
- E. Oh, M. Y. Hong, D. Lee, S. H. Nam, H. C. Yoon and H. S. Kim, *J. Am. Chem. Soc.*, 2005, **127**, 3270-3271.
- P. C. Ray, G. K. Darbha, A. Ray, J. Walker and W. Hardy, *Plasmonics*, 2007, **2**, 173-183.
- K. Aslan, J. Zhang, J. R. Lakowicz and C. D. Geddes, *Fluoresc.*, 2004, **14**, 391-400.
- N. T. Chen, S. H. Cheng, C. P. Liu, J. S. Souris, C. T. Chen, C. Y. Mou and L. W. Lo, *Int. J. Mol. Sci.*, 2012, **13**, 16598-16623.
- M. Li, X. Zhou, S. Guo and N. Q. Wu, *Biosens. Bioelectron.*, 2013, **43**, 69-74.
- H. Chang, L. Tang, Y. Wang, J. Jiang and J. Li, *Anal. Chem.*, 2010, **82**, 2341-2346.
- H. Dong, W. Gao, F. Yan, H. Ji and H. Ju, *Anal. Chem.*, 2010, **82**, 5511-5517.
- C. Zhang, Y. Yuan, S. Zhang, Y. Wang and Z. Liu, *Angew. Chem. Int. Ed. Engl.*, 2011, **50**, 6851-6854.
- M. Li, H. L. Gou, I. Al-Ogaidi and N. Q. Wu, *ACS Sustainable Chem. Eng.*, 2013, **1**, 713-723.
- J. S. Lee, H. A. Joung, M. G. Kim and C. B. Park, *ACS Nano*, 2012, **6**, 2978-2983.
- C. Chen and B. Li, *J. Mater. Chem. B*, 2013, **1**, 2476-2481.
- X. Huang, L. Li, H. Qian, C. Dong and J. Ren, *Angew. Chem. Int. Ed. Engl.*, 2006, **45**, 5140-5143.
- R. Freeman, X. Liu and I. Willner, *J. Am. Chem. Soc.*, 2011, **133**, 11597-11604.
- X. Huang and J. Ren, *Anal. Chim. Acta.*, 2011, **686**, 115-120.
- X. Huang and J. Ren, *Trends Anal. Chem.*, 2012, **40**, 77-89.
- M. Li, S. K. Cushing, J. Zhang, S. Suri, R. Evans, W. P. Petros, L. F. Gibson, D. Ma, Y. Liu and N. Q. Wu, *ACS Nano*, 2013, **7**, 4967-4976.
- C. Dodeigne, L. Thunus and R. Lejeune, *Talanta*, 2000, **51**, 415-439.