ChemComm

Accepted Manuscript



This is an *Accepted Manuscript*, which has been through the Royal Society of Chemistry peer review process and has been accepted for publication.

Accepted Manuscripts are published online shortly after acceptance, before technical editing, formatting and proof reading. Using this free service, authors can make their results available to the community, in citable form, before we publish the edited article. We will replace this Accepted Manuscript with the edited and formatted Advance Article as soon as it is available.

You can find more information about *Accepted Manuscripts* in the **Information for Authors**.

Please note that technical editing may introduce minor changes to the text and/or graphics, which may alter content. The journal's standard <u>Terms & Conditions</u> and the <u>Ethical guidelines</u> still apply. In no event shall the Royal Society of Chemistry be held responsible for any errors or omissions in this *Accepted Manuscript* or any consequences arising from the use of any information it contains.



Cite this: DOI: 10.1039/c0xx00000x

www.rsc.org/xxxxxx

ARTICLE TYPE

Vesicular Aptasensor for the Detection of Thrombin

Andreas Müller and Burkhard König*

Received (in XXX, XXX) Xth XXXXXXXXX 20XX, Accepted Xth XXXXXXXXX 20XX DOI: 10.1039/b000000x

5 Self-assembled phospholipid vesicles are functionalized with thrombin-binding aptamers using a thiol-click reaction. The resulting aptasensors signal the binding of the analyte to the vesicle surface by changes of the emission properties of membrane co-embedded reporter dyes.

10 The last two decades have seen considerable progress in the design of biopolymer-based receptors for high-affinity and -specificity analyte binding. The first reports on the evolutionary selection of RNA molecules binding to specific ligands in 1990¹ laid the foundation for the development of numerousness nucleic 15 acid aptamers. A broad variety of ligands were targeted ranging from small organic molecules^{2a} to proteins^{2b} and pathogens^{2c} and the technique was applied in environmental and food analytics.³ With an increasing number of available aptamers, suitable methods for transforming the microscopic binding event into a 20 macroscopic signal were established. Fluorescence-based and electrochemical^{5b} detection methods are among the most important ones due to their high sensitivities. They both typically rely on covalent modification of the aptamers with a luminescent or an electroactive moiety. Another commonly used technique is 25 based on hybridization of the aptamers with antisense oligonucleotides whose displacement by the target analytes induces a fluorescent or electrical response. The requirement of covalent reporter attachment at the aptamer or the need of accordingly labeled complementary oligonucleotide strands can 30 make such assays laborious and costly. For that reason, label-free strategies were investigated, but their general applicability is often limited as they rely on specific structural characteristics of the applied aptamers such as certain folding modes upon binding to the target. Formation of a double helix domain, for example, 35 allows the intercalation of a suitable dye accompanied by a change of its emission properties. Previously, our group reported on the modular construction of luminescent chemosensors by means of unilamellar vesicular membranes as self-assembled supporting frames for metal complex receptors and fluorescent 40 reporter sites. This strategy facilitates an effortless sensor preparation based on self-assembly by simple mixing of different functional amphiphiles in aqueous solution. As a consequence, our method allows an easy variation of the sensor composition, i.e., of the binding and signaling units, their ratios and 45 concentration on the surface and the physical properties of the supporting vesicle membranes determined by the nature of the phospholipids. We now expand the pool of available binding

origin. As a model system, we immobilized the very well 50 established thrombin-binding 15-mer DNA aptamer¹⁰ on fluorescent phospholipid vesicle surfaces. DNA-decorated liposomes have been used as signal amplification tags, 11 for controlled liposome adhesion or fusion¹² and for specific cell targeting with regard to drug delivery. 13 Moreover, the 55 conjugated polymeric backbone of oligonucleotide-functionalized polydiacetylene vesicles was utilized as colorimetric sensing matrix for target detection.¹⁴ These reports prompted us to combine the concept of liposome-anchored aptamers with our strategy of non-covalent receptor and fluorophor co-embedding 60 into phospholipid membranes.

As attachment site for thiol-modified aptamers, we incorporated the amphiphilic maleimide Mal-C₁₆ (Fig. 1, 5 mol-%) into membranes of small unilamellar DSPC vesicles. Amphiphilic pyrene Pyr- C_{18} (\rightarrow vesicles V-Pyr) or rhodamine ₆₅ **Rho-C**₁₈¹⁵ (\rightarrow vesicles **V-Rho**) were co-embedded as reporter dyes (1-5 mol-%). The synthesis of Mal-C₁₆ is described in the ESI†, preparation of Pyr-C₁₈ was performed by in situ click reaction of commercially available N-(1-pyrenyl)maleimide and 1-octadecanethiol. The vesicle preparation was performed in 70 25 mm HEPES buffer at pH = 7.4 according to previously reported procedures (see ESI† for details).





Fig. 1 Structures of the functional amphiphiles used for vesicle preparation and of the immobilized aptamers.

Next, we functionalized the vesicle membrane covalently with the aptamer by nucleophilic thiol-maleimide addition. After addition of thiolated Apt-Thr (0.4-0.5 equivalents with respect to the total amount of Mal-C₁₆) to V-Pyr, we observed a timedependent decrease of the fluorescence signal indicating the 80 reaction progress on the vesicle surface. Phase separation processes in the membrane result in high local concentrations of the embedded functional amphiphiles as it is evidenced by the

sites to more complex structures based on molecules of biological

presence of the pyrene excimer signal of V-Pyr at 470 nm (Fig. 2A, inset). We assume that the surface-anchored oligonucleotides strongly impact the physical properties at the membrane interface by their high negative charge and steric demand. This local 5 interference results in a partial fluorescence emission quenching of the nearby pyrene aggregates (Fig. 3, top). Moreover, it is well documented that nucleotides are able to quench the fluorescence of fluorophores, such as pyrene¹⁶ or rhodamine¹⁷ derivatives, by photoinduced electron transfer. We believe that this mechanism 10 also contributes to the observed effect in our case. Rhodaminemodified liposomes V-Rho exhibit a similar behavior during functionalization with the aptamer **Apt-Thr** (Fig. 2B).



15 Fig. 2 Fluorescence response monitoring the reaction progress of pyrene-(A) and rhodamine-modified (B) vesicles during surface functionalization with aptamer Apt-Thr.

The thrombin-binding aptamer folds into a highly compact Gquadruplex structure upon binding to thrombin. 18 This chair-like 20 conformation is stabilized by a metal cation, particularly potassium cations. Analogously to the sensing mechanism for the membrane functionalization process, we believe that binding of the target to the surface-bound oligonucleotide, accompanied by its conformational change, has a drastic impact on the membrane 25 structure which triggers then the dynamic emission response of the fluorophor molecules in close proximity (Fig. 3, bottom). At last, rather the presence of the aptamer bound analyte at the membrane-water interface causes the change of the emission signal than a defined conformation of the immobilized 30 oligonucleotide molecules.




Fig. 3 Principle of vesicle surface functionalization (I.) and fluorescence response after analyte binding (II.).

High concentrations of potassium cations are known to induce 35 the transition of the oligonucleotide from random coil to the folded form with a reported binding constant of $\lg K \approx 2.1^{.19}$ Therefore, we first tested the suitability of the thrombin aptamerfunctionalized vesicles V-Pyr-Thr for K⁺ detection. The fluorescence titration revealed an excimer signal change of 40 almost 60 % (Fig. 4, red squares). Due to the relatively low binding affinity of the aptamer to potassium cations, KCl concentrations of up to 140 mm were required to obtain a reasonable binding isotherm. In the same concentration range, non-functionalized vesicles showed a smaller non-specific 45 response with linear relationship (Fig. 4, blue dots). We explain this effect by the huge change of ionic strength in the aqueous buffer. It is well documented that high salt concentrations have a strong influence on the membrane's physical properties such as surface polarity and hydration and thus can alter the emission of 50 membrane-embedded fluorophors. 20 Non-linear curve fitting of the obtained binding isotherm resulted in a logarithmic binding constant of 1.3 (Table 1). The lower value of the potassium cation binding constant compared to the reported one can be attributed to the interference of specific binding and non-specific salt 55 effects.




Fig. 4 Fluorescence emission titrations of KCl versus aptamerfunctionalized (red squares) and non-functionalized (blue dots) vesicles

Next, we investigated the detection of thrombin. Pyrenefunctionalized vesicles V-Pyr-Thr showed a relatively small

fluorescence increase upon addition of thrombin (≈ 20 % at a 2.0 µM concentration of thrombin, see Fig. S4, ESI†), whereas rhodamine-modified liposomes V-Rho-Thr exhibited a considerable luminescence response (Fig. 5, red squares). As 5 expected, non-functionalized vesicles V-Rho did not significantly alter their emission properties upon the addition of thrombin (Fig. 5, blue dots).

Table 1 summarizes the apparent binding constants of the vesicular aptasensors to different analytes. All calculations were 10 performed by non-linear curve fitting using the equation for 1:1 binding.²¹ The affinity constants of **V-Rho-Thr** to both bovine and human thrombin are in good agreement with reported literature values. The limit of detection was determined to be 25 nM (on basis of a signal to noise ratio of 3), while the sensors 15 span a dynamic range of up to 2 µM concentration of thrombin (see ESI† for fluorescence titration data). The analytic results are reproducible as shown by repeated measurements. Since thrombin is a crucial enzyme in the blood coagulation cascade, its concentration in circulating blood is low and often undetectable. 20 During clotting initiation, it becomes significantly elevated at a low nanomolar concentration level and peaks at concentrations in the low micromolar region during the propagation phase. 11b,22 Thus, the dynamic recognition range of our sensor matches the physiological concentration range.


25 Table 1 Binding affinities of fluorescent, aptamer-functionalized vesicles to different analytes.

Vesicles	Analyte	lg K	Reference
V-Pyr-Thr	KCl	1.3 (± 8 %)	2.119
V-Pyr-Thr	Bovine thrombin	$6.3 (\pm 3 \%)$	7.1^{23}
V-Rho-Thr	Bovine thrombin	$7.2 (\pm 8 \%)$	7.1^{23}
V-Rho-Thr	Human thrombin	7.7 ^a	$6.7 - 7.6^{24}$

^a Single experiment.

To demonstrate that the interaction of thrombin with the covalently membrane-attached aptamer is responsible for the 30 change of the rhodamine emission intensity, we prepared fluorescent vesicles lacking the amphiphilic binding anchor Mal-C₁₆ and incubated them with the thiolated thrombin-binding aptamer Apt-Thr leading to no significant response upon analyte addition (Fig. 5, orange diamonds).

We tested the specificity of our aptasensors by immobilizing a non-binding control oligonucleotide Apt-Ctrl (Fig. 1) with comparable length on the vesicle surface. The slight response after thrombin addition can be attributed to non-specific binding (Fig. 5, magenta triangles). Although thrombin has a small 40 negative overall charge in the used buffer system (pH: 7.4, isoelectric point of thrombin: 7.05²⁵), its numerous negatively and positively charged residues are not distributed uniformly over the whole molecule.²⁶ They are clustered in a sandwich-like structure of two highly positive poles, whose electrostatic fields spread far 45 into the extramolecular space. We suppose that the electropositive exosites are capable to displace counter ions from the negatively charged oligonucleotide backbone by electrostatic attractions.27



50 Fig. 5 Binding isotherm of thrombin to aptamer-functionalized vesicles V-Rho-Thr (red squares) and fluorescence response of control experiments: non-functionalized vesicles V-Rho (blue dots), malemidelacking V-Rho vesicles incubated with Apt-Thr (orange diamonds) and vesicles V-Rho-Ctrl functionalized with non-binding oligonucleotide 55 Apt-Ctrl (magenta triangles).

Emission titration using the protein elastase, which is besides thrombin a member of the serine proteases enzyme family, versus V-Rho-Thr indicates the aptamer-protein non-specific interaction. The fluorescence intensity decreases during titration (Fig. 60 6), which is in contrast to the emission intensity increase observed with thrombin. Elastase has an isoelectric point of 9.5.28 and it is assumed that its high positive net charge possibly results in strong non-specific oligonucleotide-protein interactions,²⁷ leading to different fluorescence emission intensity changes.




Fig. 6 Fluorescence emission titration of control protein elastase to V-Rho-Thr.

In conclusion, we have developed a novel type of aptasensor based on self-assembled, luminescent vesicle membranes, which 70 are functionalized with the thrombin-binding aptamer by nucleophilic thiol-maleimide addition. The high-affinity binding of the aptamer to thrombin is indicated by a dynamic response of the co-embedded fluorescent reporter molecules in the physiological concentration range. The obtained binding 75 constants are in good agreement with the reported values. Despite of the observed non-specific interactions in case of mismatched aptamer-protein combinations, the specific binding to thrombin can be clearly distinguished in the fluorescence output. As the reporter group is neither covalently attached nor coordinated to 80 the aptamer, an inexpensive and easy variation of aptamerfluorophor combinations is possible. The signaling mechanism is independent of specific conformational changes of the receptor molecules, which may allow a broader application compared to previously reported label-free aptasensors. An extension to other 85 aptamers, enzymes or proteins and the immobilization of the vesicular aptasensors on solid supports can be readily envisaged.

A.M. thanks the Bayerische Eliteförderung for a research

scholarship.

Notes and references

Institut für Organische Chemie, Universität Regensburg, 93040 Regensburg, Germany. Fax: +49 943 1717; Tel: +49 943 4575; E-mail: 5 burkhard.koenig@ur.de

- † Electronic Supplementary Information (ESI) available: Experimental details, synthesis of amphiphiles, preparation of functionalized vesicles, emission spectra and binding curves. See DOI: 10.1039/b000000x/
- (a) C. Tuerk and L. Gold, Science, 1990, 249, 505-510; (b) A. D. Ellington and J. W. Szostack, Nature, 1990, 346, 818-822.
- (a) M. McKeague and M. C. DeRosa, J. Nucleic Acids, 2012, 2012, 748913; (b) A. D. Keefe, S. Pai and A. Ellington, Nat. Rev. Drug Discov., 2010, 9, 537-550; (c) E. Torres-Chavolla and E. C. Alocilja, Biosens. Bioelectron., 2009, 24, 3175-3182.
- (a) B. Strehlitz, C. Reinemann, S. Linkorn and R. Stoltenburg, Bioanal. Rev., 2012, 4, 1-30; (b) S. Amaya-González, N. de-los-Santos-Álvarez, A. J. Miranda-Ordieres and M. J. Lobo-Castañón, Sensors, 2013, 13, 16292-16311.
- For reviews, see: (a) J. Liu, Z. Cao and Y. Lu, Chem. Rev., 2009, 109, 1948–1998; (b) B. Deng, Y. Lin, C. Wang, F. Li, Z. Wang, H. Zhang, X.-F. Li and X. C. Le, Anal. Chim. Acta, 2014, 837, 1-15.
- (a) R. E. Wang, Y. Zhang, J. Cai, W. Cai and T. Gao, Curr. Med. Chem., 2011, 18, 4175-4184; (b) I. Willner and M. Zayats, Angew. Chem. Int. Ed., 2007, 46, 6408-6418; Angew. Chem., 2007, 119, 6528-6538.
- (a) B. Li, H. Wei and S. Dong, Chem. Commun., 2007, 73-75; (b) Z. Xu, K. Morita, Y. Sato, Q. Dai, S. Nishizawa and N. Teramae, Chem. Commun., 2009, 6445-6447; (c) M. Zayats, Y. Huang, R. Gill, C.-an Ma and I. Willner, J. Am. Chem. Soc., 2006, 128, 13666-13667.
- (a) L. Hu, Z. Bian, H. Li, S. Han, Y. Yuan, L. Gao and G. Xu, Anal. Chem., 2009, 81, 9807-9811; (b) Z. Zhu, C. Yang, X. Zhou and J. Qin, Chem. Commun., 2011, 47, 3192-3194.
- (a) B. Gruber, S. Stadlbauer, A. Späth, S. Weiss, M. Kalinina and B. König, Angew. Chem. Int. Ed., 2010, 49, 7125-7128; Angew. Chem., 2010, **122**, 7280–7284; (b) S. Banerjee, M. Bhuyan and B. König, Chem. Commun., 2013, 49, 5681-5683.
- For examples on how the modular construction principle can be utilized for the preparation of highly selective chemosensors, see: (a) B. Gruber, S. Balk, S. Stadlbauer and B. König, Angew. Chem. Int. Ed., 2012, 51, 10060-10063; Angew. Chem., 2012, 124, 10207-10210; (b) S. Banerjee and B. König, J. Am. Chem. Soc., 2013, 135, 2967-2970.
- 10 L. C. Bock, L. C. Griffin, J. A. Latham, E. H. Vermaas and J. J. Toole, Nature, 1992, 355, 564-566.
- 11 (a) K. A. Edwards and A. J. Baeumner, Anal. Bioanal. Chem., 2006, 386, 1613-1623; (b) K. A. Edwards, Y. Wang and A. J. Baeumner, Anal. Bioanal. Chem., 2010, 398, 2645-2654.
- 12 (a) P. A. Beales and T. K. Vanderlick, J. Phys. Chem. A, 2007, 111, 12372-12380; (b) Y.-H. M. Chan, B. van Lengerich and S. G. Boxer, Proc. Natl. Acad. Sci. U. S. A., 2009, 106, 979–984.
- 13 (a) Z. Cao, R. Tong, A. Mishra. W. Xu, G. C. L. Wong, J. Cheng and Y. Lu, Angew. Chem. Int. Ed., 2009, 48, 6494-6498; Angew. Chem., 2009, 121, 6616-6620; (b) H. Kang. M. B. O'Donoghue, H. Liu and W. Tan, Chem. Commun., 2010, 46, 249-251.
- 14 (a) C. Wang and Z. Ma, Anal. Bioanal. Chem., 2005, 382, 1708-1710; (b) J. Lee, H.-J. Kim and J. Kim, J. Am. Chem. Soc., 2008, 130, 5010-5011; (c) J. Lee, H. Jun and J. Kim, Adv. Mater, 2009, 21, 3674-3677; (d) Y. K. Jung, T. W. Kim, H. G. Park and H. T. Soh, Adv. Funct. Mater., 2010, 20, 3092-3097; (e) W. Wu, J. Zhang, M. Zheng, Y. Zhong, J. Yang, Y. Zhao, W. Wu, W. Ye, J. Wen, Q. Wang and J. Lu, PLOS ONE, 2012, 7, e48999.
- 15 P. M. Keller, S. Person and W. Snipes, J. Cell. Sci., 1977, 28, 167-
- 16 M. Manoharan, K. L. Tivel, M. Zhao, K. Nafisi and T. L. Netzel, J. Phys. Chem., 1995, 99, 17461–17472.
- S. A. E. Marras, F. R. Kramer and S. Tyagi, Nucleic Acids Res., 2002, 30, e122.

- 18 (a) R. F. Macaya, P. Schultze, F. W. Smith, J. A. Roe and J. Feigon, Proc. Natl. Acad. Sci. U. S. A., 1993, 90, 3745-3749; (b) K. Y. Wang, S. McCurdy, R. G. Shea, S. Swaminathan and P. H. Bolton, Biochemistry, 1993, 32, 1899-1904; (c) K. Padmanabhan, K. P. Padmanabhan, J. D. Ferrara, J. E. Sadler and A. Tulinsky, J. Biol. Chem., 1993, 268, 17651-17654.
- 19 S. Nagatoishi, T. Nojima, B. Juskowiak and S. Takenaka, Angew. Chem. Int. Ed., 2005, 44, 5067-5070; Angew. Chem., 2005, 117, 5195-5198.
- (a) G. Cevc, Biochim. Biophys. Acta, Rev. Biomembr., 1990, 1031, 311-382; (b) Y. Zhang and P. S. Cremer, Curr. Opin. Chem. Bio., 2006, 10, 658-663.
- 21 A. E. Hargrove, Z. Zhong, J. L. Sessler and E. V. Anslyn, New J. Chem., 2010, 34, 348-354.
- M. Lee and D. R. Walt, Anal. Biochem., 2000, 282, 142–146.
- 23 M. C. R. Buff, F. Schäfer, B. Wulffen, J. Müller, B. Pötzsch, A. Heckel and G. Mayer, Nucleic Acids Res., 2010, 38, 2111-2118.
- 24 P. Baaske. C. J. Wienken, P. Reineck, S. Duhr and D. Braun, Angew. Chem. Int. Ed., 2010, 49, 2238-2241; Angew. Chem., 2010, 122, 2286-2290.
- 25 P. G. Righetti, G. Tudor and K. Ek, J. Chromatogr. A, 1981, 220, 115-194
- 26 W. Bode, D. Turk and A. Karshikov, Protein Sci., 1992, 1, 426-471.
- P. L. Privalov, A. I. Dragan and C. Crane-Robinson, Nucleic Acids Res., 2011, 39, 2483-2491.
- U. J. Lewis, D. E. Williams and N. G. Brink, J. Biol. Chem., 1956, **222**, 705–720.