## JAAS

Accepted Manuscript



This is an *Accepted Manuscript*, which has been through the Royal Society of Chemistry peer review process and has been accepted for publication.

Accepted Manuscripts are published online shortly after acceptance, before technical editing, formatting and proof reading. Using this free service, authors can make their results available to the community, in citable form, before we publish the edited article. We will replace this Accepted Manuscript with the edited and formatted Advance Article as soon as it is available.

You can find more information about *Accepted Manuscripts* in the **Information for Authors**.

Please note that technical editing may introduce minor changes to the text and/or graphics, which may alter content. The journal's standard <u>Terms & Conditions</u> and the <u>Ethical guidelines</u> still apply. In no event shall the Royal Society of Chemistry be held responsible for any errors or omissions in this *Accepted Manuscript* or any consequences arising from the use of any information it contains.



www.rsc.org/jaas

### Journal Name

Cite this: DOI: 10.1039/x0xx00000x

Received 10th January 2014,

Accepted ooth January 2014

DOI: 10.1039/x0xx00000x

#### ARTICLE

# LA-ICP-MS Pb-U dating of young zircons from the Kos-Nisyros volcanic centre, SE Aegean arc. M. Guillong,<sup>a</sup> A. von Quadt,<sup>a</sup> S. Sakata,<sup>b</sup> I. Peytcheva<sup>a</sup> and O. Bachmann<sup>a</sup> Zircon Pb-U dating has become a key technique for answering many important questions in geosciences. This paper describes a new LA-ICP-MS approach. We show, using previously dated samples of a large quaternary rhyolitic eruption in the Kos-Nisyros volcanic centre (the 161 ka Kos Plateau Tuff) that the precision of our LA-ICP-MS method is as good as via SHRIMP, while ID-TIMS measurements confirm the accuracy. Gradational age distribution over >140 ka of the Kos zircons and the near-absence of inherited cores indicate near-continuous crystallisation in a growing magma reservoir with little input from wall rocks. Previously undated silicic eruptions from Nisyros volcano (Lower Pumice, Nikia Flow, Upper Pumice), which are stratigraphically constrained to have happened after the Kos Plateau Tuff, are dated to be younger than respectively 124 ± 35 ka, 111 ± 42 ka and 70 ± 24 ka. Samples younger than 1 Ma were corrected for initial thorium disequilibrium using a new formula that also accounts for disequilibrium in <sup>230</sup>Th decay.

#### Introduction

www.rsc.ora/

Laser ablation inductively coupled plasma mass spectrometry (LA-ICP-MS) is a relatively cheap and fast method to obtain age information from zircons<sup>1</sup>. Recent improvements in instrumentation and methods <sup>2-5</sup>, standard reference materials<sup>6-9</sup>, and data processing<sup>10-12</sup> have made this method very popular<sup>13-16</sup>. Better understanding of elemental fractionation processes<sup>17-19</sup>, which previously were a limitation on accuracy, has produced improved results<sup>11</sup>. Alternative methods, like sensitive high resolution ion microprobe (SHRIMP) and isotope dilution thermal ionisation mass spectrometry (ID-TIMS), still benefit respectively, from lower sample consumption and better precision and accuracy, but are more expensive and slower than LA-ICP-MS, limiting the amount of samples that can be analysed for a given project.

Due to the complexity<sup>17, 18</sup> of LA-ICP-MS for U-Pb zircon dating, the method requires care and results must be validated. However, it can now be shown that it is possible to generate accurate results with a good precision over extended periods of time using matrix matched standard reference materials and secondary standards<sup>4, 6, 7, 9, 14, 19-21</sup>.

As dating techniques improve, young samples are of increasing interest; they allow (1) improved stratigraphic constraints on active volcanoes, (2) investigating magma chamber processes that occur over timescales less than a few 100 kyr, and (3) inter-comparison of different dating methods at high resolution. The analysis of young zircons in the range of 50 ky to 10 Ma is challenging due to low radiogenic Pb content, problems with common Pb contamination and the necessity to correct for initial Th disequilibrium<sup>22</sup>. However, new results presented in this paper on young zircons (in the range of 50 to 500 ka) from the Kos-Nisyros volcanic centre (Aegean arc) show that we can reproduce SHRIMP precision and accuracy, while analysing a

lot more zircons with relative ease. We also show that the LA-ICP-MS technique can provide a powerful means of finding the youngest zircon within a population to estimate the eruption age of silicic units that would not be easily dateable otherwise. We stress, however, that this youngest zircon within a population can still be much older than the true eruption age.

#### Samples

Five samples from the active Kos-Nisyros volcanic center (Eastern Aegean Arc) were chosen for this study (KPT04-24, KPT04-36, NS07, NS24 NISP2(1), Table 1). The two samples from the Kos Plateau Tuff ( $a > 60 \text{ km}^3$  rhyolite ignimbrite, erupted at 161  $\pm$  1 ka<sup>23</sup>, KPT04-24, KPT04-36), studied petrologically<sup>24</sup> and already dated by SHRIMP<sup>24</sup>, were used to compare the LA-ICP-MS results to those from the SHRIMP and ID-TIMS. The last three samples (NS07, NS24, NISP2(1),) are from (1) the Lower Pumice (NISP2(1)) (2) The Nikia Flow (NS24), and (3) the Upper Pumice (NS07), all erupted on Nisyros Island<sup>25</sup> after the Kos Plateau Tuff. However, at present, these rhyo-dacitic to rhyolitic units, all with volumes of a few km<sup>3</sup>, are poorly-dated<sup>26</sup> to undated. As the youngest of tens of Pb-U zircon analyses typically overlap with the Ar/Ar ages when both methods have been applied<sup>27, 28</sup>, we used the youngest zircons to give an estimate of the maximum eruption age.

#### Instrumentation

The LA-ICP-MS system used for this work was a Resonetics Resolution 155 laser ablation system coupled to a Thermo Element XR Sector-field ICP-MS. The LA system is equipped with some improvements compared to similar systems previously described<sup>3, 29-31</sup>. The design of the 2 volume constant geometry ablation cell has changed: the stage movement

ournal of Analytical Atomic Spectrometry Accepted Manusc

2

3

4

5

6

7

8

9

10

11

12

13

14

15

16

17

18

19

59 60

mechanism is now within the larger volume ablation cell making the ball joint and arm that moves through the ball joint in and out redundant. Cell volume and gas flow thus remain truly constant while moving the samples in the sample holder. The small volume ablation cell (funnel) is fixed on the large cell and the Argon make up gas is mixed to the Helium aerosol within the funnel, decreasing the effective ablation volume where the ablated material is dispersed to less than 1 cm<sup>3</sup>. This improves the wash out and decreases the transport time from ablation cell to ICP. The sample holder can take up to 15 one inch mounts. Due to the implementation of a z-stage, small changes in the height of a sample relative to the objective can be corrected. A back lash correction is implemented in the Geostar software that controls the stage movement pushing the precision and reproducibility of the stage in the range of 3 micrometres. The only drawback compared to the original  $HelEx^{3, 29-31}$  ablation cell is a longer purge process to remove the air after sample exchange (15 min.). Parameters of the system are summarized in Table 2 following the guidelines presented by the Earthtime initiative<sup>32</sup>. The laser-ablated aerosols are introduced into a Thermo element XR mass

spectrometer. Coupled to the LA system, the spectrometer is used in the fast scan mode with 5 pre-scans. A separate set of cones is used for zircon analysis to minimize cross contamination. A jet-sampling cone in combination with an H skimmer cone is used for best sensitivity with the guard electrode off. Neither N<sub>2</sub> nor H<sub>2</sub> addition<sup>33</sup> generally enhanced sensitivity with or without the use of the guard electrode with this ICP-MS. Instrument parameters are daily optimized after instrument warm up, and for quality control the sensitivity of selected elements as well as the signal stability and the gas blank intensities for most masses are recorded. The instrument is optimized ablating NIST 612 with a 43-micron crater at 10 Hz and 3.5 J/cm<sup>2</sup> energy density in scanning mode (3  $\mu$ m/s). Typical sensitivity for <sup>238</sup>U is 1.5M cps. Typical gas bank levels are 0-3 CPS for <sup>238</sup>U, 5-10 cps for <sup>206</sup>Pb and 2000-3000 cps for <sup>202</sup>Hg although all gases pass a mercury trap. ID-TIMS measurements of 7 zircons from Kos Plateau Tuff were performed on a Thermo Triton instrument. Details of the instrument settings and the method can be found elsewhere.<sup>34</sup>

Laboratory nameDept. of Earth Science, ETH ZurichSample typeparationZirconsMake, Model & typeConventional mineral separation, 1 inch resin mount, 1µm polish to finishMake, Model & typeLaser ablation systemMake, Model & typeLaser ablation cell & volumeLaser valength193 mmPulse width25 nsPulse width25 nsSpot size30 µmAblation rate5 112Sampling mode / patternSingle hole drilling, 5 cleaning pulsesCarrier gas and flow100% He, 0.7 / minAblation duration40 secondsWake, Model & typeThermo Element XR SF-ICP-MSSampling mode / patternIsolowCarrier gas and flow100% He, 0.7 / minAblation duration40 secondsWake, Model & typeThermo Element XR SF-ICP-MSSample introductionXhafiton acrosol only, squid aerosol homogenisation deviceRF power1500WMake, Model ac typeThermo Element XR gas amised to the carrier inside ablation cell funnel)Detection system202, 204, 206, 207, 206, 232, 235, 238 anuIntegration time per peak12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206Carlbration time per reading0.20 secondsData ProcessingGarl how acrossing patkage usedUncertainty level & propagationGarl aread approximation of all ratios nomalized to primary reference materialOran integration time per reading0.20 secondsData Processing patkage usedHeypho <sup>2410</sup> (0.976) <sup>14</sup> (0.9076) <sup>140</sup>	Laboratory & Sample Preparation	
Sample preparation   Conventional mineral separation, 1 inch resin mount, 1µm polish to finish     Laser avaletion system   Resonetics Resolution 155     Ablation cell & volume   Laurin Technics 155, constant geometry, aerosol dispersion volume < 1 cm <sup>3</sup> Pulse width   25 ns     Energy density / Fluence   ~ 2.0 Lm <sup>2</sup> Repetition rate   5 Hz     Sampling mode / pattern   Single hole drilling, 5 cleaning pulses     Carrier gas and flow   100% He, 0.7 Junin     Ablation ration   Single hole drilling, 5 cleaning pulses     Carrier gas and flow   100% He, 0.7 Junin     Ablation atration   40 seconds     ICP-MS Instrument   Thermo Element XR SF-ICP-MS     Make, Model & type   Thermo Element XR SF-ICP-MS     Sample introduction   Single detcot triple mode SEM, analog, Faraday     Makes measured   12 ms (masses 202, 204), 20, 207, 208, 232, 235, 238, 140 ms (masses 206     Total integration time per reading   0.202 seconds     Detection system   8 ns     Typical oxide rate (ThO/Th)   0.18%     Typical dubely charged rate (Ba <sup>11</sup> /Ba <sup>1</sup> )   3.5%     Data processing   10 second prior to each ablation spot     Calibration strategy <t< td=""><td></td><td></td></t<>		
Laser ablation system   Resonetics Resolution 155     Make, Model & type   Laurin Technics 155, constant geometry, aerosol dispersion volume < 1 cm <sup>3</sup> Laser wavelength   193 nm     Pulse width   25 ns     Energy density / Fluence   ~ 2.0 Lm <sup>2</sup> Repetition rate   5 Hz     Sampling mode / pattern   Single hole drilling, 5 cleaning pulses     Carrier gas and flow   100% He, 0.7 l/min     Ablation rate   ~ 75 nm pulse <sup>4</sup> Sampling mode / pattern   Single hole drilling, 5 cleaning pulses     Carrier gas and flow   100% He, 0.7 l/min     Ablation duration   40 seconds     ICP-MS Instrument   Thermo Element XR SF-ICP-MS     Make-, Model & type   Thermo Element XR SF-ICP-MS     Sample introduction   Ablation aerosol only, squid aerosol homogenisation device     RF power   ~ 0.95 l/min Ar (gas mixed to He carrier inside ablation cell funnel)     Detection system   Single detector triple mode SEM, analog, Faraday     Make-up gas flow   0.95 l/min Ar (gas mixed to He carrier inside ablation cell funnel)     Detection system   10 second prior to each ablation spot     Total integration time per peak   10 second yrior to each ablation spot		
Make, Model & type   Resonetics Resolution 155     Ablation cell & volume   Laurin Technics 155, constant geometry, aerosol dispersion volume < 1 cm <sup>3</sup> Laser wavelength   193 nm     Pulse width   25 ns     Energy density / Fluence   ~ 2.0 Lm <sup>2</sup> Repetition rate   5 Hz     Spot size   30 µm     Ablation rate   ~ 75 nm pulse <sup>-1</sup> Sampling mode / pattern   Single hole drilling, 5 cleaning pulses     Carrier gas and flow   100° He, 0.7 Umin     Ablation duration   40 seconds <b>ICP-MS Instrument</b> Make, Model & type     Make, Model & type   Thermo Element XR SF-ICP-MS     Sample introduction   Ablation aerosol only, squid aerosol homogenisation device     RF powr   1500W     Make-up gas flow   ~ 0.95 Umin Ar (gas mixed to He carrier inside ablation ell funnel)     Detection system   12002, 204, 206, 207, 208, 232, 235, 238 anu     Integration time per peak   12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206     Total integration time per reading   0 second prior to each ablation spot     Calibration strategy   0 second syr for quality control     Bota Processing   10 second prior to each		Conventional mineral separation, 1 inch resin mount, 1µm polish to finish
Ablation cell & volumeLaurin Technics 155, constant geometry, aerosol dispersion volume < 1 cm³Laser wavelength193 nmPulse width25 nsEnergy density / Fluence $\sim 2.0  {\rm Lmr}^3$ Repetition rate5 HzSpot size30 µmAblation rate $\sim 75  {\rm nm}  {\rm pulse}^{-1}$ Sampling mode / patternSingle hole drilling, 5 cleaning pulsesCarrier gas and flow100% He, 0.7 UminAblation duration40 seconds <b>ICP-MS Instrument</b> Thermo Element XR SF-ICP-MSMake, Model & typeThermo Element XR SF-ICP-MSSample introductionAblation aerosol only, squid aerosol homogenisation deviceRF power1500WAdsesse measured202, 204, 206, 207, 208, 232, 235, 238, muIntegration time per peak12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206Typical oxide rate (ThO/Th)0.18%Typical oxide rate (Ba <sup>TT</sup> /Ba <sup>T</sup> )3.5% <b>Data Processing</b> 10 second prior to each ablation spotCalibration strategyG-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationG-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality 0.0002 (Wid mean of TIMS analysis) <sup>35</sup> Data processing package usedlolite 2.5 with VizualAgeMass discriminationG-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlQuality control / ValidationPlesovice; Wid ave: <sup>200</sup> Ph/ <sup>201</sup> U age = 313, 3 + 1, 7 (2 E		
Laser wavelength193 nmPulse width25 nsEnergy density / Fluence~ 2.0 Lm²Repetition rate5 HzSpot size30 µmAblation rate~ 75 nm pulse¹Sampling mode / patternSingle hole drilling, 5 cleaning pulsesCarrier gas and flow100% He, 0.7 1µminAblation duration40 secondsICP-MS InstrumentThermo Element XR SF-ICP-MSMake, Model & typeThermo Element XR SF-ICP-MSSample introductionAblation aerosol only, squid aerosol homogenisation deviceRF power1500WMasses measured202, 204, 206, 207, 208, 232, 235, 238, anuIntegration time per peak12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206Dead time (ns)8 nsTypical doubley charged rate (Ba*/Ba*)3.5%Data processing package used10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 used as primary reference material, pleavice, 91500 & Temora used as secondary's for quality controlQuality control / ValidationPleavice: Wid ave: <sup>309</sup> Ph/ <sup>281</sup> U age = 38.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), 790 (Wid ave: <sup>309</sup> Ph/ <sup>281</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)The disquilibrium correction and error propagationPlesovice: Wid ave: <sup>309</sup> Ph/ <sup>281</sup> U age = 1064.8 ± 3.3 (2 SE		
Pulse width25 ns - 2.0 Jcm²Energy density / Fluence $\sim 2.0 Jcm²$ Repetition rate5 HzSpot size30 µmAblation rate $\sim 75$ nm pulse¹Sampling mode / patternSingle hole drilling, 5 cleaning pulsesCarrier gas and flow100% He, 0.7 VminAblation duration40 seconds <b>ICP-MS Instrument</b> Thermo Element XR SF-ICP-MSMake, Model & typeThermo Element XR SF-ICP-MSSample introductionAblation aerosol only, squid aerosol homogenisation deviceRF power $\sim 0.95 Vinin Ar (gas mixed to He carrier inside ablation cell funnel)Detection systemSingle detector triple mode SEM, analog, FaradayMakes, Model & typeSingle detector triple mode SEM, analog, FaradayMasses measured202, 204, 206, 207, 208, 232, 235, 238 nuIntegration time per peak12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206Total integration time per reading0.8%Dead time (ns)8 nsTypical doubley charged rate (Ba"/Ba')3.5%Data processingGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlQuality control / ValidationAgas are quoted at 2 SF absolute, propagation is by quadratic addition. Reproduc or freeme taetrial incertainty is pr$		, , , , , , , , , , , , , , , , , , , ,
Energy density / Fluence   ~ 2.0 Jcm <sup>2</sup> Repetition rate   5 Hz     Spot size   30 µm     Ablation rate   ~ 75 nm pulse <sup>-1</sup> Sampling mode / pattern   Single hole drilling, 5 cleaning pulses     Carrier gas and flow   100% He, 0.7 l/min     Ablation duration   40 seconds     ICP-MS Instrument   Make, Model & type     Make, Model & type   Thermo Element XR SF-ICP-MS     Sample introduction   Ablation aerosol only, squid aerosol homogenisation device     RF power   1500W     Make-up gas flow   ~ 0.95 l/min Ar (gas mixed to He carrier inside ablation cell funnel)     Detection system   Single detector triple mode SEM, analog, Faraday     Masses measured   102, 204, 206, 207, 208, 232, 235, 238, aru     Integration time per peak   12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206     Total integration time per reading   0.202 seconds     Dead time (ns)   8 ns     Typical oxide rate (ThO/Th)   3.5%     Data Processing   10     Gas blank   10 second prior to each ablation spot     Calibration strategy   GL-1 used as primary reference material, Plesovice, 91500 & Temora used as secoredary's for qua		
Repetition rate5 HzSpot size30 µmAblation rate $30 \mu m$ Sampling mode / patternSingle hole drilling, 5 cleaning pulsesCarrier gas and flow100% He, 0.7 l/minAblation duration40 seconds <b>ICP-MS Instrument</b> Thermo Element XR SF-ICP-MSMake, Model & typeThermo Element XR SF-ICP-MSSample introductionAblation aerosol only, squid aerosol homogenisation deviceRF power1500WMake-up gas flow $\sim 0.95$ l/min Ar (gas mixed to He carrier inside ablation cell funnel)Detection systemSingle detector triple mode SEM, analog, FaradayMasess measured202, 204, 206, 207, 208, 232, 235, 238 anuIntegration time per reading0.202 secondsDead time (ns)8 nsDead time (ns)8 nsTypical oxide rate (ThO/Th)0.18%Typical oxide rate (ThO/Th)0.18%Typical oxide rate (Ba**/Ba*)3.5%Data ProcessingIo second prior to each ablation spotCalibration strategyGI-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGI-1 <sup>208</sup> Pb/ <sup>238</sup> U 0.09761 $\pm$ 0.0002 (Wid mean of TIMS analysis) <sup>35</sup> Data processing package usedNo common-Pb correction appliedQuality control / ValidationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduc of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. <sup>209</sup> Pb/ <sup>238</sup> U age = 338, 3 $\pm$ 1.7 (2 SE, MSWD = 3.1, n = 16 in 3 sessio		
Spt size   30 µm     Ablation rate   ~75 nm pulse <sup>1</sup> Sampling mode / pattern   Single hole drilling, 5 cleaning pulses     Carrier gas and flow   100% He, 0.7 1/min     Ablation duration   40 seconds     ICP-MS Instrument   Thermo Element XR SF-ICP-MS     Sample introduction   Ablation aerosol only, squid aerosol homogenisation device     RF power   1500W     Make-up gas flow   ~0.95 1/min Ar (gas mixed to He carrier inside ablation cell funnel)     Detection system   Single detector triple mode SEM, analog, Faraday     Masses measured   202, 204, 206, 207, 208, 232, 235, 238, amu     Integration time per peak   20 masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206     Total integration time per reading   0.202 seconds     Dead time (ns)   8 ns     Typical oubely charged rate (Ba <sup>+*</sup> /Ba <sup>+</sup> )   3.5%     Data Processing   10 second prior to each ablation spot     Calibration strategy   GJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary''s for quality control     Reference Material information   GJ-1 used as correction or all ratios normalized to primary reference material     Common-Pb correction   Mass discorrection applied		
Åblation rate $\sim 75$ nm pulse <sup>-1</sup> Sampling mode / patternSingle hole drilling, 5 cleaning pulsesCarrier gas and flow100% He, 0, 7 l/minAblation duration40 secondsICP-MS InstrumentThermo Element XR SF-ICP-MSSample introductionAblation aerosol only, squid aerosol homogenisation deviceRF power1500WMake, Model & typeThermo Element XR SF-ICP-MSSample introductionAblation aerosol only, squid aerosol homogenisation deviceRF power1500WMake-up gas flow $\sim 0.95$ l/min Ar (gas mixed to He carrier inside ablation cell funnel)Detection systemSingle detector triple mode SEM, analog, FaradayMasses measured12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238 amuIntegration time per peak12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238, 340 ms (masses 206Total integration time per reading0.202 secondsBoad time (ns)8 nsTypical doubely charged rate (Ba <sup>+</sup> /Ba <sup>+</sup> )3.5%Data Processing10Gas blank10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 <sup>200</sup> Pp/ <sup>238</sup> U que = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 secondary)Mass discriminationMass discorrection appliedUncertainty level & propagationAges are quoted at 2 SE abolute, propagation is by quadratic addition. Reproduc of reference material uncertainty is propagated. <sup>10</sup> Quality control / Validation <t< td=""><td></td><td></td></t<>		
Sampling mode / patternSingle hole drilling, 5 cleaning pulsesCarrier gas and flow100% He, 0.71 WiniAblation duration40 secondsICP-MS InstrumentThermo Element XR SF-ICP-MSSample introductionAblation aerosol only, squid aerosol homogenisation deviceRF power1500WMake-up gas flow~ 0.95 1/min Ar (gas mixed to He carrier inside ablation cell funnel)Detection systemSingle detector triple mode SEM, analog, FaradayMasses measured202, 204, 206, 207, 208, 232, 235, 238, 40 ms (masses 206Total integration time per peak12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206Dead time (ns)8 nsTypical oxide rate (ThO/Th)0.18%Typical oxide rate (ThO/Th)0.18%Typical oxide rate (Ba <sup>++</sup> /Ba <sup>+</sup> )3.5%Data ProcessingGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlData processing package usedIolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference materialUncertainty level & propagationPlesovice: W1 ave. <sup>200</sup> Ph <sup>238</sup> U age = 21.6 ± 2.0 (25E, MSWD = 3.1, n=16 in 3 sessions), 91500: Wt dave. <sup>200</sup> Ph <sup>238</sup> U age = 1064.8 ± 3.3 (25E, MSWD = 3.4, n 3 sessions), 91500: Wt dave. <sup>200</sup> Ph <sup>238</sup> U age = 1064.8 ± 3.3 (25E, MSWD = 3.4, n 3 sessions), 91500: Wt dave. <sup>200</sup> Ph <sup>238</sup> U age = 1064.8 ± 3.3 (25E, MSWD = 3.4, n 3 sessions), 91500: Wt dave. <sup>200</sup> Ph <sup>238</sup> U age of the samples were c		
Carrier gas and flow   100% He, 0.7 1/min     Ablation duration   40 seconds     ICP-MS Instrument   Make, Model & type     Sample introduction   Ablation acrosol only, squid acrosol homogenisation device     RF power   1500W     Makes-up gas flow   ~ 0.95 1/min Ar (gas mixed to He carrier inside ablation cell funnel)     Detection system   Single detector triple mode SEM, analog, Faraday     Masses measured   12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206     Total integration time per peak   12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206     Total integration time per reading   0.202 seconds     Bead time (ns)   8 ns     Typical doubely charged rate (Ba**/Ba*)   3.5%     Data Processing   Gas blank     Calibration strategy   GJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality control     Bata processing package used   Mass bias correction for all ratios normalized to primary reference material     Common-Pb correction   Mass bias correction for all ratios normalized to primary reference material     Quality control / Validation   Plesovice: Wit ave: <sup>300</sup> Pb/ <sup>38</sup> U age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 session), Temora: Wit ave: <sup>300</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0		
Ablation duration   40 seconds     ICP-MS Instrument   Thermo Element XR SF-ICP-MS     Sample introduction   Ablation aerosol only, squid aerosol homogenisation device     RF power   1500W     Make, Model & type   Single detector triple mode SEM, analog, Faraday     Detection system   Single detector triple mode SEM, analog, Faraday     Masses measured   202, 204, 206, 207, 208, 232, 235, 238 amu     Integration time per reading   0.202 seconds     Data time (ns)   8 ns     Typical oxide rate (ThO/Th)   0.18%     Typical doubely charged rate (Ba <sup>++</sup> /Ba <sup>+</sup> )   3.5%     Data Processing   Io second prior to each ablation spot     Calibration strategy   Gi-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality control     Calibration strategy   Gi-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality control     Quality control / Validation   Mass bias correction for all ratios normalized to primary reference material     Quality control / Validation   Plesovice: Wtd. are. <sup>206</sup> Pb/ <sup>238</sup> U age = 338.3 ± 1.7 (25E, MSWD = 3.1, n=16 in 3 session), 1500: Wtd are. <sup>306</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (25E, MSWD = 3.4, n 3 session), 1500: Wtd are. <sup>306</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (25E, MSWD = 3.4, n 3 session), 1500: Wtd are. <sup>306</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (25E, MSW		
ICP-MS Instrument     Make, Model & type   Thermo Element XR SF-ICP-MS     Sample introduction   Ablation aerosol only, squid aerosol homogenisation device     RF power   1500W     Make-up gas flow   ~ 0.95 Virmin Ar (gas mixed to He carrier inside ablation cell funnel)     Detection system   202, 204, 206, 207, 208, 232, 235, 238 mu     Integration time per peak   12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206     Total integration time per reading   0.202 seconds     Dead time (ns)   8 ms     Typical oxide rate (ThO/Th)   0.18%     Typical doubely charged rate (Ba <sup>++</sup> /Ba <sup>+</sup> )   3.5%     Data Processing   10 second prior to each ablation spot     Calibration strategy   GJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality control     Gamon-Pb correction   Mass bias correction for all ratios normalized to primary reference material     Uncertainty level & propagation   Ages are quoted at 2 SE absolute, propagatel. <sup>10</sup> Quality control / Validation   Plesovice: Wtd ave. <sup>306</sup> Pb/ <sup>28U</sup> age = 1064.8 ± 3.3 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. <sup>306</sup> Pb/ <sup>28U</sup> age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)     The isequilibrium correction and error propagation <sup>306</sup> Pb/ <sup>28U</sup> ages of the samples were corrected usi		,
Make, Model & typeThermo Element XR SF-ICP-MSSample introductionAblation aerosol only, squid aerosol homogenisation deviceRF power1500WMake-up gas flow~ 0.95 l/min Ar (gas mixed to He carrier inside ablation cell funnel)Detection systemSingle detector triple mode SEM, analog, FaradayMasses measured202, 204, 206, 207, 208, 232, 235, 238 amuIntegration time per peak12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206Total integration time per reading0.202 secondsDead time (ns)8 nsTypical oxide rate (ThO/Th)0.18%Typical doubely charged rate (Ba**/Ba*)3.5%Data Processing10 second prior to each ablation spotGas blank10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlGeneratinty level & propagationMass discriminationCommon-Pb correctionMass correction for all ratios normalized to primary reference materialUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduc of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 3.4, n 3 session), 91500: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age of the samples were corrected using equation (3). Errors from (20		40 seconds
Sample introductionAblation aerosol only, squid aerosol homogenisation deviceRF power1500WMake-up gas flow $0.95$ L/min Ar (gas mixed to He carrier inside ablation cell funnel)Detection systemSingle detector triple mode SEM, analog, FaradayMasses measured202, 204, 206, 207, 208, 232, 235, 238 amuIntegration time per peak12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206Dead time (ns)8 nsTypical oxide rate (ThO/Th)0.18%Typical oxide rate (ThO/Th)0.18%Typical doubely charged rate (Ba**/Ba*)3.5%Data Processing10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 $^{200}$ Pb/ $^{238}$ U 0.09761 ± 0.0002 (Wtd mean of TIMS analysis) <sup>35</sup> Data processing package usedIolite 2.5 with VizualAgeMass bias correction onNo common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduc of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. $^{206}$ Pb/ $^{238}$ U age = 1064.8 ± 3.3 (2 SE, MSWD = 3.4, n 3 session), 91500: Wtd ave. $^{206}$ Pb/ $^{238}$ U age of the samples were corrected using equation (3). Errors from (20		
RF power   1500W   - 0.95 L/min Ar (gas mixed to He carrier inside ablation cell funnel)     Make-up gas flow   Single detector triple mode SEM, analog, Faraday     Masses measured   202, 204, 206, 207, 208, 232, 235, 238 anu     Integration time per peak   12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206     Total integration time per reading   8 ns     Dead time (ns)   0.18%     Typical doubely charged rate (Ba <sup>++</sup> /Ba <sup>+</sup> )   0.18%     Data Processing   0 second prior to each ablation spot     Galibration strategy   GJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality control     Reference Material information   GJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality control     Quality control / Validation   Mass bias correction for all ratios normalized to primary reference material normalized to reference material normalized to primary is propagated. <sup>10</sup> Quality control / Validation   Plesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 session), 91500: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)     Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		
Make-up gas flow~ 0.95 l/min Ar (gas mixed to He carrier inside ablation cell funnel)Detection systemSingle detector triple mode SEM, analog, FaradayMasses measured202, 204, 206, 207, 208, 232, 235, 238 amuIntegration time per peak12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206Dead time (ns)8 nsTypical doubely charged rate (Ba**/Ba*)0.18%Data Processing10 second prior to each ablation spotGas blank10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 <sup>206</sup> Pb/ <sup>238</sup> U 0.09761 ± 0.0002 (Wtd mean of TIMS analysis) <sup>35</sup> Data processing package usedIol secore to for all ratios normalized to primary reference materialMass discriminationMass bias correction for all ratios normalized to primary reference materialUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduc of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 318. 3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), 9.1500: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		
Detection systemSingle detector triple mode SEM, analog, FaradayMasses measured202, 204, 206, 207, 208, 232, 235, 238 anuIntegration time per peak202, 204, 206, 207, 208, 232, 235, 238 anuTotal integration time per reading202 secondsDead time (ns)8 nsTypical oxide rate (ThO/Th)0.18%Typical doubely charged rate (Ba <sup>++</sup> /Ba <sup>+</sup> )3.5%Data Processing6J-1 used as primary reference material, Plesovice, 91500 & Temora used asGas blank10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used assecondary's for quality controlGJ-2 secondsMass discriminationGJ-1 <sup>206</sup> Pb/ <sup>238</sup> U 0.09761 ± 0.0002 (Wtd mean of TIMS analysis) <sup>35</sup> Iolite 2.5 with VizualAgeMass bias correction for all ratios normalized to primary reference material No common-Pb correctionQuality control / ValidationPlesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 338, 3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 3 sessions), 91500: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		
Masses measured202, 204, 206, 207, 208, 232, 235, 238 amuIntegration time per peak202, 204, 206, 207, 208, 232, 235, 238, 40 ms (masses 206Total integration time per reading202, 204, 206, 207, 208, 232, 235, 238, 40 ms (masses 206Dead time (ns)8 msTypical oxide rate (ThO/Th)0.18%Typical doubely charged rate (Ba <sup>++</sup> /Ba <sup>+</sup> )3.5%Data Processing3.5%Gas blank10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used asReference Material informationGJ-1 used as primary reference material, Plesovice, 91500 & Temora used asData processing package usedIolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference materialUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduc of reference material uncertainty is propagated.10Quality control / ValidationPlesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 3.4, n 3 sessions), 91500: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation2 <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		
Integration time per peak12 ms (masses 202, 204), 20 ms (masses 208, 232, 235, 238), 40 ms (masses 206Total integration time per reading0.202 secondsDead time (ns)8 nsTypical oxide rate (ThO/Th)0.18%Typical doubely charged rate (Ba <sup>++</sup> /Ba <sup>+</sup> )3.5%Data Processing0.10 second prior to each ablation spotGas blank10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 <sup>206</sup> Pb/ <sup>238</sup> U 0.09761 ± 0.0002 (Wtd mean of TIMS analysis) <sup>35</sup> Data processing package usedIolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference materialUncertainty level & propagationNo common-Pb correction appliedQuality control / ValidationPlesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		
Total integration time per reading $0.202$ secondsDead time (ns)8 nsTypical oxide rate (ThO/Th) $0.18\%$ Typical doubely charged rate (Ba <sup>++</sup> /Ba <sup>+</sup> ) $3.5\%$ Data Processing $3.5\%$ Gas blank10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 $^{206}Pb/^{238}U$ 0.09761 ± 0.0002 (Wtd mean of TIMS analysis) <sup>35</sup> Data processing package usedIolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference materialUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduct of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. $^{206}Pb/^{238}U$ age = $338.3 \pm 1.7$ (2 SE, MSWD = $3.1$ , n=16 in 3 sessions), 1500: Wtd ave. $^{206}Pb/^{238}U$ age = $1064.8 \pm 3.3$ (2 SE, MSWD = $3.4$ , n 3 sessions), 2106Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		
Dead time (ns)8 nsTypical oxide rate (ThO/Th)0.18%Typical doubely charged rate (Ba <sup>++</sup> /Ba <sup>+</sup> )3.5%Data Processing0Gas blank10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 <sup>206</sup> Pb/ <sup>238</sup> U 0.09761 ± 0.0002 (Wtd mean of TIMS analysis) <sup>35</sup> lolite 2.5 with VizualAgeMass discriminationMass discriminationCommon-Pb correctionNo common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduc of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), 91500: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20	Integration time per peak	
Typical oxide rate (ThO/Th)0.18%Typical doubely charged rate (Ba <sup>++</sup> /Ba <sup>+</sup> )3.5%Data Processing Gas blank10 second prior to each ablation spotCalibration strategy10 second prior to each ablation spotReference Material informationGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlData processing package usedIolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference materialCommon-Pb correctionNo common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduct of reference material uncertainty is propagated.10Quality control / ValidationPlesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		
Typical doubely charged rate (Ba <sup>++</sup> /Ba <sup>+</sup> )3.5%Data Processing Gas blank10 second prior to each ablation spotCalibration strategy10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material information Data processing package used10 second prior to each ablation spotMass discrimination Common-Pb correctionGJ-1 <sup>206</sup> Pb/ <sup>238</sup> U 0.09761 ± 0.0002 (Wtd mean of TIMS analysis) <sup>35</sup> Uncertainty level & propagationMass bias correction for all ratios normalized to primary reference material No common-Pb correction appliedQuality control / ValidationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduct of reference material uncertainty is propagated. <sup>10</sup> Plesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), 91500: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		
Data Processing Gas blank10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 $^{206}Pb/^{238}U 0.09761 \pm 0.0002$ (Wtd mean of TIMS analysis)^{35} Iolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference material No common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduct of reference material uncertainty is propagated.10Quality control / ValidationPlesovice: Wtd. ave. $^{206}Pb/^{238}U$ age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. $^{206}Pb/^{238}U$ age = 1064.8 ± 3.3 (2 SE, MSWD = 3.4, n 3 sessions), 91500: Wtd ave. $^{206}Pb/^{238}U$ age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation $^{206}Pb/^{238}U$ ages of the samples were corrected using equation (3). Errors from (20		
Gas blank10 second prior to each ablation spotCalibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 <sup>206</sup> Pb/ <sup>238</sup> U 0.09761 ± 0.0002 (Wtd mean of TIMS analysis) <sup>35</sup> Iolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference material No common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduce of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		3.5%
Calibration strategyGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlReference Material informationGJ-1 used as primary reference material, Plesovice, 91500 & Temora used as secondary's for quality controlData processing package usedGJ-1 <sup>206</sup> Pb/ <sup>238</sup> U 0.09761 ± 0.0002 (Wtd mean of TIMS analysis) <sup>35</sup> lolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference material No common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduce of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. <sup>206</sup> Pb/ <sup>238</sup> U age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20	9	
Secondary's for quality controlReference Material informationGJ-1Data processing package usedIolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference materialCommon-Pb correctionNo common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduce of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. $^{206}Pb/^{238}U$ age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. $^{206}Pb/^{238}U$ age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation $^{206}Pb/^{238}U$ ages of the samples were corrected using equation (3). Errors from (20		
Reference Material information $GJ-1^{206}Pb/^{238}U \ 0.09761 \pm 0.0002 \ (Wtd mean of TIMS analysis)^{35}$ Data processing package usedIolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference materialCommon-Pb correctionNo common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduce of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. $^{206}Pb/^{238}U$ age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. $^{206}Pb/^{238}U$ age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation $^{206}Pb/^{238}U$ ages of the samples were corrected using equation (3). Errors from (20	Canoradon strategy	
Data processing package usedIolite 2.5 with VizualAgeMass discriminationMass bias correction for all ratios normalized to primary reference materialCommon-Pb correctionNo common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. ReproduceQuality control / ValidationPlesovice: Wtd. ave. $^{206}Pb/^{238}U$ age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. $^{206}Pb/^{238}U$ age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation $^{206}Pb/^{238}U$ ages of the samples were corrected using equation (3). Errors from (20	Pafaranaa Matarial information	Secondary S 10F quality control $CI = 1^{206} Db^{238} U = 0.00761 \pm 0.0002$ (Wtd mean of TIMS analysis) <sup>35</sup>
Mass discrimination Common-Pb correctionMass bias correction for all ratios normalized to primary reference material No common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduce of reference material uncertainty is propagated. <sup>10</sup> Quality control / ValidationPlesovice: Wtd. ave. $^{206}Pb/^{238}U$ age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 		
Common-Pb correctionNo common-Pb correction appliedUncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduce of reference material uncertainty is propagated.10Quality control / ValidationPlesovice: Wtd. ave. $^{206}Pb/^{238}U$ age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. $^{206}Pb/^{238}U$ age = 421.6 ± 2.0 (2 SE, MSWD = 3.4, n 3 sessions), 91500: Wtd ave. $^{206}Pb/^{238}U$ age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation $^{206}Pb/^{238}U$ ages of the samples were corrected using equation (3). Errors from (20		
Uncertainty level & propagationAges are quoted at 2 SE absolute, propagation is by quadratic addition. Reproduct of reference material uncertainty is propagated.10Quality control / ValidationPlesovice: Wtd. ave. $^{206}Pb/^{238}U$ age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. $^{206}Pb/^{238}U$ age = 421.6 ± 2.0 (2 SE, MSWD = 3.4, n 3 sessions), 91500: Wtd ave. $^{206}Pb/^{238}U$ age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation $^{206}Pb/^{238}U$ age of the samples were corrected using equation (3). Errors from (20)		
Quality control / Validationof reference material uncertainty is propagated.Quality control / ValidationPlesovice: Wtd. ave. $^{206}Pb/^{238}U$ age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. $^{206}Pb/^{238}U$ age = 421.6 ± 2.0 (2 SE, MSWD = 3.4, n 3 sessions), 91500: Wtd ave. $^{206}Pb/^{238}U$ age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation $^{206}Pb/^{238}U$ ages of the samples were corrected using equation (3). Errors from (20		
Quality control / ValidationPlesovice: Wtd. ave. ${}^{206}Pb/{}^{238}U$ age = 338.3 ± 1.7 (2 SE, MSWD = 3.1, n=16 in 3 sessions), Temora: Wtd ave. ${}^{206}Pb/{}^{238}U$ age = 421.6 ± 2.0 (2 SE, MSWD = 3.4, n 3 sessions), 91500: Wtd ave. ${}^{206}Pb/{}^{238}U$ age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session)Th disequilibrium correction and error propagation ${}^{206}Pb/{}^{238}U$ age s of the samples were corrected using equation (3). Errors from (20	Oncertainty level & propagation	Ages are quoted at 2 SE absolute, propagation is by quadratic addition. Reproductor
sessions), Temora: Wtd ave. ${}^{206}Pb/{}^{238}U$ age = 421.6 ± 2.0 (2 SE, MSWD = 3.4, n 3 sessions), 91500: Wtd ave. ${}^{206}Pb/{}^{238}U$ age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session) ${}^{206}Pb/{}^{238}U$ ages of the samples were corrected using equation (3). Errors from (20	Quality control / Validation	Discrete indicate indicating is propagated. Discrete with an $2^{206}$ Db/ $^{238}$ Lago = 229.2 ± 1.7 (2 SE MSWD = 2.1 m=16 in 2
Th disequilibrium correction and error propagation $3 \text{ sessions}$ , 91500: Wtd ave. ${}^{206}\text{Pb}/{}^{238}\text{U}$ age = 1064.8 ± 3.3 (2 SE, MSWD = 0.44 in 1 session) ${}^{206}\text{Pb}/{}^{238}\text{U}$ ages of the samples were corrected using equation (3). Errors from (20	Quality control / vandation	resolute. with ave. $r_{0} = 0.025 = 336.5 \pm 1.7$ (2.5E, MSWD = 3.1, II=10 III 5 sessions). Temora: Wtd ave. ${}^{206}Pb/{}^{238}U$ are = 421.6 + 2.0 (2.5E, MSWD = 3.4, n=19)
Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		Sessions), remota. Wild ave. $10^{-0}$ of age = 421.0 ± 2.0 (2.5E, MSWD = 0.4, n=10) 3 sessions), 01500; Wild ave. $2^{206}$ Db/ $2^{238}$ L age = 1064.8 ± 3.3 (2.5E, MSWD = 0.44, n=10)
Th disequilibrium correction and error propagation <sup>206</sup> Pb/ <sup>238</sup> U ages of the samples were corrected using equation (3). Errors from (20		
$10^{-10} \text{ are sequence for a condition and on or propagation}$	Th disequilibrium correction and error propagation	
	in disequinorium concention and error propagation	$107 = 0$ ages of the samples were confected using equation (5). Efforts from (200/2) measured from $\lambda = (1.55125 \times 10^{-10} \text{ 0} \cdot 1106 \text{ 28})^{36}$ and $\lambda = (0.17055 \times 10^{-06} \text{ 0} \cdot 1506 \text{ 2})^{36}$
are propagated.		

16

17

18

19

20

21

22

23

24

25

26

27

28

29

30

31

32

33

34

35

36

37

38

39

40

41

42

43

44

45

46

47

48

49

50

51

52

53

54

55

56

57

58 59 60 Journal Name

Sample	Coordinates	Th/U magma $\pm 2$ SE (30%) <sup>1</sup>	Source <sup>2</sup>	Th/U zircons $\pm 2SE$	$f(Th/U) \pm 28$
NS07	N 36°36.728 E 27°09.407	4.0 ± 0.1 (1.2)	WR analysis	$0.95 \pm 0.11$	$0.236 \pm 0.02$
NS24	N 36°34.630 E 27°10.560	3.85 ± 0.43 (1.15)	WR analysis (= NISP2)	$0.89\pm0.16$	$0.231 \pm 0.04$
NISP2(1)	N 36°36.986 E 27°09.814	$3.85 \pm 0.43$ (1.15)	WR analysis	$0.95 \pm 0.22$	$0.246 \pm 0.05$
KPT04-36	N 36°46.097 E 27°06.990	3.3 ± 0.3 (1)	WR analysis	$0.83\pm0.18$	$0.252 \pm 0.03$
KPT04-24	N 36°57.791 E 26°57.294	3.3 ± 0.3 (1)	WR analysis	$0.59\pm0.05$	$0.181 \pm 0.01$

#### Methods

All samples were mounted in 1 inch epoxy mounts, polished and prior to analysis, cathode luminescence (CL) images were acquired to identify possible inherited cores. Laser analysis spots on the zircon grains were programmed using the Geostar software with a precision of 3 microns in the x and y direction. Each spot consists of: 1. Movement of the stage to the location. 2. Surface cleaning with 5 laser pulses 3. Wash out of the aerosol from cleaning during 17 seconds. 4. Triggering the ICP-MS to start the next measurement 5. Acquire gas blank signal for 10 seconds. 6. Ablating the sample for 40 seconds (30 micron crater, 5Hz repetition rate, 2 J/cm<sup>2</sup> energy density). The primary standard reference sample GJ-1 is measured 3 times at the beginning and end of each sequence and also 3 times after 15-20 samples. Total numbers of primary SRM were >15. Between the primary SRM measurements at least 2 secondary SRM (Plesovice, Temora, 91500) were included for quality control. Raw data was imported into Iolite<sup>10, 11</sup> and with the use of the VizualAge<sup>12</sup> data reduction scheme reduced to obtain ages and ratios corrected for instrumental drift and down hole fractionation. Downhole fractionation was found to be very similar between SRM, secondary SRM and the samples.

Zircon ages can be corrected for common Pb contamination using different methods. The most commonly used approach is via measuring <sup>204</sup>Pb and using common Pb isotope ratios to subtract the common Pb component from the radiogenic Pb isotopes. It is also widely used to calculate the intercept of the measured ratios with the concordia curve assuming a common Pb composition<sup>38</sup> or applying the routine proposed by Andersen<sup>39</sup>. These common Pb corrections were considered, but due to the very low 204 counts and the isobaric interference from <sup>204</sup>Hg, it was not possible to measure the <sup>204</sup>Pb counts with sufficient precision. The 207Pb signal was also very low and therefore not precise enough to reliably correct for common Pb. All age results presented in this work are therefore not common Pb corrected. With the help of VizualAge and a live Concordia diagram common Pb contaminated signals are easily recognised as a shift of the data point ellipse away from the Concordia towards common Pb composition as shown in figures ESI 1-7. Integration intervals are, where possible, set to exclude any common Pb. As few as 1 analysis (KPT samples) to up to 50% of analyses (NS07) with remaining common Pb were discarded based on live Concordia diagrams and the offset to the Concordia as is shown for all samples in the electronic supplementary information, Figures ESI 1-7. These Concordia plots show results prior to the Th disequilibrium correction, and are therefore expected to be discordant<sup>40</sup>. Although all masses

were measured, only the <sup>206</sup>Pb/<sup>238</sup>U age is considered precise and accurate enough. The other ratios (<sup>207</sup>Pb/<sup>235</sup>U, <sup>208</sup>Pb/<sup>232</sup>Th and <sup>207</sup>Pb/<sup>206</sup>Pb) could not be measured with sufficient precision and accuracy, partly due to the influence of common Pb on these ratios and the low abundance in such young zircons.

#### Thorium disequilibrium correction

The U/Pb age of zircons can be calculated using the following equation, originally suggested by Schärer (1984)<sup>2</sup>

$$\frac{\lambda_{238}}{2^{38}U} = \left(e^{\lambda_{238}t} - 1\right) + \frac{\lambda_{238}}{\lambda_{230}}\left(f_{Th/U} - 1\right)$$
(1)

As zircon grows from a melt, Th gets fractionated from U, imparting a disequilibrium in <sup>230</sup>Th (an intermediate product in the <sup>238</sup>U decay series) that has to be corrected to get an accurate age. The initial Th disequilibrium correction has been described in detail in several papers<sup>22, 40-43</sup>. This fractionation is described by  $f_{(Th/U)}$  (equation 2) that needs to be determined.

$$f_{Th/U} = \left(\frac{(Th/U)_{zircon}}{(Th/U)_{magma}}\right)$$
(2)

There are several approaches to quantify the fractionation factor f<sub>(Th/U)</sub>. With LA-ICP-MS and SHRIMP, the Th/U in the mineral is directly measured. Th/U in the magma is typically considered as a constant and can be estimated by whole rock or volcanic glass analysis. However, the assumption of a homogeneous magma with constant Th/U is questionable. Instead of assuming a constant Th/U in the magma, it is possible to assume a constant fractionation and variations in Th/U in the magma. An investigation of these two approaches<sup>43</sup> found larger scatter in the data using the constant fractionation compared to the constant magma value. A constant magma value implies varying fractionation, while a constant fractionation would imply variations in the Th/U ratio within the magma. Reality might be in-between and the assumption of a constant Th/U in the magma was weighted more<sup>43</sup> and this approach was also used in this work. The uncertainty on Th/U in the magma was found to be 14% estimated from isotope dilution Th/U measurements<sup>43</sup>. Another approach to overcome the problem of unknown Th/U in the magma uses the difference between the measured  ${}^{208}\text{Pb}/{}^{232}\text{Th}$  age which is not affected by disequilibrium and the  ${}^{206}\text{Pb}/{}^{238}\text{U}$  age to get  $f_{(Th/U)}{}^{44}$ . Our data did not allow this approach as the  $^{208}$ Pb/ $^{232}$ Th age has too high uncertainty and variability possibly due to common Pb contamination.

2

3

4

5

6

7

8

9

10

11 12

13 14 15

16

17

18

19

20 21

22 23

24 25

26

27

28

29 30

31 32

33

34

35

36 37

42

43 44 45

46

47

48

49

50

51

52

53

54

55

56

57

58 59 60 Page 4 of 9

The growth of the  ${}^{206}$ Pb/ ${}^{238}$ U ratio can be calculated correctly with this equation (1) only after  ${}^{230}$ Th reaches radioactive

$$\frac{{}^{206}Pb}{{}^{238}U} = \left(e^{\lambda_{238}t} - 1\right) + \frac{\lambda_{238}}{\lambda_{230}}\left(f_{\frac{Th}{U}} - 1\right)\left(1 - e^{-\lambda_{230}t}\right)e^{\lambda_{238}t}$$

As the ages measured in this work are between 70 ka and 400 ka, equation 3 was used to get accurate results. A comparison of ages determined using both equations is discussed below. Th/U whole rock data available on the Kos Plateau Tuff is limited and the precision based on few samples and bulk chemistry analysis is likely overestimating the uncertainty on

the Th/U in the magma itself. Therefore we assume an uncertainty on the Th/U in the magma of 30% and propagated this uncertainty to the corrected age.

(3)

equilibrium. This equation does not accurately correct when

<sup>230</sup>Th is still in disequilibrium which is the case up to about 300

ka. In fact, the isotopic growth curve defined from equation (1)

does not pass the origin point when t = 0 and 206Pb = 0.

Therefore, for the chronology of the zircon crystals with ages

below 1 Ma, a more accurate correction must be applied as

shown in equation (3). The initial assumptions and the

derivation of equation (3) can be found in the electronic

supplementary information and in Sakata et al. 2013<sup>44</sup>




Figure 1 <sup>206</sup>Pb/<sup>238</sup>U age results for SHRIMP<sup>27</sup> (narrow error bars), ID-TIMS (medium error bars,) and LA-ICP-MS (wide error bars) for Kos Plateau Tuff samples including 2 standard errors. Only LA-ICP-MS and ID-TIMS results have the errors from the Th disequilibrium correction propagated. 2 old cores found with LA-ICP MS and integrations with common Pb are not plotted.

#### Results

Only  $^{206}$ Pb/ $^{238}$ U ages are considered for the interpretation of the young samples from Kos and Nisyros, because of the low intensities of  $^{207}$ Pb and  $^{208}$ Pb and the larger influence from small amounts of common Pb on the  $^{208}$ Pb/ $^{232}$ Th and  $^{207}$ Pb/ $^{235}$ U ages. The accuracy of the method in general is shown using secondary standards. Unfortunately no suitable young secondary SRM was available at the time of measurements. Resulting ages are compared with accepted (ID-TIMS) ages. All three secondary standards (Plesovice: Measured 338.3 ± 1.7 Ma ref: 337.13 ± 0.37 ma<sup>8</sup>, Temora: 421.6 ± 2.0 Ma ref: 416.8 ± 1.1 ma<sup>45</sup>, 91500: 1064.8 ± 3.3 Ma ref:1065 ± 2 ma<sup>46</sup>) agree reasonably well with ID-TIMS reference values. Precision of

these secondary SRM's was found to be between 0.5% and 1%, 2 standard error for 8-18 replicate analyses and about 1% (2 SE) on individual point analysis. The MSWDs obtained for the secondary standards (Table 1) are as high as 3.4 which could either be due to standard heterogeneity or an overestimated precision. All data can be found in the electronic supplementary information following the guidelines suggested by the Earthtime initiative<sup>32</sup>.

The approach of using an old SRM to quantify these very young zircons might be affected by small differences in the ablation behaviour between the SRM and the samples, especially the downhole fractionation. Parameters that can have and influence include the amount of radiation damage, trace element composition, transparency, crystal orientation and carrier gas composition and are currently under investigation

2

3

4

5

6

7

8

9

10

11

12

13

14

15

16

17

18

19

20

21

22

23

24

25

26

27

28

29

30

31

32

33

34

35

36

37

38

39

40

41

42

43

44

45

46

47

48

49

50

51

52

53

54

55

56


57

58

59 60

The comparison between age data acquired with LA-ICP-MS in this study with the published data using SHRIMP<sup>27</sup> shows a remarkable agreement. Zircon ages cover the same range between the eruption age of 161 ka (40Ar/39Ar sanidine dating<sup>23</sup>) and approximately 350 ka (Figure 1), with very similar uncertainties (although errors on Th-corrections were not propagated in the SHRIMP results<sup>27</sup>). Ages are evenly distributed in this range, with no major gaps in crystallization that can be seen within the precision of the method (Figures ESI 8 and 9). It reinforces previous interpretations that zircons have been crystallizing over an extended period of time (> 140 ka), recording the growth and maturation of the silicic magma reservoir in the upper crust<sup>27</sup>. Based on the precision we obtain, it is not possible to distinguish between several discrete crystallisation events and a real continuous crystallisation in the given time frame.

ID-TIMS measurements on 7 selected grains (Table ESI 2) show again a remarkably overlap with SHRIMP and LA-ICP-MS ages. The age of 6 grains falls well within the range of all SHRIMP and LA-ICP-MS data. One zircon with an age of 410  $\pm$  27 ka can be treated as outlier. The distribution of ID-TIMS ages from 187  $\pm$  15 ka to 262  $\pm$  16 ka support both the accuracy of the LA-ICP-MS measurements and the interpretation of zircon crystallization over an extended time. All details on the ID-TIMS method and measurements is given in von Quadt et. al 2014.<sup>34</sup>



Although no old grains where found using SHPIMP two

Although no old grains where found using SHRIMP, two inherited cores stand out in the LA-ICP-MS analyses. Precise ages on those cores cannot be given, as they are a mixture of young overgrowth and old core, but they are very old, with ages in the range of 1 to 2 Ga (see ESI, Table 1). In general, LA-ICP-MS is much more sensitive in finding inherited cores as LA removes about 10 times more material per analysis than SHRIMP (crater depth for LA is 10-15 microns while SHRIMP craters are only 1-2 microns deep).

As our LA-ICP-MS method has shown its high degree of accuracy and precision especially in comparison to ID-TIMS data, we dated 3 younger, hitherto undated samples from the nearby Nisyros volcano. Using the youngest zircons as the best estimate of the eruption age, results yield (1)  $124 \pm 35$  ka for the Lower Pumice (NISP2(1)), (2)  $111 \pm 42$  ka for Nikia Flow

ARTICLE

(NS24), and 70  $\pm$  24 ka for Upper Pumice (NS07) as shown in Figure 2. All errors are given as 2 SE. Only a limited number of zircons could be separated from these samples, especially NS24, as they may have been zircon-undersaturated prior to eruption<sup>25, 26, 49</sup>. By analysing more zircon grains from these samples it is likely that the spread increases, therefore the youngest zircon found can only be interpreted as maximum ages.

#### Impact of data reduction on final age

We investigated the effects of data reduction on final ages. The selection of integration intervals in U/Pb dating appears to be a critical aspect of the data reduction process. Modern data reduction software can give instant feedback on the resulting age while selecting integration intervals. These instant feedback tools can be used to recognize heterogeneities, inclusions and inherited cores when drilling with the laser through a zircon grain. Figure 3 provides an example of such a procedure (with full data set in the electronic supplementary information). The ablation consisted of two parts: the first integration of 10 seconds has a resulting age of  $210 \pm 31$  ka with a corresponding Th/U ratio of  $0.440 \pm 0.003$ , while the second integration interval is  $260 \pm 29$  ka old with a Th/U ratio of  $0.93 \pm 0.03$ . Based on the variable Th/U ratio, we assume that this zircon grain was growing over an extended period, during which the composition and/or the fractionation  $f_{(Th/U)}$  changed. Several zircon grains showed similar characteristics, and were similarly treated in the data reduction process. All ages that do not show enhanced common Pb were taken into account for figures 1 and 2 and are listed in the electronic supplementary information with a comment. As a detailed and general description of a data reduction protocol may have to be adjusted to each sample, method, instrumentation and software used, it is beyond the focus of this work.




Figure 3: LA-ICP-MS signal of a single zircon grain ablation showing variable U, Th intensities and Th/U and  $^{206}$ Pb/ $^{238}$ U ratios when drilling into the sample. It is important to set integration intervals carefully and correctly. Th/U ratio changes between integration intervals from 0.44 to 0.93 while the corrected  $^{206}$ Pb/ $^{238}$ U age changes from 210 ka to 260 ka. It is interpreted that this zircon has seen 2 crystallization periods in the magma chamber.

The different Th disequilibrium corrections (equations 1 and 3) also have a significant impact on the final age for young samples. Using sample NS07, we show that the higher the measured Th/U, the less is the correction (Figure 4). The correction is as previously reported in the range of 70-110 ka<sup>41</sup> towards older age compared with the uncorrected age. For ages above 300 ka, the correction model of either equation (1) or equation (3) has limited influence on the final age. For ages

2

3

4

5

6

7

8

9

10

11

12

13

14

15

16

17

18

19

20

21

22

23

24

25

26

27

28

29

30

31

32

33

34

35

36

37

38

39

40

41

42

43

44

45

46

47

48

49

50

51

52

53

54

55

56

57

58

59 60 Journal Name

younger than 300 ka, the correction model becomes important. A zircon with an age as young as 70 ka calculated using equation (3), is more than 50% older (110 ka) when using equation (1), which is outside the error bar in that case.




Figure 4: Influence of the initial Th disequilibrium correction model on the final age of sample NS 07 in comparison with the uncorrected age and the Th/U ratio. The younger the zircons, the larger are the differences between the two correction models. Grey lines show the age range when the Th/U ratio in the magma changes from 2 to 8. Th/U = 4 was measured in whole rock for this sample.

The Th/U ratio in the magma measured by whole rock data or assumed, influences in a limited way the final age as was investigated and is shown in Figure 4. The grey lines represent the corrected ages using two different Th/U ratios for the magma. The measured Th/U ratio was found to be 4 for this sample (whole rock analysis, table 1). Changing the Th/U in the magma to either 8 or 2, instead of 4, changes the corrected age by about +10 or -20 ka respectively, which is always within the uncertainty. As the variation of Th/U in the zircons is lower than a factor 4, all corrected ages assuming a constant fractionation factor  $f_{(Th/U)}$  and varying Th/U ratios in the magma would fall between the two grey lines. Based on these findings, the method including the Th disequilibrium correction gives accurate results for the LA-ICPMS technique. However, more precise and accurate Th/U ratios in the magma are needed for high precision ID-TIMS dating. A possibility might be analysing melt inclusions in zircons that represent the local magma composition and fractionation factors can be determined for individual zircons.

#### Conclusions

This work shows that the in-situ LA-ICP-MS method is well suited to date zircons as young as 50 ka. Applying our improved initial thorium disequilibrium correction, results are accurate based on comparison between three different methods: SHRIMP, ID-TIMS and LA-ICP-MS applied to zircons from KPT. Precision of ages acquired via LA-ICP-MS for zircons in the range of 70 ka - 500 ka is comparable with data collected with SHRIMP. ID-TIMS results for such young zircons are a factor of 2-3 more precise. We stress, however, that as more material is consumed with LA-ICP-MS, much more information is acquired which obviously can also be a drawback when analyses of very small features like thin rims are required. In addition, running costs are low, and hundreds of zircons can easily be analysed per day. The limiting step becomes sample preparation as well as data reduction and interpretation of the results after analysis. Due to the large sample consumption (about 75 nm per laser pulse), zircons with inclusions not detected by CL images can lead to common Pb contamination. Hence, we recommend, when possible, to

measure additional elements (<sup>31</sup>P, <sup>27</sup>Al, <sup>56</sup>Fe) to recognize the presence of these inclusions.

To further improve the method, a reliable common Pb correction is necessary. Possibilities include better detection of <sup>207</sup>Pb and the <sup>207</sup>Pb/<sup>206</sup>Pb ratio by increasing repetition rate and/or crater size and increasing the dwell time on mass 207, keeping <sup>238</sup>U in the pulse counting mode to avoid detector cross calibration problems. Due to the low count rate on Pb isotopes, a bias in calculated ratios is possible as the mean-of-ratios can be different from the ratio-of-means. A simple comparison between the two calculations showed a bias of the <sup>206</sup>Pb/<sup>238</sup>U ratio smaller than 3% for sample NS07 well below the precision of these measurements and without down-hole fractionation correction.

Th disequilibrium correction is essential for zircons younger than 10 Ma and the improved equation (3) should be used for zircons younger 0.5 Ma. The influence from the uncertainty of the  $f_{(Th/U)}$  is relatively small in comparison to the overall uncertainty for analysis with LA-ICP-MS. Nevertheless errors from the correction should be propagated.

The evenly spread age distribution of the Kos Plateau Tuff samples (from 320 ka to 180 ka) is interpreted as recording near-continuous crystallisation of zircons in a persistent magma chamber. A Probability density plot (Figure ESI 8) does not show a normal distribution of the KPT ages that would be expected from a sharp single crystallisation age. Therefore the calculation of a weighted mean of all the 132 zircons from all 3 techniques (SHRIMP, ID-TIMS and LA-ICP-MS, figure ESI 9) is meaningless in this case, as it results in an weighted mean age of  $238.3 \pm 7.1$  ka with a MSWD of 7.8 (see also previous discussion<sup>50</sup>).

The fact that only 2 inherited cores were found, despite the very large number of analyses, suggests that the silicic magma reservoir was mostly isolated from its upper crustal wall rocks. Such a near closed-system fractionation is confirmed by the unradiogenic isotopic ratios of the Kos Plateau Tuff<sup>25, 27, 51</sup> and is consistent with thermal models of upper crustal reservoirs that interacted in a limited way with their wall rocks<sup>27, 50, 52, 53</sup>. The fact that only very few zircons were found with ages in the range of 20 ka before eruption is also consistent with the interpretation that the magma was rejuvenated prior to eruption by fresh recharge<sup>24</sup>, reducing the growth and even causing resorption of zircons.

The ages of the youngest zircons found in the Nisyros samples agree with previously-determined stratigraphy<sup>25, 54</sup> and provide the first direct geochronological estimates of the eruption ages for the following three deposits: (1) Lower Pumice, (2) Nikia Flow, and (3) Upper Pumice. We stress that these are maximum ages, as they may all be zircon-undersaturated prior to eruption<sup>26</sup>. It demonstrate that the Kos-Nisyros system took at least several tens of thousands of years (from 160 ka, eruption of KPT to ~ 120-130 ka, maximum eruption age of the lower Pumices on Nisyros)) to generate another explosive silicic eruption after the caldera-forming Kos Plateau Tuff. All three units also display an extended range of zircon ages, suggesting similar crystal recycling as observed for the Kos Plateau Tuff<sup>2</sup> and many other eruptions (e.g., see reviews by Costa, 2008 and Simon et al., 2008)<sup>41, 55</sup>. A comparison between zircons with and without chemical abrasion using LA-ICP-MS are ongoing projects with the Kos Plateau Tuff and related units<sup>34</sup>.

#### Acknowledgements

We like to thank Tilman Jeske for the preparation of the Nisyros samples. Critical reading by Ben Ellis helped to

Accepted Manu

rnal of Analytical Atomic

46

47

48

49

50

51

52

53

54

55

56

57

58 59 60 Journal Name improve the manuscript. Comments especially on the common Pb correction by two unknown referees further improve the work. This work was in part supported by Swiss SNF grant 200021 146268 (to OB). Notes and references <sup>a</sup> Department of Earth Sciences, ETH Zurich, 8092 Zurich, Switzerland. <sup>b</sup> Department of Geology and Mineralogy, Kyoto University, Kyoto 606-28 8502. Japan Electronic Supplementary Information (ESI) available: [Table with all results of LA-ICP-MS and ID-TIMS analyses, Figures ESI 1-9, disequilibrium Th Derivation of the correction]. See DOI: 10.1039/b00000x/ 1. B. J. Fryer, S. E. Jackson and H. P. Longerich, Chemical Geology, 1993, 109, 1-8. DOI: 10.1016/0009-2541(93)90058-q. S. Johnston, G. Gehrels, V. Valencia and J. Ruiz, Chemical Geology, 2. 2009, 259, 218-229. DOI: 10.1016/j.chemgeo.2008.11.004 W. Muller, M. Shelley, P. Miller and S. Broude, Journal of Analytical 3 Atomic Spectrometry, 2009, 24, 209-214. DOI: 10.1039/b805995k. D. Frei and A. Gerdes, Chemical Geology, 2009, 261, 261-270. DOI: 4. 10.1016/j.chemgeo.2008.07.025. 35. 5. B. Shaulis, T. J. Lapen and A. Toms, Geochemistry Geophysics Geosystems, 2010, 11. DOI: 10.1029/2010gc003198. 6. U. Kloetzli, E. Kloetzli, Z. Guenes and J. Kosler, Geochimica Et Cosmochimica Acta, 2007, 71, A495-A495. 7. Y. Orihashi, S. i. Nakai and T. Hirata, Resource Geology, 2008, 58, 101-123. DOI: 10.1111/j.1751-3928.2008.00052.x. 8. J. Slama, J. Kosler, D. J. Condon, J. L. Crowley, A. Gerdes, J. M. Hanchar, M. S. A. Horstwood, G. A. Morris, L. Nasdala, N. Norberg, U. Schaltegger, B. Schoene, M. N. Tubrett and M. J. Whitehouse, Chemical Geology, 2008, 249, 1-35. DOI: 10.1016/j.chemgeo.2007.11.005. U. Kloetzli, E. Kloetzli, Z. Guenes and J. Koslar, Geostandards and 9. Geoanalytical Research, 2009, 33, 5-15. DOI: 10.1111/j.1751-908X.2009.00921.x. 10. C. Paton, J. Hellstrom, B. Paul, J. Woodhead and J. Hergt, Journal of Analytical Atomic Spectrometry, 2011, 26, 2508-2518. DOI: 40 10.1039/c1ja10172b. 11. C. Paton, J. D. Woodhead, J. C. Hellstrom, J. M. Hergt, A. Greig and R. Maas, Geochemistry Geophysics Geosystems, 2010, 11. DOI: 10.1029/2009gc002618. 42. 12. J. A. Petrus and B. S. Kamber, Geostandards and Geoanalytical Research, 2012, 36, 247-270. DOI: 10.1111/j.1751-908X.2012.00158.x. 13. E. Butler, W. Winkler and M. Guillong, Swiss Journal of Geosciences, 2011, 104, 225-236. DOI: 10.1007/s00015-011-0065-1.

- 14. J. Kosler, J. Slama, E. Belousova, F. Corfu, G. E. Gehrels, A. Gerdes, M. S. A. Horstwood, K. N. Sircombe, P. J. Sylvester, M. Tiepolo, M. J. Whitehouse and J. D. Woodhead, Geostandards and Geoanalytical Research, 2013, 37, 243-259. DOI: 10.1111/j.1751-908X.2013.00245.x.
- 15. E. Buetler, W. Winkler and M. Guillong, Swiss Journal of Geosciences, 2011, 104, 225-236. DOI: 10.1007/s00015-011-0065-1.
- 16. T. D. Yokoyama, T. Suzuki, Y. Kon and T. Hirata, Analytical Chemistry, 2011, 83, 8892-8899. DOI: 10.1021/ac2012448.
- 17. J. Kosler, M. Wiedenbeck, R. Wirth, J. Hovorka, P. Sylvester and J. Mikova, Journal of Analytical Atomic Spectrometry, 2005, 20, 402-409. DOI: 10.1039/b41629b.
- 18. B. K. Kuhn, K. Birbaum, Y. Luo and D. Guenther, Journal of Analytical Atomic Spectrometry, 2010, 25, 21-27. DOI: 10.1039/b917261k.
- 19. G. E. Gehrels, V. A. Valencia and J. Ruiz, Geochemistry Geophysics Geosystems, 2008, 9. DOI: 10.1029/2007gc001805.
- 20. L. A. Solari, A. Gomez-Tuena, J. Pablo Bernal, O. Perez-Arvizu and M. Tanner, Geostandards and Geoanalytical Research, 2010, 34, 5-18.
- 21. E. Kooijman, J. Berndt and K. Mezger, European Journal of Mineralogy, 2012, 24, 5-21. DOI: 10.1127/0935-1221/2012/0024-2170.
- 22. U. Schärer, Earth and Planetary Science Letters, 1984, 67, 191-204. DOI: 10.1016/0012-821x(84)90114-6.
- 23. P. E. Smith, D. York, Y. Chen and N. M. Evensen, Geophysical Research Letters, 1996, 23, 3047-3050. DOI: 10.1029/96gl02759.

- 24. O. Bachmann, B. Schoene, C. Schnyder and R. Spikings, Geochemistry Geophysics Geosystems, 2010, 11. DOI: 10.1029/2010gc003073
- 25. L. Francalanci, J. C. Varekamp, G. Vougioukalakis, M. J. Defant, F. Innocenti and P. Manetti, Bulletin of Volcanology, 1995, 56, 601-620.
- 26. O. Bachmann, C. D. Deering, J. S. Ruprecht, C. Huber, A. Skopelitis and C. Schnyder, Contributions to Mineralogy and Petrology, 2012, 163, 151-166. DOI: 10.1007/s00410-011-0663-y.
- 27. O. Bachmann, B. L. A. Charlier and J. B. Lowenstern, Geology, 2007, 35, 73-76. DOI: 10.1130/g23151a.1.
- S. J. A. Brown and I. R. Fletcher, Geology, 1999, 27, 1035-1038. DOI: 10.1130/0091-7613(1999)027<1035:supdot>2.3.co;2.
- 29. S. Eggins, R. Grun, A. W. G. Pike, M. Shellev and L. Tavlor, Quaternary Science Reviews, 2003, 22, 1373-1382. DOI: 10.1016/s0277-3791(03)00064-7.
- 30. W. Muller, J. M. G. Shelley and S. O. Rasmussen, Journal of Analytical Atomic Spectrometry, 2011, 26, 2391-2395. DOI: 10.1039/c1ja10242g.
- 31. J. Woodhead, J. Hergt, M. Shelley, S. Eggins and R. Kemp, Chemical Geology, 2004, 209, 121-135. DOI: 10.1016/j.chemgeo.2004.04.026.
- 32. https://cirdles.org/LA-ICP\_MS\_Data\_Handling, ed. M. S. A. Horstwood, J. Kosler, S. Jackson, N. Pearson and P. Sylvester. 2013.
- 33. M. Guillong and C. A. Heinrich, Journal of Analytical Atomic Spectrometry, 2007, 22, 1488-1494. DOI: 10.1039/b709489b.
- 34. A. von Quadt, D. Gallhofer, M. Guillong, I. Peytcheva, M. Waelle and S. Shuhei, Journal of Analytical Atomic Spectrometry, 2014, in Review, this volume.
- S. E. Jackson, N. J. Pearson, W. L. Griffin and E. A. Belousova, Chemical Geology, 2004, 211. 47-69. DOI: 10.1016/j.chemgeo.2004.06.017.
- 36. A. H. Jaffey, K. F. Flynn, Glendeni.Le, W. C. Bentley and A. M. Review *C*, 1971, **4**, 1889-&. Essling. Physical DOE 10.1103/PhysRevC.4.1889.
- 37. H. Cheng, R. L. Edwards, C. C. Shen, V. J. Polyak, Y. Asmerom, J. Woodhead, J. Hellstrom, Y. J. Wang, X. G. Kong, C. Spotl, X. F. Wang and E. C. Alexander, Earth and Planetary Science Letters, 2013, 371, 82-91. DOI: 10.1016/j.epsl.2013.04.006.
- 38. A. Cocherie, C. M. Fanning, P. Jezequel and M. Roberta, Geochimica Et 1095-1108. Cosmochimica Acta. 2009, 73, DOI: 10.1016/j.gca.2008.11.028.
- 39. T. Andersen, Chemical Geology, 2002, 192, 59-79. DOI: 10.1016/s0009-2541(02)00195-x.
- M. Chiaradia, U. Schaltegger, R. Spikings, J. F. Wotzlaw and M. Ovtcharova, Economic Geology, 2013, 108, 565-584.
- 41. J. I. Simon, P. R. Renne and R. Mundil, Earth and Planetary Science Letters, 2008, 266, 182-194. DOI: 10.1016/j.epsl.2007.11.014.
- N. M. McLean, J. F. Bowring and S. A. Bowring, Geochemistry Geophysics Geosystems, 2011, 12. DOI: 10.1029/2010gc003478.
- 43. M. Rioux, C. J. Lissenberg, N. M. McLean, S. A. Bowring, C. J. MacLeod, E. Hellebrand and N. Shimizu, Nature Geoscience, 2012, 5, 275-278. DOI: 10.1038/ngeo1378.
- 44. S. Sakata, S. Hirakawa, H. Iwano, T. Danahara and T. Hirata, 2013, Mineralogical Magazine, 77, 2116. DOI 10.1180/minmag.2013.077.5.19.
- L. P. Black, S. L. Kamo, C. M. Allen, J. N. Aleinikoff, D. W. Davis, R. 45. J. Korsch and C. Foudoulis, Chemical Geology, 2003, 200, 155-170. DOI: 10.1016/s0009-2541(03)00165-7.
- 46. M. Wiedenbeck, P. Alle, F. Corfu, W. L. Griffin, M. Meier, F. Oberli, A. Vonquadt, J. C. Roddick and W. Speigel, Geostandards Newsletter, 1995, 19, 1-23. DOI: 10.1111/j.1751-908X.1995.tb00147.x.
- 47. E. Marillo Sialer, J. Woodhead, J. M. Hergt, A. Greig, M. Guillong, A. Gleadow, N. Evans and C. Paton, Journal of Analytical Atomic Spectrometry, 2014. DOI: 10.1039/C4JA00008K.
- 48. J. Kosler, S. Jackson, Z. Yang and R. Wirth, Journal of Analytical Atomic Spectrometry, 2014. DOI: 10.1039/C3JA50386K.
- 49. L. Vanderkluysen, A. Volentik, C. Principe, J. C. Hunziker and J. Hernandez, in The petrology and geochemistry of lavas and tephras of Nisyros Volcano (Greece), ed. J. C. Hunziker and L. Marini, Lausanne. 2006, vol. 44, pp. 100-106.
- 50. O. Bachmann, F. Oberli, M. A. Dungan, M. Meier, R. Mundil and H. Fischer, Chemical Geology, 2007, 236, 134-166. DOI: 10.1016/j.chemgeo.2006.09.005.
- 51. A. Buettner, I. C. Kleinhanns, D. Rufer, J. C. Hunziker and I. M. Villa, Lithos, 2005, 83, 29-46. DOI: 10.1016/j.lithos.2005.01.001.

Journal of Analytical Atomic Spectrometry Accepted Manusc

- 52. J. Dufek and G. W. Bergantz, Journal of Petrology, 2005, 46, 2167-2195. DOI: 10.1093/petrology/egi049.
- 53. S. E. Gelman, F. J. Gutierrez and O. Bachmann, Geology, 2013, 41, 759-762. DOI: 10.1130/g34241.1.
- 54. A. Volentik, L. Vanderkluysen and C. Principe, Eclogae Geologicae Helvetiae, 2002, 95, 223-235.
- 55. F. Costa, R. Dohmen and S. Chakraborty, in Minerals, Inclusions and Volcanic Processes, ed. K. D. Putirka and F. J. Tepley. 2008, vol. 69, pp. 545-594.



Dating of zircons in the 100 ka range is possible with LA-ICP-MS applying Th-disequilibrium correction.

