

# Lab on a Chip

Accepted Manuscript



This is an *Accepted Manuscript*, which has been through the Royal Society of Chemistry peer review process and has been accepted for publication.

*Accepted Manuscripts* are published online shortly after acceptance, before technical editing, formatting and proof reading. Using this free service, authors can make their results available to the community, in citable form, before we publish the edited article. We will replace this *Accepted Manuscript* with the edited and formatted *Advance Article* as soon as it is available.

You can find more information about *Accepted Manuscripts* in the [Information for Authors](#).

Please note that technical editing may introduce minor changes to the text and/or graphics, which may alter content. The journal's standard [Terms & Conditions](#) and the [Ethical guidelines](#) still apply. In no event shall the Royal Society of Chemistry be held responsible for any errors or omissions in this *Accepted Manuscript* or any consequences arising from the use of any information it contains.

# Droplet microfluidic system for sequential generation of lipid bilayers and transmembrane electrical recordings

Magdalena A. Czekalska<sup>a</sup>, Tomasz S. Kaminski<sup>a</sup>, Slawomir Jakiela<sup>a</sup>, K. Tanuj Sapra<sup>b</sup>, Hagan Bayley<sup>b\*</sup> and Piotr Garstecki<sup>a\*</sup>

<sup>5</sup> Received (in XXX, XXX) Xth XXXXXXXXX 20XX, Accepted Xth XXXXXXXXX 20XX

DOI: 10.1039/b000000x

This paper demonstrates a microfluidic system that automates i) formation of a lipid bilayer at the interface between a pair of nanoliter-sized aqueous droplets in oil, ii) exchange of one droplet of the pair to form a new bilayer, and iii) current measurements on single proteins. A new microfluidic architecture is introduced - a set of traps designed to localize the droplets with respect to each other and with respect to the recording electrodes. The system allows for automated execution of experimental protocols by active control of the flow on chip with the use of simple external valves. Formation of stable artificial lipid bilayers, incorporation of  $\alpha$ -hemolysin into the bilayers and electrical measurements of ionic transport through the protein pore are demonstrated.

## 15 Introduction

Efficient screening of the function of membrane proteins against physical and chemical factors demands a reproducible and cost-effective method for the generation of lipid bilayers and for measuring electrical currents through channels and pores inserted in them. The widely used patch clamp analysis enables recordings of electric currents through protein channels in their natural cellular environment or in artificial lipid vesicles. However, classical patch clamp demands skilled manual operation thereby limiting its throughput in routine drug screening assays. The high demand for screens of protein function has prompted several attempts to increase throughput and to automate the measurements.<sup>1</sup> These methods require control of protein expression in cells, which can be challenging,<sup>2</sup> and provide limited information on activity, usually limited to IC<sub>50</sub> values.

Methods for in vitro formation of bilayer lipid membranes (BLM) – the variations of the Montal-Mueller<sup>3</sup> or painting technique are, in turn, limited by the low stability of the bilayers.<sup>4</sup> Formation of bilayers on a solid support significantly increases the stability, however the format is prone to the formation of incomplete membranes.<sup>5,6</sup> Moreover proteins may interact unfavorably with the solid substrate and their mobility and function can be modified or hampered making electrical measurements difficult, especially at the single molecule level.<sup>6</sup>

A convenient method for forming a stable lipid bilayer is by bringing the lipid monolayer-coated interfaces of two aqueous droplets into contact in oil.<sup>7–9</sup> Compared to the classical planar bilayer method, Droplet Interface Bilayers (DIBs) show critical advantages of exceptional stability that supports long-term measurements, and minute, micro-liter consumption of samples.<sup>4</sup> The stability of the interface and the ability to easily re-form the bilayer make the DIB method an attractive strategy for implementation in robust inhibitor screening schemes.<sup>9,10</sup> The main challenge is in automation of the technique as implementing automation may evolve the technique to a high-throughput screening platform. Here we show, for the first time, repetitive

formation of DIBs, separation of the droplets and exchange of at least one droplet of a pair to form a new DIB, all in an automated sequence.

Droplet microfluidics has successfully been used in high-throughput screening to address various problems in biochemistry<sup>11,12</sup> with primary applications in protein crystallization,<sup>13</sup> single cell screening,<sup>14</sup> directed evolution of enzymes<sup>15</sup> and the clonal selection of antibodies.<sup>16</sup> Droplet microfluidics is an ideal tool for reproducible formation of droplets with controlled composition of their interfaces,<sup>17</sup> including lipid monolayers.

Still, the research on droplet microfluidic techniques for in vitro reconstitution and electrophysiological studies with lipid bilayers has so far been limited. In a pioneering report, Funakoshi *et al.*<sup>8</sup> built microfluidic chips in which a bilayer is formed between two aqueous streams surrounded with an oily fluid. The formation of a true DIB was confirmed by capacitance measurements of the bilayer and voltage-clamp recording indicating incorporation of  $\alpha$ -hemolysin pores. The DIB system was proved to be useful in investigation of other proteins, for example bacterial potassium channels<sup>10,18,19</sup> or eukaryotic mammalian menthol receptor or equinatoxin<sup>22</sup>.

So far many microfluidic platforms and other chip technologies have been developed, usually lacking the possibility of automated serial formation of lipid bilayers.<sup>23,24,18,25–35,20,36–38,21,39,40</sup> Highly parallelized systems capable of the creation of arrays of lipid bilayers usually rely solely on measurements of fluorescence<sup>32,35,41–43</sup> that does not provide for complete characterization of the function of the pores.

Here, we present an automated microfluidic system that combines the active manipulation of droplets<sup>44,45</sup> with passive trapping,<sup>46–48</sup> and the capability to insert electrodes into the droplets for in situ electrical measurements on membrane proteins incorporated into a DIB.




Figure 1. Geometry of the microfluidic device. a) Schematic diagram of the microfluidic chip: A, E, J, K, P – inlets for oil, B, I – inlets for aqueous samples, C – aspiration module with pure buffer for washing of electrodes, D, G – T-junctions, F – tubing with a sequence of droplets containing inhibitors, H – aspiration module with a solution of proteins, M – ground electrode, N – hydrodynamic trap, O – working electrode, S – inlet for oil in a control channel, L, R – outlets b) Schematic picture of the microfluidic trap. The DIB is formed at the interface of two droplets locked in the trap filled with oil. The light-grey area represents so called ‘bypasses’ – shallow channels on the sides of the trap (dark grey). The continuous liquid (oil) passes through the bypasses while the droplets are locked in the position. (For details of the trap architecture see Figure S2). c) Fragment of current trace (20 s) reflecting the incorporation of  $\alpha$ -hemolysin into a lipid bilayer, clamped at + 50 mV. Stepwise increase of current from 0 to 50 pA reflects the incorporation of a single pore. The presence of inhibitor ( $\gamma$ -cyclodextrin, 5  $\mu$ M) within the pore is seen as a decrease of current by app. 60%. The presence of next  $\alpha$ HL heptamer in a lipid bilayer is seen as a further stepwise increase of current. BL – baseline, O – open pore level, 2x O – 2 open pores, B – blocked pore, 2x b – 2 blocked pores.




Figure 2. Micrographs illustrating the orientation of the electrodes in the microfluidic trap and the exchange of one of the droplets. a) Two droplets are trapped. b) Bilayer (indicated with an arrow) is formed at the interface. c-e) Removal of one of the droplets. f) Introduction of the new droplet. The droplet wets the electrode immediately after being placed in a trap (see details in ESI). Scale bar is 200  $\mu$ m.

For screening applications it is essential that the microfluidic system is capable of automated generation and manipulation of multiple droplets. Droplet on demand systems<sup>49</sup> provide a high degree of control over the composition and location of each droplet. Although, the generation of DIBs and networks of droplets has been demonstrated on a chip,<sup>32,35,50,51</sup> a remaining challenge was to interface electrodes for repetitive measurements of bilayer capacitance and of conductance through single channels and pores.

## Results

### Microfluidic device

The microfluidic device (Figure 1) comprises 2 polycarbonate plates, milled with an Ergwind msc4025 CNC milling station and

subsequently thermally bonded in a hydraulic press. The width of the channels is 400  $\mu$ m (excluding channels no. 14 and 15 [Figure S1] which have width of 200  $\mu$ m) The depth of all channels is 400  $\mu$ m. The device consists of i) a microfluidic trap (N, Fig 1a) in which droplets are contacted to form a DIB, ii) two microfluidic T-junctions (D and G) with additional sample-ports, and iii) an inlet port for a sequence of droplets containing various inhibitors (F). Two silver wires coated with AgCl and a thin layer of agarose gel serve as an electrode pair. One electrode is inserted into each of the two chambers of the trap prior to running an assay.

### Operation of a chip

We first filled the whole system with a continuous organic phase containing lipids (1 mg mL<sup>-1</sup> 1,2-diphytanoyl-*sn*-glycero-3-phosphocholine [DPhPC] in 75% hexadecane and 25% silicone oil AR20). We then opened the port (B) and transferred a small volume (tens of microliters) of buffer (1 M KCl, 10 mM Tris-HCl, pH 7.0) from a glass microsyringe connected to the chip via a capillary. We used the port (J) to introduce a small volume of the protein solution ( $\alpha$ -hemolysin ( $\alpha$ HL) wild-type, 5  $\mu$ g mL<sup>-1</sup>, diluted in the same buffer). After closing the deposition ports we opened the flow of oil from inlets (A and J) into the channel containing the sample to push the front of the aqueous aliquot to the T-junction. A sequence of droplets containing inhibitors was deposited on the chip in a similar manner (F and E).

We used valves to control the flow of oil from pressurized containers (Figure S3). Droplets were generated in T-junctions – a defined volume of the aqueous sample was pushed into the orthogonal channel and then broken-off with the flow of oil<sup>52</sup>

(Fig. 1a,  $K$  and  $P$ ) and transported into the trap ( $N$ ). The system (see Fig. 1b) comprises two cavities (dark grey in Fig. 1b) that lock the droplets by capillary forces in place. The shallow bypasses (marked with the light grey in Fig. 1b) allow for small flow of oil around the droplets without distorting them. In the experiments, we first generated and placed in one trap a droplet containing  $\alpha$ -hemolysin, then sequentially pushed droplets containing different concentrations of an inhibitor into the second trap.




Figure 3. Dependence of the measured bilayer capacitance on the volume of droplets. Each data point represent the average capacitance values of five droplet pairs; error bars represent the standard deviation (SD). The trend in the capacitance reflects the systematic dependence of the surface area of the bilayer on the volume of the droplets. The standard deviation of the recorded capacitance roots in small differences in the positioning of the droplets in the traps.

The pair of chambers in the hydrodynamic traps allows us to lock in place two droplets and observe the kinetics of formation of the DIB (Figure 2). In addition to tracking the process visually with a video camera we measured the capacitance of the DIB in absence of any membrane proteins. To this end we applied a triangular potential (10 Hz, 50 mV peak-to-peak) and recorded the resulting square wave current at a 1 kHz sampling rate. After the droplets first touch each other, it takes about 60 s for the bilayer to form.

### Characterization of the bilayer

Formation of a bilayer causes a rapid increase in the capacitive current  $I_C$ , that reaches typically  $394 \pm 13$  pA for two droplets of volume 300 nl each.

$$I_C = C_T \times \frac{dV}{dt} \quad (1)$$

Capacitive current  $I_C$  (Equation 1) is directly proportional to the capacitance  $C_T$  and the rate of change of potential, which was set to 1 V/s. Therefore, we find the value of the capacitance to be  $394 \pm 13$  pF for 300 nl droplets. The area of the bilayer can be determined through the measurement of the capacitive current<sup>53</sup>. Using the specific capacitance for a bilayer formed in pure hexadecane<sup>53</sup> ( $0.64 \mu\text{F cm}^{-2}$ ), we estimated the contact area between the 300 nl droplets to be  $61500 \pm 2500 \mu\text{m}^2$ . Our system allows us to trap droplets with volumes ranging between 250 nl and 350 nl and form lipid bilayers between them. This range of volumes allows for a nearly 4-fold range in the surface area of the DIB (Figure 3). Internal flows of oil (e.g. convection) cause

fluctuations of the capacitive current which is proportional to bilayer size, yet these changes are small and do not impact the process of pore insertion. Relative standard deviation (RSD) of single bilayer area is in the range of 1-6%. Additionally, subsequent pairs of droplets can have slightly different positions during formation of bilayer, thus the overall RSD for a given droplet volume is slightly higher (in the range of 4-10% - see detailed analysis in SI – Table S1).




Figure 4. Measurements of capacitive current, reflecting the size of the bilayer. The application of a constant flow of oil (respectively: a)  $5 \text{ nl s}^{-1}$ , b)  $10 \text{ nl s}^{-1}$  and c)  $15 \text{ nl s}^{-1}$ ) results in a decrease of the bilayer surface area. The vertical dotted red lines indicate the moment when the flow of oil was applied and when it was stopped. d) We were able to reduce the surface area of the bilayer from about  $85900 \mu\text{m}^2$  to  $31400 \mu\text{m}^2$ . The areas were estimated by the analysis of images with ImageJ with the assumption that the contact area is circular. The ion current across the bilayer was constant and equal to  $\sim 2330$  pA at a potential of +50 mV, which indicated that the number of pores in the membrane remained constant throughout the measurement. The vertical arrows in inset d) indicate the beginning and ending of a period of 45 s during which the droplets were partially separated by the flow of oil ( $15 \text{ nl s}^{-1}$ ).

### $\alpha$ -hemolysin measurements

In order to test the functionality of the DIB we trapped a pair of  $\sim 300$  nl droplets, one containing a solution of WT  $\alpha$ -hemolysin and the second with only the buffer. After formation of a DIB, we conducted a voltage clamp measurement at +50 mV. We observed the incorporation of single heptameric protein pores as step-wise (50 pA) increases in the current. Subsequently, we replaced the buffer droplet with a droplet containing a  $50 \mu\text{M}$   $\gamma$ -cyclodextrin ( $\gamma\text{CD}$ ), which is a reversible non-covalent blocker of the  $\alpha$ -hemolysin pore. Pore blockades were observed as transient decreases in the current to  $\sim 40\%$  of its open-pore value. Next, we checked whether residual inhibitor remained on the electrode after the droplet containing  $\gamma\text{CD}$  was replaced by a droplet with buffer alone. Rare blockades (on average 1.5 events per minute compared to  $\sim 80 \text{ min}^{-1}$  prior to washing) were recorded. Repetition of the exchange of the droplet completely eliminated the blockades (Figure S5). These results illustrate efficient washing of the electrodes.

5




Figure 5. a) Rapid screening of inhibitors against a single  $\alpha$ -hemolysin channel. Short segments of current traces in the presence of  $\gamma$ -cyclodextrin at various concentrations (1  $\mu$ M - 50  $\mu$ M). b) Concentration dependence of the inhibition of a single  $\alpha$ HL pore by  $\gamma$ -cyclodextrin.  $P_{open}$  was plotted versus the concentration of the blocker (c). The error bars represent the SD from at least three trials. The sum of all the measurement intervals (recordings of current) included in the calculation of the probability of blocking of the pore at each concentration was at least 145 s (see details in table S2). The curve was plotted using the saturation function (Equation 2) and the dissociation constant ( $K_d$ ) found to be  $61 \pm 7 \mu$ M.

$$P_{open} = \left(1 + \frac{c}{K_d}\right)^{-1} \quad (2)$$

### Control of the surface area

We are able to change the surface area of the DIB formed between the two droplets by varying the rate of flow of oil injected from the perpendicularly oriented additional channel (Fig. 1a, S). This flow pushed the interfaces apart with the magnitude of the effect depending on the rate of flow (Figure 4). The control over the surface area of the DIB allows us to influence the surface concentration of the inserted pores. After a high number (~45) of  $\alpha$ HL pores incorporated into the bilayer, we decreased the contact area. Despite decreasing the area of the bilayer, the current produced by the inserted pores remained the same (Fig. 4d, S8 and S9). Therefore, the number of functional protein molecules per unit area of the DIB had increased<sup>51</sup>.

### Screening experiment

We demonstrated the ability of our system to perform automated screening. We first prepared a sequence of 12 droplets presenting 6 different concentrations (duplicates of 1, 10, 20, 30, 40 and 50  $\mu$ M) of the inhibitor  $\gamma$ -cyclodextrin. The droplets were formed inside a 0.3 mm diameter PTFE tube using a syringe pump by alternating the aspiration of portions (300 nl) of aqueous solutions of different concentrations of inhibitor and oil admixed with lipids (700 nl) stored in a standard 96-well plate<sup>54</sup> (as shown in the Figure S6), demonstrating compatibility of the system with standard screening interfaces. Routing of the droplets containing inhibitors on the chip was controlled by optical feedback and external valves<sup>44,45</sup>.

In the screening experiments, we first trapped a droplet containing  $\alpha$ -hemolysin heptamers. Then we introduced the first droplet from the sequence of inhibitors. We voltage-clamped the bilayer at +50 mV and performed measurements of the current through the pore. Then we exchanged the droplet containing the

inhibitor with a fresh one, presenting a new concentration (preceded by the washing of electrode with a droplet of pure buffer). An exemplary trace of 4 subsequent exchanges is depicted in Figure S7. It is important to note, that the exchange of the droplets disrupts the bilayer, and dislodges the protein. As a result, most likely different molecules are studied in the measurements on subsequent droplet pairs. We observed, with a judiciously chosen concentration of protein (5  $\mu$ g mL<sup>-1</sup>), clear intervals between incorporation of the first and subsequent pores. This allows for the analysis of kinetics of inhibition at a single-pore level. After the recording, we removed the droplet containing the inhibitor and washed the electrode with a droplet comprising buffer alone generated automatically at a T-junction. The single cycle of removing a droplet, washing the electrode, placing a fresh droplet with new concentration of the inhibitor and performing the measurements took 210 s. Screening experiments in which we measured the current through the pores in the presence of inhibitors at different concentration lasted a multiplicity of the interval.

We were able to perform 2 repetitions of the screening with 6 concentrations of the blocker in each screening (Figure 5a) by using a single droplet containing  $\alpha$ -hemolysin. In about 50% of individual DIBs we observed the successful incorporation of at least one pore, which is consistent with the results of a previous studies<sup>38</sup>. Over time we noticed a decreased rate of pore insertion. We suppose that the decrease in the insertion rate is associated with aggregation of the diluted protein. After the 2 screening repetitions (40 min) we observed almost no insertion of pores within the arbitrarily set interval of 210 s.

Our system allows the replacement of the droplet containing the protein with a fresh one. In order to demonstrate this, we repeated the whole screening procedure five times, giving in total ten repetitions (2 repetitions in each screening experiment) of the

sequence of six different concentrations (a total of 60 measurements). Using the current traces obtained for 10-50  $\mu\text{M}$   $\gamma\text{CD}$ , we analyzed the duration and frequency of blocking and calculated the dissociation constant ( $K_d$ ) value for the binding of  $\gamma$ -cyclodextrin to  $\alpha$ -hemolysin to be  $61 \pm 7 \mu\text{M}$  (lit.,<sup>55</sup>  $47 \pm 9$ , see Figure 5b and description in SI).

## Discussion

We demonstrated for the first time an automated microdroplet system that enables fast, simple and reproducible screening of inhibitors against membrane proteins.

Up to now, droplet microfluidics have been used to generate droplet interface bilayers at a high yield<sup>32,35</sup> but without the ability to perform electrophysiological measurements on samples encapsulated in the droplets. In comparison to high-throughput systems relying on assay of diffusion of fluorescent dye through the pore,<sup>41-43</sup> our system is able to carry out reproducible single-pore electrophysiological recordings with high resolution including the on-off kinetics. The technique that we report here allows for automation of these measurements and for additional operations, such as washing and exchange of both the inhibitor solution and of the protein. The exclusion of manual handling of droplets is an important feature that increases the reproducibility of the protocol.

The droplet microfluidic system also allows the use of a minimal volume of sample. In our system this is achieved by the direct aspiration of tiny aqueous droplets (300 nl) from a micro-well plate and by modules that accept the microliter batches of liquids for further processing on chip. Our system is open for easy exchange and for automated processing of sequences of samples, an outstanding challenge that inspired a number of studies in the recent years.<sup>18-21,23,25,30,36,39</sup>

In addition, we demonstrated control of the area of a bilayer, both statically by tuning the volume of the droplets and dynamically by adjusting the rate of flow of an additional stream of oil. Interestingly, the decrease in bilayer area did not result in a decrease in the number of inserted pores, consistent with the results reported by Gross *et al.*<sup>56</sup> In the future, our system may be improved to permit the precise, automated control of the bilayer area and its use in the control of protein density in a bilayer.<sup>56</sup>

Additionally, the electrodes in our device did not show loss of activity of despite their intensive use – including repetitive wetting and de-wetting of electrodes with droplets - during a period of 3 weeks. Neither the layer of agarose gel, nor the layer of AgCl on the surface of silver wire, were depleted. We did not observe any changes in the performance of the chip over extended periods of operation, in particular we did not observe any wetting of the walls of the channels with the aqueous phase.

The increasing number of studies on various membrane proteins incorporated into droplet interface bilayers<sup>10,18-22</sup>, suggests that DIB may constitute an alternative to the BLM method, with the cautionary note, that each type of protein must be tested for compatibility with the droplet system.

## Experimental

### Reagents

1,2-Diphytanoyl-*sn*-glycero-3-phosphocholine (DPhPC, Avanti

Polar Lipids), hexadecane (Sigma Aldrich), silicone oil AR20 (Sigma Aldrich),  $\gamma$ -cyclodextrin (Cyclolab) were used as received. The buffer for washing electrodes, and protein and  $\gamma$ -cyclodextrin dilution, consisted of 1 M KCl (Sigma Aldrich), 10 mM TRIS.HCl (Roth), pH 7.0. The lipid solution was prepared by dissolving DPhPC (200 mg) in chloroform (10 mL, Chempur). The chloroform was evaporated under vacuum and the lipid film was re-solubilized in a mixture of 75 % v/v hexadecane and 25 % v/v silicone oil AR20.

### Protein expression and purification

WT  $\alpha\text{HL}$  was expressed in *Staphylococcus aureus*, converted to heptamer with deoxycholate and purified by preparative SDS-polyacrylamide gel electrophoresis as described.<sup>57</sup> The stock solution ( $50 \mu\text{g mL}^{-1}$ ) was kept on ice at all times and diluted 10-fold in buffer immediately before introduction onto the microfluidic chip. The rate of protein incorporation into a bilayer decreased with time. We did not notice any difference in performance when a diluted sample was kept at room temperature ( $20^\circ\text{C}$ ) for 2 h. We suppose that the decrease in activity is due to the aggregation of protein molecules after dilution.

### Electrical recordings

Ag/AgCl electrodes were used for electrical measurements. Silver wires of 100  $\mu\text{m}$  in diameter (Sigma) were treated overnight (12 h) with sodium hypochlorite solution (Sigma). The tips of the electrodes were submerged in melted agarose (1% w/v, Roth), diluted in 1M KCl, 10 mM Tris.HCl, pH 7.0 buffer. The AgCl and agarose coatings allowed for sufficient wetting and electrical connectivity when a chip was tested intensively over a period of 3 weeks.

A patch-clamp amplifier Axopatch200B (Molecular Devices) was used for recording the electrical current, which was acquired with a 1 kHz low-pass Bessel filter at a sampling rate of 10 kHz. The electrical recordings were post-filtered with a 400 Hz Gaussian filter for the analysis of single channels and capacitive current, and with 100 Hz filter for display only.

### Microchip fabrication

The polycarbonate chip was fabricated from 5 mm thick plates (Macrolon, Bayer) using a CNC milling machine (MSG4025, Ergwind). The milled two plates were then thermally bonded by compressing them together for 30 min at  $130^\circ\text{C}$ . 11 steel needles (~4-cm long, O.D. 0.82 mm, I.D. 0.65 mm, Fishman Corporation) served as inlets. 7 of them were connected to resistive steel capillaries (O.D. 400  $\mu\text{m}$ , I.D. 205  $\mu\text{m}$ , length 100 cm, Mifam) using segments of Tygon tubing (~2 cm, O.D. 0.91 mm, I.D. 0.25 mm, Ismatec). We used the capillaries to connect the device to the external electromagnetic valves. The use of steel capillaries of high fluidic resistance allows for precise control of the flow on the chip and the performance of iterative operations, including generation of droplets on demand. One of the capillaries was extended with a short piece of HPLC tubing (10 cm, PEEK Tubing 1/16 x 0,0025, Upchurch Scientific) in order to increase the resistance to flow. 2 out of 9 needles were connected through 10-cm long capillaries with 500  $\mu\text{l}$  syringes, each equipped with a built-in valve (1750SL, Gastight Hamilton), which were used to store and introduce onto the chip pure buffer and the protein solution. 2 of the ports, which served as outlets, were connected

with valves via 50-cm PTFE tubing (O.D. 1,6 mm, I.D. 0.8 mm, Bola).

We generated the sequence of droplets containing inhibitors off the chip (see the procedure in Supplementary Information). We then introduced these droplets onto the chip through a PTFE tubing, rather than a needle. Therefore, a connection was made directly between the channel and the tubing: 3 cm of PTFE tubing (O.D. 1.0 mm, I.D. 0.5 mm, Bola) was sealed to the chip with Hysol glue.

#### 10 Calculations of $K_d$

All measurements of  $\alpha$ HL activity were performed at +50 mV. 5 concentrations (10 – 50  $\mu$ M) of  $\gamma$ CD were used. We discarded the results of  $\alpha$ HL blocking by 1  $\mu$ M  $\gamma$ CD since the number of blockades was very low. The protein was contained in a droplet positioned at working electrode. The buffer was 1M KCl, 10 mM Tris.HCl, pH 7.0. For more detailed explanation see the Supplementary Information.

#### Conclusions

The system presented here provides numerous opportunities for further development. The use of in vitro transcription and translation (IVTT) for protein expression directly in the droplets may allow to avoid the steps of purification, freezing and dilution.<sup>10</sup> Alternatively the addition of a microfluidic module for the dilution<sup>48</sup> of protein stock on chip will allow for more efficient screening protocols. The use of on-demand droplet screening techniques<sup>49,52</sup> may provide for fast screening of the composition of buffers and their effects on the function of protein pores. We hope that the technique reported here is a step towards the construction of systems for high-throughput in vitro tests of activity of membrane proteins, in which sequential screens might be combined with parallelization.<sup>19,58,59</sup>

#### Notes and references

<sup>a</sup> Institute of Physical Chemistry, Polish Academy of Sciences Kasprzaka 44/52, 01-224 Warsaw, Poland

<sup>35</sup> E-mail: garst@ichf.edu.pl

<sup>b</sup> Chemical Biology Sub-Department, University of Oxford, Chemistry Research Laboratory

12 Mansfield Road, Oxford OX1 3TA, United Kingdom

E-mail: hagan.bayley@chem.ox.ac.uk

<sup>40</sup> † Electronic Supplementary Information (ESI) available: description microchip design and operation, screening experiments, analysis of bilayer size and videos. See DOI: 10.1039/b000000x/

#### Acknowledgements

K.T. Sapra present address: Biochemisches Institut, Universität Zürich, Winterthurerstrasse 190, 8057, Zürich

M.A. Czekalska and T.S. Kaminski contributed equally to this work.

The authors declare no competing financial interests.

We thank S. Mantri, L. Höfler, G. Villar, T. Mach, M. Laskowski and L. Kozon for helpful discussions. The  $\alpha$ HL was made by E. Mikhailova.

<sup>50</sup> Project co-financed by the European Research Council Starting Grant 279647 (to P.G.) and European Regional Development Fund under the Operational Programme Innovative Economy - within the Foundation for Polish Science grant VENTURES/2012-10/4 (to T.S.K.) and START fellowship (to T.S.K.). S.J. acknowledges financial support from the Polish Ministry of Science under the grant Iuventus Plus - IP2012 015172. K.T.S. was supported by the Marie-Curie Intra European fellowship.

#### References

1. J. Dunlop, M. Bowlby, R. Peri, D. Vasilyev, and R. Arias, *Nat Rev Drug Discov*, 2008, **7**, 358–368.
2. T. J. Dale, C. Townsend, E. C. Hollands, and D. J. Trezise, *Mol. Biosyst.*, 2007, **3**, 714–722.
3. M. Montal and P. Mueller, *Proc. Natl. Acad. Sci.*, 1972, **69**, 3561–3566.
4. H. Bayley, B. Cronin, A. Heron, M. A. Holden, W. L. Hwang, R. Syeda, J. Thompson, and M. Wallace, *Mol. Biosyst.*, 2008, **4**, 1191–1208.
5. L. K. Tamm and H. M. McConnell, *Biophys. J.*, 1985, **47**, 105–113.
6. E. T. Castellana and P. S. Cremer, *Surf. Sci. Rep.*, 2006, **61**, 429–444.
7. L. Tsofina, E. Liberman, and A. Babakov, *Nature*, 1966, **212**, 681–&.
8. K. Funakoshi, H. Suzuki, and S. Takeuchi, *Anal. Chem.*, 2006, **78**, 8169–8174.
9. M. A. Holden, D. Needham, and H. Bayley, *J. Am. Chem. Soc.*, 2007, **129**, 8650–8655.
10. R. Syeda, M. A. Holden, W. L. Hwang, and H. Bayley, *J. Am. Chem. Soc.*, 2008, **130**, 15543–15548.
11. Y. Schaerli and F. Hollfelder, *Mol. Biosyst.*, 2009, **5**, 1392–1404.
12. A. B. Theberge, F. Courtois, Y. Schaerli, M. Fischlechner, C. Abell, F. Hollfelder, and W. T. S. Huck, *Angew. Chem. Int. Ed.*, 2010, **49**, 5846–5868.
13. L. Li, D. Mustafi, Q. Fu, V. Tereshko, D. L. Chen, J. D. Tice, and R. F. Ismagilov, *Proc. Natl. Acad. Sci.*, 2006, **103**, 19243–19248.
14. E. Brouzes, M. Medkova, N. Savenelli, D. Marran, M. Twardowski, J. B. Hutchison, J. M. Rothberg, D. R. Link, N. Perrimon, and M. L. Samuels, *Proc. Natl. Acad. Sci.*, 2009, **106**, 14195–14200.
15. J. J. Agresti, E. Antipov, A. R. Abate, K. Ahn, A. C. Rowat, J.-C. Baret, M. Marquez, A. M. Klibanov, A. D. Griffiths, and D. A. Weitz, *Proc. Natl. Acad. Sci.*, 2010, **107**, 4004–4009.
16. B. E. Debs, R. Utharala, I. V. Balyasnikova, A. D. Griffiths, and C. A. Merten, *Proc. Natl. Acad. Sci.*, 2012.
17. J. E. Kreutz, L. Li, L. S. Roach, T. Hatakeyama, and R. F. Ismagilov, *J. Am. Chem. Soc.*, 2009, **131**, 6042–6043.
18. S. Aghdaei, M. E. Sandison, M. Zagnoni, N. G. Green, and H. Morgan, *Lab. Chip*, 2008, **8**, 1617–1620.
19. R. Kawano, Y. Tsuji, K. Sato, T. Osaki, K. Kamiya, M. Hirano, T. Ide, N. Miki, and S. Takeuchi, *Sci Rep*, 2013, **3**.
20. A. M. El-Arabi, C. S. Salazar, and J. J. Schmidt, *Lab. Chip*, 2012, **12**, 2409–2413.
21. S. A. Acharya, A. Portman, C. S. Salazar, and J. J. Schmidt, *Sci Rep*, 2013, **3**.
22. N. Rojko, B. Cronin, J. S. H. Danial, M. A. B. Baker, G. Anderluh, and M. I. Wallace, *Biophys. J.*, 2014, **106**, 1630–1637.
23. N. Malmstadt, M. A. Nash, R. F. Purnell, and J. J. Schmidt, *Nano Lett.*, 2006, **6**, 1961–1965.
24. M. Zagnoni, M. E. Sandison, P. Marius, A. G. Lee, and H. Morgan, *Lab. Chip*, 2007, **7**, 1176–1183.
25. M. Zagnoni, M. E. Sandison, and H. Morgan, *Biosens. Bioelectron.*, 2009, **24**, 1235–1240.
26. J. L. Poulos, W. C. Nelson, T.-J. Jeon, C.-J. ‘CJ’ Kim, and J. J. Schmidt, *Appl. Phys. Lett.*, 2009, **95**, -.
27. M. Zagnoni, M. E. Sandison, P. Marius, and H. Morgan, *Anal. Bioanal. Chem.*, 2009, **393**, 1601–1605.
28. S. A. Sarles and D. J. Leo, *Lab. Chip*, 2010, **10**, 710–717.

29. J L Poulos and S A Portonovo and H Bang and J J Schmidt, *J. Phys. Condens. Matter*, 2010, **22**, 454105.
30. R. Kawano, T. Osaki, H. Sasaki, and S. Takeuchi, *Small*, 2010, **6**, 2100–2104.
- 5 31. M. Takinoue, H. Onoe, and S. Takeuchi, *Small*, 2010, **6**, 2374–2377.
32. C. E. Stanley, K. S. Elvira, X. Z. Niu, A. D. Gee, O. Ces, J. B. Edel, and A. J. deMello, *Chem. Commun.*, 2010, **46**, 1620–1622.
- 10 33. T. Thapliyal, J. L. Poulos, and J. J. Schmidt, *Biosens. Bioelectron.*, 2011, **26**, 2651–2654.
34. G. Baaken, N. Ankri, A.-K. Schuler, J. R u he, and J. C. Behrends, *ACS Nano*, 2011, **5**, 8080–8088.
35. Y. Elani, A. J. deMello, X. Niu, and O. Ces, *Lab. Chip*, 15 2012, **12**, 3514–3520.
36. V. C. Stimberg, J. G. Bomer, I. van Uitert, A. van den Berg, and S. Le Gac, *Small*, 2013, **9**, 1076–1085.
37. M. Lein, J. Huang, and M. A. Holden, *Lab. Chip*, 2013, **13**, 2749–2753.
- 20 38. Y. Tsuji, R. Kawano, T. Osaki, K. Kamiya, N. Miki, and S. Takeuchi, *Anal. Chem.*, 2013, **85**, 10913–10919.
39. Y. Tsuji, R. Kawano, T. Osaki, K. Kamiya, N. Miki, and S. Takeuchi, *Lab. Chip*, 2013, **13**, 1476–1481.
40. T. Tonooka, K. Sato, T. Osaki, R. Kawano, and S. 25 Takeuchi, *Small*, 2014, n/a–n/a.
41. M. Urban, A. Kleefen, N. Mukherjee, P. Seelheim, B. Windschiegl, M. Vor der Br uggen, A. Ko cer, and R. Tamp e, *Nano Lett.*, 2014, **14**, 1674–1680.
42. I. Kusters, A. M. van Oijen, and A. J. M. Driessen, *ACS Nano*, 2014, **8**, 3380–3392.
- 30 43. O. K. Castell, J. Berridge, and M. I. Wallace, *Angew. Chem.-Int. Ed.*, 2012, **51**, 3134–3138.
44. M. E. Dolega, S. Jakiela, M. Razew, A. Rakszewska, O. Cybulski, and P. Garstecki, *Lab. Chip*, 2012, **12**, 4022– 35 4025.
45. S. Jakiela, T. S. Kaminski, O. Cybulski, D. B. Weibel, and P. Garstecki, *Angew. Chem. Int. Ed.*, 2013, **52**, 8908–8911.
46. Y. Bai, X. He, D. Liu, S. N. Patil, D. Bratton, A. Huebner, F. Hollfelder, C. Abell, and W. T. S. Huck, *Lab. Chip*, 2010, **10**, 1281–1285.
- 40 47. M. Zagnoni and J. M. Cooper, *Lab. Chip*, 2010, **10**, 3069–3073.
48. P. M. Korczyk, L. Derzsi, S. Jakiela, and P. Garstecki, *Lab. Chip*, 2013.
- 45 49. K. Churski, P. Korczyk, and P. Garstecki, *Lab. Chip*, 2010, **10**, 816–818.
50. P. H. King, G. Jones, H. Morgan, M. R. R. de Planque, and K.-P. Zauner, *Lab. Chip*, 2014, **14**, 722–729.
51. S. Thutupalli, J.-B. Fleury, A. Steinberger, S. Herminghaus, and R. Seemann, *Chem. Commun.*, 2013, **49**, 1443–1445.
- 50 52. K. Churski, T. S. Kaminski, S. Jakiela, W. Kamysz, W. Baranska-Rybak, D. B. Weibel, and P. Garstecki, *Lab. Chip*, 2012, **12**, 1629–1637.
53. L. C. M. Gross, A. J. Heron, S. C. Baca, and M. I. Wallace, *Langmuir*, 2011, **27**, 14335–14342.
- 55 54. T. S. Kaminski, S. Jakiela, M. A. Czekalska, W. Postek, and P. Garstecki, *Lab. Chip*, 2012, **12**, 3995–4002.
55. S. Mantri, K. T. Sapra, S. Cheley, T. H. Sharp, and H. Bayley, *Nat. Commun.*, 2013, **4**.
- 60 56. L. C. M. Gross, O. K. Castell, and M. I. Wallace, *Nano Lett.*, 2011, **11**, 3324–3328.
57. S. Cheley, L.-Q. Gu, and H. Bayley, *Chem. Biol.*, 2002, **9**, 829–838.
58. B. Lu, G. Kocharyan, and J. J. Schmidt, *Biotechnol. J.*, 65 2014, **9**, 446–451.
59. B. Le Pioufle, H. Suzuki, K. V. Tabata, H. Noji, and S. Takeuchi, *Anal. Chem.*, 2007, **80**, 328–332.