JAAS

Accepted Manuscript



This is an *Accepted Manuscript*, which has been through the Royal Society of Chemistry peer review process and has been accepted for publication.

Accepted Manuscripts are published online shortly after acceptance, before technical editing, formatting and proof reading. Using this free service, authors can make their results available to the community, in citable form, before we publish the edited article. We will replace this Accepted Manuscript with the edited and formatted Advance Article as soon as it is available.

You can find more information about *Accepted Manuscripts* in the **Information for Authors**.

Please note that technical editing may introduce minor changes to the text and/or graphics, which may alter content. The journal's standard <u>Terms & Conditions</u> and the <u>Ethical guidelines</u> still apply. In no event shall the Royal Society of Chemistry be held responsible for any errors or omissions in this *Accepted Manuscript* or any consequences arising from the use of any information it contains.



www.rsc.org/jaas

1		
2 3		
4 5 6	1	
7 8 9	2	
10 11 12 13	3	Molybdenum Isotopic Analysis with Negative
14 15 16	4	Thermal Ionization Mass Spectrometry (N-TIMS):
17 18 19	5	Effects on Oxygen Isotopic Composition
20 21 22	6	
23 24 25	7	
26 27	8	Yuichiro NAGAI ^{*1} and Tetsuya YOKOYAMA ¹
29 30	9	
31 32	10	1: Department of Earth and Planetary Sciences, Tokyo Institute of Technology, Ookayama,
33	11	Tokyo 152-8551, Japan
34 35	12	
36 37	13	* Corresponding author: (yuichiro.nagai@geo.titech.ac.jp)
38	14	September 8th, 2015
40	15	Revised version: January 30th, 2016
41 42	16	
43 44	17	Submitted to Journal of Analytical Atomic Spectrometry
45 46	18	
40	19	6924 words
48 49	20	
50 51		
52		
53 54		
55 56		
57		
58 59		
60		

21 Abstract

We developed a new, highly precise, and accurate Mo isotope analysis with thermal ionization mass spectrometry in negative ionization mode (N-TIMS). We discovered that the optimal condition to ionize Mo most efficiently into MoO₃⁻ was to load the Mo sample on a Re filament and cover the sample with $La(NO_3)_3$, thus yielding La/Mo = 5. To achieve highly precise Mo isotope analysis, determining the oxygen isotopic composition of MoO₃⁻ ion in each measurement by monitoring masses 149 (¹⁰⁰Mo¹⁶O₂¹⁷O⁻) and 150 $(^{100}Mo^{16}O_2^{-18}O_2^{$ important. After correcting the O isotopic interferences and performing mass-dependent fractionation during the TIMS measurement, the acquired Mo isotopic ratios yielded the following reproducibilities (2 SD): 47, 16, 10, 13, and 33 ppm for ⁹²Mo/⁹⁶Mo, ⁹⁴Mo/⁹⁶Mo, ⁹⁵Mo/⁹⁶Mo, ⁹⁷Mo/⁹⁶Mo, and ¹⁰⁰Mo/⁹⁶Mo, respectively. The reproducibilities have been improved by 1.3-2.7 times compared to those obtained in previous studies using multi-collector inductively coupled plasma mass spectrometry. The accuracy of our technique was confirmed by measuring two synthesized solutions with enriched ⁹²Mo, ⁹⁷Mo, and ¹⁰⁰Mo abundances and two iron meteorites, i.e., Henbury (IIIAB) and Albion (IVA). Moreover, we determined positive Mo isotope anomalies for a new iron meteorite, Tambo Quemad (IIIAB). Our N-TIMS technique can be applied to the studies of nucleosynthetic isotope anomalies in extraterrestrial materials as well as mass-dependent Mo isotopic shift in environmental samples.

1. Introduction

43	Considerable improvements in mass spectrometry techniques over the last 15 years
44	have enabled high precision measurements of nontraditional stable isotopes, including a
45	unique transition element such as molybdenum (Mo). It has seven stable isotopes, ⁹² Mo,
46	⁹⁴ Mo, ⁹⁵ Mo, ⁹⁶ Mo, ⁹⁷ Mo, ⁹⁸ Mo, and ¹⁰⁰ Mo with averaged abundances of 14.8%, 9.23%,
47	15.9%, 16.7%, 9.56%, 24.2%, and 9.67% ¹ , respectively ¹ . The Mo isotope system is useful
48	in geochemistry, biogeosciences, and environmental sciences because of the potential
49	isotopic fractionation during redox reactions. For example, mass-dependent Mo isotope
50	fractionation at a level of a few permil was detected with multi-collector inductively
51	coupled plasma mass spectrometry (MC-ICP-MS) in various types of marine sediments,
52	including Mn-oxide-enriched sediments in oxic conditions ^{2,3} , anoxic sediments ^{4,5} , and
53	euxinic sediments ^{2,6-8} . The individual Mo isotopic behaviors of black shales and banded
54	iron formations under each of the redox conditions are studied to reveal a complicated
55	oxidation history for the surface environment in the Archean and Paleoproterozoic ^{9,10} .
56	Moreover, the isotopic fractionation of Mo was caused by N ₂ -fixing organisms because Mo
57	plays an important role as a metal cofactor in enzymes ¹¹ . This type of isotopic fractionation
58	was reproduced by an experiment for Mo assimilation during nitrate reduction and
59	atmospheric N_2 fixation in fresh water cyanobacterium cultures ¹² .

Journal of Analytical Atomic Spectrometry Accepted Manuscript

In contrast, nucleosynthetic isotope anomalies in extraterrestrial materials (i.e.,

mass-independent isotopic shifts from the terrestrial value) have been another intriguing application for Mo isotopes. Molybdenum isotopes in the universe were synthesized via stellar nucleosynthesis of the s-process (trace ⁹⁴Mo, ⁹⁵Mo, ⁹⁶Mo, ⁹⁷Mo, and ⁹⁸Mo), the *r*-process (95 Mo, 97 Mo, 98 Mo, and 100 Mo), and the *p*-process (92 Mo and 94 Mo)¹³. Multiple studies utilizing MC-ICP-MS have discovered nucleosynthetic Mo isotope anomalies in various types of meteorites^{14,15}, which suggest the existence of isotope heterogeneity for Mo in the early Solar System. Compared to Mo isotope anomalies for carbonaceous chondrites and iron meteorites, those for other non-carbonaceous meteorites such as Cr¹⁶ were undetected because of the small degree of their Mo isotope deviations from terrestrial materials. Therefore, highly precise Mo isotope analysis is required to discriminate these marginal anomalies even if large amounts of samples are used.

Although Mo isotopic analyses have generally been performed with MC-ICP-MS, thermal ionization mass spectrometry (TIMS) is an alternative choice for enabling high precision Mo isotopic compositions in terrestrial and extraterrestrial materials. One of the advantages of TIMS is that the kinetic energy distribution of the thermally ionized beam is much smaller (~0.5 eV) than that of the ion beam generated via plasma ionization (~5 eV), minimizing the fluctuation of the ion beam intensity compared to MC-ICP-MS. This enables high-precision isotope ratio measurements for some elements at levels ≤ 10 ppm using new generation TIMS instruments (e.g., Nd^{17} and W^{18}).

In the 1960s, Mo isotope measurements were performed by TIMS with a Mo⁺ ion

81	beam, resulting in an analytical precision of $\pm 0.6\%$ for each Mo ratio (ⁱ Mo/ ¹⁰⁰ Mo). ¹⁹ Such a
82	large analytical uncertainty was caused by poor Mo ionization efficiency owing to its
83	higher first ionization potential (7.1 eV) compared to elements such as Sr^{20} (5.7 eV) and
84	Nd^{17} (5.5 eV) whose isotopic compositions are analyzed with high precision by TIMS. To
85	avoid this problem, TIMS has been devised and successfully applied to a variety of
86	elements including not only Mo but also those with higher first ionization potentials (e.g.,
87	W^{21} and Os^{22}). The N-TIMS technique has been applied to Mo isotope analysis since the
88	early 2000s, resulting in the analytical precision of >44 ppm for ${}^{95}Mo/{}^{98}Mo^{23}$; however,
89	these measurements were performed with previous generation TIMS instruments, which
90	have an inferior analytical precision compared to current techniques such as
91	MC-ICP-MS. ^{15,24} To date, Mo isotope analysis has rarely been performed using the latest
92	generation TIMS instrument, nor has it been compared to the results of MC-ICP-MS
93	measurements. As demonstrated by W isotope analysis ¹⁸ , higher precision can be achieved
94	with TIMS when stronger beam intensities are generated by loading the target element on a
95	filament much larger than that which is normally applied.

In this study, we developed a method to enable highly precise Mo isotope analysis utilizing the latest generation TIMS instrument in negative ion mode. To improve the accuracy and precision of isotope analysis, we specifically focused on the variability of O isotopic compositions of Mo trioxide ion (MoO_3^-) during the N-TIMS measurements. Moreover, we present Mo isotope compositions in an extraterrestrial sample (iron

ว
Ζ
3
4
5
5
6
7
8
0
9
10
11
12
12
13
14
15
16
10
17
18
10
00
20
21
22
22
23
24
25
26
20
27
28
20
23
30
31
32
202
33
34
35
36
50
37
38
30
40
40
41
42
13
43
44
45
46
17
47
48
49
50
50
51
52
53
53
53 54
53 54 55
53 54 55 56
53 54 55 56 57

58 59 60

1

101 meteorite) for comparison with recent works conducted with MC-ICP-MS.

102	
103	2. Experimental
104	2.1. Reagents and activators
105	Deionized water (18.2 M Ω cm ⁻¹) was prepared using a Milli-Q Integral 5
106	(Merck-Millipore Corp.) water purification system. Electronics industry (EL) grade 12 M
107	HCl, 16 M HNO ₃ (Mitsubishi Chem.) and 30 M HF of atomic absorption spectrometry
108	(AAS) grade (Kanto Chem. Co. Ltd.) were distilled once using a two-bottle Teflon
109	distillation system (referred to as 1D HCl, 1D HNO ₃ , and 1D HF, respectively).
110	Ca, La, and Cd standard solutions prepared in aqueous nitrate media for atomic
111	absorption spectrometry (1000 μ g/g; Kanto Chem. Co. Ltd.) were used to make activators,
112	which helped enhance the ionization efficiency of MoO ₃ ⁻ ions in the TIMS. In this study, we
113	tested the following four activators: i) 1000 μ g/g Ca(NO ₃) ₂ in 0.4 M HNO ₃ , ii) 1000, 5000,
114	10000, 15000, 20000, and 30000 μ g/g La(NO ₃) ₃ in 1 M HNO ₃ , iii) 5000 μ g/g La(NO ₃) ₃ +
115	5000 μ g/g Gd(NO ₃) ₃ in 1 M HNO ₃ , and iv) 5000 μ g/g La(NO ₃) ₃ + 5000 μ g/g Gd(NO ₃) ₃ in
116	1 M HNO ₃ , coupled with 1000 μ g/g Ca(NO ₃) ₂ in 0.4 M HNO ₃ .
117	

2.2. Mo isotope standard and enriched isotopic solutions

119 A Mo standard solution for atomic absorption spectrometry (1000 μ g/g in 1.3 M 120 HNO₃-0.1 M HCl; Kanto Chemical Co., Inc.) was used as a running standard for Mo 121 isotope analysis with N-TIMS. This solution, referred to as "Kanto-Mo," is considered to have a representative terrestrial Mo isotopic composition, excluding the effect ofmass-dependent fractionation.

To evaluate the accuracy of Mo isotope analysis using N-TIMS, two artificial samples with nonterrestrial Mo isotopic compositions (Mix-A and Mix-B) were prepared using three spike solutions that were enriched with ⁹²Mo, ⁹⁷Mo, and ¹⁰⁰Mo, respectively. The ⁹²Mo spike solution was prepared by dissolving 10.4 mg of ⁹²Mo-enriched (99.93%) metal powder, which was obtained from Isoflex USA Inc. The metal powder was first dissolved in 30 M HF and completely dried, and then re-dissolved in 16 M HNO₃ and heated to dryness. The residual was finally dissolved in 1 M HF-1 M HNO₃. The Mo concentration of the ⁹²Mo-enriched solution was determined by isotope dilution using a reference isotopic composition for ⁹²Mo-enriched metal, reported by Isoflex, and that for Kanto-Mo defined by IUPAC, resulting in a Mo concentration of 933 ng/g. Similarly, the ⁹⁷Mo and ¹⁰⁰Mo spike solutions were prepared from 9.14 and 8.04 mg of ⁹⁷Mo-enriched (94.19%) and ¹⁰⁰Mo-enriched (98.59%) metal powders, which were obtained from the Oak Ridge National Laboratory, and resulted in Mo concentrations of 656 and 582 ng/g, respectively. The isotopically enriched samples, Mix-A and Mix-B, were prepared by gravitationally mixing the Kanto-Mo and a mixture of three spike solutions (${}^{92}Mo$; ${}^{97}Mo$; ${}^{100}Mo = 1:0.4:0.5$) in different proportions. The Mix-B sample contained approximately twice the amount of the spike mixture than that of the Mix-A sample. Note that the true isotopic compositions of Mix-A and Mix-B are difficult to determine by direct measurement with TIMS because the

effect of mass fractionation cannot be corrected without a known isotope ratio as a reference (e.g., 98 Mo/ 96 Mo). Therefore, we estimated the isotopic compositions of Mix-A and Mix-B by the gravimetric calculations using the isotopic compositions of individual Mo spikes reported by Isoflex and Oak Ridge National Laboratory. Because the individual spike compositions reported are noncertified reference values, the accuracy of the estimated Mo isotope compositions in Mix-A and Mix-B are not as precise as ppm (10⁻⁶) levels but are at percent (10⁻²) levels.

2.3. Samples

The Mo isotopic composition of three terrestrial rocks, JB-3 (basalt, Mt. Fuji), JA-3 (andesite, Mt. Asama), and JR-2 (rhyolite, Wada Toge obsidian), as well as three iron meteorites, Tambo Quemado (IIIAB), Henbury (IIIAB), and Albion (IVA), were determined. The terrestrial rocks (~2 g) were dissolved in 90 mL Savillex Teflon beakers using HF-HNO₃ (2:1) at 120°C for a day, 150°C for 2 days, and 180°C for several days, followed by drying at 120°C. The Tambo Quemado sample was digested in a 1:2 mixture of concentrated HCl and HNO₃. From the sample solution, W was separated by a two-stage Hf-W chemical separation technique²⁵ and used in another project²⁶. All the residual fractions excluding W were combined and dried at 120°C, which was dissolved in 0.5 M HF for 2 h in an ultrasonic bath. After centrifuging the sample solution, the supernatant was dried and dissolved in 1 mL 0.4 M HCl-0.5 M HF. The other iron meteorites excluding IIE

Journal of Analytical Atomic Spectrometry Accepted Manuscript

Spectrometry Accepted Manuscript

Journal of Analytical Atomic

irons (~2g) were leached in 6 M HCl for 30 min at 120°C to obtain more purified fractions. The residual fractions were digested in 16 M HNO₃-12 M HCl at 120°C overnight. In contrast, IIE irons were dissolved in 16 M HNO₃-12 M HCl at 120°C overnight, with additional digestion in 16 M HNO₃-30 M HF to dissolve the silicates. Mo was further purified by two-step anion exchange column chromatography using an Eichrom AG1X8 (200–400 mesh) with HF-HCl and HF-HNO₃ mixtures²⁷. In the first step, the resin was charged in a polyethylene column, Muromac mini-column S-size (5 $mm\phi$, Muromachi Technos co.); cleaned sequentially with 6 M HCl, purified H₂O, 6 M HNO₃-3M HF, and purified H₂O; and finally conditioned using 0.4 M HCl-0.5M HF (2 mL). The sample solution (1 mL) was loaded on the column. The major elements (e.g., Fe), Ti–Zr–Hf and W, which could not be removed in the previous procedure, were successively removed by adding 0.4 M HCl-0.5 M HF (3 mL), 9 M HCl-0.05 M HF (5 mL), and 9 M HCl-1 M HF (10 mL), respectively. Subsequently, Mo was eluted with 6 M HNO₃-3 M HF (5 mL). The Mo fraction was dried and dissolved in 2 M HF (0.3 mL). In the second step, the Mo fraction was purified using a polypropylene column filled with 0.1 mL Eichrom 1X8. The resin was cleaned sequentially with 6 M HCl, purified H₂O, 6M HNO₃-3M HF, and purified H₂O. The resin bed was then conditioned using 0.4 M HCl-0.5 M HF. The Mo fraction was loaded on the cleaned resin bed, followed by the addition of 2 M HF (0.4 mL). Nb was subsequently eluted using 6 M HCl-0.1 M HF (2 mL), and Mo was finally obtained by adding 6 M HNO₃-3 M HF (1 mL).

The blank of this digestion and the chemical procedure was ~ 1 ng.
2.4. Mass Spectrometry
Molybdenum isotopic analysis was performed with N-TIMS using TRITON plus
(Thermo Fisher Scientific Inc.) installed at the Tokyo Institute of Technology (Tokyo Tech).
This instrument is equipped with nine Faraday cups with $10^{11} \Omega$ amplifiers. A secondary
electron multiplier (SEM) is placed behind the center Faraday cup.
2.4.1. Sample Loading
A zone-refined 99.999% Re filament (thickness = 0.0305 mm, width = 0.750 mm; H.
Cross) was used in this study. Prior to sample loading, the Re filament was outgassed for 30
min at 4.5 A in a vacuum degassing system. The Mo sample (ca. 1000-4000 ng) was
dissolved in 1 μ L of 1.3 M HNO ₃ –0.1 M HCl and loaded on the filament. The sample
solution was dried at 0.6 A and subsequently covered by 1 μ L of the activator. As described
above, we tested the following four activators, which were used in previous N-TIMS
studies ^{18,23,28-29} : i) Ca(NO ₃) ₂ , ii) La(NO ₃) ₃ , iii) (La, Gd)(NO ₃) ₃ , and iv) Ca(NO ₃) ₂ + (La,
Gd)(NO ₃) ₃ . The optimal choice of the activator will be discussed in Section 3.1.
2.4.2. Measurement of Mo isotope ratios
The sample filament was placed in an ion source under vacuum at $<1.5 \times 10^{-7}$ mbar

without oxygen bleeding. The sample filament was first heated at 250 mA/min until 1075°C, and then heated at 20 mA/min until 1150°C. After focusing the MoO₃⁻ion beam, the filament continued to heat at 20 mA/min. The MoO₃⁻ion beam intensity decreased slightly around 1160–1180°C, but started to increase above 1230–1280°C. The acquisition of Mo isotopes was started when the ⁹⁴MoO₃⁻ion beam intensity exceeded 2.0 V (2.0 × 10^{-11} A).

In this study, we examined the static multicollection method. The cup configuration of the N-TIMS measurement is summarized in Table 1. Each measurement consists of 360 cycles (20 cycles/block \times 18 blocks) of data acquisition with 16.777 s integration time. The amplifier gains were calibrated once at the start of each day. The baselines of amplifiers were measured for 30 s at the beginning of each block by deflecting the beam away from the detectors. The amplifier rotation system was applied to reduce the error due to amplifier gain calibration, in which an array of nine $10^{11} \Omega$ amplifiers connected behind the nine Faraday cups was electrically rotated in each block using a relay matrix. The typical total analytical time including the sample heating was 3-4 h.

2.5. Data processing

In the N-TIMS measurement, Mo isotopic compositions were obtained by collecting Mo trioxide ions (MoO_3^-) using nine Faraday cups (Table 1). The intensities of the $MoO_3^$ ions collected by individual detectors do not reflect the true Mo isotopic composition primarily because of the following two reasons: mass fractionation during the analysis (e.g., evaporation from filament) and oxygen isotopic interferences for ${}^{i}Mo^{16}O_{3}$ -occurring as isobaric molecular ions such as ${}^{i-1}Mo^{16}O_2{}^{17}O^-$ and ${}^{i-2}Mo^{16}O_2{}^{18}O^-$. The latter issue is problematic because these isobars have very low abundances (0.2% and 0.04% for ⁱ⁻¹Mo¹⁶O₂¹⁷O⁻and ⁱ⁻²Mo¹⁶O₂¹⁸O⁻relative to ⁱMo¹⁶O₃⁻) and are difficult to precisely determine; however, they are not negligible for achieving highly precise Mo isotope analysis. In some previous studies for Os $(OsO_3^{-})^{30}$ and Nd $(NdO^{+})^{31}$ isotope analysis, the correction of O isotopic interferences was conducted by assuming a uniform O isotopic composition of Nier³²: ${}^{17}O/{}^{16}O = 0.000375$ and ${}^{18}O/{}^{16}O = 0.002044$. Conversely, Luguet et al.³³ determined the in-run O isotope composition for each Os isotope measurement with N-TIMS by sampling ${}^{192}\text{Os}{}^{16}\text{O}{}_{2}{}^{17}\text{O}^{-}$ and ${}^{192}\text{Os}{}^{16}\text{O}{}_{2}{}^{18}\text{O}^{-}$ ions. The authors concluded that (1) the ${}^{18}O/{}^{16}O$ ratio can vary across multiple Os isotope measurements, and (2) the Nier's ¹⁷O/¹⁶O and ¹⁸O/¹⁶O ratios obviously deviated from the mass-dependent fractionation trend they obtained.

Journal of Analytical Atomic Spectrometry Accepted Manuscript

In this study, we conducted both correction methods for Mo isotope analysis, which are hereafter referred to as the "Nier's O" correction and the "In-run O" correction. In addition, we examined the "Mean O" correction method in which a uniform O isotopic composition determined by averaging multiple "In-run O" compositions was applied for correcting O isobaric interferences. The O isotope composition of MoO_3^- ions during each isotope measurement was determined by monitoring masses 148 ($^{100}Mo^{16}O_3^-$), 149

242 (
$$^{100}Mo^{16}O_2^{17}O^{-}$$
), and 150 ($^{100}Mo^{16}O_2^{18}O^{-}$), and conducting the iterative calculations as
243 described below. First, apparent isotopic ratios of $^{17}O/^{16}O(R_1)$ and $^{18}O/^{16}O(R_2)$ were
244 determined:
245
246 $R_1 = V_{149}/V_{148}$ (1)
247 $R_2 = V_{149}/V_{148}$ (2)
248
249 where V_i represents the total ion beam intensity for the molecular mass of *i*. Next, the
250 intensities of $^{1}Mo^{16}O_3^{-}$ ions (= I_j) were determined as follows.
251
252 $I_{92} = V_{140}$ (3)
253 $I_{94} = V_{142} - I_{92} \times (3R_1^2 + 3R_2)$ (4)
254 $I_{95} = V_{143} - I_{92} \times (R_1^3 + 6R_1R_2) - I_{94} \times 3R_1$ (5)
255 $I_{96} = V_{145} - I_{92} \times (3R_1^2R_2 + 3R_2^2) - I_{94} \times (3R_1^2 + 3R_2) - I_{95} \times 3R_1$ (6)
256 $I_{97} = V_{145} - I_{92} \times 3R_1R_2^2 - I_{94} \times (R_1^3 + 6R_1R_3) - I_{95} \times (3R_1^2 + 3R_2) - I_{96} \times (3R_1^2 + 3R_2) - I_{97}$
258 $\times 3R_1$ (8)
259 $I_{100} = V_{148} - I_{94} \times R_2^3 - I_{95} \times 3R_1R_2^2 - I_{96} \times (3R_1^2R_2 + 3R_2^2) - I_{97} \times (R_1^3 + 6R_1R_2) - I_{98} \times 260$ ($3R_1^2 + 3R_2$) (9).
261

4		
5 6 7	262	Similarly, the ion beam intensities of ${}^{100}Mo{}^{16}O{}_{2}{}^{17}O{}^{-} (=I{}_{100}{}^{17})$ and ${}^{100}Mo{}^{16}O{}_{2}{}^{18}O{}^{-} (=I{}_{100}{}^{18})$
7 8 9	263	were determined as follows:
10 11	264	
12 13 14	265	$I_{100}^{17} = V_{149} - I_{95} \times R_2^{3} - I_{96} \times 3 R_1 R_2^{2} - I_{97} \times (3R_1^{2}R_2 + 3R_2^{2}) - I_{98} \times (R_1^{3} + 6R_1R_2) - I_{100} \times (R_1^{2}R_2 + 3R_2^{2}) - I_{98} \times (R_1^{3} + 6R_1R_2) - I_{100} \times (R_1^{2}R_2 + 3R_2^{2}) - I_{98} \times (R_1^{3} + 6R_1R_2) - I_{100} \times (R_1^{2}R_2 + 3R_2^{2}) - I_{98} \times (R_1^{3} + 6R_1R_2) - I_{100} \times (R_1^{2}R_2 + 3R_2^{2}) - I_{98} \times (R_1^{3} + 6R_1R_2) - I_{100} \times (R_1^{3} + 6R_1R_2) - I_{10} \times (R_1^{3} + 6R_1R_2) - I_{10} \times $
15 16 17	266	$3R_1$ (10)
18 19	267	$I_{100}{}^{18} = V_{150} - I_{96} \times R_2{}^3 - I_{97} \times 3 R_1 R_2{}^2 - I_{98} \times (3R_1{}^2R_2 + 3R_2{}^2) - I_{100} \times (3R_1{}^2 + 3R_2) - I_{100}{}^{17}$
20 21 22	268	$\times 3R_1$ (11).
23 24 25	269	
26 27	270	Here, new oxygen isotope ratios were given by the following equations:
28 29 30	271	
31 32	272	$R_{1}' = I_{100}^{17} / I_{100} $ (12)
33 34 35	273	$R_2' = I_{100}^{18} / I_{100} $ (13).
36 37 38	274	
39 40	275	The new values R_1' and R_2' were compared with R_1 and R_2 .
41 42 43	276	
44 45 46	277	$\Delta_{17} = R_1' - R_1 \tag{14}$
40 47 48	278	$\Delta_{18} = R_2' - R_2 \tag{15}.$
49 50 51	279	
52 53	280	If either Δ_{17} or Δ_{18} was larger than 1×10^{-6} , then the calculations from equation (3)
55 56	281	to (15) were repeated by replacing R_1 and R_2 with R_1' and R_2' , respectively. In general, this
57 58 59 60		

procedure was repeated three times to achieve $<1 \times 10^{-6}$. After the iterative calculations, the final R₁' and R₂' values were regarded as true O isotopic ratios, and the Mo isotope ratios were determined from I₉₂/I₉₆, I₉₄/I₉₆, I₉₅/I₉₆, I₉₇/I₉₆, I₉₈/I₉₆, and I₁₀₀/I₉₆. After the correction of O isotopic compositions, all Mo data were normalized to ⁹⁶Mo and corrected for mass-dependent fractionation via the exponential law of MoO₃ ions by assuming ⁹⁸Mo¹⁶O₃/⁹⁶Mo¹⁶O₃ = 1.453173³⁴.

In the correction scheme for oxide interferences mentioned above, we assume that the isotopic composition of oxygen, which forms the molecular ion MoO_3^- , detected by Faraday cups is identical among different Mo isotopic species (e.g., ${}^{92}Mo^{16}O_2{}^{18}O/{}^{92}Mo^{16}O_3$ $= {}^{100}Mo^{16}O_2{}^{18}O/{}^{100}Mo^{16}O_3$). This is strictly incorrect because fractionation of Mo oxide ions occurs depending on their molecular masses until they are detected by Faraday cups. In the case of exponential law, we derive the following equation:

295
$$({}^{92}\text{Mo}{}^{16}\text{O}_{2}{}^{18}\text{O}/{}^{92}\text{Mo}{}^{16}\text{O}_{3}) \times (m_{142}/m_{140})^{\alpha} = ({}^{100}\text{Mo}{}^{16}\text{O}_{2}{}^{18}\text{O}/{}^{100}\text{Mo}{}^{16}\text{O}_{3}) \times (m_{150}/m_{148})^{\alpha},$$
 (16)
296

where *m* stands for the molecular masses of MoO₃ and α is the fractionation factor. However, in our N-TIMS measurements of MoO₃⁻ ions, the magnitude of mass fractionation was in the range of $-0.3 < \alpha < 0.3$. With this range of mass fractionation factors, $(m_{150}/m_{148})^{\alpha} / (m_{142}/m_{140})^{\alpha}$ varies from 0.9998 to 1.0002, which does not significantly affect the result of Mo isotope ratios when their analytical precisions are

302 considered.

303 After the correction scheme described above, Mo isotope ratios were obtained by 304 averaging 360 ratios with 2σ rejection (rejecting 5% of the data). The precision of Mo 305 isotopic ratios in a single isotope run is evaluated by 2 standard error (2SE) of the ratios 306 within the 95% confidence interval. The Mo isotope ratios in samples are reported in the μ 307 notation as follows:

309
$$\mu^{i} Mo = \left\{ \frac{\left({}^{i} Mo / {}^{96} Mo \right)_{sample}}{\left({}^{i} Mo / {}^{96} Mo \right)_{std}} - 1 \right\} \times 10^{6}$$
(17)

311 where *i* is the mass number and the subscript *std* represents the Kanto-Mo standard.

Spectrometry Accepted Manuscript

Journal of Analytical Atomic

3. Results and Discussion

315 3.1. Optimization of MoO₃⁻ ionization

316 In N-TIMS, negative ions are produced on the surface of a hot metal filament. The 317 ion yield β is theoretically described by the Langmuir–Saha equation for N-TIMS mode³⁵:

$$\beta^{-} = \frac{1}{1 + \frac{g_0}{g_-} exp\left(\frac{W - E_A}{\hbar T}\right)} \tag{18}$$

where $g_0/g_{\rm c}$ is the ratio of the statistical weights of the neutral species and the ion, $E_{\rm A}$ is the electron affinity of the atom or molecule to be analyzed, W is the electron work function of the filament material used, T is the temperature in Kelvin, and k is the Boltzmann constant. According to the equation, the ionization efficiency (ion yield) can be improved by reducing the work function by choosing an optimal filament material and activator.

At first, we chose an activator that strengthened the MoO_3^- ion beam of four types of activators: $Ca(NO_3)_2$, $La(NO_3)_3$, $La(NO_3)_3 + Gd(NO_3)_3$, and $La(NO_3)_3 + Gd(NO_3)_3 +$ $Ca(NO_3)_2$. The efficiencies were determined by integrating the total MoO_3^- ion beam intensities (500 ng of Mo) acquired with the initial filament temperature at 1210–1280 °C until all of the Mo on the filament has been exhausted. $La(NO_3)_3$ generated MoO_3^- ions more efficiently (at least 3.5 times) than the other activators. Although the $Gd(NO_3)_3$ activator was not particularly efficient to ionized MoO_3^- ions, it is most useful for

333 generating WO_3^{-1} ions in the N-TIMS analysis¹⁸.

To achieve the highest precision and analytical accuracy, selecting an appropriate activator/sample ratio that can sufficiently maintain the beam intensity is important. The detection efficiency of MoO_3^- drastically varied with the change in the La/Mo ratio on the filament (Figure 1). Irrespective of the amount of Mo used, the detection efficiency increased linearly at La/Mo <5 and stabilized at 5< La/Mo <7.5, although it decreased rapidly at La/Mo >7.5. Based on these results, we decided to use the La(NO₃)₃ solution as the activator, which achieved a ratio of La/Mo = 5 in the following experiments.

During the N-TIMS analysis, bleeding of O_2 gas increases the beam intensity for Os^{30,36}, Ru,³⁷ and W¹⁸. However, in the case of Mo, the more the partial pressure of O_2 in the sample chamber was increased (e.g., $> 5 \times 10^{-7}$ mbar), the more the MoO₃⁻ion beam intensities deteriorated. Therefore, we did not apply O_2 gas bleeding for the analysis of Mo isotopes with N-TIMS. Journal of Analytical Atomic Spectrometry Accepted Manuscript

3.2. Variation of oxygen isotopic compositions

The O isotopic interferences on MoO_3^- were traditionally corrected using a constant O isotopic composition such as the Nier's composition²³ or the IUPAC "best" composition²⁸. However, these studies did not report the O isotopic composition actually observed during the individual Mo isotope measurement. As shown in Figure 2 and Table 2, our "In-run" O isotope composition (circles) varied slightly across different sample

measurements: $(3.874 \pm 0.002) - (3.885 \pm 0.002) \times 10^{-4}$ for ${}^{17}\text{O}/{}^{16}\text{O}$ and (2.0583 ± 0.0003) $-(2.0685 \pm 0.0003) \times 10^{-3}$ for ¹⁸O/¹⁶O (errors are 2SE of the individual measurement). The average of 21 "In-run" O isotopic compositions obtained during the period of this study resulted in the "Mean" O isotopic composition (double circles) of which ${}^{17}O/{}^{16}O = (3.880 \pm$ 0.006) × 10⁻⁴ and ¹⁸O/¹⁶O = (2.064 ± 0.005) × 10⁻³ (errors are 2SD). A much larger variation of the O isotopic composition was observed in some Os isotope studies in which oxygen was bled in the sample chamber while maintaining a constant gas pressure^{33,38}. Oxygen isotopic variations were also observed in the study of W isotope analysis in which W isotopic compositions were controlled by the mass-dependent fractionation of oxygen isotopes¹⁸.

The "In-run" O isotope ratios determined for the oxide interference correction in the calculation scheme of equations (1) to (15) represent the isotope ratios of oxygen that form MoO₃-ions. They do not necessarily match the isotope ratios of the source oxygen for MoO_3 including O_2 gas remaining in the vacuumed sample chamber, the activator, and/or the sample solvent. This is because the equilibrium fractionation of O isotopes can occur between the source O_2 and MoO_3 , the fractionation of which is controlled by isotope masses of atomic O (m = 16, 17, and 18). This type of fractionation is illustrated in Figure 2 as a dashed line passing through the "Mean" O data point. Similarly, fractionation of O isotopes would occur via kinetic processes of the source oxygen (e.g., evaporation) before the formation of MoO₃, in which case the trend of fractionation is indistinguishable from that of equilibrium fractionation in Figure 2. In contrast, once MoO₃ has formed in the TIMS ion source, they fractionate as molecular ions of MoO_3^- until they are detected by Faraday cups, which is controlled by molecular masses of MoO₃ (m = 140-150). This type of fractionation is illustrated in Figure 2 as a thin line passing through the "Mean O" data point, which has a slope marginally different but nearly equal to the dashed line. In summary, the variation of "In-run" O isotopic composition observed in this study is attributed to the combination of three potential mass fractionation processes described above, which have mass fractionation trends nearly indistinguishable from one another in Figure 2.

Spectrometry Accepted Manuscript

Journal of Analytical Atomic

Our "Mean" O isotope composition is slightly different from those of previous studies, including the Nier's O. We advocate that the deviation between the "Mean" O and the Nier's O compositions is possibly caused by large analytical uncertainties or inaccurate quantification of O isotopic compositions by Nier's study. In fact, O isotopic compositions, which are published by IUPAC, are plotted on our mass-fractionation line passing through the "Mean" O within the uncertainty (the error of $^{17}O/^{16}O$ for the IUPAC "best" value is ~9 $\times 10^{-7}$) (Figure 2). Therefore, we believe that mass-independent fractionation or isobaric interference did not produce the deviation.

3.3. Molybdenum isotopic compositions in standard material

The Mo isotopic composition of our in-house standard Kanto-Mo has been

repeatedly analyzed (n = 21) with N-TIMS over the course of this study (Jan–March 2015).
Table 3 summarizes the results obtained by applying the "Nier's O," "Mean O," and
"In-run O" corrections.

397 3.3.1. "Nier's O" and "Mean O" corrections

In regards to the "Nier's O" correction, the reproducibilities (2 SD) of Mo isotope ratios for Kanto-Mo were 48, 13, 15, 23, and 46 ppm for ⁹²Mo/⁹⁶Mo, ⁹⁴Mo/⁹⁶Mo, ⁹⁵Mo/⁹⁶Mo, ⁹⁷Mo/⁹⁶Mo, and ¹⁰⁰Mo/⁹⁶Mo, respectively. The consistent reproducibilities were obtained for the "Mean-O" correction. In contrast, all of the Mo isotopic ratios obtained by the "Mean-O" correction resulted in significant shifts from those of the "Nier's O" correction. This indicates that the "Mean-O" corrections are simply a linear shift of the Mo isotope ratios from those of the "Nier's O" correction. The Mo isotopic shift between the two correction methods is evaluated by modifying equation (17), in which case $({}^{i}Mo/{}^{96}Mo)_{sample}$ and $({}^{i}Mo/{}^{96}Mo)_{std}$ are replaced by $({}^{i}Mo/{}^{96}Mo)_{Mean-O}$ and $({}^{i}Mo/{}^{96}Mo)_{Nier's O}$, respectively. As shown in Table 3, the μ^i Mo(Mean/Nier) were 98, -11, 54, -104, and -105 ppm for ${}^{i}Mo = {}^{92}Mo$, ${}^{94}Mo$, ${}^{95}Mo$, ${}^{97}Mo$, and ${}^{100}Mo$, respectively. The shifts were clearly caused by the difference in the O isotopic composition used in the correction. However, geochemical applications of Mo isotopes commonly utilize the δ , ε , or μ notation in which the Mo isotope ratio in the sample of interest is expressed by the relative deviation ($\times 10^3$, 10^4 , or 10^6 , respectively) from that of a standard material¹⁴. In such cases, the two

413 correction methods are essentially identical.

- 415 3.3.2. "In-run O" correction

As noted above, we observed variable O isotopic compositions of MoO_3^- ions across multiple isotopic measurements, which is consistent with the case of Os isotope analysis with N-TIMS^{33,38}. The Mo isotope ratios in Kanto-Mo obtained by the "In-run O" correction resulted in the following reproducibilities (2 SD): 47, 15, 10, 13, and 32 ppm for $^{92}Mo/^{96}Mo$, $^{94}Mo/^{96}Mo$, $^{95}Mo/^{96}Mo$, $^{97}Mo/^{96}Mo$, and $^{100}Mo/^{96}Mo$, respectively (Table 3). These are 1.0–1.8 times smaller than those obtained by the "Nier's O" and "Mean O" corrections, except for $^{94}Mo/^{96}Mo$.

In Figure 3, the ⁹⁷Mo/⁹⁶Mo ratios of Kanto-Mo obtained by the "Mean O" and "In-run O" corrections are plotted against the "In-run" ¹⁸O/¹⁶O values. Clearly, the "Mean O" ⁹⁷Mo/⁹⁶Mo increases as the ¹⁸O/¹⁶O increases (Figure 3a), whereas the variation of "In-run O" 97 Mo/ 96 Mo is independent to the 18 O/ 16 O value (Figure 3b). This indicates that the Mo isotopic composition measured using MoO₃ions with N-TIMS is strongly controlled by the "In-run" O isotopic compositions. We conclude that the Mo isotope measurement with N-TIMS requires the correction of O isotopic interferences using the O isotope composition obtained in individual isotope runs.

431 In addition to the three correction methods mentioned above, we examined the432 "Line-by-line O" correction in which the correction of O isotope interferences was

433	performed in individual data acquisition cycles (16.777 s) within a single isotopic run
434	consisting of 360 cycles. The reproducibilities (2SD) of Mo isotope ratios in Kanto-Mo by
435	this approach were 47, 16, 10, 13, and 33 ppm for ⁹² Mo/ ⁹⁶ Mo, ⁹⁴ Mo/ ⁹⁶ Mo, ⁹⁵ Mo/ ⁹⁶ Mo,
436	⁹⁷ Mo/ ⁹⁶ Mo, and ¹⁰⁰ Mo/ ⁹⁶ Mo, respectively. These are generally consistent with those
437	obtained by the "In-run O" correction, but the "Line-by-line O" correction in theory
438	provides a more accurate Mo isotopic composition. Therefore, in the following, we use the
439	Mo isotope ratios in various samples obtained by the "Line-by-line O" correction. Note that
440	the analytical uncertainties for all the μ Mo values were improved by 1.3–2.7 times
441	compared to those of the standard measurements in previous studies conducted by
442	MC-ICP-MS. ^{14,15} These studies reported that the external reproducibilities (2SD) of the
443	standard (Alfa Aesar Mo) are 72, 43, 26, 21, and 43 ppm for ⁹² Mo/ ⁹⁶ Mo, ⁹⁴ Mo/ ⁹⁶ Mo,
444	95 Mo/ 96 Mo, 97 Mo/ 96 Mo, and 100 Mo/ 96 Mo, respectively (normalized to 98 Mo/ 96 Mo).

3.3.3. Mass interferences and blanks

Potential isobaric interferences in the Mo isotope analysis with N-TIMS (e.g., Zr and Ru) would cause inaccurate results. Thus, special care is required to ionize Ru because (i) Ru can produce a strong RuO_3^{-1} ion beam when using a Pt filament and $Ba(OH)_2$ activator³⁷, and (ii) the complete removal of Ru from Mo via ion exchange chromatography is difficult to achieve^{15,27}. To evaluate the effect of isobaric interferences, the mixed solution (Mo standard solution containing 2% Zr or 2% Ru, which was assumed to be separated by the

453 two-stage chemical separation procedure²⁷) was measured (Table 4). Within the uncertainty, 454 the molybdenum isotopic ratios in the mixed solution were indistinguishable with those of 455 the Kanto-Mo standard solution. The result indicates that the ionization of ZrO_3^- and RuO_3^- 456 ions with the combination of Re filament and La(NO3)₃ activator is negligible compared to 457 the ionization of MoO₃⁻ ions.

Molybdenum blanks during the N-TIMS analysis should cause unexpected isotopic shifts in Mo isotopic compositions, especially when analyzing isotopically anomalous extraterrestrial samples. The amount of Mo blank during the N-TIMS analysis was evaluated by loading 20 pg of the ⁹⁷Mo-enriched spike on the Re filament together with the $La(NO_3)_3$ activator. The filament temperature was increased as described in section 2.4.2 until the condition of sample measurement was met, and MoO₃⁻ ions were detected in the ion-counting mode using the central SEM. The amount of Mo blank during the N-TIMS analysis was calculated to be <56 pg by the isotope dilution method, assuming that the blank Mo has a terrestrial isotopic composition. The Mo blank has a negligible effect on the sample measurements examined in this study in which ~ 3000 ng of Mo was applied.

Journal of Analytical Atomic Spectrometry Accepted Manuscript

 3.3.4. Accuracy of Mo isotope analysis

To evaluate the accuracy of our Mo isotope analysis, we measured the Mo isotope compositions of two synthesized solutions, Mix-A and Mix-B. The two solutions were artificially enriched in ⁹²Mo, ⁹⁷Mo, and ¹⁰⁰Mo compared to the Kanto-Mo solution. As

mentioned earlier, the absolute isotopic compositions of the two synthesized samples are difficult to determine precisely; however, the two samples were synthesized by gravimetrically mixing the Kanto-Mo solution and the mixed spike solution enriched in 92 Mo, 97 Mo, and 100 Mo in different proportions such that the μ^{92} Mo, μ^{97} Mo, and μ^{100} Mo values in Mix-A can be predicted from those obtained by the analysis of Mix-B and the weights of the starting materials. As shown in Table 4, the μ Mo values obtained by the analysis of Mix-A match those of the predicted values within the uncertainty, confirming the accuracy of our Mo isotope analysis.

To further evaluate the accuracy of our mass spectrometry technique, the Mo isotope composition in three terrestrial rocks (JB-3, JA-3, JR-2), two iron meteorites (Henbury and Albion), as well as new Mo isotopic data for IIIAB Tambo Quemado, was investigated (Table 4). Molybdenum isotopic compositions of terrestrial rocks were not distinct from those of Kanto-Mo within the uncertainties. In contrast, the iron meteorites are known to possess positive values in μ^{92} Mo, μ^{94} Mo, μ^{95} Mo, μ^{97} Mo, and μ^{100} Mo relative to the terrestrial standard^{14,15}. As shown in Table 4 and Figure 4, we reproduced the positive uMo values associated with this meteorite. Molybdenum isotope anomalies for two iron meteorites, (Henbury and Albion) were consistent with previously reported values. In addition, new Mo isotopic compositions of IIIAB Tambo Quemado were determined here, which were identical to other IIIAB iron meteorites within uncertainty, including Henbury.

3.4. Application of the method

Extremely high precision isotope analysis for nontraditional transition metals (e.g., Mo. W) in terrestrial and extraterrestrial materials could provide new insights into the studies of dynamical and chemical processes in the early Solar System. For example, W isotopic analyses with less than 10 ppm of analytical precision revealed that the lunar mantle had a well-resolved ¹⁸²W excess of 20 ppm relative to the modern terrestrial mantle, which was caused by a giant impact and subsequent late veneer events^{39,40}. Conversely, excluding some carbonaceous chondrites and iron meteorites. Mo isotope compositions in extraterrestrial materials are difficult to resolve from the terrestrial material at the current level of analytical precision with MC-ICP-MS¹⁵. Our mass spectrometric technique developed in this study has the potential to better discriminate the Mo isotope compositions in a variety of meteorites that would reflect the diversity of their origin in terms of time and space in the early Solar System.

506	
507	4. Concluding Remarks
508	Based on this study, the following conclusions were reached:
509	(1) We examined four types of activators to evaluate the detection efficiencies of MoO_3^-
510	ions with N-TIMS. The optimal condition to ionize Mo was to load Mo on a Re single
511	filament together with $La(NO_3)_3$ so that $La/Mo = 5$.
512	(2) We discovered that the oxygen isotope composition of MoO_3^- ions significantly varied
513	across different isotopic measurements. To achieve highly precise Mo isotope analysis,
514	determining the in situ O isotope composition for each measurement and use the data
515	to correct for the O isotope interferences is important, rather than utilizing a fixed O
516	isotope composition throughout all measurements (e.g., Nier's O isotope
517	composition ³²).
518	(3) The Mo isotopic ratios in the Kanto-Mo standard solution obtained by the in-situ O
519	isotope correction resulted in the following reproducibilities: 47, 16, 10, 13, and 33
520	ppm for ⁹² Mo/ ⁹⁶ Mo, ⁹⁴ Mo/ ⁹⁶ Mo, ⁹⁵ Mo/ ⁹⁶ Mo, ⁹⁷ Mo/ ⁹⁶ Mo and ¹⁰⁰ Mo/ ⁹⁶ Mo, respectively.
521	The reproducibilities have been improved at least several times compared to those
522	obtained in previous MC-ICP-MS studies.

(4) The accuracy of our technique was confirmed by measuring two synthesized solutions
enriched with an abundance of ⁹²Mo, ⁹⁷Mo, and ¹⁰⁰Mo. In addition, the Mo isotope
ratios in two iron meteorites Henbury (IIIAB) and Albion (IVA) obtained in this study

1 2		
3 4 5	596	were consistent with these reported in a previous study ¹⁵ . In addition Ma isotone ratios
6 7	520	were consistent with those reported in a previous study . In addition, wo isotope fatios
8 9 10 11 12	527	in a new iron meteorite, Tambo Quemado (IIIAB), were determined. Our N-TIMS
	528	technique is suitable for the studies of nucleosynthetic isotope anomalies in
13 14	529	extraterrestrial materials as well as the mass-dependent Mo isotopic shift in
15 16 17	530	environmental samples.
18 19	531	
20 21 22		
23 24		
25 26 27		
28 29		
30 31		
32 33		
34 35		
36 37 38		
39 40		
41 42		
43 44		
45 46		
47 48		
49 50 51		
51 52 53		
54 55		
56 57		
58 59		
60		

1
2
3
4
5
6
1
8
9
10
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30 24
31 32
32 33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50 51
52
52
54
55
56
57
58
59
60

533 Acknowledgements

We are grateful to R.J. Walker for useful discussion. We thank T. Hirata and Y. Fukami for providing an iron meteorite sample. This research was supported by a grant for the Global COE Program, "From the Earth to Earths" from the Ministry of Education, Culture, Sports, Science, and Technology of Japan, and KAKENHI Grant Number 23340171 from the Japan Society for the Promotion of Science (JSPS).

539

2		
3 4 5	540	
6 7 8	541	References
9 10	542	(1) Böhkle, JK.; De Laeter, J. R.; De Bièvre, P.; Hidaka, H.; Peiser, H. S.; Rosman, K. J. R.
11 12 13 14 15 16 17 18 19 20 21 22 23 24 25 26 27 28 29 30 31 32 33 4 35 36 37 38 39 40	543	Taylor, P. D. P. J. Phys. Chem. Ref. Data, 2005, 34, 57-67.
	544	(2) Barling, J.; Arnold, G. L.; Anbar, A. D. Earth Planet. Sci. Lett., 2001, 193, 447-457.
	545	(3) Siebert, C.; Nägler, T. F.; von Blanckenburg, F.; Kramers, J. D. Earth Planet. Sci. Lett.,
	546	2003 , <i>211</i> , 159-171.
	547	(4) Siebert, C.; McManus, J.; Bice, A.; Poulson, R.; Berelson, W. Earth Planet. Sci. Lett.,
	548	2006 , <i>241</i> , 723-733.
	549	(5) Poulson, R. L.; Siebert, C.; McManus, J.; Berelson, W. M. Geology, 2006, 34, 617-620.
	550	(6) Arnold, G. L.; Anbar, A. D.; Barling, J.; Lyons, T. W. Science, 2004, 304, 87-90.
	551	(7) Nägler, T. F.; Siebert, C.; Lüschen, H.; Böttcher, M. E. Chem. Geol., 2005, 219,
	552	283-295.
	553	(8) Neubert, N.; Nägler, T. F.; Böttcher, M. E. Geology, 2008, 36, 775-778.
41 42 43	554	(9) Duan, Y.; Anbar, A. D.; Arnold, G. L.; Lyons, T. W.; Gordon, G. W.; Kendall, B.
44 45	555	Geochim. Cosmochim. Acta, 2010, 74, 6655-6668.
46 47 48	556	(10) Kendall, B.; Gordon, G. W.; Poulton, S. W.; Anbar, A. D. Earth Planet. Sci. Lett., 2011,
49 50 51	557	307, 450-460.
52 53	558	(11) Zhang, Y.; Gladyshev, V. N. J. Mol. Biol., 2008, 379, 881-899.
54 55 56 57 58 59 60	559	(12) Zerkle, A. L.; Scheiderich, K.; Maresca, J. A.; Liermann, L. J.; Brantley, S. L.

560	<i>Geobiology</i> , 2011 , <i>9</i> , 94-106.
561	(13) Bisterzo, S.; Gallino, R.; Straniero, O.; Cristallo, S.; Käppeler, F. Mon. Not. R. Astron.
562	Soc., 2011 , 418, 284-319.
563	(14) Dauphas, N.; Marty, B.; Reisberg, L. ApJ, 2002, 565, 640-644.
564	(15) Burkhardt, C.; Kleine, T.; Oberli, F.; Pack, A.; Bourdon, B.; Wieler, R. Earth Planet.
565	Sci. Lett., 2011, 312, 390-400.
566	(16) Trinquier, A.; Birck, J L.; Allègre, C. J. ApJ, 2007, 655, 1179-1185.
567	(17) Caro, G.; Bourdon, B.; Birck, JL.; Moorbath, S. Geochim. Cosmochim. Acta, 2006,
568	70, 164-191.
569	(18) Touboul, M.; Walker, R. J. Int. J. Mass Spectrom., 2011, 309, 109-117.
570	(19) Murthy, V., R. Geochim. Cosmochim. Acta, 1963, 27, 1171-1178.
571	(20) Yokoyama, T.; Fukami, Y.; Okui, W.; Ito, N.; Yamazaki, H. Earth Planet. Sci. Lett.,
572	2015 , <i>416</i> , 46-55.
573	(21) Völkening, J.; Köppe, M.; Heumann, K., G. Int. J. Mass Spectrom. Ion Proc., 1991,
574	107, 147-159.
575	(22) Völkening, J.; Köppe, M.; Heumann, K., G. Int. J. Mass Spectrom. Ion Proc., 1991,
576	107, 361-368.
577	(23) Becker, H.; Walker, R. J. Nature, 2003, 425, 152-155.
578	(24) Burkhardt, C.; Kleine, T.; Dauphas, N.; Wieler, R. Earth Planet. Sci. Lett., 2012,
579	357-358, 298-307.

2	
3	
4	
Ē	
Э	
6	
7	
و	
0	
9	
10	
11	
10	
12	
13	
14	
15	
10	
16	
17	
18	
10	
19	
20	
21	
22	
~~	
23	
24	
25	
26	
20	
27	
28	
29	
20	
30	
31	
32	
33	
33	
34	
35	
36	
27	
31	
38	
39	
40	
44	
41	
42	
43	
ΔΛ	
44	
45	
46	
47	
10	
40	
49	
50	
51	
51	
52	
53	
54	
55	
55	
56	
57	
58	
50	
<u>nu</u>	

60

580	(25) Quitte, G.; Birck,	J.,-L.; Capman,	F.; Allègre,	C., J.	Geostand.	Newsl.,	2002,	26
581	149-160.							

582 (26) Fukami, Y.; Kimura, J.; Irisawa, K.; Yokoyama, T.; Hirata, T. *41st Lunar and* 583 *Planetary Science Conference*, The Woodlands, Texas, March 1–5, 2010, #1649.

- 584 (27) Nagai, Y.; Yokoyama, T. Anal. Chem., **2014**, *86*, 4856-4863.
- 585 (28) Wieser, M. E.; De Laeter, J. R.; Varner, M. D. Int. J. Mass Spectrom., 2007, 265,
 586 40-48.
- 587 (29) Nagai, Y.; Yokoyama, T. *44st Lunar and Planetary Science Conference*, The
 588 Woodlands, Texas, March 1–5, 2013, #2373.
- 589 (30) Yokoyama, T.; Rai, V. K.; Alexander, C. M. O.; Lewis, R. S.; Carlson, R. W; Shirey, S.
- 590 B.; Thiemens, M. H.; Walker, R. J. *Earth Planet. Sci. Lett.*, **2007**, *259*, 567-580.
- 591 (31) DePaolo, D. J.; Wasserburg, G. J. *Geophys. Res. Lett.*, **1976**, *3*, 249-252.
- 592 (32) Nier A. O. *Phys. Rev.*, **1950**, *77*, 789-793.
- 593 (33) Lugdet, A.; Nowell, G. M.; Pearson, D. G. Chem. Geol., 2008, 248, 342-362.
- 594 (34) Lu, Q.; Masuda, A. Int. J. Mass Spectrom. Ion Process., 1994, 130, 65-72.
- 595 (35) Heumann, K. G.; Eisenhut, S.; Gallus, S.; Hebeda, E. H.; Nusko, R.; Vengosh, A.;
 596 Walczyk, T. *Analyst*, **1995**, 120, 1291-1299.
- 597 (36) Walczyk, T.; Hebeda E. H.; Heumann, K. G. Fresenius J. Anal. Chem., 1991, 341,
 598 537-541.
- 599 (37) Becker, H.; Walker, R. J. Chem. Geol., 2003, 196, 43-56.

(38) Liu, Y.; Huang, M.; Masuda, A.; Inoue, M. Int. J. Mass Spectrom. Ion Process., 1998, 173, 163-175. (39) Touboul, M.; Puchtel, I. S.; Walker, R. J. Nature, 2015, 520, 530-533. (40) Kruijer, T. S.; Kleine, T.; Fischer-Gödde, M.; Sprung, P. Nature, 2015, 520, 534-537. (41) Wasserburg, G. J.; Jacobsen, S. B.; DePaolo, D. J.; McCulloch, M.T.; Wen, T. Geochim. Cosmochim. Acta, 1981, 45, 2311-2323. (42) Reisberg, L.; Zindler, A. Earth Planet. Sci. Lett., 1986, 81, 29-45. (43) Thirlwall, M. F. Chem. Geol., 1991, 94, 13-22. (44) Griselin, M.; van Belle, J. C.; Pomies, C.; Vroon, P. Z.; van Soest, M. C.; Davies, G. R. Chem. Geol., 2001, 172, 347-359. (45) Harvey, J.; Baxter, E. F. Chem. Geol., 2009, 258, 251-257.

2		
3 ⊿		
5	612	
7 8	613	Figure Captions
9 10 11	614	Figure 1. Detection efficiencies of MoO_3^- ions as a function of La/Mo ratio on the filament
12 13	615	when loading 1000–4000 ng of the Mo standard with $La(NO_3)_3$.
14 15 16	616	Figure 2. Variation of O isotope ratios across multiple Mo isotope measurements with
17 18	617	N-TIMS ("In-run" O, $n = 21$). The "Mean" O isotope ratios were obtained by
20 21	618	averaging "In-run" O data. Diamond symbols represent the O isotope ratios from
22 23 24	619	N-TIMS measurements in previous studies. ^{1,33,38,41-45} The Nier's O isotope ratios are
25 26	620	from ref. 32. Dashed (gray) and thin (black) lines passing through the "Mean" O
27 28 29	621	data point are mass fractionation trends of O isotopes for atomic oxygen ($m = 16, 17,$
30 31 32	622	18) and 100 MoO ₃ (<i>m</i> = 148, 149, 150), respectively.
33 34	623	Figure 3. Mo isotopic ratios (⁹⁷ Mo/ ⁹⁶ Mo) plotted against the in-situ O isotope ratios, i.e.,
35 36 37	624	$(^{18}O/^{16}O)_{In-run}$ used for the O interference corrections. (a) Mo isotopic ratios
38 39 40	625	corrected by the "Mean" O, and (b) Mo isotopic ratios corrected by the "In-run" O.
41 42	626	The light blue line represents the regression line calculated using ISOPLOT 3.00,
43 44 45	627	including the errors of O isotopic ratios (2SE) and Mo isotopic ratios (2SE).
46 47 48	628	Figure 4. μ^{95} Mo values for three iron meteorites (Henbury, Alibion, and Tambo Quemado)
49 50	629	measured in this study. Those of the same iron meteorites (Henbury, Albion)
51 52 53	630	obtained in a previous study (ref. 15) are also plotted. The upper gray band is the
54 55 56 57 58	631	reproducibility (2SD) of μ^{95} Mo in the Kanto-Mo standard obtained by the

1
2
3
4
5
6
7
0
0
9
10
11
12
13
14
15
16
17
18
19
20
20
∠ I 22
22
23
24
25
26
27
28
29
30
31
32
22
22
04 05
35
36
37
38
39
40
41
42
43
44
45
40 76
40 17
47
48
49
50
51
52
53
54
55
56
57
57
Эð

632 "Line-by-line" O correction at 10 ppm. The lower gray band represents the 633 reproducibility of μ^{95} Mo in a standard obtained in ref. 15 at 26 ppm.













						1	Table 1. Cup configurati	on of Mo isotopic analy	sis						
cup	L4		L3	L2	L1	Center	H1		H2	H3	H4				
mass	140	141	142	143	144	145	146	147	148	149	150	151	152	153	154
	⁹² Mo ¹⁶ O ₃		94Mo16O3	⁹⁵ Mo ¹⁶ O ₃	⁹⁶ Mo ¹⁶ O ₃	⁹⁷ Mo ¹⁶ O ₃	⁹⁸ Mo ¹⁶ O ₃		¹⁰⁰ Mo ¹⁶ O ₃						
							O isotope i	interferences							
		92Mo16O217O		94Mo ¹⁶ O ₂ ¹⁷ O ⁻	95Mo16O217O	⁹⁶ Mo ¹⁶ O ₂ ¹⁷ O	97Mo16O217O	⁹⁸ Mo ¹⁶ O ₂ ¹⁷ O ⁻		100Mo16O217O					
			92Mo ¹⁶ O ₂ ¹⁸ O [°]		94Mo16O218O	⁹⁵ Mo ¹⁶ O ₂ ¹⁸ O	⁹⁶ Mo ¹⁶ O ₂ ¹⁸ O	97Mo16O218O	⁹⁸ Mo ¹⁶ O ₂ ¹⁸ O [°]		100Mo16O218O				
			92Mo16O17O2		94Mo16O17O2	95Mo16O17O2	96Mo16O17O2	97Mo16O17O2	98Mo16O17O2		100Mo16O17O2				
				92Mo16O17O18O		94Mo16O17O18O	95Mo16O17O18O	96Mo16O17O18O	97Mo16O17O18O	98Mo16O17O18O		100Mo16O17O18O			
				⁹² Mo ¹⁷ O ₃		94Mo ¹⁷ O ₃	⁹⁵ Mo ¹⁷ O ₃	⁹⁶ Mo ¹⁷ O ₃	⁹⁷ Mo ¹⁷ O ₃	⁹⁸ Mo ¹⁷ O ₃		100Mo ¹⁷ O ₃			
					92Mo16O18O2		94Mo ¹⁶ O ¹⁸ O ₂	95Mo16O18O2	96Mo16O18O2	97Mo16O18O2	98Mo16O18O2		100Mo ¹⁶ O ¹⁸ O ₂		
					92Mo17O218O		94Mo17O218O	95Mo17O218O	96Mo17O218O	97Mo17O218O	98Mo17O218O		100Mo17O218O		
					-	92Mo17O18O2	-	94Mo17O18O2	95Mo17O18O2	96Mo17O18O2	97Mo17O18O2	98Mo17O18O2	-	100Mo ¹⁷ O ¹⁸ O ₂	
							⁹² Mo ¹⁸ O ₃		⁹⁴ Mo ¹⁸ O ₃	⁹⁵ Mo ¹⁸ O ₃	⁹⁶ Mo ¹⁸ O ₃	⁹⁷ Mo ¹⁸ O ₃	⁹⁸ Mo ¹⁸ O ₃		100Mo ¹⁸ O ₃

Journal of Analytical Atomic Spectrometry

	¹⁸ O/ ¹⁶ O	2SE	¹⁷ O/ ¹⁶ O	2SE	reference
this study					
T0374F01	0.0020639	0.0000003	0.0003884	0.0000002	
T0374F02	0.0020641	0.0000002	0.0003882	0.0000002	
T0374F03	0.0020653	0.0000003	0.0003880	0.0000002	
T0374F04	0.0020640	0.0000003	0.0003881	0.0000002	
T0374F05	0.0020639	0.0000002	0.0003883	0.0000002	
T0374F06	0.0020664	0.0000002	0.0003883	0.0000002	
T0374F07	0.0020640	0.0000002	0.0003883	0.0000001	
T0385F01	0.0020583	0.0000003	0.0003876	0.0000002	
T0385F02	0.0020599	0.0000002	0.0003874	0.0000002	
T0385F03	0.0020657	0.0000002	0.0003881	0.0000002	
T0385F04	0.0020622	0.0000002	0.0003880	0.0000002	
T0385F05	0.0020640	0.0000003	0.0003879	0.0000002	
T0385F06	0.0020668	0.0000002	0.0003879	0.0000002	
T0385F07	0.0020629	0.0000002	0.0003878	0.0000002	
T0385F08	0.0020633	0.0000002	0.0003878	0.0000001	
T0385F09	0.0020661	0.0000002	0.0003875	0.0000002	
T0385F10	0.0020685	0.0000002	0.0003885	0.0000002	
T0390F01	0.0020615	0.0000002	0.0003880	0.0000002	
T0390F02	0.0020606	0.0000002	0.0003877	0.0000002	
T0390F03	0.0020625	0.0000002	0.0003880	0.0000002	
T0390F04	0.0020642	0.0000002	0.0003883	0.0000002	
average (2SD)	0.0020637	0.0000048	0.0003880	0.0000006	
previous studies					
	0.0020439		0.0003749		Nier (1950) ³²
	0.0021100		0.0003870		Wasserburg et al. (1981) ⁴¹
	0.0021600		0.0003960		Nyquist in Wasserburg (1981) ⁴¹
	0.0020120		0.0003650		Reisberg and Zindler (1986) ⁴²
	0.0021140		0.0003920		Thirlwall (1991) ⁴³
	0.0021171		0.0003936		Liu et al. (1998) ³⁸
	0.0020400		0.0003860		Griselin et al. (2001) ⁴⁴
	0.0020052		0.0003799		Böhlke et al. (2005) ¹ *
	0.0020520		0.0003930		Harvey and Baxter (2009) ⁴⁵
	0.0020349		0.0003835		Luguet et al. $(2008)^{33}$

Journal of Analytical Atomic Spectrometry

-105

-104

	92		94		95		97		100	
Nier O correction	Mo/ Mo		Mo/ Mo		Mo/ Mo		Mo/ Mo		Mo/ Mo	
T0374F01	0.883137	0.00008	0.552482	0.000003	0.953195	0.000004	0.574017	0.000002	0.581539	0.00
T0374F02	0.883141	0.000006	0.552484	0.000003	0.953201	0.000004	0.574022	0.000002	0.581536	0.000
T0374F03	0.883138	0.000006	0.552484	0.000003	0.953198	0.000004	0.574025	0.000002	0.581544	0.00
T0374F04	0.883153	0.000005	0.552484	0.000002	0.953204	0.000003	0.574019	0.000002	0.581537	0.00
T0374F05	0.883143	0.000006	0.552483	0.000002	0.953199	0.000003	0.574022	0.000001	0.581534	0.00
T0374F06	0.883160	0.000006	0.552493	0.000002	0.953199	0.000003	0.574026	0.000001	0.581549	0.00
T0374F07	0.883149	0.000006	0.552488	0.000002	0.953201	0.000003	0.574020	0.000002	0.581541	0.00
T0385F01	0.883187	0.000006	0.552483	0.000003	0.953221	0.000004	0.574011	0.000002	0.581532	0.00
T0385F02	0.883131	0.000006	0.552477	0.000003	0.953217	0.000003	0.574017	0.000002	0.581530	0.00
T0385F03	0.883156	0.000006	0.552491	0.000003	0.953210	0.000003	0.574029	0.000003	0.581555	0.00
T0385F04	0.883173	0.000006	0.552487	0.000002	0.953215	0.000003	0.574021	0.000001	0.581544	0.00
T0385F05	0.883136	0.000006	0.552485	0.000003	0.953206	0.000003	0.574027	0.000002	0.581559	0.00
T0385F06	0.883114	0.000006	0.552488	0.000003	0.953204	0.000004	0.574036	0.000002	0.581564	0.00
T0385F07	0.883140	0.000006	0.552483	0.000003	0.953212	0.000003	0.574024	0.000002	0.581546	0.00
T0385F08	0.883137	0.000006	0.552485	0.000002	0.953202	0.000003	0.574023	0.000001	0.581550	0.00
T0385F09	0.883102	0.000005	0.552480	0.000002	0.953196	0.000003	0.574031	0.000001	0.581551	0.00
T0385F10	0.883099	0.000006	0.552487	0.000003	0.953196	0.000003	0.574038	0.000002	0.581565	0.00
T0390F01	0.883136	0.000007	0.552485	0.000003	0.953207	0.000003	0.574017	0.000001	0.581520	0.00
T0390F02	0.883138	0.000006	0.552482	0.000003	0.953211	0.000003	0.574017	0.000002	0.581517	0.00
T0390F03	0.883122	0.000006	0.552482	0.000003	0.953208	0.000003	0.574023	0.000001	0.581525	0.00
T0390F04	0.883115	0.000005	0.552482	0.000002	0.953200	0.000003	0.574021	0.000001	0.581535	0.00
average*	0.883138	0.000043	0.552485	0.000007	0.953205	0.000015	0.574023	0.000013	0.581542	0.00
reproducibility [ppm]	-	48	-	13	-	15	-	23	-	4
the Mean O correction										
T0374F01	0.883224	0.000008	0.552476	0.000003	0.953247	0.000004	0.573958	0.000002	0.581478	0.00
T0374F02	0.883228	0.000006	0.552478	0.000003	0.953253	0.000004	0.573962	0.000002	0.581475	0.00
T0374F03	0.883225	0.000006	0.552478	0.000003	0.953250	0.000004	0.573965	0.000002	0.581483	0.00
T0374F04	0.883239	0.000005	0.552478	0.000002	0.953256	0.000003	0.573959	0.000002	0.581476	0.00
T0374F05	0.883230	0.000006	0.552477	0.000002	0.953251	0.000003	0.573962	0.000001	0.581473	0.00
T0374F06	0.883248	0.000006	0.552487	0.000002	0.953250	0.000003	0.573967	0.000001	0.581488	0.00
T0374F07	0.883236	0.000006	0.552482	0.000002	0.953253	0.000003	0.573960	0.000002	0.581479	0.00
T0385F01	0.883274	0.000006	0.552477	0.000003	0.953272	0.000004	0.573951	0.000002	0.581471	0.00
T0385F02	0.883218	0.000006	0.552471	0.000003	0.953269	0.000003	0.573957	0.000002	0.581469	0.00
T0385F03	0.883243	0.000006	0.552485	0.000003	0.953262	0.000003	0.573969	0.000003	0.581493	0.00
T0385F04	0.883260	0.000006	0.552481	0.000002	0.953267	0.000003	0.573962	0.000001	0.581483	0.00
T0385F05	0.883223	0.000006	0.552479	0.000003	0.953257	0.000003	0.573968	0.000002	0.581498	0.00
T0385F06	0.883201	0.000006	0.552482	0.000003	0.953256	0.000004	0.573976	0.000002	0.581503	0.00
T0385F07	0.883227	0.000006	0.552477	0.000003	0.953263	0.000003	0.573965	0.000002	0.581484	0.00
T0385F08	0.883224	0.000006	0.552479	0.000002	0.953254	0.000003	0.573963	0.000001	0.581489	0.00
T0385F09	0.883188	0.000005	0.552474	0.000002	0.953247	0.000003	0.573972	0.000001	0.581490	0.00
T0385F10	0.883185	0.000006	0.552481	0.000003	0.953247	0.000003	0.573979	0.000002	0.581504	0.00
T0390F01	0.883223	0.000007	0.552479	0.000003	0.953259	0.000003	0.573957	0.000001	0.581459	0.00
T0390F02	0.883225	0.000006	0.552476	0.000003	0.953263	0.000003	0.573957	0.000002	0.581455	0.00
T0390F03	0.883209	0.000006	0.552476	0.000003	0.953259	0.000003	0.573964	0.000001	0.581463	0.00
T0390F04	0.883202	0.000005	0.552476	0.000002	0.953252	0.000003	0.573961	0.000001	0.581473	0.00
average*	0.883225	0.000043	0.552479	0.000007	0.953257	0.000015	0.573964	0.000013	0.581480	0.00
reproducibility [ppm]	-	48	-	13	-	15	-	23	-	
µ ⁱ Mo(Mean/Nier) [ppm]	98		-11		54		-104		-105	
the in-run O correction										
T0374F01	0.883226	0.000008	0.552477	0.000003	0.953248	0.000004	0.573957	0.000002	0.581490	0.00
T0374F02	0.883229	0.000006	0.552478	0.000003	0.953254	0.000004	0.573961	0.000002	0.581486	0.00
T0374F03	0.883226	0.000006	0.552475	0.000003	0.953252	0.000004	0.573962	0.000002	0.581491	0.00
T0374F04	0.883240	0.000005	0.552478	0.000002	0.953257	0.000003	0.573958	0.000002	0.581487	0.00
T0374F05	0.883232	0.000006	0.552478	0.000002	0.953252	0.000003	0.573962	0.000001	0.581485	0.00
T0374F06	0.883251	0.000006	0.552483	0.000002	0.953255	0.000003	0.573961	0.000001	0.581493	0.00
T0374F07	0.883238	0.000006	0.552482	0.000002	0.953254	0.000003	0.573959	0.000002	0.581491	0.00
T0385F01	0.883268	0.000006	0.552486	0.000003	0.953264	0.000004	0.573961	0.000002	0.581499	0.00
T0385F02	0.883212	0.000006	0.552477	0.000003	0.953262	0.000003	0.573965	0.000002	0.581492	0.00
		0.000006	0.552481	0.000003	0.953265	0.000003	0.573966	0.000003	0.581500	0.00
T0385F03	0.883245	0.000006	0.552484	0.000002	0.953265	0.000003	0.573964	0.000001		
T0385F03 T0385F04	0.883245 0.883259	0.000000							0.581500	0.00
T0385F03 T0385F04 T0385F05	0.883245 0.883259 0.883223	0.000006	0.552479	0.000003	0.953258	0.000003	0.573967	0.000002	0.581500 0.581509	0.00
T0385F03 T0385F04 T0385F05 T0385F06	0.883245 0.883259 0.883223 0.883202	0.000006	0.552479 0.552476	0.000003 0.000003	0.953258 0.953260	0.000003 0.000004	0.573967 0.573971	0.000002	0.581500 0.581509 0.581506	0.00 0.00 0.00
T0385F03 T0385F04 T0385F05 T0385F06 T0385F06 T0385F07	0.883245 0.883259 0.883223 0.883202 0.883225	0.000006 0.000006 0.000006	0.552479 0.552476 0.552478	0.000003 0.000003 0.000003	0.953258 0.953260 0.953262	0.000003 0.000004 0.000003	0.573967 0.573971 0.573967	0.000002 0.000002 0.000002	0.581500 0.581509 0.581506 0.581499	0.00 0.00 0.00
T0385F03 T0385F04 T0385F05 T0385F06 T0385F07 T0385F08	0.883245 0.883259 0.883223 0.883202 0.883225 0.883225	0.000006 0.000006 0.000006 0.000006	0.552479 0.552476 0.552478 0.552479	0.000003 0.000003 0.000003 0.000002	0.953258 0.953260 0.953262 0.953253	0.000003 0.000004 0.000003 0.000003	0.573967 0.573971 0.573967 0.573965	0.000002 0.000002 0.000002 0.000002	0.581500 0.581509 0.581506 0.581499 0.581502	0.00 0.00 0.00 0.00
T0385F03 T0385F04 T0385F05 T0385F06 T0385F07 T0385F08 T0385F09	0.883245 0.883259 0.883223 0.883202 0.883225 0.883223 0.883223 0.883187	0.000006 0.000006 0.000006 0.000006 0.000006	0.552479 0.552476 0.552478 0.552479 0.552468	0.000003 0.000003 0.000003 0.000002 0.000002	0.953258 0.953260 0.953262 0.953253 0.953250	0.000003 0.000004 0.000003 0.000003 0.000003	0.573967 0.573971 0.573967 0.573965 0.573969	0.000002 0.000002 0.000002 0.000002 0.000001 0.000001	0.581500 0.581509 0.581506 0.581499 0.581502 0.581495	0.00 0.00 0.00 0.00 0.00
70385F03 70385F04 70385F05 70385F06 70385F06 70385F08 70385F09 70385F10	0.883245 0.883259 0.883223 0.883202 0.883202 0.883225 0.883223 0.883223 0.883187 0.883191	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005	0.552479 0.552476 0.552478 0.552479 0.552468 0.552473	0.000003 0.000003 0.000003 0.000002 0.000002 0.000002	0.953258 0.953260 0.953262 0.953253 0.953250 0.953255	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003	0.573967 0.573971 0.573967 0.573965 0.573969 0.573970	0.000002 0.000002 0.000002 0.000001 0.000001 0.000001	0.581500 0.581509 0.581506 0.581499 0.581502 0.581495 0.581502	0.00 0.00 0.00 0.00 0.00 0.00
T0385F03 T0385F05 T0385F05 T0385F06 T0385F07 T0385F08 T0385F09 T0385F09 T0385F09	0.883245 0.883259 0.883223 0.883202 0.883225 0.883223 0.883187 0.883191 0.883221	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006	0.552479 0.552476 0.552478 0.552479 0.552468 0.552473 0.552484	0.000003 0.000003 0.000003 0.000002 0.000002 0.000003 0.000003	0.953258 0.953260 0.953262 0.953253 0.953250 0.953255 0.953255 0.953256	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003	0.573967 0.573971 0.573967 0.573965 0.573969 0.573970 0.573970	0.000002 0.000002 0.000002 0.000001 0.000001 0.000002 0.000002	0.581500 0.581509 0.581506 0.581499 0.581502 0.581495 0.581502 0.581502 0.581478	0.00 0.00 0.00 0.00 0.00 0.00 0.00
T0385F03 T0385F04 T0385F05 T0385F06 T0385F07 T0385F09 T0385F09 T0385F10 T0390F01 T0390F02	0.883245 0.883259 0.883223 0.883202 0.883225 0.883223 0.883187 0.883191 0.883191 0.883221	0.000006 0.000006 0.000006 0.000006 0.000005 0.000006 0.000007 0.000006	0.552479 0.552476 0.552478 0.552479 0.552468 0.552473 0.552484 0.552484	0.000003 0.000003 0.000003 0.000002 0.000002 0.000003 0.000003 0.000003	0.953258 0.953260 0.953262 0.953253 0.953250 0.953255 0.953255 0.953256 0.953256	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003	0.573967 0.573971 0.573967 0.573965 0.573969 0.573970 0.573961 0.573963	0.000002 0.000002 0.000002 0.000001 0.000001 0.000002 0.000001 0.000002	0.581500 0.581509 0.581506 0.581499 0.581502 0.581495 0.581502 0.581478 0.581477	0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.0
T0285F03 T0285F05 T0285F05 T0285F05 T0285F07 T0285F09 T0385F09 T0385F10 T0390F01 T0390F01 T0390F03	0.883245 0.883259 0.883223 0.883202 0.883225 0.883223 0.883223 0.883187 0.883191 0.883221 0.883221 0.883221	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.000006 0.000006	0.552479 0.552476 0.552478 0.552479 0.552468 0.552473 0.552484 0.552481 0.552481	0.000003 0.000003 0.000002 0.000002 0.000003 0.000003 0.000003 0.000003	0.953258 0.953260 0.953262 0.953253 0.953250 0.953255 0.953256 0.953258 0.953258	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003	0.573967 0.573971 0.573967 0.573965 0.573969 0.573970 0.573961 0.573963 0.573966	0.000002 0.000002 0.000002 0.000001 0.000001 0.000002 0.000001 0.000002 0.000001	0.581500 0.581509 0.581506 0.581499 0.581502 0.581495 0.581502 0.581478 0.581477 0.581479	0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.0
TU385F03 TU385F05 TU385F05 TU385F05 TU385F08 TU385F09 TU385F09 TU385F10 TU386F01 TU390F01 TU390F02 TU390F03 TU390F04	0.883245 0.883259 0.883223 0.883202 0.883202 0.883223 0.883223 0.883223 0.883191 0.883221 0.883221 0.883208 0.883204	0.000006 0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.000006 0.000006 0.000006	0.552479 0.552476 0.552478 0.552478 0.552479 0.552468 0.552473 0.552484 0.552481 0.552481 0.552479 0.552476	0.000003 0.000003 0.000002 0.000002 0.000002 0.000003 0.000003 0.000003 0.000003 0.000003	0.953258 0.953260 0.953262 0.953253 0.953255 0.953255 0.953255 0.953258 0.953258 0.953258 0.953258	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003	0.573967 0.573971 0.573967 0.573965 0.573969 0.573970 0.573961 0.573963 0.573966 0.573960	0.000002 0.000002 0.000002 0.000001 0.000001 0.000002 0.000001 0.000001 0.000001	0.581500 0.581509 0.581506 0.581499 0.581502 0.581502 0.581502 0.581478 0.581477 0.581479 0.581479	0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.0
T0385F03 T0385F04 T0385F05 T0385F05 T0385F08 T0385F09 T0385F09 T0385F10 T0390F03 T0390F03 T0390F04 avrage*	0.883245 0.883259 0.883223 0.883222 0.883225 0.883223 0.883191 0.883221 0.883221 0.883221 0.883224 0.883204 0.883204	0.000006 0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.000006 0.000006 0.000005 0.000005	0.552479 0.552476 0.552478 0.552478 0.552479 0.552468 0.552484 0.552484 0.552481 0.552481 0.552479 0.552479	0.000003 0.000003 0.000002 0.000002 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003	0.953258 0.953260 0.953262 0.953253 0.953253 0.953255 0.953255 0.953258 0.953258 0.953258 0.953253 0.953253 0.953253	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003	0.573967 0.573971 0.573967 0.573965 0.573969 0.573961 0.573963 0.573966 0.573960 0.573960 0.573960	0.000002 0.000002 0.000002 0.000001 0.000001 0.000002 0.000001 0.000002 0.000001 0.000001 0.000001	0.581500 0.581509 0.581506 0.581499 0.581502 0.581495 0.581478 0.581477 0.581479 0.581479 0.581484 0.581493	0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.0
T0385F03 T0385F04 T0385F05 T0385F05 T0385F08 T0385F08 T0385F09 T0385F10 T0390F01 T0390F02 T0390F03 T0390F03 T0390F03 T0390F03 T0390F04 r0790f04	0.885245 0.885259 0.883220 0.883202 0.883202 0.883202 0.883223 0.883187 0.883211 0.883221 0.883208 0.883204 0.883204 0.883204	0.000006 0.000006 0.000006 0.000006 0.000006 0.000005 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000005 0.0000041 47	0.552479 0.552476 0.552478 0.552478 0.552479 0.552473 0.552484 0.552484 0.552484 0.552484 0.552479 0.552476 0.552479	0.000003 0.000003 0.000002 0.000002 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 15	0.953258 0.953260 0.953262 0.953253 0.953253 0.953255 0.953255 0.953255 0.953258 0.953258 0.953253 0.953253 0.953257	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.0000010	0.573967 0.573971 0.573967 0.573965 0.573969 0.573960 0.573960 0.573960 0.573960 0.573960	0.000002 0.00002 0.00002 0.00001 0.00001 0.00001 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000008 13	0.581500 0.581509 0.581506 0.581499 0.581502 0.581495 0.581495 0.581477 0.581477 0.581479 0.581484 0.581484	
T0385F03 T0385F04 T0385F06 T0385F07 T0385F07 T0385F09 T0385F09 T0385F09 T0385F0 T0380F02 T0390F02 T0390F03 T0390F04 average* reproducibility (ppm]	0.883245 0.883259 0.883202 0.883202 0.883202 0.883225 0.883223 0.883187 0.883191 0.883291 0.883208 0.883204 0.883204 0.883204 0.883204 0.883204 0.883204 0.883204 0.883204 0.883204 0.883204 0.883204 0.883204 0.883205 0.883206 0.88206 0.8900000000000000000000000000000000000	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.000006 0.000006 0.000005 0.000005 0.0000041 47	0.552479 0.552476 0.552478 0.552478 0.552479 0.552468 0.552473 0.552481 0.552479 0.552479 0.552479 - - -	0.00003 0.00003 0.00002 0.00002 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00002 0.00008 15	0.953258 0.953260 0.953262 0.953253 0.953255 0.953255 0.953256 0.953258 0.953258 0.953258 0.953253 0.953253 0.953257 - -	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.0000010 10	0.573967 0.573971 0.573965 0.573969 0.573969 0.573961 0.573961 0.573966 0.573966 0.573964 	0.000002 0.000002 0.000002 0.000001 0.000001 0.000002 0.000001 0.000001 0.000001 0.000001 0.000008 13	0.581500 0.581509 0.581506 0.581499 0.581502 0.581495 0.581495 0.581479 0.581479 0.581479 0.581484 0.581484 0.581484 0.581484	
10385F03 10385F04 10385F05 10385F05 10385F07 10385F07 10385F07 10385F0 10390F01 10390F02 10390F03 10390F04 average* reproducibility (ppm] μ ⁴ Mo((h-Run/Nicr) (ppm)	0.885245 0.883253 0.883223 0.883202 0.883202 0.883223 0.883187 0.883187 0.883210 0.883221 0.883221 0.883221 0.883222 0.883224 0.883225 	0.000006 0.000006 0.000006 0.000006 0.000005 0.000006 0.000006 0.000006 0.000006 0.000006 0.000005 0.0000041 47	0.552479 0.552476 0.552478 0.552478 0.552479 0.552484 0.552484 0.552484 0.552481 0.552479 0.552479 0.552479 0.552479	0.00003 0.00003 0.00002 0.00002 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00002 0.00008 15	0.953258 0.953260 0.953262 0.953253 0.953250 0.953255 0.953255 0.953258 0.953258 0.953258 0.953253 0.953253 0.953253 0.953257 - - 54	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.0000010 10	0.573967 0.573971 0.573967 0.573969 0.573969 0.573970 0.573961 0.573960 0.573960 0.573960 0.573960 0.573964 - -	0.000002 0.000002 0.000002 0.000001 0.000001 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 13	0.581500 0.581509 0.581502 0.581499 0.581502 0.581495 0.581470 0.581477 0.581477 0.581479 0.581484 0.581484 0.581484 0.581489 	
10285F03 10285F05 10285F05 10385F05 10385F07 10285F09 10285F09 10285F09 10385F0 10390F03 10390F03 10390F03 10390F04 average* reproducibility (ppm] μ ¹ Mo(In-Run/Nicr) (ppm] the line-by-line O correction	0.885245 0.882259 0.883223 0.883202 0.883202 0.883223 0.883187 0.883221 0.883221 0.883221 0.883221 0.883221 0.883220 0.883208 0.883204 0.883205 	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.000006 0.000006 0.000005 0.000005 0.000005 0.000005	0.552479 0.552476 0.552478 0.552478 0.552479 0.552484 0.552481 0.552481 0.552481 0.552479 0.552479 0.552479 0.552479	0.000003 0.000003 0.000002 0.000002 0.000002 0.000003 0.000003 0.000003 0.000003 0.000002 0.000008 15	0.953258 0.953260 0.953262 0.953253 0.953255 0.953255 0.953255 0.953258 0.953258 0.953258 0.953253 0.953257 - - 54	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.0000010 10	0.573967 0.573971 0.573965 0.573965 0.573969 0.573961 0.573966 0.573966 0.573966 0.573966 0.573966 0.573966 0.573966 	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 13	0.581500 0.581506 0.581506 0.581502 0.581502 0.581495 0.581475 0.581477 0.581477 0.581477 0.581473 0.581484 0.581493 - -	
T0385F03 T0385F04 T0385F05 T0385F07 T0385F07 T0385F09 T0385F09 T0385F09 T0390F01 T0390F02 T0390F02 T0390F03 T0390F04 average* reproducibility (ppm] p ⁴ Mo(In-RunNier) [ppm]	0.883245 0.883259 0.883223 0.883223 0.883222 0.883223 0.883191 0.883221 0.883221 0.883221 0.883224 0.883225 	0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006	0.552479 0.552476 0.552478 0.552479 0.552479 0.552473 0.552484 0.552484 0.552484 0.552479 0.552476 0.552479 - -11 0.552477	0.000003 0.000003 0.000002 0.000002 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000008 15	0.953258 0.953262 0.953253 0.953253 0.953255 0.953255 0.953258 0.953258 0.953258 0.953258 0.953258 0.953257 - 54	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.0000003 0.0000003 0.0000003 0.0000003 0.0000003 0.0000004	0.573967 0.573971 0.573971 0.573967 0.573969 0.573969 0.573963 0.573963 0.573960 0.573960 0.573964 - - -104	0.000002 0.000002 0.000001 0.000001 0.000001 0.000002 0.000001 0.000001 0.000001 0.000001 0.000008 13	0.581500 0.581506 0.581506 0.581502 0.581502 0.581502 0.581502 0.581477 0.581477 0.5814843 	0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.0
T0385F03 T0385F04 T0385F05 T0385F05 T0385F07 T0385F07 T0385F09 T0385F0 T0390F01 T0390F02 T0390F03 T0390F04 average* reproducibility (ppm] μ ⁴ Mo((hRun/Nier) (ppm] the line-by-line O correction T0374F01	0.885245 0.883253 0.883223 0.883202 0.883202 0.883223 0.883187 0.883204 0.883221 0.883221 0.883224 0.883225 - - 98 0.883225	0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.000007	0.552479 0.552478 0.552478 0.552478 0.552473 0.552473 0.552476 0.552476 0.552476 0.552477 0.552477	0.00003 0.00003 0.00003 0.00002 0.000002 0.000003 0.000003 0.000003 0.000003 15	0.953258 0.953260 0.953260 0.953253 0.953253 0.953255 0.953256 0.953258 0.953258 0.953258 0.953253 0.953257 - 54	0.00003 0.00004 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.000003 0.000001 10	0.573967 0.573967 0.573965 0.573965 0.573969 0.573969 0.573963 0.573963 0.573963 0.573966 0.573966 0.573969 0.573964 - - -104	0.00002 0.00002 0.00002 0.00001 0.00001 0.00001 0.00001 0.00001 0.000001 1.000001 1.000001 0.000001 1.000001 0.000002	0.581500 0.581509 0.581506 0.581502 0.581502 0.581495 0.581495 0.581477 0.581474 0.581474	
T0385F03 T0385F04 T0385F05 T0385F05 T0385F07 T0385F09 T0385F09 T0385F0 T0390F01 T0390F02 T0390F03 T0390F04 average* reproducibility (ppm] p ⁴ Mo(th-Run/Nicr) (ppm] the line-by-line O correction T0374F01 T0374F02	0.885245 0.882259 0.883223 0.883202 0.883202 0.883203 0.883187 0.883203 0.883221 0.883221 0.883220 	0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00007 0.000007 0.000007	0.552479 0.552479 0.552478 0.552478 0.552479 0.552484 0.552479 0.552479 0.552479 - -11 0.552477 0.552477 0.552477 0.552477 0.552477	0.00003 0.00003 0.00003 0.00002 0.00002 0.00003 0.00003 0.00003 0.00003 15	0.953258 0.953260 0.953260 0.953253 0.953253 0.953255 0.953258 0.953258 0.953258 0.953258 0.953258 0.953258 0.953258	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000001 10	0.573967 0.573967 0.573965 0.573965 0.573969 0.573969 0.573963 0.573966 0.573966 0.573966 0.573966 0.573966 0.573964 0.573957 0.573957	0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000002 0.000001 0.000001 0.000001 0.000001 13	0.581500 0.581506 0.581506 0.581499 0.581502 0.581499 0.581502 0.581478 0.581478 0.581479 0.581484 0.581479 0.581474 0.581474	0.000 0.0000 0.000000
T0385F03 T0385F04 T0385F06 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0380F01 T0390F02 T0390F03 T0390F03 T0390F04 average* reproducibility (ppm) µ' Mo(In-Run/Ner) (ppm) the line-by-line O correction T0374F02 T0374F03	0.883245 0.883253 0.883223 0.883223 0.883222 0.883223 0.883191 0.883221 0.883221 0.883221 0.883221 0.883224 0.883225 - 98 0.883227 0.883227 0.883227 0.883227 0.883227	0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00007 0.000007 0.000007	0.552479 0.552476 0.552478 0.552478 0.552473 0.552473 0.552473 0.552479 0.552479 0.552477 0.552477 0.552477	0.00003 0.00003 0.00003 0.00002 0.00002 0.000002 0.000003 0.000003 0.000003 15	0.953258 0.953260 0.953262 0.953262 0.953253 0.953255 0.953256 0.953256 0.953258 0.953258 0.953257 	0.00003 0.00004 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00004 10 0.00004 0.00004 0.00004	0.573967 0.573971 0.573965 0.573965 0.573969 0.573969 0.573969 0.573961 0.573964 0.573964 - -104 0.573957 0.573951	0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002	0.581500 0.851509 0.581506 0.581499 0.581499 0.581495 0.581495 0.581478 0.581479 0.581479 0.581473 - 	0.00
T0385F03 T0385F04 T0385F05 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0385F0 T0390F04 T0390F04 average* reproducibility [pm] µ ¹ Mo(In-Run/Nicr) [pm] the line-by-line O correction T0374F03 T0374F03 T0374F04 T0374F04	0.885245 0.885253 0.883223 0.883202 0.883223 0.883223 0.883187 0.883221 0.883221 0.883224 0.883225 - - 98 0.883227 0.883227 0.883227 0.883227 0.883227	0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006	0.552479 0.552478 0.552478 0.552478 0.552473 0.552473 0.552484 0.552476 0.552476 0.552477 0.552477 0.552477 0.552477 0.552475 0.552475 0.552478	0.000003 0.000003 0.000003 0.000002 0.000002 0.000003 0.000003 0.000003 0.000003 15 0.000003 0.000003 0.000003 0.000003 0.0000003 0.0000003	0.953258 0.953260 0.953260 0.953260 0.953253 0.953250 0.953256 0.953258 0.953258 0.953258 0.953258 0.953258 0.953254 0.953254	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000004	0.573967 0.573967 0.573967 0.573965 0.573969 0.573969 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573961 0.573961 0.573962 0.573961	0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.000002	0.581500 0.581506 0.581506 0.581502 0.581502 0.581478 0.581478 0.581477 0.581474 0.581477 0.581477 0.581477 0.581477 0.581477	0.000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.000000
T0:85F03 T0:85F04 T0:85F06 T0:85F07 T0:85F07 T0:85F07 T0:85F07 T0:85F07 T0:90F02 T0:90F02 T0:90F03 T0:90F04 average* reproducibility (pm] μ ⁴ Mo(n-Run/Nier) (ppm] the line-by-line O correction T0:374F01 T0:374F03 T0:374F03 T0:374F03	0.885245 0.882259 0.883223 0.883202 0.883223 0.883203 0.883203 0.883221 0.883221 0.883220 0.883226 	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000005 0.000006 0.000006 0.000006 0.000006 0.000007 0.000006 0.000006 0.000006 0.000006	0.552479 0.552478 0.552478 0.552479 0.552473 0.552473 0.552473 0.552484 0.552479 0.552477 0.552477 0.552477 0.552477 0.552477	0.00003 0.00003 0.00002 0.00002 0.000002 0.000003 0.000003 0.000003 0.000003 15 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003	0.953258 0.953260 0.953262 0.953262 0.953253 0.953256 0.953256 0.953258 0.953258 0.953258 0.953258 0.953258 0.9532548 0.9533248 0.9533248 0.9533248	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000003 0.000004	0.573967 0.573967 0.573967 0.573969 0.573969 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573957 0.573957 0.573957 0.573957	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000001 0.000001 0.000001 0.000001	0.581500 0.851506 0.581409 0.581409 0.581402 0.581402 0.581402 0.581478 0.581478 0.581479 0.581474 0.581474 0.581474 0.581474 0.581477 0.581475 0.581475 0.581472	0.000 0.0000 0.0000 0.0000 0.0000 0.0000 0.0000 0.000000
T0385F03 T0385F04 T0385F06 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0380F02 T0390F02 T0390F02 T0390F03 T0390F03 T0390F03 T0390F03 t0390F03 t0390F03 t0390F03 t0390F03 T0374F03	0.883245 0.883225 0.883223 0.883223 0.883223 0.883223 0.883197 0.883197 0.883221 0.883221 0.883221 0.883225 - - 98 0.883225 0.883225 0.883227 0.883227 0.883227 0.883227 0.883227 0.883223	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006	0.552479 0.552476 0.552478 0.552473 0.552473 0.552473 0.552473 0.552479 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477	0.00003 0.00003 0.00003 0.00002 0.00002 0.00003 0.00003 0.00003 0.000003 0.000003 15 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003	0.953258 0.953260 0.953262 0.953262 0.953253 0.953250 0.953256 0.953256 0.953258 0.953258 0.953258 0.953254 0.953254 0.953254 0.9533254 0.9533254	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.0000003 0.000004 0.000004 0.0000004 0.0000004 0.0000004 0.0000004	0.573967 0.573967 0.573967 0.573969 0.573969 0.573963 0.573963 0.573960 0.573960 0.573960 0.573966 0.573964 0.573957 0.573961 0.573961 0.573961	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000002 0.000001 0.000002 13	0.581500 0.851509 0.581506 0.581499 0.581499 0.581495 0.581495 0.581477 0.581477 0.581473 0.581477 0.581477 0.581477 0.581477 0.581477 0.581478	0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.0
T0385F03 T0385F04 T0385F05 T0385F07 T0385F07 T0385F07 T0385F07 T0385F0 T0385F0 T0390F04 T0390F04 average* reproducibility (pm) µ ¹ Mo(h-Run/Nicr) (pm) the line-by-line O correction T0374F03 T0374F03 T0374F04 T0374F04 T0374F05 T0374F05	0.885245 0.885253 0.883223 0.883223 0.883223 0.883223 0.883187 0.883221 0.883221 0.883221 0.883224 0.883225 - - 98 0.883225 0.883227 0.883227 0.883227 0.883227 0.883225 0.883225 0.883221 0.883221 0.883221 0.883221	0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006 0.00006	0.552479 0.552476 0.552476 0.552479 0.552473 0.552473 0.552484 0.552479 0.552476 0.552479 0.552477 0.552477 0.552477 0.552477 0.552477 0.552478 0.552478 0.552478	0.00003 0.00003 0.00002 0.00002 0.000002 0.000003 0.000003 0.000003 0.000003 15 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000002 0.000002 0.000002	0.953258 0.953260 0.953260 0.953260 0.953250 0.953250 0.953256 0.953258 0.953258 0.953258 0.953258 0.953257 - 54 0.953254 0.953252 0.953252 0.953252 0.953252	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000004 0.000004 0.000004 0.000003 0.000003 0.000003	0.573967 0.573967 0.573967 0.573969 0.573969 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573961 0.573950 0.573951 0.573961 0.573951	0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.000002 0.000002	0.581500 0.881506 0.581506 0.581499 0.581502 0.581478 0.581478 0.581478 0.581474 0.581474 0.581474 0.581477 0.581474 0.581477 0.581477 0.581477	0.00 0.00 0.00 0.00 0.00 0.00 0.00 0.0
T0385F03 T0385F06 T0385F06 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0390F01 T0390F03 T0390F03 T0390F04 average* reproducibility (ppm] p ⁴ Mo(In-Run/Ner) (ppm] the line-by-line O correction T0374F03 T0374F03 T0374F04 T0374F04 T0374F04 T0374F04 T0374F04 T0374F04 T0374F04 T0374F04	0.883245 0.883253 0.883223 0.883223 0.883223 0.883127 0.883191 0.883221 0.883221 0.883224 0.883225 	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.000006 0.000006 0.000007 0.000007 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006	0.552479 0.552479 0.552479 0.552479 0.552473 0.552473 0.552479 0.552479 0.552476 0.552479 0.552477 0.552477 0.552477 0.552477 0.552477 0.552473 0.552473 0.552483 0.552483 0.552483	0.00003 0.00003 0.00002 0.00002 0.000002 0.000003 0.000003 0.000003 0.000003 15 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000002 0.000003 0.000002	0.953258 0.953260 0.953262 0.953262 0.953253 0.953256 0.953256 0.953256 0.953258 0.953258 0.953258 0.953257 - 54 0.953254 0.9533254 0.9533254 0.9533264	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000003 0.000004 0.00004 0.0004 0	0.573967 0.573967 0.573967 0.573969 0.573963 0.573963 0.573966 0.573966 0.573966 0.573966 0.573966 0.573967 0.573961 0.573961 0.573961 0.573961 0.573961	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001	0.581500 0.851506 0.581409 0.581409 0.581402 0.581402 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477	
T0385F03 T0385F04 T0385F06 T0385F07 T0385F07 T0385F07 T0385F07 T0385F00 T0390F02 T0390F02 T0390F02 T0390F03 T0390F03 T0390F03 T0390F04 average* reproducibility (ppm) μ⁴ Mo(In-Run/Nier) (ppm) the line-by-line O correction T0374F03 T0374F03 T0374F04 T0374F05 T0374F05 T0374F05 T0374F07 T0385F01 T0385F01 T0385F01	0.883245 0.883253 0.883223 0.883223 0.883223 0.883223 0.883223 0.883221 0.883221 0.883221 0.883221 0.883221 0.883225 - - 98 0.883225 0.883225 0.883225 0.883225 0.883227 0.883229 0.883229 0.883221 0.883221 0.883221 0.883221 0.883226 0.883246	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006	0.552479 0.552479 0.552478 0.552473 0.552473 0.552473 0.552473 0.552479 0.552479 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552478 0.552477	0.00003 0.00003 0.00003 0.00002 0.00002 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003	0.953258 0.953260 0.953262 0.953262 0.953253 0.953256 0.953256 0.953258 0.953258 0.953258 0.953258 0.953257 54 0.9533254 0.9533254 0.9533252 0.9533254 0.9533254 0.9533254	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000004 0.000004 0.000004 0.000003 0.00003 0.0000 0.000 000	0.573967 0.573967 0.573967 0.573969 0.573969 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573965 0.573965 0.573965 0.573965	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000002 0.000001 0.000001 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.0000001	0.581500 0.851506 0.581506 0.581409 0.581402 0.581402 0.581477 0.581477 0.581477 0.581473 0.581474 0.581474 0.581477 0.581474 0.5814	
T0385F03 T0385F04 T0385F05 T0385F05 T0385F08 T0385F09 T0385F09 T0385F09 T0390F01 T0390F02 T0390F03 T0390F04 average* reproducibility (pmp1) µ ⁴ Mo(1n-Run/Ner) (pm1) the line-by-line O correction T0374F03 T0374F03 T0374F04 T0374F05	0.885245 0.885253 0.883223 0.883223 0.883223 0.883223 0.883221 0.883221 0.883221 0.883221 0.883224 0.883225 - - 98 0.883225 0.883225 0.883225 0.883225 0.883225 0.883225 0.883225 0.883225 0.883225 0.883225 0.883225 0.883225 0.883221 0.883225 0.883225 0.883221 0.883225 0.883225 0.883221 0.883225	0.00006 0.00006	0.552479 0.552476 0.552478 0.552478 0.552479 0.552473 0.552473 0.552479 0.552479 0.552479 0.552477 0.5552477 0.5552477 0.5552477 0.5552477 0.5552477 0.5552473 0.5552477 0.5552473	0.00003 0.00003 0.000002 0.000002 0.000002 0.000003 0.00003 0.0003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.0003 0.0000 0.0003 0.00003 0	0.953288 0.953260 0.953262 0.953262 0.953253 0.953250 0.953256 0.953256 0.953258 0.953258 0.953257 	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003	0.573967 0.573967 0.573967 0.573969 0.573969 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573957 0.573951 0.573954 0.573955555555555555555555555555555555555	0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000001	0.581500 0.581506 0.581409 0.581502 0.581409 0.581502 0.581478 0.581478 0.581478 0.581478 0.581478 0.581474 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581474	
T0385F03 T0385F04 T0385F06 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0380F02 T0390F02 T0390F03 T0390F03 T0390F03 T0390F03 T0390F03 T0390F03 T0390F03 T0390F03 T0390F03 T0390F03 T0390F03 T0390F03 T0374F02 T0374F03 T0374F03 T0374F03 T0374F04 T0374F03 T0374F04 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0385F01 T0385F02 T0385F02	0.883245 0.883253 0.883223 0.883223 0.883222 0.883223 0.883191 0.883221 0.883221 0.883224 0.883225 - 98 0.883225 0.883227 0.883227 0.883225 0.883225 0.883225 0.883225 0.883225 0.883226 0.883225 0.883226 0.883226 0.883226 0.883226 0.883226 0.883226 0.883226 0.883226 0.883226 0.883226 0.883226 0.883240 0.883226 0.883240 0.88340000000000000000000000000000000000	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006	0.552479 0.552479 0.552478 0.552479 0.552473 0.552473 0.552479 0.552479 0.552479 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552481 0.552477 0.552482 0.552482 0.552484 0.552484 0.552484	0.00003 0.00003 0.00002 0.00002 0.00002 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003	0.953258 0.953260 0.953262 0.953262 0.953253 0.953255 0.953256 0.953258 0.953258 0.953258 0.953258 0.953257 0.953254 0.953254 0.953254 0.953252 0.953252 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953256 0.953266 0.953265	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003	0.573967 0.573967 0.573967 0.573965 0.573965 0.573963 0.573960 0.573966 0.573966 0.573966 0.573966 0.573966 0.573961 0.573961 0.573961 0.573961 0.573961 0.573964 0.573964 0.573964 0.573964	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.0000001 0.0000002 0.0000002 0.0000002 0.0000001 0.0000001 0.0000002 0.0000002 0.0000002 0.0000002 0.0000002	0.581500 0.851506 0.581499 0.581499 0.581499 0.581492 0.581472 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581478 0.581477 0.581472 0.581472 0.581472 0.581484 0.581484 0.581487	
T0385F03 T0385F04 T0385F04 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0380F01 T0390F04 average* reproducibility (ppm) μ ⁴ (bo(in-Run/Nicr) (ppm) the line-by-line O correction T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03	0.883245 0.883223 0.883223 0.883223 0.883223 0.883223 0.883223 0.883221 0.883221 0.883221 0.883221 0.883221 0.883225 - - 98 0.883225 0.883225 0.883225 0.883225 0.883225 0.883225 0.883227 0.883227 0.883227 0.883223 0.883226 0.883221 0.883226 0.883221 0.883226 0.883221 0.883226 0.883226 0.883221 0.883226 0.88326 0.8836 0.88326 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0.8836 0	0.000006 0.000006	0.552479 0.552476 0.552476 0.552479 0.552479 0.552473 0.552484 0.552476 0.552476 0.552476 0.552476 0.552477 0.552478 0.552477 0.552478 0.552477 0.552483 0.552483 0.552484 0.552477	0.00003 0.00003 0.00003 0.00002 0.00002 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.000	0.953258 0.953260 0.953262 0.953262 0.953253 0.953256 0.953256 0.953258 0.953258 0.953258 0.953257 54 0.953252 0.953252 0.953252 0.953252 0.953252 0.953252 0.953252 0.953252 0.953252 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953255 0.953254 0.953255 0.95355 0.95555 0.955555 0.9555555 0.95555555 0.95555555555	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000003 0.00003 0	0.573967 0.573967 0.573967 0.573967 0.573969 0.573963 0.573961 0.573960 0.573960 0.573960 0.573960 0.573964 0.573961 0.573961 0.573961 0.573961 0.573961 0.573965 0.573964 0.573964 0.573964 0.573964	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000002 0.000001 0.000001 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.0000002	0.581500 0.881506 0.581409 0.581502 0.581470 0.581477 0.581477 0.581473 0.581474 0.581474 0.581474 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581477 0.581474 0.581477 0.581474 0.581477 0.581474 0.581477 0.581477 0.581474 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581478 0.581477 0.581478 0.581477 0.581478 0.581477 0.581478 0.581477 0.581478 0.581477 0.581478 0.581477 0.581478 0.581477 0.581478 0.581477 0.581478 0.581479 0.581478 0.581479 0.581478 0.581479 0.581478 0.581479 0.581478 0.581479 0.581478 0.581479 0.5814	0.00 0.
T0385F03 T0385F04 T0385F06 T0385F07 T0385F07 T0385F07 T0385F07 T0380F01 T0390F01 T0390F03 T0390F04 average* productiolity (pm] p ⁴ Mo(h-Run/Nier) (pm] de line-by-line O correction T0374F01 T0374F01 T0374F03 T0374F03 T0374F03 T0374F03 T0374F04 T0374F03 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0374F05 T0385F03 T0385F03 T0385F04 T0385F05	0.885245 0.885253 0.885253 0.885223 0.885223 0.885223 0.885223 0.885221 0.885221 0.885221 0.885221 0.885221 0.885225 0.885225 0.885225 0.885225 0.885225 0.885225 0.885225 0.885225 0.885251 0.885251 0.885251 0.885251 0.885251 0.885251 0.885251 0.885264 0.885268 0.885268	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000005 0.000006 0.00000000	0.552479 0.552479 0.552478 0.552479 0.552473 0.552473 0.552473 0.552479 0.552479 0.552479 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552483 0.552483 0.552483	0.00003 0.00003 0.000003 0.000002 0.000002 0.000003 0.000003 0.000003 0.000003 0.000003 15 0.000003 0.000003 0.000003 0.000003 0.000002 0.000003 0.00003 0.0003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.	0.953258 0.953260 0.953262 0.953262 0.953253 0.953255 0.953256 0.953258 0.953258 0.953258 0.953257 	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000003 0.00003 0.0	0.573967 0.573967 0.573967 0.573969 0.573969 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573961 0.573961 0.573961 0.573961 0.573964 0.573965 0.573966 0.57396 0.57396	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000001 0.000001 0.000002 0.000001 0.000002 0.000002 0.000002 0.000002	0.581500 0.581506 0.581409 0.581409 0.581402 0.581402 0.581478 0.581478 0.581479 0.581484 0.581479 0.581474 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581478 0.581477 0.581478 0.581478 0.581478 0.581478 0.581478 0.581478 0.581480 0.581478 0.581480 0.581480 0.581488 0.581488	
T0385F03 T0385F04 T0385F06 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0380F02 T0390F02 T0390F03 T0390F03 T0390F04 average* reproducibility (ppm] the line-by-line O correction T0374F02 T0374F03 T0374F03 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F05 T0374F05 T0385F03 T0385F03 T0385F03 T0385F05 T0385F05	0.883245 0.883225 0.883223 0.883223 0.883223 0.883223 0.883221 0.883221 0.883221 0.883221 0.883221 0.883221 0.883225 - - 98 0.883225 0.88325 0.8835 0.8835 0.8835 0.8835 0.8835 0.8835 0.8835 0.8955 0.8955 0.8955 0.89555 0.895555 0.895555555555555	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.000006 0.000006 0.000007 0.000007 0.000007 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006 0.000006	0.552479 0.552479 0.552478 0.552479 0.552473 0.552473 0.552473 0.552476 0.552477 0.552477 0.552477 0.552477 0.552478 0.552477 0.552478 0.5525	0.00003 0.00003 0.00002 0.00002 0.00002 0.00003 0.00003 0.00003 0.00003 0.000003 15 0.000003 0.000003 0.000003 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002	0.953258 0.953260 0.953262 0.953262 0.953253 0.953255 0.953256 0.953256 0.953257 0.953258 0.953257 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953255 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0.95555 0.	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000004 0.000003 0.00003 0.000003 0.0003 0.00003 0	0.573967 0.573967 0.573967 0.573969 0.573963 0.573963 0.573960 0.573960 0.573960 0.573966 0.573966 0.573964 0.573959 0.573961 0.573961 0.573959 0.573964 0.573967 0.573964 0.573964 0.573967 0.5739	0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002	0.581500 0.581506 0.581499 0.581499 0.581490 0.581492 0.581472 0.581477 0.581477 0.581477 0.581477 0.581474 0.581474 0.581474 0.581474 0.581474 0.581474 0.581486 0.581488 0.581487 0.581487	
T0385F03 T0385F04 T0385F04 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0380F01 T0390F01 T0390F04 average* reproducibility (ppm) μ ¹ (Mo(In-Run/Nicr) (ppm) the line-by-line O correction T0374F03 T0374F03 T0374F04 T0374F04 T0374F04 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0385F03 T0385F03 T0385F03 T0385F04 T0385F05 T0385F06 T0385F06 T0385F06 T0385F06	0.883245 0.883253 0.883223 0.883223 0.883223 0.883223 0.883223 0.883221 0.883221 0.883221 0.883221 0.883223 0.883225 - - 98 0.883225 0.883225 0.883225 0.883225 0.883225 0.883225 0.883227 0.883225 0.883223 0.883226 0.883227 0.883226 0.883227 0.883228 0.883226 0.883227 0.883228 0.883228 0.88326 0.883228 0.88326 0.883228 0.88326 0.88326 0.88326 0.88327 0.883228 0.88326 0.88327 0.88326 0.88327 0.88326 0.88327 0.8937 0.8937 0.8937 0.8937 0.8937 0.8937 0.89	0.000006 0.00000000	0.552479 0.552476 0.552476 0.552479 0.552473 0.552473 0.552473 0.552479 0.552476 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552478 0.552477 0.552482 0.552477 0.552482 0.552477 0.552482 0.552477 0.552482 0.552477 0.552478 0.552477 0.552482 0.552477 0.552477 0.552482 0.552477 0.552479 0.5524	0.00003 0.00003 0.00002 0.00002 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000002 0.000003 0.000002 0.000002 0.000002 0.000003 0.000002 0.000002 0.000002 0.000003 0.000002 0.000002 0.000002 0.000003 0.000002 0.000002 0.000003 0.000002 0.000002 0.000002 0.000003 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.000003 0.000002 0.000002 0.000002 0.000003 0.000002 0.000002 0.000003 0.000002 0.000002 0.000003 0.000002 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.00003 0.00002 0.00002 0.00003 0.00002 0.00002 0.00003 0.00002 0.00002 0.00003 0.00003 0.00002 0.00003 0.00003 0.00002 0.00003 0.00002 0.0000 0.00003 0.00002 0.00003 0.00002 0.0000 0.0000 0.0000 0.0000 0.000 0.00	0.953258 0.953260 0.953262 0.953262 0.953253 0.953256 0.953256 0.953258 0.953258 0.953258 0.953258 0.953257 54 0.953252 0.953252 0.953252 0.953252 0.953252 0.953252 0.953252 0.953254 0.953256 0.953256 0.953264 0.953265 0.953258	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000003 0.00003 0.0	0.573967 0.573967 0.573967 0.573969 0.573969 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573961 0.573950 0.573961 0.573951 0.573951 0.573951 0.573954 0.573954 0.573954 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573968 0.573964 0.573968 0.573976 0.573968 0.573968 0.573968 0.573976 0.573968 0.5739	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001	0.581500 0.581506 0.581502 0.581502 0.581502 0.581478 0.581477 0.581478 0.581478 0.581478 0.581477 0.581474 0.581477 0.581474 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581480 0.581477 0.581480 0.581478 0.581478 0.581478 0.581478 0.581478 0.581479 0.581480 0.581478 0.581480 0.581478 0.581480 0.581478 0.581480 0.581478 0.581480 0.581478 0.581480 0.581478 0.581488 0.581479 0.581488 0.581479 0.581488 0.581479 0.581488 0.581479 0.581488 0.581479 0.5814888 0.5814888 0.581488 0.581488 0.581488 0.581488 0.581488 0	
T0385F03 T0385F04 T0385F06 T0385F07 T0385F07 T0385F07 T0385F07 T0390F01 T0390F01 T0390F03 T0390F04 average* reproducibility (pm] p ⁴ Mo(tn-Run/Nicr) (pm] the line-by-line O correction T0374F01 T0374F03 T0374F04 T0374F03 T0374F04 T0374F04 T0374F03 T0374F04 T0374F04 T0374F03 T0374F04 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F04 T0374F05 T0374F04 T0374F05 T0385F03 T0385F03 T0385F04 T0385F04 T0385F05 T0385F05 T0385F05 T0385F07	0.883245 0.883253 0.883223 0.883223 0.883223 0.883223 0.883191 0.883221 0.883221 0.883221 0.883224 0.883225 98 0.883224 0.883225 0.883224 0.883225 0.883240 0.883225 0.883240 0.883224 0.883224 0.883224 0.883224 0.883223 0.883224 0.883223 0.883223 0.883223 0.883223 0.883223 0.883223 0.883223 0.883223 0.883223 0.883223	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.00000000	0.552479 0.552479 0.552478 0.552479 0.552479 0.552473 0.552479 0.552479 0.552479 0.552479 0.552479 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552478 0.552477 0.552478 0.552478 0.552482 0.552482 0.552482 0.552484 0.552484 0.552484 0.552476 0.552478	0.00003 0.00003 0.000002 0.000002 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000003 0.000002 0.000003 0.000003 0.000003 0.000002 0.000003 0.000003 0.000003 0.000002 0.000003 0.000003 0.000002 0.000003 0.000003 0.000002 0.000003 0.000003 0.000003 0.000002 0.000003 0.00003 0.000003 0.000	0.953258 0.953260 0.953262 0.953262 0.953253 0.953256 0.953256 0.953258 0.953258 0.953258 0.953257 	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000003 0.00003 0	0.573967 0.573967 0.573967 0.573967 0.573969 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573961 0.573961 0.573961 0.573961 0.573961 0.573961 0.573961 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573965 0.573964	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.000001 0.000002 0.000002 0.000002 0.000002 0.000001 0.000002 0.000002 0.000001 0.000002	0.581500 0.581506 0.581409 0.581409 0.581402 0.581402 0.581477 0.581478 0.581479 0.581479 0.581473 0.581473 0.581474 0.581477 0.581474 0.581477 0.581474 0.581474 0.581474 0.581478 0.581474 0.581478 0.581480 0.581470 0.581480 0.5	
T0385F03 T0385F04 T0385F06 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0390F02 T0390F02 T0390F02 T0390F03 T0390F04 average* reproducibility (ppm) µ' Mo(Ia-Run/Ner) (ppm) the line-by-line O correction T0374F02 T0374F03 T0374F03 T0374F03 T0374F04 T0374F03 T0374F03 T0374F04 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0385F03 T0385F03 T0385F03 T0385F05 T0385F05 T0385F08 T0385F08	0.883245 0.883225 0.883223 0.883223 0.883223 0.883223 0.883221 0.883221 0.883221 0.883221 0.883221 0.883221 0.883221 0.883225 - - 98 0.883225 0.883225 0.883225 0.883225 0.883225 0.883225 0.883224 0.883226 0.883227 0.883226 0.883227 0.883226 0.883227 0.883226 0.883227 0.883227 0.883226 0.883227 0.883226 0.883227 0.883227 0.883226 0.883227 0.883227 0.883227 0.883227 0.883227 0.883227 0.883227 0.883227 0.883227 0.883228 0.883228 0.883228 0.883228 0.883228 0.883228 0.883228 0.883228 0.883228 0.883228 0.883228 0.883228 0.883223 0.88323 0.8832	0.000006 0.000006 0.000006 0.000005 0.000005 0.000005 0.000006 0.00000000	0.552479 0.552479 0.552478 0.552479 0.552473 0.552473 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552478 0.552477 0.552478 0.552478 0.552478 0.552478 0.552478 0.552478 0.552478 0.552478 0.552478 0.552477 0.552478 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552477 0.552478 0.552477 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552477 0.552478 0.552477 0.5524	0.00003 0.00003 0.00002 0.00002 0.00003 0.00003 0.00003 0.00003 0.00003 0.000003 15 0.000003 0.000003 0.000003 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002	0.953258 0.953260 0.953262 0.953262 0.953253 0.953255 0.953256 0.953256 0.953256 0.953258 0.953257 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953255 0.953266 0.953266 0.953266 0.953266 0.953258 0.953261 0.953258 0.953251 0.953251 0.953252 0.953251 0.953252 0.953251 0.953252 0.953255 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0.	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000004 0.000004 0.000003 0.000004 0.000003 0.00003 0.000003 0.00	0.573967 0.573967 0.573967 0.573965 0.573963 0.573963 0.573960 0.573960 0.573960 0.573960 0.573966 0.573964 0.573961 0.573961 0.573961 0.573961 0.573964 0.573967 0.573967 0.573967 0.573967 0.573967 0.573967 0.573967 0.573967 0.573968 0.5739	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.000002 0.000002 0.000002 0.0000002 0.0000002 0.0000002 0.0000002 0.0000002 0.0000002 0.0000002 0.0000001 0.0000002 0.0000001 0.0000002 0.0000001 0.0000002 0.0000001 0.0000002	0.581500 0.581506 0.581499 0.581499 0.581492 0.581492 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581478 0.581477 0.581478 0.581478 0.581478 0.581480 0.581487 0.581487 0.581487 0.581487 0.581487 0.581487 0.581487 0.581487 0.581487	
T0385F03 T0385F04 T0385F04 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0390F01 T0390F01 T0390F03 T0390F04 average* reproducibility (pmp) μ ⁴ (Mo(In-Run/Nicr) (pmp) the line-by-line O correction T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0374F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F04 T0385F05 T0385F05 T0385F06 T0385F07 T0385F09 T0385F09 T0385F09	0.885245 0.885253 0.885223 0.885223 0.885223 0.885223 0.885223 0.885221 0.885221 0.885221 0.885221 0.885221 0.885225 - - 98 0.885225 0.885225 0.885225 0.885225 0.885225 0.885225 0.885226 0.885227 0.885226 0.885227 0.885226 0.885227 0.885226 0.885227 0.885226 0.885227 0.885226 0.885227 0.885226 0.885227 0.885226 0.885226 0.885226 0.885226 0.885226 0.885226 0.885226 0.8856 0.8856	0.000006 0.00000000	0.552479 0.552476 0.552478 0.552478 0.552479 0.552473 0.552473 0.552479 0.552479 0.552476 0.552477 0.552477 0.552477 0.552477 0.552478 0.552477 0.552478 0.552477 0.552484 0.552477 0.552484 0.552478 0.552478 0.552478	0.00003 0.00003 0.00002 0.00002 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.000003 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.00003 0.0003 0.0003 0.00003 0.00003 0.000	0.953288 0.953260 0.953262 0.953262 0.953253 0.953253 0.953253 0.953258 0.953258 0.953258 0.953258 0.953258 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953255 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0	0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000004 0.000003 0.00003 0	0.573967 0.573967 0.573967 0.573969 0.573969 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573961 0.573951 0.573961 0.573951 0.573964 0.573956 0.573964 0.573965 0.573964 0.573965 0.573964 0.573965 0.573964 0.573965 0.5739	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.0000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.0000001 0.0000001 0.0000001 0.0000001 0.0000001 0.0000002 0.0000001 0.0000001 0.0000001 0.0000001 0.0000001 0.0000001 0.000001 0.000001 0.000001 0.0000001 0.000001 0.0000002 0.0000001 0.0000001 0.0000001 0.0000001 0.0000002 0.0000001 0.00000001 0.0000001 0.00000000	0.581500 0.581506 0.581502 0.581502 0.581479 0.581477 0.581477 0.581477 0.581474 0.581477 0.581474 0.581474 0.581477 0.581474 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581477 0.581478 0.581477 0.581478 0.581477 0.581480 0.581478 0.581488 0.5	
T0385F03 T0385F04 T0385F06 T0385F07 T0385F07 T0385F07 T0385F07 T0380F02 T0390F02 T0390F03 T0390F04 average* reproducibility (pm] p' Mo(In-Run/Nicy) (pm) the line-by-line O correction T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F03 T0374F03 T0385F03 T	0.883245 0.883253 0.883223 0.883223 0.883223 0.883223 0.883191 0.883221 0.883221 0.883224 0.883225 	0.000006 0.000006 0.000006 0.000006 0.000005 0.000005 0.000006 0.00000000	0.552479 0.552479 0.552478 0.552479 0.552473 0.552473 0.552473 0.552479 0.552479 0.552479 0.552476 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552478 0.552477 0.552478 0.552478 0.552482 0.552482 0.552484 0.552476 0.552476 0.552476 0.552476 0.552476 0.552476 0.552476 0.552476 0.552476 0.552478 0.552476 0.552477 0.552480 0.552476 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552478 0.552477 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.552477 0.552478 0.5	0.00003 0.00003 0.000002 0.000002 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.00003 0.000003 0.0003 0.000	0.953258 0.953260 0.953262 0.953262 0.953253 0.953256 0.953256 0.953258 0.953258 0.953258 0.953258 0.953257 	0.000003 0.000004 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000003 0.000004 0.000003 0.00003 0	0.573967 0.573967 0.573967 0.573967 0.573969 0.573963 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573960 0.573961 0.573961 0.573961 0.573961 0.573961 0.573961 0.573961 0.573963 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573964 0.573965 0.573964 0.573964 0.573966 0.573966 0.573966 0.573966 0.573967 0.573968	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.000002 0.000002 0.000001 0.000002 0.000002 0.000002 0.000001 0.000002 0.000001 0.000002 0.000001 0.000002	0.581500 0.581506 0.581409 0.581409 0.581402 0.581402 0.581477 0.581479 0.581479 0.581479 0.581473 0.581473 0.581474 0.581474 0.581474 0.581474 0.581474 0.581475 0.581475 0.581475 0.581475 0.581474 0.581480 0.581480 0.581480 0.581480 0.581480 0.581483 0.581483 0.581483 0.581483 0.581484 0.5	
T0385F03 T0385F04 T0385F04 T0385F07 T0385F07 T0385F07 T0385F07 T0385F07 T0380F02 T0390F02 T0390F03 T0390F03 T0390F04 average* reproducibility (ppm] µ* (hot(=.kuw/xe) (ppm] µ* (hot(=.kuw/xe) (pm) the line-y-line O correction T0374F03 T0374F03 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F04 T0374F03 T0374F03 T0374F03 T0374F04 T0374F03 T0374F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03 T0385F03 T03855F03 T0385F03 T0374F03 T0375F03 T	0.883245 0.883225 0.883223 0.883223 0.883223 0.883223 0.883221 0.883221 0.883221 0.883221 0.883225 - - 98 0.883225 0.883225 0.883225 0.883225 0.883225 0.883223 0.883224 0.88324 0.88325 0.88325 0.88325 0.88325 0.88325 0.88325 0.88324 0.88325 0.88324 0.88325 0.88324 0.88325 0.88325 0.88325 0.88325 0.88325 0.88326 0.88324 0.88325 0.88324 0.88325 0.8835 0.8855 0.8855 0.8855 0.8855 0.8855 0.8855 0.8855 0.8855	0.000006 0.000006 0.000006 0.000005 0.000005 0.000005 0.000006 0.00000000	0.552479 0.552479 0.552478 0.552479 0.552473 0.552481 0.552476 0.552477 0.552477 0.552477 0.552477 0.552477 0.552477 0.552478 0.552477 0.552478 0.552478 0.552477 0.552478 0.552477 0.5524	0.00003 0.00003 0.00002 0.00002 0.00003 0.00002 0.00003 0.00003 0.00003 0.00003 0.000003 0.000003 0.000003 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.000002 0.000003 0.00003 0.0003 0.	0.953258 0.953260 0.953262 0.953262 0.953253 0.953255 0.953256 0.953256 0.953256 0.953258 0.953257 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953254 0.953258 0.953266 0.953266 0.953266 0.953265 0.953258 0.953258 0.953258 0.953251 0.953252 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0.95325 0.9	0.00003 0.00004 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00003 0.00004 0.00003 0.00004 0.00004 0.00004 0.00003 0.00004 0.00003 0.0000	0.573967 0.573967 0.573967 0.573965 0.573963 0.573963 0.573963 0.573960 0.573966 0.573966 0.573966 0.573966 0.573966 0.573961 0.573961 0.573961 0.573964 0.573959 0.573964 0.573966 0.573966 0.573966	0.000002 0.000002 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000001 0.000002 0.000002 0.000002 0.0000002 0.0000001 0.0000001 0.0000001 0.0000002 0.0000001 0.0000002 0.0000001 0.0000001 0.0000001 0.0000001 0.0000002 0.0000001 0.0000001 0.0000001 0.0000001 0.0000002 0.0000001 0.000001 0.0000002 0.000002 0.0000002 0.000002 0.000002 0.0000002 0.0000000 0.0000000 0.0000000 0.0000000	0.581500 0.851506 0.581506 0.581409 0.581402 0.581402 0.581477 0.581477 0.581473 0.581473 0.581474 0.581474 0.581474 0.581474 0.581474 0.581474 0.581475 0.581480 0.581480 0.581485 0.581483 0.581483 0.581483 0.581483 0.581484 0.581483 0.581483 0.581484 0.581483 0.581484 0.581485 0.581485 0.581465 0.58145 0.58	

> μⁱMo(Line-by-Line/Nier) [ppm] 99 -11 * Erros of each measurement were 2 standard errors (2SE), but those of averages were 2 standard deviations (2SD).

Journal of Analytical Atomic Spectrometry

		Table 4. Molybdenum isotopic ratios of various kinds of samples											
	п	μ ⁹² Mo ^a	error ^b	µ ⁹⁴ Mo ^a	error ^b	μ ⁹⁵ Mo ^a	error ^b	µ ⁹⁷ Mo ^a	error ^b	μ^{100} Mo ^a	error ^b		
standard (the line-by-line O)				· · · · · · · · · · · · · · · · · · ·									
Junuary-March, 2015	21	-	± 47	-	± 16	-	± 10	-	± 13	-	± 33		
standard Mo with 2% interference													
Mo + Zr	3	45	± 47	19	±16	6	± 10	-8	± 13	-2	± 33		
Mo + Ru	3	31	± 47	14	± 16	2	± 10	-5	± 13	5	± 33		
artificial-spiked samples													
Mix-A (small)	4	140	± 47	8	± 16	5	± 10	47	± 13	66	± 33		
Mix-B (large)	4	258	± 65	10	± 24	3	± 11	81	± 15	125	± 33		
predict Mix-A from Mix-B		130	± 33	5	± 12	1	± 6	40	± 8	63	± 17		
terrestrial rocks													
JA-3 (andesite)	1	17	± 36	2	± 13	1	± 8.7	3	± 0.9	13	± 15		
JB-3 (basalt)	1	17	± 31	6	± 14	1	± 9.9	0	± 6.7	15	± 4.8		
JR-2 (rhyolite)	1	12	± 36	4	± 13	2	± 8.7	0	± 0.9	-1	± 15		
JR-2 duplicate	1	-6	± 31	0	± 14	0	± 9.9	-9	± 6.7	-8	± 4.8		
terrestrial average ^c	4	10	± 22	3	± 5.1	1	± 1.5	-1	± 10	5	± 22		
iron meteorites													
Tambo Quemade (IIIAB iron)	3	99	± 47	93	± 21	35	± 10	26	± 13	30	± 33		
Henbury (IIIAB iron)	2	128	± 36	103	± 16	36	± 11	23	± 13	47	± 46		
Albion (IVA iron)	2	88	± 36	74	± 16	31	± 11	12	± 13	16	± 46		

Journal of Analytical Atomic Spectrometry Accepted Manuscript

^a μ iMo = [(iMo/96Mo)sample/(iMo/96Mo)std - 1] × 10⁶: normalized to ⁹⁸Mo/⁹⁶Mo = 1.413173 by using the exponential law. The Mo isotopic ratios of the standard are the ones measured at measurement day.

^b For samples measured more than twice, errors are the standard deviation from repeated analysis of standard solution or samples, whichever is larger. For samples measured once, errors are estimated by the standard deviation (2SD) from repeated analysis of the standard solution obtained at measurement day, which is larger than those of the 2SE.

^c Errors were the standard deviations of terrestrial rocks (n=4: 2SD).