

RSC Advances



This is an *Accepted Manuscript*, which has been through the Royal Society of Chemistry peer review process and has been accepted for publication.

Accepted Manuscripts are published online shortly after acceptance, before technical editing, formatting and proof reading. Using this free service, authors can make their results available to the community, in citable form, before we publish the edited article. This *Accepted Manuscript* will be replaced by the edited, formatted and paginated article as soon as this is available.

You can find more information about *Accepted Manuscripts* in the [Information for Authors](#).

Please note that technical editing may introduce minor changes to the text and/or graphics, which may alter content. The journal's standard [Terms & Conditions](#) and the [Ethical guidelines](#) still apply. In no event shall the Royal Society of Chemistry be held responsible for any errors or omissions in this *Accepted Manuscript* or any consequences arising from the use of any information it contains.



Journal Name

ARTICLE

Disposable (bio)chemical integrated optical waveguide sensors implemented on roll-to-roll produced platforms

Sanna Aikio,^{*a} Martin Zeilinger,^b Jussi Hiltunen,^a Leena Hakalahti,^a Johanna Hiitola-Keinänen,^a Marianne Hiltunen,^a Ville Kontturi,^c Samuli Siitonen,^c Jarkko Puustinen,^d Peter Lieberzeit,^b and Pentti Karioja^a

Received 00th January 20xx,
Accepted 00th January 20xx

DOI: 10.1039/x0xx00000x

www.rsc.org/

To enable wide spread dissemination of sensors in cost-critical applications and resource poor settings, methods to implement sensor chips using low-cost materials and mass-manufacturing methods are developed. In this paper we demonstrate that disposable polymeric integrated Young interferometer (YI) sensor chips, implemented on roll-to-roll mass-manufactured waveguides, are applicable for analyte specific sensing of small molecules and for multi-analyte detection of biomolecules. For the chemical sensing of small molecules, sensor chip was functionalised with a molecularly imprinted polymer (MIP). We demonstrate that MIP receptor layer is compatible with polymer-based evanescent wave sensor for direct refractive index sensing. For multi-analyte detection of biomolecules, antibody-based receptor layers were patterned by inkjet printing onto the sensor surface demonstrating the applicability of the method with integrated YI chips. Demonstration of the analyte specific chemical- and biosensing with disposable polymeric YI sensor chips opens new possibilities to implement low-cost (bio)chemical sensors.

Introduction

Research on micro total analysis and optofluidic systems is advancing rapidly developing sensor technologies for a wide range of applications including point-of-care diagnostics, environmental monitoring and food safety.^{1,2} To enable wide spread dissemination of the sensors, also in resource poor settings, there is a need for sensor chips that are sensitive, robust, and reliable, and that can be implemented using low-cost fabrication methods and materials.

Waveguide based integrated interferometric sensors, such as integrated Young interferometers (YIs), show potential for low-cost sensor applications.³ Integrated YI sensors utilize evanescent wave field of the guided optical wave to directly sense the refractive index (RI) changes at or within few hundred nanometres of the waveguide surface. The label-free sensing method enables simpler sample handling, and reduced complexity and cost of the sensor chip.³ YI sensor chips are reported to detect RI differences 10^{-5} - 10^{-8} refractive index units (RIU) showing some of the lowest limit of detection values among integrated interferometric sensors.³⁻⁶

The fabrication of YI sensor chips has relied on batch-based

manufacturing methods³, for example micromachining of silicon wafers, using inorganic^{4,5,7-10} or low-cost polymeric^{6,11-13} materials. Recently, the authors have demonstrated that single-mode waveguides for polymeric integrated YI sensor chips can be fabricated on flexible foils with length of hundreds of meters by using continuous ultra-high volume roll-to-roll (R2R) manufacturing methods making the chips inherently disposable.¹⁴ Quality of the waveguides was assessed by measuring the attenuation properties of hundreds of samples showing that waveguides with repeatable performance were manufactured. It was also demonstrated that the unfunctionalized integrated YI sensor chips were capable to detect ambient unspecific RI changes in the level of 10^{-6} RIU and quantify them in the level of 10^{-5} RIU by using aqueous glucose solutions, showing thus rather comparable performance than reported earlier for polymeric YI sensors.^{6,13}

For analyte specific sensing, YI sensor chips need to be functionalized with a receptor layer that binds analyte molecules selectively to the sensor surface. When analyte molecule binds to a receptor it replaces sample matrix that typically has lower RI, and induces a RI change. The magnitude of the RI change depends on the RIs of the chosen analyte molecule and sample matrix. Of biological receptors, antibodies have been used to functionalize inorganic and polymeric YI sensors for detection of proteins and viruses including demonstration of multi-analyte sensing.^{6,8,9,15,16} Artificial receptor layers, such as molecularly imprinted polymers (MIPs), provide another route for sensor functionalization. MIP receptor layers may be advantageous for low-cost sensors due to their potential to be low-cost,

^a VTT Technical Research Centre of Finland, Kaitoväylä 1, FI-90590, Oulu, Finland

^b University of Vienna, Faculty for Chemistry, Department of Analytical Chemistry, Währinger Straße 38, 1010 Vienna, Austria

^c Nanocomp Oy Ltd, Ensolantie 6, FI-80710 Lehmo, Finland

^d University of Oulu, Faculty of Information Technology and Electrical Engineering, Department of Electrical Engineering, Microelectronics and Materials Physics Laboratories, Pentti Kaiteran katu 1, FI-90014 Oulu, Finland

*Corresponding author: sanna.aikio@vtt.fi

robust, repeatable, and mass-manufacturable.¹⁷ MIP receptor layers have been applied with silicon-based interferometric sensor chips to detect vapor phase trinitrotoluene and proteins.^{18,19} In this work we study the applicability of MIP receptor layer with polymeric YI sensor chips.

The antibody receptor layers of integrated YI sensors are applied by either incubating antibody solution on the chip surface⁹, or by flowing the antibody solution through the flow-cell assembled on top of the sensor chip enabling the binding of the antibodies onto the sensor surface^{6,8,15,16}. The MIP receptor layers are formed by either dip coating the YI sensor chip¹⁸ or injecting the polymer solution onto the sensor surface, and performing the polymerization under glass plate resulting in a thin receptor layer¹⁹. Patterned functionalisation by inkjet printing of antibodies has been demonstrated for multi-analyte functionalisation of silicon-based ring resonator sensors²⁰, however that method has not been utilized to functionalise YI sensor chips. Functionalisation by inkjet printing is especially interesting for low-cost sensors since, as an additive method, it minimizes the consumption of potentially expensive receptor materials, and is also considered to be compatible with mass-production methods.²¹

In this paper, we demonstrate the applicability of the disposable polymeric multi-analyte YI sensor chips having R2R fabricated waveguides for analyte specific chemical- and biochemical sensing. Due to strict surface quality requirements of optical waveguides, it is not self-evident that receptor layers can be applied onto polymeric waveguides without deteriorating their capability to guide light and act as a sensor, for example due to increased scattering. We use melamine as a model analyte to demonstrate the chemical sensing of small molecules and the applicability of the MIP receptor layer with a polymeric evanescent wave sensor chip for direct RI sensing. The multi-analyte sensor design and patterning of the MIP and the reference non-imprinted polymer (NIP) layer onto a single sensor chip enables the simultaneous measurement of the analyte specific binding into the MIP layer and the non-specific sample matrix effects. The multi-analyte detection of biomolecules is demonstrated by using C-reactive protein (CRP) and human chorionic gonadotropin (hCG) as model analytes. We also demonstrate the applicability of the inkjet printed antibody-based receptor and reference layers with integrated YI sensor chips for multi-analyte detection. The chosen functionalisation schema enables real-time detection of the specific binding of analytes and suppression of non-specific sample matrix effects.

Sensor chip embodiments and sensing principle

Illustration of the sensor chip is shown in Fig. 1(a). Sensor chip has an input waveguide that branches into four waveguides forming two parallel Young interferometers, named YI1 and YI2, integrated into one sensor chip. Both of the interferometers contain a reference and a measurement waveguide. The separation between the reference and measurement waveguides is increased at the sensing area enabling patterning of the measurement windows and

receptor layers with loose positional tolerances. Measurement window is an opening in the passivating overcladding layer of the waveguides enabling the interaction of the sample with the evanescent wave field of the light propagating in the underlying waveguide.

Illustration of the cross-section of the sensor chip structure for the chemical sensing within the sensing area is shown in Fig. 1(b). The reference waveguides of both YIs were passivated by the overcladding layer and were thus not exposed to sample solutions. The measurement windows were formed by leaving a length of 10 mm of the measurement waveguides without the overcladding layer. The MIP receptor layer was applied into the measurement window of YI1, and the reference NIP layer into the measurement window of YI2.

Cross-section of the sensor chip for multi-analyte biosensing is illustrated in Fig. 1(c). In the sensing area, all the waveguides were left without the overcladding layer and they were sharing a single measurement window. Thus, all the waveguides were similarly exposed to sample solutions. Inkjet printing was used to pattern analyte specific antibody layers, i.e. anti-human CRP antibodies and anti-hCG antibodies, onto their respective measurement waveguides and the reference antibody layer, i.e. mouse immunoglobulin G (IgG), onto the reference waveguides. It is worth noticing that with this sensor embodiment the sensing length of the sensor is defined by the inkjet printed receptor layers, not by the measurement window patterned in the overcladding layer. The overcladding layer was applied onto the waveguides outside the sensing area as required by the single mode operation of the waveguides.¹⁴

In the measurements, laser light was end-fire coupled into the input waveguide and was divided into the measurement and reference waveguides. The four light beams emerging from the outcoupling end of the chip diverge, overlap and interfere forming a fringed interferogram that was imaged onto the camera by using a microscope objective as illustrated in Fig. 2(a). Since the imaging was done close to the chip outcoupling end, the beams emanating from the measurement and reference waveguide of YI1 interfered only with each other. Similarly, the beams of YI2 interfered mutually. This resulted in two separate 2-beam interferograms that were imaged simultaneously onto a single camera detector as shown in Fig. 2(b) and 2(c). Another approach to implement multi-analyte integrated YI sensors is to use sensor layouts that have multiple measurement waveguides and a single reference waveguide with unique waveguide pair separations at the outcoupling end, and letting all the emanating beams interfere to form a multi-beam interferogram.^{16,22,23}

Sensing method is based on the analysis of the sample induced changes in the interferogram fringe positions.^{22,23} When RI on top of the waveguide is changed by the sample, the effective RI of the waveguide is changed due to the interaction of the sample with the evanescent wave field of the light propagating in the waveguide. The binding of the molecules changes the RI of the layer on top of the waveguides since water is replaced by a material having typically a higher RI. In the designed sensor layouts, sample

induced effective RI change differs between the measurement and reference waveguides. This changes their mutual optical path length difference that again shifts the interferogram fringes. The sample induced phase change of the fringes, $\Delta\phi$, is²³

$$\Delta\phi = 2\pi \frac{\ell(\Delta n_{\text{eff,m}} - \Delta n_{\text{eff,r}})}{\lambda}, \quad (1)$$

where ℓ is the sensing length of the sensor, $\Delta n_{\text{eff,m}}$ and $\Delta n_{\text{eff,r}}$ are the sample induced effective refractive index changes of the measurement and reference waveguides, respectively, and λ the wavelength of the light. As shown by Eq. 1, the induced phase change is inversely proportional to the wavelength that also affects to the probing depth of the evanescent wave field and to the effective refractive index changes of the waveguides. Interferograms captured during the experiments were analyzed by using two dimensional fast Fourier transform (FFT) yielding the phases of the fringes and enabling the calculation of the sensor responses as phase changes.

Materials and methods

Fabrication of sensor chips

Illustration of the cross-section of the inverted ridge waveguide structure used in the sensor chips is shown in Fig. 3(a). The waveguide fabrication is shortly outlined in this paper, and detailed description is given in the earlier paper by the authors.¹⁴

The fabrication of the waveguides was done in two phases using two R2R units. In the first phase, the waveguide grooves were patterned into the UV-curable acrylate based undercladding material (Nalax2, Nanocomp) using UV-imprinting method, and in the second phase the grooves were filled with the epoxy-based waveguide core material (Epocore, Micro resist technology). In the first phase, the undercladding material was gravure coated onto polycarbonate foil. The waveguide grooves were patterned by pressing a nickel stamp, having a ridge structure, into the uncured undercladding material and curing the material by applying UV-light through the transparent foil. In the second phase, gravure printing method was used to fill the waveguide grooves with the UV-curable waveguide core material followed by curing with UV-light. A roll of R2R produced sensor platforms is shown in Fig. 3(b).

The sensor platforms were cut from the roll for post-processing to sensor chips as shown in Fig. 3(c). The in- and outcoupling facets of the waveguides were formed by cleaving. The overcladding layer (OP-4-20632, Dymax) was patterned onto the chips as required by the different sensor embodiments.

Functionalisation with MIP layer

The MIP- and NIP-layers were applied into the measurement window of Y11 and Y12, respectively, by spin-coating the

respective oligomer solutions at 3000 rpm using custom-built spin coater. Oligomers leading to MIP were prepared by dissolving 17 mg dihydroxyethene bisacrylamide (DHEBA; cross-linker), 8 mg methacrylic acid (MAA; functional monomer) and 1.5 mg melamine (template) in 1 ml de-ionized water. Synthesis of NIP followed the same experimental protocol but for adding melamine. This solution was then heated to 55°C followed by adding 5 mg sodium peroxodisulfate (PODS; radical initiator). Then, polymerization took place until the solution turned slightly turbid. The coated length of the waveguides was 10 mm defined by an opening in the overcladding layer. Prior to interferometric experiments, the melamine template molecules were removed from the MIP layer by immersing the sensor chip into water and shaking it overnight.

Functionalisation with antibodies

For multi-analyte biosensing, antibodies for hCG and human-CRP were inkjet patterned onto the measurement waveguides of Y11 and Y12, respectively, and mouse IgG, onto the reference waveguides. Bio-inks containing anti-human CRP antibodies (monoclonal antibody (Mab) 6404, Medix Biochemica), anti-hCG antibodies (Mab 5006, Medix Biochemica) or mouse IgG (ChromPure, Jackson ImmunoResearch Laboratories) were prepared. Each bio-ink contained 1.0 mg/ml of the respective antibody prepared in Na₂CO₃ buffer solution (50 mM, pH 9.6) containing 3% of isopropanol and 0.1% of Tween-20. Inks were filtered through a nylon membrane filter with 0.2 µm pore size to avoid clogging of the inkjet printhead nozzles. Before inkjet printing, the chips were oxygen plasma treated 1 min at 300 W to make surfaces more hydrophilic to improve the printing quality. The printing was done using a piezoelectric inkjet printer (Dimatix DMP-2800, Fujifilm Dimatix). Size of the printed rectangular patterns, shown in Fig. 7(b), was 0.6×10 mm².

After inkjet printing, the sensor surface was blocked to reduce unspecific binding during the experiments by applying 500 µl of 3% bovine serum albumin (BSA) (Sigma-Aldrich) in phosphate buffered saline (PBS) buffer (12 mM, pH 7.4) on top of the sensor chips. Chips were shaken for 10 min, after which they were aspirated. The unbound BSA was washed with 500 µl PBS with Tween-20 (PBST) applied on the sensor chips. Chips were shaken for 5 min, after which they were aspirated. During these steps the salts in the bio-inks that dry onto the chip surface during the inkjet printing were washed away.

Sample solutions and their actuation

For chemical sensor experiments, aqueous melamine solutions were prepared by dissolving melamine (Alfa Aesar) in ultrapure water at the following concentrations: 0.1, 0.25, 0.5 and 1.0 g/l. Melamine concentrations of the sample solutions were chosen to reflect the levels found in contaminated milk powders reported to contain melamine up to 2563 mg/kg.²⁴

For biosensor experiments, sample solutions containing either 2.0 µg/ml of human-CRP antigen (BBI Solutions) or 0.9 µg/ml (corresponding 100 mIU/ml) of hCG antigen (Scripps

Laboratories) were prepared in 1% bovine serum albumin (BSA) (Sigma-Aldrich) in phosphate buffered saline (PBS) buffer (12 mM, pH 7.4) as a sample matrix. To study the sensing of ambient RI changes, a sample solution was prepared containing 0.5 weight % of D-glucose (Sigma-Aldrich) in 1% BSA in PBS buffer. The BSA content of the sample matrix increases its complexity compared to mere buffered saline solution. The concentrations of hCG and CRP in sample solutions were chosen to be physiologically relevant. For hCG, pregnancy tests typically have sensitivities ~ 25 IU/l.²⁵ However, during pregnancy the concentration of hCG in serum can be $>300\,000$ IU/l and in urine >2500 IU/l although the levels vary greatly.²⁵ For CRP, the normal concentration in human serum is <8 $\mu\text{g/ml}$ but the levels can even 1000 fold in inflammation.²⁶ CRP concentration can also be used to estimate the risk of cardiovascular disease where levels <1 $\mu\text{g/ml}$ indicate low, 1-3 $\mu\text{g/ml}$ moderate, and >3 $\mu\text{g/ml}$ high risk.²⁷

Flow cell was assembled on top of the interferometer chips at least a day before the measurements, and it was filled with the sample matrix solution. The inlet and the outlet of the flow cell were connected to a sample vial and to a syringe pump (Nexus 3000, Chemyx), respectively, by a tube. Sample and flushing solutions were pipetted into the sample vial sequentially. The syringe pump was driven in withdraw mode at a flow rate of 100 $\mu\text{l/min}$ during the chemical sensing experiments and at a flow rate of 20 $\mu\text{l/min}$ during biosensing experiments. In the chemical- and biosensing experiments, the sensor response time was limited by the time it took for the sample to replace the preceding fluid within the flow cell, and by the mass transportation of the analyte molecules from the sample volume to the sensor surface. The response time was about 2 min in chemical sensing experiments, and about 5 min in biosensing experiments. The fluid replacement times were different for the chemical- and biosensing experiments due to different flow rates and different flow cell volumes.

Fluorescent staining

For the fluorescent staining of CRP analytes, 13.2 $\mu\text{g/ml}$ of anti-CRP antibody (Mab 6405, Medix Biochemica) labelled with Alexa Fluor 546 (Thermo Fisher Scientific) fluorescent label was prepared in 1% BSA in 12 mM PBS buffer (pH 7.4). For the staining of hCG analytes, 6.6 $\mu\text{g/ml}$ of anti-hCG antibody (Mab 5014, Medix Biochemica) labelled with Alexa Fluor 546 fluorescent label was prepared in 1% BSA in 12 mM PBS buffer (pH 7.4). For IgG staining, 4.0 $\mu\text{g/ml}$ of goat anti-mouse IgG antibodies (Thermo Fisher Scientific) labelled with Alexa Fluor 647 (Thermo Fisher Scientific) fluorescent label was prepared in 1% BSA in 12 mM PBS buffer (pH 7.4).

Fluorescent staining was done by using the following procedure: 500 μl of staining solution was pipetted onto the sensor surface, and the chip was placed onto a shaker for 10 minutes. The staining solution was aspirated and 400 μl of PBST solution was pipetted onto the chip to wash the unbound labelled antibodies. The chip was placed onto a shaker for 10 minutes, and PBST solution was aspirated.

Optical setup and stability

Light from a laser source (QFBGLD-980-5, QPhotonics) emitting at the wavelength of 975 nm was end-fire coupled into the input waveguide of a sensor chip by using a tapered polarization maintaining fibre (spot size of ~ 2.8 μm). The chosen wavelength enabled the single-mode operation of the waveguides¹⁴ as well as the use of a silicon-based camera. An optical isolator was used to prevent the disturbing reflections back to the laser. The sensor experiments were conducted using TM polarized light reported to have a higher sensitivity than TE polarization.⁶

Interferograms were imaged onto a camera detector (UI-3240CP-NIR-GL, IDS Imaging Development Systems) using a 40 \times microscope objective. The imaging was done at the distance of ~ 200 μm from the out-coupling end of the chip. The camera was located ~ 17 cm from the chip. The interval between the captured interferograms was 2 s during the chemical sensing experiments and 10 s during biosensing experiments.

Stability of the sensing system was studied by measuring phase changes while the sample matrix was flowing through the flow cell. Measured phase change curves for MIP and antibody functionalized chips are shown in Fig. 4 as well as lines fitted to the data points showing the trend of the system drift. In the curves, the noise of the system is seen as the small scale variation of the phase changes in the range of milliradians. The larger fluctuations and the overall drift of the phase change curves were attributed to the mechanical and thermal instabilities of the sensor chip and the experimental setup. The phase change fluctuations were less than 0.1 radians being thus much smaller than the analyte induced phase changes that were in the range of radians. It can be also seen that in all of the cases the trend of the drift was below 0.01 rad/min that is less than responses induced by the analytes in the sensing experiments.

Results and discussion

Chemical sensor

To demonstrate the chemical sensing, a MIP-functionalized sensor chip was exposed to aqueous melamine solutions. Each concentration was measured three times by flowing 500 μl of the melamine solution through the flow cell followed by flushing with water until the end of the measurement leading to 5 min analyte exposure. Timing pattern of the solutions in the flow cell is indicated in Fig. 5.

Due to the patterned functionalization of the chip, the MIP functionalized interferometer was measuring the specific binding and the NIP coated interferometer the non-specific binding of melamine to the polymer layers. Both interferometers were also measuring the bulk effects of the sample solutions. The shift of the interference fringes during an experiment is shown in Fig. 6. The phase change curves of the MIP and NIP coated interferometers, i.e. $\Delta\phi_{\text{MIP}}$ and $\Delta\phi_{\text{NIP}}$, respectively, are shown in Fig. 5(a). The phase change curves were first analysed from the interferograms and were

subsequently baseline corrected based on the values between 0-1.1 min. The curves clearly show that the MIP coated interferometer undergoes significantly higher phase changes on exposure to melamine: they are exceeding those of the corresponding NIP by roughly a factor of four. This gives clear evidence that incorporation of the melamine into the cavities of the MIP leads to substantial measurable phase change.

The measured phase changes also clearly exceed the values that would be produced by the mere bulk refractive index differences of the melamine solutions to pure water. The phase change $\Delta\phi_{\text{bulk}}$ induced by the bulk RI difference of the sample solutions to pure water, Δn , was evaluated by using formula

$$\Delta\phi_{\text{bulk}} = S\Delta n = S \left[\underbrace{\frac{C}{\rho_m} n_m + \left(1 - \frac{C}{\rho_m}\right) n_w}_{\text{RI of melamine solution}} - n_w \right], \quad (2)$$

where S is the sensitivity of the sensor in radians/RIU, C the concentration of melamine in g/cm^3 , ρ_m the density of melamine in g/cm^3 , and n_m and n_w the RI values of melamine and water, respectively. The RI of melamine solution was evaluated using the volume fractions of solution components.²⁸ The ratio C/ρ_m gives the volume fraction of melamine in the solution. The volume fraction of water was calculated assuming that the volumes of the components are additive with the melamine concentrations used in the experiments. For the concentration of 1.0 g/l, $\Delta\phi_{\text{bulk}}$ was calculated to be 0.4 rad by using the RI value of 1.872 to melamine²⁹ and 1.333 to water²⁹ (RIs were measured at the wavelength of 589 nm and at the temperature of 20 °C), and the melamine density²⁹ 1.573 g/cm^3 (measured at 16 °C). The sensor sensitivity¹⁴ 1251 rad/RIU was obtained from the earlier paper by the authors defined for similarly manufactured non-functionalized YI sensor chips for bulk RI changes. The calculated value of $\Delta\phi_{\text{bulk}}$ is indicative of the magnitude of the phase change due to the uncertainty of the RI value of the melamine solution and the uncertainty of the sensor chip response to bulk RI changes through the MIP and NIP layers. Since the phase changes measured with the highest melamine concentration shown in Fig. 5(a) are roughly a magnitude larger than the calculated values of $\Delta\phi_{\text{bulk}}$, the phase changes are enhanced by interaction of the molecules with MIP and NIP layers.

The actual imprinting effect can be assessed by calculating the net difference of the phase change values, $\Delta\phi_{\text{diff}}$, between the MIP and NIP coated interferometers, i.e. $\Delta\phi_{\text{diff}} = \Delta\phi_{\text{MIP}} - \Delta\phi_{\text{NIP}}$, shown in Fig. 5(b). It is worth noticing that the calculation of the difference signal eliminates the phase changes related to internal changes at the chip that both YIs see similarly. These kinds of changes may be induced, for example, by temperature variations. However, the capability of the method to eliminate the chip internal changes was not separately assessed by experiments.

Responses were determined from $\Delta\phi_{\text{diff}}$ curves by 1) calculating the height of the phase change step by taking the

average of the values within 5.0-5.3 min, and 2) by determining the slope of the phase change curves within 2.0-2.8 min. The responses and the second order polynomials fitted to the data points are shown in Fig. 7(a) and 7(b). In the fitting, the limit of detection was not taken into account since it was not assessed by the experiments. As can be seen, the sensor responses in overall increase with increasing melamine concentration demonstrating the capability of MIP functionalized chip to chemical sensing. The slope-based method was capable to differentiate between the two lowest concentrations whereas the step height –based method could not differentiate between these two. The results show beginning saturation for melamine concentrations above 0.5 g/l, which is in line with previous MIP studies aiming at detecting small molecules.³⁰ The reason for such saturation behavior can be found in the recognition mechanisms of MIP: analyte incorporation is the result of thermodynamic equilibrium between analyte molecules bound in recognition sites of the MIP and those in solution. Both the amount of binding sites within a MIP layer and accessibility of diffusion pathways within the matrix are limited hence limiting the dynamic range of the sensors. The slope-based method enables faster signal quantification than the method based on the step height that is an advantage for rapid diagnostics. The slight decrease in the sensor responses seen in the $\Delta\phi_{\text{MIP}}$ and $\Delta\phi_{\text{NIP}}$ curves of Fig. 5(a) with time can most probably be appointed loss of a few percent of the film as well as to minimal irreversible binding of melamine by the films. However, this does not affect the difference signal and hence can be neglected.

Multi-analyte biosensor

To demonstrate the multi-analyte biosensing, three antibody-functionalized chips were sequentially exposed to 400 μl of hCG and CRP sample solutions separated by a flushing step with 300 μl of the sample matrix leading to 20 min analyte exposure followed by 15 min sample matrix exposure. The timing of the solutions in the flow cell is indicated in Fig. 8(a).

With the chosen antibody functionalization schema, the antibody coating on the measurement waveguide makes it specifically bind its target analyte onto the sensor surface whereas the IgG-coated reference waveguide takes into account unspecific binding. Since all the sensor waveguides are exposed similarly to the solutions in the measurement window, in theory, the ambient RI change should not induce any phase difference between the measurement and the reference waveguides, and thus also not lead to measurable signal. Therefore, the measured phase change signal can be assumed to directly show the difference between the binding of the analytes on the measurement and the reference waveguides in real-time. In addition, the similarity of the measurement and reference waveguide structures makes them react similarly to temperature changes reducing the temperature sensitivity of the sensor.

The phase change graphs, shown in Fig. 8(a), were analysed from the captured interferograms and were subsequently

baseline corrected based on the values between 0 and 5.7 min following the same procedure with melamine measurements. In each of the three experiments, Y11 shows a clear phase change signal when the sensor was exposed to hCG sample, whereas the curve of Y12 remains nearly unchanged. Similarly, when the sensor was exposed to CRP sample, the phase change curve of Y12 shows a clear phase change signal whereas the curve of Y11 remains unchanged. This indicates that analytes were selectively binding to the receptor layers on the measurement waveguides. During the following flushing steps, the phase change curves level off, indicating that no more analyte molecules were binding onto the surface.

Sensor responses to the hCG and CRP samples were determined from the phase change curves by calculating the average value within 35-38 min for Y11, and within 68-71 min for Y12, i.e. during the flushing step following the analyte exposure. The average response of the three measurements of Y11 to hCG samples was 1.0 rad and the range of the response values was 0.6 rad. Similarly the average response of Y12 to CRP samples was 2.5 rad and the range 1.7 rad. The responses to both analytes varied substantially from chip to chip. This was attributed to variations in the inkjet printed functionalization layers on the waveguides, such as the layer thickness and the amount of capture antibodies, causing also variation in intensity of the fluorescence signal in Fig. 8(b).

As discussed above, the chosen antibody functionalization schema should desensitize the sensor to ambient RI changes. To study sensitivity to ambient RI changes, chips were exposed to 400 μ l of 0.5 wt. % glucose solution followed by flushing with the sample matrix. Timing of the solutions is shown in Fig. 8(a). It can be seen that in all but one of the phase change curves, the response to the ambient RI change is significantly smaller than the value of \sim 1 rad reported earlier by the authors for polymeric R2R fabricated non-functionalized YI sensor chips for aqueous glucose solution having the same concentration. The larger phase change signal of one of the curves, \sim 0.5 rad, is most likely caused by the different thicknesses of the inkjet-printed antibody layers on the measurement and reference waveguides causing a different penetration of the evanescent wave fields into the sample volume and thus different change of the effective RIs. Although the chosen functionalization schema could not render the sensors completely insensitive to the ambient RI changes, the effect was nonetheless suppressed.

After the interferometric measurements, fluorescent staining and imaging was used to verify the specific binding of the analytes to the correspondingly functionalized areas, as well as the presence of IgG on the reference waveguides. Fluorescently labelled antibodies were used to stain CRP and hCG analytes (green fluorescence excitation) and IgG (red fluorescence excitation) in three sequential steps. The fluorescence images of Fig. 8(b) show no fluorescence signal from the printed areas before staining. During each staining step, the corresponding inkjet functionalized area became fluorescent, as shown in Fig. 8(b), proving the presence of CRP, hCG and IgG at these areas, and confirming the results of interferometric measurements. However, when investigating

the fluorescent areas in detail, it can be observed that fluorescent intensity is not uniform. This indicates that the variations in the inkjet printed functionalization layers lead to fluctuating amounts of bound analyte molecules resulting in variation of the response values of biosensor experiments.

Discussion

In this paper we reported for the first time analyte specific sensing using integrated YI sensor chips having R2R mass-fabricated polymeric single-mode waveguides. We demonstrated proof-of-concept for chemical sensing of small molecules using MIP receptor layer, and for multi-analyte detection of biomolecules using inkjet printed antibody-based receptor layers. The model analytes chosen for the demonstrations have substantial interest: melamine has become an important analyte in food adulteration, whereas both hCG and CRP are important clinical parameters: the former is an early stage pregnancy marker, the latter an unspecific inflammation marker and an indicator for the assessment of the risk of cardiovascular disease.

The MIP functionalized chip was shown to be capable to measure melamine concentrations quantitatively having linear responses below the concentration of 0.5g/l and beginning saturation at the higher concentrations. In the case of the antibody functionalized chip, the responses to both analytes varied substantially from chip to chip. Thus the sensor responses at the moment only qualitatively indicate the presence or absence of the analytes and no quantitative responses were obtained. Deviating responses with the antibody-functionalized chips were attributed to variations in the inkjet printed functionalization layers on the waveguides. These are related to reproducibility of layer thickness and the amount of active capture antibodies. The repeatability of the sensor responses is most likely improved if the capture antibodies are bound to waveguide surface oriented and are forming mono layers.

To further characterize the sensor chip properties, the limit of detection, limit of quantification and repeatability of the sensor responses need to be studied by conducting experiments with large number of sensor chips. In addition, to assess the applicability of the sensors for real-life applications, experiments need to be conducted with real samples in realistic experimental conditions.

Conclusions

In conclusion, we demonstrated that disposable polymeric YI sensor chips are applicable for analyte specific chemical sensing of small molecules, and for multi-analyte detection of biomolecules. The MIP functionalized sensor chip indeed revealed responses depending on melamine concentration, which is the first demonstration of MIP on a low-cost, polymer-based sensor device for direct RI sensing. The antibody-functionalized YI sensor chips demonstrated the applicability of patterned, inkjet printed receptor layers for

specific multi-analyte detection showing qualitative indication of the presence or absence of the analytes.

Proving the concept of disposable, low-cost sensor systems thus inherently opens up scientifically and economically interesting strategies for portable sensing and analysis. This ranges from determining the quality of food at home or providing low-cost diagnostic tests for rapid evaluation of important clinical parameters directly at the point-of-care.

Development of low-cost sensor solutions is one of the main challenges in the sensor field. Demonstration of the analyte specific chemical- and biosensing with disposable polymeric YI sensor chips using artificial and biological receptor layers opens new avenues for sensor implementation.

Acknowledgements

Authors acknowledge financial support by the European Commission via the Seventh Framework Programme under the grant agreement No. 263382 "PHOTOSENS", and by Academy of Finland grant No. 284907. Authors also acknowledge the assistance of J. Ollila, VTT, in preparing the sensor chips.

Notes and references

- D. E. W. Patabadige, S. Jia, J. Sibbitts, J. Sadeghi, K. Sellens and C. T. Culbertson, *Anal. Chem.*, 2016, 88, 320.
- Y. Chen, L. Jiang, M. Mancuso, A. Jain, V. Oncescu and D. Erickson, *Nanoscale*, 2012, 4, 4839.
- P. Kozma, F. Kehl, E. Ehrentreich-Förster, C. Stamm and F. F. Bier, *Biosens. Bioelectron.*, 2014, 58, 287.
- A. Ymeti, J. S. Kanger, R. Wijn, P. V. Lambeck and J. Greve, *Sens. Actuator B-Chem.*, 2002, 83, 1.
- K. Schmitt, B. Schirmer, C. Hoffmann, A. Brandenburg and P. Meyrueis, *Biosens. Bioelectron.*, 2007, 22, 2591.
- M. Wang, J. Hiltunen, C. Liedert, L. Hakalahti and R. Myllylä, *J. Eur. Opt. Soc., Rapid Publ.*, 2012, 7, 12019.
- A. Brandenburg and R. Henninger, *Appl. Opt.*, 1994, 33, 5941.
- G. H. Cross, A. A. Reeves, S. Brand, J. F. Popplewell, L. L. Peel, M. J. Swann and N. J. Freeman, *Biosens. Bioelectron.*, 2003, 19, 383.
- J. Xu, D. Suarez and D. S. Gottfried, *Anal. Bioanal. Chem.*, 2007, 389, 1193.
- A. Schimpf, D. Bucci, M. Nannini, A. Magnaldo, L. Couston and J.-E. Broquin, *Sens. Actuator B-Chem.*, 2012, 163, 29.
- G. H. Cross and Y. Ren, *J. Appl. Phys.*, 1999, 86, 6483.
- M. Wang, S. Uusitalo, C. Liedert, J. Hiltunen, L. Hakalahti and R. Myllylä, *Appl. Opt.*, 2012, 51, 1886.
- M. Hiltunen, J. Hiltunen, P. Stenberg, S. Aikio, L. Kurki, P. Vahimaa and P. Karioja, *Opt. Express*, 2014, 22, 7229.
- S. Aikio, J. Hiltunen, J. Hiitola-Keinänen, M. Hiltunen, V. Kontturi, S. Siitonen, J. Puustinen and P. Karioja, *Opt. Express*, 2016, 24, 2527.
- A. Ymeti, J. S. Kanger, J. Greve, G. A. J. Besselink, P. V. Lambeck, R. Wijn and R. G. Heideman, *Biosens. Bioelectron.*, 2005, 20, 1417.
- A. Ymeti, J. Greve, P. V. Lambeck, T. Wink, S. W. F. M. van Hövell, T. A. M. Beumer, R. R. Wijn, R. G. Heideman, V. Subramaniam and J. S. Kanger, *Nano Lett.*, 2007, 7, 394.
- L. Uzun and A. P. F. Turner, *Biosens. Bioelectron.*, 2016, 76, 131.
- P. L. Edmiston, D. P. Campbell, D. S. Gottfried, J. Baughman and M. M. Timmers, *Sens. Actuator B-Chem.*, 2010, 143, 574.
- S. M. Reddy, D. M. Hawkins, Q. T. Phan, D. Stevenson and K. Warriner, *Sens. Actuator B-Chem.*, 2013, 176, 190.
- J. T. Kirk, G. E. Fridley, J. W. Chamberlain, E. D. Christensen, M. Hochberg and D. M. Ratner, *Lab Chip*, 2011, 11, 1372.
- N. Komuro, S. Takaki, K. Suzuki and D. Citterio, *Anal. Bioanal. Chem.*, 2013, 405, 5785.
- A. Ymeti, J. S. Kanger, J. Greve, P. V. Lambeck, R. Wijn and R. G. Heideman, *Appl. Opt.*, 2003, 42, 5649.
- J. S. Kanger, V. Subramaniam, P. H. J. Nederkoorn and A. Ymeti, in *Advanced Photonic Structures for Biological and Chemical Detection*, ed. X. Fan, Springer, Dordrecht Heidelberg London New York, 2009, 10, 265-295.
- J. Liu, A. Ren, L. Yang, J. Gao, L. Pei, R. Ye, Q. Qu and X. Zheng, *Can. Med. Assoc. J.*, 2010, 182, 439.
- M. Montagnana, T. Trenti, R. Aloe, G. Cervellin and G. Lippi, *Clin. Chim. Acta*, 2011, 412, 1515.
- K. D. McClatchey, S. Alkan and E. Hackel, *Clinical Laboratory Medicine (2nd Edition)*, LWW, Philadelphia (PA), 2001.
- P. M. Ridker, *Tex. Heart Inst. J.*, 2005, 32, 384.
- W. Heller, *J. Phys. Chem.*, 1965, 69, 1123-1129.
- D. R. Lide, *CRC Handbook of Chemistry and Physics. 83rd Edition*, CRC, Boca Raton (FL), 2002.
- R. Suedee, W. Intakong, P. A. Lieberzeit, P. Wanichapichart, P. Chooto and F. L. Dickert, *J. Appl. Polym. Sci.*, 2007, 106, 3861.

Journal Name

ARTICLE




Fig. 1 (a) Top view illustration of the sensor chip with names of the parts and dimensions. (b) Illustration of the cross-cut structure of the MIP functionalised sensor chip within the sensing area showing the MIP and NIP coatings on the respective measurement waveguides, and the passivating overcladding layer on the reference waveguides. (c) Antibody functionalized chip having analyte specific anti-hCG and anti-CRP antibody coatings on the respective measurement waveguides and reference IgG coating on the reference waveguides. The separations of the measurement and the reference waveguides at the outcoupling end of the chip, d , are shown next to the sensor illustrations in (b) and (c).




Fig. 2 (a) Schematic illustration of the experimental arrangement. Captured interferogram of (b) MIP functionalized, and (c) antibody functionalized sensor chip.




Fig. 3 (a) Schematic of the cross-section of the inverted ridge waveguide structure showing the materials and the refractive indices¹⁴, n , of the core and undercladding layers at the wavelength of 975 nm, i.e. at the wavelength used in the sensing experiments. (b) Roll of sensor waveguides. (c) Sensor platform cut from the roll.




Fig. 4 Phase change curves measured while the sample matrix flowed in the flow cell, and lines fitted to the data points for a) MIP functionalized chip (image interval 2 s), and b) for antibody functionalized chip (image interval 30 s). The slope of the fitted lines is also shown indicating the trend of the system drift.




Fig. 5 (a) Phase change curves of MIP functionalized Y11 and NIP coated Y12 measured at various melamine concentrations. Each concentration (value shown next to the curves) was measured three times. (b), $\Delta\phi_{\text{diff}}$ curves showing the net difference of the phase change values between the MIP functionalized Y11 and NIP functionalized Y12 at different melamine concentrations. Timing of the solutions in the flow cell is indicated by the arrows and the background colour.




Fig. 6 Shift of the interferogram fringes during a chemical sensing experiment conducted with the melamine concentration of 1.0 g/l. Figure is constructed by first selecting the same narrow section (height 5 pixels) across each of the time laps images captured during the experiment, and then combining these sections.




Fig. 7 Calculated responses and lines fitted to the data points based on (a) the phase change step height, and (b) the slope of the rising edge of $\Delta\Phi_{diff}$ curves.






Fig. 8 (a) Phase change curves showing the specific responses of Y11 (dashed lines) and Y12 (solid lines) to hCG and human CRP samples, respectively. Responses to glucose solution demonstrate the sensor sensitivity to ambient RI changes. (b) Left: Microscope images of the YI chip before and after the antibody functionalization by inkjet printing showing the printed areas on top of the waveguides. Printed areas are visible due to the salts in the bio-inks that dry onto the chip surface. Right: Verification of specific binding of analytes onto the inkjet printed areas by fluorescent staining and imaging. The fluorescence images (the darker the colour the higher the fluorescence intensity) taken before staining show no fluorescence from the printed areas. After each staining step, the corresponding inkjet printed area became fluorescent verifying the localized presence of the analytes and the reference IgG coating.

Demonstration of disposable multi-analyte polymeric integrated Young interferometers for analyte specific chemical- and biochemical sensing using biological and biomimetic recognition

Roll of sensor waveguides



Biomimetic recognition



Biological recognition

