


 Cite this: *Lab Chip*, 2024, 24, 924

## Label-free cell classification in holographic flow cytometry through an unbiased learning strategy

 Gioele Ciaparrone,<sup>a</sup> Daniele Pirone,<sup>b</sup> Pierpaolo Fiore,<sup>iD</sup><sup>a</sup> Lu Xin,<sup>c</sup> Wen Xiao,<sup>c</sup> Xiaoping Li,<sup>d</sup> Francesco Bardozzo,<sup>ab</sup> Vittorio Bianco,<sup>iD</sup><sup>b</sup> Lisa Miccio,<sup>b</sup> Feng Pan,<sup>iD</sup><sup>\*c</sup> Pasquale Memmolo,<sup>iD</sup><sup>\*b</sup> Roberto Tagliaferri<sup>\*ab</sup> and Pietro Ferraro<sup>iD</sup><sup>b</sup>

Nowadays, label-free imaging flow cytometry at the single-cell level is considered the stepforward lab-on-a-chip technology to address challenges in clinical diagnostics, biology, life sciences and healthcare. In this framework, digital holography in microscopy promises to be a powerful imaging modality thanks to its multi-refocusing and label-free quantitative phase imaging capabilities, along with the encoding of the highest information content within the imaged samples. Moreover, the recent achievements of new data analysis tools for cell classification based on deep/machine learning, combined with holographic imaging, are urging these systems toward the effective implementation of point of care devices. However, the generalization capabilities of learning-based models may be limited from biases caused by data obtained from other holographic imaging settings and/or different processing approaches. In this paper, we propose a combination of a Mask R-CNN to detect the cells, a convolutional auto-encoder, used to the image feature extraction and operating on unlabelled data, thus overcoming the bias due to data coming from different experimental settings, and a feedforward neural network for single cell classification, that operates on the above extracted features. We demonstrate the proposed approach in the challenging classification task related to the identification of drug-resistant endometrial cancer cells.

 Received 4th May 2023,  
 Accepted 24th December 2023

DOI: 10.1039/d3lc00385j

[rsc.li/loc](http://rsc.li/loc)

## Introduction

Imaging flow cytometry (IFC) is the current state of the art technology for single-cell analysis,<sup>1</sup> exploited for powerful clinical applications, both diagnostic<sup>2,3</sup> and therapeutic.<sup>4,5</sup> In IFC systems, thousands of cells are individually analysed in very short times as they are illuminated by a light source while they are flowing in a microfluidic channel.<sup>6</sup> The high-throughput property allows collecting large amounts of data, thus providing a multi-scale characterization of a heterogenous cell line.<sup>7</sup> In fact, average parameters can be measured, but, at the same time, the intra-species variability is not lost since each single cell is regarded as a separate

entity. Therefore, a technological effort is underway aimed at the miniaturization of the IFCs towards on-chip systems, cheap and simple to fabricate.<sup>8</sup> The conventional IFC benefits from the use of fluorescent dyes with the aim of staining in different ways the several subcellular structures, thus making an intra-cellular analysis possible. However, the same exogenous markers play a destructive role, since they interfere with the cell (photodamaging) and alter the imaging (photobleaching), with the risk of recording an untrue signal. Besides, staining protocols can be expensive and time-consuming, and they often require trained operators. More importantly, in addition to the 2D morphological information, no quantitative measurement can be extracted from the recorded images about the cell biophysical properties. These drawbacks reduce the IFC potential in understanding more complex phenomena and point out the need to integrate into IFCs a label-free and quantitative imaging module. Among several technologies, digital holography (DH) in microscopy is a remarkable candidate as it exploits the light interference principle to image a biological specimen, thus allowing quantitative phase-contrast information without using exogenous agents.<sup>9,10</sup> The simpler and cheaper sample preparation, along with the unique capability of numerical refocusing, makes the

<sup>a</sup> *Neurone Lab, Department of Management and Innovation Systems (DISA-MIS), University of Salerno, Fisciano, Italy. E-mail: robttag@unisa.it*

<sup>b</sup> *CNR – Institute of Applied Sciences and Intelligent Systems “Eduardo Caianiello”, Pozzuoli, Italy. E-mail: pasquale.memmolo@isasi.cnr.it*

<sup>c</sup> *Key Laboratory of Precision Opto-Mechatronics Technology of Ministry of Education, School of Instrumentation Science & Optoelectronics Engineering, Beihang University, 100191 Beijing, China. E-mail: panfeng@buaa.edu.cn*

<sup>d</sup> *Department of Obstetrics and Gynecology, Peking University People's Hospital, Beijing 100044, China*



combination of DH imaging to conventional IFC systems a natural evolution,<sup>11,12</sup> further exploitable for on-chip implementations.<sup>13,14</sup> By using an appropriate numerical processing, the reconstruction of a quantitative phase map (QPM) from the recorded digital hologram can be obtained.<sup>15</sup> Unlike fluorescence imaging, in a QPM the cell biophysical information is recorded, as the cell refractive index (RI) is coupled to the cell morphology in the form of a 2D image, thus explaining the increasing applications of quantitative phase imaging (QPI) in biomedicine.<sup>16</sup> The most fascinating evolution of QPI is phase-contrast tomography (PCT), which consists in mapping the 3D RI spatial distribution at the single-cell level,<sup>17</sup> and which has been also reproduced under flow cytometry conditions.<sup>18,19</sup> Therefore, the capability of a Holographic-IFC (HIFC) to provide a quantitative biophysical cell fingerprint suggests the enormous potential in combining this technology with artificial intelligence (AI) for label-free single-cell phenotyping. While in the 3D PCT case, the use of machine learning<sup>20,21</sup> and deep learning<sup>22</sup> is starting to be tested, in the 2D QPI case the learning approaches have been widely investigated,<sup>23–27</sup> also in a flow cytometry environment.<sup>28–35</sup> The standard workflow consists of creating a training dataset, usually obtained by acquiring digital holograms of single cells in the HIFC system under investigation. Then, QPMs are retrieved and used to train classifiers exploiting the paradigms of feature selection or convolutional neural networks. The assessment of the learned classification model is made by using testing datasets with additional QPMs obtained by using the same acquisition setting and processing pipeline. However, it is demonstrated that this protocol suffers from generalization robustness with respect to data coming from other imaging systems or processed in a different way,<sup>36</sup> *i.e.*, the training stage can be considered biased by the employed recording and processing settings. In addition, heavy and long computations to process a digital hologram to recover the corresponding QPM may represent a further bottleneck for on-chip applications.<sup>37</sup> In this paper, we propose a strategy to overcome unwanted biases mainly related to the intrinsic and unpredictable changes within the experimental protocol, *e.g.*, sample preparation by different users, recording in different days, imaging system calibration, *etc.* We design a learning scheme based on the combination of a convolutional neural network for single cell hologram detection, followed by a convolutional autoencoder operating on unlabelled data and a shallow neural network for cell classification. We demonstrate that the adopted solution is robust with respect to different HIFC acquisition biases, mainly related to the recorded image brightness changes, different levels of speckle noise and fringe orientation variations, thanks to the use of the unsupervised autoencoder for the image feature extraction. It overcomes possible biases caused by different holographic reconstruction methodologies since the learning model is trained with unprocessed digital holograms. In fact, a digital hologram encodes compact and dense information about the light–cell interaction in the form of a 2D intensity

map, hence including the phase delay retrieved in the corresponding QPM. Actually, learning approaches fed by the raw holographic diffraction patterns are expected to provide even better results in terms of using QPMs, as recently demonstrated.<sup>38,39</sup>

Here we employ the proposed methodology in detecting drug resistant and non-drug resistant cells within the same Endometrial Cancer (EC) cell line, recorded in a HIFC system. EC is the sixth most diagnosed cancer in women worldwide<sup>40,41</sup> and the incidence of EC has increased in many countries.<sup>42</sup> To date, the best therapeutic approach is chemotherapy, which mainly exploits platinum-based drugs. Among them, cisplatin is believed to effectively affect EC proliferation and apoptosis.<sup>43</sup> Unfortunately, the recurrent or advanced disease that develops in a minority of patients reduces the therapeutic efficacy because of chemoresistance.<sup>44,45</sup> Hence, in such cases an accurate evaluation of the drug sensitivity of the patient can allow planning a more effective individualized chemotherapy. A promising solution to this challenging clinical issue can be found in the well assessed discovery about distinctive morphological changes in drug resistant cancer cells.<sup>46–50</sup> Then, single-cell analysis becomes the key-technique to fully exploit this property. In summary, the results reported in this paper demonstrate that the combination among raw hologram inference, the high-throughput and single-cell analysis provided by HIFC, and the fast, automatic and accurate results obtained by the proposed unbiased learning approach pushes towards the realization of lab-on-chip devices for aiding medical decision making.<sup>51</sup> Then, the label-free and non-destructive DH properties become fundamental to allow downstream more in-depth studies on drug resistant cells after a real-time AI-based cell identification.<sup>52–54</sup>

## Materials and methods

### Sample preparation

Two kinds of human EC cells, namely ISK cells and ISK-CisR cells, were considered in this study. ISK is a well-differentiated adenocarcinoma, oestrogen receptor  $\alpha$  (ER $\alpha$ ) (+), ER $\beta$ (+), and progesterone receptor (PR)(+) cell line derived from the American Type Culture Collection and stored in the Obstetrics and Gynecology Laboratory of Peking University People's Hospital. The drug-resistant characteristics of ISK-CisR cells were induced by exposing ISK cells to cisplatin for 10 months. Both cell types were cultured and grown in DMEM/F12 (Dulbecco's Modified Eagle's Medium/Ham's F-12 50/50 Mix, CORNING, USA) supplemented with L-glutamine, 15 mM HEPES and 10% fetal bovine serum (Gibco 10099-141, Australia). Then, cells were seeded onto a 100 mm glass Willco dish at a density of  $13.7 \times 10^6$  cells for 24 hours with 5% CO<sub>2</sub> and 37 °C. Finally, the cells were collected from wells using trypsin digestion, washed in phosphate-buffered saline (PBS) to dilute to half the concentration (about  $5 \times 10^6$ ), and then suspended in 500  $\mu$ L of 1 $\times$  binding buffer (10 mM



HEPES, 140 mM NaCl, 2.5 mM CaCl<sub>2</sub>, pH 7.4) with 10 μL of annexin conjugate and PI.

### HIFC system design and implementation

We built up a HIFC system to acquire holographic images of flowing cells, as shown in Fig. 1A. The recording system is equipped with a Mach-Zehnder imaging interferometer in off-axis configuration. A 100 mW coherent laser at 532 nm is used as the light source employing a single-mode fiber. The plane wave is produced by a collimating lens (L) and split into the object beam and the reference beam by a polarizing beam splitter (PBS). In the object arm we used a 20× microscope objective (MO) with a numerical aperture of 0.50 and a tube lens (TL) to amplify the object and output plane wave. Such an object wave is then combined with the reference one in a beam splitter (BS) at a small angle and the digital holograms of the object are recorded by a 1024 × 1024 CCD camera having 5.5 × 5.5 μm<sup>2</sup> pixel sizes (PointGrey, Canada) operating at 20 fps. The cell solution is flowing in a microfluidic PMMA channel with 100 μm in width and depth 58.5 mm in length (microfluidic-chipshop, Fluidic 144, Germany) driven by a syringe pump system with an injection rate of 5 μl min<sup>-1</sup>.

### Dataset construction

In order to create the dataset for solving the above-mentioned drug-resistant classification problem, 9 holographic video

sequences per class were recorded, corresponding to 9316 and 8789 holographic frames containing respectively the ISK and ISK-CisR cells flowing in the microfluidic channel, as shown in Fig. 1B and C. Among them, we randomly select 8 videos (4 for ISK and 4 for ISK-CisR) to be used for validating and testing the performance of the entire classification scheme. In this way, it can be ensured that the learning system is not influenced by the bias induced in the Mask R-CNN and concurrently in the classification learning system. Two possible strategies were tested, with the aim of understanding how the holographic diffraction pattern biases can influence the classifier's performances. The first one consists in taking all the diffraction information by selecting a bounding box around the unfocused cell. The second one is based on the segmentation of the cell with respect to its background by means of a binary mask, thus reducing the diffraction content. In order to perform a fast and automatic selection of the region of interests (ROIs) containing the single cells used for feeding the classifier, a Mask R-CNN<sup>55</sup> with ResNet-101 (ref. 56) backbone was trained to extract both bounding boxes and cell masks from each holographic frame. The detection network was trained on a subset of 4 videos (2 per class), while 2 videos (1 per class) were used as a validation set for hyperparameter tuning. Other 4 videos (2 per class) were instead used as an independent test set. More precisely, the Computer Vision Annotation Tool (CVAT)<sup>57</sup> was used to label 2 different datasets made of the same 120 × 120 ROIs, *i.e.* a bounding box (*bbbox*) dataset and a *mask* dataset. In this way, a total of 71 528 cells were labelled, *i.e.*, 44 456 cells belonging to the ISK class and 27 072 cells belonging to the ISK-CisR class. In addition, 14 890 dubious objects were identified. They were considered false positives during the training step and manually removed because we wanted the model to extract only real cells, while they were disregarded at evaluation time for computing the Average Precision (AP). AP is used to evaluate the precision (the ratio of correctly identified positives to the total identified positives) across varying thresholds of recall (the ratio of correctly identified positives to the total actual positives). AP is essentially an average measure of precision at different levels of recall, providing an overall performance metric for the precision-recall curve. We used transfer learning by initializing the convolutional neural network (CNN) with weights trained on the COCO object detection dataset.<sup>58</sup> Frames were cropped to 1024 × 625 by removing the channel's borders, as displayed in Fig. 1B and C. The Mask R-CNN was trained for 25 epochs using Stochastic Gradient Descent with momentum, with a learning rate of 10<sup>-3</sup>, a momentum of 0.9, and 4 images per batch. We chose the best epoch using the standard AP metric computed at Intersection over Union IoU = 0.5 on the validation set. Finally, two different test analyses are needed to ultimately evaluate both learning systems. The output of the Mask R-CNN is reported in Fig. 2A and C, along with some bounding boxes and cell masks displayed as examples in Fig. 2B and D about the ISK and ISK-CisR cells, respectively. Then, for the cell classification experiments, we created

- a training set made of the 10 videos (5 per class) used to train, validate, and test the cell detector;



**Fig. 1** Opto-fluidic recording system. A) Sketch of the DH flow cytometer. L – lens; M – mirror; HWP – half-wave plate; PBS – polarizing beam splitter; MO – microscope objective; TL – tube lens; BS – beam splitter; CCD – camera. B and C) report 1024 × 1024 DH frames with flowing ISK and ISK-CisR cells, respectively, with highlighted in red the 1024 × 625 cut performed for feeding the Mask R-CNN. The scale bar is 20 μm.





**Fig. 2** Mask R-CNN detection of (A and B) ISK and (C and D) ISK-CisR flowing cells inside the recorded holograms. (A and C) Digital holograms given in input to the detection CNN, with highlighted in orange the predicted ROIs. (B and D) Bounding boxes (up) and cell masks (down) selected by the detection CNN and used as input of the classification CNN. The ROI's side is 27.6  $\mu\text{m}$ .

- a validation set and a test set each made of 4 videos (2 per class) not used to train, validate, and test the cell detector.

The training set included 17 236 cells (10 881 ISK and 6355 ISK-CisR), the validation set included 8842 cells (5843 ISK and 2999 ISK-CisR), while the test set included 9126 cells (6171 ISK and 2955 ISK-CisR). Notice that the validation and test sets are created from the 8 videos set aside previously and never seen by the network during the training stage.

### CNN classification

The CNN model used to discriminate ISK and ISK-CisR cells is made of two main sections, as sketched in Fig. 3. The former, which includes the convolution and pooling layers, extracts features from the images, while the latter (*i.e.*, the

fully-connected head) classifies them. The input layer takes  $120 \times 120$  images normalized in the  $[0, 1]$  interval. It is followed by 4 blocks, each made of a  $3 \times 3$  convolution layer with different number of filters (32, 64, 128, and 128) and a max pooling layer with  $2 \times 2$  size, in order to reduce the number of parameters (*i.e.*, the computational complexity) and to prevent over-fitting. All convolutional layers use Rectified Linear Unit (ReLU) as activation function. The classifier head has two fully-connected (dense) layers (512-d and 1-d) and a dropout layer again to prevent over-fitting.

In particular, the final 1-d fully connected layer predicts the probability that a cell belongs to one of the two classes by using a sigmoid activation function. We trained the network using a binary cross-entropy loss, the RMSprop optimization algorithm<sup>59</sup> with the learning rate set to  $10^{-4}$ , and a batch size of 32. We used an early stopping strategy in order to stop the training process when no improvement on the validation loss was obtained for at least 5 consecutive epochs.



**Fig. 3** CNN classifier. Sketch of the CNN used to classify ISK and ISK-CisR cells.

## Results

We firstly evaluate the performance of the Mask-RCNN by using 4 test videos. The network obtained 95.9%, 95.7%, 97.8%, and 94.4% AP in the bounding box detection task, with a corresponding miss rate between 5.6% and 2.1%, and 96.0%, 95.7%, 97.9%, and 94.4% AP in the cell mask prediction task with a corresponding miss rate between 5.6% and 2.2%. The abovementioned miss rate refers to the proportion of actual positive cases (cells, in this context) that the system fails to detect. It is equivalent to the false negative rate, or  $1 - \text{recall}$ . The first two scores refer to the 2 ISK videos, while the last two scores refer to the 2 ISK-CisR



videos. The same CNN classification architecture and training procedure were used for identifying ISK and ISK-CisR cells in both the *mask* and *bbox* datasets, and 99.93% and 100% accuracies were reached, respectively. Notice that the presence of the diffraction content included in the *bbox* case seems to push up the classification performance. Moreover, the almost equal accuracies can be explained by referring to Fig. 2, in the case of cells recorded very close to their focus plane, since their sharpness appears very low as in the numerical refocusing. However, high classification performances with simple CNN architectures, as in our case, are typically achieved when the classification problem is relatively easy to address. Definitely, by considering that the *bbox* strategy does not introduce any significant performance's gain, it is more effective to consider the *mask* detection method only for evaluating the overall performance of the proposed unbiased learning strategy due to its ability to extract only the smallest region containing the cell, thus reducing the number of informative pixels and becoming more suitable for the subsequent steps. As claimed in the introduction, the aim of this study is to test an unbiased learning strategy to identify drug-resistant EC cells, namely ISK-CisR, with respect to the background cell line, *i.e.*, ISK. Essentially, the identification problem can be seen as a phenotyping of a cell sub-class within the same cell line, that is usually a hard issue. Therefore, one can expect that the CNN has learned intrinsic biases of the training dataset.

The main causes of biases in HIFC data are related to different recording and processing settings. To avoid the latter, we implement a classifier that works on raw hologram images, thus completely skipping the processing step. Instead, there are several intrinsic recorded video properties that could be sources of bias and their removal can be a hard procedure. For example, the holographic sequences used for creating the training sets were recorded in different days, showing variation in frame brightness. In this case, the

brightness may be removed by rescaling the intensity of all the holographic videos to the same average value. In addition, although training sets contain only suitable cells, the network classifies also non-cell objects as belonging to one of the two possible classes. Thus, the network may be using more than just intrinsic cell properties to classify input images. In order to extract actual cell features less likely to be biased by intrinsic video properties, we employed the convolutional autoencoder<sup>60–62</sup> sketched in Fig. 4A. The autoencoder is a type of encoder/decoder neural network that allows for the extraction of a compact representation capable of providing a reasonably accurate reconstruction of the input, while eliminating various types of noise and perturbations at multiple levels of abstraction, such as in microscopy<sup>63</sup> or, for example, in smoke removal from soft tissues in surgical videos,<sup>64</sup> *etc.* We trained the network using the Adam optimizer<sup>65</sup> with a learning rate of 0.01, a batch size of 32 images, and the mean squared error as loss function. We used the same training, validation, and test sets as the classification experiment. We only performed the experiments on the *mask* dataset, in order to remove any information outside the cell boundaries that could induce a further bias. We used an early stopping strategy to stop training when the validation loss did not decrease by at least  $10^{-7}$  for 30 consecutive epochs. A 512-d bottleneck produced a suitable cell reconstruction, shown as an example in Fig. 4B. It is clear that the network is able to preserve the main informative content at low-frequency, while the fine details at high-frequency are instead mostly lost. We thus expect that this approach should discard the residual experimental bias due to its unsupervised nature. Therefore, we trained a shallow fully-connected neural network with 2 hidden layers with 32 neurons each to perform the classification task according to the 512 features extracted by the CNN autoencoder. We used binary cross-entropy loss, Adam optimizer with 0.01 learning rate, and early stopping



Fig. 4 Feature extraction by means of a convolutional autoencoder. A) Architecture of the convolutional autoencoder. B) 9 EC cells (on the left) along with the corresponding autoencoder reconstructions (on the right).



with patience of 30 epochs, and we obtained a test set accuracy of  $92.26 \pm 0.67\%$  computed by averaging the results obtained by repeating 10 times the training of the shallow network, in order to take into account the random weight initialization. Of course, while always remaining high, the performances of the unsupervised approach are lower than the supervised one, but with the advantage of having removed possible biases due to the experiment that could reduce the classification robustness.

By using tSNE,<sup>66</sup> we projected the 512 unsupervised features of the training, validation, and test sets, as well as those of the entire dataset, in order to produce the human-readable 2D plots reported in Fig. 5A–D, respectively. As further confirmation of the advantages in exploiting the holographic diffraction patterns for solving classification problems, it is evident that the two classes are well separated even by using the autoencoder features.

However, a major overlap area can be observed, as highlighted in red in Fig. 5D. As visible in Fig. 5E and F, manual inspection showed that points in this area correspond to a high number of non-cell objects, resulting from detection errors of the Mask R-CNN. Therefore, we filtered the dataset by removing this cluster, thus obtaining a

cleaner dataset including 17 064 cells in the training set (10 709 ISK, 6355 ISK-CisR), 7779 cells in the validation set (4894 ISK, 2885 ISK-CisR), and 8046 cells in the test set (5411 ISK, 2635 ISK-CisR). Then, we trained the shallow network using the 512-d autoencoder features of the new clean dataset. As reported by the confusion matrix in Table 1, the accuracy on the test set rose to  $96.68 \pm 0.61\%$ . In order to assess the robustness of the CNN model with respect to the brightness changes, we both decreased or increased the image pixel values of the test set by 50%. In the first case, the network obtained 99.65% and 99.41% accuracies on the mask and *bbox* test sets, respectively, while in the second case it obtained 99.44% and 100% accuracy on the mask and *bbox* test sets, respectively. The almost unchanged performances point out that the network is able to generalize with respect to the brightness of the DH experiment.

### CNN validation in the presence of biases never seen by the trained model

In order to evaluate the effectiveness and the robustness of the proposed method with respect to further sources of biases never seen by the trained CNN model, two test sets



**Fig. 5** tSNE applied to the 512 features computed by the convolutional autoencoder. A–C) 2D projections of the training set, validation set, and test set features about the ISK cells (violet) and the ISK-CisR cells (yellow). D) 2D projection of the whole dataset, with highlighted in red the overlap region between the ISK and ISK-CisR cells. E and F) Non-cell objects detected by the Mask R-CNN from the ISK and ISK-CisR holographic videos, respectively, randomly selected from the overlap region in D.



**Table 1** Confusion matrix related to the prediction of ISK and ISK-CisR classes by a shallow network trained through the clean dataset

		Predicted	
		ISK	ISK-CisR
7	ISK	96.76 ± 0.84%	3.24 ± 0.84%
	ISK-CisR	3.49 ± 0.99%	96.51 ± 0.99%

have been created numerically for simulating the presence of higher image noise level and interference fringe variation. In general, the noise level of the holographic images as well as the direction, orientation and frequency of the interference fringe patterns can remarkably change when the digital hologram recording is made by different imaging systems. In addition, noise increasing and/or fringe pattern variation can be observed on the same optical setup if, for example, unwanted misalignment occurs. To test and validate the proposed unbiased learning method in the above cases, we randomly selected 1000 images per class from the test set used in the previous section and we simulated the biases on them. In Table 2 the resulting subset of 2000 images is named “Originally Recorded”. In particular, we create three new test sets by adding speckle noise numerically with three different standard deviations of noise, namely  $\sigma = \{0.01, 0.03, 0.05\}$ . Moreover, we simulate the fringe orientation variation in the ranges  $[-3, +3]$ ,  $[-6, +6]$ ,  $[-10, +10]$  around the recorded fringe direction. In this way, we create 6 new datasets and used them for testing the performance of the proposed CNN model.

Fig. 6 reports an example of an ISK cell (Fig. 6A) and the related simulated biases, *i.e.*, the image with increased noise level obtained by adding the speckle noise with  $\sigma = 0.05$  (Fig. 6B) and a fringe rotation of 10 degrees (Fig. 6C). The evaluation of the performance of the CNN model is made by calculating the average accuracy in classification, as reported in Table 2.

As expected, the convolutional autoencoder is highly robust with respect to the increased noise level. In fact, even with more noisy images, the classification performance remains almost the same. Instead, the fringe variation can cause a remarkable image distortion if compared to the originally recorded one (as shown in Fig. 6A and C) but the proposed CNN model still provides a very good average

**Table 2** Average accuracy in predicting the ISK and ISK-CisR classes for the simulated test sets

Dataset	Accuracy (%)
Originally recorded	99.70
Speckle noise with $\sigma = 0.01$	99.60
Speckle noise with $\sigma = 0.03$	99.75
Speckle noise with $\sigma = 0.05$	99.70
Fringe rotation in the range $[-3, +3]$	99.15
Fringe rotation in the range $[-6, +6]$	98.40
Fringe rotation in the range $[-10, +10]$	94.35

**Fig. 6** Simulation of experimental biases. A) Originally recorded ISK cell. In B a speckle noise is added to the image in A. In C a rotation of the fringe pattern is simulated on the image in A.

accuracy. This result confirms that the convolutional autoencoder is capable of strongly mitigating different types of “noise” (*i.e.*, biases) under specific conditions and with several experimental configurations. Finally, we consider the merged dataset obtained from the cases “Speckle noise with  $\sigma = 0.05$ ” and “Fringe rotation in the range  $[-10, +10]$ ”, to simulate multiple biases simultaneously, achieving an average accuracy of 92.84%.

## Conclusions

Currently, deep learning for image classification from flow cytometry data is one of the most interesting strategies for cell phenotyping. The possibility to perform cell identification in label-free modality has been established by using a HIFC system. However, the numerical processing needed to recover quantitative phase images from recorded holograms, to be further used to train and infer the deep CNN, is often cumbersome due to the computation cost. Here we investigate the possibility of using raw holographic images as input of a CNN, thus completely skipping the processing step. Although this solution was already demonstrated in previous papers, the generalization capabilities of such models may be limited by intrinsic biases within the recording stage, thus limiting their exploitation only for data coming from the same imaging system and acquired exactly with the same setting. In this paper, we propose a learning pipeline able to counteract experimental and numerical processing biases, thus making the CNN inference robust and independent from the origin of datasets. In particular, we used a convolutional autoencoder, whose aim is to remove noise, and mitigate the biases effect, assuming them as a generic “noise”. The reported results show that the proposed solution is able to effectively overcome the effect of different experimental biases, related to image brightness, noise level and fringe orientation. We successfully apply the proposed methodology for the identification of drug resistant and non-drug resistant cells within the same EC cell line, thus demonstrating the effectiveness of the unbiased learning strategy. Of course, a generalized unbiased classification framework needs to be further investigated to include other sources of biases, also considering data collected with other HIFC systems and/or data recording settings. This would be obtained only by



remarkably increasing the training set by adding data from different recording systems. Therefore, based on the achieved proof-of-concept results, this strategy is found to be promising and encouraging for further investigations in other applications of cytopathology studies. We anticipate that a fully automatic pipeline is one of the main goals to be aimed at. In fact, the miss rate of the Mask R-CNN needs to be pushed down, also avoiding the necessity of manually removing false positive predictions. Finally, further types of bias need to be considered and, in case of low performance of the current CNN model, other more complex models and strategies might be used.

## Author contributions

G. C., D. P. and F. B. were in charge of the data analysis/numerical results and wrote the original draft; P. Fi. was in charge of the data analysis/numerical results in the revised draft; L. X., W. X. and X. L. took care of biological samples and performed the experiment; V. B. and L. M. revised the original draft; P. M. supervised, wrote and revised the original draft; F. P., R. T. and P. Fe. revised the original draft, conceptualized and supervised the research.

## Conflicts of interest

There are no conflicts to declare.

## Acknowledgements

This work was supported by project PRIN 2017, MORphological biomarkers For Early diagnosis in Oncology (MORFEO) Prot. 2017N7R2CJ funded by the Italian Ministry of University & Research.

## References

- 1 T. Blasi, H. Hennig, H. D. Summers, F. J. Theis, J. Cerveira, J. O. Patterson, D. Davies, A. Filby, A. E. Carpenter and P. Rees, *Nat. Commun.*, 2016, **7**, 10256.
- 2 M. Doan, I. Vorobjev, P. Rees, A. Filby, O. Wolkenhauer, A. E. Goldfeld, J. Lieberman, N. Barteneva and A. E. Carpenter, *Trends Biotechnol.*, 2018, **36**, 649–652.
- 3 L. Miccio, F. Cimmino, I. Kurelac, M. M. Villone, V. Bianco, P. Memmolo, F. Merola, M. Mugnano, M. Capasso, A. Iolascon, P. L. Maffettone and P. Ferraro, *View*, 2020, **1**, 20200034.
- 4 C. Probst, Y. Zeng and R.-R. Zhu, *J. Pharm. Sci.*, 2017, **106**, 1952–1960.
- 5 J. Hazin, G. Moldenhauer, P. Altevogt and N. R. Brady, *J. Immunol. Methods*, 2015, **423**, 70–77.
- 6 Y. G. Yuanyuan Han, Y. Gu, A. Ce Zhang and Yu.-H. Lo, *Lab Chip*, 2016, **16**, 4639–4647.
- 7 N. S. Barteneva, E. Fasler-Kan and I. A. Vorobjev, *J. Histochem. Cytochem.*, 2012, **60**(10), 723–733.
- 8 M. I. Lapsley, L. Wang and T. J. Huang, *Biomarkers Med.*, 2013, **7**, 75–78.
- 9 P. Ferraro, A. Wax and Z. Zalevsky, *Coherent Light Microscopy, Springer Series in Surface Sciences*, Springer, 2011, vol. 46.
- 10 M. K. Kim, *SPIE Rev.*, 2010, **1**, 018005.
- 11 J. Min, B. Yao, V. Trendafilova, S. Ketelhut, L. Kastl, B. Greve and B. Kemper, *J. Biophotonics*, 2019, **12**(9), e201900085.
- 12 N. A. Turko and N. T. Shaked, *Biomed. Opt. Express*, 2020, **11**, 6649–6658.
- 13 S. Seo, T. W. Su, D. K. Tseng, A. Erlinger and A. Ozcan, *Lab Chip*, 2009, **9**, 777–787.
- 14 F. Merola, P. Memmolo, L. Miccio, V. Bianco, M. Paturzo and P. Ferraro, *Proc. IEEE Inst. Electr. Electron Eng.*, 2015, **103**(2), 192–204.
- 15 M. K. Kim, *SPIE Rev.*, 2010, **1**, 018005.
- 16 Y. K. Park, C. Depeursinge and G. Popescu, *Nat. Photonics*, 2018, **12**, 578–589.
- 17 W. Choi, C. Fang-Yen, K. Badizadegan, S. Oh, N. Lue, R. R. Dasari and M. S. Feld, *Nat. Methods*, 2007, **4**, 717–719.
- 18 F. Merola, P. Memmolo, L. Miccio, R. Savoia, M. Mugnano, A. Fontana, G. D'Ippolito, A. Sardo, A. Iolascon, A. Gambale and P. Ferraro, *Light: Sci. Appl.*, 2017, **6**, e16241.
- 19 D. Pirone, J. Lim, F. Merola, L. Miccio, M. Mugnano, V. Bianco, F. Cimmino, F. Visconte, A. Montella, M. Capasso, A. Iolascon, P. Memmolo, D. Psaltis and P. Ferraro, *Nat. Photonics*, 2022, **16**, 851–859.
- 20 J. Yoon, Y. Jo, M.-H. Kim, K. Kim, S. Y. Lee, S.-J. Kang and Y. K. Park, *Sci. Rep.*, 2017, **7**, 6654.
- 21 D. Pirone, L. Xin, V. Bianco, L. Miccio, W. Xiao, L. Che, X. Li, P. Memmolo, F. Pan and P. Ferraro, *Sens. Actuators, B*, 2023, **375**, 132963.
- 22 D. Ryu, J. Kim, D. Lim, H.-S. Min, I. Y. Yoo, D. Cho and Y. K. Park, *BME Front.*, 2021, **2021**, 9893804.
- 23 Y. Jo, H. Cho, S. Y. Lee, G. Choi, G. Kim, H.-S. Min and Y. K. Park, *IEEE J. Sel. Top. Quantum Electron.*, 2018, **25**(1), 1–14.
- 24 C. X. Chen, H. S. Park, H. Price and A. Wax, *Front. Phys.*, 2021, **9**, 759142.
- 25 P. Memmolo, G. Aprea, V. Bianco, R. Russo, I. Andolfo, M. Mugnano, F. Merola, L. Miccio, A. Iolascon and P. Ferraro, *Biosens. Bioelectron.*, 2022, **201**, 113945.
- 26 G. Kim, Y. Jo, H. Cho, H.-S. Min and Y. K. Park, *Biosens. Bioelectron.*, 2019, **123**, 69–76.
- 27 Y. H. Lin, K. Y. K. Liao and K. B. Sung, *J. Biomed. Opt.*, 2020, **25**(11), 116502.
- 28 Y. Li, B. Cornelis, A. Dusa, G. Vanmeerbeeck, D. Vercruyse, E. Sohn, K. Blaszkiewicz, D. Prodanov, P. Schelkens and L. Lagae, *Comput. Biol. Med.*, 2018, **96**, 147–156.
- 29 M. Ugele, M. Weniger, M. Stanzel, M. Bassler, S. W. Krause, O. Friedrich, O. Hayden and L. Richter, *Adv. Sci.*, 2018, **5**(12), 1800761.
- 30 M. Ugele, M. Weniger, M. Leidenberger, Y. Huang, M. Bassler, O. Friedrich and L. Richter, *Lab Chip*, 2018, **18**(12), 1704–1712.
- 31 D. Roitshtain, L. Wolbromsky, E. Bal, H. Greenspan, L. L. Satterwhite and N. T. Shaked, *Cytometry, Part A*, 2017, **91**(5), 482–493.
- 32 D. K. Singh, C. C. Ahrens, W. Li and S. A. Vanapalli, *Lab Chip*, 2017, **17**(17), 2920–2932.



- 33 N. Nissim, M. Dudaie, I. Barnea and N. T. Shaked, *Cytometry, Part A*, 2021, **99**(5), 511–523.
- 34 M. Rubin, O. Stein, N. A. Turko, Y. Nygate, D. Roitshtain, L. Karako, I. Barnea, R. Giryes and N. T. Shaked, *Med. Image Anal.*, 2019, **57**, 176–185.
- 35 L. Xin, W. Xiao, L. Che, J. Liu, L. Miccio, V. Bianco, P. Memmolo, P. Ferraro, X. Li and F. Pan, *ACS Omega*, 2021, **6**(46), 31046–31057.
- 36 H. Chen, L. Huang, T. Liu and A. Ozcan, *Light: Sci. Appl.*, 2022, **11**, 254.
- 37 D. Pirone, D. Sirico, L. Miccio, V. Bianco, M. Mugnano, P. Ferraro and P. Memmolo, *Lab Chip*, 2022, **22**, 793–804.
- 38 M. Delli Priscoli, P. Memmolo, G. Ciaparrone, V. Bianco, F. Merola, L. Miccio, F. Bardozzo, D. Pirone, M. Mugnano, F. Cimmino, M. Capasso, A. Iolascon, P. Ferraro and R. Tagliaferri, *IEEE J. Sel. Top. Quantum Electron.*, 2021, **27**(5), 1–9.
- 39 Y. Zhu, C. H. Yeung and E. Y. Lam, *JPhys Photonics*, 2021, **3**, 024013.
- 40 F. Bray, J. Ferlay, I. Soerjomataram, R. L. Siegel, L. A. Torre and A. Jemal, *Ca-Cancer J. Clin.*, 2018, **68**, 394–424.
- 41 M. T. Doherty, O. B. Sanni, H. G. Coleman, C. R. Cardwell, W. G. McCluggage, D. Quinn, J. Wylie and U. C. McMenamin, *PLoS One*, 2020, **15**(4), e0232231.
- 42 J. Lortet-Tieulent, J. Ferlay, F. Bray and A. Jemal, *J. Natl. Cancer Inst.*, 2018, **110**(4), 354–361.
- 43 D. Cella, H. Huang, H. D. Homesley, A. Montag, R. Salani, K. D. Geest, R. Lee and N. M. Spirtos, *Gynecol. Oncol.*, 2010, **119**, 538–542.
- 44 Y. Xiao, L. Jin, C. Deng, Y. Guan, E. Kalogera, U. Ray, P. Thirusangu, J. Staub, S. S. Bhattacharya, H. Xu, X. Fang and V. Shridhar, *Oncogene*, 2021, **40**(8), 1409–1424.
- 45 T. Yao, R. Cao, W. Xiao, F. Pan and X. Li, *J. Biophotonics*, 2019, **12**(7), e201800443.
- 46 T. S. Gerashchenko, M. V. Zavyalova, E. V. Denisov, N. V. Krakhmal, D. N. Pautova, N. V. Litviakov, S. V. Vtorushin, N. V. Cherdyntseva and V. M. Perelmuter, *Acta Nat.*, 2017, **9**, 56–67.
- 47 A. Pasqualato, A. Palombo, A. Cucina, M. A. Marigliò, L. Galli, D. Passaro and M. Bizzarri, *Exp. Cell Res.*, 2012, **318**(7), 835–846.
- 48 Y. Kanintronkul, R. Worayuthakarn, N. Thasana, P. Winayanuwattikun, K. Pattanapanyasat, R. Surarit and J. Svasti, *Anticancer Res.*, 2011, **31**(3), 921–927.
- 49 N. A. Puspita and A. Bedford, *Int. J. Health Sci.*, 2017, **5**(1), 8–14.
- 50 R. Domura, R. Sasaki, Y. Ishikawa and M. Okamoto, *J. Funct. Biomater.*, 2017, **8**(2), 18.
- 51 A. Isozaki, J. Harmon, Y. Zhou, S. Li, Y. Nakagawa, M. Hayashi, H. Mikami, C. Lei and K. Goda, *Lab Chip*, 2020, **20**(17), 3074–3090.
- 52 Y. Li, A. Mahjoubfar, C. L. Chen, K. R. Niazi, L. Pei and B. Jalali, *Sci. Rep.*, 2019, **9**, 11088.
- 53 A. Isozaki, H. Mikami, K. Hiramatsu, S. Sakuma, Y. Kasai, T. Iino, T. Yamano, A. Yasumoto, Y. Oguchi, N. Suzuki, Y. Shirasaki, T. Endo, T. Ito, K. Hiraki, M. Yamada, S. Matsusaka, T. Hayakawa, H. Fukuzawa, Y. Yatomi, F. Arai, D. Di Carlo, A. Nakagawa, Y. Hoshino, Y. Hosokawa, S. Uemura, T. Sugimura, Y. Ozeki, N. Nitta and K. Goda, *Nat. Protoc.*, 2019, **14**(8), 2370–2415.
- 54 Y. J. Heo, D. Lee, J. Kang, K. Lee and W. K. Chung, *Sci. Rep.*, 2017, **7**, 11651.
- 55 K. He, G. Gkioxari, P. Dollar and R. Girshick, in *Proceedings of the IEEE International Conference on Computer Vision (ICCV)*, 2017, pp. 2961–2969.
- 56 K. He, X. Zhang, S. Ren and J. Sun, in *Proceedings of the IEEE conference on computer vision and pattern recognition*, 2016, pp. 770–778.
- 57 “opencvtoolkit/cvat: Powerful and efficient Computer Vision Annotation Tool (CVAT).” <https://github.com/opencvtoolkit/cvat> (accessed Jul. 23, 2021).
- 58 T.-Y. Lin, M. Maire, S. Belongie, J. Hays, P. Perona, D. Ramanan, P. Dollár and C. Lawrence Zitnick, in *European conference on computer vision*, 2014, pp. 740–755.
- 59 G. Hinton, “Neural Networks for Machine Learning. Lecture 6a: Overview of mini-batch gradient descent,” Coursera. <https://www.cs.toronto.edu/~hinton/coursera/lecture6/lec6.pdf> (accessed Jul. 29, 2021).
- 60 P. Baldi, in *Proceedings of ICML Workshop on Unsupervised and Transfer Learning*, PMLR, 2012, vol. 27, pp. 37–49.
- 61 I. Goodfellow, Y. Bengio and A. Courville, *Deep Learning*, MIT Press, Cambridge, MA, USA, 2016.
- 62 W. Liu, Z. Wang, X. Liu, N. Zeng, Y. Liu and F. E. Alsaadi, *Neurocomputing*, 2017, **234**, 11–26.
- 63 F. Bardozzo, T. Collins, A. Forgione, A. Hostettler and R. Tagliaferri, *Med. Image Anal.*, 2022, **77**, 102380.
- 64 F. H. G. Zuluaga, F. Bardozzo, J. I. R. Patino and R. Tagliaferri, in *2021 43rd Annual International Conference of the IEEE Engineering in Medicine & Biology Society (EMBC)*, 2021, pp. 3483–3486.
- 65 D. P. Kingma and J. Ba, *arXiv*, 2014, preprint, arXiv:1412.6980, DOI: [10.48550/arXiv.1412.6980](https://doi.org/10.48550/arXiv.1412.6980).
- 66 L. van der Maaten and G. Hinton, *J. Mach. Learn. Res.*, 2008, **9**, 2579–2605.

