



Cite this: *J. Mater. Chem. C*, 2021, 9, 11619

Colorimetric metasurfaces shed light on fibrous biological tissue

Zaid Haddadin,[†] Trinity Pike,[†] Jebin J. Moses and Lisa V. Poulikakos *

Fiber-affecting diseases – encompassing fibrosis, heart disease, neurological disease and cancer – are directly linked to the density and reorganization of fibrous media in biological tissue. Polarized light has unique capabilities to probe the structural reorganization of fibrous biological media, yet naturally occurring anisotropic light–matter interactions lie below thresholds relevant for disease visualization. Nano-optical metasurfaces have demonstrated immense capabilities to enhance light–matter interactions. However, beyond the initial focus of the field on energetic enhancement, the study of nano-optical materials which selectively enhance anisotropic light–matter interactions is at its infancy. This perspective highlights how colorimetric metasurfaces, with abilities to quantitatively map the polarization state of light onto structural color, hold great potential to visualize disease in fibrous biological tissue. We first discuss a range of major fiber-affecting diseases, where the role of tissue microstructure in disease progression has been characterized with polarized light, with a focus on Mueller Matrix polarimetry. Subsequently, an overview of state-of-the-art polarization-sensitive colorimetric metasurfaces in plasmonic, dielectric, hybrid and tunable platforms is presented. The insights provided forge a path to bridge the fields of tissue polarimetry and colorimetric metasurfaces for miniaturized, all-optical and label-free disease visualization enabled by the nanoscale science of light.

Received 2nd May 2021,
Accepted 5th August 2021

DOI: 10.1039/d1tc02030g

rsc.li/materials-c

Introduction

The presence and orientation of fibrous media in biological tissue is indicative of the origin and progression of a range of serious diseases, affecting organs across the human body from

fibrosis^{1,2} to heart disease^{3,4} and neurological disease^{5,6} or various cancers.^{7–11} While current practice in the clinic and biomedical industry has established technologies to understand and characterize disease on the cellular, genetic and molecular level,¹² the visualization of tissue microstructure – a critical indicator of fiber-affecting diseases – remains challenging.^{13,14}

Polarized light interacts selectively with anisotropic matter, including fibrous biological tissue, where the presence and organization of fibrous media, *e.g.* ordered structural proteins

Department of Mechanical and Aerospace Engineering, UC San Diego, 9500 Gilman Drive, La Jolla, California, 92093-0411, USA. E-mail: lpoulikakos@eng.ucsd.edu

[†] Authors contributed equally.



Zaid Haddadin

Zaid Haddadin is a graduate student in Electrical Engineering at the University of California, San Diego. He holds a BSc degree in Computer Science and Biology from Purdue University. As part of the Poulikakos Lab, Zaid is working towards his interests in medical diagnostics by designing, optimizing, and fabricating nanophotonic surfaces for the purposes of detecting fiber-affecting diseases and discriminating between diseased & healthy tissue states.



Trinity Pike

Trinity Pike holds a BSc degree in Mechanical Engineering from the University of California, San Diego. She has previously served as an undergraduate research assistant for Project Lim[b]itless, an initiative that aims to make prosthetic limbs more accessible and affordable. Her work with the Poulikakos Lab investigates the optics and spectroscopy of nanostructures and organic materials.

in the extracellular matrix, distinctly alters the polarization state of light.^{15,16} Polarized light thus presents a unique opportunity to selectively visualize the microstructural properties of fibrous biological tissue, with potential to elucidate the origin and progression of fiber-affecting diseases in a label-free, non-invasive manner.

However, the inherent weakness of anisotropic optical properties in fibrous biological tissue has necessitated sophisticated optical configurations to probe and analyze the selective interaction with polarized light, as simpler methods such as polarized light microscopy face challenges in sensitive and quantitative readout.¹⁷ Examples of existing technologies include confocal microscopy,^{18–20} second harmonic generation (SHG) microscopy,^{21,22} polarization-sensitive optical coherence tomography (PS-OCT),^{23–25} or Stokes and Mueller Matrix polarimetry,^{15,26,27} the last of which is discussed in detail in this work, as it fully quantifies the interaction of a medium of interest with polarized light. While these and other methodologies have made significant strides toward revealing the role of tissue microstructure in disease, the cost and complexity of the required optical components and configurations, including modulated optoelectronics, lasers and extensive data post-processing, pose significant barriers to clinical and industrial applications.^{21,28} Indeed, hand-held devices in PS-OCT²⁹ and Mueller Matrix polarimetry^{30–32} have made steps toward clinical translation, yet their underlying experimental design faces the same challenges as their benchtop counterparts.²⁸

Here we provide a perspective on how the nanoscale science of light delivers uniquely suitable tools to address the challenge of polarization-sensitive imaging in fibrous biological tissue. In particular, nano-optical metasurfaces, composed of sub-wavelength-periodically arranged nanostructures with versatile

optical properties, have been demonstrated to scale down and enhance the complex manipulation of light on a single optical surface, thus enabling a host of miniaturized, mobile and translatable applications.^{33–35}

The concentrated electromagnetic near fields generated by optical nanostructures exhibit high sensitivity to changes in the refractive index environment, which has been leveraged for a variety of biosensing applications,^{36–38} while at infrared frequencies surface-enhanced spectroscopy enabled by optical nanoantennas has achieved ultrasensitive monitoring of molecular composition.^{39–41}

Nano-optical metasurfaces have additional abilities to selectively detect the structural properties of biological matter. The chirality of biomolecules, embodied in the left- and right-handed arrangement of their molecular components, has been detected with enhanced sensitivity and selectivity by means of metasurfaces which enhance chiral light–matter interactions either by their chiral geometric configuration,^{42–53} or their distinct interaction with chiral light.^{54–58}

Expanding on this concept, metasurfaces have been designed to exhibit material anisotropy in the geometry or assembly of their nanoresonator components, resulting in full phase and polarization control of incident light,^{59,60} while further work demonstrated full Stokes polarimetry on a single optical metasurface,^{61–64} paving the way toward a host of miniaturized bioimaging applications.^{65–67} While many of the abovementioned metasurfaces showcase abilities to probe light polarization with high sensitivity and selectivity, their readout remains encoded in the physics of polarimetry, posing barriers to the extraction of the material properties of interest and subsequently to widespread clinical and industrial applications.



Jebin J. Moses

Jebin J. Moses holds a MSc degree in Nanoengineering from the University of California, San Diego. He previously earned his BTech degree in Chemical Engineering from Anna University. In the Poulidakos Lab, his research interests include the design and nanofabrication of polarization-tunable, colorimetric, dielectric metasurfaces for tissue imaging applications.



Lisa V. Poulidakos

Lisa V. Poulidakos is an Assistant Professor in Mechanical and Aerospace Engineering at UC San Diego. Her research furthers the science of nanophotonic materials to controllably enhance, probe, and influence nanoscale light–matter interactions. She received her PhD from ETH Zürich, where she introduced an original theoretical and experimental technique enabling rational design of chiral nanophotonic systems. Her post-

doctoral research at Stanford University focused on developing functional nanophotonic surfaces for cancer tissue diagnostics. She received the ETH Medal for outstanding doctoral theses, the L'Oréal USA For Women in Science Postdoctoral Fellowship, the Swiss National Science Foundation Early Postdoc Mobility Fellowship and she is a Fellow of the RCSA Sciolog on Advancing BioImaging.

To address this challenge, this perspective focuses on a further, formative property of nano-optical metasurfaces: their ability to generate structural color, which to date has been proposed for applications in display technology,^{68,69} miniaturized color printing^{70,71} or refractive-index sensing.^{68,72} In the context of biomedical imaging, colorimetry is arguably the most intuitive and translatable optical sensing metric, circumventing the need for spectroscopic analysis.

Here, we outline why structural color in metasurfaces has great potential for polarization-sensitive imaging of fibrous biological tissue. We first introduce the basic quantitative principles of polarized light–matter interactions, focusing on Mueller Matrix formalism. Subsequently, we discuss a range of representative fiber-affecting diseases, applicable to organs across the human body (Fig. 1). We highlight how the tissue microstructural properties of various disease states have been characterized in terms of their interaction with polarized light by Mueller Matrix polarimetry, thus showcasing how polarization-sensitive analysis provides critical information on disease state and progression. An overview of polarization-sensitive colorimetric metasurfaces is then presented, highlighting progress in this field for metallic, dielectric, hybrid and tunable material platforms. Finally, we discuss the opportunities and challenges of bridging these two fields where colorimetric, polarization-sensitive metasurfaces provide an accessible, miniaturized and translatable platform to visualize fiber-affecting disease.

Quantifying polarized light–matter interactions

Stokes parameters

The polarization state of light can be quantified by the Stokes parameters, written in vector form as

$$\mathbf{S} = \begin{pmatrix} S_0 \\ S_1 \\ S_2 \\ S_3 \end{pmatrix} = \begin{pmatrix} E_{0x}^2 + E_{0y}^2 \\ E_{0x}^2 - E_{0y}^2 \\ 2E_{0x}E_{0y}\cos(\phi) \\ 2E_{0x}E_{0y}\sin(\phi) \end{pmatrix} \quad (1)$$

where $E_{0x(y)}$ represents the amplitude of the incident electric field in x and y orientation, and $\phi = \phi_x - \phi_y$ represents the phase difference between x and y components of the electric field. The zeroth Stokes parameter, S_0 , describes the total intensity of the optical beam and the first Stokes parameter, S_1 , represents the differential intensity of horizontally *vs.* vertically polarized light. The differential intensity of linearly polarized light at $\pm 45^\circ$ is quantified by the second Stokes parameter, S_2 . Finally, the third Stokes parameter, S_3 , describes the differential intensity of left- and right-handed circularly polarized light.^{26,73}

Mueller matrix

The linear interaction between polarized light and a medium of interest is fully quantified by means of the Mueller Matrix

$$\mathbf{M} = \begin{pmatrix} m_{11} & m_{12} & m_{13} & m_{14} \\ m_{21} & m_{22} & m_{23} & m_{24} \\ m_{31} & m_{32} & m_{33} & m_{34} \\ m_{41} & m_{42} & m_{43} & m_{44} \end{pmatrix} \quad (2)$$

a 4×4 matrix which relates the input and output polarization state of light, represented by the Stokes' vectors \mathbf{S}_{in} and \mathbf{S}_{out} , as^{26,27,74}

$$\mathbf{S}_{\text{out}} = \mathbf{M}\mathbf{S}_{\text{in}}. \quad (3)$$

In complex, optically anisotropic media such as biological tissue, the Mueller Matrix elements simultaneously embody a variety of phenomena which can pose challenges to their interpretation.²⁷ Thus, Mueller Matrix decomposition methods have been successfully implemented to facilitate physical interpretation, as described below.

Mueller Matrix Polar Decomposition (MMPD)¹⁵ describes the breakdown of the Mueller Matrix into three constituent matrices that describe the physical phenomena of depolarization (\mathbf{M}_A), diattenuation (\mathbf{M}_D), and retardance (\mathbf{M}_R)^{15,75–77} as

$$\mathbf{M} = \mathbf{M}_A\mathbf{M}_D\mathbf{M}_R \quad (4)$$

all of which have relevant occurrence in biological tissue. Depolarization represents a decrease in the degree of polarization of light upon traversing a medium, quantified by the net depolarization factor $\Delta \in [0,1]$, where $\Delta = 1$ corresponds to fully polarized light.⁷⁸ In biological tissue, depolarization commonly occurs due to multiple scattering within the turbid medium, an effect which increases with penetration depth, becoming particularly significant for thicknesses above 0.5 mm.^{79,80}

Decreases in the degree of linear and circular polarization, can be directly obtained from the Mueller Matrix elements as^{77,81}

$$\text{LDOP} = \frac{m_{21} + m_{22}}{m_{11} + m_{12}} \quad (5)$$

$$\text{CDOP} = \frac{m_{41} + m_{44}}{m_{11} + m_{14}} \quad (6)$$

where LDOP and CDOP denote the linear and circular degree of depolarization, respectively. Accordingly, the linear depolarization is written as $1 - \text{LDOP}$, while $1 - \text{CDOP}$ is the circular depolarization.

Diattenuation occurs when orthogonal polarization states exhibit a differential attenuation upon interaction with a medium, which can be defined accordingly for orthogonal linear or circular polarization states. From the first row of the Mueller Matrix (eqn (2)) the diattenuation value can be obtained as⁷⁷

$$D = \frac{1}{m_{11}}\sqrt{m_{12}^2 + m_{13}^2 + m_{14}^2}. \quad (7)$$

Retardance describes a phase shift occurring between orthogonal polarization states of light, where δ denotes




Fig. 1 A visual summary of how disease states can be differentiated and determined using polarimetric methods. Parts (a) through (m) correspond to data entries in Table 1, which summarizes the experimental setup specifications for the following diseases: (a) amyloid fibrils in the retina,⁹⁰ (b-1) lung cancer,⁷⁷ (b-2) lung adenocarcinoma and lung squamous carcinoma,⁷⁷ (c) coronary plaque collagen,²⁴ (d) Crohn's disease and gastrointestinal luminal tuberculosis,⁸⁶ (e) giant cell tumor of the bone,⁹¹ (f) brain tumor,⁹² (g) papillary thyroid carcinoma,⁷⁵ (h) breast ductal carcinoma,⁹³ (i) liver fibrosis,⁸⁵ (j) colon adenocarcinoma,⁹⁴ (k) basal cell carcinoma,⁷⁵ (l) cervical cancer,⁸⁷ (m) prostate cancer.⁸³ Created with BioRender.com.

linear retardance, also known as birefringence, and ψ denotes circular retardance, also known as optical rotation, while the optical axis orientation angle of the medium is written as θ .^{26,76} The total retardance $R \in [0, \pi]$ is obtained from

the Mueller Matrix and allows for further material classification.^{26,76}

Mueller Matrix Transformation (MMT)^{82,83} is a further method by which a physically-relevant set of parameters, which

provide information about the microstructural properties of anisotropic media, are extracted from the Mueller Matrix. The MMT parameters A , b , t and x are written as^{75,82,84–87}

$$A = \frac{2(m_{22} + m_{33})\sqrt{(m_{22} - m_{33})^2 + (m_{23} + m_{32})^2}}{(m_{22} + m_{33})^2 + (m_{22} - m_{33})^2 + (m_{23} + m_{32})^2} \in [0, 1] \quad (8)$$

$$b = \frac{m_{22} + m_{33}}{2} \quad (9)$$

$$t = \sqrt{\frac{(m_{22} - m_{33})^2 + (m_{23} + m_{32})^2}{2}} \quad (10)$$

$$x = \frac{1}{2} \arctan\left(\frac{m_{42}}{m_{34}}\right) \quad (11)$$

wherein A represents the degree of anisotropy and characterizes the order of alignment of fibrous media. Isotropic and anisotropic media yield A values of 0 and 1, respectively. The parameter b is related to the depolarization ability of bulk tissues while t is sensitive to the degree of anisotropy in scattering matter. Finally, x is related to orientation directions of aligned fibrous structures.

Measurement of the Mueller Matrix can be achieved with Mueller Matrix polarimetry. This has been realized in transmission and backscattering configurations and, in certain cases, has been integrated in commercial microscopes (Fig. 2(a)),^{86,88} where the polarization state of the incident light is controlled by a polarization state generator (PSG) and the output polarization state is measured by a polarization state analyzer (PSA), both of which are composed of a frequency-modulated element, *e.g.* a photo-elastic modulator (PEM), and polarizer arranged sequentially. The 16 Mueller Matrix elements are obtained in a series of measurements, where the input polarization state of light is varied between four polarization states and the four Stokes parameters are obtained by analysis of the output polarization state in each case.^{26,89–91}

Characterizing fibrous tissue with polarized light

This section highlights applications of polarimetric imaging methods for the detection of diseased states in fibrous biological tissue. The first section titled Characterizing Individual Disease States focuses on research that attempts to differentiate a diseased state from its healthy counterpart. The second section titled Differentiating Among Disease Stages highlights how Muller Matrix polarimetry has been used to describe the progression of a single diseased state. The third section titled Discriminating Between Disease States features polarimetric applications for discerning between different diseases that display similar diagnostic characteristics. Finally, the Miscellaneous Applications section showcases some unique applications of polarimetry beyond the categories of the previous sections. Relevant equations and variables used to quantify polarized light-matter interactions have been defined in the previous

section. Fig. 1 summarizes the diseases discussed in the context of their polarizing optical properties, while Table 1 provides an overview of the experimental specifications required for the presented research.

Characterizing individual disease states

Basal cell carcinoma (BCC) Fig. 1(k). Du *et al.* differentiated healthy and cancerous tissues affected by basal cell carcinoma (BCC) by using parameters from three different polarization imaging techniques: linear degree of polarization (LDOP), Mueller Matrix polar decomposition (MMPD), and Mueller Matrix transformation (MMT).⁷⁵

The backscattering Mueller Matrix of unstained BCC samples (Fig. 2(b)) measured nondiagonal elements to be close to zero, indicating isotropic optical properties which may be attributed to an absence of well-aligned fibers in cancerous regions.^{75,84,92} This indication was confirmed with measurements approaching zero for the linear retardance (δ) and the anisotropic MMT parameter A . Moreover, cancerous regions presented larger values of LDOP, D , and b than those in healthy regions, while the Δ value was smaller than in healthy regions.⁷⁵ These findings imply that cancerous regions of BCC tissues exhibit lower depolarization values than healthy regions.⁷⁵ As such, cancerous regions of BCC tissue samples can be characterized as less anisotropic and less depolarizing than their healthy tissue counterparts.

Cervical cancer (CEC) Fig. 1(l). An examination by Wang *et al.* of human cervical cancerous tissues (CEC) using an ordinary, commercially available light microscope with a polarization state generator (PSG) and analyzer (PSA) (Fig. 2(a)) showed that while cancerous regions are more birefringent, they exhibit a lower uniform directionality in their fibers than their healthy counterparts.⁸⁸ The purpose of this study was two-fold: (1) to show how a clinically-available microscope can be altered to measure the Mueller Matrix elements of a tissue sample, and (2) to differentiate the characteristic microstructural features of the examined cancerous tissues through measurements of the Mueller Matrix polar decomposition (MMPD) and Mueller Matrix transformation (MMT) parameters.⁸⁸

The δ and t values, obtained from MMPD and MMT, respectively, were observed to be smaller in cancerous regions of CEC tissue samples than in healthy regions, indicative of decreased birefringence.^{88,92} Moreover, parameters characteristic of the direction of aligned fibers were lower in cancerous regions than in healthy regions, implying that CEC tissues exhibit lower optical anisotropy than their healthy counterparts.^{84,88,93}

Colon adenocarcinoma (COA) Fig. 1(j). Ahmad *et al.* aimed to study whether Mueller Matrix polarimetry can be used to differentiate between *ex vivo* normal and adenocarcinoma colon tissue samples in the 425–725 nm visible spectral range by characterizing the Mueller Matrix polar decomposition (MMPD).⁹⁸ It was found that total (Δ) and linear depolarization (δ), diattenuation (D), and retardance (R) of the MMPD parameters were significantly higher for the adenocarcinoma colon samples than the healthy colon samples.⁹⁸




Fig. 2 (a) A Mueller Matrix microscope created by adding a polarization state generator (PSG) and analyzer (PSA) to a commercially available light microscope. (b) The normalized Mueller Matrix of basal cell carcinoma tissues.⁷⁵ (c) The m_{24} and m_{34} elements of the normalized Mueller Matrix of Crohn's disease and gastrointestinal luminal tuberculosis (ITB).⁸⁷ (d) The average experimental values of MMPD δ (blue squares) and MMT t (green circles) parameters in different stages of liver cancer.⁸⁶ (e) MMPD δ & θ and MMT t & x parameters of Crohn's disease and gastrointestinal luminal tuberculosis. The left image of any parameter is of a Crohn's disease sample. The right image of any parameter is a gastrointestinal luminal tuberculosis sample. (f) Box plots of an imaging analysis technique performed on 10 Crohn's disease (CD) samples (red dots) and 10 gastrointestinal luminal tuberculosis (ITB) samples (blue dots).⁸⁷ (g) Mueller Matrix plot for healthy lung cells (red), lung squamous carcinoma cells (green), lung adenocarcinoma cells (blue) and a 50/50 mixture of squamous carcinoma & adenocarcinoma cells (purple).⁷⁷ (b–d) Reprinted (adapted) from ref. 75, 86 and 87 with permission from SPIE Digital Library. (g) Reprinted/adapted from ref. 77 with permission from Elsevier.

These results are in accordance with previous studies showing how alterations to the adenocarcinoma tissue architecture result in malignant glands of varying sizes and disorganized fibrous stroma. These physical changes in turn increased the

depolarization and affected the anisotropic properties of diseased colon tissues.^{93,98–100} Using statistical and algorithmic analysis,^{98,101–103} depolarization and retardance were determined as good candidates for the discrimination of the normal

Table 1 The experimental polarimetric setup specifications for different disease states

Tissue sample	Thickness (mm)	Light source	Wavelength (nm)	Power (mW)	Experimental geometry	Available data in literature
a Retina ⁹⁴	—	—	—	—	—	Intensity map: linear retardance & linear anisotropy
b Normal lung fibroblast ⁷⁷	0.005	Laser diode	1065	—	Backscattering	Numerical values: NMM
b Lung squamous cell carcinoma (LUSC) ⁷⁷	0.005	Laser diode	1065	—	Backscattering	Numerical values: NMM
b Lung adenocarcinoma (LUAC) ⁷⁷	0.005	Laser diode	1065	—	Backscattering	Numerical values: NMM
b Mixture of LUSC & LUAC ⁷⁷	0.005	Laser diode	1065	—	Backscattering	Numerical values: NMM
d Crohn's disease ⁸⁷	0.012	LED	633	3000	—	Intensity map: MM, MMPD, & MMT
d Gastrointestinal luminal tuberculosis ⁸⁷	0.012	LED	633	3000	—	Intensity map: MM, MMPD, & MMT
e Giant cell tumor of bone ⁹⁵	—	Halogen lamp	—	—	—	Intensity map: MM
f Human brain ⁹⁶	60 × 90 × 10	—	550	—	Reflection	Intensity map: depolarization & retardance
f Calf brain ⁹⁶	90 × 70 × 10	—	550	—	Reflection	Intensity map: depolarization & retardance
g Papillary thyroid carcinoma ⁷⁵	0.028	LED	650	3000	Backscattering	Intensity map: NMM, LDOP, MMPD, & MMT
h Breast ductal carcinoma ⁹⁷	0.012	LED	632	3000	Backscattering	Intensity map: MMPD & MMT
i Liver fibrosis ⁸⁶	0.008	LED	632	3000	Backscattering	Intensity map: MMPD & MMT
j Colon adenocarcinoma ⁹⁸	0.003	Xenon lamp	425	150	—	Numerical values: MM & MMPD
k Basal cell carcinoma ⁷⁵	0.028	LED	650	3000	Backscattering	Intensity map: NMM, LDOP, MMPD, & MMT
l Cervical cancer ⁸⁸	0.02	LED	632	3000	Backscattering	Intensity map: MMPD & MMT
m Prostate cancer ⁸³	—	Laser diode	632.8	—	Backscattering	Intensity map: NMM & MMPD

The letters in the leftmost column correspond to the disease state described in Fig. 1. Abbreviations: LDOP: linear degree of polarization; MM: Mueller Matrix; MMPD: Mueller Matrix polar decomposition; MMT: Mueller Matrix transformation; NMM: normalized Mueller Matrix.

and adenocarcinoma colon tissues.⁹⁸ Normal and adenocarcinoma colon samples were algorithmically characterized with 100% accuracy, sensitivity, and specificity for total (A) and linear (δ) depolarization as well as retardance (R).⁹⁸ Through this work, Ahmad and colleagues showed how the Mueller Matrix can be used in combination with statistical and algorithmic analysis for the identification and classification of normal and adenocarcinoma colon tissue samples, permitting for future research to utilize this application toward automated and non-invasive cancer diagnosis.⁹⁸

Giant cell tumor of bone (GCTB) Fig. 1(e). Giant cell tumor of bone (GCTB) diagnosis has low efficiency as it depends on experience of histopathologists and requires a triple diagnosis with input from orthopedic surgeons, radiologists, and pathologists.^{95,104,105} To combat this drawback, Zhao *et al.* developed a deep learning model¹⁰⁶ that jointly uses histopathological samples, Mueller Matrix polar decomposition (MMPD) parameters and Mueller Matrix transformation (MMT) parameters to predict whether a histological bone sample is healthy or contains a cancerous lesion.⁹⁵

When characterizing images of cancerous lesions in histological bone samples, the deep-learning model showed the highest accuracy of ~96% when compared to characterization through MMT (~91% accuracy) or MMPD (~73% accuracy).⁹⁵ Through this research effort, Zhang and colleagues demonstrated a proof-of-concept for the use of deep learning models in the detection of cancerous *versus* healthy regions of tissues.⁹⁵ Moreover, their ability to quantify how well MMT and MMPD characterize the differences in diseased and healthy states elucidates how, despite the abundance of information contained in both types of Muller Matrix decomposition, one polarization imaging technique may be more sensitive and

accurate to a certain disease state than the other.⁹⁵ For example, MMT is said to be orientation-insensitive, potentially explaining its improved performance with respect to MMPD in the study of Zhang *et al.*^{84,88,95}

Papillary thyroid carcinoma (PTC) Fig. 1(g). Polarimetric parameters obtained from the linear degree of polarization (LDOP), Mueller Matrix polar decomposition (MMPD), and Mueller Matrix transformation (MMT) were used to characterize and differentiate papillary thyroid carcinoma (PTC) tissues from their healthy tissue counterparts.⁷⁵ Du *et al.* utilized the backscattering configuration of Mueller Matrix polarimetry to obtain their results.⁷⁵

The Mueller Matrix of unstained PTC samples measured nondiagonal elements to be non-negligible, indicative that the PTC tissue samples are optically anisotropic.^{75,84} This conclusion was further confirmed by Mueller Matrix analysis where LDOP, the MMT parameter b and the MMPD parameters did not provide an indication of the cancerous region. However, the MMT parameter A revealed the cancerous region.⁷⁵ This finding suggested the existence of well-aligned structures at the cancerous regions of the tissues, a result that is in conjunction with previous research showing that a morphological feature of PTC is fibroplasia, the development of fibrous structures.^{75,107} Moreover, it highlights that the MMT parameter A may be used as a discriminatory variable to differentiate between cancerous and healthy regions in tissues afflicted by papillary thyroid carcinoma.⁷⁵

Prostate cancer (PRC) Fig. 1(m). Backscattered Mueller Matrix imaging was utilized by Badieyan *et al.* to extract the characteristic optical features of bulk prostate tissues retrieved immediately after surgery. The aim of this study was to quantify a difference between prostate cancer (PRC) tissues and healthy

prostate tissues through measurements of the Mueller Matrix polar decomposition (MMPD) parameters and the Mueller Matrix transformation (MMT) parameters, alongside statistical analysis. The obtained MMPD and MMT values can be seen in Table 2.

The cancerous prostate tissues displayed higher values for the MMPD diattenuation (D) and depolarization (A) parameters than their healthy counterpart. Lower values were obtained for the MMPD retardance (R) and the MMT A & b parameters of cancerous samples than healthy prostate tissue.⁸³ These values are in accordance with the rapid growth of cancerous tissues leading to a reduction in the order of alignment of the fibers. The subsequent lowering of the optical anisotropy is seen in the reduced MMPD retardance (R) and MMT A values. An increased absorption of light, leading to higher diattenuation, then follows with an increased cell density.^{83,108} Overall, these results indicated how changes in optical anisotropy can be powerful differentiators of healthy and cancerous prostate tissues, providing distinct information on the role of microstructure in disease progression.⁸³

Differentiating among disease stages

Breast ductal carcinoma (BDC) Fig. 1(h). The gold standard for breast cancer diagnosis has relied on histopathological observations that depend on the experience of pathologists.^{105,109} However, previous studies have shown that the progression of breast cancer affects the collagen structure around ducts and tumors in intact, unfixed, and non-sectioned mammary glands.⁸ In their study, Dong *et al.* characterized the progression of invasive breast ductal carcinoma (BDC) across disease stages 1 through 3 by segmenting the ductal areas in images of the Mueller Matrix polar decomposition (MMPD) and measuring their textural features.^{97,110}

Dong and colleagues found, when compared to stage 1 cancerous tissues, stage 2 cancerous tissues showed MMPD linear retardance (δ) images with higher texture contrast and lower texture uniformity outside the ductal region. Inside and outside the ductal areas, images of the MMPD orientation (θ) in stage 2 tissue showed decreased texture contrast and increased linear dependence of grey tones across neighboring pixel areas.⁹⁷ For stage 3 tissue, the MMPD δ parameter exhibited an increase in texture contrast inside the duct and an increase in texture uniformity outside the duct, in comparison to stage 2. Additionally, images of the MMPD θ parameter for stage 3 tissue showed an increase in texture contrast and a decrease in the linear dependence of grey tones outside the ducts.⁹⁷ This study

thus demonstrated how the Mueller Matrix can be assessed with an image textural analysis technique. The resulting quantitative characteristics aid pathologists to better characterize stages of progression in ductal breast cancer.⁹⁷

Liver fibrosis (LVF) Fig. 1(i). An examination of human liver cancer tissues by Wang *et al.* showed that cancerous regions were more birefringent with higher fibrous density than their healthy counterparts.⁸⁸ Their measurements of the Mueller Matrix polar decomposition (MMPD) and the Mueller Matrix transformation (MMT) parameters exhibited higher values in cancerous regions for the MMPD retardance (R) and orientation (θ) parameters as well as the MMT parameters t and x .^{86,88} The values of R and t characterize the linear retardance of the media while θ and x characterize the direction of the aligned fibers; thus implying the presence of anisotropic properties, such as birefringence, in the cancerous tissue.^{84,88,92}

The progression of liver cancer is accompanied by an increased development of fibrosis resulting from inflammatory reactions.^{86,111} As such, Wang *et al.* investigated the effects of this heightened fibrotic accumulation on the MMPD and MMT parameters across the four stages of liver cancer (F1, F2, F3, & F4): the orientation of the fibers, as described by the MMPD θ and MMT x values, remained relatively consistent across the four stages. In contrast, the R and t parameters showed a marked increase from F1 to F4⁸⁶ (Fig. 2(d)). The R parameter of MMPD changed more prominently than the t parameter of MMT, implying that MMPD may be more sensitive to the birefringence of a sample.⁸⁶ Nonetheless, through their study, Wang *et al.* showed that liver fibrosis disease stages can be differentiated by Mueller Matrix polarimetry.

Discriminating between disease states

Crohn's disease (CRD) vs. gastrointestinal luminal tuberculosis (ITB) Fig. 1(d). There remains a need to find a clinically distinguishing factor for the diagnosis of Crohn's disease and intestinal tuberculosis, which exhibit similar radiological, endoscopic, and histological features.¹¹² As no single pathognomonic trait exists for either disease, misdiagnosis can result in inappropriate treatments with side effects that risk toxicity to the patient and a delay in proper treatment. Thus, it becomes imperative to find a standard by which to discriminate between both diseases.^{87,112} Liu *et al.* applied Mueller Matrix polar decomposition (MMPD) and Muller Matrix transformation (MMT) imaging techniques to Crohn's disease (CRD) and gastrointestinal luminal tuberculosis (ITB) tissue samples. The normalized Mueller Matrix of their samples is shown in

Table 2 MMPD & MMT values of cancerous prostate tissues and healthy prostate tissues

Sample	Diattenuation	MMPD		MMT			
		Retardance	Depolarization	A	b	t	x
Cancerous prostate tissue	0.7419	0.1648	0.8864	0.2050	0.0955	—	—
Healthy prostate tissue	0.7176	0.5094	0.8636	0.3140	0.1565	—	—

Cancerous prostate tissues have higher MMPD diattenuation and depolarization values. Healthy prostate tissues have higher MMPD retardance and MMT A & b values. MMT values x and t were not provided in the literature.

Fig. 2(c). They determined that birefringence is a dominant feature of both diseases by their retardance (MMPD parameter R and MMT parameter t) and orientation (MMPD parameter θ and MMT parameter x)⁸⁷ (Fig. 2(e)). It was previously shown that the MMPD parameters are more sensitive to birefringence than MMT parameters.¹¹³

As such, Liu and colleagues applied an image processing method to the MMPD linear retardance δ and orientation θ to quantitatively describe the samples across six physical and textural features (Fig. 2(f)): coarseness, contrast, directionality, line-likeness, regularity, and roughness.^{87,114} These indicated that fibrous structures in ITB tissues exhibit approximately the same orientation (line-likeness) and are of higher imaging quality (contrast) than CRD tissues. In contrast, fibrous structures in CRD tissues showed a larger tendency towards a certain direction (directionality). This supported the finding that ITB tissues have smaller, more aligned fibers while CRD tissues contain longer non-uniformly distributed fibers.^{87,112,114} These preliminary results indicate how the Mueller Matrix can be used to discriminate between two pathologically similar diseases with potential to provide quantitative information to clinical diagnosis.⁸⁷

Lung squamous carcinoma (LUSC) vs. lung adenocarcinoma (LUAC) Fig. 1(b). Shrestha *et al.* aimed to incorporate the Mueller Matrix in a label-free, polarimetric imaging technique at near-infrared wavelengths for the classification of different types of histopathological lung cancer samples.⁷⁷ Specifically, they utilized an imaging modality technique known as Whole Slide Imaging, wherein remote scanning of conventional histopathological samples on glass slides produces digital images for clinical examination. Whole Slide Imaging is believed to lead to increased efficiency and improved patient outcomes if implemented across the healthcare industry.^{77,115} The authors introduced new optical polarimetric quantities to existing digital data analysis techniques.^{77,116–118} The four examined histopathological samples included (1) healthy fetal lung fibroblast epithelial cells, (2) human stage II squamous carcinoma cells (LUSC), (3) human stage II adenocarcinoma cells (LUAC), and (4) a 50%:50% mixture of the human stage II squamous carcinoma cells and of the human stage II adenocarcinoma cells.⁷⁷

A normalized intensity plot of the Mueller Matrix elements of histopathological samples (1)–(4) establishes the diagonal matrix elements as the best discriminators⁷⁷ (Fig. 2(g)). Overall, the diattenuation (D) and depolarization (Δ) values of LUAC cells were higher than LUSC cells. Further, the D and Δ values of LUSC cells were higher than for healthy cells.⁷⁷ Additional polarimetric analysis showed an increase in backscattered polarized intensities from healthy cells to LUSC cells to LUAC cells.⁷⁷ This increase in polarized intensity is correlated to an increase in ratio of nucleus-to-cytoplasmic sizes of the cells – from healthy to LUSC and from LUSC to LUAC. A larger ratio contributes an overall more complex morphology. This trend correlates with the fact that cancerous cells exhibit a higher depolarization of light than healthy cells.^{77,93,119,120} This study indicates how the Mueller Matrix along with other optical

polarimetric tools can be utilized for both the discrimination of healthy and cancerous lung tissues as well as the differentiation between types of lung cancers.⁷⁷

Miscellaneous applications

Amyloid detection in the retina Fig. 1(a). The accumulation of amyloid plaques and neurofibrillary tangles are hallmark symptoms of Alzheimer's disease. However, although their accumulation begins before symptoms of cognitive impairment appear, a definitive diagnosis based on these two symptoms can only be obtained post-mortem.^{94,121} Early detection of these hallmark symptoms may be needed for the development of novel treatments. To date, no seemingly effective cure exists for patients who have developed cognitive deficiencies due to Alzheimer's disease.^{94,122,123} Qiu *et al.* demonstrated a promising route toward early detection of Alzheimer's disease by utilizing machine learning methods to detect amyloid presence in the retina as a biomarker for amyloid build-up in the brain.^{94,124} They utilized three different machine learning algorithms for this purpose,^{94,125–127} where Mueller Matrix polarimetry was incorporated to distinguish the presence of amyloid deposits.

Qiu and colleagues performed post-mortem analyses to determine whether detection of amyloid deposits required the fluorescent marker thioflavin.⁹⁴ Their findings indicated that it may be possible to distinguish between fluorescent and non-fluorescent retinas by their polarizing optical properties (*i.e.*, linear anisotropy and linear retardance), which implies that the ordered arrangement of the fibrils may be key to distinguishing between them.^{84,93,94} Moreover, a previous study has shown that *in vivo* detection of retinal amyloid deposits stained with curcumin is possible as a proof-of-concept imaging trial in Alzheimer's disease.¹²⁸ The data presented by Qiu *et al.* provide a step towards the development of non-invasive, dye-free clinical methods of imaging retinal amyloid deposits as a predictor of amyloid build-up in the brain.^{94,121}

Assessment of coronary plaque collagen for acute coronary syndromes Fig. 1(c). Optical coherence tomography (OCT) operates analogously to ultrasound imaging but uses light, instead of sound, to obtain higher image resolution by measuring the back reflection of near-infrared light at the micron scale.^{24,129} When expanded to polarization sensitive OCT (PS-OCT), the birefringence of tissues can be measured.^{23,24,129} Giattina *et al.* used this imaging technique for *in vivo* assessment of coronary plaque collagen to detect vulnerable artery plaques that may lead to acute coronary syndromes.^{24,130}

Giattina and colleagues showed that PS-OCT has significant correlation in identifying birefringent properties of collagen in comparison to Picrosirius staining, a dye used for assessment of organized collagen.^{24,131} Moreover, it was shown that PS-OCT images can correlate the concentration of fibers with the type of fibers: type I fibers, which are denser & larger, show higher birefringence than the smaller type III fibers.^{24,132} The difference in birefringence and resultant appearance of type I and type III fibers was also observed when examining cardiac tissue under polarized light.³ Although only 60% of the light had diagnostic value due to partially polarized illumination,²⁴ a relation

between PS-OCT images and Picosirius stained images was identified regarding the build-up and presence of collagen in the coronary arteries. This demonstrated the capability of PS-OCT images to assess plaque collagen content, thus making an argument for its potential in the clinic.²⁴

White matter visualization during neurosurgical procedures

Fig. 1(f). The extent of surgical brain tumor removal is an important factor in determining the post-operative prognosis a patient may receive.^{96,133} However, no more than 36% of patients achieve full resection of their brain tumors.¹³⁴ Challenges lie in the difficulty of delineating the boundary between solid brain tumor and white brain matter during surgery, even with intraoperative microscopes. Incomplete removal can result when too little resection occurs, while too much resection can give rise to neurological deficits.^{96,134} Cancer invasion leads to substantial tissue remodeling that destroys the ordered structures of fibrillary tracts, characteristic of healthy white brain matter, resulting in a loss of optical anisotropy that can be used as a delineating factor.^{96,135} In their 2020 paper, Schucht *et al.* utilized wide-field Mueller polarimetry to detect healthy white matter by its anisotropic properties, thus revealing the location of tumor tissue.⁹⁶ The visualization of white brain matter was performed by deriving the azimuth of the tissue optical axis from Mueller Matrix polar decomposition (MMPD) images of thick specimens measured in reflection.⁹⁶

In fixed human brain tissue and fresh animal tissue, it was found that the orientation of fiber tracts in healthy brain white matter could be visualized by the azimuth of their optical axis, obtained from backscattered Mueller Matrix images.⁹⁶ This proof-of-concept study illustrates how Mueller Matrix polarimetry can be used in real time during surgical procedures for the detection of the tissue–tumor interface. Though Schucht *et al.* emphasized its use in neurosurgery, this technique could potentially be applicable to any fiber-affecting disease where a boundary in the tissue can be delineated with polarized light, such as those highlighted earlier in this perspective.

Polarization-sensitive colorimetric metasurfaces

Structural color is a phenomenon that occurs when nanostructures selectively scatter and absorb incident light due to variations in the index of refraction. Prime examples from nature include wings of the morpho butterfly^{136–138} and bird of paradise,¹³⁹ feathers of hummingbirds¹³⁸ and peacocks,¹³⁸ or the scales of jewel beetles¹⁴⁰ and peacock spiders.¹⁴¹ Compared to color arising from the absorption of light in chemical pigments, structural color offers vivid saturation, improved resolutions, and increased resistance to high temperatures.¹⁴²

Recent advances in nanotechnology have enabled researchers to emulate the effects observed in nature by realizing artificial structural color. A first glimpse into this world was demonstrated by colloidal assembly of so-called photonic pigments.¹⁴² The disordered arrangement of these colloidal particles can constitute photonic glasses which

generate angle-independent, non-iridescent structural color.^{143–145} In contrast, when optical nanostructures are arranged in ordered, periodic assemblies to form photonic crystals^{146,147} or metasurfaces, the symmetry or asymmetry of their geometric configuration enables additional degrees of tunability in the generation of structural color.¹⁴⁸ For example, structural colors enable inkless color printing through controlled crazing^{149,150} which leverages miniature material patterns in transparent polymers to produce color, and photonic crystals, which contain periodic structures that manipulate light to generate bright, tunable colors.¹⁵¹

In the following sections, we provide a broad overview of structural color response in nanostructures composed of plasmonic, dielectric and hybrid material platforms. Plasmonic metasurfaces generate color response from the resonant properties of surface plasmons in metallic nanostructures.^{148,152,153} Dielectric metasurfaces leverage the electric and magnetic dipolar resonances of high-refractive-index dielectric nanostructures or guided-mode-resonant properties for optical output.^{148,154–157} Hybrid systems utilize and combine the strengths of purely plasmonic and dielectric counterparts. After introducing each material platform, we highlight examples of corresponding polarization-tunable metasurfaces. For a comprehensive review of structural color in metasurfaces we refer the reader to prior articles.^{148,153,158,159}

Structural color in plasmonic nanostructures

Surface plasmons arise from the collective oscillation of the free-electron cloud at the surface of conductive materials.¹⁶⁰ Noble metal nanoparticles, with surface plasmon resonances (SPRs) in the visible wavelength range, are effective generators of structural color.¹³⁹ A diverse set of plasmonic nanostructures including nanoholes, nanogratings and metal–insulator–metal systems have been arranged in periodic lattices to form metasurfaces.^{141,161–164} The optical properties of plasmonic nanostructures exhibit a high sensitivity to their refractive index environment.^{152,165} The resulting tunability of SPRs has enabled a variety of applications in refractive-index-sensing which can be leveraged in biomedical applications.¹⁶⁶ With improving nanofabrication processes, enhancement in color saturation capabilities, and high customization, plasmonic nanostructures have a wide scope of applications which utilize structural color.¹⁴⁸

The SPR effect renders plasmonic nanostructures particularly suitable for high resolution color generation at the optical diffraction limit.¹⁴⁸ Kumar *et al.* achieved bright-field color prints with diffraction-limited resolution by raising silver (Ag) or gold (Au) nanodisks above a plasmonic backreflector film *via* nanopillars made of hydrogen silsesquioxane (HSQ). The resulting pixels of 250 nm pitch were able to generate individual colors.¹⁶⁷ Subsequent work by Tan *et al.* expanded on this system and demonstrated photorealistic color printing by varying nanodisk size, spacing and dual-size variations within a single pixel of 800 × 800 nm.¹⁶⁸ By utilizing aluminum (Al) as the plasmonic material, a low-cost, environmentally friendly,

and minimally lossy system was achieved in comparison to its Ag and Au counterparts.

Exploring various geometries of plasmonic structures has expanded possibilities for customization. For instance, configurations of nanohole arrays, where extraordinary optical transmission can hinder loss,¹⁶⁹ have been designed to control and enhance structural color output (Fig. 3(a)).^{160,170,171} Additionally, nanoclusters can be leveraged to achieve an engineered color response. To this end, King *et al.* fabricated ultraviolet and visible Fano resonant aluminum nanoclusters to colorimetrically sense changes in refractive index in the particle surroundings (Fig. 3(b)).¹⁷² Though most nanohole color filters are polarization insensitive due to the symmetric nature of the nanohole lattice,^{161,173} the work of Dai *et al.* showed an added benefit of utilizing a triangular lattice array, which enabled improved color purity and a resolution of 12 700 dpi.¹⁶¹

Polarization-sensitive plasmonic metasurfaces

Many metasurfaces are composed of symmetric nanoresonator unit-cell elements arranged in regular, periodic patterns to achieve vibrant color output. However, breaking symmetry in metasurface lattice or nanoresonator geometry can enable a polarization-sensitive color response.¹⁵⁹ In plasmonic nanostructures, polarization of the incident light directly impacts the excitation of localized surface plasmon resonances (LSPRs).¹⁷⁴ Several researchers have leveraged polarization-dependent

plasmonic nanostructures for colorimetric applications. For example, Goh *et al.* fabricated tunable color pixels with linear polarization dependence using nanoscale ellipses and squares. Different polarization directions revealed two sets of color patterns in the same nanostructure, one square-shaped and the other star-shaped.⁷¹ Similarly, Duempelmann *et al.* achieved directional structural colors with tunable, symmetry-breaking aluminum plasmonic nanowires. Due to the asymmetry, plasmon resonances were only excited by a certain direction of incident light. The structure appeared colorful in one angle of observation, but colorless in the opposite direction.¹⁷⁴

High aspect ratio geometries have also been implemented to achieve polarization-dependent color response. Duempelmann *et al.* developed a plasmonic phase retarder using silver nanowires. The nanowire geometry was optimized to decrease angle and orientation dependence and tune the LSPR effect into the visible spectrum. Rotating a polarizer displayed four distinct colors (Fig. 3(c)).¹⁷⁵ Olson *et al.* constructed vibrant plasmonic color pixels from Al nanorod arrays, which produced a spectrally narrow, polarization-tunable optical response (Fig. 3(d)).⁶⁸

Cross-shaped geometries have activated polarization-based phenomena in plasmonic metasurfaces. The work of Li *et al.* presented periodically arranged cross-shaped plasmonic nanoapertures fabricated in an aluminum film, which exhibited a polarization-switchable color response in transmission.¹⁷⁶




Fig. 3 (a) Top-to-bottom: Schematic of color thin-film transistor by Heo *et al.* with color response by transmitted incident light shown to the left. SEM images of nanohole structure with scale bar 500 nm.¹⁷⁰ (b) Left-to-right: Theoretical spectra by King *et al.* of a nanocluster in air ($n = 1.0$), PDMS ($n = 1.4$), PMMA ($n = 1.5$), and diphenyl ether ($n = 1.6$). Schematic of refractive index sensing setup, with labeled angle of incidence and polarization of excitation light. Aluminum nanocluster with charge plot (negative, red; positive, blue) at the subradiant mode of the Fano resonance.¹⁷² (c) Left-to-right: Schematic of the angle variation of the incident polarizer (θ_1) and the analyzing polarizer (θ_2) in the experimental setup by Duempelmann *et al.* Color response for varying θ_2 angles, resulting in four different images of Marilyn Monroe.¹⁷⁵ (d) Top-to-bottom: Image of Al pixels with polarized image of pixel array (part A), image of the array with polarizer parallel to x-axis (part B), unpolarized image of pixels sorted according to D_y/D_x ratio and length (l) taken with scattering DSLR camera (part C). Bottom: Schematic showing the Al pixel unit cell by Olson *et al.* with labeled physical parameters (length l , width w , height h , and edge-to-edge spacing Γ). Diagram of excitation geometry for the pixel, where p-polarized white light propagates across the yz plane.⁶⁸ (e) Left-to-right: Transmission images of localized surface plasmon (LSP) pattern sample (above) and twisted chromatic plasmonic polarizers sample (below) by Ellenbogen *et al.* for varying angles of polarized light. SEM image of LSP sample showing that the background nanoantennas (region "2") are perpendicular to the letter ("LSP") nanoantennas (region "1").¹⁵² (a–c and e) Reprinted (adapted) with permission from ref. 152, 170, 172 and 175. Copyright 2020, 2015, 2016, 2012 American Chemical Society. (d) Copyright 2014 National Academy of Sciences ref. 68.

Ellenbogen *et al.* developed plasmonic optical nanoantennas with high sensitivity to polarized light. This system, termed chromatic plasmonic polarizers (CPP), exhibited a visible color response to birefringent chicken breast tissue and plastic foil due to their cross-shaped structure (Fig. 3(e)). Based on selective excitation of either the vertical or horizontal LSPR, incident light polarized along the short cross axis displayed a blue color, while light polarized along the long axis showed yellow.¹⁵² This finding offers key insights into biomedical applications, where nanophotonics solutions can contribute to tissue imaging without invasive staining procedures.

Material challenges in plasmonic metasurfaces

Plasmonic metasurfaces effectively produce structural colors. However, the resonance mechanism of each unit-cell nanostructure, material limitations, and production methods pose notable challenges. For instance, refractive-index-based LSPR sensing relies on specific environmental conditions and requires surface contact, which is not always possible in clinical settings.¹⁷² Plasmonic diffraction gratings produce vivid colors, yet they provide limited spatial resolution.¹⁵³ However, certain variations in the nanostructures have improved plasmonic resolution. For example, disk-hole coupling has achieved $\sim 100\,000$ dpi.¹⁴⁸ Nonetheless, fabrication of such nanostructures is challenging to replicate for large scale manufacturing.

Moreover, intrinsic material loss in plasmonic metasurfaces can significantly limit color saturation.^{177–179} The cost of fabrication and environmental degradation present further hurdles to high-throughput, durable applications. Typical plasmonic metals include gold, copper, silver, and aluminum. Gold and silver are too costly to be considered practical, while copper, silver, and aluminum exhibit rapid oxidation that renders them non-enduring candidates.^{180–183} Gold is more robust, yet does not produce blue light due to its strong transition band. Aluminum, which forms a thin, finite oxide layer upon environmental contact, is a low-cost, relatively stable plasmonic material. However, it exhibits notable loss due to its interband transition.¹⁸⁴ Though imperfect, aluminum has been frequently utilized by researchers as a lightweight material with favorable optical properties and could be an option for wide scale production in the future.^{148,158}

Structural color in dielectric nanostructures

All-dielectric nanostructures have been demonstrated as efficient generators of saturated structural color in comparison to their plasmonic counterparts. This is attributed to their low dissipative losses and high refractive index contrast with respect to the surrounding environment.^{41,185,186} When spatially confined in three dimensions to form subwavelength nanoparticles, these all-dielectric systems exhibit spectrally distinct electric and magnetic Mie resonances that are tuned by altering resonator geometry and material crystallinity.^{187,188}

Examples of high-refractive-index dielectric materials exhibiting strong scattering behavior include silicon (Si), gallium arsenide (GaAs), and gallium phosphide (GaP).¹⁸⁹ Further systems which exhibit low losses at visible frequencies

encompass silicon nitride (Si_3N_4) or titanium dioxide (TiO_2). A multitude of periodic arrangements can be leveraged to construct all-dielectric metasurfaces. The configuration of Mie-resonant dielectric nanoparticles in two-dimensional lattices of different geometries can alter the metasurface optical response.³⁵ Therein, for example, changes in periodic nanostructure spacings can alter color response due to partial reflection of the incident light and changes in the index of refraction. Further, two-dimensionally-confined dielectric nanostructures that form sub-wavelength gratings can be patterned periodically to generate high-quality-factor resonances^{190–192} with potential to generate high-purity structural color.^{193–195}

A variety of silicon metasurfaces have generated vibrant colors over a range of frequencies, for which we highlight examples here. To leverage the potential of low-order electric and magnetic Mie resonances, Proust *et al.* developed an all-dielectric silicon nanopillar array with variation in the nanostructure diameter and aspect ratio to achieve a colorimetric optical response by reproducing a Mondrian painting on the micrometer scale (Fig. 4(a)).¹⁸⁷ Højlund-Nielsen *et al.* tested the color output of square silicon nanostructures in a square lattice with varying periodicity. The resulting silicon diffraction gratings produced strong, uniform structural colors over a large angular interval.¹⁹⁶ Cross-shaped patterns also elicit compelling color effects from all-dielectric metasurfaces. Vashistha *et al.* utilized arrays of equilateral cross-shaped Si nanoresonators to obtain narrow resonances, high hue, and high saturation in the visible spectrum. The nanoresonators were polarization-insensitive, yet allowed ample color control with the dimensional parameters in the cross structure. A range of colors was produced by adjusting the length, width, and periodicity of the crosses.¹⁸⁸

In addition to geometry, the surrounding environment of the metasurface can play a crucial role in the customization of the color response. Yang *et al.* fabricated silicon nanostructures and showed that increasing the refractive index of the surrounding medium will improve vibrancy of the metasurface structural color response (Fig. 4(b)).⁷² The increase in the refractive index causes the electric dipole resonance to undergo a redshift, whereas the magnetic dipole resonance is insensitive to environmental changes. The narrower-bandwidth spectral response in transmission and reflection, resulting from subsequent destructive and constructive interference of the electric and magnetic resonances, generated structural color of improved brightness and purity.^{72,197}

Polarization-sensitive dielectric metasurfaces

The optical properties of all-dielectric metasurfaces provide a stable platform for polarization-sensitive nanophotonics. Researchers have leveraged symmetry breaking to achieve polarization-based color response. For example, Vashistha *et al.* developed an all-dielectric color filter using non-equilateral cross-shaped Si nanoantennas. The color filter was sensitive to polarization angle changes in transmission and reflection modes due to the asymmetric design (Fig. 4(c)). Furthermore, the team achieved an array of high-quality colors




Fig. 4 (a) Mondrian micropainting by Proust *et al.* made with silicon nanoparticles on glass and labeled schematic of all-dielectric metasurface structure.¹⁸⁷ (b) Left-to-right: Top-view SEM image by Yang *et al.* of a Si metasurface patterned to show a peacock with orchid (scale bar of 100 μm). Bright-field microscope image of the peacock with orchid in air.⁷² (c) Top-to-bottom: Dielectric metasurface by Vashitha *et al.* containing nonsymmetric cross-shaped Si nanoantennas with inset schematic of wave propagation for x-polarized ($\phi = 0^\circ$) and y-polarized ($\phi = 90^\circ$) normally incident light. To the right, an SEM image showing the top view of the dielectric metasurface with scale bar 250 nm. Below, a graph of polarization-based color changes from magenta to yellow for transmission mode and from green to blue for reflection mode (simulation results for polarization angle of incident white light (ϕ) from 0° to 90° with $\Delta\phi = 10^\circ$ are shown, and blue numbers correspond to ideal color wavelengths).¹⁸⁴ (d) Left-to-right: Spectral response of the red-yellow polarization-controlled tunable color filters by Uddin *et al.*, with a yellow filter for TM polarization and green filter for TE polarization. Schematic of final GMR color filter structure, with perceived colors calculated using the experimental reflectance values of the pixels.¹⁹⁴ (e) Left-to-right: Schematic illustration by Poulidakos *et al.* of the metasurface colorimetric response for a localized tumor and a metastasized tumor. Reflectance spectra showing colorimetric sensing of the metasurface for isotropic media with varying relative permittivity $\epsilon = 1:0.2:2$, with insets and line colors corresponding to reflected structural color. Colorimetric sensing capabilities for an anisotropic sample matching linearly birefringent properties of fibrous biological tissue for $\epsilon = 2$ and varying orientation angle θ of the tissue optical axis.¹⁹⁵ (a) Reprinted (adapted) with permission from ref. 187. Copyright 2016 American Chemical Society. (b and c) Creative Commons Attribution 4.0 International License. (d) [Reprinted/adapted] with permission from ref. 194. Copyright The Optical Society. (e) Reprinted (adapted) with permission from ref. 195. Copyright 2016 American Chemical Society.

by varying the size and aspect ratio of the crosses.¹⁸⁴ Instead of cross shapes, Wang *et al.* tuned the dimensions of rectangular nanoblocks to produce vibrant, polarization-sensitive colors in transmission. His team fabricated asymmetric monocrystalline silicon nanoblocks that were arranged in arrays on a flexible substrate. Different polarization angles shifted the transmission wavelength, giving access to a wide range of colors. The color filter demonstrated subwavelength resolution in the visible wavelength region.¹⁹⁸ In another study, Shibanuma *et al.* developed dielectric metasurfaces that shifted optical response from high transmission to high reflection due to changes in incident light polarization. The metasurfaces contained arrays of high-refractive-index dielectric dimers. The excited electric and magnetic dipole modes of the dimers were coupled to form hybridized modes. Incident polarization manipulated the hybridized modes, thus varying the optical response.¹⁸⁹ Asymmetric, elliptical nanopillars fabricated from TiO_2 were developed by Yang *et al.* to circumvent Ohmic loss in all-dielectric colorimetric metasurfaces.¹⁹⁹ The observed narrow-bandwidth reflectance spectra resulted in polarization-sensitive structural colors of high saturation, efficiency ($>90\%$) and resolution.

Guided-mode-resonant dielectric metasurfaces provide an additional versatile avenue to generate polarization-sensitive structural color. The Guided Mode Resonance (GMR) effect occurs in thin-film structures with diffractive elements and an optical waveguide layer. Therein, leaky modes, arising from

periodic modulation in the waveguide, can generate a narrow-bandwidth resonant response.^{156,157} In GMR metasurfaces, the resonance wavelength can be tuned using refractive index, pattern periodicity, waveguide thickness, and incident angle. Notably, dielectric GMR reflective filters have achieved structural colors with high efficiency, purity, and a narrow bandwidth. For example, Uddin *et al.* designed a GMR-based color filter array using subwavelength silicon nitride gratings on a glass substrate with an efficiency of $\sim 95\%$ and a narrow bandwidth of ~ 12 nm. The filter separated the incident transverse electric (TE)-polarized white light into its three primary colors: red, green, and blue.¹⁹³

Manipulating metasurfaces to create an anisotropic geometry allows for selective response to polarized light. Uddin *et al.* fabricated polarization-controlled tunable color filters using subwavelength gratings etched into a thin Si_3N_4 film on a glass substrate. Depending on grating dimensions, the authors demonstrated two color filters which exhibited red-yellow and green-blue reflected structural color upon transverse electric (TE) or transverse magnetic (TM) polarized light excitation, respectively (Fig. 4(d)).¹⁹⁴ In a further study which explored the scope of potential applications, Magnusson *et al.* utilized similar guided-mode-resonant grating structures for refractive-index-based biochemical sensing with high sensitivity and selectivity.²⁰⁰ Guided-mode-resonant dielectric metasurfaces have also been leveraged to quantitatively map

the presence and orientation of birefringent media onto structural color. Poulikakos *et al.* developed a guided-mode-resonant dielectric metasurface from dispersion-engineered Si_3N_4 and SiO_2 layers on a SiO_2 substrate. The reflective and rotational symmetry breaking achieved by the rhombohedral metasurface unit-cell-element enabled each incident polarization state to be attributed to a distinct colorimetric response (Fig. 4(e)).¹⁹⁵ The results of this numerical study hold potential for future applications in polarization-sensitive biomedical applications.

Material challenges in dielectric metasurfaces

High-refractive-index dielectric materials have enabled a new class of metasurfaces with versatile advantages. However, common material platforms such as silicon (Si), silicon nitride (Si_3N_4) or titanium dioxide (TiO_2) also present drawbacks that should be considered for nanophotonics applications. For example, silicon exhibits intrinsic, non-negligible material loss at visible frequencies, which can limit color saturation and range.^{201,202} Several engineering methods have been put forward to address this challenge, such as thermal annealing, geometric tuning, or addition of an antireflective layer.^{203–205} Si_3N_4 and TiO_2 have been explored as alternatives with low material loss in the visible spectrum. Contrary to Si_3N_4 , both TiO_2 and monocrystalline silicon (c-Si), are not readily compatible with Complementary Metal Oxide Semiconductor (CMOS) fabrication and would therefore face challenges in mass production.^{148,202,206–209}

In general, high refractive index dielectric nanostructures generate Mie resonances with high-quality reflection spectra, leading to vibrant colors.^{154,210} However, current research demonstrates that producing additive colors from dielectric metasurfaces in transmission mode remains challenging. As a potential CMOS-compatible alternative, amorphous silicon (a-Si) can be used, yet high material loss would hinder color saturation.²⁰⁶ In the future, hybrid platforms that integrate plasmonic and dielectric nanostructures could provide another solution for additive color generation.¹⁴⁸

Polarization-tunable structural color in hybrid and dynamic metasurfaces

In addition to metasurfaces composed purely of metallic or dielectric nanostructures, hybrid metasurfaces, which combine the strengths of various material platforms, have been demonstrated as effective generators of structural color, offering new avenues toward dynamic color tunability.^{69,162,211–215} We now focus on examples of hybrid metasurfaces that have produced polarization-tunable structural color.

Metasurface gratings, an inherently polarization-tunable optical system, are a versatile platform to demonstrate the combined advantages of metallic and dielectric metasurfaces. Kim *et al.* developed one-dimensional poly(methyl methacrylate) (PMMA) gratings on a lossy metallic substrate with the ability to generate structural color of high vibrancy and contrast.²¹¹ Their hybrid nanograting leveraged phase shifts between spectrally separate resonances arising from orthogonal TE and TM grating

modes to achieve a sharp spectral response when observed between crossed polarizers. For silver (Ag), tungsten (W) and platinum (Pt) substrates, this study was able to achieve coverage of the sRGB color gamut at 55%, 69% and 90%, respectively. Metallic nanogratings on dielectric films have been presented by Koirala *et al.*, where an ultrathin aluminum (Al) grating was fabricated on a subwavelength silicon nitride waveguide. The resulting broad color palette, observed in transmission mode upon tuning incident light polarization and grating geometry, was attributed to reinforcement of electric and magnetic field profiles in the silicon nitride waveguide layer.²¹²

Metal-insulator-metal grating metasurfaces, which act as plasmonic nanoresonators, were demonstrated by Xu *et al.* as highly efficient, high-resolution color filters across the visible spectrum.²¹³ In the work by Jung *et al.* an etalon structure composed of a subwavelength nanoscale silicon dioxide (SiO_2) and Ag layers was fabricated, where the upper Ag layer was structured as a nanograting to enable a polarization-tunable color filter.¹⁶² Therein, changing the refractive index of the dielectric environment and the thickness of the SiO_2 spacer layer varied the generated color. As an additional benefit, an increased transmission and higher resolution color images were observed with this system. Further studies^{163,164} showed numerically that stacked multilayered thin films of metal and dielectric grating structures were able to create an optical coupling of the resonance fields at the metal-dielectric interface. This effect suppresses the noise caused by higher order diffraction, which would then narrow the Full Width at Half Maximum (FWHM), leading to an improved color purity and transmission. Additional noteworthy material systems include chalcogenides, such as germanium antimony telluride (GST), where transitions between amorphous and crystalline phases yield dielectric to plasmonic material properties, respectively. To this end, Gholipour *et al.* put forward a GST nanograting where additional tailoring of structural color was achieved by phase change.²¹⁶

Liquid crystals exhibit optical properties which are tunable through variations in electric and magnetic fields, chemical composition, or temperature. These parameters offer an additional versatile platform to enhance the polarization tunability of metasurfaces. Olson *et al.* utilized periodic arrays of Al nanostructures as switchable pixels in liquid crystal displays with high chromaticity.⁶⁹ Plasmonic nanohole arrays, assembled in a rectangular lattice, were interfaced with a twisted nematic liquid crystal by Lee *et al.* to obtain an electrically- and polarization-tunable metasurface color filter.²¹⁴ Sharma *et al.* included a liquid crystal layer on top of an Al metasurface between two indium tin oxide (ITO) coated glass substrates which behaved as an electrode in their device.²¹⁵ By changing the voltage that is applied across the electrode, the device was able to both gradually and abruptly change its color. The observed color-tunability occurred as the liquid crystal varied its orientation along the electric field, thus altering the incident light polarization.

Colorimetric metasurfaces in tissue imaging: challenges and opportunities

The ability of polarization-sensitive, colorimetric metasurfaces to quantitatively map the polarization state of light onto structural color holds great promise for biomedical imaging and disease visualization, as shown schematically in Fig. 5. As was highlighted for specific examples in this work (Fig. 1), the presence and orientation of fibrous tissue is an indicator of a range of serious diseases affecting major organs of the human body (Fig. 5(a)). In contrast to biochemical markers, which can provide disease-specific information at the cellular, molecular or genetic level,¹² the ability of polarized light to probe tissue microstructure is translatable across diseases and organs, thus rendering a powerful versatility to all-optical analysis. Moreover, recent developments in biocompatible and biodegradable inorganic nanostructures for biomedical applications, such as black phosphorous nanosheets, present potential prospects for the manipulation of tissues in diagnostic and theranostic applications.^{217–220} The anisotropic nature of these material platforms offers opportunities for analysis *via* polarimetry and may provide a direct impact in the design and development of nano-optical metasurfaces in the future. Additionally, metasurfaces enable label-free visualization, thus circumventing several drawbacks encountered with fluorescent probes such as photoblinking, photobleaching and result convolution.²²¹

Leveraging colorimetric metasurfaces to visualize tissue microstructure (Fig. 5(b)) presents several advantages over existing high-sensitivity optical techniques, where this work focused on Mueller Matrix polarimetry. The ability of nano-optical structures to enhance the intensity and selectivity of light-matter interactions^{222,223} holds potential to overcome detection challenges arising from the naturally weak optical anisotropy of biological tissue. Encoding the polarization state of light onto the structural color response of a metasurface

allows for a user-friendly readout of the polarizing properties of a medium of interest, thus circumventing extensive data post-processing required in Mueller Matrix polarimetry and alternative methodologies.²²⁴ The facile tunability of metasurface optical properties, based on the material, geometry and arrangement of their nanoresonator unit-cell elements, allows for the design of customizable metasurfaces which can be adapted for various applications. Additionally, when metasurfaces are leveraged to probe the polarization state of light, direct contact between the metasurface and the material of interest is not required, as shown schematically in Fig. 5(b). Importantly, this circumvents artefacts arising from tissue roughness, which can realistically range around $1\mu\text{m}$ for sectioned tissue samples.²²⁵ The opportunities of contact-free nanophotonic detection lie in contrast to alternative metasurface sensing techniques. These methods frequently rely on sensitivity to the refractive-index environment, resulting from concentrated electromagnetic near fields, which decay exponentially from the sensing interface. Finally, perhaps the most compelling property of metasurfaces lies in their ability to scale down the complex manipulation of light to a single optical surface. This enables a host of affordable, mobile and readily translatable applications in the clinic and biomedical industry including *ex vivo* platforms in pathology and bioimaging or *in vivo* devices which could be utilized intra-operatively. For example, metasurfaces could be integrated into *in vivo* imaging systems as metalenses on optical fibers.^{226–229} Pahlevaninezhad *et al.* attached a metalens to an optical coherence tomography catheter to perform endoscopic imaging *in vivo*. Utilizing the high resolution of the metalens, the smallest FWHMs of the nano-optic endoscope were $6.37\mu\text{m}$ (tangential) and $6.53\mu\text{m}$ (sagittal).²³⁰ Metasurface-based devices can achieve subwavelength resolution at submillimeter thickness. Thus, metasurfaces provide a uniquely precise, portable bioimaging solution compared to traditional spectrometers.^{228,229,231}

Along the path to clinically- and industrially-translatable colorimetric metasurfaces, challenges must be overcome. The complex optical properties of fibrous biological tissue play a central role when leveraging metasurfaces to visualize tissue microstructure. The superposition of depolarization, diattenuation and retardance (eqn (4)) which constitute the polarizing properties of a tissue of interest, must be accounted for when designing a metasurface probe or assessing its optical response. For example, depolarization, which varies based on the refractive index and size of scatterers present in a medium,⁷⁹ significantly increases at tissue thicknesses greater than 0.5 mm .²³² While tissue biopsies are typically below thicknesses where depolarization plays a significant role,¹² these effects may become more significant in backscattering applications on bulk tissue.²⁸ Fortunately, the tunability of metasurface design presents opportunities to optically separate competing physical effects in a sensing or imaging application.¹⁹⁵

Nanofabrication methodologies which are cost-effective at an industrial scale, preferably compatible with high throughput roll-to-roll processes,²³³ are required to fabricate translatable metasurface devices. While electron-beam lithography has




Fig. 5 Schematic illustration of the opportunities for colorimetric metasurfaces to visualize the origin and progression of disease. Fibrous tissue from a range of fiber-affecting diseases exhibits a selective interaction with polarized light (part a). Colorimetric metasurfaces can be designed to quantitatively map the polarizing properties of fibrous biological tissue onto structural color (part b). Part a created with BioRender.com.

proven to be an accurate and reliable technique to demonstrate proof-of-concept metasurfaces at the research scale (*ca.* $100 \times 100 \mu\text{m}^2$), metasurfaces at a larger scale, *e.g.* expanding over the surface of a 1×3 inch microscope coverslip, would require long electron-beam exposure times which are not economically viable or serial stitching processes.¹⁷⁴ Alternative, scalable technologies include nanoimprint lithography,²³⁴ where high-throughput and low-cost pattern transfer of high-resolution nanostructures is achieved by compression molding from a reusable template. Laser interference lithography (IL)²³⁵ presents a further opportunity for large-scale patterning with laser sources, where nanoscale features can be achieved by illumination at ultraviolet wavelengths with extreme ultraviolet interference lithography (EUV-IL),²³⁶ or immersion interference lithography.^{235,237} These cost-effective, scalable fabrication techniques could increase the accessibility of metasurface-based technology, enable widespread implementation of metasurfaces into bioimaging devices, and reduce the need for expensive optical equipment in biomedical and clinical environments. The optical complexity of biological tissue and challenges of mass production present obstacles to immediate clinical use. However, with demonstrated tunability and emerging scalable processes, metasurfaces have the potential to overcome existing limitations and transform the future of medical diagnostics.

Outlook

Precisely and rationally designed material anisotropy and colorimetry in nano-optical metasurfaces present exciting opportunities for quantitative imaging of biological tissue. While the enhanced electromagnetic near fields in metasurfaces have enabled a host of successful refractive-index-based or surface-enhanced sensing applications, the avenue of enhanced detection of material anisotropy with metasurfaces remains largely unexplored. In combination with their ability to generate polarization-sensitive structural color, such metasurfaces hold promise to enable a new class of quantitative, miniaturized, cost-effective and user-friendly bioimaging platforms with facile implementation in the clinic and biomedical industry.

Author contributions

Following CRediT definitions: conceptualization: Z. H., T. P., J. J. M., L. V. P.; funding acquisition: L. V. P.; investigation: Z. H., T. P., J. J. M., L. V. P.; project administration: L. V. P.; supervision: L. V. P.; validation: Z. H., T. P., J. J. M., L. V. P.; visualization: Z. H., T. P., L. V. P.; writing – original draft: Z. H., T. P., J. J. M., L. V. P.; writing – review & editing: Z. H., T. P., J. J. M., L. V. P.

Conflicts of interest

There are no conflicts of interest to declare.

Acknowledgements

The authors thank Aniket S. Puri for insightful discussions, as well as Elissa A. Klopfer and I-Hsun Wu for artistic rendering. Lisa V. Poulidakos gratefully acknowledges funding from the UC San Diego Jacobs School of Engineering, 13991-1023256.

References

- 1 S. L. Friedman, *Gastroenterology*, 2008, **134**, 1655–1669.
- 2 D. S. James, A. N. Jambor, H.-Y. Chang, Z. Alden, K. B. Tilbury, N. K. Sandbo and P. J. Campagnola, *J. Biomed. Opt.*, 2019, **25**, 1.
- 3 P. Whittaker, R. A. Kloner, D. R. Boughner and J. G. Pickering, *Basic Res. Cardiol.*, 1994, **89**, 397–410.
- 4 N. Ghosh, N. Ghosh, M. A. Wallenburg, S.-H. Li, R. D. Weisel, B. C. Wilson, R.-K. Li and I. A. Vitkin, *J. Biomed. Opt.*, 2010, **15**, 047009.
- 5 K. Hsiao, P. Chapman, S. Nilsen, C. Eckman, Y. Harigaya, S. Younkin, F. Yang and G. Cole, *Science*, 1996, **274**, 99–102.
- 6 L. W. Jin, K. A. Claborn, M. Kurimoto, M. A. Geday, I. Maezawa, F. Sohraby, M. Estrada, W. Kaminsky and B. Kahr, *Proc. Natl. Acad. Sci. U. S. A.*, 2003, **100**, 15294–15298.
- 7 S. L. Jacques, J. C. Ramella-Roman and K. Lee, *J. Biomed. Opt.*, 2002, **7**, 329–340.
- 8 P. P. Provenzano, K. W. Eliceiri, J. M. Campbell, D. R. Inman, J. G. White and P. J. Keely, *BMC Med.*, 2006, **4**, 38.
- 9 M. W. Conklin, J. C. Eickhoff, K. M. Riching, C. A. Pehlke, K. W. Eliceiri, P. P. Provenzano, A. Friedl and P. J. Keely, *Am. J. Pathol.*, 2011, **178**, 1221–1232.
- 10 Y. K. Tao, D. Shen, Y. Sheikine, O. O. Ahsen, H. H. Wang, D. B. Schmolze, N. B. Johnson, J. S. Brooker, A. E. Cable, J. L. Connolly and J. G. Fujimoto, *Proc. Natl. Acad. Sci.*, 2014, **111**, 15304–15309.
- 11 O. Nadiarynykh, R. B. LaComb, M. A. Brewer and P. J. Campagnola, *BMC Cancer*, 2010, **10**, 94.
- 12 H. Mohan, *Textbook of Pathology*, Jaypee Brothers Medical Publishers, New Dehli, NCT Dehli, 6th edn, 2010.
- 13 M. W. Pickup, J. K. Mouw and V. M. Weaver, *EMBO Rep.*, 2014, **15**, 1243–1253.
- 14 P. Sivakumar, C. Kitson and G. Jarai, *Connect. Tissue Res.*, 2019, **60**, 62–70.
- 15 S.-Y. Lu and R. A. Chipman, *J. Opt. Soc. Am. A*, 1996, **13**, 1106.
- 16 N. Ghosh, M. F. G. Wood, S. Li, R. D. Weisel, B. C. Wilson, R.-K. Li and I. A. Vitkin, *J. Biophotonics*, 2009, **2**, 145–156.
- 17 R. Oldenbourg and G. Mei, *J. Microsc.*, 1995, **180**, 140–147.
- 18 K. R. Levental, H. Yu, L. Kass, J. N. Lakins, M. Egeblad, J. T. Erler, S. F. T. Fong, K. Csiszar, A. Giaccia, W. Weninger, M. Yamauchi, D. L. Gasser and V. M. Weaver, *Cell*, 2009, **139**, 891–906.
- 19 J. Kim, J. Feng, C. A. R. Jones, X. Mao, L. M. Sander, H. Levine and B. Sun, *Nat. Commun.*, 2017, **8**, 842.

- 20 P. V. Taufalele, J. A. VanderBurgh, A. Muñoz, M. R. Zanotelli and C. A. Reinhart-King, *PLoS One*, 2019, **14**, e0216537.
- 21 P. J. Campagnola and C.-Y. Dong, *Laser Photonics Rev.*, 2011, **5**, 13–26.
- 22 X. Chen, O. Nadiarynkh, S. Plotnikov and P. J. Campagnola, *Nat. Protoc.*, 2012, **7**, 654–669.
- 23 M. R. Hee, E. A. Swanson, J. G. Fujimoto and D. Huang, *J. Opt. Soc. Am. B*, 1992, **9**, 903–908.
- 24 S. D. Giattina, B. K. Courtney, P. R. Herz, M. Harman, S. Shortkroff, D. L. Stamper, B. Liu, J. G. Fujimoto and M. E. Brezinski, *Int. J. Cardiol.*, 2006, **107**, 400–409.
- 25 F. A. South, E. J. Chaney, M. Marjanovic, S. G. Adie and S. A. Boppart, *Biomed. Opt. Express*, 2014, **5**, 3417.
- 26 J. Schellman and H. P. Jensen, *Chem. Rev.*, 1987, **87**, 1359–1399.
- 27 N. Ghosh, M. Wood and A. Vitkin, *Handbook of Photonics for Biomedical Science*, 2010, pp. 253–282.
- 28 J. Qi and D. S. Elson, *J. Biophotonics*, 2017, **10**, 950–982.
- 29 M. Villiger, D. Lorensen, R. A. McLaughlin, B. C. Quirk, R. W. Kirk, B. E. Bouma and D. D. Sampson, *Sci. Rep.*, 2016, **6**, 28771.
- 30 J. Qi, M. Ye, M. Singh, N. T. Clancy and D. S. Elson, *Biomed. Opt. Express*, 2013, **4**, 2433.
- 31 N. T. Clancy, S. Arya, J. Qi, D. Stoyanov, G. B. Hanna and D. S. Elson, *Biomed. Opt. Express*, 2014, **5**, 4108.
- 32 J. Qi and D. S. Elson, *Sci. Rep.*, 2016, **6**, 25953.
- 33 S. Harris, *Nat. Photonics*, 2010, **4**, 748–749.
- 34 S. Jahani and Z. Jacob, *Nat. Nanotechnol.*, 2016, **11**, 23–36.
- 35 P. Genevet, F. Capasso, F. Aieta, M. Khorasaninejad and R. Devlin, *Optica*, 2017, **4**, 139–152.
- 36 B. Liedberg, C. Nylander and I. Lunström, *Sens. Actuators*, 1983, **4**, 299–304.
- 37 J. N. Anker, W. P. Hall, O. Lyandres, N. C. Shah, J. Zhao and R. P. Van Duyne, *Nat. Mater.*, 2008, **7**, 442–453.
- 38 F. Yesilkoy, E. R. Arvelo, Y. Jahani, M. Liu, A. Tittl, V. Cevher, Y. Kivshar and H. Altug, *Nat. Photonics*, 2019, **13**, 390–396.
- 39 F. Neubrech, C. Huck, K. Weber, A. Pucci and H. Giessen, *Chem. Rev.*, 2017, **117**, 5110–5145.
- 40 A. Tittl, A. Leitis, M. Liu, F. Yesilkoy, D. Y. Choi, D. N. Neshev, Y. S. Kivshar and H. Altug, *Science*, 2018, **360**, 1105–1109.
- 41 M. L. Tseng, Y. Jahani, A. Leitis and H. Altug, *ACS Photonics*, 2021, **8**, 47–60.
- 42 J. K. Gansel, M. Thiel, M. S. Rill, M. Decker, K. Bade, V. Saile, G. von Freymann, S. Linden and M. Wegener, *Science*, 2009, **325**, 1513–1515.
- 43 M. Hentschel, M. Schäferling, T. Weiss, N. Liu and H. Giessen, *Nano Lett.*, 2012, **12**, 2542–2547.
- 44 E. S. A. Goerlitzer, R. Mohammadi, S. Nechayev, K. Volk, M. Rey, P. Banzer, M. Karg and N. Vogel, *Adv. Mater.*, 2020, **32**, 2001330.
- 45 M. Hentschel, M. Schäferling, X. Duan, H. Giessen and N. Liu, *Sci. Adv.*, 2017, **3**, e1602735.
- 46 B. Frank, X. Yin, M. Schäferling, J. Zhao, S. M. Hein, P. V. Braun and H. Giessen, *ACS Nano*, 2013, **7**, 6321–6329.
- 47 A. G. Mark, J. G. Gibbs, T. C. Lee and P. Fischer, *Nat. Mater.*, 2013, **12**, 802–807.
- 48 K. M. McPeak, C. D. van Engers, S. Bianchi, A. Rossinelli, L. V. Poulikakos, L. Bernard, S. Herrmann, D. K. Kim, S. Burger, M. Blome, S. V. Jayanti and D. J. Norris, *Adv. Mater.*, 2015, **27**, 6244–6250.
- 49 C. Zhou, X. Duan and N. Liu, *Acc. Chem. Res.*, 2017, **50**, 2906–2914.
- 50 L. V. Poulikakos, P. Thureja, A. Stollmann, E. De Leo and D. J. Norris, *Nano Lett.*, 2018, **18**, 4633–4640.
- 51 H.-E. Lee, H.-Y. Ahn, J. Mun, Y. Y. Lee, M. Kim, N. H. Cho, K. Chang, W. S. Kim, J. Rho and K. T. Nam, *Nature*, 2018, **556**, 360–365.
- 52 J. Karst, N. H. Cho, H. Kim, H. E. Lee, K. T. Nam, H. Giessen and M. Hentschel, *ACS Nano*, 2019, **13**, 8659–8668.
- 53 Q. Zhang, T. Hernandez, K. W. Smith, S. A. H. Jebeli, A. X. Dai, L. Warning, R. Baiyasi, L. A. McCarthy, H. Guo, D.-H. Chen, J. A. Dionne, C. F. Landes and S. Link, *Science*, 2019, **365**, 1475–1478.
- 54 A. García-Etxarri and J. A. Dionne, *Phys. Rev. B: Condens. Matter Mater. Phys.*, 2013, **87**, 235409.
- 55 C. S. Ho, A. Garcia-Etxarri, Y. Zhao and J. Dionne, *ACS Photonics*, 2017, **4**, 197–203.
- 56 M. L. Solomon, J. A. Dionne, J. M. Abendroth, L. V. Poulikakos and J. Hu, *J. Am. Chem. Soc.*, 2020, **142**, 18304–18309.
- 57 E. Mohammadi, K. L. Tsakmakidis, A. N. Askarpour, P. Dehkoda, A. Tavakoli and H. Altug, *ACS Photonics*, 2018, **5**, 2669–2675.
- 58 M. L. Solomon, A. A. E. Saleh, L. V. Poulikakos, J. M. Abendroth, L. F. Tadesse and J. A. Dionne, *Acc. Chem. Res.*, 2020, **53**, 588–598.
- 59 A. Arbabi, Y. Horie, M. Bagheri and A. Faraon, *Nat. Nanotechnol.*, 2015, **10**, 937–943.
- 60 J. Zhou, H. Qian, C.-F. Chen, J. Zhao, G. Li, Q. Wu, H. Luo, S. Wen and Z. Liu, *Proc. Natl. Acad. Sci. U. S. A.*, 2019, **116**, 11137–11140.
- 61 N. A. Rubin, A. Zaidi, M. Juhl, R. P. Li, J. P. B. Mueller, R. C. Devlin, K. Leósson and F. Capasso, *Opt. Express*, 2018, **26**, 21455.
- 62 N. A. Rubin, G. D'Aversa, P. Chevalier, Z. Shi, W. T. Chen and F. Capasso, *Science*, 2019, **364**, eaax1839.
- 63 Y. Hu, X. Wang, X. Luo, X. Ou, L. Li, Y. Chen, P. Yang, S. Wang and H. Duan, *Nanophotonics*, 2020, **9**, 3755–3780.
- 64 S. M. Kamali, E. Arbabi, A. Arbabi and A. Faraon, *Nanophotonics*, 2018, **7**, 1041–1068.
- 65 F. Yesilkoy, R. A. Terborg, J. Pello, A. A. Belushkin, Y. Jahani, V. Pruneri and H. Altug, *Light: Sci. Appl.*, 2018, **7**, 17152.
- 66 H. Kwon, E. Arbabi, S. M. Kamali, M. Faraji-Dana and A. Faraon, *Nat. Photonics*, 2020, **14**, 109–114.
- 67 J. Zhou, H. Qian, J. Zhao, M. Tang, Q. Wu, M. Lei, H. Luo, S. Wen, S. Chen and Z. Liu, *Natl. Sci. Rev.*, 2020, nwaa176.
- 68 J. Olson, A. Manjavacas, L. Liu, W. S. Chang, B. Foerster, N. S. King, M. W. Knight, P. Nordlander, N. J. Halas and

- S. Link, *Proc. Natl. Acad. Sci. U. S. A.*, 2014, **111**, 14348–14353.
- 69 J. Olson, A. Manjavacas, T. Basu, D. Huang, A. E. Schlather, B. Zheng, N. J. Halas, P. Nordlander and S. Link, *ACS Nano*, 2016, **10**, 1108–1117.
- 70 K. Kumar, H. Duan, R. S. Hegde, S. C. W. Koh, J. N. Wei and J. K. W. Yang, *Nat. Nanotechnol.*, 2012, **7**, 557–561.
- 71 X. M. Goh, Y. Zheng, S. J. Tan, L. Zhang, K. Kumar, C.-W. Qiu and J. K. W. Yang, *Nat. Commun.*, 2014, **5**, 5361.
- 72 W. Yang, S. Xiao, Q. Song, Y. Liu, Y. Wu, S. Wang, J. Yu, J. Han and D.-P. Tsai, *Nat. Commun.*, 2020, **11**, 1864.
- 73 E. Collett, *Field Guide to Polarization*, SPIE, 2005.
- 74 C. F. Bohren and D. R. Huffman, *Absorption and scattering of light by small particles*, 1983, vol. 35.
- 75 E. Du, H. He, N. Zeng, M. Sun, Y. Guo, J. Wu, S. Liu and H. Ma, *J. Biomed. Opt.*, 2014, **19**, 076013.
- 76 N. Ghosh, M. F. G. Wood and I. A. Vitkin, *J. Biomed. Opt.*, 2008, **13**, 044036.
- 77 S. Shrestha, A. Deshpande, T. Farrahi, T. Cambria, T. Quang, J. Majeski, Y. Na, M. Zervakis, G. Livanos and G. C. Giakos, *Biomed. Signal Process. Control*, 2018, **40**, 505–518.
- 78 N. Ghosh and A. I. Vitkin, *J. Biomed. Opt.*, 2011, **16**, 110801.
- 79 N. Ghosh, P. K. Gupta, A. Pradhan and S. K. Majumder, *Phys. Lett. A*, 2006, **354**, 236–242.
- 80 V. Sankaran and J. T. Walsh, *Photochem. Photobiol.*, 1998, **68**, 846–851.
- 81 D. H. Goldstein, *Appl. Opt.*, 1992, **31**, 6676.
- 82 H. He, N. Zeng, E. Du, Y. Guo, D. Li, R. Liao and H. Ma, *Photonics Lasers Med.*, 2013, **2**, 129–137.
- 83 S. Badiyan, A. Ameri, M. R. Razzaghi, H. Raffi-Tabar and P. Sasanpour, *Photodiagn. Photodyn. Ther.*, 2019, **26**, 90–96.
- 84 M. Sun, H. He, N. Zeng, E. Du, Y. Guo, S. Liu, J. Wu, Y. He and H. Ma, *Biomed. Opt. Express*, 2014, **5**, 4223.
- 85 S. Badiyan, A. Ameri, M. R. Razzaghi, H. Raffi-Tabar and P. Sasanpour, *Photodiagn. Photodyn. Ther.*, 2019, **26**, 90–96.
- 86 Y. Wang, H. He, J. Chang, C. He, S. Liu, M. Li, N. Zeng, J. Wu and H. Ma, *J. Biomed. Opt.*, 2016, **21**, 071112.
- 87 T. Liu, M. Lu, B. Chen, Q. Zhong, J. Li, H. He, H. Mao and H. Ma, *J. Biophotonics*, 2019, **12**, e201900151.
- 88 Y. Wang, H. He, J. Chang, N. Zeng, S. Liu, M. Li and H. Ma, *Micron*, 2015, **79**, 8–15.
- 89 R. Anderson, *Appl. Opt.*, 1992, **31**, 11.
- 90 G. E. Jellison and F. A. Modine, *Appl. Opt.*, 1997, **36**, 8190.
- 91 N. Ghosh, M. Wood and A. Vitkin, in *Handbook of Photonics for Biomedical Science*, ed. V. V. Tuchin, CRC Press/Taylor & Francis, Boca Raton, 2010, pp. 253–282.
- 92 Y. Guo, N. Zeng, H. He, T. Yun, E. Du, R. Liao, Y. He and H. Ma, *Opt. Express*, 2013, **21**, 18361.
- 93 H. He, N. Zeng, E. Du, Y. Guo, D. Li, R. Liao and H. Ma, *Photonics Lasers Med.*, 2013, **2**, 129–137.
- 94 Y. Qiu, T. Jin, E. Mason and M. C. W. Campbell, *Transl. Vis. Sci. Technol.*, 2020, **9**, 1–15.
- 95 Y. Zhao, M. Reda, K. Feng, P. Zhang, G. Cheng, Z. Ren, S. G. Kong, S. Su, H. Huang and J. Zang, *IEEE Sens. J.*, 2020, **20**, 7208–7215.
- 96 P. Schucht, H. R. Lee, H. M. Mezouar, E. Hower, A. Raabe, M. Murek, I. Zubak, J. Goldberg, E. Kövari, A. Pierangelo and T. Novikova, *IEEE Trans. Med. Imaging*, 2020, **39**, 4376–4382.
- 97 Y. Dong, J. Qi, H. He, C. He, S. Liu, J. Wu, D. S. Elson and H. Ma, *Biomed. Opt. Express*, 2017, **8**, 3643.
- 98 I. Ahmad, M. Ahmad, K. Khan, S. Ashraf, S. Ahmad and M. Ikram, *J. Biomed. Opt.*, 2015, **20**, 056012.
- 99 N. Ghosh, M. Wood and A. Vitkin, *Handbook of Photonics for Biomedical Science*, 2010, pp. 253–282.
- 100 A. Gillessen, B. Voß, J. Rauterberg and W. Domschke, *Scand. J. Gastroenterol.*, 1993, **28**, 688–689.
- 101 B. Rosner, R. J. Glynn and M. Ting, Lee, *Biometrics*, 2003, **59**, 1089–1098.
- 102 W. S. Noble, *Nat. Biotechnol.*, 2006, **24**, 1565–1567.
- 103 B. Zhou, J. Xuan, H. Zhao, G. J. Chepko, M. T. Freedman and K. Y. Zou, *2007 IEEE/NIH Life Science Systems and Applications Workshop, LISA*, IEEE Computer Society, 2007, pp. 217–220.
- 104 M. Kotru and N. Singh, *Indian J. Orthop.*, 2007, **41**, 97–100.
- 105 B. Saikia, K. Gupta and U. N. Saikia, *Diagn. Pathol.*, 2008, **3**, 25.
- 106 Y. Lecun, Y. Bengio and G. Hinton, *Nature*, 2015, **521**, 436–444.
- 107 F. Limaiem, A. Rehman and T. Mazzoni, *Cancer, Papillary Thyroid Carcinoma (PTC)*, StatPearls Publishing, 2020.
- 108 H. He, R. Liao, N. Zeng, P. Li, Z. Chen, X. Liu and H. Ma, *J. Light Technol.*, 2019, **37**, 2534–2548.
- 109 D. B. Strader, T. Wright, D. L. Thomas and L. B. Seeff, *Hepatology*, 2004, **39**, 1147–1171.
- 110 R. M. Haralick, I. Dinstein and K. Shanmugam, *IEEE Trans. Syst. Man Cybern.*, 1973, **SMC-3**, 610–621.
- 111 J. M. O'Rourke, V. M. Sagar, T. Shah and S. Shetty, *World J. Gastroenterol.*, 2018, **24**, 4436–4447.
- 112 A. Sood, V. Midha and A. Singh, *Curr. Gastroenterol. Rep.*, 2014, **16**, 418.
- 113 W. Sheng, W. Li, J. Qi, T. Liu, H. He, Y. Dong, S. Liu, J. Wu, D. Elson and H. Ma, *Photonics*, 2019, **6**, 34.
- 114 H. Tamura, S. Mori and T. Yamawaki, *IEEE Trans. Syst. Man Cybern.*, 1978, **8**, 460–473.
- 115 M. C. Montalto, *J. Pathol. Inform.*, 2016, **7**, 18.
- 116 G. C. Giakos, *Int. J. Signal Imaging Syst. Eng.*, 2010, **3**, 3–12.
- 117 J. W. Turkey, *Exploratory data analysis*, Pearson, 1977.
- 118 G. C. Giakos, S. Marotta, S. Shrestha, A. Deshpande, T. Farrahi, L. Zhang, T. Cambria, A. Blinzler, T. Quang, Y. Na, G. Livanos, M. Zervakis and S. Musa, *IST 2015 - 2015 IEEE International Conference on Imaging Systems and Techniques, Proceedings*, Institute of Electrical and Electronics Engineers Inc., 2015.
- 119 K. Kadota, K. Suzuki, S. S. Kachala, E. C. Zabor, C. S. Sima, A. L. Moreira, A. Yoshizawa, G. J. Riely, V. W. Rusch, P. S. Adusumilli and W. D. Travis, *Mod. Pathol.*, 2012, **25**, 1117–1127.
- 120 H. A. Carvalho, P. H. N. Saldiva, T. Y. Takagaki and V. L. Capelozzi, *Histopathology*, 1997, **31**, 420–429.
- 121 B. T. Hyman, C. H. Phelps, T. G. Beach, E. H. Bigio, N. J. Cairns, M. C. Carrillo, D. W. Dickson,

- C. Duyckaerts, M. P. Frosch, E. Masliah, S. S. Mirra, P. T. Nelson, J. A. Schneider, D. R. Thal, B. Thies, J. Q. Trojanowski, H. V. Vinters and T. J. Montine, *Alzheimer's Dementia*, 2012, **8**, 1–13.
- 122 C. R. Jack, D. A. Bennett, K. Blennow, M. C. Carrillo, B. Dunn, S. B. Haeberlein, D. M. Holtzman, W. Jagust, F. Jessen, J. Karlawish, E. Liu, J. L. Molinuevo, T. Montine, C. Phelps, K. P. Rankin, C. C. Rowe, P. Scheltens, E. Siemers, H. M. Snyder, R. Sperling, C. Elliott, E. Masliah, L. Ryan and N. Silverberg, *Alzheimer's Dementia*, 2018, **14**, 535–562.
- 123 C. Ising, M. Stanley and D. Holtzman, *Clin. Pharmacol. Ther.*, 2015, **98**, 469–471.
- 124 M. N. Sabbagh and K. Blennow, *Neurol. Ther.*, 2019, **8**, 33–36.
- 125 A. J. Izenman, Modern multivariate statistical techniques. *Regression, classification, and Manifold learning (Springer texts in statistics)*, Springer, New York, NY, 2013.
- 126 G. Biau and E. Scornet, *Test*, 2016, **25**, 197–227.
- 127 W. Iba and P. Langley, *Machine Learning Proceedings 1992*, Elsevier, 1992, pp. 233–240.
- 128 Y. Koronyo, D. Biggs, E. Barron, D. S. Boyer, J. A. Pearlman, W. J. Au, S. J. Kile, A. Blanco, D.-T. Fuchs, A. Ashfaq, S. Frautschy, G. M. Cole, C. A. Miller, D. R. Hinton, S. R. Verdooner, K. L. Black and M. Koronyo-Hamaoui, *JCI Insight*, 2017, **2**, e93621.
- 129 M. R. Hee, C. A. Puliafito, C. Wong, E. Reichel, J. S. Duker, J. S. Schuman, E. A. Swanson and J. G. Fujimoto, *Am. J. Ophthalmol.*, 1995, **120**, 65–74.
- 130 M. J. Davies and A. C. Thomas, *Br. Heart J.*, 1985, **53**, 363–373.
- 131 L. C. U. Junqueira, G. Bignolas and R. R. Brentani, *Histochem. J.*, 1979, **11**, 447–455.
- 132 D. Chapman, K. T. Weber and M. Eghbali, *Circ. Res.*, 1990, **67**, 787–794.
- 133 M. Salvati, A. Pichierri, M. Piccirilli, G. M. F. Brunetto, A. D'Elia, S. Artizzu, F. Santoro, A. Arcella, F. Giangaspero, A. Frati, L. Simione and A. Santoro, *J. Neurosurg.*, 2012, **117**, 204–211.
- 134 R. Soffietti, B. G. Baumert, L. Bello, A. Von Deimling, H. Duffau, M. Frénay, W. Grisold, R. Grant, F. Graus, K. Hoang-Xuan, M. Klein, B. Melin, J. Rees, T. Siegal, A. Smits, R. Stupp and W. Wick, *Eur. J. Neurol.*, 2010, **17**, 1124–1133.
- 135 P. G. Gritsenko, O. Ilina and P. Friedl, *J. Pathol.*, 2012, **226**, 185–199.
- 136 T. F. Anderson and A. G. Richards, *J. Appl. Phys.*, 1942, **13**, 748–758.
- 137 H. Ghiradella, *Appl. Opt.*, 1991, **30**, 3492.
- 138 Y. Fu, C. A. Tippets, E. U. Donev and R. Lopez, *Wiley Interdiscip. Rev.: Nanomed. Nanobiotechnol.*, 2016, **8**, 758–775.
- 139 D. G. Stavenga, H. L. Leertouwer, N. J. Marshall and D. Osorio, *Proc. R. Soc. B*, 2011, **278**, 2098–2104.
- 140 J. Lee, T. Terlier, Y. J. Jang, K. Lee and Y. Lee, *Surf. Interface Anal.*, 2020, **52**, 656–664.
- 141 B.-K. Hsiung, R. H. Siddique, D. G. Stavenga, J. C. Otto, M. C. Allen, Y. Liu, Y.-F. Lu, D. D. Deheyn, M. D. Shawkey and T. A. Blackledge, *Nat. Commun.*, 2017, **8**, 2278.
- 142 J. D. Forster, H. Noh, S. F. Liew, V. Saranathan, C. F. Schreck, L. Yang, J.-G. Park, R. O. Prum, S. G. J. Mochrie, C. S. O'Hern, H. Cao and E. R. Dufresne, *Adv. Mater.*, 2010, **22**, 2939–2944.
- 143 S. Magkiriadou, J.-G. Park, Y.-S. Kim and V. N. Manoharan, *Opt. Mater. Express*, 2012, **2**, 1343.
- 144 S. Magkiriadou, J.-G. Park, Y.-S. Kim and V. N. Manoharan, *Phys. Rev. E: Stat., Nonlinear, Soft Matter Phys.*, 2014, **90**, 062302.
- 145 J.-G. Park, S.-H. Kim, S. Magkiriadou, T. M. Choi, Y.-S. Kim and V. N. Manoharan, *Angew. Chem., Int. Ed.*, 2014, **53**, 2899–2903.
- 146 Y. Shen, V. Rinnerbauer, I. Wang, V. Stelmakh, J. D. Joannopoulos and M. Soljač, *ACS Photonics*, 2015, **2**, 23.
- 147 Y. Liu, H. Wang, J. Ho, R. C. Ng, R. J. H. Ng, V. H. Hall-Chen, E. H. H. Koay, Z. Dong, H. Liu, C.-W. Qiu, J. R. Greer and J. K. W. Yang, *Nat. Commun.*, 2019, **10**, 4340.
- 148 B. Yang, H. Cheng, S. Chen and J. Tian, *Mater. Chem. Front.*, 2019, **3**, 750–761.
- 149 S. H. Ko, *Nature*, 2019, **570**, 312–313.
- 150 M. M. Ito, A. H. Gibbons, D. Qin, D. Yamamoto, H. Jiang, D. Yamaguchi, K. Tanaka and E. Sivaniah, *Nature*, 2019, **570**, 363–367.
- 151 Y. Wang, Q. Zhao and X. Du, *Mater. Horiz.*, 2020, **7**, 1341–1347.
- 152 T. Ellenbogen, K. Seo and K. B. Crozier, *Nano Lett.*, 2012, **12**, 1026–1031.
- 153 A. Kristensen, J. K. W. Yang, S. I. Bozhevolnyi, S. Link, P. Nordlander, N. J. Halas and N. A. Mortensen, *Nat. Rev. Mater.*, 2016, **2**, 16088.
- 154 A. B. Evlyukhin, S. M. Novikov, U. Zywietz, R. L. Eriksen, C. Reinhardt, S. I. Bozhevolnyi and B. N. Chichkov, *Nano Lett.*, 2012, **12**, 3749–3755.
- 155 P. D. Terekhov, K. V. Baryshnikova, Y. Greenberg, Y. H. Fu, A. B. Evlyukhin, A. S. Shalin and A. Karabchevsky, *Sci. Rep.*, 2019, **9**, 3438.
- 156 G. Quaranta, G. Basset, O. J. F. Martin and B. Gallinet, *Laser Photonics Rev.*, 2018, **12**, 1800017.
- 157 Y. Zhou, B. Wang, Z. Guo and X. Wu, *Nanomaterials*, 2019, **9**, 837.
- 158 M. Keshavarz Hedayati and M. Elbahri, *Plasmonics*, 2017, **12**, 1463–1479.
- 159 Y. Gu, L. Zhang, J. K. W. Yang, S. P. Yeo and C.-W. Qiu, *Nanoscale*, 2015, **7**, 6409–6419.
- 160 S. A. Maier, *Plasmonics: Fundamentals and Applications*, Springer Science + Business Media LLC New York, NY, 2007.
- 161 P. Dai, Y. Wang, X. Zhu, H. Shi, Y. Chen, S. Zhang, W. Yang, Z. Chen, S. Xiao and H. Duan, *Nanotechnology*, 2018, **29**, 395202.
- 162 Y. Jung, H. Jung, H. Choi and H. Lee, *Nano Lett.*, 2020, **20**, 6344–6350.

- 163 S. Ren, L. Zhou, J. Yang and Y. Zhou, *Plasmonics*, 2021, **16**, 167–173.
- 164 C. Jiang, S. Liang, L. Wan, K. Du, W. Zhang, Z. Li, S. Jin Chua and T. Mei, *Appl. Opt.*, 2020, **59**, 4432.
- 165 P. K. Jain, X. Huang, I. H. El-Sayed and M. A. El-Sayed, *Acc. Chem. Res.*, 2008, **41**, 1578–1586.
- 166 J. N. Anker, W. P. Hall, O. Lyandres, N. C. Shah, J. Zhao and R. P. Van Duyne, *Nat. Mater.*, 2008, **7**, 442–453.
- 167 K. Kumar, H. Duan, R. S. Hegde, S. C. W. Koh, J. N. Wei and J. K. W. Yang, *Nat. Nanotechnol.*, 2012, **7**, 557–561.
- 168 S. J. Tan, L. Zhang, D. Zhu, X. M. Goh, Y. M. Wang, K. Kumar, C.-W. Qiu and J. K. W. Yang, *Nano Lett.*, 2014, **14**, 4023–4029.
- 169 T. W. Ebbesen, H. J. Lezec, H. F. Ghaemi, T. Thio and P. A. Wolff, *Nature*, 1998, **391**, 667–669.
- 170 S. J. Heo, S. Kim, G. Pyo, J. H. Yang, J. H. Shin, H.-J. Kwon, H. Kang, G. H. Lee and J. E. Jang, *ACS Appl. Electron. Mater.*, 2020, **2**, 2489–2497.
- 171 H. Yun, S.-Y. Lee, K. Hong, J. Yeom and B. Lee, *Nat. Commun.*, 2015, **6**, 7133.
- 172 N. S. King, L. Liu, X. Yang, B. Cerjan, H. O. Everitt, P. Nordlander and N. J. Halas, *ACS Nano*, 2015, **9**, 10628–10636.
- 173 F. F. Mahani, A. Mokhtari and M. Mehran, *Nanotechnology*, 2017, **28**, 385203.
- 174 L. Duempelmann, D. Casari, A. Luu-Dinh, B. Gallinet and L. Novotny, *ACS Nano*, 2015, **9**, 12383–12391.
- 175 L. Duempelmann, A. Luu-Dinh, B. Gallinet and L. Novotny, *ACS Photonics*, 2016, **3**, 190–196.
- 176 Z. Li, A. W. Clark and J. M. Cooper, *ACS Nano*, 2016, **10**, 492–498.
- 177 J. B. Khurgin and G. Sun, *Appl. Phys. Lett.*, 2010, **96**, 181102.
- 178 A. Raman, W. Shin and S. Fan, *Phys. Rev. Lett.*, 2013, **110**, 183901.
- 179 O. D. Miller, A. G. Polimeridis, M. T. H. Reid, C. W. Hsu, B. G. DeLacy, J. D. Joannopoulos, M. Soljačić and S. G. Johnson, *Opt. Express*, 2015, **24**, 3329–3364.
- 180 D. Inoue, A. Miura, T. Nomura, H. Fujikawa, K. Sato, N. Ikeda, D. Tsuya, Y. Sugimoto and Y. Koide, *Appl. Phys. Lett.*, 2011, **98**, 093113.
- 181 A. Henglein, *Chem. Mater.*, 1998, **10**, 444–450.
- 182 P. M. Schwab, C. Moosmann, K. Dopf and H.-J. Eisler, *Opt. Express*, 2015, **23**, 26533.
- 183 H. Wang, F. Tam, N. K. Grady and N. J. Halas, *J. Phys. Chem. B*, 2005, **109**, 18218–18222.
- 184 V. Vashistha, G. Vaidya, P. Gruszecki, A. E. Serebryannikov and M. Krawczyk, *Sci. Rep.*, 2017, **7**, 8092.
- 185 L. Cao, P. Fan, E. S. Barnard, A. M. Brown and M. L. Brongersma, *Nano Lett.*, 2010, **10**, 2649–2654.
- 186 S. R. Amanaganti, M. Ravnik and J. Dontabhaktuni, *Sci. Rep.*, 2020, **10**, 15599.
- 187 J. Proust, F. Bedu, B. Gallas, I. Ozerov and N. Bonod, *ACS Nano*, 2016, **10**, 7761–7767.
- 188 V. Vashistha, G. Vaidya, R. S. Hegde, A. E. Serebryannikov, N. Bonod and M. Krawczyk, *ACS Photonics*, 2017, **4**, 1076–1082.
- 189 T. Shibanuma, S. A. Maier and P. Albella, *Appl. Phys. Lett.*, 2018, **112**, 063103.
- 190 M. Lawrence, D. R. Barton and J. A. Dionne, *Nano Lett.*, 2018, **18**, 1104–1109.
- 191 E. Klopfer, M. Lawrence, D. R. Barton, J. Dixon and J. A. Dionne, *Nano Lett.*, 2020, **20**, 5127–5132.
- 192 M. Lawrence, D. R. Barton, J. Dixon, J.-H. Song, J. van de Groep, M. L. Brongersma and J. A. Dionne, *Nat. Nanotechnol.*, 2020, **15**, 956–961.
- 193 M. J. Uddin and R. Magnusson, *Opt. Express*, 2013, **21**, 12495.
- 194 M. J. Uddin, T. Khaleque and R. Magnusson, *Opt. Express*, 2014, **22**, 12307.
- 195 L. V. Poulikakos, M. Lawrence, D. R. Barton, S. S. Jeffrey and J. A. Dionne, *ACS Photonics*, 2020, **7**, 3216–3227.
- 196 E. Højlund-Nielsen, J. Weirich, J. Nørregaard, J. Garnæs, N. Asger Mortensen and A. Kristensen, *J. Nanophotonics*, 2014, **8**, 083988.
- 197 C.-S. Park, I. Koirala, S. Gao, V. R. Shrestha, S.-S. Lee and D.-Y. Choi, *Opt. Express*, 2019, **27**, 667.
- 198 X. Wang, J. Chen, T. Guo and Y. Shi, *Opt. Express*, 2020, **28**, 21704.
- 199 B. Yang, W. Liu, Z. Li, H. Cheng, S. Chen and J. Tian, *Adv. Opt. Mater.*, 2018, **6**, 1701009.
- 200 R. Magnusson, D. Wawro, S. Zimmerman and Y. Ding, *Sensors*, 2011, **11**, 1476–1488.
- 201 C. Wu, N. Arju, G. Kelp, J. A. Fan, J. Dominguez, E. Gonzales, E. Tutuc, I. Brener and G. Shvets, *Nat. Commun.*, 2014, **5**, 3892.
- 202 D. Sell, J. Yang, S. Doshay and J. A. Fan, *Adv. Opt. Mater.*, 2017, **5**, 1700645.
- 203 Z. Wang, X. Xu, D. Fan, Y. Wang, H. Subbaraman and R. T. Chen, *Sci. Rep.*, 2016, **6**, 24106.
- 204 S. V. Makarov, A. N. Tsyppkin, T. A. Voytova, V. A. Milichko, I. S. Mukhin, A. V. Yulin, S. E. Putilin, M. A. Baranov, A. E. Krasnok, I. A. Morozov and P. A. Belov, *Nanoscale*, 2016, **8**, 17809–17814.
- 205 Z. Dong, J. Ho, Y. F. Yu, Y. H. Fu, R. Paniagua-Dominguez, S. Wang, A. I. Kuznetsov and J. K. W. Yang, *Nano Lett.*, 2017, **17**, 7620–7628.
- 206 C.-S. Park, V. R. Shrestha, W. Yue, S. Gao, S.-S. Lee, E.-S. Kim and D.-Y. Choi, *Sci. Rep.*, 2017, **7**, 2556.
- 207 D. J. Moss, R. Morandotti, A. L. Gaeta and M. Lipson, *Nat. Photonics*, 2013, **7**, 597–607.
- 208 P. Muellner, E. Melnik, G. Koppitsch, J. Kraft, F. Schrank and R. Hainberger, *Procedia Eng.*, 2015, **120**, 578–581.
- 209 Z. Zhou, J. Li, R. Su, B. Yao, H. Fang, K. Li, L. Zhou, J. Liu, D. Stellinga, C. P. Reardon, T. F. Krauss and X. Wang, *ACS Photonics*, 2017, **4**, 544–551.
- 210 Y. Kivshar and A. Miroshnichenko, *Opt. Photonics News*, 2017, **28**, 24–31.
- 211 Y. Kim, K. Jung, J. Cho and J. K. Hyun, *ACS Nano*, 2019, **13**, 10717–10726.
- 212 I. Koirala, V. R. Shrestha, C.-S. Park, S. Gao, S.-S. Lee and D.-Y. Choi, *Sci. Rep.*, 2017, **7**, 13574.
- 213 T. Xu, Y.-K. Wu, X. Luo and L. J. Guo, *Nat. Commun.*, 2010, **1**, 59.

- 214 Y. Lee, M.-K. Park, S. Kim, J. H. Shin, C. Moon, J. Y. Hwang, J.-C. Choi, H. Park, H.-R. Kim and J. E. Jang, *ACS Photonics*, 2017, **4**, 1954–1966.
- 215 M. Sharma, N. Hendler and T. Ellenbogen, *Adv. Opt. Mater.*, 2020, **8**, 1901182.
- 216 B. Gholipour, A. Karvounis, J. Yin, C. Soci, K. F. MacDonald and N. I. Zheludev, *NPG Asia Mater.*, 2018, **10**, 533–539.
- 217 C. Xing, S. Chen, M. Qiu, X. Liang, Q. Liu, Q. Zou, Z. Li, Z. Xie, D. Wang, B. Dong, L. Liu, D. Fan and H. Zhang, *Adv. Healthcare Mater.*, 2018, **7**, 1701510.
- 218 M. Qiu, A. Singh, D. Wang, J. Qu, M. Swihart, H. Zhang and P. N. Prasad, *Nano Today*, 2019, **25**, 135–155.
- 219 M. Luo, T. Fan, Y. Zhou, H. Zhang and L. Mei, *Adv. Funct. Mater.*, 2019, **29**, 1808306.
- 220 Z. Tang, N. Kong, J. Ouyang, C. Feng, N. Y. Kim, X. Ji, C. Wang, O. C. Farokhzad, H. Zhang and W. Tao, *Matter*, 2020, **2**, 297–322.
- 221 T. Y. Chang, M. Huang, A. A. Yanik, H. Y. Tsai, P. Shi, S. Aksu, M. F. Yanik and H. Altug, *Lab Chip*, 2011, **11**, 3596–3602.
- 222 Y. Lee, S. J. Kim, H. Park and B. Lee, *Sensors*, 2017, **17**, 1726.
- 223 M. L. Tseng, Y. Jahani, A. Leitis and H. Altug, *ACS Photonics*, 2021, **3**, 21.
- 224 N. Ghosh, M. F. G. Wood and I. A. Vitkin, *J. Biomed. Opt.*, 2008, **13**, 044036.
- 225 H. K. Graham, N. W. Hodson, J. A. Hoyland, S. J. Millward-Sadler, D. Garrod, A. Scothern, C. E. M. Griffiths, R. E. B. Watson, T. R. Cox, J. T. Erler, A. W. Trafford and M. J. Sherratt, *Matrix Biol.*, 2010, **29**, 254–260.
- 226 X. Yu, S. Zhang, M. Olivo and N. Li, *Photonics Res.*, 2020, **8**, 1703.
- 227 P. Vaiano, B. Carotenuto, M. Pisco, A. Ricciardi, G. Quero, M. Consales, A. Crescitelli, E. Esposito and A. Cusano, *Laser Photonics Rev.*, 2016, **10**, 922–961.
- 228 S. Zhang, C. L. Wong, S. Zeng, R. Bi, K. Tai, K. Dholakia and M. Olivo, *Nanophotonics*, 2021, **10**, 259–293.
- 229 D. Lee, J. Gwak, T. Badloe, S. Palomba and J. Rho, *Nano-scale Adv.*, 2020, **2**, 605–625.
- 230 H. Pahlevaninezhad, M. Khorasaninejad, Y.-W. Huang, Z. Shi, L. P. Hariri, D. C. Adams, V. Ding, A. Zhu, C.-W. Qiu, F. Capasso and M. J. Suter, *Nat. Photonics*, 2018, **12**, 540–547.
- 231 B. Li, W. Piyawattanametha and Z. Qiu, *Micromachines*, 2019, **10**, 310.
- 232 V. Sankaran, J. T. Walsh and D. J. Maitland, *J. Biomed. Opt.*, 2002, **7**, 300.
- 233 S. H. Ahn and L. J. Guo, *ACS Nano*, 2009, **3**, 2304–2310.
- 234 S. Y. Chou, P. R. Krauss and P. J. Renstrom, *Science*, 1996, **272**, 85–87.
- 235 D. Xia, Z. Ku, S. C. Lee and S. R. J. Brueck, *Adv. Mater.*, 2011, **23**, 147–179.
- 236 F. Juillerat, H. H. Solak, P. Bowen and H. Hofmann, *Nanotechnology*, 2005, **16**, 1311–1316.
- 237 R. Sidharthan, V. M. Murukeshan and K. Sathiyamoorthy, *J. Nanosci. Nanotechnol.*, 2012, **12**, 6428–6431.